



# Standard Test Method for pH of Water Extractions of Halogenated Organic Solvents and Their Admixtures<sup>1</sup>

This standard is issued under the fixed designation D2110; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This test method covers the measurement of the pH of water extractions of halogenated organic solvents and admixtures thereof. This test method is applicable for the determination of the pH of water extractions of virgin, reclaimed, or used solvents.

1.2 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

## 2. Summary of Test Method

2.1 The solvent sample is shaken with freshly boiled neutral distilled or deionized water and the pH of the water extract determined by one of two procedures:

2.1.1 *Procedure A*, by comparing the color formed upon the addition of Universal Indicator with that of a calibrated pH Indicator Color Chart, or

2.1.2 *Procedure B*, using a glass electrode pH meter.

## 3. Significance and Use

3.1 The pH of halogenated solvents varies according to the nature of stabilizers in the particular solvent. Solvents with alkaline stabilizers (amine-types) generally have pHs in the 7 to 11 range, whereas solvents containing neutral stabilizers (epoxide-types) generally have pHs in the 5 to 7 range.

3.2 Virgin and reclaimed solvents should have pHs within these ranges, dependent upon the claimed method of stabilization. These values should be within the accepted limits as proposed by the producers of the original virgin solvent.

3.3 A lowering of pH from the above values, on receipt or use, is cause for concern. In this event, solvents having epoxide stabilizers should be tested for acid acceptance; solvent having alkaline stabilizers should be tested for alkalinity.

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D26 on Halogenated Organic Solvents and Fire Extinguishing Agents and is the direct responsibility of Subcommittee D26.04 on Test Methods.

Current edition approved March 1, 2012. Published May 2012. Originally approved in 1962 as D2110-62 T. Last previous edition approved in 2006 as D2110-00 (2006). DOI: 10.1520/D2110-00R12.

## 4. Apparatus

4.1 *Universal Indicator*<sup>2</sup> and pH Indicator Color Chart (Procedure A),<sup>2</sup> or

4.2 *pH Meter* (Procedure B).

## 5. Water

5.1 *Neutral Distilled or Deionized Water* shall be prepared as follows: Boil 1 L of distilled or deionized water for 5 min in a borosilicate glass or stainless steel container, then cover and cool to room temperature.

5.2 *Procedure A*—To a 50-mL aliquot of the cooled, boiled water, add 2 mL of Universal Indicator. Titrate the aliquot to a pH of 7 (as indicated by comparing the color of the aliquot to the color chart per manufacturer's instructions) with either 0.01 N HCl if the pH is above 7 or 0.01 N NaOH if the pH is below 7, and record the titer. From this titer, calculate and add the volume of 0.01 N NaOH or 0.01 N HCl required to adjust the remaining 950 mL of the boiled distilled or deionized water to a pH of 7. The amount of 0.01 N NaOH or 0.01 N HCl needed to neutralize the 950 mL water will be 19 X titer. Determine the pH of a second aliquot. If further adjustment is necessary to obtain a pH of 7, repeat the above procedure.

5.3 *Procedure B*—Using a pH meter, titrate a 50 mL aliquot to a pH of 7.0 to 7.3 with either 0.01 N HCl if the pH is above 7.0 or 0.01 N NaOH if the pH is 7.0 and record the titer. From this titer, calculate and add the volume of 0.01 N NaOH or 0.01 N HCl required to adjust the remaining 950 mL of the boiled distilled or deionized water to a pH of 7.0 to 7.3. The amount of 0.01 N NaOH or 0.01 N HCl needed to neutralize the 950 mL water will be 19 X titer. Determine the pH of a second aliquot. If further adjustment is necessary to obtain a pH of 7.0 to 7.3, repeat the above procedure.

## 6. Procedure

6.1 Transfer 50 mL of the halogenated organic solvent to a 250-mL separatory funnel containing 50 mL of freshly boiled, neutral, distilled or freshly boiled, neutral, deionized water. Shake the mixture for 2 min; then let stand and allow the layers

<sup>2</sup> Available from suppliers of reagent chemicals such as Fisher Scientific, Houston, TX or Aqua Solutions, Deer Park, TX.

to separate. Drain and discard the halogenated solvent layer and transfer the water extract to a 100-mL beaker. Determine the pH of the water extract by either Procedure A or B.

6.2 *Procedure A*—Pipet 2 mL of Universal Indicator into the water extract, stir, and compare the resultant color to the color chart per manufacturer’s instructions. The pH of the water extract is that of the pH of the closest matching color on the color chart.

6.3 *Procedure B*—Determine the pH of the water extract to the nearest 0.1 pH unit with a pH meter.

## 7. Precision and Bias<sup>3</sup>

7.1 The precision for Procedure A, Universal Indicator, is  $\pm 0.5$  pH unit; for Procedure B, pH meter,  $\pm 0.2$  pH unit; and for multilaboratory,  $\pm 0.9$  pH unit.

7.2 Bias has not been determined.

## 8. Keywords

8.1 acidity; alkalinity; halogenated organic solvents; pH; solvents

---

<sup>3</sup> Supporting data have been filed at ASTM headquarters and may be obtained by requesting Research Report RR D26-1007.

*ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.*

*This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.*

*This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the ASTM website (www.astm.org/COPYRIGHT/).*