



FIG. 1 Apparatus

where such specifications are available.³ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

5.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean Type III reagent water conforming to Sections 1, 2, and 3 of Specification D1193.

5.3 *Carbon Dioxide or Nitrogen*, cylinder gas.

5.4 *Lubricant* for ground-glass joints.

5.5 *Dibutyl Phthalate*.

5.6 *Glass Beads*.

5.7 *Hydriodic Acid (55 to 58 %) (suitable for methoxyl determinations)*—The hydriodic acid (HI) should not contain hypophosphorus acid (H₃PO₂) stabilizer. The free iodine content should be less than 10 mL of 0.1 N sodium thiosulfate solution/5 mL of HI.

5.8 *Hydrochloric Acid (sp gr 1.19)*—Concentrated hydrochloric acid (HCl).

5.9 *Methanol*.

5.10 *Potassium Dichromate (K₂Cr₂O₇)*, primary standard.

5.11 *Potassium Iodide Solution (200 g/L)*—Dissolve 200 g of potassium iodide (KI) in distilled water and dilute to 1000 mL.

³ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

5.12 *Sodium Thiosulfate, Standard Solution (0.1 N)*—Dissolve 24.8 g of sodium thiosulfate (Na₂S₂O₃) in water and dilute to 1000 mL. Standardize by accurately weighing 0.16 to 0.22 g of finely ground and dried K₂Cr₂O₇ into a 500-mL iodine flask. Dissolve in 25 mL of water, add 5 mL of HCl and 15 mL of KI solution and swirl to mix. Allow to stand for 5 min, then add 100 mL of distilled water. Titrate with standard Na₂S₂O₃ solution, shaking continuously until the yellow color has almost disappeared. Add a small amount of Paragon indicator⁴ or 1 to 2 mL of starch indicator solution and continue the titration, adding the Na₂S₂O₃ solution slowly until the blue color has just disappeared. Calculate the normality of the Na₂S₂O₃ solution as follows:

$$N = C / (0.04904 \times D) \quad (1)$$

where:

N = normality of the Na₂S₂O₃ solution,

C = grams of K₂Cr₂O₇ used, and

D = millilitres of Na₂S₂O₃ solution required for titration of the solution.

5.13 *Starch Indicator Solution*—Make a homogeneous paste of 10 g of soluble starch in water. To this add with rapid stirring 1 L of boiling water and then cool. As a preservative, add 1.25 g of salicylic acid per litre. Store the indicator in a refrigerator at 4 to 10°C. Prepare fresh indicator when the titration end point from blue to colorless fails to be sharp.

6. Procedure

6.1 Assemble the apparatus as shown in Fig. 1. Lubricate the ground-glass joints with a very thin film of lubricant. The dual apparatus is designed to permit a sample and a blank to be run simultaneously; however, it should not be necessary to make more than two blank determinations each day.

6.2 Accurately weigh a sample of such size that 4 to 5 meq of iodine will be liberated (about 0.15 g of a nonionic surfactant that contains 60 % ethylene oxide) into a 100-mL round bottom flask. Pipet exactly 5.0 mL of HI into the flask, and add a glass bead. Connect the vertical condenser to the flask using a thin seal of lubricant at the outer edge. Prepare a blank by pipetting 5.0 mL of HI into a second flask containing a glass bead. Connect the condenser as described above, and allow water to circulate through the condensers.

6.3 Attach the tubes from the carbon dioxide manifold to the sidearms on the flasks. Open the valve on the gas cylinder and adjust the flow to a rate of 1 to 5 bubbles per second through the bubbler, blanketing the solutions in the reaction flasks. The flow of carbon dioxide should continue for 10 min before heat is applied to the reaction flasks. After 10 min turn on the heat, and reflux the contents gently for 90 min.

6.4 After the 90-min boiling period, wash the walls of each condenser with 15 mL of 20 % KI solution to dissolve any free iodine in the condensers. Wash the condensers with two 10-mL

⁴ The sole source of supply of the Paragon indicator for iodine titrations known to the committee at this time is Eastern Chemical Corp., 34 Spring St., Newark, NJ 07104. If you are aware of alternative suppliers, please provide this information to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee,¹ which you may attend.

portions of water and disconnect. Wash the tips of the condensers with water, adding the washings to the flasks. Transfer the contents of the reaction flasks to Erlenmeyer flasks by washing with water.

NOTE 4—Rinse any iodine that adheres to the condensers into the flasks with a small amount of methanol. On rare occasions some samples leave a tarry residue as a button. Add methanol to dissolve the residue, and wash into the Erlenmeyer flask.

6.5 Titrate with standard $\text{Na}_2\text{S}_2\text{O}_3$ solution, shaking continuously until the yellow color has almost disappeared. Add a small amount of Paragon indicator or 1 to 2 mL of starch indicator solution and continue the titration, adding the $\text{Na}_2\text{S}_2\text{O}_3$ solution slowly until the blue color has just disappeared.

7. Calculation

7.1 Calculate the ethylene oxide content as follows:

$$\text{Ethylene oxide, weight \%} = [(A - B) \times N \times 2.203] / W \quad (2)$$

where:

- A* = millilitres of $\text{Na}_2\text{S}_2\text{O}_3$ solution required to titrate the sample,
- B* = millilitres of $\text{Na}_2\text{S}_2\text{O}_3$ solution required to titrate the blank,
- N* = normality of $\text{Na}_2\text{S}_2\text{O}_3$ solution, and
- W* = grams of sample.

8. Precision and Bias

8.1 The following criteria should be used for judging the acceptability of results (**Note 5**):

8.1.1 *Repeatability (Single Analyst)*—The standard deviation of results obtained by the same analyst was estimated to be 0.47 % absolute at 9 df. Two such values should be considered suspect (95 % confidence level) if they differ by more than 1.5 % absolute.

8.1.2 *Reproducibility (Multilaboratory)*⁵—The standard deviation of results (each the average of duplicate determinations), obtained by analysts in different laboratories, has been estimated to be 1.81 % absolute at 7 df. Two such values should be considered suspect (95 % confidence level) if they differ by more than 6.0 % absolute.

NOTE 5—The precision estimates are based on an interlaboratory study on Igepal CO 630 by nine laboratories. However, this procedure as written calls for the use of 100-mL reaction flasks (instead of 50-mL flasks which were used in the interlaboratory study). Data obtained by Continental Oil Co. indicate an improvement in precision when the larger flask is used.

9. Keywords

9.1 ethylene oxide content; nonionic surfactants

⁵ Data supporting the precision statements are on file at ASTM Headquarters. Request Research Report No. RR:D12-1000.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; http://www.copyright.com/