



# Standard Test Method for Detection of Holes in Medical Gloves<sup>1</sup>

This standard is issued under the fixed designation D5151; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This test method covers the detection of holes in medical gloves.

1.2 This test method is limited to the detection of holes that allow water leakage under the conditions of the test.

1.3 The smallest hole size that will allow water leakage in a medical glove has not been determined and is beyond the scope of this test method.

1.4 The safe and proper use of medical gloves is beyond the scope of this test method.

1.5 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

## 2. Terminology

2.1 *Definitions of Terms Specific to This Standard:*

2.1.1 *medical gloves, n*—medical gloves, as used in this test method, refer to both surgical and examination gloves.

## 3. Significance and Use

3.1 This test method is designed to detect holes that leak water and thereby compromise the usefulness of the glove.

3.2 This test method is suitable and designed as a reference method to evaluate samples of medical gloves.

## 4. Apparatus

4.1 *Mandrel*—A clear plastic tube, nominally  $6 \pm 0.5$  cm outside diameter, 5 cm inside diameter or 50 mm schedule 80 poly(vinyl chloride) (PVC) pipe,  $38 \pm 1$  cm long. The mandrel shall have a clamp or rubber band for securing the glove to the mandrel.

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D11 on Rubber and is the direct responsibility of Subcommittee D11.40 on Consumer Rubber Products.

Current edition approved Nov. 1, 2015. Published December 2015. Originally approved in 1990. Last previous edition approved in 2011 as D5151 – 06 (2011). DOI: 10.1520/D5151-06R15.

4.2 *Graduated Cylinder or Beaker*—As a minimum, the graduated cylinder or beaker shall be marked in increments of  $50 \text{ cm}^3$  to a capacity of  $1000 \text{ cm}^3$ . The markings shall be accurate to within  $\pm 10 \text{ cm}^3$  at volumes greater than  $500 \text{ cm}^3$ .

4.3 *Stopwatch*.

4.4 *Various Clamps, Hangers, and Stands*.

## 5. Reagents

5.1 *Tap Water*—The water used in this test method shall be clean and free from any visible contaminants.

## 6. Procedure

6.1 Throughout 6.2 and 6.3, ensure that the exterior of the glove remains dry.

6.2 Mount the mandrel in a vertical position using appropriate stands, clamps, and hangers. Affix the glove to the mandrel by stretching the cuff of the glove around the mandrel. A maximum 40 mm of glove cuff should fit over the bottom end of the mandrel. Use the securing device, as necessary, to hold the glove in place. The remainder of the glove should hang freely from the mandrel when filled with water.

6.3 Pour a minimum of  $1000 \text{ cm}^3$  of water having a room temperature of 15 to  $30^\circ\text{C}$  into the top of the mandrel. The water shall pass freely into the glove.

6.4 Visually inspect the glove for immediate water leakage. Let the glove hang for 2 min and again inspect for water leakage.

## 7. Interpretation of Results

7.1 Any glove that shows a droplet, stream, or other type of water leakage shall be considered to have failed the test.

7.2 Water leakage above the junction of the glove cuff and mandrel is not a test failure. If this leakage occurs, remove and dry the outside of the glove and retest.

## 8. Report

8.1 The report shall include the number of samples tested, the number of failures, and a general description of the location of each hole.

## **9. Precision and Bias**

9.1 No statement is made about either the precision or bias of this test method. This is a “pass-fail” test procedure not subject to the usual analysis for quantitative variable precision.

*ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.*

*This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.*

*This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or [service@astm.org](mailto:service@astm.org) (e-mail); or through the ASTM website ([www.astm.org](http://www.astm.org)). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; <http://www.copyright.com/>*

## **10. Keywords**

10.1 detection; gloves; holes; medical