



# Standard Specification for Poly(vinyl chloride) Gloves for Medical Application<sup>1</sup>

This standard is issued under the fixed designation D5250; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This specification covers certain requirements for poly(vinyl chloride) gloves used in conducting medical examinations and diagnostic and therapeutic procedures. It also covers poly(vinyl chloride) gloves used in handling contaminated medical material.

1.2 This specification provides for poly(vinyl chloride) gloves that fit either hand, paired gloves, and gloves by size. It also provides for packaged sterile or nonsterile or bulk nonsterile poly(vinyl chloride) gloves.

1.3 This specification does not cover two-dimensional heat sealed poly(vinyl chloride) gloves.

1.4 This specification is similar to that of Specification [D3578](#) for rubber examination gloves.

## 2. Referenced Documents

### 2.1 *ASTM Standards*:<sup>2</sup>

[D412 Test Methods for Vulcanized Rubber and Thermoplastic Elastomers—Tension](#)

[D573 Test Method for Rubber—Deterioration in an Air Oven](#)

[D3578 Specification for Rubber Examination Gloves](#)

[D3767 Practice for Rubber—Measurement of Dimensions](#)

[D5151 Test Method for Detection of Holes in Medical Gloves](#)

[D6124 Test Method for Residual Powder on Medical Gloves](#)

### 2.2 *Other Document*:

[ISO 2859 Sampling Procedures and Tables for Inspection by Attributes](#)<sup>3</sup>

[U. S. Pharmacopeia](#)<sup>4</sup>

<sup>1</sup> This specification is under the jurisdiction of [D11](#) on Rubber and is the direct responsibility of [D11.40](#) on Consumer Rubber Products.

Current edition approved Nov. 1, 2015. Published December 2015. Originally approved in 1992. Last previous edition approved in 2011 as D5250 – 06(2011). DOI: 10.1520/D5250-06R15.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

<sup>3</sup> Available from American National Standards Institute, 25 W. 43rd St., 4th Floor, New York, NY 10036.

<sup>4</sup> U.S. Pharmacopeia, latest edition, Mack Publishing Co., Easton, PA, 19175.

## 3. Materials

3.1 Any poly(vinyl chloride) polymer compound may be used that permits the glove to meet the requirements of this standard.

3.2 A lubricant that meets the current requirements of the U.S. Pharmacopeia for absorbable dusting powder may be applied to the glove. Other lubricants may be used if their safety and efficacy have been previously established.

3.3 The inside and outside surface of the poly(vinyl chloride) examination gloves shall be free of talc.

## 4. Significance and Use

4.1 The specification is intended as a referee procedure for evaluating the performance and safety of poly(vinyl chloride) examination gloves. The safe and proper use of poly(vinyl chloride) examination gloves is beyond the scope of this standard.

## 5. Sampling

5.1 For referee purposes, gloves shall be sampled from finished product, after sterilization, and inspected in accordance with ISO 2859. The inspection levels and acceptable quality levels (AQL) shall conform to those specified in [Table 1](#), or as agreed upon between the purchaser and seller, if the latter is more comprehensive.

## 6. Performance Requirements

6.1 Gloves, sampled in accordance with Section 5, shall meet the following referee performance requirements:

6.1.1 Comply with requirements for sterility when tested in accordance with [7.2](#).

6.1.2 Be free from holes when tested in accordance with [7.3](#).

6.1.3 Have consistent physical dimensions in accordance with [7.4](#).

6.1.4 Have acceptable physical property characteristics in accordance with [7.5](#).

6.1.5 Have a powder residue limit of 2.0 mg in accordance with [7.6](#).

6.1.6 Have a recommended maximum powder limit of 10 mg/dm<sup>2</sup> in accordance with [7.7](#).

## 7. Referee Test Methods

7.1 The following tests shall be conducted to ensure the requirements of Section 6, as prescribed in [Table 1](#).

**TABLE 1 Performance Requirements**

Characteristic	Related Defects	Inspection Level	Acceptable Quality Levels
Sterility	falls sterility	A	N/A
Freedom from holes	holes	I	2.5
Dimensions	width, length, and thickness	S-2	4.0
Physical requirements	before aging, after accelerated aging	S-2	4.0
Powder Free Residue	Exceeds Maximum Limit	N=5	N/A
Powder Amount	Exceeds Recommended Maximum Limit	N=2	N/A

<sup>A</sup> See U.S. Pharmacopeia.

7.2 *Sterility Test*—Testing for sterility shall be conducted in accordance with the latest edition of The U.S. Pharmacopeia.

7.3 *Freedom from Holes*—Testing for freedom from holes shall be conducted in accordance with Test Method **D5151**.

7.4 *Physical Dimensions Test:*

7.4.1 The gloves shall comply with the dimension requirements prescribed in **Table 2**.

7.4.2 The length shall be expressed in millimetres as measured from the tip of the second finger to the outside edge of the cuff.

7.4.3 The width of the palm shall be expressed in millimetres as measured at a level between the base of the index finger and the base of the thumb. Values of width per size other than listed shall meet the stated tolerance specified in **Table 2**.

7.4.4 The minimum thickness shall be expressed in millimetres as specified in **Table 2** when using a dial micrometer described in Test Methods **D412**, and in the locations indicated in **Fig. 1**. For referee tests, cutting the glove is necessary to obtain single-thickness measurements. (See Practice **D3767** for more information.)

7.5 *Physical Requirements Test:*

7.5.1 Before and after accelerated aging, the gloves shall conform to the physical requirements specified in **Table 3**. Tests shall be conducted as specified in Test Methods **D412**. Die C is recommended.

7.5.2 Accelerated aging tests shall be conducted on samples cut from the glove in accordance with Test Method **D573** by exposing the glove to  $70 \pm 2^\circ\text{C}$  for  $72 \pm 2$  h. The glove shall withstand these conditions without evidence of tackiness, exudation, or other deterioration.

7.6 *Powder Free Gloves:*

7.6.1 Determine the powder residue using Test Method **D6124**.

7.7 *Powdered Gloves:*

7.7.1 Determine the recommended maximum powder limit using Test Method **D6124** for powdered gloves.

7.7.2 Determine the square decimetres for the glove size as in 8.7.3 in Specification **D3578**.

**8. Acceptance**

8.1 Gloves will be considered to meet the referee performance requirements when test results conform to the requirements prescribed in **Table 1**.

8.2 Retests or reinspections are permissible under the provisions of the U.S. Pharmacopeia and ISO 2859.

**9. Packaging and Package Marking**

9.1 *Sterile Packaging:*

9.1.1 The unit of packaging shall normally be one glove or one pair of gloves.

9.1.2 A glove or pair of gloves, normally, shall be enclosed in an inner wallet or wrapper. The wrapper shall be of sufficient size when opened to provide a field for glove-donning purposes.

9.1.3 The glove or pair, and accompanying wrapper if utilized, shall be totally enclosed in an outer package that will allow sterilization of the product.

9.1.4 The outer package shall have a method of closure sufficient to ensure the sterility of the product until opened or damaged.

9.1.5 The outer package shall have sufficient strength and integrity to withstand normal transportation and storage within the intermediate or shipping cartons, or both.

9.1.6 The method of closure of the outer package shall be such that prior opening will be detectable by the user.

9.1.7 None of the packaging material shall contain any material likely to impair the quality and use of the gloves.

9.1.8 Intermediate cartons and shipping cases shall be of sufficient strength to maintain the quality and sterility of the product during normal transportation and storage.

9.2 *Nonsterile and Bulk Packaging:*

9.2.1 The unit of packaging shall normally be more than one glove and of a specific amount.

9.2.2 The gloves shall be enclosed in an outer package that has sufficient strength to withstand normal transportation and storage within the cartons or shipping cases, or both.

**TABLE 2 Dimensions and Tolerances**

Designation	Size							Tolerance, mm
	6	6.5	7	7.5	8	8.5	9	
Width by size, mm	76	83	89	95	102	108	114	6
Width by small, medium, large, and extra large, mm	small		medium		large		x-large	5
Length, mm	85		95		105		115	
Thickness, mm				230 for all sizes				min
Finger				0.05				min
Palm				0.08				min




FIG. 1 Location of Thickness Measurements

TABLE 3 Physical Requirements

Tensile Strength, MPa,	Ultimate Elongation, % ,
min	min
11	300

9.2.3 None of the packaging material shall contain any material likely to impair the quality and use of the gloves.

9.2.4 Cartons and shipping cases shall be of sufficient strength to maintain the quality of the product during normal transportation and storage.

9.3 Package Marking:

9.3.1 Sterile packages shall bear markings for the contents to include the glove size, instructions for opening, the legend “sterile,” and a manufacturing lot number.

9.3.2 Nonsterile and bulk packages shall bear markings for the contents to include the glove size and a manufacturing lot number.

9.3.3 The outermost case shall be labeled with the glove size and a manufacturing lot number. Sterile product cases shall also be marked with the legend “sterile.”

9.3.4 All levels of packaging shall conform to all appropriate government labeling regulations.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; http://www.copyright.com/