



# Standard Test Methods for Bromate, Bromide, Chlorate, and Chlorite in Drinking Water by Suppressed Ion Chromatography<sup>1</sup>

This standard is issued under the fixed designation D6581; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 These multi-test methods cover the determination of the oxyhalides—chlorite, bromate, and chlorate, and bromide, in raw water, finished drinking water and bottled (non-carbonated) water by chemically and electrolytically suppressed ion chromatography. The ranges tested using these test methods for each analyte were as follows:

	Range	Sections
Test Method A: Chemically Suppressed Ion Chromatography		8 to 20
Chlorite	5 to 500 $\mu\text{g/L}$	
Bromate	1 to 25 $\mu\text{g/L}$	
Bromide	5 to 250 $\mu\text{g/L}$	
Chlorate	5 to 500 $\mu\text{g/L}$	
Test Method B: Electrolytically Suppressed Ion Chromatography		21 to 31
Chlorite	20 to 1000 $\mu\text{g/L}$	
Bromate	1 to 30 $\mu\text{g/L}$	
Bromide	20 to 200 $\mu\text{g/L}$	
Chlorate	20 to 1000 $\mu\text{g/L}$	

1.1.1 The upper limits may be extended by appropriate sample dilution or by the use of a smaller injection volume. Other ions of interest, such as fluoride, chloride, nitrite, nitrate, phosphate, and sulfate may also be determined using these test methods. However, analysis of these ions is not the object of these test methods.

1.2 It is the user’s responsibility to ensure the validity of these test methods for waters of untested matrices.

1.3 These test methods are technically equivalent with Part B of U.S. EPA Method 300.1,<sup>2</sup> titled “The Determination of Inorganic Anions in Drinking Water by Ion Chromatography.”

1.4 The values stated in either SI or inch-pound units are to be regarded as the standard. The values given in parentheses are for information only.

<sup>1</sup> These test methods are under the jurisdiction of ASTM Committee D19 on Water and are the direct responsibility of Subcommittee D19.05 on Inorganic Constituents in Water.

Current edition approved March 1, 2012. Published April 2012. Originally approved in 2000. Last previous edition approved in 2008 as D6581 – 08. DOI: 10.1520/D6581-12.

<sup>2</sup> U.S. EPA 300.1, Cincinnati, OH, 1997.

1.5 This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

## 2. Referenced Documents

### 2.1 ASTM Standards:<sup>3</sup>

- D1129 Terminology Relating to Water
- D1193 Specification for Reagent Water
- D2777 Practice for Determination of Precision and Bias of Applicable Test Methods of Committee D19 on Water
- D3370 Practices for Sampling Water from Closed Conduits
- D3856 Guide for Management Systems in Laboratories Engaged in Analysis of Water
- D5810 Guide for Spiking into Aqueous Samples
- D5847 Practice for Writing Quality Control Specifications for Standard Test Methods for Water Analysis

## 3. Terminology

3.1 *Definitions*—For definitions of terms used in the test methods, refer to Terminology D1129.

### 3.2 Definitions of Terms Specific to This Standard:

3.2.1 *analytical column*—the ion exchange column used to separate the ions of interest according to their retention characteristics prior to detection.

3.2.2 *analytical column set*—a combination of one or more guard columns, followed by one or more analytical columns used to separate the ions of interest. All of the columns in series then contribute to the overall capacity and resolution of the analytical column set.

3.2.3 *eluent*—the ionic mobile phase used to transport the sample through the chromatographic system.

3.2.4 *guard column*—a column used before the analytical column to protect it from contaminants, such as particulates or irreversibly retained material.

<sup>3</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For Annual Book of ASTM Standards volume information, refer to the standard’s Document Summary page on the ASTM website.

3.2.5 *ion chromatography*—a form of liquid chromatography in which ionic constituents are separated by ion exchange then detected by an appropriate detection means, typically conductance.

3.2.6 *resolution*—the ability of an analytical column to separate the method analytes under specific test conditions.

3.2.7 *suppressor device*—an ion exchange based device that is placed between the analytical column set and the conductivity detector. Its purpose is to minimize detector response to the ionic constituents in the eluent, in order to lower background conductance; and at the same time enhance the conductivity detector response of the ions of interest.

3.2.7.1 *chemical suppression*—the use of an acid solution to the suppressor in order to suppress the background conductivity.

3.2.7.2 *electrolytic suppressor device*—electrolytic suppression is an ion exchange device that is placed between the analytical column and the conductivity detector. Its purpose is similar to a suppressor device, however, it does not require addition of acid. Instead the electrolytic suppressor generates protons electrolytically and plugs into an electrical power source on typically located on the chromatography device.

#### 4. Significance and Use

4.1 The oxyhalides chlorite, chlorate, and bromate are inorganic disinfection by-products (DBPs) of considerable health risk concern worldwide. The occurrence of chlorite and chlorate is associated with the use of chlorine dioxide, as well as hypochlorite solutions used for drinking water disinfection. The occurrence of bromate is associated with the use of ozone for disinfection, wherein naturally occurring bromide is oxidized to bromate. Bromide is a naturally occurring precursor to the formation of bromate.

#### 5. Reagents and Materials

5.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.<sup>4</sup> Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without reducing the accuracy of the determination.

5.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Specification **D1193**, Type I. Other reagent water types may be used, provided it is first ascertained that the water is of sufficiently high purity to permit its use without adversely affecting the bias and precision of the determination.

#### 6. Precautions

6.1 These test methods address the determination of very low concentrations of selected anions. Accordingly, every

precaution should be taken to ensure the cleanliness of sample containers as well as other materials and apparatus that come in contact with the sample.

#### 7. Sampling and Sample Preservation

7.1 Collect the sample in accordance with Practice **D3370**, as applicable.

7.2 Immediately upon taking the sample, sparge it with an inert gas (for example, nitrogen, argon, or helium) for 5 minutes to remove active gases such as chlorine dioxide or ozone. Add 1.00 mL of EDA Preservation Solution (see **15.3**) per 1.000 litre of sample to prevent conversion of residual hypochlorite or hypobromite to chlorate or bromate. This also prevents metal catalyzed conversion of chlorite to chlorate. The oxyhalides in samples preserved in this manner are stable for at least 14 days when stored in amber bottles at 4°C.<sup>5</sup>

#### Test Method A

#### Chemically Suppressed Ion Chromatography

#### 8. Scope

8.1 This test method covers the determination of the oxyhalides—chlorite, bromate, and chlorate, and bromide, in raw water, finished drinking water and bottled (non-carbonated) water by chemically suppressed ion chromatography. The ranges tested using this test method for each analyte were as follows:

Chlorite	5 to 500 µg/L
Bromate	1 to 25 µg/L
Bromide	5 to 250 µg/L
Chlorate	5 to 500 µg/L

8.1.1 The upper limits may be extended by appropriate sample dilution or by the use of a smaller injection volume. Other ions of interest, such as fluoride, chloride, nitrite, nitrate, phosphate, and sulfate may also be determined using this test method. However, analysis of these ions is not the object of this test method.

8.2 It is the user's responsibility to ensure the validity of this test method for waters of untested matrices.

8.3 This test method is technically equivalent with Part B of U.S. EPA Method 300.1,<sup>2</sup> titled "The Determination of Inorganic Anions in Drinking Water by Ion Chromatography."

8.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

#### 9. Summary of Test Method A

9.1 Oxyhalides (chlorite, bromate, and chlorate) and bromide in raw water, finished drinking water and bottled water are determined by ion chromatography. A sample (200 µL) is injected into an ion chromatograph and the pumped eluent (sodium carbonate) sweeps the sample through the analytical

<sup>4</sup> "Reagent Chemicals, American Chemical Society Specifications," Am. Chemical Soc., Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see "Analar Standards for Laboratory Chemicals," by BDH Ltd., Poole, Dorset, U.K., and the "United States Pharmacopoeia."

<sup>5</sup> Hautman, D. P., and Bolyard, M., *Journal of Chromatography*, Vol 602, 1992, p. 65.

column set. Here, anions are separated from the sample matrix according to their retention characteristics, relative to the anions in the eluent.

9.1.1 The separated anions in the eluent stream then pass through a suppressor device, where all cations are exchanged for hydronium ions. This converts the eluent to carbonic acid, thus reducing the background conductivity. This process also converts the sample anions to their acid form, thus enhancing their conductivity. The eluent stream then passes through a conductivity cell, where they are detected. A chromatographic integrator or appropriate computer-based data system is typically used for data presentation.

9.2 The anions are identified based on their retention times compared to known standards. Quantification is accomplished by measuring anion peak areas and comparing them to the areas generated from known standards.

**10. Interferences**

10.1 Positive errors can be caused by progressive oxidation of residual hypochlorite or hypobromite, or both, in the sample to the corresponding chlorate and bromate. Furthermore, chlorite can also be oxidized to chlorate, causing negative errors for chlorite and positive errors for chlorate. These interferences are eliminated by the sample preservation steps outlined in 15.3. Chloride present at >200 mg/L can interfere with bromate determination.

**11. Apparatus**

11.1 *Ion Chromatography Apparatus*—Analytical system complete with all required accessories, including eluent pump, injector, syringes, columns, suppressor, conductivity detector, data system, and compressed gasses.

11.1.1 *Eluent Pump*—Capable of delivering 0.10 to 5.0 mL/min of eluent at a pressure of up to 4000 psi (27600 kPa).

11.1.2 *Injection Valve*—A low dead-volume switching valve that will allow the loading of a sample into a sample loop and subsequent injection of the loop contents into the eluent

stream. A loop size of up to 50 µL may be used without compromising the resolution of early eluting peaks, such as chlorite and bromate.

11.1.3 *Guard Column*—Anion exchange column typically packed with the same material used in the analytical column. The purpose of this column is to protect the analytical column from particulate matter and irreversibly retained material.

11.1.4 *Analytical Column*—Anion exchange column capable of separating the ions of interest from each other, as well as from other ions which commonly occur in the sample matrix. The separation shall be at least as good as that shown in Fig. 1. Conditions of the eluent may vary by column manufacturer.

NOTE 1—The Analytical Column Set (see 3.2.2) should be able to give baseline resolution of all anions, even for a 50 µL injection containing up to 200 mg/L, each, of common anions, such as chloride, bicarbonate, and sulfate.

11.1.5 *Suppressor Device*—A suppressor device based upon cation exchange principles. In this test method, simultaneously regenerating suppressor device with sequential carbonate remover was used. An equivalent suppressor device may be used provided that comparable method detection limits are achieved and that adequate baseline stability is attained.

11.1.6 *Conductivity Detector*—A low-volume, flow through, temperature stabilized conductivity cell equipped with a meter capable of reading from 0 to 15 000 µS/cm on a linear scale.

11.1.7 *Data System*—A chromatographic integrator or computer-based data system capable of graphically presenting the detector output signal versus time, as well as presenting the integrated peak areas.

**12. Example of Chromatogram—IC Conditions—1**

12.1 See Fig. 1, Fig. 2, and Table 1.

**13. Example of Chromatogram—IC Conditions—2**

13.1 A carbonate removal device is developed to remove the majority of the carbonate from the eluent and allow hydroxide-like performance with improved detection sensitivity. This




FIG. 1 Chromatogram of a Standard Containing Low µg/L Oxyhalides, and Bromide, in the Presence of Common Inorganic Anions (See Table 1 for Analysis Conditions)




FIG. 2 Overlay Chromatogram Spiked and Unspiked of Low µg/L Oxyhalides, and Bromide, in Houston, TX, Drinking Water (See Table 1 for Analysis Conditions)

TABLE 1 Instrumentation<sup>A</sup> and Operating Conditions for the Determination of Oxyhalides and Bromide and by Ion Chromatography, as shown in Figs. 1 and 2

Ion Chromatograph:	Metrohm 850 Professional IC <sup>A</sup> (or equivalent)
Guard Column:	Metrosep ASUPP4/5 <sup>A</sup> (or equivalent)
Analytical Column:	Metrosep ASUPP7 <sup>A</sup> (or equivalent)
Eluent:	3.5 mM Sodium carbonate
Flow-Rate:	0.7 mL/min
Injection Volume:	50 µL
Suppressor:	Metrohm MSM-II <sup>A</sup> (Tri-Chamber Micro-packed) or equivalent coupled sequentially with MCS <sup>B</sup>
Detector:	Conductivity Detector stabilized at 40°C

<sup>A</sup> Metrohm AG, Switzerland.

<sup>B</sup> MCS is a carbonate suppressor device that permits use of carbonate/bicarbonate buffer based eluent in ion chromatography to achieve greater sensitivity and better detection limits.

device, the CRD-300, was used with the IonPac AS23 to determine bromate in a bottled mineral water samples. This data shows the improved detection sensitivity when using the CRD-300 compared to chromatography without the CRD-300. Scientists responsible for water analysis can chose the column and eluent chemistry that best meets their needs to reliably determine bromate at concentrations below the common 10 µg/L regulatory limit.

	Conditions
	<b>Condition A with and without CRD-300</b>
Column:	IonPac AS23 (4 × 250 mm) IonPac AG23 (4 × 50 mm)
Eluent:	4.5 mM K <sub>2</sub> CO <sub>3</sub> / 0.8 mM KHCO <sub>3</sub>
Flow rate:	1.0 mL/min
Suppressor:	Suppressed conductivity ASRS-300, 4 mm External water mode, CRD-300 4 mm, Vacuum mode
Background:	<1.5 µS
Noise:	~0.3 nS

13.2 See Fig. 3, Fig. 4, Table 2, Table 3, and Table 4.

#### 14. Preparation of Apparatus

14.1 Set up the ion chromatograph according to the manufacturer’s instructions. If an Anion Self Regenerating Suppres-

or is used, operate the device at 100 mA in the external water mode. The conductivity detector cell should be thermally stabilized at 35°C.

14.2 The recommended operating conditions for the ion chromatograph are summarized in Table 1.

14.3 The detector ranges are variable. Normal operating ranges for quantifying the low level of oxyhalides encountered in treated drinking water are in the 0.2 to 2 µS/cm full scale range. Choose a range consistent with the concentration range in the expected samples and with the operating requirements of the chromatographic system used.

14.4 Equilibrate the chromatographic system by pumping the analysis eluent (see 15.2) through the system until a stable baseline is obtained (approximately 20 minutes). Typical baseline characteristics necessary to obtain the method detection limits required for this analysis are: (1) a background conductance of 20 to 25 µS/cm and (2) a peak-to-peak (noise) variation of no greater than 5 nS/cm per minute of monitored baseline response.

#### 15. Reagents and Materials


15.1 *Eluent, Concentrate (90.0 mM Sodium Carbonate)*—Dissolve 9.540 g of sodium carbonate in 1000 mL of water.

15.2 *Eluent, Analysis (9.0 mM Sodium Carbonate)*—Dilute 100.0 mL of Eluent Concentrate (see 15.1) to 1.000 L with water.

15.2.1 The Eluent Analysis solution (9.0 mM Sodium Carbonate) must be purged for 10 minutes with helium prior to use to remove dissolved gasses in order to ensure optimal system performance.

15.3 *Ethylenediamine (EDA) Preservation Solution (50.0 g/L)*—Dilute 11.2 mL of ethylenediamine (99 %) to 200 mL with reagent water. Prepare this solution fresh monthly. Add 1.00 mL of this solution per 1.000 L of blank, standard or sample to produce a final EDA concentration of 50 mg/L.

15.4 *SPE Sample Pretreatment Cartridges*—Chloride present at >200 mg/L and carbonate present at >300 mg/L can



NOTE 1—Peaks:

- |                        |                        |                         |
|------------------------|------------------------|-------------------------|
| 1. Fluoride (0.1 mg/L) | 5. Nitrite (0.3 mg/L)  | 8. Nitrate (0.5 mg/L)   |
| 2. Chlorite (0.3 mg/L) | 6. Chlorate (0.5 mg/L) | 9. Phosphate (0.8 mg/L) |
| 3. Bromate (0.6 mg/L)  | 7. Bromide (0.5 mg/L)  | 10. Sulfate (0.6 mg/L)  |
| 4. Chloride (0.2 mg/L) |                        |                         |

FIG. 3 Chromatography of a Mixed Anion Standard with a CRD-300 (A) and without a CRD-300 (B)

interfere with bromate determination.  $H^+$  form and  $Ag^+$  form cation exchange SPE cartridges can be used to minimize the carbonate and chloride interferences, respectively, if required. OnGuard-H and OnGuard-Ag cartridges have been shown to be suitable for this application.<sup>6</sup> The use of these pretreatment cartridges will effect recoveries for bromide, requiring that it be analyzed in a separate run.

15.5 *Suppressor Regenerant Solution*—If a suppressor requiring chemical regeneration is used, the regenerant solution is prepared by cautiously adding 3.00 mL of concentrated sulfuric acid (sp. gr. 1.84) to 4.000 L of water. If an Anion Self Regenerating Suppressor is used, it should be operated in the external water mode.

15.6 *Standard Solutions, Stock (1.00 mL = 1.00 mg)*—Purchase certified solutions or prepare stock standard solutions from the following salts, as described below:


15.6.1 *Bromate ( $BrO_3^-$ ) Solution, Stock (1.00 mL = 1.00 mg  $BrO_3^-$ )*—Dissolve 1.180 g of sodium bromate ( $NaBrO_3$ ) in water and dilute to 1.000 L.

15.6.2 *Bromide ( $Br^-$ ) Solution, Stock (1.00 mL = 1.00 mg  $Br^-$ )*—Dissolve 1.288 g of sodium bromide ( $NaBr$ ) in water and dilute to 1.000 L.

15.6.3 *Chlorate ( $ClO_3^-$ ) Solution, Stock (1.00 mL = 1.00 mg  $ClO_3^-$ )*—Dissolve 1.275 g of sodium chlorate ( $NaClO_3$ ) in water and dilute to 1.000 L.

15.6.4 *Chlorite ( $ClO_2^-$ ) Solution, Stock (1.00 mL = 1.00 mg  $ClO_2^-$ )*—Dissolve 1.680 g of sodium chlorite ( $NaClO_2$ ) in water and dilute to 1.000 L. Note that as sodium chlorite is

<sup>6</sup> Joyce, R. J., and Dhillon, H. J., *Chromatography*, Vol 671, 1994, p. 165.



NOTE 1—Peaks:

- |                                  |                                  |                       |
|----------------------------------|----------------------------------|-----------------------|
| 1. Fluoride (0.5 mg/L)           | 4. Chloride (50 mg/L)            | 6. Unknown            |
| 2. Chlorite (10, 20 and 40 µg/L) | 5. Chlorate (10, 20 and 40 µg/L) | 7. Sulfate (100 mg/L) |
| 3. Bromate (10, 20 and 40 µg/L)  |                                  |                       |

**FIG. 4 Overlay of Chromatograms of Three Concentration Levels of Chlorite, Bromate, and Chlorate in a Mixed Anion Standard with a CRD-300 (A) and without a CRD-300 (B)**

**TABLE 2 Calibration Report for Chlorite, Bromate, and Chlorate with a CRD-300 (A) and without a CRD-300 (B)**

Peak Name	Points	R-Square (%)	
		A	
Chlorite	3	99.9961	99.9748
Bromate	3	100.0000	99.9986
Chlorate	3	99.9995	99.9637

weight of salt used after determining the exact percentage of NaClO<sub>2</sub>, which can be done using an iodometric titration procedure.<sup>2,7</sup>

15.7 *Reagent Blank*—Add 1.00 mL of EDA Preservation Solution (see 15.3) to 1.000 L of reagent water.

usually available only as an 80 % technical grade salt, the 80 % purity is accounted for in the 1.680 g weight cited above. If an alternate purity is used, make an appropriate adjustment in the

<sup>7</sup> Method 4500-ClO<sub>2</sub>-C in A. E. Greenberg, L. S. Clesceri, A. D. Eaton (Eds.), *Standard Methods for the Examination of Water and Wastewater*, 18th Ed., APHA, Washington, DC, 1992.

**TABLE 3 Determination of Bromate and Chlorate in a Bottled Mineral Water Sample**

Injection No.	A ( $\mu\text{g/L}$ )		B ( $\mu\text{g/L}$ )	
	Bromate	Chlorate	Bromate	Chlorate
1	11.0	1.52	5.33	ND
2	10.9	1.55	6.23	ND
3	10.9	1.35	5.02	ND
4	10.1	1.91	6.25	ND
5	11.3	1.48	5.89	ND
Average	10.8	1.56	5.74	—
RSD	4.34	13.42	9.61	—

**TABLE 4 Spike Recovery of Bromate, Chlorate, and Chlorite in Mineral Water Using a System with a CRD-300 (A) and without a CRD-300 (B)**

Injection No.	A ( $\mu\text{g/L}$ )			B ( $\mu\text{g/L}$ )		
	Chlorite	Bromate	Chlorate	Chlorite	Bromate	Chlorate
Sample	ND <sup>A</sup>	10.83	1.56	ND	5.74	ND
Spike	10	10	10	10	10	10
Measured <sup>B</sup> Amount	9.88	20.51	12.02	8.58	15.30	8.50
RSD	2.39	1.60	2.45	2.39	1.60	2.45
Recovery (%)	98.8	98.5	104	85.8	97.2	85.0

<sup>A</sup> ND = Not Detected

<sup>B</sup> The average of five injections.

## 16. Calibration and Standardization

16.1 *Typical Range of Applicability*—This test method is applicable to the determination of bromate, bromide, chlorate, and chlorite in raw water, finished drinking water and bottled (non-carbonated) water. The application ranges tested for each analyte are as follows: bromate; 5–30  $\mu\text{g/L}$ , bromide; 20–200  $\mu\text{g/L}$ , chlorite; 20–500  $\mu\text{g/L}$ , and chlorate; 20–500  $\mu\text{g/L}$ .

16.2 *Calibration Standards*—For each individual calibration curve, prepare calibration standards, at a minimum of three concentration levels, by accurately adding measured volumes of the stock standards (see 15.6) to a volumetric flask(s). Add 50 mg/L of EDA (the equivalent of 1.00 mL of EDA Preservation Solution (see 15.3) per 1.000 L of solution) to the volumetric flask(s) and dilute to volume with reagent water. A minimum of five concentration levels is recommended if the curve covers two orders of magnitude.

16.3 *Calibration Curve*—To establish the calibration curve, analyze a reagent blank and the calibration standards in accordance to the procedure in Section 17, using a 200  $\mu\text{L}$  injection (with a 4 mm ID column) or a 50  $\mu\text{L}$  injection (with a 2 mm ID column). Tabulate peak area responses against concentration. These results are used to prepare a calibration curve using a linear least squares fit for each analyte. The squared correlation coefficient of the regression ( $r^2$ ) should be  $\geq 0.995$  for accurate results. Once the calibration curves have been established, verification must be performed on each analysis day, whenever fresh eluent is prepared, and twice each batch of samples, as outlined in 20.4 and 20.5.

## 17. Procedure

17.1 Inject the reagent blank, calibration standard or sample into the eluent stream and record the chromatogram. In the case of a manual injector, flush an excess of the sample (minimum

of 5x loop volume) through the sample injection port using a syringe prior to injection. A 200- $\mu\text{L}$  injection is required when using a 4 mm ID column, a 50  $\mu\text{L}$  injection is required when using a 2 mm ID column, in order to achieve the required detection limits for this analysis. An example of a chromatogram of low level oxyhalides and bromide is shown in Fig. 1. An example chromatogram of low level oxyhalides and bromide in a modest ionic strength, simulated drinking water is shown in Fig. 2.

## 18. Calculation

18.1 Compare the peak areas for the anions in the sample to the calibration curves prepared in 16.3 to calculate and report the anion concentration in  $\mu\text{g/L}$ :

$$\text{Anion concentration, } \mu\text{g/L} = A \times F \quad (1)$$

where:

$A$  = reading from the appropriate calibration plot, in  $\mu\text{g/L}$ , and

$F$  = dilution factor if the sample was diluted prior to analysis.

18.1.1 Computing integrators and computer based chromatographic data systems can be programmed to perform these calculations automatically.

18.2 Report only those values that fall between the lowest and highest calibration standards. Samples exceeding the highest standard should be diluted and reanalyzed.

## 19. Precision and Bias

19.1 The precision and bias data presented in this test method meets the requirements of Practice D2777 – 98, which was in place at the time of collaborative testing. Under the allowances made in 1.4 of D2777 – 06, these precision and bias

data do meet existing requirements for interlaboratory studies of Committee D19 test methods. The full research report can be obtained from ASTM Headquarters.<sup>8</sup>

19.2 The interlaboratory study that generated the precision and bias data in this test method was performed in reagent water, municipal drinking water, and bottled (non-carbonated) water by ten laboratories using one operator each. Six levels of concentration were used for four analytes, producing three Youden pairs. The Youden pair data was used to calculate the single operator precision ( $S_o$ ). The analytes were supplied separately as six (mixed) concentrates. The reagent water, municipal drinking water and bottled water were supplied by the participating laboratories. Six reagent water samples, six bottled water samples, and six municipal drinking water samples (for a total of 18 samples) were prepared by pipetting 1.0 mL aliquots of the concentrates labeled A1-2 (×3), B1-2 (×3), C1-2 (×3) into volumetric flasks (18 total); adding 50 mg/L EDA preservation solution, as detailed in 7.2, and diluting to a total of 100 mL with reagent water (×6), bottled water (×6), and drinking water (×6), as appropriate.

19.2.1 A quality control (QC) sample was supplied (as a concentrate) to serve as initial, and on-going, calibration verification. A separate method detection limit (MDL) sample was supplied (as a concentrate) for the determination of the pooled MDL values. The QC sample was prepared by pipetting a 1.0 mL aliquot of the QC concentrate into a clean volumetric flask; adding 50 mg/L EDA, and diluting to a total of 100 mL with reagent water. The MDL sample was prepared by pipetting a 1.0 mL aliquot of the MDL concentrate into a clean volumetric flask; adding 50 mg/L EDA, and diluting to a total of 100 mL with reagent water.

19.3 All the precision and bias data presented in this test method was obtained using the IonPac AS9-HC column listed in Table 1.

19.4 The precision and bias of this test method for each analyte for reagent, drinking, and bottled water are shown in Tables 5-8.

19.5 The results of the interlaboratory study can also be summarized as regression equations, as shown in Table 9 for reagent water and in Table 10 for a typical sample matrix of drinking water.

19.6 In addition to performing the analyses required to generate the precision and bias data shown in Tables 5-8, the participating laboratories each analyzed seven replicates of an MDL sample. The MDLs were derived for each laboratory using the students *t*-test at six degrees of freedom, as follows:

$$MDL = (t) \times (S) \quad (2)$$

where:

- t* = students *t* value for a 99 % confidence level and a standard deviation estimate with n-1 degrees of freedom [*t* = 3.14 for seven replicates], and
- S* = standard deviation of the replicate analysis.<sup>8</sup>

19.6.1 True amounts injected, mean value determined, and pooled MDL values (10 laboratories × 7 replicates) are shown in Table 11.

## 20. Quality Control

20.1 Before this test method is applied to analyzing unknown samples, the analyst should establish quality control procedures as recommended in Guide D3856.

20.2 The laboratory using this test should perform an initial demonstration of laboratory capability. Analyze seven replicates of an Initial Demonstration of Performance (IDP) solution. The IDP solution contains method analytes of known concentration, prepared from a different source to the calibration standards, used to fortify reagent water, which also contains a final EDA concentration of 50 mg/L (see 15.3). Ideally, the IPD solution should be prepared by an independent source from reference materials. The level 3 standard used for the method precision and bias study is recommended as an IDP solution.

<sup>8</sup> Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D19-1164. Contact ASTM Customer Service at service@astm.org.

**TABLE 5 Determination of Precision and Bias for Chlorite**

Water	Amount Added (µg/L)	Amount Found (µg/L)	Number Retained Parts	$S_o$ (µg/L)	$S_t$ (µg/L)	Bias (%)
Reagent	20	19.94	8	1.40	1.25	-0.3
	25	25.06			1.05	0.2
	180	178.29			5.64	-1.0
	220	214.73	8	4.37	6.18	-2.4
	400	394.36			8.29	-1.4
	450	440.53			21.81	-2.1
Drinking	20	19.19	8	1.52	6.58	-4.1
	25	23.77			6.17	-4.9
	180	174.30			9.29	-3.2
	220	216.89	8	5.06	14.76	-1.4
	400	398.30			15.56	-0.4
	450	439.85			19.59	-2.3
Bottled	20	20.94	8	4.41	3.65	4.7
	25	22.74			4.64	-9.0
	180	177.71			8.76	-1.3
	220	216.16	8	2.95	8.74	-1.7
	400	390.14			13.65	-2.5
	450	433.72			15.30	-3.6



**TABLE 6 Determination of Precision and Bias for Bromate**

Water	Amount Added (µg/L)	Amount Found (µg/L)	Number Retained Parts	S <sub>o</sub> (µg/L)	S <sub>i</sub> (µg/L)	Bias (%)
Reagent	5	4.95	9	0.99	1.19	-0.9
	7	7.84			1.29	12.0
	10	9.98	9	0.66	0.99	-0.2
	12	11.76			0.55	-2.0
	20	19.56	9	2.17	2.37	-2.2
	25	24.18			1.53	-3.3
Drinking	5	4.41	6	0.62	1.20	-11.8
	7	6.44			0.78	-8.0
	10	8.54	8	0.72	2.88	-14.6
	12	10.20			2.88	-15.0
	20	17.31	8	2.80	2.85	-13.4
	25	20.51			4.77	-17.9
Bottled	5	4.95	6	1.09	1.51	-1.1
	7	7.21			1.80	3.0
	10	9.07	6	1.67	4.83	-9.3
	12	10.35			3.41	-13.7
	20	20.16	6	4.34	3.91	0.8
	25	20.99			7.75	-16.0

**TABLE 7 Determination of Precision and Bias for Bromide**

Water	Amount Added (µg/L)	Amount Found (µg/L)	Number Retained Parts	S <sub>o</sub> (µg/L)	S <sub>i</sub> (µg/L)	Bias (%)
Reagent	20	20.75	9	1.94	1.91	3.8
	25	25.51			2.32	2.1
	75	74.52	9	3.80	2.94	-0.6
	100	99.42			4.84	-0.6
	150	143.50	9	5.79	5.82	-4.3
	180	176.38			5.24	-2.0
Drinking	20	20.68	8	1.30	4.39	3.4
	25	25.49			3.31	2.0
	75	71.89	8	4.67	5.67	-4.2
	100	97.05			6.90	-3.0
	150	145.81	8	1.23	8.39	-2.8
	180	173.40			9.12	-3.7
Bottled	20	20.25	7	2.21	1.79	1.3
	25	26.23			1.11	4.9
	75	74.35	8	6.45	4.69	-0.9
	100	98.49			5.00	-1.5
	150	142.67	8	6.57	6.08	-4.9
	180	172.71			9.24	-4.1

**TABLE 8 Determination of Precision and Bias for Chlorate**

Water	Amount Added (µg/L)	Amount Found (µg/L)	Number Retained Parts	S <sub>o</sub> (µg/L)	S <sub>i</sub> (µg/L)	Bias (%)
Reagent	20	20.69	7	2.73	2.43	3.5
	25	26.64			3.79	6.6
	180	176.05	8	11.96	3.70	-2.2
	220	215.39			7.47	-2.1
	400	393.00	7	18.27	5.85	-1.7
	450	443.47			16.50	-1.5
Drinking	20	19.94	9	1.81	3.95	-0.3
	25	23.93			5.13	-4.3
	180	175.10	9	6.92	9.05	-2.7
	220	216.14			7.02	-1.8
	400	396.74	9	4.74	16.55	-0.8
	450	441.69			16.55	-1.8
Bottled	20	21.72	8	2.86	3.88	8.6
	25	25.75			3.21	3.0
	180	179.82	8	3.59	5.37	-0.1
	220	217.58			9.26	-1.1
	400	389.51	7	6.72	15.83	-2.6
	450	443.70			10.00	-1.4

20.2.1 The mean and standard deviation of the seven values should then be calculated and compared, according to Practice **D5847**, to the single operator precision and recovery estab-

lished for this test method. The upper limit for acceptable precision and the range of acceptable recoveries are detailed below:

**TABLE 9 Summary of Precision and Bias Results for Reagent Water**

	Chlorite Precision and Bias Summary	Bromate Precision and Bias Summary	Bromide Precision and Bias Summary	Chlorate Precision and Bias Summary
Number of Laboratories	10	10	10	10
Range Tested	20–450 µg/L	5–25 µg/L	20–180 µg/L	20–450 µg/L
Mean Recovery	$y = 0.9805x + 0.5261$	$y = 0.9432x + 0.6272$	$y = 0.9629x + 1.7475$	$y = 0.9809x + 0.8245$
S <sub>o</sub>	$y = 0.0465x - 1.4801$	$y = 0.0878x + 0.1281$	$y = 0.0282x + 1.3087$	$y = 0.0389x + 2.7278$
S <sub>t</sub>	$y = 0.0332x + 0.3294$	$y = 0.046x + 0.721$	$y = 0.0246x + 1.6352$	$y = 0.0226x + 1.8244$

**TABLE 10 Summary of Precision and Bias Results for Drinking Water**

	Chlorite Precision and Bias Summary	Bromate Precision and Bias Summary	Bromide Precision and Bias Summary	Chlorate Precision and Bias Summary
Number of Laboratories	10	10	10	10
Range Tested	20–450 µg/L	5–25 µg/L	20–180 µg/L	20–450 µg/L
Mean Recovery	$y = 0.9872x - 1.0243$	$y = 0.9432x + 0.6272$	$y = 0.9583x + 1.2113$	$y = 0.9868x - 0.7347$
S <sub>o</sub>	$y = 0.0068x + 2.2164$	$y = 0.1721x - 0.5532$	$y = -0.0022x + 2.6$	$y = 0.0066x + 3.0956$
S <sub>t</sub>	$y = 0.0289x + 5.8552$	$y = 0.1934x + 0.3866$	$y = 0.0357x + 3.1189$	$y = 0.03x + 3.3368$

**TABLE 11 Pooled MDL Values Obtained for This Test Method**

Analyte	Injected Amount	Mean Value	Pooled MDL
Chlorite	3.0 µg/L	3.32 µg/L	2.39 µg/L
Bromate	4.0 µg/L	3.98 µg/L	2.73 µg/L
Bromide	4.0 µg/L	3.96 µg/L	2.91 µg/L
Chlorate	4.0 µg/L	3.74 µg/L	3.49 µg/L

20.4 A continuing CVS should be analyzed after every tenth field sample and an end CVS should analyzed at the end of the sample batch (maximum of 20 samples) to verify the previously established calibration curves. After initially meeting the requirements of 20.3, the levels selected for the continuing and end CVS should be varied between a middle calibration level and the highest calibration level standard. If the continuing and end CVS values are not within ±15 % of the known amounts, the analyst should reanalyze the CVS. If the analyte concentrations still fall outside acceptable limits (±15 %) that analyte is judged out of control, and the source of the problem should be identified before continuing with on-going analyses. All samples following the last acceptable CVS should be reanalyzed.

20.5 A reagent blank (see 15.7) should be run when generating the initial calibration curves. A blank should also be run with each sample batch (maximum of 20 samples) to check for sample or system contamination.

20.6 One Laboratory Control Sample (LCS) should be used with each sample batch (maximum of 20 samples). The LCS is a solution of method analytes of known concentration added to a matrix which sufficiently challenges the Test Method. A synthetic drinking water matrix, containing fluoride at 1.0 mg/L, chloride at 50 mg/L, nitrite at 0.1 mg/L, nitrate at 10 mg/L, phosphate at 0.1 mg/L and sulfate at 50 mg/L, spiked with the four method analytes at the level of the IDP solution would be an example of an appropriate LCS. The LCS shall also contain 50 mg/L of EDA (the equivalent of 1.00 mL of EDA Preservation Solution (see 15.3) per 1.000 L of solution).

20.6.1 The analyte recoveries for the LCS should fall within the control limits of  $x \pm 3S$ , where  $x$  is the mean recovery and ( $S$ ) is the standard deviation of the mean recovery established from the interlaboratory precision and bias study data at the IDP levels, as shown below:

Analyte	LCS Amount	Lower Recovery Limit	Upper Recovery Limit
Chlorite	180 µg/L	165 µg/L	191 µg/L
Bromate	10 µg/L	8.0 µg/L	12.0 µg/L
Bromide	75 µg/L	63 µg/L	86 µg/L
Chlorate	180 µg/L	140 µg/L	219 µg/L

20.7 One Matrix Spike (MS) should be run with each sample batch (maximum of 20 samples) to test method

20.2.2 The S<sub>o</sub> and mean recovery values can be calculated for different IDP solution concentrations using the regression equations for each analyte shown in Table 9. If the values obtained for the IDP precision and recovery do not meet the criteria described above, initial demonstration of performance must be repeated until the results fall within these criteria.

20.3 When beginning use of this method, an initial Calibration Verification Standard (CVS) should be used to verify the calibration standards and acceptable instrument performance. This verification should be performed on each analysis day or whenever fresh eluent has been prepared. As this method is intended for use at trace levels, a low level CVS (that is, equivalent to the lowest calibration standard) should initially be analyzed before beginning use of this method. The CVS is a solution of method analytes of known concentration used to fortify reagent water, which also contains a final EDA concentration of 50 mg/L (see 15.3). If the determined low level CVS values are not within ±25 % of the known amounts, the low level CVS should be reanalyzed. If the values still fall outside acceptable limits, a new calibration curve is required which must be confirmed by a successful low level CVS before continuing with on-going analyses.

recovery. The MS should be prepared in accordance with Guide [D5810](#). Spike a portion of a drinking water (or other) sample from each batch with the four method analytes at the level of the IDP solution. The % recovery of the spike should fall within limits established from the interlaboratory precision and bias study data (assuming a background level of zero), according to Practice [D5847](#), as shown below:

Analyte	MS Amount	Lower Recovery Limit (%)	Upper Recovery Limit (%)
Chlorite	180 µg/L	90.9 %	109.1 %
Bromate	10 µg/L	73.1 %	126.9 %
Bromide	75 µg/L	80.8 %	119.2 %
Chlorate	180 µg/L	88.3 %	111.7 %

20.8 One Matrix Duplicate (MD) should be run with each sample batch (maximum of 20 samples) to test method precision. If non-detects are expected in all the samples to be analyzed, a Matrix Spike Duplicate should be run instead. The precision of the duplicate analysis should be compared, according to Practice [D5847](#), to the nearest tabulated  $S_o$  value established from the interlaboratory precision and bias study data for each analyte.

20.9 In order to verify the quantitative values produced by the test method, an Independent Reference Material (IRM), submitted to the laboratory as a regular sample (if practical), should be analyzed once per quarter. The concentration of the IRM should be within the scope of the method, as defined in [1.1](#). The values obtained must fall within the limits specified by the outside source.

20.10 The laboratory may perform additional quality control as desired or appropriate, for instance the use of a surrogate as outlined in Section 9.4.2 of U.S. EPA Method 300.1. In addition, it is recommended that a laboratory determine the method detection limits, as discussed in [19.6](#), before using this test method.

## Test Method B Electrolytically Suppressed Ion Chromatography

### 21. Scope

21.1 This test method is technically consistent with US EPA Method 300.1 (Part B), as cited above, except that it uses analyte separation on a hydroxide-selective anion exchange column, using a hydroxide eluent followed by electrolytically suppressed conductivity detection of the analytes in a deionized water matrix, thus improving method detection limits. The range tested for each analyte were as follows:

Chlorite	20 to 1,000 µg/L
Bromate	1 to 30 µg/L
Chlorate	20 to 1,000 µg/L
Bromide	20 to 200 µg/L

21.1.1 It should be noted that the U.S. EPA maintains that the use of hydroxide-based separation is consistent with Method 300.1 (Parts A and B) for compliance monitoring.

### 22. Summary of Test Method B

22.1 Oxyhalides (chlorite, bromate, and chlorate) and bromide in raw water, finished drinking water and bottled water are determined by ion chromatography. A sample (200 µL) is injected into an ion chromatograph and the pumped hydroxide

eluent sweeps the sample through the analytical column set. Here, anions are separated from the sample matrix according to their retention characteristics, relative to the anions in the eluent.

22.1.1 The separated anions in the eluent stream then pass through a suppressor device, where all cations are exchanged for hydronium ions. This converts the eluent to water, thus reducing the background conductivity. This process also converts the sample anions to their acid form, thus enhancing their conductivity. The eluent stream then passes through a conductivity cell, where they are detected. An appropriate computer-based data system is typically used for data presentation.

22.2 The anions are identified based on their retention times compared to known standards. Quantification is accomplished by measuring anion peak areas and comparing them to the areas generated from known standards.

### 23. Interferences

23.1 Positive errors can be caused by progressive oxidation of residual hypochlorite or hypobromite, or both, in the sample to the corresponding chlorate and bromate. Furthermore, chlorite can also be oxidized to chlorate, causing negative errors for chlorite and positive errors for chlorate. These interferences are eliminated by the sample preservation steps outlined in [26.3](#). Chloride present at >200 mg/L and carbonate present at >300 mg/L can interfere with bromate determination. These interferences can be minimized, or eliminated, by the sample pretreatment steps outlined in [26.4](#). Fluoride and low molecular weight monocarboxylic acids, present at mg/L concentrations, may interfere with the quantitation of chlorite and bromate.

### 24. Apparatus

24.1 *Ion Chromatography Apparatus*—Analytical system complete with all required accessories, including eluent pump, injector, syringes, columns, suppressor, conductivity detector, data system and compressed gasses.

24.1.1 *Eluent Pump*—Capable of delivering 0.25 to 5 mL/min of eluent at a pressure of up to 4000 psi.

24.1.2 *Injection Valve*—A low dead-volume switching valve that will allow the loading of a sample into a sample loop and subsequent injection of the loop contents into the eluent stream. A loop size of up to 250 µL may be used without compromising the resolution of early eluting peaks, such as chlorite and bromate.

24.1.3 *Guard Column*—Anion exchange column typically packed with the same material used in the analytical column, for example, AG19, or equivalent. The purpose of this column is to protect the analytical column from particulate matter and irreversibly retained material.

24.1.4 *Analytical Column*—Anion exchange column capable of separating the ions of interest from each other, as well as from other ions which commonly occur in the sample matrix, for example, IonPac AS19 (4 mm ID), or equivalent. The separation shall be at least as good as that shown in [Fig. 5](#). The use of 2 mm ID AS19 column, in conjunction with a 50 µL sample loop, may improve the peak shape for early eluting anions, such as chlorite and bromate.

NOTE 2—The Analytical Column Set ([24.1.4](#)) should be able to give

Column:	IonPac AG19, AS19, 4 mm	Peaks:	1. Fluoride
Eluent:	10 mM KOH 0-10 min, 10-45 mM 10-25 min, 45 mM 25-30 min		2. Chlorite 1.0 µg/L
Eluent Source:	EGC-KOH with CR-ATC		3. Bromate 0.5
Temperature:	30 °C		4. Chloride
Flow Rate:	1.0 mL/min		5. Nitrite
Inj. Volume:	250 µL		6. Chlorate 1.0
Detection:	Suppressed conductivity ASRS ULTRA II, recycle mode		7. Bromide
			8. Nitrate
			9. Carbonate
			10. Sulfate
			11. Phosphate




FIG. 5 Chromatogram of Mineral Water A Spiked with 1 µg/L Each Chlorite and Chlorate and 0.5 µg/L Bromate

baseline resolution of all anions, even for a 200 µL injection containing up to 200 mg/L, each, of common anions, such as chloride, bicarbonate, and sulfate.

24.1.5 *Suppressor Device*—A suppressor device based upon cation exchange principles. An ASRA ULTRA II (4 mm) electrolytic suppressor device was used that does not require the addition of an acid but is a plug in electrolytic device. The suppressed eluent (water) is simply recirculated from the conductivity cell back to the electrolytic suppressor to back flush the suppressor device. Alternative pumps are also typically not required.

24.1.6 *Conductivity Detector*—A low-volume, flow through, temperature stabilized conductivity cell equipped with a meter capable of reading from 0 to 1000 µS/cm on a linear scale.

24.1.7 *Data System*—A computer-based data system capable of graphically presenting the detector output signal versus time, as well as presenting the integrated peak areas.

## 25. Preparation of Apparatus

25.1 Set up the ion chromatograph according to the manufacturer's instructions. If an Anion Self Regenerating Suppressor is used, operate the at the appropriate current setting. The conductivity detector cell should be thermally stabilized at 35°C.

25.2 The recommended operating conditions for the ion chromatograph are summarized in Table 12.

**TABLE 12 Instrumentation<sup>A</sup> and Operating Conditions for the Determination of Oxyhalides and Bromide by Ion Chromatography, as shown in Fig. 5**

Ion Chromatograph	ICS-2000 (or equivalent)
Guard Column	IonPac AG19 (or equivalent)
Analytical Column	IonPac AS19 (or equivalent)
Eluent	Hydroxide
Flow-Rate	1.0 mL/min
Injection Volume	250 $\mu$ L
Suppressor	ASRS-ULTRA II (or equivalent)
Detector	Conductivity Detector (or equivalent), stabilized at 35°C

<sup>A</sup> Dionex Corporation, Sunnyvale, CA.

25.3 The detector ranges are variable. Normal operating ranges for quantifying the low level of oxyhalides encountered in treated drinking water are in the 0.2 to 2  $\mu$ S/cm full scale range.

25.4 Equilibrate the chromatographic system by pumping the analysis eluent (see 26.1) through the system until a stable baseline is obtained (approximately 20 minutes). Typical baseline characteristics necessary to obtain the method detection limits required for this analysis are: (1) a background conductance of 0 to 2  $\mu$ S/cm and (2) a peak-to-peak (noise) variation of no greater than 5 nS/cm per minute of monitored baseline response.

## 26. Reagents and Materials

26.1 *Hydroxide Eluent*—If NaOH is manually prepared use 50 % (w/w) NaOH using degassed, deionized water (18.2 megaohm-cm) to a final volume of 1000  $\mu$ L using a volumetric flask. Avoid the introduction of carbon dioxide from the air into the 50 % NaOH or the distilled water being used to make the eluent. Do not shake the 50 % NaOH or pipette the required aliquot from the top of the solution where sodium carbonate may have formed. Eight grams or 5.25 mL of 50 % NaOH makes a 100 mM solution. A positive pressure of an inert gas should be maintained over the headspace to avoid carbon dioxide contamination. The used of electrolytically generated hydroxide by Reagent Free Ion Chromatography® to generate carbonate free hydroxide is also acceptable.<sup>9</sup>

26.2 *Test Method B Eluent Conditions*—10 mM Hydroxide from 0 to 10 minutes and 45 mM from 10 to 25 minutes at 1 mL/min, 30°C (see Fig. 5).

26.3 *Ethylenediamine (EDA) Preservation Solution (50.0 g/L)*—Dilute 11.2 mL of ethylenediamine (99 %) to 200 mL with reagent water. Prepare this solution fresh monthly. Add 1.00 mL of this solution per 1.000 L of blank, standard or sample to produce a final EDA concentration of 50 mg/L.

26.4 *SPE Sample Pretreatment Cartridges*—Chloride present at > 200 mg/L and carbonate present at > 300 mg/L can interfere with bromate determination. H<sup>+</sup> form and Ag<sup>+</sup> form cation exchange SPE cartridges can be used to minimize the

<sup>9</sup> The sole source of supply of the apparatus known to the committee at this time is Reagent Free Ion Chromatography® Systems available from Dionex Corporation. If you are aware of alternative suppliers, please provide this information to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee,<sup>1</sup> which you may attend.

carbonate and chloride interferences, respectively, if required. OnGuard-H and OnGuard-Ag cartridges have been shown to be suitable for this application. The use of these pretreatment cartridges will effect recoveries for bromide, requiring that it be analyzed in a separate run.<sup>6</sup>

26.5 *Suppressor Regenerant Solution*—Not required if using electrolytic suppression. Refer to 24.1.5, if a suppressor requiring chemical regeneration is used.

26.6 *Standard Solutions, Stock (1.00 mL = 1.00 mg)*—Purchase certified solutions or prepare stock standard solutions from the following salts, as described below:

26.6.1 *Bromate (BrO<sub>3</sub><sup>-</sup>) Solution, Stock (1.00 mL = 1.00 mg BrO<sub>3</sub><sup>-</sup>)*—Dissolve 1.180 g of sodium bromate (NaBrO<sub>3</sub>) in water and dilute to 1.000 L.

26.6.2 *Bromide (Br<sup>-</sup>) Solution, Stock (1.00 mL = 1.00 mg Br<sup>-</sup>)*—Dissolve 1.288 g of sodium bromide (NaBr) in water and dilute to 1.000 L.

26.6.3 *Chlorate (ClO<sub>3</sub><sup>-</sup>) Solution, Stock (1.00 mL = 1.00 mg ClO<sub>3</sub><sup>-</sup>)*—Dissolve 1.275 g of sodium chlorate (NaClO<sub>3</sub>) in water and dilute to 1.000 L.

26.6.4 *Chlorite (ClO<sub>2</sub><sup>-</sup>) Solution, Stock (1.00 mL = 1.00 mg ClO<sub>2</sub><sup>-</sup>)*—Dissolve 1.680 g of sodium chlorite (NaClO<sub>2</sub>) in water and dilute to 1.000 L. Note that as sodium chlorite is usually available only as an 80 % technical grade salt, the 80 % purity is accounted for in the 1.680 g weight cited above. If an alternate purity is used, make an appropriate adjustment in the weight of salt used after determining the exact percentage of NaClO<sub>2</sub>, which can be done using an iodometric titration procedure.<sup>2,10</sup>

26.7 *Reagent Blank*—Add 1.00 mL of EDA Preservation Solution (see 26.3) to 1.000 L of reagent water.

## 27. Calibration and Standardization

27.1 *Typical Range of Applicability*—This test method is applicable to the determination of bromate, bromide, chlorate, and chlorite in raw water, finished drinking water, and bottled (non-carbonated) water. The application ranges tested for each analyte are as follows: bromate; 1–30  $\mu$ g/L, bromide; 20–200  $\mu$ g/L, chlorite; 2–500  $\mu$ g/L, and chlorate; 2–500  $\mu$ g/L.

27.2 *Calibration Standards*—For each individual calibration curve, prepare calibration standards, at a minimum of three concentration levels, by accurately adding measured volumes of the stock standards (see 26.6) to a volumetric flask(s). Add 50 mg/L of EDA (the equivalent of 1.00 mL of EDA Preservation Solution (see 26.3) per 1.000 L of solution) to the volumetric flask(s) and dilute to volume with reagent water. A minimum of five concentration levels is recommended if the curve covers two orders of magnitude.

27.3 *Calibration Curve*—To establish the calibration curve, analyze a reagent blank and the calibration standards in accordance to the procedure in Section 28, using a 250  $\mu$ L injection (with a 4 mm ID column) or a 50  $\mu$ L injection (with a 2 mm ID column). Tabulate peak area responses against

<sup>10</sup> Method 4500-ClO<sub>2</sub>-C in A. E. Greenberg, L. S. Clesceri, A. D. Eaton (Eds.), *Standard Methods for the Examination of Water and Wastewater*, 18th Ed., APHA, Washington, DC, 1992.

concentration. These results are used to prepare a calibration curve using a linear least squares fit for each analyte. The squared correlation coefficient of the regression ( $r^2$ ) should be  $\geq 0.995$  for accurate results. Once the calibration curves have been established, verification must be performed on each analysis day, whenever fresh eluent is prepared, and twice each batch of samples.

## 28. Procedure

28.1 Inject the reagent blank, calibration standard or sample into the eluent stream and record the chromatogram. In the case of a manual injector, flush an excess of the sample (minimum of 5x loop volume) through the sample injection port using a syringe prior to injection. A 250- $\mu$ L injection is required when using a 4 mm ID column, a 50- $\mu$ L injection is required when using a 2 mm ID column, in order to achieve the required detection limits for this analysis. An example chromatogram of low level oxyhalides and bromide in a modest ionic strength, drinking water is shown in [Fig. 5](#).

## 29. Calculations

29.1 Compare the peak areas for the anions in the sample to the calibration curves prepared in [27.3](#) to calculate and report the anion concentration in  $\mu$ g/L:

$$\text{Anion concentration, } \mu\text{g/L} = A \times F \quad (3)$$

where:

$A$  = reading from the appropriate calibration plot, in  $\mu$ g/L, and

$F$  = dilution factor if the sample was diluted prior to analysis.

29.1.1 Computing integrators and computer based chromatographic data systems can be programmed to perform these calculations automatically.

29.2 Recoveries were determined in drinking and bottled waters using the following equation:

$$R = \frac{C_s - C}{S} \times 100 \quad (4)$$

where:

$R$  = percent recovery,

$C_s$  = sample concentration,

$C$  = background concentration, and

$S$  = concentration added.

## 30. Precision and Bias

30.1 Standard test methods under the jurisdiction of the ASTM committee D19 may be published for a maximum of five years to the completion of a full collaborative study validation. Such standards are deemed to have met all other D19 qualifying requirements but have not completed the required validation studies to fully characterize the performance of the test methods across multiple laboratories and matrices. Publication of standards that have not been fully validated is done to make current technology accessible to users of standards and to solicit additional input from the user community.

30.2 A quality control (QC) sample served as initial, and on-going, calibration verification. A separate method detection limit (MDL) sample was used for the determination of MDL values. The MDL sample was prepared by pipetting a 1.0 mL aliquot of the MDL concentrate into a clean volumetric flask; adding 50 mg/L EDA, and diluting to a total of 100 mL with reagent water.

30.3 All the single operator precision and bias data presented in this test method was obtained using the IonPac AS19 column listed in [Table 12](#).

30.4 The single operator precision and bias of this test method for each analyte for reagent, drinking and bottled water are shown in [Tables 13-19](#).

30.5 In addition to performing the analyses required to generate the precision and bias data shown in [Tables 13-19](#), the participating laboratory analyzed seven replicates of an MDL sample from reagent water and drinking water. The MDLs were derived using the students  $t$ -test at six degrees of freedom, as follows:

$$MDL = (t) \times (S) \quad (5)$$

where:

$t$  = students  $t$  value for a 99 % confidence level and a standard deviation estimate with  $n-1$  degrees of freedom [ $t = 3.14$  for seven replicates], and

$S$  = standard deviation of the replicate analysis.<sup>8</sup>

30.5.1 True amounts injected, mean value determined, and pooled MDL values (10 laboratories  $\times$  7 replicates) are shown in [Tables 16 and 17](#).

## 31. Quality Control

31.1 Before this test is applied to analyzing unknown samples, the analyst should establish quality control procedures as recommended in [Guide D3856](#).

31.2 The laboratory should repeat the test chromatogram in accordance with the column manufacturer's instructions as an Initial Demonstration of Performance (IDP) using the recommended analytes and concentrations. The user should perform at least 7 replicate injections showing retention time and peak area % RSD's below 5 %.

31.3 When beginning use of this test method, an initial Calibration Verification Standard (CVS) should be used to verify the calibration standards and acceptable instrument performance. This verification should be performed on each analysis day or whenever fresh eluent has been prepared. As this test method is intended for use at trace levels, a low level CVS (that is, equivalent to the lowest calibration standard)

**TABLE 13**

Analyte	Range	Linearity $r^2$	Retention Time Precision % RSD <sup>A</sup>	Peak Area Precision % RSD
Chlorite	20–1000	0.9997	< 0.03	0.44
Bromate	1–30	0.9995	< 0.03	1.09
Chlorate	20–1000	0.9996	< 0.03	0.12
Bromide	20–200	0.9997	< 0.03	0.11

<sup>A</sup>  $n = 10$  injections.

**TABLE 14 Recoveries of Chlorite, Chlorate, Bromate and Bromide from Drinking Waters**

Drinking Water 1				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	8.765	9.993	18.1983	94.39908
			18.4371	96.78875
			18.2464	94.88042
BrO <sub>3</sub>	0	4.88	4.1483	95.35608
			4.7173	85.00615
			4.6325	96.66598
ClO <sub>3</sub>	81.93	105.7	94.92828	92.20014
			184.0871	96.64816
			184.3314	96.87928
Br	26.26	29.86	184.6171	97.14957
			55.7069	96.89234
			56.219	98.61654
			56.0772	100.3315
				99.85666
				99.60159
Drinking Water 2				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	20.983	21.7133	103.480437
			21.6218	103.044369
			21.548	102.692656
BrO <sub>3</sub>	7.98	9.867	103.072487	95.9004763
			17.4425	100.73984
			17.92	103.762035
ClO <sub>3</sub>	26.81	28.668	18.2182	100.134117
			53.3068	92.4263988
			53.5405	93.2415934
Br	190.3	198.34	53.5761	93.3657737
			396.6508	93.0112553
			396.1512	104.038923
			396.6626	103.787032
				104.044872
				103.956943
Drinking Water 3				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	20.946	22.1486	105.7414
			22.0941	105.4812
			22.0861	105.443
BrO <sub>3</sub>	0	5.124	105.5552	98.25917
			5.0348	92.24434
			4.7266	96.21194
ClO <sub>3</sub>	119.6	143.92	4.9299	95.57182
			269.3937	104.0812
			269.7678	104.3412
Br	201.5	199.14	270.4989	104.8492
			401.3183	104.4238
			399.9261	100.3406
			399.4704	99.64151
				99.41267
				99.79827
Drinking Water 4				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	19.982	20.0213	100.196677
			20.2444	101.313182
			20.1344	100.762686
BrO <sub>3</sub>	0	4.979	100.757515	67.4292027
			3.3573	67.3990761
			3.3558	63.6031332
ClO <sub>3</sub>	10.6	10.282	3.1668	66.143804
			19.2819	84.4378526
			19.0115	81.808014
Br	451.9	398.04	19.2987	84.6012449
			853.1127	83.6157038
			850.6708	100.797081
			851.8068	100.1836
				100.468998
				100.483226

**TABLE 14** *Continued*

Drinking Water 5				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	11.6	9.99	20.9888	93.98198
			21.1907	96.003
			21.314	97.23724
BrO <sub>3</sub>	0	4.982	4.679	95.74074
			4.7584	93.91811
			5.0302	95.51184
ClO <sub>3</sub>	85.3	90.668	173.5636	100.9675
			173.5467	96.79914
			174.2912	97.34813
Br	1.23	24.892	25.1916	97.32949
			24.3046	98.15062
			24.5128	97.60941
Drinking Water 6				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	20.754	20.9891	101.132794
			21.1041	101.686904
			21.0476	101.414667
BrO <sub>3</sub>	0	4.827	4.1992	101.411455
			4.1658	86.9939921
			4.1592	86.302051
ClO <sub>3</sub>	4.83	20.646	24.653	86.1653201
			24.2799	86.4871211
			24.2464	96.0137557
Br	206.1	229.68	437.5857	94.206626
			436.8931	94.0443669
			437.7055	94.7549162
Drinking Water 7				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	19.695	21.6019	109.6822
			21.4894	109.1109
			20.7189	105.1988
BrO <sub>3</sub>	1.29	4.907	5.8527	107.9973
			6.0023	92.98349
			5.839	96.0322
ClO <sub>3</sub>	73.62	79.387	151.6691	92.7043
			151.596	93.90666
			151.4981	98.31471
Br	9.74	9.79	19.9167	98.22263
			20.2736	98.09931
			20.5738	98.21222
Drinking Water 8				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	4.56	13.99	17.7194	103.9499
			17.8656	107.5955
			17.3008	110.6619
BrO <sub>3</sub>	0	4.98	4.927	107.4025
			4.9493	99.38353
			5.1358	103.1285
ClO <sub>3</sub>	136.36	151.04	285.7773	100.4826
			285.6934	98.92565
			286.1799	98.8701
Br	0	20.304	5.3299	99.1922
			4.3673	98.99598
			5.4314	26.25049
				21.50955
				26.75039
				24.83681



**TABLE 14** *Continued*

Drinking Water 9				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	18.992	18.5731	97.79433
			18.406	96.91449
			18.3725	96.7381
BrO <sub>3</sub>	0	4.982	4.7743	97.14898
			4.7031	95.83099
			4.6853	94.40185
				94.04456
ClO <sub>3</sub>	0	20.146	21.3346	94.75913
			21.3523	105.8999
			21.888	105.9878
Br	0	19.912	20.5743	108.6469
			20.78	106.8449
			20.3922	103.3261
				104.3592
				102.4116
				103.3656

**TABLE 15 Summary of Table 14 Recoveries of Chlorite, Chlorate, Bromate and Bromide from Drinking Waters**

Analyte	Tap H <sub>2</sub> O Source	Amount Found	Amount Added	Mean % Recovery
Chlorite	Sunnyvale	8.8	10	95.3
	Union City	< MDL	20.9	105.6
	Palo Alto	11.6	10	95.7
	Vacaville	< MDL	19.7	108
	Twain Heart Valley surface water	4.6	14	93.4
	shallow well water	< MDL	19	97.1
	well water	< MDL	21	103.1
		< MDL	20	101.4
Bromate	Sunnyvale	< MDL	4.9	92.2
	Union City	< MDL	5.1	95.6
	Palo Alto	< MDL	5	96.8
	Vacaville	1.3	4.9	93.9
	Twain Heart Valley surface water	< MDL	5	100.5
	shallow well water	< MDL	5	94.7
	well water	16	9.8	100.1
		< MDL	5	86.5
Chlorate	Sunnyvale	81.9	105.7	96.9
	Union City	119.6	143.9	104.4
	Palo Alto	85.3	90.7	97.6
	Vacaville	73.6	79.4	98.2
	Twain Heart Valley surface water	136.4	151	99
	shallow well water	< MDL	20	106.8
	well water	53.6	28.7	93
		4.8	20	94.8
Bromide	Sunnyvale	26.3	29.9	99.6
	Union City	201.5	199	99.8
	Palo Alto	1.2	24.9	94.2
	Vacaville	9.7	9.8	107.4
	Twain Heart Valley surface water	< MDL	20.3	24.8
	shallow well water	< MDL	20	103.3
	well water	380.6	198.3	104
		206	230	100.7

**TABLE 16 Calculated MDL Results Using the Standards Indicated from Reagent Water**

# Injection	Chlorite (1 ppb std)	Bromate (1.5 ppb std)	Chlorate (1.3 ppb std)	Bromide (2 ppb std)
1	0.8816	1.1983	1.0774	2.2766
2	0.7794	0.9713	1.0293	2.1381
3	0.7161	0.9511	1.0428	2.1113
4	0.7881	0.8846	1.0766	1.9853
5	0.7557	0.8962	1.0222	2.0751
6	0.7156	0.9069	1.2095	1.7211
7	0.8961	0.9170	0.8636	2.0353
Avg	0.7903	0.9607	1.0459	2.0489
SD	0.0729	0.1090	0.1022	0.1713
MDL (ppb)	0.2290	0.3425	0.3211	0.5378

**TABLE 17 Calculated MDL Results Using the Standards Indicated from Drinking Water**

# Injection	Chlorite (1 ppb std)	Bromate (1.5 ppb std)	Chlorate (1.3 ppb std)	Bromide (2 ppb std)
1	0.5662	0.6833	1.3809	3.6176
2	0.4475	1.044	1.2177	3.6838
3	0.6901	0.7641	1.4476	3.7967
4	0.6665	0.7189	1.3978	4.0928
5	0.6247	0.6235	1.4755	3.7301
6	0.5612	0.7254	1.4513	3.602
7	0.6325	0.7539	1.2782	3.6958
Avg	0.5983	0.7590	1.3784	3.7455
SD	0.0817	0.1342	0.0964	0.1666
MDL (ppb)	0.2566	0.4214	0.3027	0.5233

should initially be analyzed before beginning use of this test method. The CVS is a solution of method analytes of known concentration used to fortify reagent water, which also contains a final EDA concentration of 50 mg/L (see 26.3). If the determined low level CVS values are not within  $\pm 25\%$  of the known amounts, the low level CVS should be reanalyzed. If the values still fall outside acceptable limits, a new calibration curve is required which must be confirmed by a successful low level CVS before continuing with on-going analyses.

31.4 A continuing CVS should be analyzed after every tenth field sample and an end CVS should be analyzed at the end of the sample batch (maximum of 20 samples) to verify the previously established calibration curves. After initially meeting the requirements of 31.3, the levels selected for the continuing and end CVS should be varied between a middle calibration level and the highest calibration level standard. If the continuing and end CVS values are not within  $\pm 15\%$  of the known amounts, the analyst should reanalyze the CVS. If the analyte concentrations still fall outside acceptable limits ( $\pm 15\%$ ) that analyte is judged out of control, and the source of the problem should be identified before continuing with on-going analyses. All samples following the last acceptable CVS should be reanalyzed.

31.5 A reagent blank (see 26.7) should be run when generating the initial calibration curves. A blank should also be run with each sample batch (maximum of 20 samples) to check for sample or system contamination.

31.6 One Laboratory Control Sample (LCS) should be used with each sample batch (maximum of 20 samples). The LCS is a solution of method analytes of known concentration added to a matrix which sufficiently challenges the test method. A synthetic drinking water matrix, containing fluoride at 1.0 mg/L, chloride at 50 mg/L, nitrite at 0.1 mg/L, nitrate at 10 mg/L, phosphate at 0.1 mg/L and sulfate at 50 mg/L, spiked with the four method analytes at the level of the IDP solution would be an example of an appropriate LCS. The LCS shall also contain 50 mg/L of EDA (the equivalent of 1.00 mL of EDA Preservation Solution (see 26.3) per 1.000 L of solution).

31.6.1 The analyte recoveries for the LCS should fall within the control limits of  $x \pm 3S$ , where  $x$  is the mean recovery and ( $S$ ) is the standard deviation of the mean recovery established from the single operator precision and bias study data.

31.7 One Matrix Spike (MS) should be run with each sample batch (maximum of 20 samples) to test method recovery. The MS should be prepared in accordance with Guide D5810. Spike a portion of a drinking water (or other) sample from each batch with the four method analytes at the level of the IDP solution. The % recovery of the spike should fall within limits established from the interlaboratory precision and bias study data (assuming a background level of zero), according to Practice D5847.

31.8 One Matrix Duplicate (MD) should be run with each sample batch (maximum of 20 samples) to test method precision. If non-detects are expected in all the samples to be analyzed, a Matrix Spike Duplicate should be run instead.

31.9 In order to verify the quantitative values produced by the test method, an Independent Reference Material (IRM), submitted to the laboratory as a regular sample (if practical), should be analyzed once per quarter. The concentration of the IRM should be within the scope of the test method, as defined in 21.1. The values obtained must fall within the limits specified by the outside source.

31.10 The laboratory may perform additional quality control as desired or appropriate, for instance the use of a surrogate as outlined in Section 9.4.2 of U.S. EPA Method 300.1 or as recommended by the manufacturer. In addition, it is recommended that a laboratory determine the method detection limits, as discussed in 30.5, before using this test method.

## 32. Keywords

32.1 bottled and finished drinking water; bromate; bromide; chemical suppression; chlorate; chlorite; ion chromatography; raw water; sample preservation

**TABLE 18 Recoveries of Chlorite, Chlorate, Bromate, and Bromide from Bottled Waters**

Bottled Water 1				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	19.92	21.5416	108.1406
			21.5525	108.1953
			21.5296	108.0803
BrO <sub>3</sub>	0	4.983	4.6816	108.1387
			4.6958	93.95143
			4.982	94.2364
ClO <sub>3</sub>	2.4	20.072	23.932	99.97993
			23.8838	96.05592
			24.2592	107.2738
Br	7.5	19.992	28.6255	107.0337
			28.3698	107.0337
			28.5085	108.9039
Bottled Water 2				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	19.93	20.9161	104.9478
			20.3823	102.2694
			20.2116	101.4129
BrO <sub>3</sub>	0	4.985	4.902	102.8767
			5.0747	98.33501
			5.0815	101.7994
ClO <sub>3</sub>	0	20.08	21.3216	101.9358
			21.7464	100.6901
			21.072	106.1833
Br	0	20.002	21.2098	108.2988
			21.2935	104.9402
			21.3975	106.4741
Bottled Water 3				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	20.929	20.7492	99.1409
			21.0457	100.5576
			20.8743	99.73864
BrO <sub>3</sub>	10.2	9.872	20.4379	99.81238
			20.4159	103.7064
			20.718	103.4836
ClO <sub>3</sub>	0	21.088	21.7205	106.5438
			21.618	104.5779
			21.6927	102.9993
Br	19.37	21.004	38.9859	102.5133
			38.7882	102.8675
			38.8527	102.7934
Bottled Water 4				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	20.983	19.0318	90.70104
			19.1275	91.15713
			18.6043	88.66368
BrO <sub>3</sub>	0	4.97	4.272	90.17395
			4.2707	85.95573
			3.9144	85.92958
ClO <sub>3</sub>	10.16	20.138	30.442	78.76056
			30.9605	83.54863
			31.2967	100.7151
Br	95.49	104.5	198.2911	103.2898
			197.7618	104.9593
			196.7893	102.988
				98.37426
				97.86775
				96.93713
				97.72638

**TABLE 18** *Continued*

Bottled Water 5				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	19.952	20.3706	102.098
			20.195	101.2179
			20.0104	100.2927
BrO <sub>3</sub>	0	4.872	4.7786	101.2029
			4.5397	98.08292
			4.692	93.17939
ClO <sub>3</sub>	1.55	20.106	4.692	96.30542
			23.3528	95.85591
			23.3969	108.4393
Br	1.21	19.862	23.438	108.6586
			20.0039	108.863
			20.2719	108.6536
			20.3306	94.62239
				95.9717
				96.26724
				95.62045
Bottled Water 6				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	19.841	20.0341	100.9732
			20.1373	101.4934
			20.2725	102.1748
BrO <sub>3</sub>	9.23	9.789	101.5471	103.3425
			19.3462	107.7955
			19.7821	108.5473
ClO <sub>3</sub>	374.5	149.96	19.8557	106.5618
			520.2669	97.20385
			520.7441	97.52207
Br	2.49	19.762	520.3303	97.24613
			22.3591	97.32402
			22.4557	100.5419
			22.4863	101.0308
				101.1856
				100.9194
Bottled Water 7				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	19.978	21.1067	105.6497
			21.4455	107.3456
			21.4251	107.2435
BrO <sub>3</sub>	0	4.978	106.7463	106.7463
			4.5736	91.87626
			4.6553	93.51748
ClO <sub>3</sub>	0	25.165	4.56	91.60305
			22.6989	92.33226
			22.6964	90.20028
Br	31.8	29.847	23.0201	91.47665
			61.2379	90.62243
			61.0686	98.62934
			61.6879	100.137
				98.94283
Bottled Water 8				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	19.99	20.3662	101.8819
			20.3674	101.8879
			20.5673	102.8879
BrO <sub>3</sub>	0	4.981	102.2193	102.2193
			4.7302	94.96487
			4.5656	91.66031
ClO <sub>3</sub>	0	20.146	4.7116	94.59145
			21.2721	93.73887
			21.2247	105.5897
Br	18.67	19.912	21.1961	105.3544
			37.3404	105.2124
			37.3579	105.3855
			37.3659	93.76456
				93.85245
				93.89263
				93.83655

**TABLE 18** *Continued*

Bottled Water 9				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	19.94	21.1555	106.0958
			20.9677	105.154
			21.3584	107.1133
BrO <sub>3</sub>	0	4.87	4.5977	106.121
			4.5697	94.40862
			5.2069	93.83368
ClO <sub>3</sub>	0	20.094	21.2774	106.9179
			21.3351	98.38672
			21.1095	105.8893
Br	2.66	19.86	23.2262	106.1765
			23.0562	105.0537
			23.6949	105.7065
Bottled Water 10				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	20.985	20.5197	97.7827
			20.6457	98.38313
			20.6274	98.29593
BrO <sub>3</sub>	4.4	4.98	9.388	98.15392
			9.5339	100.1606
			9.3899	103.0904
ClO <sub>3</sub>	0	20.14	21.501	100.1988
			21.7981	101.1499
			21.787	106.7577
Br	0	19.906	20.8241	108.2329
			21.0055	108.1778
			21.0659	107.7228
Bottled Water 11				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	20.494	21.3516	104.1846
			21.5828	105.3128
			21.5015	104.9161
BrO <sub>3</sub>	0	5.107	5.0536	104.8045
			4.7337	98.95438
			4.9806	92.69042
ClO <sub>3</sub>	0	22.947	22.7441	97.52497
			22.1578	96.38992
			22.7628	99.11579
Br	6.34	22.68	27.9533	96.56077
			27.5945	99.19728
			27.7531	98.29128
Bottled Water 12				
Analyte	Amount Found (µg/L)	Amount Added (µg/L)	Final Concentration (µg/L)	%Recovery
ClO <sub>2</sub>	0	21.905	20.9851	95.8005
			20.7819	94.87286
			20.8136	95.01758
BrO <sub>3</sub>	0.984	4.882	5.7522	95.23031
			6.1172	97.66899
			6.0343	105.1454
ClO <sub>3</sub>	4.19	22.072	25.9114	103.4474
			25.9705	102.0873
			25.9366	98.41156
Br	0	21.985	21.8261	98.67932
			21.8205	98.52573
			21.7536	98.53887
				99.27723
				99.25176
				98.94746
				99.15882

**TABLE 19 Summary of Table 18 Recoveries of Chlorite, Chlorate, Bromate and Bromide from Bottled Waters**

Analyte	Bottled H <sub>2</sub> O Source	Amount Found	Amount Added	Mean % Recovery
Chlorite	1	< MDL	19.9	108.1
	2	< MDL	19.9	102.9
	3	< MDL	20.9	99.8
	4	< MDL	21	90.2
	5	< MDL	20	101.2
	6	< MDL	19.8	101.5
	7	< MDL	20	106.7
	8	< MDL	20	102.2
	9	< MDL	19.9	106.1
	10	< MDL	21	98.2
	11	< MDL	20.5	104.8
	12	< MDL	21.9	95.2
Bromate	1	< MDL	5	96.1
	2	< MDL	5	100.7
	3	10.2	9.8	104.6
	4	< MDL	5	83.5
	5	< MDL	4.9	95.9
	6	9.2	9.8	106.6
	7	< MDL	5	92.3
	8	< MDL	5	93.7
	9	< MDL	4.9	98.4
	10	4.4	5	101.1
	11	< MDL	5.1	96.4
	12	0.98	4.89	102.1
Chlorate	1	2.4	20.1	107.7
	2	< MDL	20.1	106.5
	3	< MDL	21.1	102.8
	4	10.2	20	103
	5	1.6	20.1	108.6
	6	374.5	150	97.3
	7	< MDL	25.2	90.6
	8	< MDL	20.1	105.4
	9	< MDL	20.1	105.7
	10	< MDL	20.1	107.7
	11	< MDL	22.9	98.3
	12	4.2	22.1	98.5
Bromide	1	7.5	20	105
	2	< MDL	20	106.5
	3	19.4	21	92.9
	4	95.5	104.5	97.7
	5	1.2	19.9	95.6
	6	2.5	19.8	100.9
	7	31.8	29.8	98.9
	8	18.7	19.9	93.8
	9	2.7	19.9	104.1
	10	< MDL	19.9	105.3
	11	6.3	22.7	94.5
	12	< MDL	22	99.2

**TABLE 20 Disinfection Procedure for the Bottled Water Sources Used in This Study**

Bottled Water Source	Disinfection Procedure Used for Bottled Waters
1	natural spring water, no treatment
2	UV light, reverse osmosis, ozonation
3	ozonation
4	natural mineral water
5	filtered by reverse osmosis, minerals added
6	micro-filtration, ozonation, UV light treatment
7	filtration
8	microfiltration, ozonation
9	natural spring water
10	microfiltered, ozonated
11	microfiltration, reverse osmosis, deionization, ozonation
12	ozonation

*ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.*

*This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.*

*This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or [service@astm.org](mailto:service@astm.org) (e-mail); or through the ASTM website ([www.astm.org](http://www.astm.org)). Permission rights to photocopy the standard may also be secured from the ASTM website ([www.astm.org/COPYRIGHT/](http://www.astm.org/COPYRIGHT/)).*