



Designation: D6674 – 01 (Reapproved 2017)

## Standard Guide for Proficiency Test Program for Fabrics<sup>1</sup>

This standard is issued under the fixed designation D6674; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

### INTRODUCTION

Proficiency testing is the use of interlaboratory test comparisons to determine the performance of individual laboratories for specific tests and to monitor the consistency and comparability of a laboratory's test data.

Participation in proficiency testing programs provides laboratories with an objective means of assessing and demonstrating the reliability of the data they are producing. Although there are several types of proficiency testing programs, they all share the common feature of the comparison of test results obtained by two or more laboratories.

One of the main uses of proficiency testing programs is to assess laboratories' ability to perform tests competently. This will involve the preparation of the test specimens, calibrating or validating the testing equipment, performing the tests and reporting the data.

Bodies assessing the technical competence of testing laboratories normally require or expect satisfactory participation in proficiency testing as evidence of a laboratory's ability to produce reliable test results.

### 1. Scope

1.1 This guide outlines the Proficiency Test Program for Fabrics. Elements for planning the proficiency test program, selecting the sample fabrics to be used, the testing protocol, and the calculations for the data to be reported are included in this practice.

1.2 The planning of the proficiency test program requires a general knowledge of testing of textile fabrics and statistical principles included in the analysis of the data.

1.3 This guide is designed to meet the quality systems proficiency and competence requirements of participating laboratories. This program is not accredited to any international standard.

1.4 The instructions in this guide follow the logic of full scale laboratory tests as described in Practice D2904 and Guide E1301, except with this new guide placing its emphasis on proficiency testing.

1.5 Procedures given in this guide are applicable to methods based on the measurement of discrete measurement data and grades or scores.

1.6 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

### 2. Referenced Documents

2.1 *ASTM Standards:*<sup>2</sup>

D123 Terminology Relating to Textiles

D737 Test Method for Air Permeability of Textile Fabrics

D1230 Test Method for Flammability of Apparel Textiles

D2261 Test Method for Tearing Strength of Fabrics by the Tongue (Single Rip) Procedure (Constant-Rate-of-Extension Tensile Testing Machine)

D2904 Practice for Interlaboratory Testing of a Textile Test Method that Produces Normally Distributed Data (Withdrawn 2008)<sup>3</sup>

D2906 Practice for Statements on Precision and Bias for Textiles (Withdrawn 2008)<sup>3</sup>

D3775 Test Method for Warp (End) and Filling (Pick) Count of Woven Fabrics

<sup>1</sup> This guide is under the jurisdiction of ASTM Committee D13 on Textiles and is the direct responsibility of Subcommittee D13.60 on Fabric Test Methods, Specific.

Current edition approved July 15, 2017. Published August 2017. Originally approved in 2001. Last previous edition approved in 2013 as D6674-01(2013). DOI: 10.1520/D6674-01R17.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

<sup>3</sup> The last approved version of this historical standard is referenced on [www.astm.org](http://www.astm.org).

- [D5034 Test Method for Breaking Strength and Elongation of Textile Fabrics \(Grab Test\)](#)
- [D5035 Test Method for Breaking Force and Elongation of Textile Fabrics \(Strip Method\)](#)
- [D6545 Test Method for Flammability of Textiles Used in Children's Sleepwear](#)
- [E1301 Guide for Proficiency Testing by Interlaboratory Comparisons \(Withdrawn 2012\)<sup>3</sup>](#)

### 3. Terminology

3.1 The following terms are relevant to this standard: accuracy, *of a test method*; batch sample; bias, *in statistics*; calibrate; precision; precision, *under conditions of between-laboratory precision*; proficiency testing; program coordinator; repeatability, *in statistics*; reproducibility, *in statistics*; sample; specimen; test result; testing laboratory.

3.2 For definitions of textile and statistical terms used in this practice, and discussions of their use, refer to Terminology [D123](#) and appropriate textbooks on statistics.

### 4. Significance and Use

4.1 Proficiency testing is a means of securing estimates of the variability of results obtained by different laboratories testing homogeneous materials taken from batch samples when following procedures prescribed in a specific test method.

4.1.1 For the purpose of this guide, homogeneous materials are considered to be laboratory samples cut from the same batch sample (roll or garment lot) and selected at random for the participant laboratories.

4.2 This proficiency test program is to be considered a full-scale interlaboratory test, in which a reasonably large number of laboratories participate by testing a series of materials using one or more operators per laboratory and report the data for analysis.

4.2.1 For the purposes of this guide, Full Scale Laboratory Test is defined in Practice [D2904](#). This is not to be confused with the full scale testing terminology and definitions which appear in other test methods, such as flammability, that describe the size of the specimens being tested.

4.3 The statistical data generated by this practice provide information needed to exhibit participation in a formal proficiency test study.

4.4 All data are submitted to the program coordinator at ASTM Headquarters for the preparation and distribution of the proficiency testing program reports. All laboratory data are confidential with no disclosure of lab identity except for each participant's own laboratory. Published reports contain all laboratory test data (coded), statistical analysis of test data, charts plotting test results versus lab code, and other information.

### 5. Materials to be Used in Study

5.1 *Mechanical Testing Program*—Three types of materials are to be included in the study. The materials shall be described as light, medium and heavy as defined by the breaking strengths. The materials to be used for this study shall be selected by the Subcommittee D13.93 with assistance from

participant laboratories and pertinent subcommittees. An accredited laboratory selected by the Subcommittee D13.93 shall act as host and maintain these materials. The host laboratory shall not be held responsible for the performance of the test materials at the participant laboratories.

5.2 *Flammability Test Program*—At least seven types of fabrics are to be available for the study, not all of which will be used at the same time. The materials shall be described by fiber type, construction type (woven, knit, nonwoven), surface characteristics and fabric weight as defined by the flammability characteristics. The fabrics to be used for this study shall be selected by the Subcommittee D13.93. An accredited laboratory selected by the Subcommittee D13.93 shall act as the host and maintain these fabrics. The host laboratory shall not be held responsible for the performance of the materials at the participant laboratories.

### 6. Distribution of Test Materials

6.1 The proficiency test program is designed for the test materials to be distributed at a minimum two times each calendar year—February and August. Each participating laboratory shall apply through ASTM Headquarters for inclusion in the program.

6.2 The materials will be distributed by having the host laboratory cut fabric samples from each batch roll or garment lot. The samples will be identified, labeled and mailed to each participating laboratory. ASTM Headquarters will provide detailed instruction sheets and mailing labels to the host laboratory.

6.3 Instructions for testing, recording data and return; forms and fabric samples will be mailed together as instructed.

### 7. Selection of Test Methods

7.1 *For Mechanical Testing*—The test methods shall be selected and monitored by the Subcommittee D13.93. The protocol of this guide may not include the entire scope of methods used for testing fabric, but will include at least one mechanical test from the fields used to define woven fabric properties such as construction, air permeability and strength. Other methods may be added or deleted as directed by D13.93 with input from ASTM staff, participating laboratories, and pertinent Sub-Committee.

7.1.1 The methods to be used for this guide are as follows:

- [D737](#) Air Permeability of Textile Fabrics
- [D2261](#) Tearing Strength By Tongue Method (CRE Type Instrument)
- [D3775](#) Fabric Count of Woven Fabric
- [D5034](#) Breaking Strength and Elongation of Textile Fabrics (Grab Test)
- [D5035](#) Breaking Strength and Elongation of Textile Fabrics (Strip Test)

7.2 *For Flammability Testing*—The test methods shall be selected and monitored by the subcommittee D13.93. The protocol of this guide may not include the entire scope of methods used for testing fabrics, but will include at least one flammability test from the fields used to define characteristics of the end use product.

NOTE 1—D13.52 is preparing two additional testing protocols to be used in conjunction with the flammability tests. The criteria are being

established for lab personnel to identify the types of product being tested and a comprehensive exam for the scope of testing.

7.2.1 The methods to be used for this guide are as follows:

- D1230 Flammability of Apparel Fabrics
- D6545 Flammability of Textiles Used in Children's Sleepwear

## 8. Sampling

8.1 *Sampling of Materials*—The host laboratory shall allocate enough material to provide for all participating laboratories to perform the complete scope of testing for the program.

8.2 *Numbers of Determinations*—Each participating lab shall prepare the test specimens and number of determinations (replicates) as directed by the proficiency testing program.

## 9. Data Reporting

9.1 All data shall be reported as instructed using the forms supplied with the material package.

9.2 *Outlying Observations*—All data shall be reported. Exceptions to this general policy should be made only when assignable causes for deletion of a test value are present. In cases where there is no assignable cause for a test value being out of line, the test value shall be retained and reported.

## 10. Report

10.1 Table of Results pages provides a listing of laboratories with corresponding test results. Each laboratory is identified by a randomly selected code with the same laboratory code applying throughout the entire report. To maintain confidentiality of test data, the identification of lab codes will not be released from ASTM. Therefore, program participants will be able to identify only their own test data. The laboratory identification code may change with subsequent reports. Each table page contains a heading identifying the method and test fabric.

10.2 Lab Average column contains the lab average as reported to ASTM by the participating laboratories.

10.3 The Lab Range column represents the difference between the lab's lowest and highest of the raw data results.

10.4 *Description of Z-Score Column*—The Z-Score column (comparable to the Student's t statistic) reports each lab's deviation in units of standard deviations. The Z Score is the calculated number of standard deviations a sample mean is

above or below the population mean. The criteria for the Z Scores will be as follows:

- [Z] <2 = satisfactory
- 2 < [Z] <3 = questionable
- [Z] >3 = unsatisfactory

Z-Scores outside this range should occur only about one time in twenty, if a laboratory has average capability running the method. Laboratories should strive to obtain Z Score values close to zero. The Z-Score is calculated as follows:

$$Z \text{ Score} = (\text{lab average} - \text{total average}) / \text{total standard deviation}$$

10.5 *Identification of Outlier Data*—During the analytical review of the data it was observed that when including all data in the calculation of the Overall Average and Std. Deviation, in some cases, the averages and standard deviations are adversely affected by outlier data. Therefore, through two iterations of calculations, labs who had a Z-Score Value of  $\pm 3$  or greater have been asterisked in the table. These data, while still included in the tables, have not been included in the statistical analysis. This step of excluding data is only performed through two iterations. After the second iteration, all remaining data have been included in the resulting statistical analysis.

10.6 *Flags*—This column contains reference to data outside either three standard deviations or the upper limit of the Range Chart. A "1" indicates that the lab's average exceeds 3 standard deviations, a "2" indicates that the lab's range exceeds the upper range limit.

10.7 *High-Low Charts*—These are provided for each sample as a visual comparison of laboratory performance. The "Y" scale represents the test results and "X" scale the laboratory codes. Each laboratory's data is plotted. The length of the whiskers represents the high-low range of the five test results for each lab. Horizontal lines are provided at the Overall Average and Overall Average plus/minus three standard deviations.

10.8 *Range Graph*—This is provided to chart individual Lab Ranges against the Total Range Average. The Upper Control Limit is determined by multiplying 2.114 times the Total Range Average.

## 11. Keywords

11.1 calibration; proficiency testing; program coordinator; testing laboratory; Z score

APPENDIX

(Nonmandatory Information)

**X1. A CHECK LIST FOR INVESTIGATING THE ROOT CAUSE OF UNSATISFACTORY ANALYTICAL PERFORMANCE**

**INTRODUCTION**

To identify why a laboratory's data may have been considered a statistical outlier and/or improve the precision, the following action items are suggested. There may be additional ways to improve the performance.

- 1) Check the results for typos, calculation errors and transcription errors.
- 2) Reanalyze the sample: check for repeatability.
- 3) Check the sample for homogeneity or contamination and that a representative specimen has been analyzed.
- 4) Review the test method and ensure that the latest version of the ASTM test method is being used. Check the procedure step by step with the analyst.
- 5) Check the instrument calibration.
- 6) Check the statistical quality control chart to see if the problem has been developing earlier.
- 7) Check the quality of the reagents and standards used and whether they are expired or contaminated.
- 8) Check the equipment for proper operation against vendor's operating manual.
- 9) Call the vendor for maintenance and/or repairs.
- 10) After the problem has been resolved, analyze a certified reference material if one is available or the laboratory quality control sample to ascertain that the analytical operation is under control.
- 11) Provide training to new analysts and if necessary, refresher training to experienced analysts.
- 12) Document the incident and the learnings for use in future similar problem.

*ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.*

*This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.*

*This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; <http://www.copyright.com/>*