



# Standard Test Method for Particle Count in Mineral Insulating Oil Using Automatic Optical Particle Counters<sup>1</sup>

This standard is issued under the fixed designation D6786; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This test method covers the determination of particle concentration and particle size distribution in mineral insulating oil. It is suitable for testing oils having a viscosity of 6 to 20 mm<sup>2</sup>/s at 40°C. The test method is specific to liquid automatic particle analyzers that use the light extinction principle.

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

## 2. Referenced Documents

2.1 *ASTM Standards:*<sup>2</sup>

[D923 Practices for Sampling Electrical Insulating Liquids](#)

2.2 *ISO Standards:*

[4406:1999 Hydraulic Fluid Power—Fluids—Method for Coding the Level of Contamination by Solid Particles](#)<sup>3</sup>

[11171:2010 Hydraulic Fluid Power—Calibration of Automatic Particle Counters for Liquids](#)<sup>3</sup>

## 3. Terminology

3.1 *Definitions:*

3.1.1 *coincidence*—the presence of more than one particle in the sensing zone of a particle analyzer at the same time, causing mis-sizing and mis-counting of the particle present.

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D27 on Electrical Insulating Liquids and Gases and is the direct responsibility of Subcommittee D27.07 on Physical Test.

Current edition approved Oct. 1, 2015. Published October 2015. Originally approved in 2002. Last previous edition approved in 2008 as D6786–08. DOI: 10.1520/D6786-15.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

<sup>3</sup> Available from American National Standards Institute (ANSI), 25 W. 43rd St., 4th Floor, New York, NY 10036, <http://www.ansi.org>.

The coincidence limit of the counter is determined by the maximum acceptable concentration of particles in the sensing zone and is supplied by the instrument manufacturer.

3.1.2 *concentration limit*—a direct function of coincidence and electronic saturation. The concentration limit of the system is determined by the maximum acceptable concentration of particles in the given sample and is supplied by the instrument manufacturer.

3.1.3 *electronic saturation level*—particle concentration at which the electronic circuitry of the analyzer ceases to function properly due to excessive counting rates.

3.1.4 *light extinction*—the reduction in intensity of a light beam passing through the sensing zone of a particle analyzer, caused by the absorption and/or scattering of the light by particles. Synonyms: light obscuration, light interruption, light blockage.

## 4. Summary of Test Method

4.1 Samples are taken in particle-clean bottles that are suitable for particle analysis. The sample bottle is agitated to redistribute particles in the oil, then the oil is placed in an automatic particle counter, where the number of particles and their size distribution are determined by the light extinction principle.

4.2 As particles pass through the sensing zone of the instrument, the quantity of light reaching the detector is obscured. This signal is translated to an equivalent projected area diameter based on calibration with a NIST-traceable fluid (ISO Medium Test Dust suspension).

## 5. Significance and Use

5.1 Particles in insulating oil can have a detrimental effect on the dielectric properties of the fluid, depending on the size, concentration, and nature of the particles. The source of these particles can be external contaminants, oil degradation by-products, or internal materials such as metals, carbon, or cellulose fibers.

5.2 Particle counts provide a general degree of contamination level and may be useful in assessing the condition of specific types of electrical equipment. Particle counts can also be used to determine filtering effectiveness when processing oil.

5.3 If more specific knowledge of the nature of the particles is needed, other tests such as metals analysis or fiber identification and counting must be performed.

## 6. Interferences

6.1 Dirty environmental conditions and poor handling techniques can easily contaminate the sample and/or test specimen. Care must be taken to ensure test results are not biased by introduced particles.

6.2 Air bubbles in the oil may be counted as particles giving false positive readings. Mixing or agitating the sample introduces bubbles into the oil. These are dissipated using vacuum or ultrasonic bath.

6.3 Suspended or free water in the oil will interfere with particle counts.

6.4 Excessive concentrations of particles in the oil will cause coincidence and/or electronic saturation errors. Limits are determined by ISO 11171 and are generally supplied by the instrument manufacturer. These errors may be avoided by diluting the sample with particle-clean dilution oil or particle-clean solvent.

6.5 Odd-shaped particles and fibers may be improperly analyzed, depending on their orientation as they pass through the sensing zone of the instrument.

## 7. Apparatus

7.1 *Automatic Particle Counter*, liquid optical particle counter based on the light extinction principle. The instrument shall be capable of recording the size and number of particles as they pass across the detector. The particle counter shall include a bottle sampling apparatus that automatically delivers a predetermined volume of specimen at a controlled flow rate to the sensing zone of the analyzer.

7.2 *Mechanical Shaker*, paint shaker, table shaker, or other mechanical device to vigorously agitate sample bottles.

7.3 *Ultrasonic Bath*, with a power density of 3000 to 10 000 W per m<sup>2</sup> of bottom area.

## 8. Materials

8.1 *Particle-clean Bottles*, recommended sample containers are cylindrical bottles made of polypropylene, polystyrene, PET, or glass with flat bottoms, fitted with a suitable non-shedding threaded cap. Bottles should be at least 100-mL capacity. The bottles shall meet the cleanliness criteria of contributing less than 1 % of the total particles expected in the cleanest sample.

8.2 *Particle-clean Solvent*, petroleum spirits, hexane, kerosene, or other suitable solvent, filtered through a 0.45 μm membrane filter.

8.3 *Calibration Fluid*, suspension of ISO Medium Test Dust in oil or hydraulic fluid, either a primary sample obtained directly from NIST<sup>4</sup> (SRM 2806) or a secondary sample prepared in accordance to ISO 11171 and traceable to NIST.

8.4 *Dilution Oil*, insulating oil that has been filtered to contain fewer than 1 % of the total particles expected in the cleanest sample.

## 9. Sampling

9.1 Proper sampling is crucial to particulate analysis. Obtain a sample of the liquid to be tested in accordance with Practice D923.

## 10. Calibration

10.1 Calibration of the instrument shall be with a NIST-traceable calibration fluid in accordance with ISO 11171.

10.2 Calibrate the instrument annually, unless experience justifies longer or shorter intervals.

10.3 Interim calibration checks should be made regularly by using a particle count standard fluid prepared using the procedures of ISO 11171. If the average number of counts is at least 100, the check values should be within 20 % of the standard values.

## 11. Procedure

11.1 If necessary, wipe the outside of the sample bottle with a clean lint-free towel.

11.2 If other tests are to be run on the same sample, shake the sample bottle as described in 11.3 and transfer a specimen for particle count to another particle-clean bottle. Do this before running any other tests in order to avoid contamination.

11.3 Ultrasonically disperse for 30 to 40 s then agitate the sample bottle vigorously for 30-60 s to redistribute particles, the length of time depending on the effectiveness of the method of agitation. Preferably use a mechanical shaker. Do not use a magnetic stir bar or any other device that comes in contact with the oil.

11.4 Apply vacuum to the bottle, or use the ultrasonic bath, or both, until bubbles dissipate.

11.5 Immediately analyze the sample with the automatic particle counter according to the manufacturer's operating instructions. Usually the analyzer is flushed with a specimen of the oil prior to testing. After the initial flush, 2-3 runs on each sample are recommended.

11.6 When particle count runs are finished, flush the instrument with particle-clean solvent or particle-clean oil in accordance with the instrument manufacturer's recommendations.

## 12. Report

12.1 Report the average of the particle count runs as the cumulative number of particles per mL  $\geq 4$ ,  $\geq 6$ ,  $\geq 10$ ,  $\geq 14$ ,  $\geq 21$ ,  $\geq 38$ , and  $\geq 70$  μm (c). The "(c)" after the size indicates that the particle counter was calibrated using ISO 11171. These particle sizes correspond approximately to >2, >5, >10, >15, >25, >50, and >100 μm for particle counters that were calibrated with the old ISO 4402 calibration standard. If only one particle count run is analyzed, report the results of the single run.

12.2 Optionally, also report the ISO solid contaminant code corresponding to the number of particles per mL  $\geq 4$ , 6, and 14 μm (c), as prescribed in ISO 4406:1999.

<sup>4</sup> Available from National Institute of Standards and Technology, 100 Bureau Drive, Gaithersburg, MD 20899.

**13. Precision and Bias**

13.1 A round robin is planned to determine the precision and bias of this test method.

**14. Keywords**

14.1 optical particle counter; particle count; particulate contamination

*ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.*

*This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.*

*This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or [service@astm.org](mailto:service@astm.org) (e-mail); or through the ASTM website ([www.astm.org](http://www.astm.org)). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; <http://www.copyright.com/>*