



Standard Test Method for Available Cyanides with Ligand Displacement and Flow Injection Analysis (FIA) Utilizing Gas Diffusion Separation and Amperometric Detection¹

This standard is issued under the fixed designation D6888; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This method is used to determine the concentration of available inorganic cyanide in an aqueous wastewater or effluent. The method detects the cyanides that are free (HCN and CN⁻) and metal-cyanide complexes that are easily dissociated into free cyanide ions. The method does not detect the less toxic strong metal-cyanide complexes, cyanides that are not “amenable to chlorination.”

1.2 Total cyanide can be determined for samples that have been distilled as described in Test Methods [D2036](#), Test Method A, Total Cyanides after Distillation. The cyanide complexes are dissociated and absorbed into the sodium hydroxide capture solution, which can be analyzed with this test method; therefore, ligand exchange reagents from Sections [8.12](#) and [8.13](#) would not be required when determining total cyanide after distillation.

1.3 This procedure is applicable over a range of approximately 2 to 400 $\mu\text{g/L}$ (parts per billion) available cyanides. Higher concentrations can be analyzed by dilution or lower injection volume.

1.4 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* Specific hazard statements are given in [8.6](#) and Section [9](#).

¹ This test method is under the jurisdiction of ASTM Committee [D19](#) on Water and is the direct responsibility of Subcommittee [D19.06](#) on Methods for Analysis for Organic Substances in Water.

Current edition approved Feb. 1, 2016. Published June 2016. Originally approved in 2003. Last previous addition approved in 2009 as D6888 – 09. DOI: 10.1520/D6888-16.

2. Referenced Documents

2.1 ASTM Standards:²

- [D1129 Terminology Relating to Water](#)
- [D1193 Specification for Reagent Water](#)
- [D2036 Test Methods for Cyanides in Water](#)
- [D3856 Guide for Management Systems in Laboratories Engaged in Analysis of Water](#)
- [D5847 Practice for Writing Quality Control Specifications for Standard Test Methods for Water Analysis](#)
- [D6696 Guide for Understanding Cyanide Species](#)
- [D7365 Practice for Sampling, Preservation and Mitigating Interferences in Water Samples for Analysis of Cyanide](#)

3. Terminology

3.1 Definitions:

3.1.1 For definitions of terms used in this standard, refer to Terminology [D1129](#) and Guide [D6696](#).

3.2 Definitions of Terms Specific to This Standard:

3.2.1 *available cyanides, n*—Inorganic cyanides that are free (HCN and CN⁻) and metal-cyanide complexes that are easily dissociated into free cyanide ions.

3.2.1.1 *Discussion*—Available cyanide does not include the less toxic strong metal-cyanide complexes, cyanides that are not “amenable to chlorination” and includes weak acid dissociable or weak and dissociable (WAD) cyanides for use in the implementation of International Cyanide Management Code.

4. Summary of Test Method

4.1 Complex cyanides bound with nickel or mercury are released by ligand displacement by the addition of a ligand displacement agent, when necessary.

4.2 Other weak and dissociable cyanide species do not require ligand displacement.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard’s Document Summary page on the ASTM website.




FIG. 1 Flow Injection Analysis Apparatus 1

4.3 The sample is introduced into a flow injection analysis (FIA) system where it is acidified to form hydrogen cyanide (HCN). The hydrogen cyanide gas diffuses through a hydrophobic gas diffusion membrane, from the acidic donor stream into an alkaline acceptor stream.

4.4 The captured cyanide is sent to an amperometric flowcell detector with a silver-working electrode. In the presence of cyanide, silver in the working electrode is oxidized at the applied potential. The anodic current measured is proportional to the concentration of cyanide in the standard or sample injected.

4.5 Calibrations and data are processed with the instrument's data acquisition software.

5. Significance and Use

5.1 Cyanide and hydrogen cyanide are highly toxic. Regulations have been established to require the monitoring of cyanide in industrial and domestic wastes and surface waters.³

5.2 This test method is applicable for natural water, saline waters, metallurgical process solutions, and wastewater effluent.

5.3 The method may be used for process control in wastewater treatment facilities.

6. Interferences

6.1 High levels of carbonate can release CO₂ into the acceptor stream and cause an interference with the amperometric detector that result in a slight masking effect (15 % negative bias with 20 ppb cyanide in 1500 ppm carbonate). Refer to 11.2 for sample pretreatment.

6.2 Sulfide above 50 mg/L will diffuse through the gas diffusion membrane and can be detected in the amperometric flowcell. Oxidized products of sulfide can also rapidly convert CN⁻ to SCN⁻ at a high pH. Refer to Practice D7365 for sulfide removal procedures.

6.3 Refer to Practice D7365 for further information on mitigating interferences in water samples for the analysis of cyanide.

7. Apparatus

7.1 The instrument should be equipped with a precise sample introduction system, a gas diffusion manifold with hydrophobic membrane, and an amperometric detection system to include a silver working electrode, a Ag/AgCl reference electrode, and a Pt or stainless steel counter electrode. Examples of the apparatus schematics are shown in Figs. 1-3. Example instrument settings are shown in Table 1.⁴

NOTE 1—The instrument settings in Table 1 are only examples. The analyst may modify the settings as long as performance of the method has not been degraded. Contact the instrument manufacturer for recommended instrument parameters.

7.2 An autosampler is recommended but not required to automate sample injections and increase throughput. Autosamplers are usually available as an option from the instrument's manufacturer.

7.3 *Data Acquisition System*—Use the computer hardware and software recommended by the instrument manufacturer to control the apparatus and to collect data from the detector.

⁴ OI Analytical CNSolution 3100, FS3100, or Flow Solution IV and Lachat Instruments QuikChem Automated Ion Analyzer using Method 10-204-00-5-A have been found to be suitable for this analysis. If you are aware of alternative suppliers, please provide this information to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee,¹ which you may attend.

³ 40 CFR Part 136.




FIG. 2 Flow Injection Apparatus 2 with Automated Ligand Injection




FIG. 3 Flow Injection Analysis Apparatus 3

TABLE 1 Flow Injection Analysis Parameters

FIA Instrument Parameter	Recommended Method Setting
Pump Flow Rates	0.5 to 2 mL/min
Cycle period (total)	90 to 250 s/sample
Sample load period	At least enough time to completely fill the sample loop
Reagent water rinse time between samples	At least 15 seconds
Peak Evaluation	Peak height or area
Working Potential	0.0 V vs Ag/AgCl

7.4 *Pump Tubing*—Use tubing recommended by instrument manufacturer. Replace pump tubing when worn, or when precision is no longer acceptable.

7.5 *Gas Diffusion Membranes*—A hydrophobic membrane which allows gaseous hydrogen cyanide to diffuse from the donor to the acceptor stream at a sufficient rate to allow

detection. The gas diffusion membrane should be replaced when the baseline becomes noisy or every 1 to 2 weeks.⁵

7.6 Use parts and accessories as directed by instrument manufacturer.

8. Reagents and Materials

8.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the American Chemical Society, where such specifications are available.⁶ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

⁵ PALL Life Sciences Part Number M5PU025, OI Analytical Part Number A001520, and Lachat Instruments Part Number 50398 have found to be suitable for this analysis.

⁶ *Reagent Chemicals, American Chemical Society Specifications*, Am. Chemical Soc., Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia*.

8.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Type II grade of Specification **D1193**.

8.3 *Sodium Hydroxide Solution (1.00 M)*—Dissolve 40 g NaOH in laboratory water and dilute to 1 L.

8.4 *Acceptor Solution A (0.10 M NaOH)*—Dissolve 4.0 g NaOH in laboratory water and dilute to 1 L.

8.5 *Acceptor Solution B, Carrier B (0.025 M NaOH)*—Dissolve 1.0 g NaOH in laboratory water and dilute to 1 L.

8.6 *Stock Cyanide Solution (1000 µg/mL CN⁻)*—Dissolve 2.51 g of KCN and 2.0 g of NaOH in 1 L of water. Standardize with silver nitrate solution as described in Test Methods **D2036**, Section 16.2. Store the solution under refrigeration and check concentration approximately every 6 months and correct if necessary.⁷ (**Warning**—Because KCN is highly toxic, avoid contact or inhalation.)

8.7 *Intermediate Cyanide Standards:*

8.7.1 *Intermediate Cyanide Standard 1 (100 µg/mL CN⁻)*—Pipette 10.0 mL of stock cyanide solution (see 8.6) into a 100-mL volumetric flask containing 1 mL of 1.0 M NaOH (see 8.3). Dilute to volume with laboratory water. Store under refrigeration. The standard should be stable for 6 months.

8.7.2 *Intermediate Cyanide Standard 2 (10 µg/mL CN⁻)*—Pipette 10.0 mL of Intermediate Cyanide Standard 1 (see 8.7.1) into a 100-mL volumetric flask containing 1.0 mL of 1.00 M NaOH (see 8.3). Dilute to volume with laboratory water. Store under refrigeration. The standard should be stable for 6 months.

8.8 *Working Cyanide Calibration Standards*—Prepare fresh weekly as described in 8.8.1 and 8.8.2 ranging in concentration from 2 to 400 µg/L CN⁻.

8.8.1 *Calibration Standards (20, 50, 100, 200, and 400 µg/L CN⁻)*—Pipette 20, 50, 100, 200, and 400 µL of Intermediate Cyanide Standard 1 (see 8.7.1) into separate 100-mL volumetric flasks containing 1.0 mL of 1.00 M NaOH (see 8.3). Dilute to volume with laboratory water.

8.8.2 *Calibration Standards (2 and 10 µg/L CN⁻)*—Pipette 20 and 100 µL of Intermediate Cyanide Standard 2 (see 8.7.2) into separate 100-mL volumetric flasks containing 1.0 mL of 1.00 M NaOH (see 8.3). Dilute to volume with laboratory water.

8.9 *Cyanide Electrode Stabilization Solution (2 mg/L as CN⁻)*—Pipette 200 µL of Stock Cyanide (see 8.6) into a 100-mL volumetric flask containing 1.0 mL of 1.00 M NaOH (see 8.3). Dilute to volume with laboratory water. The solution should be stored under refrigeration.

8.10 *Acetate Buffer*—Dissolve 410 g of sodium acetate trihydrate (NaC₂H₃O₂·3H₂O) in 500 mL of laboratory water. Add glacial acetic acid (approximately 500 mL) to yield a pH of 4.5.

8.11 *Carrier A*—Use water as the carrier.

8.12 *Ligand Exchange Reagent 1 (TEP Solution)*—Weigh 0.10 g tetraethylenepentamine (TEP) into a 100-mL volumetric

flask. Dilute to volume with laboratory water. The solution should be stored at room temperature.

8.13 *Ligand Exchange Reagent 2 (Dithizone Solution)*—Weigh 0.010 g of dithizone into a 100-mL volumetric flask containing 1.0 mL of 1.00 M NaOH (see 8.3). Dilute to volume with laboratory water. Sonicate if necessary until all of the dithizone has dissolved. The solution should be stored at room temperature.

NOTE 2—Commercially prepared or alternative ligand exchange reagents can be used if equivalent results can be demonstrated. Commercial reagents should be used in accordance with manufacturer's instructions.⁸

8.14 *Mixed Ligand Exchange Reagent*, for automated ligand addition as shown in Fig. 2—Transfer 0.125 millilitres of WAD Reagent A and 0.250 milliliters of WAD Reagent B8 into a 100-mL volumetric flask containing 50-mL laboratory water. Dilute to volume with laboratory water and mix. The solution should be stored at room temperature

8.15 *Mercury (II) Cyanide Stock Solution*—Weigh 0.4854 g Hg(CN)₂ into a 100-mL volumetric flask. Place 1.0 mL of 1.00 M NaOH (see 8.3) in the flask and dilute to volume with laboratory water. Hg(CN)₂ as CN⁻ = 1000 mg/L. The solution must be stored in an amber glass bottle under refrigeration. The standard should be stable for 6 months.

8.16 *Mercury (II) Cyanide Intermediate Solution*—Pipet 10.0 mL of the mercury (II) cyanide stock solution (see 8.15) into a 100-mL volumetric flask containing 1.0 mL of 1.00 M NaOH (see 8.3). Dilute to volume with laboratory grade water. Hg(CN)₂ as CN⁻ = 100 mg/L. The solution must be stored in an amber glass bottle under refrigeration. The standard should be stable for 6 months.

8.17 *Mercury (II) Cyanide Recovery Solution*—Pipet 100 µL of mercury II cyanide intermediate solution (see 8.16) into a 100-mL volumetric flask containing 1.0 mL of 1.00 M NaOH (see 8.3). Dilute to volume with laboratory water. Hg(CN)₂ as CN⁻ = 100 µg/L. Prepare fresh weekly.

8.18 *Potassium Nickel Cyanide Stock Solution*—Weigh 0.2488 g of K₂Ni(CN)₄·H₂O in a 100-mL volumetric flask. Place 1.0 mL of 1.00 M NaOH (see 8.3) in the flask and dilute to volume with laboratory water. K₂Ni(CN)₄ as CN⁻ = 1000 mg/L. The solution must be stored in an amber glass bottle under refrigeration. The standard should be stable for 6 months.

8.19 *Potassium Nickel Cyanide Intermediate Solution*—Pipet 10.0 mL of the potassium nickel cyanide stock solution (see 8.18) into a 100-mL volumetric flask containing 1.0 mL of 1.00 M NaOH (see 8.3). Dilute to volume with laboratory grade water. K₂Ni(CN)₄ as CN⁻ = 100 mg/L. The solution must be stored in an amber glass bottle under refrigeration. The standard should be stable for 6 months.

8.20 *Potassium Nickel Cyanide Recovery Solution*—Pipet 100 µL of potassium nickel cyanide intermediate solution (see 8.19) into a 100-mL volumetric flask containing 1.0 mL of 1.00

⁷ Commercial Solutions of Stock Cyanide may be substituted.

⁸ OI Analytical WAD Reagents A and B, PN A001416 and A001417 have found to be suitable for this analysis

M NaOH (see 8.3). Dilute to volume with laboratory water. $K_2Ni(CN)_4$ as $CN^- = 100 \mu\text{g/L}$. Prepare fresh weekly.

8.21 *Ag/AgCl Reference Electrode Filling Solution*—Fill the reference electrode as recommended by the instrument manufacturer.

8.22 *Sulfide Removal and Acidification Reagent*—Weigh 1.00 g bismuth nitrate pentahydrate, $Bi(NO_3)_3 \cdot 5H_2O$, into a 1-L volumetric flask. Add 55 mL of water then carefully, add 55 mL of concentrated sulfuric acid to the flask. Gently swirl the flask until the bismuth nitrate pentahydrate has dissolved in the acid solution. Carefully add water to the volumetric flask and fill to volume. (**Warning**—This is an exothermic reaction and the solution will become hot when preparing this solution.)

9. Hazards

9.1 **Warning**—Because of the toxicity of cyanide, great care must be exercised in its handling. Acidification of cyanide solutions produces toxic hydrocyanic acid (HCN). All manipulations must be done in the hood so that any HCN gas that might escape is safely vented.

9.2 **Warning**—Many of the reagents used in these test methods are highly toxic. These reagents and their solutions must be disposed of properly.

9.3 All reagents and standards should be prepared in volumes consistent with laboratory use to minimize the generation of waste.

10. Sample and Sample Preservation

10.1 Collect the sample in accordance with Practice D7365. This standard practice is applicable for the collection and preservation of water samples for the analysis of cyanide. Responsibilities of field sampling personnel and the laboratory are indicated.

11. Elimination of Interferences

11.1 Practice D7365 specifies mitigation of interference procedures for testing water samples for cyanide.

11.2 If samples contain high levels of CO_3^{2-} (above 1500 ppm), negative bias or irregular peak shapes can result. In this case, precipitate the carbonate with $Ca(OH)_2$ (hydrated lime) as specified in Practice D7365.

11.3 *Sulfide*—The acidification reagent contains bismuth nitrate that effectively mitigates sulfide up to 50-mg/L sulfide. Refer to Practice D7365 for additional sulfide mitigation procedures.

12. Calibration and Standardization

12.1 Turn on the power to the apparatus and the autosampler (if equipped). Start the data acquisition system.

12.2 Clamp the pump tube platens in place and start pumping reagents in the flow injection system. Allow the system to warm up at least 15 min or until a stable baseline is achieved. Take care not to over-tighten the pump tubes platens as this will greatly reduce their lifetime.

12.3 If recommended by the instrument manufacturer, aspirate the Cyanide Electrode Stabilization Solution (2 mg/L CN^-)

from 8.9. After at least 30 s, inject the shocking solution into the apparatus and record the amperometric response (pA value) after the cycle period has completed. Repeat this procedure until the peak responses are less than 2 % RSD. This process will ensure that the electrode system has stabilized.

12.4 After the electrode system has stabilized, aspirate the highest working standard (see 8.8) into the flow injection apparatus. Follow the instrument manufacturer's instructions to store the retention time window for cyanide using the data acquisition software.

12.5 Pipette 50 μL of Ligand Exchange Reagent 1 (TEP) and 250 μL of Ligand Exchange Reagent 2 (Dithizone, see 8.13) into 5 mL of each working standard from 8.8 and swirl to a mix. Prepare a reagent blank for background correction if necessary. If alternate ligand exchange reagents are used, follow manufacturer's instructions.

NOTE 3—TEP addition alone will yield higher recoveries of mercury cyanide species than other weak acid dissociable cyanide methods; however, complete recovery is only possible with the addition of the dithizone reagent. If the samples are known to be free of mercury or if mercury analyses indicate the absence of mercury, the dithizone ligand exchange reagent may be omitted in the samples and working standards.

12.5.1 If using the Mixed Ligand Exchange Reagent (see 8.14) and the configuration shown in Fig. 2, step 12.5 is not necessary. Inject mixed ligand reagent at a mixed ligand solution to sample ratio of at least 1:10.

12.6 Inject each working standard and the reagent blank from 12.5 into the apparatus and record the amperometric response with the data acquisition system. Plot the response versus the cyanide concentration with a straight line or a quadratic fit curve depending on the instrument and data acquisition system employed. If the calibration model is polynomial, it may be no more than third order. A second order polynomial is recommended. An example of a calibration curve is shown in Fig. 4.

12.7 Prepare a new calibration curve at least once daily.

13. Procedure

13.1 Place 5 mL of each sample to be tested in separate polyethylene containers. Pipette 50 μL of Ligand Exchange Reagent 1 (TEP, see 8.12) and 250 μL of Ligand Exchange Reagent 2 (Dithizone, see 8.13) into the sample and mix. If alternate ligand exchange reagents are used, follow manufacturer's instructions.

13.1.1 If using the Mixed Ligand Exchange Reagent (see 8.14) and the configuration shown in Fig. 2, step 13.1 is not necessary. Inject mixed ligand reagent at a mixed ligand solution to sample ratio of at least 1:10.

13.2 Inject each sample into the flow injection apparatus, and inspect for irregular peak shapes, disturbances, or detector overloads. Dilute and re-run samples if necessary.

14. Data Analysis and Calculations

14.1 Report the cyanide as parts per billion ($\mu\text{g/L}$) available cyanides using the data acquisition software.

14.2 Multiply the result by any dilution factor and round the test result to three significant figures.




FIG. 4 Example of Calibration Curve

14.3 Some instruments are capable of performing multiple injections in which the mean result for each sample can be reported. In this case, the mean result should be reported.

15. Precision and Bias⁹

15.1 Based on the results of 10 operators in 10 laboratories, the overall and single operator precision and method bias data are shown in Table 2. The precision and bias data were obtained from a synthetic wastewater and may not apply to all untested matrices. During the interlaboratory study, 0.1 M HCl was used as the acidification reagent and carrier solutions instead of the sulfide abatement reagent described in Section 8.22. This technique has been shown to be effective for this application in an independent study.¹⁰

15.2 This method was also evaluated and validated in a single laboratory. In the lab study, this procedure was found to be suitable for several natural and industrial matrices including coal strip mining water, surface water, spring water, river water, treated municipal wastewater (POTW), industrial wastewater, groundwater, and drinking water. Recoveries of potassium nickel cyanide and mercury (II) cyanide (fortified with 100 µg/L as CN⁻) ranged from 89.9 to 99.6 % and 82.9 to 99.3 %, respectively.¹¹

15.2.1 The automated ligand exchange reagent addition as shown in Fig. 2 was not included in the interlaboratory study. Single laboratory results obtained recoveries of potassium

nickel cyanide (see 8.20) and mercury (II) cyanide (see 8.17) within the specifications of 16.5.

16. Quality Assurance and Quality Control

16.1 In order to be certain that analytical values obtained using this test method are valid and accurate within the confidence limits of the test, the following QC procedures must be followed when running the test. For a general discussion of quality control and good laboratory practices, see Practice D5847 and Guide D3856.

16.2 Calibration and Calibration Verification:

16.2.1 Analyze the calibration standards daily prior to analysis to calibrate the instrument as described in Section 12.

16.2.2 Verify instrument calibration for each analytical batch of 10 samples by analyzing a mid-point standard. The recovery should be 90 to 110 % or else corrective actions should be taken.

16.3 Initial Demonstration of Laboratory Capability:

16.3.1 If a laboratory has not performed the test before or if there has been a major change in the measurement system, for example, new analyst, new instrument, etc., a precision and bias study must be performed to demonstrate laboratory capability.

16.3.2 Analyze seven replicates of a standard solution prepared from an independent reference material (IRM) containing 70 µg/L available cyanides. The matrix of the solution should be equivalent to the solution used in the collaborative study. Each replicate must be taken through the complete analytical procedure. The replicates may be interspersed with samples.

16.3.3 Calculate the mean and standard deviation of the seven values. The mean should range from 48.8 to 79.6 µg/L and the standard deviation should be less than 11.1, otherwise the study should be repeated until these criteria are met. If a

⁹ Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D19-1173. Contact ASTM Customer Service at service@astm.org.

¹⁰ Milosavljevic, S., "How to Analyze for Cyanide," *Cyanide: Social Industrial and Economic Aspects*, The Minerals, Metals & Materials Society 2001, pp. 125–128.

¹¹ Sebroski, O., "Method Comparison and Evaluation for the Analysis of Weak Acid-Dissociable Cyanide," *Environmental Science and Technology*, Vol 31, No. 1, 1997, pp. 52–57.

TABLE 2 Precision and Bias for Available Cyanides

ASTM Test Method D6888: Synthetic Wastewater—Final Statistical Summary for Available Cyanide Analyses						
Sample Number	AC19721	AC19723	AC19726	AC19722	AC19724	AC19725
Number of retained values	8	8	8	8	8	8
True concentration (C), µg/L	8.00	9.00	70.0	80.0	300	350
Mean recovery (XBAR)	7.10	8.32	64.2	71.5	266	315
Percent recovery	88.7	92.4	91.7	89.4	88.7	90.1
Overall standard deviation (S _T)	0.815	0.459	6.00	9.66	23.7	27.2
Overall relative standard deviation, %	11.5	5.52	9.35	13.5	8.92	8.63
Number of retained pairs	8		8		8	
Single standard deviation (S _o)	0.618		4.39		6.43	
Analyst relative deviation, %	8.02		6.47		2.21	

Synthetic Wastewater was prepared in 0.1 % synthetic sea salts.

concentration other than the recommended concentration is used, refer to Test Method **D5847** for information on applying the *F* test and *t* test in evaluating the acceptability of the mean and standard deviation.

16.4 Laboratory Control Sample (LCS):

16.4.1 To ensure that the test method is in control, analyze a mercury(II) cyanide or potassium nickel cyanide recovery solution (see **8.17** and **8.20**). The recoveries should be 81 to 121 % for mercury(II)cyanide and 90 to 117 % for potassium nickel cyanide or else corrective actions should be taken.

16.5 Method Blank:

16.5.1 Analyze a method blank with each batch of samples. A laboratory method blank can be prepared by adding 1.0 mL of 1.00 M NaOH (see **8.3**) into a 100-mL volumetric flask and diluting to volume with laboratory water.

16.5.2 The measured concentration of available cyanides must be less than 2 µg/L. If the concentration is found above this level, analysis of samples is halted until the contamination is eliminated and a blank shows no contamination at or above this level, or the results should be qualified with an indication that they do not fall within the performance of the test method.

16.6 Matrix Spike (MS):

16.6.1 To check for interferences in the specific matrix being tested, perform an MS on at least one sample from each batch by spiking an aliquot of the sample with a known concentration of cyanide and taking it through the analytical method. The spike must produce a concentration in the spiked sample 2 to 5 times the background concentration or 100 µg/L cyanide, whichever is greater. Cyanide matrix spikes can be prepared from the intermediate cyanide solutions-potassium cyanide (**8.7.1**), mercury cyanide (**8.16**), or potassium nickel cyanide (**8.19**). For example, partially fill a 100-mL volumetric flask with sample, add 100 µL of intermediate cyanide solution, then fill to volume with sample to produce a 100 µg/L cyanide matrix spike.

16.6.2 If the recovery is not within the limits as described in Practice **D5847**, a matrix interference may be present in the sample selected for spiking. Under these circumstances, one of the following remedies must be employed: the matrix interfer-

ence must be removed, all samples in the batch must be analyzed by a test method not affected by the matrix interference, or the results should be qualified with an indication that they do not fall within the performance criteria of the test method.

16.7 Duplicate:

16.7.1 To check the precision of sample analyses, analyze a sample in duplicate with each batch. If the concentration is less than five times the detection limit, an MS duplicate (MSD) should be used.

16.7.2 Calculate the standard deviation of the duplicate values and compare to the single operator precision from the collaborative study using an *F* test. Refer to 6.5.5 of Practice **D5847** for information on applying the *F* test.

16.7.3 If the result exceeds the precision limit, the batch must be reanalyzed or the results must be qualified with an indication that they do not fall within the performance criteria of the method.

16.8 Independent Reference Material:

16.8.1 In order to verify the quantitative value produced by the test method, analyze an IRM submitted as a regular sample (if practical) to the laboratory at least once per quarter. The concentration of the reference material should be in the range of this method. The value obtained must fall within the control limits specified by the outside source.

16.9 The analyst is permitted certain options to improve the performance of this method, provided that all performance specifications are met. These options include sample pretreatment to remove interferences and the use of alternative ligand exchange reagents. Any time such modifications are made, the Initial Demonstration of Proficiency must be successfully repeated.

17. Keywords

17.1 amperometry; available cyanides; cyanide; cyanide amenable to chlorination; flow injection analysis; free cyanide; gas diffusion membrane; ligand exchange; silver electrode; weak acid dissociable cyanide

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; <http://www.copyright.com/>