



Standard Test Method for Determination of Dipropylene Glycol Monobutyl Ether and Ethylene Glycol Monobutyl Ether in Sea Water by Liquid Chromatography/Tandem Mass Spectrometry (LC/MS/MS)¹

This standard is issued under the fixed designation D7731; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

^ε¹ NOTE—This test method was changed editorially in August 2011.

1. Scope

1.1 This procedure covers the determination of Dipropylene Glycol Monobutyl Ether (DPGBE) and Ethylene Glycol Monobutyl Ether (EGBE) in sea water by direct injection using liquid chromatography (LC) and detection with tandem mass spectrometry (MS/MS). These analytes are qualitatively and quantitatively determined by this method. This method adheres to selected reaction monitoring (SRM) mass spectrometry.

1.2 The Detection Verification Level (DVL) and Reporting Range for DPGBE and EGBE are listed in [Table 1](#).

1.2.1 The DVL is required to be at a concentration at least 3 times below the Reporting Limit (RL) and have a signal/noise ratio greater than 3:1. [Fig. 1](#) and [Fig. 2](#) display the signal/noise ratio of the single reaction monitoring (SRM) transition.

1.2.2 The reporting limit is the concentration of the Level 1 calibration standard as shown in [Table 4](#) for DPGBE and EGBE, taking into account the 20% sample preparation dilution factor.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:²

[D1193 Specification for Reagent Water](#)
[D2777 Practice for Determination of Precision and Bias of Applicable Test Methods of Committee D19 on Water](#)

2.2 Other Standards:³

[EPA publication SW-846 Test Methods for Evaluating Solid Waste, Physical/Chemical Methods](#)

3. Terminology

3.1 Definitions:

3.1.1 *detection verification level, DVL, n*—a concentration that has a signal/noise ratio greater than 3:1 and is at least 3 times below the Reporting Limit (RL).

3.1.2 *reporting limit, RL, n*—the concentration of the lowest-level calibration standard used for quantification.

3.1.2.1 *Discussion*—In this test method, a 20 mL sample aliquot is diluted to a 25 mL final volume after thoroughly rinsing the collection vial with acetonitrile for quantitative transfer. In this case, the lowest calibration level of 100 ppb for EGBE would allow for a reporting limit of 125 ppb to be achieved.

3.2 Symbols:

3.2.1 *ppb*—parts per billion, $\mu\text{g/L}$

3.2.2 *ppt*—parts per trillion, ng/L

3.2.3 *mM*—millimolar, 1×10^{-3} moles/L

3.2.4 *NA*—no addition

3.2.5 *ND*—non-detect

¹ This test method is under the jurisdiction of ASTM Committee D19 on Water and is the direct responsibility of Subcommittee D19.06 on Methods for Analysis for Organic Substances in Water.

Current edition approved May 1, 2011. Published June 2011. DOI: 10.1520/D7731-11E01.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from National Technical Information Service (NTIS), U.S. Department of Commerce, 5285 Port Royal Road, Springfield, VA, 22161 or at <http://www.epa.gov/epawaste/hazard/testmethods/index.htm>

TABLE 1 Detection Verification Level and Reporting Range

Analyte	DVL (µg/L)	Reporting Range (µg/L)
DPGBE	0.2	1–10
EGBE	25	125–1250

4. Summary of Test Method

4.1 This is a performance based method, and modifications are allowed to improve performance.

4.2 For DPGBE and EGBE analysis, samples are shipped to the lab between 0°C and 6°C and analyzed within 5 days of collection. The DOW MSDS sheet on DOWANOL* DPNB glycol ether (DPGBE) Issue Date: 06/18/2010 lists that the material is readily biodegradable. The Organisation for Economic Co-Operation and Development (OECD) 302B Test lists 96% biodegradation in 28 days.

4.3 In the lab, the entire collected 20 mL sample is spiked with surrogate and brought to a volume of 25 mL with acetonitrile. This prepared sample is then filtered using a syringe driven filter unit, and analyzed by LC/MS/MS. If visible oil is present, the prepared sample is allowed to settle resulting in an oil layer at the top of the 25 mL solution. A portion of the aqueous (bottom) layer is filtered, leaving the oil layer behind, through a syringe driven filter assembly and analyzed by LC/MS/MS.

4.4 DPGBE, EGBE and surrogate are identified by retention time and one SRM transition. The target analytes and surrogate are quantitated using the SRM transitions utilizing an external calibration. The final report issued for each sample lists the concentration of DPGBE, EGBE and the surrogate recovery.

5. Significance and Use

5.1 DPGBE and EGBE have a variety of residential and industrial applications such as cleaning formulations, surface coatings, inks and cosmetics. These analytes may be released into the environment at levels that may be harmful to aquatic life.

5.2 This method has been investigated for use with reagent and sea water.

6. Interferences

6.1 Method interferences may be caused by contaminants in solvents, reagents, glassware, and other apparatus producing discrete artifacts or elevated baselines. All of these materials are demonstrated to be free from interferences by analyzing laboratory reagent blanks under the same conditions as samples.

6.2 All glassware is washed in hot water with detergent and rinsed in hot water followed by distilled water. Detergents containing DPGBE or EGBE must not be used. The glassware is then dried and heated in an oven at 250°C for 15 to 30 minutes. All glassware is subsequently cleaned with acetone followed by methanol.

6.3 All reagents and solvents should be pesticide residue purity or higher to minimize interference problems.

6.4 Matrix interferences may be caused by contaminants in the sample. The extent of matrix interferences can vary considerably from sample source depending on variations of the sample matrix.

7. Apparatus

7.1 LC/MS/MS System

7.1.1 *Liquid Chromatography System*—A complete LC system is needed in order to analyze samples.⁴ Any system that is capable of performing at the flows, pressures, controlled temperatures, sample volumes, and requirements of the standard may be used.

7.1.2 *Analytical Column*—Waters- XBridge™, 2.1 x 150 mm, 3.5 µm particle size was used to develop this test method. Any column that achieves baseline resolution of these analytes may be used. Baseline resolution simplifies data analysis and can reduce the chance of ion suppression, leading to higher limits of detection. The retention times and order of elution may change depending on the column used and need to be monitored.

7.1.3 *Tandem Mass Spectrometer System*—A MS/MS system capable of SRM analysis.⁵ Any system that is capable of performing at the requirements in this procedure may be used.

7.2 Filtration Device:

7.2.1 *Hypodermic syringe*—A Lock Tip Glass Syringe capable of holding a Millex® HV Syringe Driven Filter Unit PVDF 0.22 µm or similar may be used.

7.2.1.1 A 25 mL Lock Tip Glass Syringe size was used in this test method.

7.2.2 *Filter*—Millex® HV Syringe Driven Filter Unit PVDF 0.22 µm (Millipore Corporation, Catalog #SLGV033NS) or similar may be used.

8. Reagents and Materials

8.1 *Purity of Reagents*—High Performance Liquid Chromatography (HPLC) pesticide residue analysis and spectrophotometry grade chemicals shall be used in all tests. Unless indicated otherwise, it is intended that all reagents shall conform to the Committee on Analytical Reagents of the American Chemical Society.⁶ Other reagent grades may be used provided they are first determined to be of sufficiently high purity to permit their use without affecting the accuracy of the measurements.

8.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming

⁴ A Waters Alliance High Performance Liquid Chromatography (HPLC) System was used to develop this test method. All parameters in this test method are based on this system and may vary depending on your instrument.

⁵ A Waters Quattro micro API tandem quadrupole mass spectrometer was used to develop this test method. All parameters in this test method are based on this system and may vary depending on your instrument.

⁶ Reagent Chemicals, American Chemical Society Specifications, American Chemical Society, Washington, D.C. For Suggestions on the testing of reagents not listed by the American Chemical Society, see Annual Standards for Laboratory Chemicals, BDH Ltd., Poole, Dorset, U.K., and the United States Pharmacopeia and National Formulators, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.




FIG. 1 Detection Verification Level Signal/Noise Ratio.




FIG. 2 Reporting Level (Calibration standard) Signal/Noise Ratio.

to ASTM Type 1 of Specification D1193. It must be demon-

strated that this water does not contain contaminants at concentrations sufficient to interfere with the analysis.

8.3 *Gases*—Ultrapure nitrogen and argon.

8.4 Acetonitrile (CAS # 75-05-8).

8.5 Methanol (CAS # 67-56-1).

8.6 Formic Acid (CAS # 64-18-6).

8.7 2-Propanol (CAS # 67-63-0).

8.8 *DPGBE*—Dipropylene Glycol Monobutyl Ether (CAS # 29911-28-2).

8.9 *EGBE*—Ethylene Glycol Monobutyl Ether (CAS# 111-76-2).

8.10 *n-NP2EO*—normal- Nonylphenol Diethoxylate (CAS# Not available).⁷

8.11 *EGBE-D₄* (2-butoxyethanol (1,1,2,2-D₄)) (Optional Surrogate, Unlabeled CAS# 111-76- 2).⁸

9. Hazards

9.1 Normal laboratory safety applies to this method. Analysts should wear safety glasses, gloves, and lab coats when working in the lab. Analysts should review the Material Safety Data Sheets (MSDS) for all reagents used in this method.

10. Sampling

10.1 *Sampling and Preservation*—Grab samples should be collected in 20 mL pre-cleaned glass vials with Teflon® lined septa caps demonstrated to be free of interferences. The vial should be filled to approximately 20 mL. This may be just below the neck of the vial, depending on the vial manufacturer. This test method is based on a 20 mL sample size per analysis. Each sample should be collected in duplicate and a quadruplicate sample must be included with each sample batch of 10 for MS/MSD quality control analyses. Store samples between 0°C and 6°C from sample collection to sample preparation. Analyze the sample within 5 days of collection.

11. Preparation of Apparatus

11.1 *Liquid Chromatograph Operating Conditions*:⁴

11.1.1 Injection volumes of all calibration standards and samples are made at 100 µL volume. The first sample analyzed after the calibration curve is a blank to ensure there is no

carry-over. The gradient conditions for the liquid chromatograph are shown in **Table 2**. Divert the column flow away from the electrospray source from 0 to 5 minutes after injection. Flow diversion to waste may be done using the mass spectrometer divert valve. Divert tubing configurations vary from manual injection. Sea water samples contain nonvolatile salts, the first 5 minute elution is diverted in order to keep the mass spectrometer source clean.

11.2 *LC Conditions*:

11.2.1 *Needle Wash Solvent*—60% Acetonitrile/40% 2-propanol

11.2.2 *Temperatures*—Column, 30°C; Sample compartment, 15°C.

11.2.3 *Seal Wash*—60% Acetonitrile/40% 2-propanol.

11.3 *Mass Spectrometer Parameters*:⁵

11.3.1 To acquire the maximum number of data points per SRM channel while maintaining adequate sensitivity, the tune parameters may be optimized according to your instrument. Each peak requires at least 10 scans per peak for adequate quantitation. This procedure contains DPGBE, EGBE and one surrogate which are in three SRM acquisition functions to optimize sensitivity. Variable parameters regarding retention times, SRM transitions, and cone and collision energies are shown in Table 3. Mass spectrometer parameters used in the development of this method are listed here:

Capillary Voltage: 3.5 kV

Cone: Variable depending on analyte (**Table 3**)

Extractor: 2 Volts

RF Lens: 0.2 Volts

Source Temperature: 120°C

Desolvation Temperature: 350°C

Desolvation Gas Flow: 800 L/hr

Cone Gas Flow: 25 L/hr

Low Mass Resolution 1: 14.5

High Mass Resolution 1: 14.5

Ion Energy 1: 0.5

Entrance Energy: -1

Collision Energy: Variable depending on analyte (**Table 3**)

Exit Energy: 1

Low Mass Resolution 2: 14.5

High Mass resolution 2: 14.5

Ion Energy 2: 0.8

Multiplier: 650

Gas Cell Pirani Gauge: 7.0 x 10⁻³ Torr

Inter-Channel Delay: 0.1 seconds

Inter-Scan Delay: 0.1 seconds

Dwell: 0.1 seconds

Solvent Delay: 5 minutes

TABLE 2 Gradient Conditions for Liquid Chromatography

Time (min)	Flow (mL/min)	Percent 95% Water/ 5% CH ₃ CN	Percent CH ₃ CN	Percent 2% Formic Acid 95% Water/ 5% CH ₃ CN
0.0	0.30	95	0	5
2.0	0.30	95	0	5
5.0	0.30	0	95	5
14.0	0.30	0	95	5
15.0	0.30	95	0	5
18.0	0.30	95	0	5

⁷ A source of *n*-NP2EO is Accustandard, Inc. 125 Market Street, New Haven, CT 06513 or Cambridge Isotope Laboratories, 50 Frontage Road, Andover, MA 01810-5413.

⁸ A source of *EGBE-D₄* is Cambridge Isotope Laboratories, 50 Frontage Road, Andover, MA 01810-5413.

TABLE 3 Retention Times, SRM transitions, and Specific Mass Spectrometer Parameters

Analyte	Retention time (min)	Cone Voltage (Volts)	Collision Energy (eV)	SRM Mass Transition (Precursor > Product)
DPGBE	8.5	19	7	191.3 > 115.1
EGBE	7.6	13	5	119.1 > 62.9
<i>n</i> -NP2EO (Surrogate)	11.2	28	10	309.3 > 89.0
EGBE-D ₄ (Optional Surrogate)	7.6	13	5	123.0 > 66.8

12. Calibration and Standardization

12.1 The mass spectrometer must be calibrated per manufacturer specifications before analysis. In order to obtain accurate analytical values through using this test method within the confidence limits, the following procedures must be followed when performing the test method. Prepare all solutions in the lab using Class A volumetric glassware.

12.2 *Calibration and Standardization*—To calibrate the instrument, analyze six calibration standards and the DVL containing (nominal concentrations in [Table 4](#)) DPGBE, EGBE and *n*-NP2EO. A calibration solution is prepared from standard materials or they are purchased as certified solutions. Level 6 calibration solution containing the targets and surrogate is prepared and aliquots of that solution are diluted to prepare Levels 1 through 5 and the DVL. The following steps will produce standards with the concentration values shown in [Table 4](#). The analyst is responsible for recording initial component weights correctly and calculating and preparing appropriate dilution calculations.

12.2.1 Prepare Level 6 calibration stock standard at 1000 ppb for EGBE, 8 ppb for DPGBE and 40 ppb for *n*-NP2EO in 80% water/20% acetonitrile. The EGBE and DPGBE concentrated stock solutions were prepared in methanol at approximately 2 g/L concentration and the *n*-NP2EO surrogate concentrated stock solution was prepared in acetonitrile at approximately 0.5 g/L. The preparation of the stock standard can be accomplished using different volumes and concentrations of stock solutions as is accustomed in the individual laboratory. Depending on the prepared stock concentrations, the solubility at that concentration will have to be ensured.

12.2.2 Aliquots of Level 6 calibration stock standard are then diluted with 80% water/20% acetonitrile to prepare the desired calibration levels in 2 mL amber glass autosampler vials. The calibration vials must be used within 24 hours to ensure optimum results. Stock calibration standards are routinely replaced every 7 days if not previously discarded for quality control failure. Calibration standards are not filtered.

12.2.3 Inject each standard and obtain its chromatogram. An external calibration technique is used to monitor the SRM transitions of each analyte. Calibration software is utilized to conduct the quantitation of the target analytes and surrogates using the SRM transition. The calibration software manual should be consulted to use the software correctly. The quanti-

tation method is set as an external calibration using the peak areas in ppb units. Concentrations may be calculated using the data system software to generate linear regression or quadratic calibration curves. Forcing the calibration curve through the origin is not recommended.

12.2.4 Linear calibration may be used if the coefficient of determination, r^2 , is >0.98 for the analyte. The point of origin is excluded and a fit weighting of 1/X is used in order to give more emphasis to the lower concentrations. If one of the calibration standards other than the high or low point causes the r^2 of the curve to be <0.98, this point must be re-injected or a new calibration curve must be regenerated. If the low and/or high point is excluded, minimally a five point curve is acceptable but the reporting range must be modified to reflect this change.

12.2.5 Quadratic calibration may be used if the coefficient of determination, r^2 , is >0.99 for the analyte. The point of origin is excluded, and a fit weighting of 1/X is used in order to give more emphasis to the lower concentrations. If one of the calibration standards causes the curve to be <0.99, this point must be re-injected or a new calibration curve must be regenerated. Minimally a six point curve is acceptable using a quadratic fit. Each calibration point used to generate the curve must have a calculated percent deviation less than 25% from the generated curve.

12.2.6 The retention time window of the SRM transitions must be within 5% of the retention time of the analyte in a midpoint calibration standard. If this is not the case, re-analyze the calibration curve to determine if there was a shift in retention time during the analysis and re-inject the sample. If the retention time is still incorrect in the sample, refer to the analyte as an unknown.

12.2.7 A calibration midpoint check standard must be analyzed at the end of each batch of 20 samples or within 24 hours after the initial calibration curve was generated. This end calibration check should be the same calibration standard that was used to generate the initial curve. The results from the end calibration check standard must have a percent deviation less than 35% from the calculated concentration for the target analytes and surrogates. If the results are not within these criteria, the problem must be corrected and either all samples in the batch must be re-analyzed against a new calibration curve or the affected results must be qualified with an indication that

TABLE 4 Concentrations of Calibration Standards (PPB)

Analyte/Surrogate	DVL	LV1	LV2	LV3	LV4	LV5	LV6
DPGBE	0.20	0.80	1.6	2.4	3.2	4.0	8.0
EGBE	25	100	200	300	400	500	1000
<i>n</i> -NP2EO (Surrogate)	1.0	4.0	8.0	12	16	20	40

they do not fall within the performance criteria of the test method. If the analyst inspects the vial containing the end calibration check standards and notices that the samples evaporated affecting the concentration, a new end calibration check standard may be made and analyzed. If this new end calibration check standard has a percent deviation less than 35% from the calculated concentration for the target analyte and surrogate, the results may be reported unqualified.

12.3 If a laboratory has not performed the test before or if there has been a major change in the measurement system, for example, new analyst, new instrument, etc., a precision and bias study must be performed to demonstrate laboratory capability.

12.3.1 Analyze at least four replicates of a sample solution containing the targets and surrogate at a concentration in the calibration range of Levels 3 to 5. The Level 3 concentration of the 6 point calibration curve was used to set the QC acceptance criteria in this method. The matrix and chemistry should be similar to the solution used in this test method. Each replicate must be taken through the complete analytical test method including any sample pre-treatment steps.

12.3.2 Calculate the mean (average) percent recovery and relative standard deviation (RSD) of the four values and compare to the acceptable ranges of the QC acceptance criteria for the Initial Demonstration of Performance in **Table 5**.

12.3.3 This study should be repeated until the single operator precision and mean recovery are within the limits in **Table 5**.

12.3.3.1 The QC acceptance criteria for the Initial Demonstration of Performance in **Table 5** is preliminary until more data and multi-laboratory study is completed. Data generated from a single-laboratory validation from reagent and sea water matrices are shown in the Precision and Bias Section **16**. It is recommended that the laboratory generate their own in-house QC acceptance criteria which meet or exceed the criteria in this standard. A reference on how to generate QC acceptance criteria is in Method 8000B in EPA publication SW-846.

12.4 Surrogate Spiking Solution:

12.4.1 A surrogate spiking methanol solution containing *n*-NP2EO is added to all samples. A stock surrogate spiking solution is prepared at 2.4 ppm. Spiking 100 µL of this spiking solution into a 20 mL water sample results in a concentration of 12 ppb of the surrogate in the sample. The result obtained for the surrogate recovery must fall within the limits of **Table 5**. If the limits are not met, the affected results must be qualified with an indication that they do not fall within the performance criteria of the test method.

12.4.1.1 *n*-NP2EO has been shown to be absorbed into the oil layer yielding a non-detect as a result. If oil is present in the

sample, the recovery of the *n*-NP2EO surrogate may be very low or not detected at or above the reporting limit.

12.5 Method Blank:

12.5.1 Analyze a reagent water blank with each batch of 20 or fewer samples. The concentration of the DPGBE and EGBE found in the blank must be below the DVL. If the concentration of DPGBE or EGBE is found above this level, analysis of samples is halted until the contamination is eliminated, and a blank shows no contamination at or above this level, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method. If DPGBE or EGBE are found in a method blank at greater than the reporting limit the reporting limit must be raised to at least 2 times the concentration of the DPGBE and EGBE found in the blank. This may occur if samples are encountered that have a high concentration of DPGBE, a water blank between samples may be required to remove carry-over between samples.

12.6 Laboratory Control Sample (LCS):

12.6.1 To ensure that the test method is in control, analyze a LCS prepared with the target analytes at a concentration in the calibration range of Levels 3 to 5. The LCS is prepared following the analytical method and analyzed with each batch of 20 samples or less. Prepare a stock matrix spiking solution in methanol containing the DPGBE at 0.48 ppm and EGBE at 60 ppm. Spike 100 µL of this stock solution into 20 mL of water to yield a concentration of 2.4 ppb for the DPGBE and 300 ppb for EGBE in the sample. The LCS result must be within the limits in **Table 5**. Matrix spiking solutions are routinely replaced every 7 days if not previously discarded for quality control failure.

12.6.2 If the result is not within these limits, analysis of samples is halted until the problem is corrected, and either all samples in the batch must be re-analyzed, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

12.7 Matrix Spike/Matrix Spike Duplicate (MS/MSD):

12.7.1 To check for interferences in the specific matrix being tested, perform a MS/MSD on at least one sample from each batch of 10 or fewer samples by spiking the sample with a known concentration of DPGBE and EGBE and following the analytical method. Prepare a stock matrix spiking solution in methanol containing the DPGBE at 0.48 ppm and EGBE at 60 ppm. Spike 100 µL of this stock solution into 20 mL of water to yield a concentration of 2.4 ppb for the DPGBE and 300 ppb for EGBE in the sample. The result obtained for the MS/MSD must fall within the limits in **Table 6**. Matrix spiking solutions are routinely replaced every 7 days if not previously discarded for quality control failure.

TABLE 5 Preliminary QC Acceptance Criteria

Analyte/Surrogate	Test Conc. (µg/L) in Reagent Water	Initial Demonstration of Performance			Lab Control Sample	
		Recovery (%)		Precision	Recovery (%)	
		Lower Limit	Upper Limit	Maximum % RSD	Lower Limit	Upper Limit
DPGBE	2.4	50	150	30	50	150
EGBE	300	50	150	30	50	150
<i>n</i> -NP2EO (Surrogate)	12	25	150	30	25	150

TABLE 6 Preliminary MS/MSD QC Acceptance Criteria

Analyte	Test Conc. (µg/L)	MS/MSD		
		Recovery (%)		Precision
		Lower Limit	Upper Limit	Maximum RPD (%)
DPGBE	2.4	50	150	30
EGBE	300	50	150	30
<i>n</i> -NP2EO (Surrogate)	12	25	150	30

12.7.2 If the spiked concentration plus the background concentration exceeds that of the Level 6 calibration standard, the sample must be diluted to a level near the midpoint of the calibration curve.

12.7.3 Calculate the percent recovery of the spike (*P*) using Eq 1:

$$P = 100 \frac{|A(V_s + V) - BV_s|}{CV} \quad (1)$$

where:

- A* = concentration found in spiked sample,
- B* = concentration found in unspiked sample,
- C* = concentration of analyte in spiking solution,
- V_s* = volume of sample used,
- V* = volume of spiking solution added, and
- P* = percent recovery.

12.7.4 The percent recovery of the spike shall fall within the limits in Table 6. If the percent recovery is not within these limits, matrix interference may be present in the selected sample. Under these circumstances, one of the following remedies must be employed: the matrix interference must be removed, all samples in the batch must be analyzed by a test method not affected by the matrix interference, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

12.7.5 The matrix spike/matrix spike duplicate (MS/MSD) limits in Table 6 are preliminary until more data is acquired. The data generated by a single-laboratory using sea water samples are in the Precision and Bias Section 16. The matrix variation between the different waters may have a tendency to generate significantly wider control limits than those generated by a single-laboratory in one water matrix. It is recommended that the laboratory generate their own in-house QC acceptance criteria which meet or exceed the criteria in this standard.

12.7.5.1 The laboratory should generate their own in-house QC acceptance criteria after the analysis of 15-20 matrix spike samples of a particular surface water matrix. References on how to generate QC acceptance criteria is in Method 8000B in EPA publication SW-846.

12.8 Duplicate:

12.8.1 To check the precision of sample analyses, analyze a sample in duplicate with each batch of 20 or fewer samples. If the sample contains the analyte at a level greater than 5 times the detection limit of the method, the sample and duplicate may be analyzed unspiked; otherwise, a MSD should be used.

12.8.2 Calculate the relative percent difference (RPD) between the duplicate values (or MS/MSD values) as shown in Eq 2. Compare to the RPD limit in Table 6.

$$RPD = |MSR - MSDR| / ((MSR + MSDR) / 2) \times 100 \quad (2)$$

where:

- RPD* = relative percent difference,
- MSR* = matrix spike recovery, and
- MSDR* = matrix spike duplicate recovery.

12.8.3 If the result exceeds the precision limit, the batch must be re-analyzed or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

13. Procedure

13.1 This test method is based upon a 20 mL sample size per analysis. Any sample size may be used such as a half filled VOA vial as long as the QC spikes and sample preparation volumes are adjusted accordingly. The samples must be analyzed within 5 days of collection. If the samples are above 6°C when received or during storage, or not analyzed within 5 days of collection, the data is qualified estimated and noted in the case narrative that accompanies the data.

13.2 In the laboratory, the entire 20 mL sample that was collected in a 20 mL VOA vial is poured into a 50 mL graduated cylinder. Every sample is spiked with the surrogate as described in Section 12. The laboratory control and matrix spike samples are then spiked with the target compounds as described in Section 12. The spiking solutions are added to the sample before transfer to the 50 mL graduated cylinder and that volume is subtracted from the measured amount. The exact volume of the sample size is recorded in order to calculate the exact final concentration of DPGBE, EGBE and surrogate. The vial is rinsed with two 2 mL portions of CH₃CN to remove any residual DPGBE, EGBE and surrogate adhered to the collection vial. These 2 portions are added to the 50 mL graduated cylinder. The samples are then diluted to 25 mL final volume with CH₃CN, shaken, filtered through the syringe driven filter unit fitted with a PVDF filter cartridge into glass storage vials and then aliquoted into 2 mL amber glass LC vials for analysis.

13.3 For samples that are homogeneous, the entire 25 mL volume is filtered through the filtration device described in 7.2 into a pre-cleaned collection vial (such as a VOA vial). A portion of that filtered sample is added to an amber glass LC vial with a Teflon® lined cap which is analyzed. A new filter unit is used for each sample.

13.4 For biphasic samples, a portion of the lower aqueous layer is filtered through the filtration device described in 7.2 into a pre-cleaned collection vial (such as a VOA vial). The upper oil layer is left behind and is not added to the filtration device. A portion of that filtered sample is added to an amber glass LC vial with a Teflon® lined cap which is analyzed. A new filter unit is used for each sample.

13.5 Samples may be encountered that have more than one phase. Those samples may be prepared for analysis using one or more of the following options:

13.5.1 For a 20 mL water sample, add 0.1 mL of 2.4 ppm *n*-NP2EO surrogate spike solution to the sample in the sample container. Cap the container and mix to ensure homogeneity. Transfer the contents of the sample container to a graduated cylinder and record the sample volume. Rinse the sample collection vial twice with 2 mL portions of acetonitrile, which is added to the prepared sample and dilute to 25 mL with acetonitrile to compose ~ 20% acetonitrile solution to ensure quantitative sample transfer. If different sample sizes are used, spiking solution and acetonitrile volume shall be adjusted proportionally. (Note: Since an accurate sample volume may not be known prior to measurement into the graduated cylinder the appropriate spike may be added directly to the graduate cylinder instead of the collection vial. The collection vial must be rinsed with acetonitrile which is added to the graduated cylinder and this prepared sample must be thoroughly mixed.) The entire sample should be filtered through the syringe driven filter unit described in 7.2. If the sample is biphasic, due to a top oil layer, a portion of the aqueous fraction (bottom) should be transferred to a glass syringe for filtration. Filter the sample using a 0.22 µm PVDF filter into a glass vial and then transfer to an autosampler vial. For samples that contain an oil layer, an aliquot of the filtered solution should be diluted with 80% water/20% acetonitrile for a preliminary analysis, a 100 fold dilution should be considered until the site samples are characterized. The concentration of the target analytes, especially DPGBE, may be very high in samples that contain an oil layer.

13.5.1.1 *n*-NP2EO has been shown to be absorbed into the oil layer yielding a non-detect as a result. If oil is present in the sample the recovery of the *n*-NP2EO surrogate may be very low or not detected at or above the reporting limit.

13.5.2 Water Subsample Analysis: Collect a 4 mL subsample of the water fraction (bottom) using a needle and a glass syringe. To reduce the oil exposure, invert the vial and tap gently to cause the oil move away from the septum. Insert needle through septum and collect 4 mL of the water layer; place the aliquot removed in a graduated cylinder. Then add 20 µL of 2.4 ppm *n*-NP2EO surrogate spike solution to the sample in the graduated cylinder. Add 1 mL of acetonitrile to the prepared sample to compose ~ 20% acetonitrile solution. If different sample sizes are used, surrogate spiking solution and

acetonitrile volume shall be adjusted proportionally. Filter the sample using a 0.22 µm PVDF filter into a glass vial and then transfer to an autosampler vial.

NOTE 1—Subsamples of water quantitates only the DPGBE and EGBE in the subsample, the DPGBE and EGBE concentration of subsamples may underestimate the sample DPGBE and EGBE concentration as a result of oil partitioning and surface binding.

13.6 The syringe must be cleaned between each filtration. It is the analyst's responsibility to ensure that the syringe is clean. A suggested method for cleaning the syringe between filtrations is to first rinse with at least 5 syringe volumes of water, followed by at least 3 volumes of 50% water/50% CH₃CN.

13.7 Once a passing calibration curve is generated the analysis of samples may begin. An order of analysis may be: method blank, laboratory control sample and duplicate, up to 20 samples, matrix spike sample(s) and duplicate followed by an end calibration check which includes a midpoint calibration check standard and a method blank.

14. Calculation or Interpretation of Results

14.1 For quantitative analysis of DPGBE, EGBE and surrogate, the SRM transitions are identified by comparison of retention times in the sample to those of the standards. External calibration curves are used to calculate the amounts of DPGBE, EGBE and surrogate. Calculate the concentration in µg/L (ppb) for each analyte. The sample concentration was diluted by 20% by the addition of surrogates and CH₃CN and target compound spike where applicable, this dilution must be accounted for when reporting the concentration. DPGBE and EGBE may be reported if present at or above the reporting limit. If the concentration of the analyte is determined to be above the calibration range, the sample is diluted with reagent water to obtain a concentration near the mid-point of the calibration range and re-analyzed. This method uses one surrogate, *n*-NP2EO, to monitor performance. The surrogate recoveries are provided with all data generated from this test method.

14.1.1 A surrogate is used to monitor the performance of DPGBE and EGBE. If the surrogate meets the quality control criteria in this test method, the data may be reported unqualified for DPGBE and EGBE if all other quality control in this test method are acceptable. If the surrogate does not meet the quality control criteria of the test method, the data is qualified for DPGBE and EGBE.

TABLE 7 Single-Laboratory Recovery Data in Reagent Water

Precision and Accuracy Samples	Analyte Measured (ppb)		
	DPGBE	EGBE	<i>n</i> -NP2EO
1	2.22	330	8.90
2	2.10	382	9.09
3	2.08	334	8.14
4	2.23	352	8.66
5	2.18	327	9.02
6	2.01	331	8.84
Spike Concentration (ppb)	2.40	300	12.00
Average Recovery	2.14	343	8.78
Average Percent Recovery	89.0	114	73.1
Standard Deviation	0.09	21.2	0.35
% Relative SD	4.09	6.19	3.93

15. Report

15.1 Determine the results in units of $\mu\text{g/L}$ (ppb) in a water sample. Calculate the concentration in the sample using the linear or quadratic calibration curve generated. All data that do not meet the specifications in the test method must be appropriately qualified.

16. Precision and Bias

16.1 Standard Test Methods under the jurisdiction of the ASTM committee D19 may be published for a maximum of five years to the completion of a full collaborative study validation. Such Standards are deemed to have met all other D19 qualifying requirements but have not completed the required validation studies to fully characterize the performance of the methods across multiple laboratories and matrices. Publication of standards that have been fully validated is done to make current technology accessible to users of Standards, and to solicit additional input from the user community. The determination of precision and bias was conducted through EPA and generated applicable data to determine the precision and bias as described in Practice **D2777**.

16.2 The determination of precision and bias was conducted through US EPA Region 5 Chicago Regional Laboratory (CRL).

16.3 This test method was tested by CRL on reagent water. The samples were spiked with the DPGBE, EGBE and *n*-NP2EO to obtain a 2.4 ppb concentration of DPGBE, 300 ppb EGBE and 12 ppb concentration of *n*-NP2EO each as described in Section 12. **Table 7** contains the recoveries and standard deviation (SD) for the target compounds and surrogate in reagent water.

16.4 This test method was tested by CRL on Gulf of Mexico sea water. The samples were spiked with target compounds and surrogate as described in Section 12. **Table 8** contains the recoveries and standard deviation (SD) for the target compounds and surrogate in sea water.

16.5 This test method was tested by CRL on Gulf of Mexico sea water containing crude oil. The samples were spiked with target compounds as described in Section 12. **Table 9** contains the recoveries for the target compounds in reagent water, whole sample and subsample.

17. Keywords

17.1 dipropylene glycol monobutyl ether (DPGBE); ethylene glycol monobutyl ether (EGBE); liquid chromatography; mass spectrometry; water

TABLE 8 Single-Laboratory Recovery Data in Gulf of Mexico Sea Water

Precision and Accuracy Samples	Analyte Measured (ppb)		
	DPGBE	EGBE	<i>n</i> -NP2EO
Method Blank Sea Water 1	ND	ND	7.83
Method Blank Sea Water 2	ND	ND	8.45
Sea Water 1	1.96	353	9.09
Sea Water 2	1.77	361	8.56
Sea Water 3	2.12	303	7.64
Sea Water 4	2.14	293	8.57
Sea Water 5	2.85	429	8.41
Spike Concentration (ppb)	2.4	300	12
Average Recovery	2.17	348	8.36
Average Percent Recovery	90.3	116	69.7
Standard Deviation	0.41	54.3	0.52
% Relative SD	18.9	15.6	6.3

TABLE 9 Single-Laboratory Recovery Data in Gulf of Mexico Sea Water Containing Oil

Sample	Reagent Water (Whole Sample)		Gulf Sea Water with 1% Crude Oil (Whole Sample)		Gulf Sea Water with 1% Crude Oil (Water Subsample)	
	Analyte Measured (ppb)		Analyte Measured (ppb)		Analyte Measured (ppb)	
	DPGBE	EGBE	DPGBE	EGBE	DPGBE	EGBE
Method Blank	ND	ND	ND	ND	Not Analyzed	Not Analyzed
Sample 1	2.04	284	1.34	240	1.20	263
Sample 2	2.04	284	1.39	239	1.18	272
Sample 3	Not Analyzed	Not Analyzed	1.35	251	1.32	239
Spike Concentration (ppb)	2.4	300	2.4	300	2.4	300
Average Recovery	2.04	284.0	1.36	243.3	1.23	258.0
Average Percent Recovery	85.0	94.7	56.7	81.1	51.4	86.0

Target compound spike was added to the entire whole sample before subsampling.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; http://www.copyright.com/