



Standard Test Method for Determination of Hydrocarbons and Non-Hydrocarbon Gases in Gaseous Mixtures by Gas Chromatography¹

This standard is issued under the fixed designation D7833; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method is intended to quantitatively determine the non-condensed hydrocarbon gases with carbon numbers from C₁ to C₅₊ and non-hydrocarbon gases, such as H₂, CO₂, O₂, N₂, and CO, in gaseous samples. This test method is a companion standard test method to Test Method [D1945](#) and Practice [D1946](#) differing in that it incorporates use of capillary columns instead of packed columns and allows other technological differences.

1.2 Hydrogen sulfide can be detected but may not be accurately determined by this procedure due to loss in sample containers or sample lines and possible reactions unless special precautions are taken.

1.3 Non-hydrocarbon gases have a lower detection limit in the concentration range of 0.03 to 100 mole percent using a thermal conductivity detector (TCD) and C₁ to C₆ hydrocarbons have a lower detection limit in the range of 0.005 to 100 mole percent using a flame ionization detector (FID); using a TCD may increase the lower detection limit to approximately 0.03 mole percent.

1.3.1 Hydrocarbon detection limits can be reduced with the use of pre-concentration techniques and/or cryogenic trapping.

1.4 This test method does not fully determine individual hydrocarbons heavier than benzene, which are grouped together as C₇₊. When detailed analysis is not required the compounds with carbon number greater than C₅ may be grouped as either C₆₊, or C₇₊. Accurate analysis of C₅₊ components depends on proper vaporization of these compounds during sampling at process unit sources as well as in the sample introduction into the analyzer in the laboratory.

1.5 Water vapor may interfere with the C₆₊ analysis if a TCD detector is used.

1.6 Helium and argon may interfere with the determination of hydrogen and oxygen respectively. Depending on the

analyzer used, pentenes, if present, may either be separated or grouped with the C₆₊ components.

1.7 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.8 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:²

[D1945 Test Method for Analysis of Natural Gas by Gas Chromatography](#)

[D1946 Practice for Analysis of Reformed Gas by Gas Chromatography](#)

[D3588 Practice for Calculating Heat Value, Compressibility Factor, and Relative Density of Gaseous Fuels](#)

[E355 Practice for Gas Chromatography Terms and Relationships](#)

[E1510 Practice for Installing Fused Silica Open Tubular Capillary Columns in Gas Chromatographs](#)

[F307 Practice for Sampling Pressurized Gas for Gas Analysis](#)

2.2 ASTM Publication:

[ASTM DS 4B, 1991 Physical Constants of Hydrocarbon and Non-Hydrocarbon Compounds](#)

3. Terminology

3.1 Terminology related to the practice of gas chromatography can be found in Practice [E355](#).

3.2 Definitions:

3.2.1 *sample set*—a collection of samples taken from the same source or at similar component composition and concentrations.

¹ This test method is under the jurisdiction of ASTM Committee [D03](#) on Gaseous Fuels and is the direct responsibility of Subcommittee [D03.07](#) on Analysis of Chemical Composition of Gaseous Fuels.

Current edition approved June 1, 2014. Published June 2014. Originally approved in 2012. Last previous edition approved in 2012 as D7833-12. DOI: 10.1520/D7833-14.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

4. Summary of Test Method

4.1 Components in a representative sample are physically separated by gas chromatography (GC) and compared to calibration data obtained under identical operating conditions from a reference standard mixture of known composition. The numerous heavy-end components of a sample can be grouped into irregular peaks by reversing the direction of the carrier gas through the column at such time as to group the heavy ends either as C₅ and heavier, C₆ and heavier, or C₇ and heavier or alternatively elute them in the non-backflushed mode and summed accordingly. The composition of the sample is calculated by comparing the peak areas with the corresponding values obtained with the reference standard.

5. Significance and Use

5.1 The hydrocarbon component distribution of gaseous mixtures is often required for end-use sale of this material. Applications such as chemical feedstock or fuel require precise compositional data to ensure uniform quality. Trace amounts of some hydrocarbon impurities in these materials can have adverse effects on their use and processing. Certain regulations may require use of such method.

5.2 The component distribution data of gaseous mixtures can be used to calculate physical properties such as relative density, vapor pressure, and heating value calculations found in Practice D3588. Precision and accuracy of compositional data is extremely important when this data is used to calculate various properties of petroleum products.

6. Apparatus

6.1 *Gas Chromatograph (GC)*—This method allows the use of most gas chromatographic analyzers designed for gas analysis. Generally, any gas chromatographic instrument with a linear temperature programmable column oven or adequate temperature control to provide the required separation of gaseous compounds being analyzed may be used. The temperature control must be capable of obtaining retention time repeatability within 5% of the retention time for each component throughout the scope of this analysis for hydrocarbon and non-hydrocarbon gas analyses.

6.1.1 *Detector*—The type and number of detectors employed is dependent on gas analyzer model and vendor used. Detectors that can be used include, but are not limited to FID, TCD, AED (Atomic Emission Detector), HID (Helium Ionization Detector), and MS (Mass Spectrometer). Many systems use a 3 detector system:

(1) One FID (Flame Ionization Detector) for the determination of the hydrocarbon gases for the compounds listed in Table 1,

(2) One TCD (Thermal Conductivity Detector) dedicated to the determination of hydrogen utilizing nitrogen or argon as a carrier gas, and

(3) One TCD for the determination of all other required non-hydrocarbon gases using helium as the carrier gas.

6.1.2 A TCD may also be used for the analysis of the hydrocarbon gases (replacing the FID) when high sensitivity (< 0.03 mole percent) for trace analysis is not required.

TABLE 1 List of Components Typically Analyzed

Component	FID	TCD
C ₅ olefin / C ₆ + composite	X	X
oxygen/argon composite		X
hydrogen		X
carbon dioxide		X
hydrogen sulfide		X
nitrogen		X
carbon monoxide		X
methane	X	X
ethane	X	X
ethylene	X	X
propane	X	X
propylene	X	X
acetylene	X	X
isobutane	X	X
propadiene	X	X
n-butane	X	X
trans-2-butene	X	X
1-butene	X	X
isobutylene	X	X
cis-2-butene	X	X
neopentane	X	X
cyclopentane	X	X
isopentane	X	X
methyl acetylene	X	X
n-pentane	X	X
1,3-butadiene	X	X

6.1.3 Other detectors or combination of detectors may be used provided that they have sufficient response, linearity, and sensitivity to measure the components of interest at the concentration levels required for this application and meeting all of the quality controls specified in this method. Some analyzers, such as micro-analyzers, may contain up to 4-channels and separation systems to accomplish the analysis described in this method.

6.2 *Data Acquisition*—Any commercial computerized data acquisition system may be used for display of the chromatographic detector signal and peak area integration from all of the detectors used in the analysis. The device should be capable of generating and storing a calibration and reporting the final corrected response factor results.

6.3 *Sample Introduction and System Configurations*—Sample introduction is typically performed with automated valves containing sampling ‘loops’ of appropriate sizes. Fig. 1 gives a suggested configuration, although systems may vary slightly among gas analyzers. The combination of valve injection size and/or splitting inlet ratio must be selected such that the required sensitivity for the application is achieved and also that no component concentration in a sample is greater than the detector upper linearity limit. The sample inlet system shall be constructed of materials that are inert and non-adsorptive with respect to the components in the sample. The preferred material of construction is stainless steel. Copper, brass, and other copper-bearing alloys are unacceptable. The sample size limitation of 0.5 mL or smaller is selected relative to the linearity of the detector response, and efficiency of column separation. Larger samples may be used to determine low-quantity components to increase measurement accuracy. Sample sizes may be determined by experimentation or as recommended by analyzer vendors.




FIG. 1 Example of a Three Detector System for Analysis of Hydrocarbons and Non-Hydrocarbon Gases





FIG. 2 Example Chromatogram of Non-Hydrocarbon and Light Hydrocarbon Gases from System Configuration in Fig. 1



NOTE 1—For the hydrocarbon analysis, the Al_2O_3 PLOT was used.

FIG. 3 Example Chromatogram of Hydrocarbons from System Configuration in Fig. 1

6.3.1 Hydrogen Sulfide and Other Reactive Gases—Samples containing hydrogen sulfide and trace reactive gases may benefit from use of surface treated metal surfaces, such as Silcosteel³ or Sulfinert⁴ processes. Such specially treated surfaces are also recommended for sample containers that may contain such reactive species.

6.3.2 With Capillary Columns—The gas chromatograph must include a heated splitting type inlet that is operated isothermally, or if appropriate, direct connection to the valve may be possible as long as sample sizes are adjusted accordingly, the calibrations are linear in the range of interest, and the required resolution of the compounds of interest is maintained. When using a split injection, split ratios in the range of 5:1 to 200:1, with a typical value of 100:1, have been used successfully depending upon the sample injection volume and sensitivity required.

6.3.3 With Pre-concentrator and/or Cryogenic Trapping—Pre-concentrator and/or cryogenic trapping can be used prior to sample introduction into the gas chromatograph. These items may enable lower detection limits on the components detailed by the manufacturer to be concentrated.

³ Silcosteel is a trademarked of SilcoTek, 112 Benner Circle, Bellefonte, PA 16823.

⁴ Sulfinert is a trademarked of Restek Corporation 110 Benner Circle Bellefonte, PA 16823.

6.4 Hydrogen Gas Analysis (Thermal Conductivity Detector)—A 10-port gas-sampling valve or equivalent may be used with nitrogen or argon carrier gas. Nitrogen or argon carrier gas is used to ensure that the hydrogen ‘peak’ remains positive over the concentration range of interest. Any column or multiple columns may be used, as long as helium and hydrogen are separated and also separated from the other components. Typically, a dedicated TCD is used for this analysis. The gas-sampling valve shall provide a repeatability of at least $\pm 2\%$ relative to the sample volume introduction for major compounds present at $>5\%$ vol%.

NOTE 1—When helium is not expected to be present in samples the resolution of hydrogen from helium is not critical.

6.5 Non-Hydrocarbon and Light Hydrocarbon Gas Analysis (Except Hydrogen) (Thermal Conductivity Detector)—A 10-port gas sampling valve in combination with a 6-port switching valve or equivalent is used with helium or hydrogen carrier to analyze for CO_2 , O_2 , N_2 , CH_4 , C_2H_6 , and CO and in some cases H_2S . Any column or multiple columns may be used as long as the desired components are well separated. A TCD may also be used for the analysis of the hydrocarbon gases (replacing the FID) when high sensitivity (<300 ppm) for trace analysis is not required. The gas-sampling valve shall provide a repeatability of at least $\pm 2\%$ relative to the sample volume introduction for major compounds present at $>5\%$ vol%.

6.6 Hydrocarbon Gas Analysis (Flame Ionization Detector)—A 6-port gas-sampling valve in combination with a 6-port pre-column switching valve (backflush) for the C₆+ or C₇+ hydrocarbons is typically used. These valves shall be contained in a heated enclosure and operated at a sufficiently high temperature, and within the limits of the valve operating temperature as specified by manufacturer, to prevent condensation of the C₆+ components in the sample. The use of a frit or packed-screen type filter ahead of the sample introduction port is recommended with use of PLOT columns. The gas-sampling valve shall provide a repeatability of at least $\pm 2\%$ relative to the sample volume introduction for major compounds present at $>5 \text{ vol}\%$.

6.7 Column Series/Reversal Switching Valve—If desired, a multi-port valve may be used to provide the C₅ olefin/C₆+ or C₇+ determination for this analysis. Other switching valve configurations may be used to allow the elution of the gaseous compounds. Consult instrument manufacturer for optimum configuration.

NOTE 2—If a dimethylsilicone capillary column or equivalent is used for the hydrocarbon analysis, then the capillary column may be used in the foreflush mode (no-backflush) until all of the hydrocarbons have eluted using temperature programming or equivalent.

6.8 Gas Controls—The gas chromatograph shall be provided with suitable facilities for delivery and control of carrier gases and detector gases. This will consist of the appropriate gas supply, down-stream regulators, and supply tubing as well as the mass or pressure controls for the precise regulation of the instrument operation.

NOTE 3—Most gas chromatograph suppliers will provide these devices or recommend the proper suppliers. Ensure that the analyzer when heated and in-use does not run out of carrier gases. In addition, running out of carrier gas will require flushing out any air introduced into the sample inlet system, column and/or detector.

6.9 Columns—Condition all columns used according to the manufacturer's suggestions prior to putting the system in service.

6.10 Analytical Column for Hydrocarbon Analysis—A recommended analytical column for the hydrocarbon analysis in **Fig. 1** is a 50 m \times 0.53 mm (I.D.) deactivated alumina (Al₂O₃) porous layer open tubular (PLOT) column used with a FID detector for lowest detection limits. Relative retention order for the alumina PLOT column is dependent upon the deactivation method for the column and moisture content. **Warning**—Specifically test the alumina PLOT column to ensure that the column does not adsorb propadiene, methyl acetylene, and butadiene when such compounds need to be determined. This condition can exist depending upon the degree of column deactivation.

6.10.1 Routine re-conditioning of the alumina PLOT column may be required to maintain column performance. It is recommended that a standby method be used when the system is idle to maintain the PLOT column at a temperature of at least 130°C or as recommended by the manufacturer.

6.10.2 Alternatively, any column or combination of columns that provides the appropriate component C₁-C₅ separations may be used.

6.11 Pre-column for Hydrocarbon Gas Analysis—When using the alumina PLOT column, if an initial backflush of the

C₅+/C₆+ components through the use of the sequence reversal/backflush valve is desired, a second column is required. Any pre-column that provides separation between the components of interest and the composite heavier components may be used. Choices may include lengths of column such as a 10 to 30 m section of 0.53 mm (I.D.) 3- μm film thickness dimethyl polysiloxane or a 9 to 15 cm section of the same column material as the analytical column or any pre-column that provides the desired retention of pentenes, hexanes, and heavier components. This pre-column acts to keep the heavier components away from the analytical alumina PLOT column and to backflush the heavier components as a composite peak to the detector for quantification. If analysis of individual C₆-C₇ components is required, extend the backflush valve time until the desired components have eluted and prior to backflushing the remaining heavier compounds.

6.12 Analytical Columns for Hydrogen Analysis—Generally hydrogen analysis consists of a pre-column to remove most of the hydrocarbons, H₂S and CO₂ and a Molecular Sieve 5A or equivalent for separation of hydrogen from oxygen and nitrogen. Follow vendor's recommendations.

6.13 Analytical Columns for Other Non-Hydrocarbon Gases—Generally a series-bypass two-valve configuration is used, consisting of porous polymer-molecular sieve 5A or 13X combination. Follow vendor's recommendations.

7. Reagents and Materials

7.1 All chemicals are reagent grade unless specified otherwise, and all water used is distilled or deionized. **Warning**—Hydrogen sulfide contained in calibration standards may be flammable and harmful or fatal if ingested or inhaled. Calibration standards or samples containing hydrogen sulfide should be handled in well ventilated locations away from sparks and flames.

7.2 Carrier Gases—For carrier gases, it is strongly recommended to install commercial active oxygen scrubbers and water dryers, such as molecular sieves, ahead of the instrument to protect the chromatographic columns. Follow supplier instructions in the use of such gas purifiers and replace as necessary.

7.2.1 Chromatographic Grade Hydrogen, 99.995% minimum purity, $<0.1 \text{ ppm H}_2\text{O}$. The use of appropriate scrubbers may be sufficient to obtain the desired purity.

7.2.2 Chromatographic Grade Helium, 99.995 % minimum purity, $<0.1 \text{ ppm H}_2\text{O}$. The use of appropriate scrubbers may be sufficient to obtain the desired purity.

7.2.3 Chromatographic Grade Nitrogen, 99.995 % minimum purity, $<0.1 \text{ ppm H}_2\text{O}$. The use of appropriate scrubbers may be sufficient to obtain the desired purity. **Warning**—Improper handling of compressed gas cylinders containing air, nitrogen, hydrogen, or helium can result in an explosion. Rapid release of nitrogen or helium can result in asphyxiation.

7.3 FID Detector Gases:

7.3.1 Chromatographic Grade Hydrogen, 99.995 % minimum purity. The use of appropriate scrubbers may be sufficient to obtain the desired purity.

7.3.2 Chromatographic Grade Air, less than 10 ppm-mol each of total hydrocarbons and water. The use of appropriate scrubbers may be sufficient to obtain the desired purity. **Warning**—Compressed air supports combustion.

7.4 Reference Standards:

7.4.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, all reagents should conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available. Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

7.4.2 *Calibration Standards*—Individual and mixed component reference materials are commercially available and may be used to establish qualitative and quantitative calibration. The calibration standard mixture should be gravimetrically prepared, supplied with both gravimetric and calculated volumetric and mole% concentrations, and at least NIST traceable or Van Swinden Laboratory, National Metrology Institute of the Netherlands (VSL) certified. The calibration standard mixture should have analytical uncertainty of <1% relative.

7.4.3 *Secondary Reference Standards*—A mixture or mixtures of known composition that is (or are) independent of the calibration standards and similar wherever possible in concentration to the samples being analyzed. The secondary reference standards are used as check standards to monitor testing precision and accuracy.

8. Preparation of Apparatus

8.1 Install the valves for the hydrogen, other non-hydrocarbon gases analysis, and/or the hydrocarbon analysis. If a turn-key analyzer is used follow manufacturer's instructions. Fig. 1 gives an example of a three-detector system using the Al₂O₃ column for the hydrocarbons, and individual analyses for hydrogen and other non-hydrocarbon gases.

8.2 If using the alumina PLOT column for the hydrocarbon analysis, it is recommended that the PLOT column undergo routine baking at the maximum method temperature in order to maintain conditioning. See Practice E1510 for recommended installation procedures. If multiple columns for this analysis are placed into the same oven with the alumina PLOT, ensure that the bake out temperature does not exceed the maximum allowable temperature of any of the other columns.

8.3 Set the GC instrument to the operating parameters. Allow the instrument to stabilize before proceeding with calibration and sample injections. When using an alumina PLOT column, recommended operating conditions are listed in Appendix X1. However, conditions may vary if other column systems are used or if tubing is different.

8.4 *Gas Sampling Valves*—Set valve on and off times to comply with manufacturer's instructions. When using a three detector system, sample introduction into the analyzer for hydrogen, non-hydrocarbon gases, and hydrocarbons may be accomplished using valve injections that are synchronized such that samples are injected simultaneously into the system using computer controlled valves. Other sample introduction may be

used as long as the repeatability for these systems meets or exceeds those stated in this method. Calibration standards are injected under similar temperature and pressures as the samples to be analyzed. Generally the sampling pressure is approximately atmospheric. The temperature of sampling is such as to avoid condensation of the compounds to be analyzed. For hexanes and higher, heat the sample loop. Most modern chromatographs have valve ovens that can be temperature controlled. It is strongly recommended in the absence of valve ovens to mount the gas-sampling valve in the chromatograph oven and operate at the column temperature. Sub-ambient sampling can be performed if a pressure transducer is used to measure the sampling line pressure and make the appropriate correction.

8.5 *Switching (backflush) Valve for Alumina PLOT Hydrocarbon Analysis (optional)*—The valve rests in the "OFF" state, allowing a continuous backflush flow through the pre-column. Before or upon injection of the sample, the valve should be rotated to the "ON" position so that the pre-column is placed at the head of the flow path from the sample valve (see Fig. 1). At a time which must be empirically determined and which is dependant upon the length and type of pre-column used, the valve must be returned to the "OFF" position, causing the flow to backflush through the pre-column and flush to the detector ahead of components eluting from the analytical column (see Fig. 1). The switch time of the valve should be determined such that the areas of 2-methyl butane, pentane, and 1,3-butadiene are not diminished, but so that pentene components are not allowed to elute to and through the analytical column (where they would elute after the pentanes and 1,3-butadiene). Determining this switch time may require iterative attempts and interpolation. However, once the time has been determined, it should remain repeatable for all samples of similar composition. For other system configurations consult the analyzer vendor.

9. Calibration and Standardization

9.1 *Qualitative*—After the analyzer is installed, at least annually, after major repair, or after extended shutdown, determine the retention times of components by analyzing known reference mixtures in the same manner as samples (Section 11). Table 1 gives a summary of components commonly determined according to this standard using TCD or FID detection. The stated elution order of the hydrocarbons is that for an alumina PLOT column. In general, elution order will vary by column used and the configuration of the analyzer. Note that some PLOT columns do not separate neopentane from butenes.

9.2 *Verification of Linearity*—Ensure that the response of each component of interest is linear bracketing the expected concentration of analytes. Use at least three concentration ranges for each component of interest. This is typically accomplished using a multi-component gas standard(s) containing analytes of interest analyzed with and without dilution or by using gas standards with components at each of the concentrations needed to prepare a calibration curve. Ensure that the linearity regression coefficient is at least 0.999 for each compound calibrated. Special care should be used to verify that

low concentration values do not deviate significantly from the calibration curve. If low concentration points do not conform to linearity as defined by higher concentration points, it is necessary to prepare a second calibration curve for low concentration analytes when samples contain components at levels approaching the low level of the initial curve are expected. The linearity is confirmed upon analyzer start-up and whenever major changes are made to the analyzer, such as changing columns, replacing detectors etc. Ensure the TCD analyzing hydrogen exhibits a positive response (peak area) over the range of interest.

9.2.1 *Introduction of Standards*—Open the outlet valve of the sample cylinder and purge the sample through the inlet system and sample loop or tube. The amount of purging required must be established and verified for each instrument.

9.2.1.1 When purging the valve, sample loops, and the general inlet system with calibration standard do not expose the calibration standards to high flows as this may change the composition of the standard over time. Initiating flows of approximately 10–30 mL/min of gas standard should be adequate. Use appropriate regulators on standard cylinders and slowly increase the flow from a zero flow. Once the correct flow has been established the cylinder regulator may be set to the required setting. The cylinder may be isolated from the system by using the shut-off valve when not in use. The sample loop pressure should be near atmospheric during injection. Close the cylinder valve and allow the pressure of the sample in the loop or tube to stabilize. Then immediately inject the contents of the loop or tube into the gas chromatograph. Such gauges may be used to adjust the pressure to close to atmospheric pressure. Flow meters and bubble meters may also be used to monitor and set the flow standards and verify pressures prior to injection. If injecting sub-ambient pressure samples, the proper equipment must be used to ensure the sample has been properly introduced into the gas chromatograph.

9.2.1.2 Standards containing hydrogen sulfide should use corrosive resistant regulators. Vents from gas sampling valves need to be vented to a well ventilated hood to avoid exposure to hydrogen sulfide. Some systems may contain a pressure gauge to measure the actual pressure in the loop.

9.2.2 If it is not possible to obtain information on the linearity of the available gas chromatograph detector for all of the test gas components, then as a minimum the linearity data must be obtained for any gas component that exceeds a concentration of 5 mol%. GC detectors have a lower detection limit and are not truly linear over wide concentration ranges. Linearity should be established over the range of interest and confirmed annually.

9.3 *Calibration and Primary Standardization*—Once linearity has been established a single standard may be used to perform a recalibration of the analyzer as necessary. Recalibration should be performed when hardware is changed, there is a failure of the QC check, or as needed to meet raw recovery requirements. Determine the experimental response factors of components detected by the detectors used by analyzing known calibration mixtures under the same conditions of pressure and

temperature as the samples (Section 11). For each component present in the calibration standard, calculate the response factor according to Eq 1.

$$RF_i = C_i/A_i \quad (1)$$

where:

RF_i = the response factor for component i,

C_i = the known concentration of i, and

A_i = the integrated area of peak i.

9.3.1 If desired for C_6 and C_7+ hydrocarbons (if split) the RF's can be estimated from isopentane and n-pentane as follows:

$$RF_{C_6} = RF_{C_{5AV}} \times 72/86 \quad (2)$$

$$RF_{C_7} = RF_{C_{5AV}} \times 72/100 \quad (3)$$

where:

$RF_{C_{5AV}}$ = average RFs of i- C_5 and n- C_5 .

9.3.2 In cases where the C_6+ are backflushed as one composite and the average molecular weight of the composite, if desired the C_6+ RF can be estimated as follows:

$$RF_{C_6+} = RF_{C_{5AV}} \times 72/93 \quad (4)$$

9.3.3 The standard should contain all of the components typically observed in the samples.

NOTE 4—Hydrocarbons with a boiling point greater than isopentane in the standard have a potential for condensation. Consult with gas standard manufacturer for proper sizing and pressure of standard cylinder to minimize condensation. In some cases where a FID is used, an adjacent eluting hydrocarbon of the same carbon number for which a calibration standard is not available may be estimated from a hydrocarbon that is present in the standard, e.g., isobutane may be estimated from the RF of n-butane, 1-butene, or isobutylene.

9.4 *Quality Monitoring*—After determining the response factor for each component using the primary standard(s), analyze a secondary check standard(s) that approximates composition of samples being analyzed (see 7.4.3) every 24 hrs or at a frequency determined through a statistical TPI assessment when samples are analyzed. Verify that the concentrations agree within the acceptable values for the standard(s). If the test method is used only occasionally, reanalyze the calibration standard(s) and secondary standard before each set of analyses. Results should agree within 2% relative of the certified value for components present at >5 vol%. Failure to compare may result from lack of injection splitter (if used) linearity or use of a standard that has not been maintained according to the standard manufacturer's recommendations. It is necessary to compare calculated results to the certified values for a known standard before accepting the calibration.

10. Sampling

10.1 Sampling at the sample source, the use of appropriate sample containers that are stored and transported properly, and introduction into a chromatograph must be done in a manner that ensures that a representative sample is being analyzed. See Practice F307 for the recommended procedures. Lack of precision and accuracy in using this method can most often be attributed to improper sampling, sample containment, or sample introduction procedures.

10.2 Higher boiling components such as C₅+ may condense during sampling of the process unit if the sample vessel is not at the same or greater temperature as the sample stream. This will result in an inaccurate collection of sample components. A representative sample must be collected to obtain valid results. Please reference API MPMS 14.1 for additional information.

10.3 It is recommended that samples be equilibrated in the laboratory at –6.7 to 10°C (20 to 50°F) above the source temperature of the field sampling. If the hydrocarbon dew point of the sample is known to be lower than the lowest temperature to which the sample has been exposed, it is not necessary to heat the sample.

10.4 Connections from the sample container to the sample inlet of the instrument should be made with stainless steel or with short pieces of TFE-fluorocarbon or inert tubing coated with Silcosteel³ or Sulfinert⁴. Copper, vinyl, or rubber connections are not acceptable. Heated lines may be necessary for samples containing C₆+ hydrocarbon content samples.

10.5 Samples should be clean and free of liquid prior to injection. It is recommended that filters be placed between the sampling device and the analytical equipment to minimize the risk of damage to the analytical equipment.

10.5.1 If the sample is known to have excessive liquid or particulates, the sampling device may need to be tilted or stood in an upright position to place the contamination away from the sample being drawn from the sampling device. Other preparatory techniques may need to be employed to reduce the potential for damaging the analytical equipment.

11. Procedure

11.1 The sample cylinder should be oriented vertically with the valve to be used for sampling located at the top. Open the outlet valve of the sample cylinder and purge the sample through the inlet system and sample loop or tube at flows <50 mL/min. The amount of purging required must be established and verified for each instrument. The sample loop pressure should be near atmospheric unless sub-ambient sampling is performed.

NOTE 5—Vents from gas sampling valves need to be directed to a well-ventilated hood to avoid exposure to hydrogen sulfide.

11.2 Close the cylinder valve and allow the pressure of the sample in the loop or tube to stabilize. Immediately inject the contents of the loop or tube into the chromatograph. Some systems may contain a pressure gauge to measure the actual pressure in the loop. Such gauges may be used to adjust pressure to close to atmospheric or to measure the injection pressure if your sampling system requires it. In addition to pressure gauges, gas bubblers and flow meters may also be used.

12. Calculations

12.1 *External Standard Calibration Calculation*—Calculate the concentration of each component according to Eq 5. Determine the total amount of all components by summing the component concentrations.

$$SC_i = RF_i \times SA_i \quad (5)$$

where:

SC_i = concentration of component i in the sample,

RF_i = response factor for component i, and

SA_i = integrated area for peak i.

12.2 The summation of all components from Eq 5 must agree within 100 ± 5 %.

12.3 Normalize the final results to 100%.

13. Reporting

13.1 Report the concentration of each component as mole percent to the nearest 0.01%.

13.2 Report the C₆+ (C₅ olefins+C₆+ components) as “C₆+ hydrocarbons.” If the C₅ olefins are resolved they may be reported separately.

13.3 For systems that are set-up to calculate a split C₆ and C₇+ hydrocarbons report them individually as separate groups as “C₆+ hydrocarbons” and “C₇+ hydrocarbons.”

14. Quality Control

14.1 Confirm the performance of the GC analysis and test procedure by analyzing a secondary reference standard(s) of known composition (see 7.4.3) according to the frequencies identified in 9.4. The secondary reference standard(s) are used as check standards to monitor testing precision and accuracy.

14.2 In addition, the analysis of other quality control (QC) samples from plant production that are representative of the samples typically analyzed is recommended. An ample supply of such QC sample material should be available and stored without undergoing composition changes for the intended period of use. It is recommended that a QC sample be analyzed with each batch of samples to ensure data quality relative to the statistical control limits established.

14.3 *Blank Analysis*—Confirmation of lack of carry-over or contamination is recommended and may be required for certain applications. This is accomplished through analysis of ultra pure hydrogen, helium, methane blank, nitrogen, or zero air blank as appropriate. Blank analysis results may be subtracted from sample results as necessary.

14.4 *Spiked Samples*—For some applications it may be necessary to analyze spiked sample for each sample set as part of a QA/QC program. Spikes are prepared by quantitative addition of a mixed component calibration gas to a known volume of sample gas. Acceptable recoveries for components present at greater than 5 ppmv should fall within 10% of the theoretical amounts to verify nominal system performance. Unacceptable recoveries indicate matrix interference or system malfunction.

14.5 *Calibration Standard Reanalysis*—For some applications a standard is reanalyzed after each sample set to assess instrument drift. All components of interest should be within 2% of the theoretical amounts based on the original standards.

14.6 It is recommended that a duplicate analysis be performed every tenth sample or one for each sample set. All components of interest should be within 2% of the averaged concentration of each component.

15. Precision and Bias

15.1 *Repeatability*—The difference between successive test results obtained by the same operator with the same apparatus under constant operating conditions on identical test materials.

15.1.1 The difference between two successive results obtained by the same operator with the same apparatus under constant operating conditions on identical test materials should be considered suspect if they differ by more than the following amounts:

Component Concentration Range (mol %)	Repeatability (mol %)
0 to 0.1	0.01
0.1 to 1.0	0.04
1.0 to 5.0	0.09
5.0 to 10	0.10
10 to 50	0.40
50 to 100	0.50

15.2 *Reproducibility*—The difference between two single and independent results obtained by different operators working in different laboratories on identical test materials.

15.2.1 Reproducibility data to be added within 5 years of method approval.

15.3 *Bias*—A statement of bias will be developed through inter-laboratory testing by the responsible study group.

16. Keywords

16.1 gas analysis; gas chromatography; gas sampling valve; hydrocarbons; non-hydrocarbon gases

APPENDIX

(Nonmandatory Information)

X1. RECOMMENDED OPERATING CONDITIONS

Typical Operating Conditions for Hydrocarbon Analysis	
Column dimensions	50 meter × 0.53 mm Al ₂ O ₃
Backflush	Yes
Initial temperature	75°C (167°F)
Initial hold time	3 min
Program rate	10°C/min (18°F/min)
Final temperature	180°C
Final hold time	2 min
Injector temperature	225°C
Sample size	200 µL
Split ratio	100:1
Typical detector	FID
Detector temperature	250°C (482°F)
Fuel gas hydrogen flow	35–45 mL/min
Oxidizing gas air flow	400–450 mL/min
Make-up gas type	He or N ₂
Make-up gas flow	35 mL/min
Capillary column carrier gas type	Constant or programmed flow or pressure helium or hydrogen at 5 mL/min
Capillary column average linear velocity	45 cm/sec
Data rate	5–10 Hz
Packed pre-column flow rate for C ₅ + or C ₆ + heart-cutting away from Al ₂ O ₃ column	30–40 mL/min
Packed pre-column temperature for C ₅ + or C ₆ + heart-cutting away from Al ₂ O ₃ column	Approximately 90°C (194°F). Pre-column usually installed in an auxiliary secondary oven
Typical Operating Conditions for Non-hydrocarbon Gas Analysis	
Backflush	Yes
Isothermal temperature	75°C (167°F)
Typical detector	TCD
Detector temperature	250°C (482°F)
Make-up gas type	He or H ₂
Column carrier gas type	Constant or programmed flow or pressure helium or hydrogen
Data rate	5–10 Hz
Typical Operating Conditions for Hydrogen Analysis	
Backflush	Yes
Isothermal temperature	75°C (167°F)
Typical detector	TCD
Detector temperature	250°C (482°F)
Column carrier gas type	Constant or programmed flow or pressure nitrogen or argon
Data rate	5–10 Hz

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; <http://www.copyright.com/>