



Standard Test Method for Initial pH (i-pH)-Value of Petroleum Products¹

This standard is issued under the fixed designation D7946; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers procedures for the determination of initial pH (i-pH) in new and in-service lubricants.

NOTE 1—In new and used oils, the constituents that may be considered to have characteristics influencing the i-pH value include organic and inorganic acids, esters, phenolic compounds, lactones, resins, salts of heavy metals, salts of ammonia and other weak bases, acid salts of polybasic acids, and addition agents such as inhibitors and detergents. “Initial” is used to differentiate from aqueous systems. The analysis is terminated after a defined time interval whenever equilibrium conditions, as known for pH measurements in aqueous systems, are not reached (see 3.1.1.2)

1.2 This test method is used to indicate relative changes that occur in oil during use under oxidizing conditions or due to contamination by blow-by gases of combustion processes of biogases regardless of the color or other properties of the in-service lubricants. Although the initial pH is made under definite equilibrium conditions, the test method is not intended to measure an absolute acidic property that can be used to predict performance of oil under service conditions. No general correlation between corrosion of non-ferrous bearing metals and initial pH value is known.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:²

D1193 Specification for Reagent Water

D4057 Practice for Manual Sampling of Petroleum and Petroleum Products

¹ This test method is under the jurisdiction of ASTM Committee D02 on Petroleum Products, Liquid Fuels, and Lubricants and is the direct responsibility of Subcommittee D02.06 on Analysis of Lubricants.

Current edition approved July 1, 2014. Published August 2014. DOI: 10.1520/D7946-14.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

D4177 Practice for Automatic Sampling of Petroleum and Petroleum Products

3. Terminology

3.1 Definitions:

3.1.1 *initial pH (i-pH), n*—the measure of the hydrogen ion (H⁺) concentration when a fixed sample amount of lubrication oil is dissolved into a fixed volume of solvent and using a specified detection system.

3.1.1.1 *Discussion*—This test method expresses the amount of dissociated acidic components as i-pH of a sample in a mixture of toluene and propan-2-ol to which a small amount of water has been added.

3.1.1.2 *Discussion*—The measured i-pH is called initial since the original definition of the pH-scale depends on aqueous solutions. Its application to organic solutions will not give stable electrode potential for all types of samples. A potential reading after a defined time is used in those cases instead.

4. Summary of Test Method

4.1 The oil sample is dissolved in solvent mixture and the i-pH value is measured using a glass indicating electrode and a reference electrode or a combination electrode. The meter readings are measured manually or automatically.

5. Significance and Use

5.1 New and used petroleum products may contain acidic constituents that are present as additives, degradation products formed during service, such as oxidation products or components formed from combustion gases. The i-pH-value is a measure of the amount of acidic substances and their acidity defined under the conditions of test. The i-pH-value is used as a measure of lubricant degradation in service.

5.2 The corrosiveness of acidic components depends on their concentration and acidity. The i-pH-value is a measure of the amount of dissociated acidic components with the potential of corrosiveness towards metals.

5.3 Since a variety of oxidation and blow-by products contribute to the i-pH-value, this test method cannot be used to predict corrosiveness of oil or blends under service conditions against metallic components. No general correlation is known between i-pH-value and the corrosive tendency of blends or oils toward metals.

6. Apparatus

6.1 *pH-Meter*—Meter, a voltmeter or a potentiometer, that will operate with an accuracy of ± 0.005 V and a sensitivity of ± 0.002 V over a range of at least ± 0.5 V when the meter is used with the electrodes specified in 6.2 to 6.4 and when the resistance between the electrodes falls within the range from 0.2 M Ω to 20 M Ω . The meter shall be protected from stray electrostatic fields so that no permanent change in the meter readings over the entire operating range is produced by touching, with a grounded lead, any part of the exposed surface of the glass electrode, the glass electrode lead, the stirrer, or the meter.

NOTE 2—A suitable apparatus could consist of a continuous-reading electronic voltmeter designed to operate on an input of less than 5×10^{-12} A, when an electrode system having a resistance of 1000-M Ω is connected across the meter terminals and provided with a metal shield connected to the ground, as well as a satisfactory terminal to connect the shielded connection wire from the glass electrode to the meter without interference from any external electrostatic field.

6.2 *Sensing Electrode*, Standard pH, suitable for nonaqueous media.

6.3 *Reference Electrode*, Silver/Silver Chloride (Ag/AgCl) Reference Electrode, filled with 1 mol/L to 3 mol/L LiCl in ethanol.

6.4 *Combination Electrodes*—Sensing electrodes may have the Ag/AgCl reference electrode built into the same electrode body, which offers the convenience of working with and maintaining only one electrode. The combination electrode shall have a sleeve junction on the reference compartment and shall use an inert ethanol electrolyte, for example, 1 mol/L to 3 mol/L LiCl in ethanol. These combination electrodes shall have the same response or better response than a dual electrode system. They shall have movable sleeves for easy rinsing and addition of electrolyte.

6.5 *Variable-Speed Mechanical Stirrer*, a suitable type, equipped with a propeller-type stirring paddle. The rate of stirring shall be sufficient to produce vigorous agitation without splattering and without stirring air into the solution. A propeller with blades 6 mm in radius and set at a pitch of 30° to 45° is satisfactory. A magnetic stirrer is also satisfactory. If an electrical stirring apparatus is used, it shall be electrically correct and grounded so that connecting or disconnecting the power to the motor will not produce a permanent change in the meter reading during the course of the determination.

6.6 *Beaker*, 250 mL capacity, made of borosilicate glass or other suitable material.

6.7 *Beaker*, 100 mL, or suitable capacity, made of borosilicate glass or other suitable material.

6.8 *Stand*, suitable for supporting the electrodes, and stirrer.

NOTE 3—An arrangement that allows the removal of the beaker without disturbing the electrodes and stirrer is desirable.

6.9 *Graduated Cylinder*, 100 mL, or dispensing device capable of delivering 70 mL ± 0.5 mL.

6.10 *Pipette*, 2.0 mL, Class A.

7. Reagents

7.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that

all reagents shall conform to the specifications of the committee on Analytical Reagents of the American Chemical Society, where such specifications are available.³ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

7.2 Commercially available solutions may be used in place of laboratory preparations provided the solutions have been certified as being equivalent.

7.3 Alternative volumes of the solutions may be prepared, provided the final solution concentration is equivalent.

7.4 *Purity of Water*—Unless otherwise indicated, reference to water shall be understood to mean reagent water that meets the requirements of either Type I, II, or III of Specification **D1193**.

7.5 *Ethanol*, (**Warning**—Flammable and toxic, especially when denatured.)

7.6 *Lithium Chloride*, LiCl.

7.7 *Lithium Chloride Electrolyte*—Prepare a 1 mol/L to 3 mol/L solution of lithium chloride (LiCl) in ethanol.

7.8 *Propan-2-ol, Anhydrous*, (less than 0.1 % H₂O). (**Warning**—Flammable.) If adequately dry reagent cannot be procured, it can be dried by distillation through a multiple plate column, discarding the first 5 % of material distilling overhead and using the 95 % remaining. Drying can also be accomplished using molecular sieves such as Linde Type 4A, by passing the solvent upward through a molecular sieve column using one part of molecular sieve per ten parts of solvent.

NOTE 4—It has been reported that, if not originally inhibited against it, propan-2-ol can contain peroxides. When this occurs, an explosion is possible when the storage of the vessel or other equipment such as a dispensing bottle, is near empty and approaching dryness.

7.9 *Toluene*, (**Warning**—Flammable.)

7.10 *Commercial Aqueous pH 4 and, pH 7 Buffer Solutions*—These solutions shall be replaced at regular intervals consistent with their stability or when contamination is suspected. Information relating to their stability should be obtained from the manufacturer.

8. Electrode System

8.1 *Preparation of Electrodes*—When a Ag/AgCl reference electrode is used for the determination and it contains an electrolyte which is not 1 mol/L to 3 mol/L LiCl in ethanol, replace the electrolyte. Drain the electrolyte from the electrode, wash away all the salt (if present) with water and then rinse with ethanol. Rinse several times with the LiCl electrolyte solution. Finally, replace the sleeve and fill the electrode with the LiCl electrolyte to the filling hole. When refitting the sleeve, ensure that there will be a free flow of electrolyte into the system. A combination electrode shall be prepared in the

³ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the United States *Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

same manner. The electrolyte in a combination electrode can be removed with the aid of a vacuum suction.

8.2 Testing of Electrodes—Test the meter-electrode combination when first put into use, or when new electrodes are installed, and retest at intervals thereafter. Rinse the electrodes with solvent then with water, and dip them into a pH 4 aqueous buffer solution. Read the mV value after stirring 1 min. Remove the electrodes and rinse with water. Dip the electrodes into a pH 7 aqueous buffer. Read the mV value after stirring 1 min. Calculate the mV difference. A good electrode system will have a difference of at least 162 mV (20 °C to 25 °C). If the difference is less than 162 mV, lift the sleeve of the electrode and insure electrolyte flow. Repeat the measurements. If the difference is still less than 162 mV, clean or replace the electrode(s).

8.3 When the sensing electrode and the reference electrode are separate, one pair of electrodes shall be considered as one unit. If one or the other is changed, it shall be considered as different pair and shall be retested.

8.4 Maintenance and Storage of Electrodes—Cleaning the electrodes thoroughly, keeping the ground-glass joint free of foreign materials, and regular testing of the electrodes are very important in obtaining repeatable potentials, since contamination may introduce uncertain erratic and unnoticeable liquid contact potentials.

NOTE 5—See **Appendix X1** for a possible procedure to check the electrode performance.

8.5 Clean the glass electrode at frequent intervals based on use and type of samples being analyzed (not less than once every week during continual use) by immersing in non-chromium containing, strongly oxidizing cleaning solution. The reference electrode shall be cleaned periodically when in use or when a new electrode is installed. Drain the reference electrode at least once each week and refill with the fresh LiCl electrolyte as far as the filling hole. Ensure that there are no air bubbles in the electrode liquid. If air bubbles are observed, hold the electrode in a vertical position and gently tap it to release the bubbles. Maintain the electrolyte level in the reference electrode above that of the liquid in the sample beaker or vessel at all times.

8.6 Prior to each i-pH measurement soak the prepared electrodes in deionized water for at least 2 min. Rinse the electrodes with propan-2-ol immediately before use, and then with the dilution solvent.

8.7 When not in use, the glass indicator electrode membrane should be kept hydrated and stored according to the manufacturers' recommendation.

8.8 Electrode Life—Typically, electrode usage is limited to 3 months to 6 months depending, upon usage. Electrodes have a limited shelf life and shall be tested before use (see **8.2**).

9. Standardization of Apparatus

9.1 Calibration of the pH-meter with the Aqueous Buffer Solutions—The electrode should be calibrated daily using aqueous pH buffer solution standards 4 and 7. When using

direct electrode potential reading, establish a linear regression for conversion of sample potential readings to i-pH.

10. Preparation of Sample of Used Oil

10.1 Strict observance of the sampling procedure is necessary since the sediment, eventually present in the sample can be acidic or basic or has absorbed acidic or basic material from the sample. Failure to obtain a representative sample causes serious errors.

10.2 When applicable, refer to Practice **D4057** (Manual Sampling) or Practice **D4177** (Automatic Sampling) for proper sampling techniques.

10.3 When sampling used lubricants, the specimen shall be representative of the system sampled and shall be free of contamination from external sources.

NOTE 6—As used oil can change appreciably in storage, test samples as soon as possible after removal from the lubricating system and note the dates of sampling and testing.

10.4 Heat the sample (see **Note 7**) of used oil to 60 °C ± 5 °C in the original container and agitate until all of the sediment is homogeneously suspended in the oil. If the original container is a can or if it is glass and more than three-fourths full, transfer the entire sample to a clear-glass bottle having a capacity at least one third greater than the volume of the sample. Transfer all traces of sediment from the original container to the bottle by vigorous agitation of portions of the sample in the original container.

NOTE 7—When samples are visibly free of sediment, the heating procedure described can be omitted.

10.5 After complete suspension of all sediment, strain the sample or a convenient aliquot through a 100-mesh screen for removal of large contaminating particles.

NOTE 8—When samples are visibly free of sediment, the straining procedure described can be omitted.

11. Dilution Solvent

11.1 Add 5 mL ± 0.2 mL of water to 495 mL ± 5 mL of anhydrous propan-2-ol and mix well. Add 500 mL ± 5 mL of toluene. The dilution solvent should be made up in large quantities.

12. Procedure i-pH Value

12.1 Into a suitable beaker of capacity 100 mL, introduce a 2.5 g ± 0.3 g of sample and add 70 mL of dilution solvent. Prepare the electrodes as directed in **8.2**. Place the beaker on the measuring stand and adjust its position so that the electrodes are about half immersed. Start the stirrer, and stir throughout the determination at a rate sufficient to produce vigorous agitation without spattering and without stirring air into the solution. Start pH or electrode potential reading after 1 min stirring period and continue reading until pH or electrode potential remains stable within ±2 mV for 20 s (±0.03 pH per 20 s), but not longer than 2 min.

12.2 Remove the sample solution, rinse the electrodes with the dilution solvent, then with propan-2-ol and if clean finally with reagent grade water. At times, it may be necessary to clean with toluene alone vigorously after solvent cleaning to remove

any residual sample coated on the electrode prior to propan-2-ol cleaning. Immerse the electrodes in water for at least 2 min before starting another measurement to restore the aqueous gel layer of the glass electrode. After 2 min in the water, rinse the electrodes with propan-2-ol, then the dilution solvent before proceeding to the next measurement. If the electrodes are found to be dirty and contaminated, proceed as in 8.5. Store electrodes according to the recommendation in 8.7.

13. Calculation

13.1 When using pH reading mode, use the stable result or the reading after 2 min as the result. When using direct electrode potential reading, convert the stable reading or the reading after 2 min to pH using the linear regression according to Section 9.

14. Quality Control Checks

14.1 Confirm the performance of the test procedure by analyzing a quality control (QC) sample that is, if possible, representative of the samples typically analyzed.

NOTE 9—Because used oils, particularly used engine oils, are known to change during storage, such samples may not be suitable for this purpose.

14.2 Prior to monitoring the measurement process, the user of the method needs to determine the average value and control limits of the QC sample.⁴

14.3 Record the QC results and analyze by control charts or other statistically equivalent technique to ascertain the statistical control status of the total testing process.⁴ Any out-of-control data should trigger investigation for root cause(s). The results of this investigation may, but not necessarily, result in instrument recalibration.

14.4 The frequency of QC testing is dependent on the criticality of the quality being measured, the demonstrated stability of the testing process, and customer requirements. Generally, a QC sample should be analyzed each testing day. The QC frequency should be increased if a large number of samples are routinely analyzed. However, when it is demonstrated that the testing is under statistical control, the QC testing frequency may be reduced. The QC precision should be

periodically checked against the precision listed in the Precision and Bias section of this test method to ensure data quality.

14.5 It is recommended that, if possible, the type of QC sample that is regularly tested be representative of the samples routinely analyzed. An ample supply of QC sample material should be available for the intended period of use, and must be homogeneous and stable under the anticipated storage conditions. Because the i-pH value can vary while the QC sample is in storage, when an out-of-control situation arises, the stability of the QC sample can be a source of the error.

15. Report

15.1 Report i-pH-value to the nearest 0.01 pH.

15.2 For used oil samples report also the date of testing and, when available, the date the sample was taken (see 10.3, Note 6).

16. Precision and Bias⁵

16.1 *Repeatability*—The difference between successive test results obtained by the same operator with the same apparatus under constant operating conditions on identical test material would in the long run, in the normal and correct operation of the test method, exceed the following values only in one case in twenty.

$$r = 2.662E - 02 \cdot (10.1353 - X) \quad (1)$$

where X is the average of the two results.

16.2 *Reproducibility*—The difference between two single and independent results obtained by different operators working in different laboratories on identical test material would, in the long run, in the normal and correct operation of the test method, exceed the following values only in one case in twenty.

$$R = 0.1207 \cdot (10.1353 - X) \quad (2)$$

where X is the average of the two results.

17. Keywords

17.1 i-pH Value; lubricants; petroleum products; potentiometric

⁴ ASTM MNL 7, *Manual on Presentation of Data Control Chart Analysis*, 6th edition, ASTM International, W. Conshohocken, PA.

⁵ Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:D02-1779. Contact ASTM Customer Service at service@astm.org.

APPENDIX**(Nonmandatory Information)****X1. CHECK FOR ELECTRODE PERFORMANCE**

X1.1 The kinetic electrode test measures the kinetic response of the electrode. Electrodes can calibrate with acceptable slope and intercept values yet still not have a stability good enough for i-pH measurement. The speed of response and subsequent stability is important for pH-electrode. A manual check is described in this appendix that can be carried out with a pH meter or titrator set to read millivolts continuously.

X1.2 The essence of this check is to challenge the electrode coming from rest in a water solution with buffers and measure the potential after 30 and 60 s. A fast electrode reaches a stable point in less than 30 s and changes little from 30 to 60 s. Use buffers pH 4, and pH 7 for this check, as needed.

X1.3 Procedure:

X1.3.1 Set the pH meter to read electrode potential continuously. Have provision for stirring the buffer solution at the same speed used for the determinations.

X1.3.2 Allow the electrode to stabilize for 1 min in distilled or equivalent deionized water.

X1.3.3 Remove the electrodes from the water, and place them in the pH 4 buffer. Start a stopwatch at about the moment when the buffer touches the electrode.

X1.3.4 After 30 s, note the potential. After an additional 30 s, note the potential again. The difference between the two potentials is termed the drift.

X1.3.5 Repeat the procedure for pH 7 buffer.

X1.3.6 Calculate the drift for each of the buffers. The electrode response may be judged as follows:

- drift < 1 excellent
- 1 < drift < 2 good
- 2 < drift < 3 acceptable
- 3 < drift < 4 questionable
- 4 < drift unacceptable

X1.3.7 The difference between the potentials for pH 4 buffer and pH 7 buffer after 60 s reading time in each buffer should be greater than 162 mV, or 54 mV/pH number. Electrodes with a slope less than 54 mV/pH number are not reliable for determination.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; http://www.copyright.com/