



# Standard Test Method for Acidity of Benzene, Toluene, Xylenes, Solvent Naphthas, and Similar Industrial Aromatic Hydrocarbons<sup>1</sup>

This standard is issued under the fixed designation D847; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

*This standard has been approved for use by agencies of the U.S. Department of Defense.*

## 1. Scope\*

1.1 This test method is intended for the detection of acidity of benzene, toluene, xylenes, solvent naphthas, and similar industrial aromatic hydrocarbons.

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements see Section 9.

## 2. Referenced Documents

2.1 *ASTM Standards:*<sup>2</sup>

D1193 Specification for Reagent Water

D3437 Practice for Sampling and Handling Liquid Cyclic Products

2.2 *Other Documents:*

OSHA Regulations, 29CFR paragraphs 1910.1000 and 1910.1200<sup>3</sup>

## 3. Terminology

3.1 *Definitions:*

3.1.1 *acidity, n*—for purposes of this standard, acidity is defined as producing a persistent pink phenolphthalein end

point with the addition of two (2) drops or more (0.1 mL) of 0.1 N sodium hydroxide solution.

3.1.2 *acid reaction, n*—a characteristic of materials producing the acid-color of the indicator used under the conditions prescribed in this test method.

3.1.3 *alkaline or basic reaction, n*—a characteristic of the materials producing the alkali-color of the indicator used under the conditions prescribed in this test method.

## 4. Summary of Test Method

4.1 The acidity of aromatic hydrocarbons is detected using a sodium hydroxide titration and a color change using a phenolphthalein indicator.

## 5. Significance and Use

5.1 This test method is suitable for setting specifications, for use as an internal quality control tool, and for use in development or research work on industrial aromatic hydrocarbons and related materials. This test method gives an indication of residual acidity and is a measure of the quality of the finished product. It is an indication of the tendency of the product to corrode equipment.

## 6. Interferences

6.1 Tests for acidity are not applicable in the presence of contaminating acidic or alkaline gases, soaps, salts, or other compounds derived from the atmosphere or apparatus. The container holding the specimen, and the apparatus, water, indicator, and other materials used in the test shall be chosen so that they themselves shall not appreciably affect the results. Since new corks used in specimen bottles often are bleached with oxalic acid, it is advisable to rinse them thoroughly and check them for neutrality with the indicator used in the test. Glassware shall be of acid-resistant and alkali-resistant glass<sup>4</sup> and shall be rinsed with neutral distilled water before use. The room in which the test is performed shall be chosen so as to prevent undue contamination by carbon dioxide, ammonia, or other interfering substances that may be present in the atmosphere.

<sup>1</sup> This test method is under the jurisdiction of ASTM Committee D16 on Aromatic Hydrocarbons and Related Chemicals and is the direct responsibility of Subcommittee D16.01 on Benzene, Toluene, Xylenes, Cyclohexane and Their Derivatives.

Current edition approved June 1, 2016. Published July 2016. Originally approved in 1945. Last previous edition approved in 2015 as D847 – 15. DOI: 10.1520/D0847-16.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

<sup>3</sup> Available from U.S. Government Printing Office Superintendent of Documents, 732 N. Capitol St., NW, Mail Stop: SDE, Washington, DC 20401, http://www.access.gpo.gov.

<sup>4</sup> Borosilicate glass or the equivalent has been found satisfactory for this purpose.

\*A Summary of Changes section appears at the end of this standard

6.2 The distilled water used in the test shall not alter the composition of the specimen nor otherwise interfere with the purpose of the work. Although inert impurities often may be neglected, care must be exercised to correct for impurities or to eliminate them entirely if they are likely to interfere. When the distilled water does not show an acid or alkaline reaction, it may be used without further adjustment to neutrality. However, if the water shows an acid or alkaline reaction, it shall be brought to a persistent pink end point before use by titration with standard 0.01 *N* NaOH solution (see 8.4) or standard 0.01 *N* H<sub>2</sub>SO<sub>4</sub>, respectively (see 8.5).

## 7. Apparatus

7.1 *Graduate*, 100-mL.<sup>4</sup>

7.2 *Bottle*, 500-mL glass-stoppered.<sup>4</sup>

## 8. Reagents

8.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.<sup>5</sup> Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

8.2 *Phenolphthalein Indicator Solution*—Dissolve 0.5 g of phenolphthalein in 100 mL of ethyl alcohol (95 %). Add 0.01 *N* NaOH solution cautiously until a faint pink color appears; then just remove the color with a drop or two of 0.01 *N* acid.

8.3 *Sodium Hydroxide, Standard Solution* (0.1 *N*).

8.4 *Sodium Hydroxide, Standard Solution* (0.01 *N*).

8.5 *Sulfuric Acid, Standard* (0.01 *N*).

8.6 *Purity of Water: Distilled Water, Neutral*—Boil vigorously for 30 min, 1 to 2 L of distilled water conforming to Type III of Specification **D1193**. Insert a stopper carrying a guard tube of soda lime. Rinse a 250-mL flask with this distilled water, add a 100-mL portion, and titrate in a closed system with 0.01 *N* NaOH solution or heat to boiling and titrate immediately, taking care that the temperature does not fall

<sup>4</sup> *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For suggestions on the testing of reagents not listed by the American Chemical Society, see *Analar Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

below 80°C during the titration. If more than 1 drop (0.05 mL) of 0.01 *N* NaOH solution is required to obtain an end point with phenolphthalein, adjust the pH of the water to be used by adding the calculated amount of NaOH solution. Repeat the blank titration, and readjust if necessary until the blank titration on 100 mL of the distilled water is 1 drop (0.05 mL) or less of the 0.01 *N* NaOH solution. The distilled water now will be neutral or very slightly acid to the phenolphthalein indicator.

## 9. Hazards

9.1 Consult current OSHA regulations, supplier's Safety Data Sheets, and local regulations for all materials used in this test method.

## 10. Sampling

10.1 Sample the material in accordance with Practice **D3437**.

## 11. Procedure

11.1 Add 100 ± 5 mL of the sample in a 500-mL bottle. Add 100 mL of neutral distilled water (see 8.6) and 2 drops of phenolphthalein indicator solution. Stopper and shake thoroughly. Without separating the layers, titrate the acidity, if any, with standard 0.1 *N* NaOH to the first persistent shade of pink, shake thoroughly after each addition. Record Pass if 2 drops or less of 0.1 *N* NaOH solution produces a persistent pink end point. If more than 2 drops are required to produce a persistent pink end point, record Fail.

## 12. Interpretation of Results

12.1 Unless otherwise indicated in the applicable specifications, the test results shall be interpreted as follows:

12.1.1 A specimen shall be said to pass and to contain no free acid, that is, show no evidence of acidity, if 2 drops or less of 0.1 *N* NaOH solution produces a persistent pink end point in the test bottle.

## 13. Precision and Bias

13.1 In the case of pass/fail data, no generally accepted method for determining precision is currently available.

## 14. Quality Guidelines

14.1 In the case of pass/fail data, no generally acceptable method for determining that this test is under statistical control is currently available.

## 15. Keywords

15.1 acidity; aromatic hydrocarbons

**SUMMARY OF CHANGES**

Committee D16 has identified the location of selected changes to this standard since the last issue (D847–15) that may impact the use of this standard. (Approved June 1, 2016.)

(1) The size of the flask was revised in 8.6.

Committee D16 has identified the location of selected changes to this standard since the last issue (D847 – 08) that may impact the use of this standard. (Approved August 1, 2015.)

(1) Made revisions to Sections 1, 2, 4, 6, 7, 8 11, and 12.                      (3) Added Section 14, Quality Guidelines.  
(2) Revised definition for acidity.

*ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.*

*This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.*

*This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or [service@astm.org](mailto:service@astm.org) (e-mail); or through the ASTM website ([www.astm.org](http://www.astm.org)). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; <http://www.copyright.com/>*