



Designation: D857 – 17

Standard Test Method for Aluminum in Water¹

This standard is issued under the fixed designation D857; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope*

1.1 This test method covers the direct flame atomic absorption determination of aluminum in the nitrous oxide-acetylene flame.

1.2 This test method is applicable to waters containing dissolved and total recoverable aluminum in the range from 0.5 to 5.0 mg/L. Aluminum concentrations as high as approximately 50 mg/L can be determined using this test method without dilution. However, no precision and bias data are available for concentrations greater than 5.0 mg/L.

1.3 This test method was tested on reagent, natural, and potable waters. It is the user's responsibility to ensure the validity of this test method for waters of untested matrices.

1.4 The same digestion procedure may be used to determine total recoverable nickel (Test Methods [D1886](#)), chromium (Test Methods [D1687](#)), cobalt (Test Methods [D3558](#)), copper (Test Methods [D1688](#)), iron (Test Methods [D1068](#)), lead (Test Methods [D3559](#)), manganese (Test Methods [D858](#)), and zinc (Test Methods [D1691](#)).

1.5 Precision and bias data have been obtained on reagent, natural, and potable waters. It is the responsibility of the user to ensure the validity of this test method on untested matrices.

1.6 The values stated in SI units are to be regarded as standard. The values given in parentheses are mathematical conversion to inch-pound units that are provided for information only and are not considered standard.

1.7 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.* For specific hazard statements, see [7.5](#), [8.4](#), and [Note 2](#).

1.8 Former Test Methods A (Fluorometric) and B and C (Spectrophotometric) were discontinued. Refer to [Appendix X1](#) for historical information.

¹ These test methods are under the jurisdiction of ASTM Committee [D19](#) on Water and are the direct responsibility of Subcommittee [D19.05](#) on Inorganic Constituents in Water.

Current edition approved June 1, 2017. Published June 2017. Originally approved in 1945. Last previous edition approved in 2012 as D857 – 12. DOI: 10.1520/D0857-17.

1.9 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:²

[D858 Test Methods for Manganese in Water](#)
[D1066 Practice for Sampling Steam](#)
[D1068 Test Methods for Iron in Water](#)
[D1129 Terminology Relating to Water](#)
[D1193 Specification for Reagent Water](#)
[D1687 Test Methods for Chromium in Water](#)
[D1688 Test Methods for Copper in Water](#)
[D1691 Test Methods for Zinc in Water](#)
[D1886 Test Methods for Nickel in Water](#)
[D2777 Practice for Determination of Precision and Bias of Applicable Test Methods of Committee D19 on Water](#)
[D3370 Practices for Sampling Water from Closed Conduits](#)
[D3558 Test Methods for Cobalt in Water](#)
[D3559 Test Methods for Lead in Water](#)
[D4841 Practice for Estimation of Holding Time for Water Samples Containing Organic and Inorganic Constituents](#)
[D5673 Test Method for Elements in Water by Inductively Coupled Plasma—Mass Spectrometry](#)
[D5810 Guide for Spiking into Aqueous Samples](#)
[D5847 Practice for Writing Quality Control Specifications for Standard Test Methods for Water Analysis](#)

3. Terminology

3.1 Definitions:

3.1.1 For definitions of terms used in this standard, refer to Terminology [D1129](#).

3.2 Definitions of Terms Specific to This Standard:

3.2.1 *continuing calibration blank, n*—a solution containing no analytes (of interest) which is used to verify blank response and freedom from carryover.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

*A Summary of Changes section appears at the end of this standard

3.2.2 *continuing calibration verification*—a solution (or set of solutions) of known concentration used to verify freedom from excessive instrumental drift; the concentration is to cover the range of calibration curve.

3.2.3 *total recoverable aluminum, n*—a descriptive term relating to the forms of aluminum recovered in the acid digestion procedure specified in this standard.

4. Summary of Test Method

4.1 Aluminum is determined by direct atomic absorption with only the addition of an ionization suppressor and sensitivity enhancer (optional).

5. Significance and Use

5.1 Although there is little information available concerning the toxicological significance of aluminum in man, the American Water Works Association has established a water quality guideline or goal of a maximum of 0.05 mg/L. Under the National Pollution Discharge Elimination System (NPDES), some permits may set aluminum discharge limits. Some evidence does exist to indicate that low levels (5 mg/L) will interfere with activated sludge processes. For the above reasons monitoring of aluminum may be desirable.

5.2 Aluminum is monitored in boiler make-up water, where alum has been used, to determine whether aluminum is present after pretreatment. Residual aluminum may consume ion exchange capacity or consume boiler water treatment chemicals added to stoichiometrically chelate hardness ions (that is, calcium and magnesium) in boiler feed water.

5.3 Aluminum is monitored in cooling water make-up, since its presence may result in deactivation of anionic substances in scale or corrosion inhibitor treatment chemicals, or both. Deactivation may result in decreased performance of inhibitors.

5.4 ICP-MS may also be appropriate but at a higher instrument cost. See Test Method [D5673](#).

6. Interferences

6.1 Aluminum ionizes slightly in the nitrous oxide-acetylene flame, but the addition of sodium chloride described in this test method suppresses this interference. By this technique, a maximum concentration of 9000 mg/L sodium, 9000 mg/L potassium, 4000 mg/L calcium, 4000 mg/L magnesium, 9000 mg/L sulfate, 9000 mg/L chloride, 9000 mg/L nitrate, and 9000 mg/L iron may be tolerated.

7. Apparatus

7.1 *Atomic Absorption Spectrophotometer*, for use at 309.3 nm.

7.2 *Aluminum Hollow-Cathode Lamp*—Multielement lamps are not recommended.

7.3 *Oxidant*—See [8.8](#).

7.4 *Fuel*—See [8.9](#).

7.5 *Pressure-Reducing Regulators*—The supplies of fuel and oxidant should be reduced by suitable regulators to the

levels recommended by the manufacturer of the spectrophotometer. (**Warning**—The nitrous oxide-acetylene flame is hazardous due to its flash-back potential. Follow the instrument manufacturer's recommended operating procedures closely.)

8. Reagents and Materials

8.1 *Purity of Reagents*—Reagent grade chemicals shall be used in all tests. Unless otherwise indicated, it is intended that all reagents shall conform to the specifications of the Committee on Analytical Reagents of the American Chemical Society where such specifications are available.³ Other grades may be used, provided it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening the accuracy of the determination.

8.2 *Purity of Water*—Unless otherwise indicated, references to water shall be understood to mean reagent water conforming to Specification [D1193](#), Type I. Other reagent water types may be used, provided it is first ascertained that the water is of sufficiently high purity to permit its use without lessening the bias and precision of the determination. Type II water was specified at the time of round-robin testing of this test method.

8.3 *Aluminum Solution, Standard (1 mL = 0.1 mg Al)*—Dissolve 1.758 g of aluminum potassium sulfate (AlK(SO₄)₂·12H₂O) in water. Add 1 mL of nitric acid and dilute to 1 L. A purchased aluminum stock solution of appropriate known purity is also acceptable.

8.4 *Bis-(2-Ethoxyethyl) Ether*. (**Warning**—Avoid inhalation. Perform all manipulation in a well-ventilated hood. This ether can form dangerous peroxides and should be inspected regularly for their presence.)

8.5 *Hydrochloric Acid (sp gr 1.19)*—Concentrated hydrochloric acid (HCl).

8.6 *Nitric Acid (sp gr 1.42)*—Concentrated nitric acid (HNO₃).

8.7 *Sodium Chloride Solution (25.4 g/L)*—Dissolve 25.4 g of sodium chloride (NaCl) in water and dilute to 1 L.

8.8 *Oxidant*:

8.8.1 *Air*, which has been suitably dried and filtered, is used to support combustion before switching to nitrous oxide.

8.8.2 *Nitrous Oxide* is the required oxidant.

8.9 *Fuel*—Standard commercially available acetylene is the required fuel. The cylinder should be replaced at a gage pressure of 517 kPa (75 psi) to minimize the carry-over of acetone. Since “purified” grades contain a solvent that softens poly(vinyl chloride) tubing, its use constitutes a safety hazard and is not recommended.

8.10 *Filter Paper*—Purchase suitable filter paper. Typically the filter papers have a pore size of 0.45- μ m membrane. Material such as fine-textured, acid-washed, ashless paper, or

³ *Reagent Chemicals, American Chemical Society Specifications*, American Chemical Society, Washington, DC. For Suggestions on the testing of reagents not listed by the American Chemical Society, see *Annual Standards for Laboratory Chemicals*, BDH Ltd., Poole, Dorset, U.K., and the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USPC), Rockville, MD.

glass fiber paper are acceptable. The user must first ascertain that the filter paper is of sufficient purity to use without adversely affecting the bias and precision of the test method.

9. Sampling

9.1 Collect samples in accordance with Practices **D1066** or **D3370**, as applicable. For dissolved aluminum, filter the samples at the time of collection through a 0.45- μm filter (8.10). Acidify the filtrate to pH 2 or less with nitric acid (8.6). For total recoverable aluminum, acidify the unfiltered sample to pH 2 or less at the time of collection. The holding time for samples may be calculated in accordance with Practice **D4841**.

NOTE 1—Alternatively, the pH may be adjusted in the laboratory within 14 days of collection. However, acid must be added at least 24 hours before analysis to dissolve any metals that adsorb to the container walls. This could reduce hazards of working with acids in the field when appropriate.

9.2 An effective way to clean glassware is to soak all glassware in hot HCl (1 + 1) for 2 h. Drain and rinse with water or reagent.

10. Preparation of Apparatus

10.1 An effective way to clean glassware is to soak all glassware in hot HCl (1 + 1) for 2 h. Drain and rinse with water or reagent.

11. Calibration and Standardization

11.1 Prepare aluminum standards in the range from 0.0 to 5.0 mg/L by dilution of the aluminum standard solution (see 8.3) with water. It is the responsibility of the user to ensure the validity of this test method for concentrations if the range is extended.

11.2 Add 1.0 mL of NaCl solution (8.7) to 10.0 mL of standard and mix thoroughly. If very low concentrations of aluminum are anticipated in the samples, 1 mL of *bis*-(2-ethoxyethyl) (8.4) ether may be added to the blank and each standard, which will act as a sensitivity enhancer. If this sensitivity enhancer is used, it must also be added to each sample (12.6).

11.3 Read directly in concentration if this capability is provided with the instrument or measure the absorbance of the standards and construct an analytical curve by plotting the absorbance of the standards versus the concentration of aluminum.

12. Procedure

12.1 If dissolved aluminum is to be determined proceed to 12.6.

12.2 Measure a volume of well-mixed acidified sample containing less than 0.5 mg of aluminum (100 mL maximum) into a 125-mL beaker.

12.3 Add 0.5 mL of HNO_3 and 5 mL of HCl.

12.4 Heat the samples (between 65°C and 95°C) on a hot plate below boiling in a well-ventilated hood until the volume has been reduced to 10 to 15 mL, making certain the samples do not boil.

TABLE 1 Overall (S_T) and Single-Operator (S_o) Interlaboratory Precision for Aluminum by Direct AAS

Reagent Water:			
Concentration (X), mg/L	0.848	2.54	4.11
S_T	0.167	0.19	0.19
S_o	0.128	0.18	0.23
Natural Water:			
Concentration (X), mg/L	0.772	2.48	4.07
S_T	0.194	0.19	0.21
S_o	0.166	0.16	0.28

NOTE 2—For brines and samples with high levels of suspended matter, the amount of reduction in volume is left to the discretion of the analyst.

NOTE 3—Many laboratories have found block digestion systems a useful way to digest samples for trace metals analysis. Systems typically consist of either a metal or graphite block with wells to hold digestion tubes. The block temperature controller must be able to maintain uniformity of temperature across all positions of the block. The digestion block must be capable of maintaining a temperature between 65°C and 95°C. For trace metals analysis, the digestion tubes should be constructed of polypropylene and have a volume accuracy of at least 0.5 %. All lots of tubes should come with a certificate of analysis to demonstrate suitability for their intended purpose.

12.5 Cool and, if necessary, filter the samples through a filter (fine-texture, acid-washed, ashless paper) into a 100-mL volumetric flask. Wash the paper several times and bring the volume to 100-mL.

12.6 To a 10-mL aliquot of sample add 1.0 mL of NaCl and mix thoroughly. If 1 mL of *bis*-(2-ethoxyethyl) ether was added to the blank and standards as a sensitivity enhancer (11.2), an equal amount must be added to each sample.

12.7 Atomize each of the standards, samples, and blank and determine its absorbance or concentration. Atomize water between samples.

13. Calculation

13.1 Determine the concentration of aluminum in each sample by referring to the calibration curve in 13.3. Alternatively, calibrate the spectrophotometer and read directly in concentration mode.

13.2 Calculate the concentration of total recoverable aluminum in milligrams per litre using Eq 1:

$$\text{Aluminum, mg/L} = C \times (100/V) \quad (1)$$

where:

C = concentration from curve, mg/L, and

V = volume of aliquot, mL.

13.3 Calculate the concentration of dissolved aluminum directly from the calibration curve.

14. Precision and Bias⁴

14.1 Precision data for this test method were obtained on reagent, natural, and potable waters. It is the user's responsibility to assure the validity of this test method for waters of untested matrices.

⁴ Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Reports RR:D19-1064 and RR:D19-1065. Contact ASTM Customer Service at service@astm.org.

TABLE 2 Recovery by Direct AAS

Amount Added, mg/L	Amount Found, mg/L	% Bias	Statistically Significant (95 % Confidence Level)
Reagent Water			
0.749	0.848	+13.2	yes
2.49	2.54	+2.0	no
4.49	4.11	-8.4	yes
Matrix Water			
0.749	0.772	+3.1	no
2.49	2.48	-0.4	no
4.49	4.07	-9.4	yes

14.2 The collaborative test of the direct atomic absorption test method for aluminum in reagent water was performed at three levels by four laboratories (seven operators) making a total of fifteen observations at each level. The test in matrix water at the same levels was performed by three laboratories (six operators) making a total of twelve observations at each level.

14.3 Precision and bias for this test method conform to Practice **D2777** – 77, which was in place at the time of collaborative testing. Under the allowances made in 1.4 of Practice **D2777** – 13, these precision and bias data do meet existing requirements for interlaboratory studies of Committee D19 test methods.

14.4 The overall and single-operator precision of this test method (S_t and S_o respectively) within its designated range for reagent water and selected water matrices varies with the quantity tested in accordance with **Table 1**.

14.5 Recoveries of known amounts of aluminum from reagent water, Type II, and selected water matrices were as shown in **Table 2**.

14.6 The sensitivity enhancer was not used in collaborative tests. It is the user's responsibility to determine its effect on precision and bias.

15. Quality Control (QC)

15.1 To ensure that analytical values obtained using this test method are valid and accurate within the confidence limits of the test, the following QC procedures must be followed when analyzing aluminum by this test method.

15.2 Calibration and Calibration Verification:

15.2.1 Analyze at least three working standards containing concentrations of aluminum that bracket the expected sample concentration prior to analysis of samples to calibrate the instrument. The calibration correlation coefficient shall be equal to or greater than 0.990.

15.2.2 Verify instrument calibration after standardization by analyzing a standard at the concentration of one of the calibration standards. The absorbance shall fall within 4 % of the absorbance from the calibration. Alternately, the concentration of a mid-range standard should fall within ± 15 % of the known concentration. Analyze a calibration blank to verify system cleanliness. The blank result should be less than the method reporting limit.

15.2.3 If calibration cannot be verified, recalibrate the instrument.

15.2.4 It is recommended to analyze a continuing calibration blank (CCB) and continuing calibration verification (CCV) at a 10 % frequency. The CCB result should be less than the method reporting limit. The CCV results should fall within the expected precision of the method or ± 15 % of the known concentration.

15.3 Initial Demonstration of Laboratory Capability:

15.3.1 If a laboratory has not performed the test before, or if there has been a major change in the measurement system, for example, new analyst, new instrument, etc., a precision and bias study must be performed to demonstrate laboratory capability.

15.3.2 Analyze seven replicates of a standard solution prepared from an Independent Reference Material containing 2.5 mg/L of aluminum. The matrix and chemistry of the solution should be equivalent to the solution used in the collaborative study. Each replicate must be taken through the complete analytical test method including any sample preservation and pretreatment steps.

15.3.3 Calculate the mean and standard deviation of the seven values and compare to the acceptable ranges of bias in **Table 1**. This study should be repeated until the recoveries are within the limits given in **Table 1**. If a concentration other than the recommended concentration is used, refer to Practice **D5847** for information on applying the F test and t test in evaluating the acceptability of the mean and standard deviation.

15.4 Laboratory Control Sample (LCS):

15.4.1 To ensure that the test method is in control, prepare and analyze a LCS containing a mid-range concentration of aluminum with each batch (laboratory-defined or 20 samples). The laboratory control samples for a large batch should cover the analytical range when possible. It is recommended, but not required to use a second source, if possible and practical for the LCS. The LCS must be taken through all of the steps of the analytical method including sample preservation and pretreatment. The result obtained for the LCS shall fall within ± 15 % of the known concentration.

15.4.2 If the result is not within these limits, analysis of samples is halted until the problem is corrected, and either all the samples in the batch must be reanalyzed, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

15.5 Method Blank:

15.5.1 Analyze a reagent water test blank with each laboratory-defined batch. The concentration of aluminum found in the blank should be less than 0.5 times the lowest calibration standard. If the concentration of aluminum is found above this level, analysis of samples is halted until the contamination is eliminated and a blank shows no contamination at or above this level, or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

15.6 Matrix Spike (MS):

15.6.1 To check for interferences in the specific matrix being tested, perform a MS on at least one sample from each laboratory-defined batch by spiking an aliquot of the sample

with a known concentration of aluminum and taking it through the analytical method.

15.6.2 The spike concentration plus the sample concentration of aluminum must not exceed the high calibration standard. The spike must produce a concentration in the spiked sample that is 2 to 5 times the analyte concentration in the unspiked sample, or 10 to 50 times the detection limit of the test method, whichever is greater.

15.6.3 Calculate the percent recovery of the spike (P) using the following formula:

$$P = \frac{100 [A(V_s + V) - B V_s]}{C V} \quad (2)$$

where:

- A = analyte known concentration ($\mu\text{g/L}$) in spiked sample,
- B = analyte known concentration ($\mu\text{g/L}$) in unspiked sample,
- C = known concentration ($\mu\text{g/L}$) of analyte in spiking solution,
- V_s = volume (mL) of sample used, and
- V = volume (mL) of spiking solution added.

15.6.4 The percent recovery of the spike shall fall within the limits, based on the analyte concentration, listed in Guide [D5810](#), Table 1. If the percent recovery is not within these limits, a matrix interference may be present in the sample selected for spiking. Under these circumstances, one of the following remedies must be employed: the matrix interference must be removed, all samples in the batch must be analyzed by a test method not affected by the matrix interference, or the

results must be qualified with an indication that they do not fall within the performance criteria of the test method.

NOTE 4—Acceptable spike recoveries are dependent on the concentration of the component of interest. See Guide [D5810](#) for additional information.

15.7 Duplicate:

15.7.1 To check the precision of sample analyses, analyze a sample in duplicate with each laboratory-defined batch. If the concentration of the analyte in the sample (and duplicate) is less than five times the detection limit for the analyte, a Matrix Spike Duplicate (MSD) should be used.

15.7.2 Calculate the standard deviation of the duplicate values and compare to the precision in the collaborative study using an F test. Refer to 6.4.4 of Practice [D5847](#) for information on applying the F test.

15.7.3 If the result exceeds the precision limit, the batch must be reanalyzed or the results must be qualified with an indication that they do not fall within the performance criteria of the test method.

15.8 Independent Reference Material (IRM):

15.8.1 In order to verify the quantitative value produced by the test method, analyze an IRM submitted as a regular sample (if practical) to the laboratory at least once per quarter. The concentration of the reference material should be in the range of 2-3 mg/L aluminum. The value obtained must fall within the control limits established by the laboratory.

16. Keywords

16.1 aluminum; analysis; atomic absorption flame; water

APPENDIX

(Nonmandatory Information)

X1. RATIONALE FOR DISCONTINUATION OF TEST METHODS

X1.1 Former Test Method A—Fluorometric

X1.1.1 This test method was discontinued in 1988. The test method may be found in its entirety in the 1988 *Annual Book of ASTM Standards*, Vol 11.01.

X1.1.2 This test method is applicable to the determination of soluble aluminum in water in concentrations below 5 $\mu\text{g/L}$. The water must be of relatively high purity. Heavy metal ions of the order of 20 $\mu\text{g/L}$ produce interferences.

X1.1.3 The fluorescence of an aluminum-morin complex formed at a pH of 3 is measured and referred to a previously prepared calibration curve to determine the aluminum concentration.

X1.1.4 This test method was discontinued because there were insufficient laboratories interested in participating in another collaborative study to obtain the necessary precision and bias as required by Practice [D2777](#).

X1.2 Former Test Method B—Spectrophotometric

X1.2.1 This test method was discontinued in 1988. The test method may be found in its entirety in the 1988 *Annual Book of ASTM Standards*, Vol 11.01.

X1.2.2 This test method is applicable to the determination of total aluminum in water in the range from 0.00 to 0.50 mg/L as Al^{3+} (based on a 100-mL sample). The range of the test may be extended by taking a suitable aliquot.

X1.2.3 The aluminum is complexed with sodium fluoride to prevent its reaction with 8-quinolinol while interfering ions are removed by a weak cationic ion exchange resin and an 8-quinolinol-chloroform extraction. The aluminum is then made reactive to the 8-quinolinol by pH adjustment and its chloroform-extracted 8-quinolate is measured spectrophotometrically.

X1.2.4 This test method was discontinued because there were insufficient laboratories interested in participating in

another collaborative study to obtain the necessary precision and bias as required by Practice **D2777**.

X1.3 Former Test Method C—Spectrophotometric

X1.3.1 This test method was discontinued in 1988. The test method may be found in its entirety in the 1988 *Annual Book of ASTM Standards*, Vol 11.01.

X1.3.2 This test method is applicable to water relatively free of interfering substances. The procedure is simple and rapid, but only soluble aluminum is determined.

X1.3.3 Ferron (8-hydroxy-7-iodo-5-quinoline sulfonic acid) reacts with aluminum to form a soluble complex that absorbs

ultraviolet light. The absorbance of the complex is proportional to the soluble aluminum and is measured by a spectrophotometer at a wavelength of 370 nm.

X1.3.3.1 A significant interference from iron is greatly minimized by adding orthophenanthroline. This has the additional advantage that iron may be simultaneously determined, but at a wavelength of 520 nm.

X1.3.4 This test method was discontinued because there were insufficient laboratories interested in participating in another collaborative study to obtain the necessary precision and bias as required by Practice **D2777**.

SUMMARY OF CHANGES

Committee D19 has identified the location of selected changes to this standard since the last issue (D857 – 12) that may impact the use of this standard. (Approved June 1, 2017.)

- | | |
|---|--|
| (1) Revised 1.6 to update the units of measurement statement. | (8) Revised Note 3 to clarify the block digestion systems. |
| (2) Revised Section 2 to include D5673 . | (9) Revised Section 11 with information on exceeding the analytical range and for direct reading instruments. |
| (3) Revised Section 3 to update and add terms. | (10) Revised Section 12 with information on the heating blocks. |
| (4) Added 5.4 to include information on using ICP-MS. | (11) Revised 15.2 and 15.4.1 . |
| (5) Added 8.10 to include information on filter paper. | |
| (6) Revised Note 1 to clarify the addition of the acid. | |
| (7) Revised 9.2 and 10.1 to include information on cleaning glassware | |

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; http://www.copyright.com/