

Standard Practice for Preparation of Soil Samples by Hotplate Digestion for Subsequent Lead Analysis¹

This standard is issued under the fixed designation E1726; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ε) indicates an editorial change since the last revision or reapproval.

1. Scope

- 1.1 This practice covers drying, homogenization, and acid digestion of soil samples and associated quality control (QC) samples using a hot plate type method for the determination of lead using laboratory atomic spectrometry analysis techniques such as Inductively Coupled Plasma Atomic Emission Spectrometry (ICP-AES), Flame Atomic Absorption Spectrometry (FAAS), and Graphite Furnace Atomic Absorption Spectrometry (GFAAS).
- 1.2 This practice is based on U.S. EPA SW 846, Test Method 3050.
- 1.3 This practice contains notes that are explanatory and are not part of the mandatory requirements of this standard.
- 1.4 The values stated in SI units are to be regarded as standard. The values given in parentheses are mathematical conversions to inch-pound units that are provided for information only and are not considered standard.
- 1.5 This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

2. Referenced Documents

2.1 ASTM Standards:²

D1129 Terminology Relating to Water

D1193 Specification for Reagent Water

E288 Specification for Laboratory Glass Volumetric Flasks

E1605 Terminology Relating to Lead in Buildings

2.2 U.S. Government Analytical Method:³

U.S. EPA SW 846 Test Methods for Evaluating Solid Waste Physical/Chemical Methods

2.3 ISO Standards:⁴

ISO Guide 30 Terms and Definitions Used in Connection with Reference Materials

ISO 1042 Laboratory glassware—One-mark volumetric flasks

ISO 8655 Piston-operated volumetric apparatus -- Part 1: Terminology, general requirements and user recommendations

3. Terminology

- 3.1 *Definitions:* For definitions of terms relating to the preparation of dried paint samples that are not given here, refer to Terminology D1129 or Terminology E1605.
- 3.1.1 *batch*—a group of field or quality control samples that are processed together using the same reagents and equipment.
- 3.1.2 *digestate*—the acidified aqueous solution that results from digestion of the sample.
- 3.1.3 *digestion*—high temperature sample preparation process that involves chemical breakdown to solubilize targeted analytes present in a sample, to result in an acidified aqueous solution called the digestate.
- 3.1.4 *method blank*—a sample, devoid of analyte, that is analyzyed to determine its contribution to the total blank (background) reading.
- 3.1.5 *non-spiked sample*—a sample, devoid of analyte, that is targeted for addition of analyte but is not fortified with all target analytes prior to sample preparation.
- 3.1.5.1 *Discussion*—Analysis results for this sample are used to correct for background levels in the blank medium that is used for spiked and spiked duplicate samples.
- 3.1.6 reagent blank—a digestate that reflects the maximum treatment given any one sample within a batch of samples, except that it has no sample placed initially into the digestion

¹ This practice is under the jurisdiction of ASTM Committee E06 on Performance of Buildings and is the direct responsibility of Subcommittee E06.23 on Lead Hazards Associated with Buildings.

Current edition approved Oct. 1, 2016. Published November 2016. Originally approved in 1995. Last previous edition approved in 2009 as E1726-01(2009). DOI: 10.1520/E1726-16.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from U.S. Government Publishing Office, 732 N. Capitol St., NW, Washington, DC 20401-0001, http://www.gpo.gov.

⁴ Available from American National Standards Institute (ANSI), 25 W. 43rd St., 4th Floor, New York, NY 10036, http://www.ansi.org.

- vessel. (The same reagents and processing conditions that are applied to field samples within a batch are also applied to the reagent blank.)
- 3.1.6.1 *Discussion*—Analysis results from this sample provide information on the level of potential contamination resulting from only laboratory sources that are experienced by samples processed within the batch.
- 3.1.7 reference material (certified reference material) (CRM)—reference material accompanied by a certificate, one or more of whose property values are certified by a procedure which establishes its traceability to an accurate realization of the unit in which the property values are expressed; each certified value is accomplished by an uncertainty at a stated level of confidence (ISO Guide 30).
- 3.1.8 *spiked sample or spiked duplicate sample*—a blank medium that contains no purposely added analyte to which a known amount of analyte is added before preparation.
- 3.1.8.1 *Discussion*—Analysis results for these samples are used to provide information on the precision and accuracy of the overall process.

4. Summary of Practice

4.1 A representative soil sample is dried and homogenized, and then digested (in a batch mode with other samples) on a hot plate using nitric acid and hydrogen peroxide. The digestate is diluted for final volume prior to lead measurement.

5. Significance and Use

- 5.1 There is a need to monitor the lead content in and around buildings and related structures in order to determine the potential lead hazard. Hence, effective and efficient methods are required for the preparation of soil samples for determination of their lead content.
- 5.2 This practice may be used for the digestion of soil samples that are collected during various construction and renovation activities associated with lead abatement in and around buildings and related structures. The practice is also suitable for the digestion of soil samples for lead analyses collected from other locations, such as near roads and steel structures.
- 5.3 This practice is intended to be used to prepare samples that have been collected for hazard assessment purposes.
- 5.4 This practice is not capable of determining lead bound within matrices, such as silica, that are not soluble in nitric acid.
- 5.5 This practice includes drying and homogenization steps in order to help assure that reported lead results are representative of the sample and are independent of potential differences in soil moisture levels among different sampling locations or changing weather conditions.

6. Apparatus

- 6.1 Equipment:
- 6.1.1 *Analytical Balance*, capable of accurately determining the mass to the nearest 0.001 g.
- 6.1.2 *Drying Oven*, capable of maintaining a temperature of 100 to 120°C.

- 6.1.3 *Electric Hot Plate*, capable of maintaining a temperature of 80 to 100°C as measured with a thermometer placed into a beaker or flask filled with water sitting on the hot plate head. When required to reduce the presence of hot spots in the electrical hot plate, place a 2 to 2.5 cm (0.75 to 1 in.) thick aluminum plate on the burner head.
- 6.1.4 *Grinding Apparatus*—Mortar and pestle (porcelain or agate), shatter box, or mixer mill.
- 6.1.5 *Micropipettors with Disposable Plastic Tips* conforming to ISO 8655, sizes needed to make reagent additions, and spiking standards (see Note 1).

Note 1—In general, the following sizes should be readily available: 1 to 5 mL adjustable, 1000, 500, 250, and 100 μ L.

6.1.6~Sieves, 4.75~mm (U.S. Standard No. 4), 2~mm (No. 10), and $500~\mu m$ (No. 35), plastic or stainless steel (see Note 2). When sieves containing soldered joints are used, then all solder joints shall be coated with epoxy resin prior to use to protect samples from potential lead contamination originating in the solder. Visually inspect prior to use for the presence of bare metal.

Note 2—Plastic or stainless steel sieves are better for use instead of brass sieves to alleviate possible lead contamination of the soil samples from contact with lead solder common to brass sieves.

- 6.1.7 *Thermometers*, red alcohol, that cover a range from 0 to 110°C.
 - 6.2 Glassware and Supplies:
- 6.2.1 *Borosilicate Glassware*—Volumetric flasks with stoppers meeting Specification E288 or conforming to ISO 1042, 100 mL; Griffin beakers, 100, 150 or 250 mL; watch glasses sized to cover Griffin beakers.
 - 6.2.2 Plastic Gloves, powderless.
- 6.2.3 *Air-Tight Sample Containers*—1 L (1 qt) or 4 L (1 gal) re-sealable plastic bags, or plastic 50 mL centrifuge tubes.
- 6.2.4 *Volumetric Flasks*—meeting Specification E288 or conforming to ISO 1042, 100 mL and other sizes as needed to make dilutions of sample digests or lead standards used for fortification of spiked samples.
 - 6.2.5 Tongs, metal.
 - 6.2.6 Spoon, stainless steel.

7. Reagents

- 7.1 *Purity of Reagents*—Reagent grade chemicals shall be used in this practice. Unless otherwise indicated, all reagents shall conform to the specifications for the Committee on Analytical Reagents of the American Chemical Society, where such specifications are available.⁵ Other grades shall not be used unless it is first ascertained that the reagent is of sufficiently high purity to permit its use without lessening accuracy of the determination.
- 7.2 Nitric Acid—Concentrated, suitable for atomic spectrometry analysis, such as spectroscopic grade.

⁵ Reagent Chemicals, American Chemical Society Specifications, American Chemical Society, Washington, DC, www.chemistry.org. For suggestions on the testing of reagents not listed by the American Chemical Society, see the *United States Pharmacopeia and National Formulary*, U.S. Pharmacopeial Convention, Inc. (USP), Rockville, MD, http://www.usp.org.

- 7.3 Hydrogen Peroxide, 30 % (w/w), suitable for atomic spectrometry analysis such as spectroscopic grade.
 - 7.4 Acetone, reagent, spectroscopic grade.
- 7.5 Water—Unless otherwise indicated, references to water shall mean reagent water as defined by Type I of Specification D1193. (ASTM Type I Water: minimum resistance of 16.7 megohm·cm, or equivalent.)
- 7.6 Calibration Stock Solution, 100 μ g/mL of lead (Pb) CRM in dilute (typically 2 %) nitric acid.

8. Sample Preparation Procedure

- 8.1 Sample Pre-Treatment:
- 8.1.1 Treat each sample in a processing batch equally.
- 8.1.2 If possible before removal, break up the soil sample within the original containers containing the samples (see Note 3).
 - Note 3—This will not be possible for wet soil samples.
- 8.1.3 Label an acid-cleaned 100, 150, or 250 mL Griffin beaker (or other vessel suitable for oven drying of soils that will not contaminate the sample with lead) with a high temperature wax pen or any other marker that will be visible after exposure to the drying oven.
- 8.1.4 Transfer the entire soil sample to the labeled Griffin beaker. Cover with a watch glass (tip to one side to permit moisture removal), and place in a drying oven for a minimum of 6 h at a temperature of $110 \pm 10^{\circ}$ C. Samples that cake or plug the sieve require additional drying (see Note 4).
- Note 4—Soil samples should not cake or exhibit packing characteristic of moisture, but should flow freely through the sieve (see 8.1.7) when broken apart.
- 8.1.5 If the received soil sample is excessively large, then any attempts to sub-sample prior to drying and sieving are likely to cause bias. If possible, use a larger beaker to contain the entire sample. If not, then use multiple beakers followed by re-combining after drying. Samples that cake or plug the sieve require additional drying (see Note 4).
- 8.1.6 Using tongs, remove the beakers containing the samples and allow them to cool to room temperature.
- 8.1.7 Don a pair of plastic gloves and push the soil sample through a clean 4.75 mm sieve (U.S. Standard No. 4) to remove any large objects or root material, or both. Discard material retained on the sieve (see Notes 5 and 6). Clean the sieve between samples by tapping or using forced air or other dry method to prevent cross-contamination. Perform this step in a location well removed from other samples in process and in an area where soil dust will not contaminate the laboratory operations such as in front of a fume hood.
- Note 5—If the samples do not appear to contain any large objects or root material, it is not necessary to perform this step with the 4.75 mm sieve
- Note 6—In order to minimize small particle size soil losses, this step should not be performed inside a fume hood.
- 8.1.8 Don a pair of plastic gloves and push the soil sample through a clean 2 mm sieve (U.S. Standard No. 10) to remove coarse material (see Note 6). Discard material retained on the sieve. Clean the sieve between samples by tapping or using

- forced air or other dry method to prevent cross-contamination. Perform this step in a location well removed from other samples in process, and in an area where soil dust will not contaminate the laboratory operations, such as in front of a fume hood.
- 8.1.9 Grind the sample using a porcelain mortar and pestle or other appropriate homogenization apparatus such as a shatter-box or mixer mill. Clean the grinding apparatus between samples to prevent cross-contamination between samples by rinsing with water and drying. When any material is retained from 8.1.10, delay cleaning until this retained material for the sample is re-ground as described in 8.1.11. An acetone rinse will facilitate drying (see Note 7).
- Note 7—Acetone should not be used on sieves since it can damage epoxy coatings which may be present to seal lead solder joints.
- 8.1.10 Place the ground up sample on a clean 500 µm sieve (U.S. Standard No. 35). Use a stainless steel spoon to help move material around until no more sample will pass through the sieve. Do not discard the retained material (see Note 8). Return any retained material for one more grinding as described in 8.1.9.
- Note 8—A second re-grinding step is included for retained material to avoid inadvertent loss of larger pieces of material that can remain as a result of inadequate grinding.
- 8.1.11 Place the ground up retained sample material back on the clean U.S. Standard No. 35 (500 μ m sieve) (see Note 6). Using a stainless steel spoon, help move material around until no more sample will pass through the sieve, adding the passed material to the previous sample material that passed through the sieve. Discard any retained material. Clean the sieve between samples by tapping or using forced air or other dry method to prevent cross-contamination. Perform this step in a location well removed from the samples in process.
- 8.1.12 Label acid-cleaned 100, 150, or 250 mL Griffin beakers and watch glasses for performing the digestion of each soil sample and associated QC samples.
- 8.1.13 Transfer sieved portion to a labeled Griffin beaker and place in a drying oven overnight or for a minimum of 12 h, or to constant mass at a temperature of $110 \pm 10^{\circ}$ C (see Note 9). Remove from oven and allow to cool to room temperature.
- Note 9—Constant mass for this procedure is defined as a less than 0.1 % change in mass for repeated measurements (a minimum of two) taken over a minimum of a 1 h interval.
- 8.1.14 Store the dried, homogenized, and sieved soil samples inside new labeled air-tight sample containers.
 - 8.2 Sample Digestion:
- 8.2.1 Turn or roll the sample container repeatedly for about 1 min. Determine the mass of each dried homogenized sample to the nearest 0.001 g and transfer a 1.0 \pm 0.10 g portion of the sample to a labeled Griffin beaker. Record the mass of each sample. For sample portions targeted for spiking, add the appropriate volume of a lead standard stock to the beaker (see Note 10). In the absence of other information, add 500 μg of lead to each beaker containing sample portions targeted for spikes and spike duplicates (5 mL of 100 $\mu g/mL$ of Pb stock solution).



Note 10—The appropriate volume will be dependent on the anticipated lead level in the targeted sample split. The optimum spike addition is one which will double the amount of lead in the targeted sample split.

- 8.2.2 Add 10 mL of 1:1 nitric acid:water to each beaker, gently swirl to mix, and cover with a watch glass. Gently heat the sample to 85 to 100°C and reflux for 10 to 15 min without boiling. Monitor the temperature by having a thermometer inside a beaker or flask containing a small volume of water on the hot plate.
- 8.2.3 Allow the sample to cool, add 5 mL of concentrated nitric acid, replace the watch glass, and reflux for 30 min without boiling.
- 8.2.4 Allow the sample to cool, add 5 mL of concentrated nitric acid, replace the watch glass, and reflux for 30 min without boiling. (This is a repeat of 8.2.3 to ensure complete oxidation.)
- 8.2.5 Remove the watch glass and allow the solution to evaporate to approximately 5 mL without boiling, while maintaining a covering of solution over the bottom of the beaker. When removing watch glasses, care must be exercised to avoid losses by rinsing them with a minimum amount of water (rinsed into the sample beaker) and avoiding contamination by placing them upside down on new clean laboratory wipes. During this step, minimize the activity in the hood area and increase the separation between beakers. Allow the sample to cool after evaporation to approximately 5 mL.
- 8.2.6 After the sample has cooled, add 2 mL of water and 3 mL of 30 % hydrogen peroxide. Cover the beaker with a watch glass and return the covered beaker to the hot plate for warming and to start the peroxide reaction. Care must be taken during heating to ensure that losses do not occur due to excessively vigorous effervescence. Heat until effervescence subsides and cool the beaker.
- 8.2.7 Continue to add 30 % hydrogen peroxide in 1 mL aliquots with warming until the effervescence is minimal or until the general sample appearance is unchanged. Do not add more than a total of 10 mL of 30 % hydrogen peroxide even if effervescence has not been reduced to a minimal level.
- 8.2.8 Remove the watch glass and continue heating the acid-peroxide digestate carefully until the volume has been reduced to approximately 5 mL. Follow the same precautions for sample evaporation as outlined in 8.2.5.
- 8.2.9 Allow the digestates to cool, rinse the beaker walls and bottom of the watch glass with water, and quantitatively transfer to a 100 mL volumetric flask. Dilute to volume with water.

8.2.10 Remove any particulate in the digestate by filtration, by centrifugation, or by allowing the sample to settle prior to instrumental measurement. The diluted digestate solution contains approximately 5% (v/v) nitric acid. Use this level of nitric acid for calibration standards during instrumental measurement.

9. Quality Assurance

- 9.1 *Quality Control Samples*—Quality control samples to process with each batch of samples are summarized in Table 1.
- 9.1.1 *Reagent Blanks*—Carry reagent blanks (water and reagents) throughout the entire sample preparation and analytical process to determine if the samples are being contaminated. Process method blanks according to the frequency listed in Table 1.
- 9.1.2 Non-Spiked Samples, Spiked Samples and Spiked Duplicate Samples—Process these samples on a routine basis to estimate method accuracy on the sample batch, expressed as percent recovery relative to the true spiked value. Choose samples to use for spikes and spike duplicates at random. After sample pre-treatment (see 8.1) steps are completed, use three portions of the selected sample (one portion for the non-spike, one portion for the spike, and one portion for the spike duplicate sample). Proportion the selected samples using the splitting procedure listed below. Process spiked samples according to the frequency listed in Table 1.
- 9.1.3 Procedures to Split Samples for Non-Spiked Samples, Spiked Samples, Spiked Duplicate Samples, and Duplicate Samples—Use the following procedure to generate sample portions:
- 9.1.3.1 To mix the dried, sieved, and homogenized soil sample stored inside the air-tight sample containers, turn or roll the container repeatedly for about one min.
- 9.1.3.2 To split the homogenized soil sample stored inside the sample container, transfer the entire homogenized soil sample to a flat dry acid cleaned glass or plastic plate. (Use of a new piece of plastic wrap or wax paper between the plate and sample will help reduce cross-contamination.) Form the soil into a conical pile using a clean plastic spoon and flatten the pile to a uniform thickness and diameter by pressing down the apex of the pile. Divide the flattened pile into quarters using a plastic spoon. Using the spoon, scoop two opposite quarters of the sample to be spiked into separate labeled sample containers. Scoop the third quarter for the non-spiked sample into a

TABLE 1 Quality Control (QC) Samples

QC Samples	Definition	Frequency
Reagent blank	Type I Water—Digest as a sample with addition of all reagents. Reflects the maximum treatment given any one sample within the batch.	1 per 20 samples, a minimum of 1 per batch
Duplicate sample	A second portion of a homogenized sample digested in the same batch as the first portion.	1 per 20 samples, a minimum of 1 per batch
Spiked sample	A portion of a homogenized sample is spiked with lead before sample digestion.	1 per 20 samples, a minimum of 1 per batch
Spiked duplicate sample	A portion of the same homogenized sample used for the spiked sample is spiked with lead before sample digestion.	1 per 20 samples, a minimum of 1 per batch
Certified reference material (standard reference)	See 3.1.7. Use a certified lead level as close as practical to the appropriate federal, state, or local action level for lead in soil.	1 per batch of samples

separate labeled sample container and store the fourth quarter in another labeled sample container for permanent storage (if needed).

- 9.1.3.3 To alleviate cross-contamination, discard the piece of plastic wrap or wax paper (if used) and clean the glass plate with acetone and a laboratory wipe following each sample. Do not clean plastic plates with acetone.
- 9.1.4 *Duplicate Samples*—Process duplicate samples on a routine basis to determine an estimate of homogenization and analysis precision, expressed as percent relative range. Process duplicate samples according to the frequency listed in Table 1.
- 9.1.5 Certified Reference Materials—Process certified reference materials (CRMs) on a routine basis to determine an estimate of method accuracy on the sample batch, expressed as percent recovery relative to the certified value. Incorporate CRMs into each analytical batch according to the frequency listed in Table 1. Use a CRM which has a matrix similar or identical to soil with a certified lead concentration level.
- 9.2 Laboratory Records—Record all information regarding the preparation of samples (both QC samples and those submitted to the analyst) in a laboratory notebook as described below.

- 9.2.1 Record all reagent sources (lot numbers) used for sample preparation in a laboratory notebook. Record any inadvertent deviations, unusual happenings, or observations on a real-time basis as samples are processed. Use these records to add supplemental lead data when reporting results.
- 9.2.2 Laboratory notebooks shall be bound with prenumbered pages. All entries on sample data forms and laboratory notebooks shall be made using ink with a signature and date of entry. Any entry errors shall be corrected by using only a single line through the incorrect entry (no scratch outs) accompanied by the initials of the person making the correction and the date of correction. The use of electronic lab notebooks is also acceptable as long as similar security and record control precautions are in place. (See Note 11.)

Note 11—These procedures are important for proper documentation and traceability of laboratory data.

10. Keywords

10.1 digestion; lead; sample preparation; soils

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; http://www.copyright.com/