



Standard Practice for Statistical Treatment of Thermoanalytical Data¹

This standard is issued under the fixed designation E1970; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reappraisal. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reappraisal.

1. Scope*

1.1 This practice details the statistical data treatment used in some thermal analysis methods.

1.2 The method describes the commonly encountered statistical tools of the mean, standard deviation, relative standard deviation, pooled standard deviation, pooled relative standard deviation, the best fit to a (linear regression of a) straight line, and propagation of uncertainties for all calculations encountered in thermal analysis methods (see Practice E2586).

1.3 Some thermal analysis methods derive the analytical value from the slope or intercept of a linear regression straight line assigned to three or more sets of data pairs. Such methods may require an estimation of the precision in the determined slope or intercept. The determination of this precision is not a common statistical tool. This practice details the process for obtaining such information about precision.

1.4 There are no ISO methods equivalent to this practice.

2. Referenced Documents

2.1 ASTM Standards:²

E177 Practice for Use of the Terms Precision and Bias in ASTM Test Methods

E456 Terminology Relating to Quality and Statistics

E691 Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method

E2161 Terminology Relating to Performance Validation in Thermal Analysis and Rheology

E2586 Practice for Calculating and Using Basic Statistics

F1469 Guide for Conducting a Repeatability and Reproducibility Study on Test Equipment for Nondestructive Testing

¹ This practice is under the jurisdiction of ASTM Committee E37 on Thermal Measurements and is the direct responsibility of Subcommittee E37.10 on Fundamental, Statistical and Mechanical Properties.

Current edition approved April 1, 2016. Published April 2016. Originally approved in 1998. Last previous edition approved in 2011 as E1970 – 11. DOI: 10.1520/E1970-16.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For Annual Book of ASTM Standards volume information, refer to the standard’s Document Summary page on the ASTM website.

3. Terminology

3.1 *Definitions*—The technical terms used in this practice are defined in Practice E177 and Terminologies E456 and E2161 including *precision*, *relative standard deviation*, *repeatability*, *reproducibility*, *slope*, *standard deviation*, *thermoanalytical*, and *variance*.

3.2 Symbols (1):³

- m = slope
- b = intercept
- n = number of data sets (that is, x_i, y_i)
- x_i = an individual independent variable observation
- y_i = an individual dependent variable observation
- Σ = mathematical operation which means “the sum of all” for the term(s) following the operator
- \bar{X} = mean value
- s = standard deviation
- s_{pooled} = pooled standard deviation
- s_b = standard deviation of the line intercept
- s_m = standard deviation of the slope of a line
- s_y = standard deviation of Y values
- RSD = relative standard deviation
- δy_i = variance in y parameter
- r = correlation coefficient
- R = gage reproducibility and repeatability (see Guide F1469) an estimation of the combined variation of repeatability and reproducibility (2)
- s_r = within laboratory repeatability standard deviation (see Practice E691)
- s_R = between laboratory repeatability standard deviation (see Practice E691)
- s_i = standard deviation of the “ i th” measurement

4. Summary of Practice

4.1 The result of a series of replicate measurements of a value are typically reported as the mean value plus some estimation of the precision in the mean value. The standard deviation is the most commonly encountered tool for estimating precision, but other tools, such as relative standard deviation or pooled standard deviation, also may be encountered in specific thermoanalytical test methods. This practice describes

³ The boldface numbers in parentheses refer to a list of references at the end of this standard.

*A Summary of Changes section appears at the end of this standard

the mathematical process of achieving mean value, standard deviation, relative standard deviation and pooled standard deviation.

4.2 In some thermal analysis experiments, a linear or a straight line, response is assumed and desired values are obtained from the slope or intercept of the straight line through the experimental data. In any practical experiment, however, there will be some uncertainty in the data so that results are scattered about such a straight line. The linear regression (also known as “least squares”) method is an objective tool for determining the “best fit” straight line drawn through a set of experimental results and for obtaining information concerning the precision of determined values.

4.2.1 For the purposes of this practice, it is assumed that the physical behavior, which the experimental results approximate, are linear with respect to the controlled value, and may be represented by the algebraic function:

$$y = mx + b \quad (1)$$

4.2.2 Experimental results are gathered in pairs, that is, for every corresponding x_i (controlled) value, there is a corresponding y_i (response) value.

4.2.3 The best fit (linear regression) approach assumes that all x_i values are exact and the y_i values (only) are subject to uncertainty.

NOTE 1—In experimental practice, both x and y values are subject to uncertainty. If the uncertainty in x_i and y_i are of the same relative order of magnitude, other more elaborate fitting methods should be considered. For many sets of data, however, the results obtained by use of the assumption of exact values for the x_i data constitute such a close approximation to those obtained by the more elaborate methods that the extra work and additional complexity of the latter is hardly justified (2 and 3).

4.2.4 The best fit approach seeks a straight line, which minimizes the uncertainty in the y_i value.

4.3 The law of propagation of uncertainties is a tool for estimating the precision in a determined value from the sum of the variance of the respective measurements from which that value is derived weighted by the square of their respective sensitivity coefficients.

4.3.1 Variance is the square of the standard deviation(s). Conversely the standard deviation is the positive square root of the variance.

4.3.2 The sensitivity coefficient is the partial derivative of the function with respect to the individual variable.

5. Significance and Use

5.1 The standard deviation, or one of its derivatives, such as relative standard deviation or pooled standard deviation, derived from this practice, provides an estimate of precision in a measured value. Such results are ordinarily expressed as the mean value \pm the standard deviation, that is, $X \pm s$.

5.2 If the measured values are, in the statistical sense, “normally” distributed about their mean, then the meaning of the standard deviation is that there is a 67 % chance, that is 2 in 3, that a given value will lie within the range of \pm one standard deviation of the mean value. Similarly, there is a 95 % chance, that is 19 in 20, that a given value will lie within the range of \pm two standard deviations of the mean. The two

standard deviation range is sometimes used as a test for outlying measurements.

5.3 The calculation of precision in the slope and intercept of a line, derived from experimental data, commonly is required in the determination of kinetic parameters, vapor pressure or enthalpy of vaporization. This practice describes how to obtain these and other statistically derived values associated with measurements by thermal analysis.

6. Calculation

6.1 Commonly encountered statistical results in thermal analysis are obtained in the following manner.

NOTE 2—In the calculation of intermediate or final results, all available figures shall be retained with any rounding to take place only at the expression of the final results according to specific instructions or to be consistent with the precision and bias statement.

6.1.1 The mean value (X) is given by:

$$X = \frac{x_1 + x_2 + x_3 + \dots + x_i}{n} = \frac{\sum x_i}{n} \quad (2)$$

6.1.2 The standard deviation (s) is given by:

$$s = \left[\frac{\sum (x_i - X)^2}{(n - 1)} \right]^{1/2} \quad (3)$$

6.1.3 The relative standard deviation (RSD) is given by:

$$RSD = (s \cdot 100\%) / X \quad (4)$$

6.1.4 The pooled standard deviation (s_p) is given by:

$$= \left[\frac{\sum \{(n_i - 1) \cdot s_i^2\}}{\sum (n_i - 1)} \right]^{1/2} \quad (5)$$

NOTE 3—For the calculation of pooled relative standard deviation, the values of s_i are replaced by RSD_i .

6.1.5 The gage repeatability and reproducibility (R) is given by:

$$R = [s_R^2 + s_r^2]^{1/2} \quad (6)$$

NOTE 4—For the calculation of relative gage repeatability and reproducibility, the values of s_r and s_R are replaced with RSD_r and RSD_R .

6.2 *Linear Regression (Best) Fit Straight Line:*

6.2.1 The slope (m) is given by:

$$m = \frac{n \sum (x_i y_i) - (\sum x_i) (\sum y_i)}{n \sum x_i^2 - (\sum x_i)^2} \quad (7)$$

6.2.2 The intercept (b) is given by:

$$b = \frac{(\sum x_i^2) (\sum y_i) - (\sum x_i) (\sum x_i y_i)}{n \sum x_i^2 - (\sum x_i)^2} \quad (8)$$

6.2.3 The individual dependent parameter variance (δy_i) of the dependent variable (y_i) is given by:

$$\delta y_i = y_i - (mx_i + b) \quad (9)$$

6.2.4 The standard deviation s_y of the set of y values is given by:

$$s_y = \left[\frac{\sum (\delta y_i)^2}{n - 2} \right]^{1/2} \quad (10)$$

6.2.5 The standard deviation (s_m) of the slope is given by:

$$s_m = s_y \left[\frac{n}{n \sum x_i^2 - (\sum x_i)^2} \right]^{1/2} \quad (11)$$

6.2.6 The standard deviation (s_b) of the intercept (b) is given by:

$$s_b = s_y \left[\frac{\sum x_i^2}{n \sum x_i^2 - (\sum x_i)^2} \right]^{1/2} \quad (12)$$

6.2.7 The denominators in Eq 7, Eq 8, Eq 11, and Eq 12 are the same. It is convenient to obtain the denominator (D) as a separate function for use in manual calculation of each of these equations.

$$D = n \sum x_i^2 - (\sum x_i)^2 \quad (13)$$

6.2.8 The linear correlation coefficient (r), a measure of the mutual dependence between paired x and y values, is given by:

$$r = \frac{n \sum xy - (\sum x_i)(\sum y_i)}{[n \sum x_i^2 - (\sum x_i)^2]^{1/2} [n \sum y_i^2 - (\sum y_i)^2]^{1/2}} \quad (14)$$

NOTE 5— r may vary from -1 to $+1$, where values of $+1$ or -1 indicate perfect (100 %) correlation and 0 indicates no (0 %) correlation, that is, random scatter. A positive (+) value indicates a positive slope and a negative (–) indicates a negative slope.

6.3 Propagation of Uncertainties:

6.3.1 The law of propagation of uncertainties, neglecting the cross terms, is given by:

$$s_z^2 = \sum [(\partial z / \partial i) s_i]^2 \quad (15)$$

or

$$s_z = \{ \sum [(\partial z / \partial i) s_i]^2 \}^{1/2} \quad (16)$$

6.3.2 For example, given the function $z = a d / c$, then the sensitivity coefficient for a is $\partial z / \partial a = d / c$, for d is $\partial z / \partial d = a / c$, and for c is $\partial z / \partial c = -ad / c^2$.

6.3.3 Eq 16 becomes:

$$s_z = \{ [(\partial z / \partial a) s_a]^2 + [(\partial z / \partial d) s_d]^2 + [(\partial z / \partial c) s_c]^2 \}^{1/2} \quad (17)$$

or

$$s_z = \{ (d s_a / c)^2 + (a s_d / b)^2 + (a d s_c / c^2)^2 \}^{1/2}$$

6.3.4 Dividing both sides of the equation by $z = a d / c$, yields:

$$s_z / z = \{ (s_a / a)^2 + (s_d / d)^2 + (s_c / c)^2 \}^{1/2} \quad (18)$$

6.3.5 The form of Eq 17 has been determined for a number of functions and is presented in Table 1.

6.4 Example Calculations:

6.4.1 Table 2 provides an example set of data and intermediate calculations which may be used to examine the manual calculation of slope (m) and its standard deviation (s_m) and of the intercept (b) and its standard deviation (s_b).

6.4.1.1 The values in Columns A and B are experimental parameters with x_i being the independent parameter and y_i the dependent parameter.

6.4.1.2 From the individual values of x_i and y_i in Columns A and B in Table 2, the values for x_i^2 and $x_i y_i$ are calculated and placed in Columns C and D.

6.4.1.3 The values in columns A, B, C, and D are summed (added) to obtain $\sum x_i = 76.0$, $\sum y_i = 86.7$, $\sum x_i^2 = 1540.0$, and $\sum x_i y_i = 1753.9$, respectively.

6.4.1.4 The denominator (D) is calculated using Eq 13 and the values $\sum x_i^2 = 1540.0$ and $\sum x_i = 76.0$ from 6.4.1.3.

$$D = (6 \cdot 1540.0) - (76.0 \cdot 76.0) = 3464.0 \quad (26)$$

6.4.1.5 The value for m is calculated using the values $n = 6$, $\sum x_i \cdot y_i = 1753.9$, $\sum x_i = 76.0$, $\sum y_i = 86.7$, and $D = 3640.0$, from 6.4.1.3 and 6.4.1.4 and Eq 7:

$$m = \frac{n \sum (x_i y_i) - \sum x_i \sum y_i}{D} \quad (27)$$

$$m = \frac{(6 \cdot 1753.9) - (76.0 \cdot 86.7)}{3464.0} = \frac{10523.4 - 6589.2}{3464.0} \quad (28)$$

$$= 1.1357$$

6.4.1.6 The value for b is calculated using the values $n = 6$, $\sum x_i \cdot y_i = 1753.9$, $\sum x_i = 76.0$, and $\sum y_i = 86.7$, from 6.4.1.3 and 6.4.1.4 and Eq 8:

$$b = \frac{(1540.0 \cdot 86.7) - (76.0 \cdot 1753.9)}{3464.0} = \frac{133518.0 - 133296.4}{3464.0} \quad (29)$$

$$= 0.064$$

6.4.1.7 Using the values for $m = 1.1357$ and $b = 0.064$ from 6.4.1.5 and 6.4.1.6, and the value $\sum x_i = 76.0$ from Table 2, the $n = 6$, values for δy_i are calculated values using Eq 9 and recorded in Column F in Table 2.

6.4.1.8 From the values in Column F of Table 2, the six values for $(\delta y_i)^2$ are calculated and recorded in Column G.

TABLE 1 Uncertainties (4)

Description	Example	Uncertainty
Addition or Subtraction	$z = a + d - c$	$s_z = [(s_a)^2 + (s_d)^2 + (s_c)^2]^{1/2} \quad (19)$
Multiplication or Division	$z = a d / c$	$s_z = [(s_a / a)^2 + (s_d / d)^2 + (s_c / c)^2]^{1/2} \quad (20)$
Exponential	$z = a^x$	$s_z / z = x(s_a / a) \quad (21)$
Logarithmic	$z = \log_{10} a$	$s_z = 0.434 s_a / a \quad (22)$
	$z = \ln a$	$s_z = s_a / a \quad (23)$
Antilogarithm	$z = 10^a$	$s_z / z = 2.303 s_a \quad (24)$
	$z = e^a$	$s_z / z = s_a \quad (25)$

TABLE 2 Example Set of Data and Intermediate Calculations ($n = 6$)

Column Experiment	A x_i	B y_i	C x_i^2	D $x_i y_i$	E $m x_i + b$	F δy_i	G $(\delta y_i)^2$	H $(y_i)^2$
1	1.0	1.2	1.0	1.2	1.1997	0.0003	0.000 000 09	1.44
2	1.0	1.3	1.0	1.3	1.1997	0.1003	0.010 060 09	1.69
3	12.0	13.7	144.0	164.0	13.6924	0.0076	0.000 057 76	187.69
4	12.0	13.5	144.0	162.0	13.6924	-0.1924	0.037 017 76	182.25
5	25.0	28.5	625.0	712.5	28.4565	0.0435	0.001 892 25	812.25
6	25.0	28.5	625.0	712.5	28.4565	0.0435	0.001 892 25	812.25
Σ	76.0	86.7	1540.0	1753.9			0.050 920 20	1997.57

6.4.1.9 The values in Column G of **Table 2** are summed to obtain $\Sigma (\delta y_i)^2$.

6.4.1.10 The value of s_y is calculated using the value from 6.4.1.9 and Eq 10:

$$s_y = [0.050\ 092\ 02/4]^{1/2} = 0.1119 \quad (30)$$

6.4.1.11 The value for s_m (expressed to two significant figures) is calculated using the values of $D = 3464.0$ and $s_y = 0.1119$ from 6.4.1.4 and 6.4.1.10, respectively.

$$s_m = 0.1119 \left[\frac{6}{3464.0} \right]^{1/2} = 0.0047 \quad (31)$$

6.4.1.12 The value for s_b (expressed to two significant figures) is calculated using the values of Σx_i^2 , $D = 3464.0$, and $s_y = 0.119$, from 6.4.1.3, 6.4.1.4, and 6.4.1.10, respectively.

$$s_b = 0.1119 \left[\frac{1540.0}{3464.0} \right]^{1/2} = 0.075 \quad (32)$$

6.4.1.13 The value of the slope along with its estimation of precision is obtained from 6.4.1.5 and 6.4.1.11 and reported as follows:

$$m \pm s_m \quad (33)$$

$$m = 1.1357 \pm 0.0047 \quad (34)$$

6.4.2 **Table 2** provides an example set of data that may be used to examine the manual calculation of the correlation coefficient (r).

6.4.2.1 The value of r is calculated using the values $n = 6$, $\Sigma x_i = 76.0$, $\Sigma y_i = 86.7$, $\Sigma x_i^2 = 1540.0$, $\Sigma x_i y_i = 1753.9$, and $\Sigma (y_i)^2 = 1997.57$ from **Table 2** and **Eq 14**.

$$r = \frac{\{(6 \cdot 1753.9) - (76.0 \cdot 86.7)\}}{\{[(6 \cdot 1540.0) - (76.0 \cdot 76.0)]^{1/2} \cdot [(6 \cdot 1997.57) - (86.7 \cdot 86.7)]^{1/2}\}} \quad (35)$$

$$= \frac{\{10523.4 - 6589.2\}}{\{[9240 - 5776]^{1/2} \cdot [11985.42 - 7516.89]^{1/2}\}}$$

$$= \frac{3934.2}{\{[3464]^{1/2} \cdot [4468.53]^{1/2}\}}$$

$$= \frac{3934.2}{\{58.856 \cdot 66.847\}}$$

$$= 0.99996$$

6.4.3 Thermal conductivity (λ) is determined by the flash method using the equation $\lambda = \rho c_p a$. One worker (5) provides values of $\rho = 8.340 \pm 0.04 \text{ g (cm)}^{-3}$, $c_p = 0.444 \pm 0.009 \text{ J g}^{-1} \text{ K}^{-1}$ and $a = 3.428 \pm 0.09 \text{ (mm)}^2 \text{ s}^{-1}$.

6.4.3.1 Since λ is of the form $z = a d c$, the value of s_z is calculated using the values from 6.4.3 and **Eq 18**.

$$s_\lambda = [(0.04 / 8.340)^2 + (0.009 / 0.444)^2 + (0.09 / 3.428)^2]^{1/2} \quad (36)$$

$$= [(0.48 \%)^2 + (2.0 \%)^2 + 2.6 \%)^2]^{1/2}$$

$$= [0.230 + 4.00 + 6.76]^{1/2} \%$$

$$= [10.96]^{1/2} \%$$

$$= 3.3 \%$$

7. Report

7.1 Report the following information:

7.1.1 All of the statistical values required to meet the needs of the respective applications method.

7.1.2 The specific dated version of this practice that is used.

8. Keywords

8.1 best fit; error; intercept; linear regression; mean; precision; propagation of uncertainties; relative standard deviation; slope; standard deviation; variance; uncertainty

REFERENCES

- (1) Taylor, J. K., *Handbook for SRM Users, Publication 260-100*, National Institute of Standards and Technology, Gaithersburg, MD, 1993.
- (2) *Measurement System Analysis*, third edition, Automotive Industry Action Group, Southfield, MI, 2003, pp. 55, 177–184.
- (3) Mandel, J., *The Statistical Analysis of Experimental Data*, Dover Publications, New York, NY, 1964.
- (4) Skoog, D. A., et al., “Standard Deviation of Calculated Results,” in *Fundamentals of Analytical Chemistry*, Thomson Asia Pte Ltd, Singapore. 2004, pp. 127–133.
- (5) Blumm, J., A. Lindemann, and B. Niedrig, “Measurement of the thermophysical properties of an NPL thermal conductivity standard Inconel 600,” *High Temperatures — High Pressures*, Vol 35/36, 2003/2007, pp. 621–626.

SUMMARY OF CHANGES

Committee E37 has identified the location of selected changes to this standard since the last issue (E1970 – 11) that may impact the use of this standard. (Approved April 1, 2016.)

- (1) Editorial changes to 1.2, 1.3, 2.1, 4.2.3, 4.2.4, 6.2, and Section 8. (2) Addition of 6.3, Table 1, 6.4.3, and 6.4.3.1.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; <http://www.copyright.com/>