



Standard Practice for Calculating Heat of Vaporization or Sublimation from Vapor Pressure Data¹

This standard is issued under the fixed designation E2071; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This practice describes the calculation of the heat of vaporization of a liquid or the heat of sublimation of a solid from measured vapor pressure data. It is applicable to pure liquids, azeotropes, pure solids, and homogenous solid solutions over the temperature range for which the vapor pressure equation fitted to the measured data is applicable.

NOTE 1—This practice is generally not applicable to liquid mixtures. For a pure liquid or azeotrope, composition does not change upon vaporization so that the integral heat of vaporization is identical to the differential heat of vaporization. Non-azeotropic liquid mixtures change composition upon vaporizing. Heat of vaporization data computed from this practice for a liquid mixture are valid only as an approximation to the mixture differential heat of vaporization; it is not a valid approximation to the mixture integral heat of vaporization.

1.2 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.3 There is no ISO standard equivalent to this practice.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 ASTM Standards:²

[D2879 Test Method for Vapor Pressure-Temperature Relationship and Initial Decomposition Temperature of Liquids by Isotenoscope](#)

[E1142 Terminology Relating to Thermophysical Properties](#)

¹ This practice is under the jurisdiction of Committee E37 on Thermal Measurements and is the direct responsibility of Subcommittee E37.10 on Fundamental, Statistical and Mechanical Properties.

Current edition approved May 1, 2015. Published May 2015. Originally approved in 2000. Last previous edition approved in 2010 as E2071 – 00 (2010). DOI: 510.1520/E2071-00R15.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

[E1194 Test Method for Vapor Pressure \(Withdrawn 2013\)](#)³

[E1719 Test Method for Vapor Pressure of Liquids by Ebulliometry](#)

[E1782 Test Method for Determining Vapor Pressure by Thermal Analysis](#)

3. Terminology

3.1 Symbols:

3.1.1 A, B, C —Antoine vapor pressure equation constants (\log_{10} , kPa, K), Antoine vapor pressure equation:

$$\log_{10}P = A - B/(T + C)$$

3.1.2 P —vapor pressure, kPa.

3.1.3 P_c —critical pressure, kPa.

3.1.4 P_r —reduced pressure = P/P_c .

3.1.5 T —absolute temperature, K.

3.1.6 T_c —critical temperature, K.

3.1.7 T_r —reduced temperature = T/T_c .

3.1.8 V —molar volume, cm^3/mol .

3.1.9 R —gas constant, 8.31433 J/mol-K; 8314330 kPa-cm³/mol-K.

3.1.10 ΔH_v —heat of vaporization, J/mol.

3.1.11 ΔZ_v —difference in compressibility factor ($Z = PV/RT$) upon vaporization. Clapeyron equation:

$$\Delta H_v = -R\Delta Z_v [d(\ln P)/d(1/T)]$$

3.1.11.1 *Discussion*—The subscript “V” will be used throughout this practice to designate the vaporization of a liquid. If the vapor pressure data were measured for a solid, substitute the subscript “S” for the sublimation of a solid.

3.2 Definitions:

3.2.1 Specialized terms used in this practice are defined in Terminology [E1142](#).

3.2.2 *sublimation*—transition from a solid phase to a gaseous phase.

3.2.3 *vaporization*—transition from a liquid phase to a gaseous phase.

³ The last approved version of this historical standard is referenced on www.astm.org.

4. Summary of Practice

4.1 Vapor pressure data are measured by other referenced ASTM standards and then correlated with the Antoine equation. The heat of vaporization or sublimation is computed at the desired temperature from the vapor-pressure temperature derivative from the fitted Antoine equation by use of the Clapeyron equation (1).⁴ In the Clapeyron equation, ΔZ_V is determined by either the Clausius-Clapeyron(2) approximation:

$$(\Delta Z_V = 1)$$

or the Haggemacher (3) approximation:

$$\left(\Delta Z_V = \left\{ 1 - \left[P_r / (T_r)^3 \right] \right\} \frac{1}{2} \right)$$

4.2 An example calculation is given in [Annex A1](#).

5. Significance and Use

5.1 If the heat of vaporization or sublimation is absorbed or liberated in a process at constant pressure, it is called enthalpy of vaporization or sublimation. Enthalpy of vaporization or sublimation is a fundamental thermodynamic property of a liquid or solid. It is an important quantity in the design of heat exchangers and other chemical process units. Enthalpy of vaporization is also used to calculate solubility parameters (4).

5.2 This practice may be used in research, regulatory compliance, and quality assurance applications.

6. Experimental Vapor Pressure Data

6.1 Vapor pressure data are measured by Test Methods [D2879](#), [E1194](#), [E1719](#), or [E1782](#). Note the safety precautions contained in the test method used.

6.1.1 Vapor pressure data from other reliable sources, for example, peer-review technical journals, may be used. The source of the vapor pressure data must be noted.

6.2 The measured vapor pressure data are fitted to an Antoine vapor pressure equation. See 10.3 in Test Method [E1719](#) for details on least-squares regression of vapor pressure data.

⁴ The boldface numbers given in parentheses refer to a list of references at the end of the text.

7. Calculation

7.1 At each temperature of interest, calculate the vapor pressure from the Antoine equation and calculate the vapor-pressure temperature derivative from the fitted Antoine equation constants from:

$$[d(\ln P)/d(1/T)] = -2.3025851[BT^2/(T+C)^2]$$

7.2 Calculate an approximation to ΔZ_V at each temperature.

7.2.1 The Clausius-Clapeyron approximation to ΔZ_V is:

$$\Delta Z_V = 1.0$$

7.2.2 The Haggemacher approximation to ΔZ_V is:

$$\Delta Z_V = \left\{ 1 - \left[P_r / (T_r)^3 \right] \right\} \frac{1}{2}$$

NOTE 2—The Clausius-Clapeyron approximation is generally used for solids and for liquids at low T_r . The Haggemacher approximation is generally used for liquids up to $T_r \approx 0.75$.

7.2.3 If equation of state (Z) data are available for both the condensed and gaseous phases, ΔZ_V may be calculated directly from the equation of state data.

7.3 Calculate the heat of vaporization or heat of sublimation at each temperature from the Clapeyron equation:

$$\Delta H_V = -R\Delta Z_V [d(\ln P)/d(1/T)]$$

8. Report

8.1 Report the following information:

8.1.1 The test method and source of the vapor pressure data used in the heat of vaporization or heat of sublimation calculation. A vapor pressure data table shall also be reported.

8.1.2 The Antoine equation constants fitted to the vapor pressure data.

8.1.3 The approximation to ΔZ_V used in the calculation.

8.1.4 The values and source of the critical temperature and critical pressure data if the Haggemacher approximation was used for ΔZ_V .

8.1.5 A table that contains temperature, vapor pressure, the vapor pressure temperature derivative $[d(\ln P)/d(1/T)]$, difference in compressibility factor (ΔZ_V), and ΔH_V , the heat of vaporization or heat of sublimation.

8.1.6 The specific dated version of this practice used.

8.2 See the sample calculations and report in [Annex A1](#).

9. Keywords

9.1 Antoine equation; Clausius-Clapeyron equation; enthalpy of sublimation; enthalpy of vaporization; Haggemacher equation; heat of sublimation; heat of vaporization; vapor pressure

ANNEX
(Mandatory Information)
A1. SAMPLE CALCULATIONS AND REPORT
A1.1 Source of Sample Vapor Pressure Data

A1.1.1 This sample calculation is performed on the sample vapor pressure data given for a toluene specimen in Annex A3 of Test Method **E1719**. Heat of vaporization is calculated in 10 K increments between 290 and 400 K. Calculations for both the Clausius-Clapeyron and Haggemacher approximations to ΔZ_V are listed.

A1.2 Sample Experimental Data

A1.2.1 These controlled pressure-boiling temperature data pairs were measured by Test Method **E1719** on a 75 cm³ specimen charged to a vapor-lift pump ebulliometer:

P (kPa)	T (K)
10.0	318.4
20.0	335.4
30.0	345.8
50.0	360.7
70.0	371.2
85.0	377.9
100.0	383.3

A1.2.2 A non-linear least-squares fit of the Antoine equation, $\log_{10}P = A - B/(T + C)$, produced these constants:

$$\begin{aligned} A \text{ (fit)} &= 6.168057 \\ B \text{ (fit)} &= 1397.23 \\ C \text{ (fit)} &= -48.10 \end{aligned}$$

A1.3 Sample Calculation

A1.3.1 The critical temperature and pressure for toluene (5) are:

$$\begin{aligned} T_c &= 591.75 \text{ K} \\ P_c &= 4108.69 \text{ kPa} \end{aligned}$$

At 290 K:

$$\begin{aligned} T_r &= 0.490071821 \\ P_r &= 0.000600191 \end{aligned}$$

$$\begin{aligned} \text{Vapor pressure} &= 10^{\left[6.168057 - \frac{1397.23}{(290 - 48.10)}\right]} \\ &= 2.465997 \text{ kPa} \end{aligned}$$

$$\begin{aligned} [d(\ln P)/d(1/T)] &= -2.3025851 \left[\frac{1397.23 \cdot 290^2}{(290 - 48.10)^2} \right] = \\ &= -4623.8938 \text{ K} \end{aligned}$$

A1.3.2 Haggemacher approximation to ΔZ_V :

$$\Delta Z_V = \left\{ 1 - \left[\frac{0.000600191}{(0.490071821)^3} \right] \right\}^{\frac{1}{2}} = 0.997447$$

A1.3.3 ΔH_V from Clausius-Clapeyron approximation:

$$\Delta H_V = (-8.31433) \cdot 1.00 \cdot (-4623.8938) = 38444.6 \text{ J/mol}$$

A1.3.4 ΔH_V from Haggemacher approximation:

$$\Delta H_V = (-8.31433) \cdot 0.997447 \cdot (-4623.8938) = 38346.4 \text{ J/mol}$$

A1.4 Sample Heat of Vaporization Report

A1.4.1 *Clausius-Clapeyron Approximation Report:*

A1.4.1.1 Data are for a toluene specimen and are listed in Annex A3 of Test Method **E1719**. These controlled pressure-

boiling temperature data pairs were measured by Test Method **E1719** on a 75 cm³ specimen charged to a vapor-lift pump ebulliometer:

P (kPa)	T (K)
10.0	318.4
20.0	335.4
30.0	345.8
50.0	360.7
70.0	371.2
85.0	377.9
100.0	383.3

A1.4.1.2 A non-linear least-squares fit of the Antoine equation:

$$\log_{10}P = A - B/(T + C)$$

produced these constants:

$$\begin{aligned} A \text{ (fit)} &= 6.168057 \\ B \text{ (fit)} &= 1397.23 \\ C \text{ (fit)} &= -48.10 \end{aligned}$$

A1.4.1.3 The Clausius Clapeyron approximation for ΔZ_V was used.

Temperature K	Pressure kPa	$[d(\ln P)/d(1/T)]$ K	ΔZ_V	ΔH_V J/mol
290	2.4659968	-4623.8938	1.00000000	38444.6
300	4.1811179	-4563.2028	1.00000000	37940.0
310	6.8089762	-4507.5026	1.00000000	37476.9
320	10.697757	-4456.2047	1.00000000	37050.4
330	16.277326	-4408.8094	1.00000000	36656.3
340	24.064868	-4364.8893	1.00000000	36291.1
350	34.668504	-4324.0774	1.00000000	35951.8
360	48.788774	-4286.0560	1.00000000	35635.7
370	67.217970	-4250.5496	1.00000000	35340.5
380	90.837442	-4217.3173	1.00000000	35064.2
390	120.61303	-4186.1482	1.00000000	34805.0
400	157.58889	-4156.8566	1.00000000	34561.5

A1.4.2 *Haggemacher Approximation Report:*

A1.4.2.1 Data are for a toluene specimen and are listed in Annex A3 of Test Method **E1719**. These controlled pressure-boiling temperature data pairs were measured by Test Method **E1719** on a 75 cm³ specimen charged to a vapor-lift pump ebulliometer:

P (kPa)	T (K)
10.0	318.4
20.0	335.4
30.0	345.8
50.0	360.7
70.0	371.2
85.0	377.9
100.0	383.3

A1.4.2.2 A non-linear least-squares fit of the Antoine equation:

$$\log_{10}P = A - B/(T + C)$$

produced these constants:

$$\begin{aligned} A \text{ (fit)} &= 6.168057 \\ B \text{ (fit)} &= 1397.23 \\ C \text{ (fit)} &= -48.10 \end{aligned}$$

A1.4.2.3 The Haggemacher approximation for ΔZ_V was used. The critical temperature and pressure used for toluene (5) are:

$$T_c = 591.75 \text{ K}$$

$$P_c = 4108.69 \text{ kPa}$$

Temperature K	Pressure kPa	$[d(\ln P)/d(1/T)]$ K	ΔZ_V	ΔH_V J/mol
290	2.4659968	-4623.8938	0.99744709	38346.4
300	4.1811179	-4563.2028	0.99608744	37791.5

Temperature K	Pressure kPa	$[d(\ln P)/d(1/T)]$ K	ΔZ_V	ΔH_V J/mol
310	6.8089762	-4507.5026	0.99421990	37260.2
320	10.697757	-4456.2047	0.99173347	36744.1
330	16.277326	-4408.8094	0.98851253	36235.2
340	24.064868	-4364.8893	0.98443961	35726.4
350	34.668504	-4324.0774	0.97939800	35211.1
360	48.788774	-4286.0560	0.97327384	34683.3
370	67.217970	-4250.5496	0.96595780	34137.4
380	90.837442	-4217.3173	0.95734617	33568.5
390	120.61303	-4186.1482	0.94734133	32972.2
400	157.58889	-4156.8566	0.93585171	32344.4

REFERENCES

- (1) Van Ness, H. C., and Abbott, M. M., *Classical Thermodynamics of Nonelectrolyte Solutions*, McGraw-Hill, New York, NY, 1982, pp. 96–100.
- (2) Van Ness, H. C., and Abbott, M. M., *Classical Thermodynamics of Nonelectrolyte Solutions*, McGraw-Hill, New York, NY, 1982, p. 100.
- (3) Haggemacher, J. E., *Journal of the American Chemical Society*, Vol 68, 1946.
- (4) Barton, A. F. M., *CRC Handbook of Solubility Parameters and Other Cohesion Parameters*, CRC Press, Boca Raton, FL, 1991.
- (5) Daubert, T. E., ed., *The DIPPR Project 801 Data Compilation*, Design Institute of Physical Property Data, AIChE, New York, NY, 1990, CAS#, 108–88–3.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; http://www.copyright.com/