



Designation: E415 – 17

Standard Test Method for Analysis of Carbon and Low-Alloy Steel by Spark Atomic Emission Spectrometry¹

This standard is issued under the fixed designation E415; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers the simultaneous determination of 21 alloying and residual elements in carbon and low-alloy steels by spark atomic emission vacuum spectrometry in the mass fraction ranges shown [Note 1](#).

Element	Composition Range, %	
	Applicable Range, Mass Fraction % ^A	Quantitative Range, Mass Fraction % ^B
Aluminum	0 to 0.093	0.006 to 0.093
Antimony	0 to 0.027	0.006 to 0.027
Arsenic	0 to 0.1	0.003 to 0.1
Boron	0 to 0.007	0.0004 to 0.007
Calcium	0 to 0.003	0.002 to 0.003
Carbon	0 to 1.1	0.02 to 1.1
Chromium	0 to 8.2	0.007 to 8.14
Cobalt	0 to 0.20	0.006 to 0.20
Copper	0 to 0.5	0.006 to 0.5
Manganese	0 to 2.0	0.03 to 2.0
Molybdenum	0 to 1.3	0.007 to 1.3
Nickel	0 to 5.0	0.006 to 5.0
Niobium	0 to 0.12	0.003 to 0.12
Nitrogen	0 to 0.015	0.01 to 0.055
Phosphorous	0 to 0.085	0.006 to 0.085
Silicon	0 to 1.54	0.02 to 1.54
Sulfur	0 to 0.055	0.001 to 0.055
Tin	0 to 0.061	0.005 to 0.061
Titanium	0 to 0.2	0.001 to 0.2
Vanadium	0 to 0.3	0.003 to 0.3
Zirconium	0 to 0.05	0.01 to 0.05

^A Applicable range in accordance with Guide [E1763](#) for results reported in accordance with Practice [E1950](#).

^B Quantitative range in accordance with Practice [E1601](#).

NOTE 1—The mass fraction ranges of the elements listed have been established through cooperative testing² of reference materials.

1.2 This test method covers analysis of specimens having a diameter adequate to overlap and seal the bore of the spark stand opening. The specimen thickness can vary significantly

¹ This test method is under the jurisdiction of ASTM Committee [E01](#) on Analytical Chemistry for Metals, Ores, and Related Materials and is the direct responsibility of Subcommittee [E01.01](#) on Iron, Steel, and Ferroalloys.

Current edition approved May 15, 2017. Published June 2017. Originally approved in 1971. Last previous edition approved in 2015 as E415 – 15. DOI: 10.1520/E0415-15.

² Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting Research Report RR:E01-1122. Contact ASTM Customer Service at service@astm.org.

according to the design of the spectrometer stand, but a thickness between 10 mm and 38 mm has been found to be most practical.

1.3 This test method covers the routine control analysis in iron and steelmaking operations and the analysis of processed material. It is designed for chill-cast, rolled, and forged specimens. Better performance is expected when reference materials and specimens are of similar metallurgical condition and composition. However, it is not required for all applications of this standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

1.5 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recommendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.*

2. Referenced Documents

2.1 ASTM Standards:³

- [E29 Practice for Using Significant Digits in Test Data to Determine Conformance with Specifications](#)
- [E135 Terminology Relating to Analytical Chemistry for Metals, Ores, and Related Materials](#)
- [E305 Practice for Establishing and Controlling Atomic Emission Spectrochemical Analytical Curves](#)
- [E350 Test Methods for Chemical Analysis of Carbon Steel, Low-Alloy Steel, Silicon Electrical Steel, Ingot Iron, and Wrought Iron](#)
- [E406 Practice for Using Controlled Atmospheres in Spectrochemical Analysis](#)
- [E1019 Test Methods for Determination of Carbon, Sulfur, Nitrogen, and Oxygen in Steel, Iron, Nickel, and Cobalt](#)

³ For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

- Alloys by Various Combustion and Fusion Techniques
- E1329 Practice for Verification and Use of Control Charts in Spectrochemical Analysis
- E1601 Practice for Conducting an Interlaboratory Study to Evaluate the Performance of an Analytical Method
- E1763 Guide for Interpretation and Use of Results from Interlaboratory Testing of Chemical Analysis Methods (Withdrawn 2015)⁴
- E1806 Practice for Sampling Steel and Iron for Determination of Chemical Composition
- E1950 Practice for Reporting Results from Methods of Chemical Analysis
- E2972 Guide for Production, Testing, and Value Assignment of In-House Reference Materials for Metals, Ores, and Other Related Materials

2.2 Other ASTM Documents

- ASTM MNL 7 Manual on Presentation of Data and Control Chart Analysis⁵

3. Terminology

3.1 For definitions of terms used in this test method, refer to Terminology E135.

4. Summary of Test Method

4.1 A capacitor discharge is produced between the flat, ground surface of the disk specimen and a conically shaped electrode. The discharge is terminated at a predetermined intensity time integral of a selected iron line, or at a predetermined time, and the relative radiant energies of the analytical lines are recorded. The most sensitive lines of arsenic, boron, carbon, nitrogen, phosphorus, sulfur, and tin lie in the vacuum ultraviolet region. The absorption of the radiation by air in this region is overcome by evacuating the spectrometer or by use of a vacuum ultraviolet (VUV) transparent gas and flushing the spark chamber with argon.

5. Significance and Use

5.1 This test method for the spectrometric analysis of metals and alloys is primarily intended to test such materials for compliance with compositional specifications. It is assumed that all who use this test method will be analysts capable of performing common laboratory procedures skillfully and safely. It is expected that work will be performed in a properly equipped laboratory.

6. Apparatus

6.1 Sampling Devices:

6.1.1 Refer to Practice E1806 for devices and practices to sample liquid and solid iron and steel.

6.2 *Excitation Source*, capable of providing electrical parameters to spark a sample. See 11.1 for details.

6.3 *Spark Chamber*, automatically flushed with argon. The spark chamber shall be mounted directly on the spectrometer

and shall be provided with a spark stand to hold a flat specimen and a lower counter electrode of rod form.

6.3.1 Follow the manufacturer's recommendations for cleaning the spark chamber. During continuous operation, this typically should be done every 24 h. Follow the manufacturer's recommendations for cleaning the entrance lens or window (verifier data or other reference sample intensity data can typically indicate when this is necessary).

6.4 *Spectral Lines*—Table 1 lists spectral lines and internal standards usable for carbon and low alloy steel. The spectrometer must be able to measure at least one of the listed spectral lines for each of the listed elements. Spectral lines other than those listed in Table 1 may be used provided it can be shown experimentally that equivalent precision and accuracy are obtained.

6.5 *Measuring System*, spectrometer capable of converting light intensities to measurable electrical signals. The measuring system may consist of one of the following configurations:

6.5.1 A photomultiplier (PMT) array having individual voltage adjustments, capacitors in which the output of each photomultiplier is stored, a voltage measuring system to register the voltages on the capacitors either directly or indirectly, and the necessary switching arrangements to provide the desired sequence of operation.

6.5.2 A semiconductor detector array (CCD or CMOS), pixel selection electronics to reset the pixels and to transport the voltage of an individual pixel to one or more output ports of the detector arrays, and a voltage measuring system to register the voltage of said output ports.

6.5.3 A hybrid design using both photomultipliers and semiconductor arrays.

6.6 *Optical Path*—If the instrument is operated using a VUV transparent gas, check the manufacturer's suggested gas purity. It may be necessary to have a gas purification system consisting of a circulation pump and a cleaning cartridge to keep the O₂ (g) residual <500 ng/g and H₂O (g) residual <1 µg/g and remove impurities of nitrogen and hydrocarbons. If the instrument is using a vacuum pump, it should be capable of maintaining a vacuum of 3.33 Pa (25 µm Hg) or less.

NOTE 2—A pump with a displacement of at least 0.23 m³/min (8 ft³/min) is usually adequate.

6.7 *Gas System*, consisting of an argon supply with pressure and flow regulation. Automatic sequencing shall be provided to actuate the flow at a given rate for a specific time interval. The flow rate may be manually or automatically set. The argon system shall be in accordance with Practice E406.

7. Reagents and Materials

7.1 *Counter Electrodes*—The counter electrodes can be silver or tungsten rods, or other material, provided it can be shown experimentally that equivalent precision and bias are obtained. The rods can vary in diameter from 1.5 mm to 6.5 mm (depending on the instrument design) and typically are machined to a 90° or 120° angled tip.

7.1.1 A black deposit will collect on the tip of the electrode. This deposit should be removed between specimens (typically with a wire brush). If not removed, it can reduce the overall

⁴The last approved version of this historical standard is referenced on www.astm.org.

⁵ASTM Manual Series, ASTM International, 8th edition, 2010.

TABLE 1 Internal Standard and Analytical Lines

Element	Wavelength, λ , nm	Line Classification ^A	Possible Interference ^B
Aluminum	396.15	I	Mo
	394.40	I	V, Mn, Mo, Ni
	308.22	I	V, Mn
Antimony	217.6	I	Ni, Nb, Mn, W
Arsenic	189.04	I	V, Cr
	197.20	I	Mo, W
	193.76	I	Mn
Boron	345.13	II	
	182.64	I	S, Mn, Mo
	182.59	I	W, Mn, Cu
Calcium	393.37	II	
	396.85	II	Nb
Carbon	165.81	I	Cr
	193.09	I	Al
Chromium	312.26	II	V
	313.21	II	
	425.44	I	
	298.92	II	Mn, V, Ni, Nb, Mo
	267.72	II	Mn, Mo, W
Cobalt	345.35	I	Cr, Mo
	228.62	II	Ni, Cr
	258.03	II	Fe, Mn, W
Copper	212.3	II	Si
	324.75	I	Mn, Nb
	327.40	I	Nb
	224.26	II	W, Ni
	213.60	II	Mo, Cr
	510.55	I	W
	136.14	II	
	157.40	II	
	172.24	II	
	174.28	II	
	179.34	I	
	182.88	II	
	205.13	I	
	216.20	I	
	217.81	I	
	218.65	II	
	226.76	II	
	235.12	II	
	239.15	I	
	277.21	I	
	281.33	I	
	285.18	I	
	296.69	II	
297.05	I		
299.95	I		
300.81	I		
303.74	I		
304.76	I		
Iron (IS)	305.91	I	
	316.79	I	
	517.16	I	
	321.33	II	
	487.21	I	
	458.38	II	
	413.70	I	
	410.75	I	
	383.63	I	
	363.83	I	
	339.93	I	
	328.68	I	
	308.37	I	
	282.33	I	
	249.59	I	

TABLE 1 *Continued*

Element	Wavelength, λ , nm	Line Classification ^A	Possible Interference ^B
	226.76	II	
	218.65	II	
	216.20	I	
	193.53	II	
	190.48	I	
	187.75	II	
	149.65	II	
	271.44	II	
	273.07	II	Co
	492.39	I	
Lead	405.75	I	Mn
Manganese	293.31	II	Cr, Mo, Ni
	255.86	II	Zr
	263.82	II	Al, W
Molybdenum	379.83	II	
	202.03	II	Mn
	277.54	I	Cu, V, Co, Mn
	281.61	II	Mn
	386.41	I	V, Cr
Nickel	471.44	I	
	227.73	II	
	341.48	I	
	352.45	I	
	231.60	II	Co, Ti
	227.02	II	Nb, W
	243.79	II	Co, Fe, Ni
Niobium	313.08	II	Ti, V
	319.50	II	Mo, Al, V
Nitrogen	149.26	I	Fe, Ti, Si, Mn, Cu, Ni and nitride forming elements such as Ti
Phosphorus	178.29	I	Mo
Silicon	288.16	I	Mo, Cr, W
	251.61	I	Fe, V
	212.41	I	Mo, Ni, V, Cu, Nb
	390.55	I	Cr, Cu, W, Ti
Sulfur	180.73	I	Mn
Tin	147.52	II	
	189.99	II	Mn, Mo, Al
Titanium	308.80	I	Cu, Co
	337.28	II	Nb
Tungsten	324.20	II	Nb
	400.88	I	
	202.99	II	Ti, V, Mn
	220.50	II	Co
Vanadium	437.92	I	
	310.23	II	Fe, Mo, Nb, Ni
Zirconium	468.78	I	
	349.62	II	
	343.82	II	W
	206.19	II	W

^A The numerals I or II in the line classification column indicate that the line has been classified in a term array and definitely assigned to the normal atom (I) or to the singly ionized atom (II).

^B Interferences are dependent upon instrument design, spectrum line choices, and excitation conditions, and those listed require confirmation based upon specimens selected especially to demonstrate suspected interferences.

intensity of the spectral radiation or transfer slight amounts of contamination between specimens, or both. The number of

acceptable burns on an electrode varies from one instrument to another, and should be established in each laboratory.

NOTE 3—It has been reported that thousands of burns can be performed on a tungsten electrode before replacement is necessary.

7.2 *Inert Gas, Argon*, in accordance with Practice E406.

8. Reference Materials

8.1 *Certified Reference Materials (CRMs)*—These are available from the National Institute of Standards and Technology (NIST) and other sources and span all or part of the mass fraction ranges listed in 1.1. They are used to calibrate the spectrometer for the elements of interest or to validate the performance of the test method. It is not recommended to use CRMs as verifiers or to establish the repeatability of the chemical measurement process.

NOTE 4—Certified Reference Materials manufactured by NIST are trademarked with the name, “Standard Reference Material.”

8.2 *Reference Materials (RMs)*—These are available from multiple suppliers or can be developed in house. Reference Materials are typically used in control procedures (verifiers) and in drift correction (standardization) of the spectrometer, and they may be useful in calibrations. These reference materials shall be homogenous and contain appropriate mass fractions of each element for the intended purpose. Refer to Guide E2972 for production of your own reference materials.

8.3 Several issues can impact the selection and use of CRMs and RMs:

8.3.1 Samples and reference materials may exhibit differences in metallurgical structure, in particular having different sizes, compositions, and distributions of inclusions. Inhomogeneous distribution of inclusions can worsen repeatability of individual measurements of elements found in the inclusions. Some inclusions may be removed during preburn steps prior to integration of intensities, causing low results. Typical samples can be used to determine repeatability of individual measurements to yield estimates consistent with performance for actual samples.

8.3.2 For certain elements, there may be no available reference materials with metallurgical structure similar to typical samples. Therefore, calibrations may be biased. It is recommended to validate results using typical samples analyzed using Test Methods E350 and E1019.

9. Preparation of Specimens and Reference Materials

9.1 The specimens and reference materials shall be prepared in the same manner. A specimen cut from a large sample section shall be of sufficient size and thickness for preparation and to properly fit the spectrometer stand. A 10-mm to 38-mm thick specimen is normally most practical.

9.2 Ensure that the specimens are free from voids and pits in the region to be measured (Note 5). Initially, grind the surface with a 50-grit to 80-grit abrasive belt or disc (wet or dry) or mill the surface. If wet grinding, perform the final grind with a dry abrasive belt or disc. A finer abrasive grinding media (for example, 120-grit) may be used for the final grind, but is not essential.

NOTE 5—Specimen porosity is undesirable because it leads to the improper “diffuse-type” rather than the desired “concentrated-type” discharge. The specimen surface should be kept clean because the specimen

is the electron emitter, and electron emission is inhibited by oily, dirty surfaces.

9.2.1 Reference materials and specimens shall be refinished dry on an abrasive belt or disc before being remeasured on the same area.

10. Preparation of Apparatus

NOTE 6—The instructions given in this test method apply to most spectrometers. However, some settings and adjustments may require modification, and additional preparation of the equipment may be required. It is not within the scope of an ASTM test method to prescribe the minute details of the apparatus preparation, which may differ not only for each manufacturer, but also for different equipment from the same manufacturer. For a description of and further details of operation for a particular spectrometer, refer to the manufacturer’s manual(s).

10.1 Program the spectrometer to use the internal standard lines and one of the analytical lines for each element listed in Table 1. Multiple lines may be used for a given element (for example, nickel) depending on the mass fraction range and the individual spectrometer software.

10.2 Test the positioning of the spectrometer entrance slit to ensure that peak radiation is entering the spectrometer chamber. This shall be done initially and as often as necessary to maintain proper entrance slit alignment. Follow the manufacturer’s recommended procedures. The laboratory will determine the frequency of positioning the alignment based on instrument performance.

10.3 Exit slit positioning and alignment is normally performed by the manufacturer at spectrometer assembly. Under normal circumstances, further exit slit alignment is not necessary (Note 7).

NOTE 7—The manner and frequency of positioning or checking the position of the exit slits will depend on factors such as the type of spectrometer, the variety of analytical problems encountered, and the frequency of use. Each laboratory should establish a suitable check procedure utilizing qualified service engineers.

11. Burn and Exposure

11.1 *Electrical Parameters:*

11.1.1 Burn parameters are normally established by the spectrometer manufacturer. The following ranges are historical guidelines and newer instruments may vary from these:

	Triggered Capacitor Discharge
Capacitance, μF	10 to 15
Inductance, μH	50 to 70
Resistance, Ω	3 to 5
Potential, V	940 to 1000
Current, A, r-f	0.3 to 0.8
Number of discharges	60

11.1.2 When parameter values are established, maintain them carefully. The variation of the power supply voltage shall not exceed $\pm 5\%$ and preferably should be held within $\pm 2\%$.

11.1.3 *Initiation Circuit*—The initiator circuit parameters shall be adequate to uniformly trigger the capacitor discharge. The following settings are historical guidelines and newer instruments may vary from these:

Capacitance, μF	0.0025
Inductance, μH	residual
Resistance, Ω	2.5
Peak voltage, V	18 000

11.1.4 *Other Electrical Parameters*—Excitation units, on which the precise parameters given in 11.1.1 and 11.1.3 are not available, may be used provided that it can be shown experimentally that equivalent precision and accuracy are obtained.

11.2 *Burn and Measurement Conditions*—The following ranges are normally adequate:

Argon flush period, s	5 to 15	
Preburn period, s	5 to 20	
Exposure period, s	3 to 30	
Argon flow	ft ³ /h	L/min
Flush	5 to 45	2.5 to 25
Preburn	5 to 45	2.5 to 25
Exposure	5 to 30	2.5 to 15

11.2.1 Select preburn and exposure periods after a study of volatilization rates during specimen burns. Once established, maintain the parameters consistently.

11.2.2 A high-purity argon atmosphere is required at the analytical gap. Molecular gas impurities, such as nitrogen, oxygen, hydrocarbons, or water vapor, either in the gas system or from improperly prepared specimens, should be minimized.

11.3 *Electrode System*—The specimen, electrically negative, serves as one electrode. The opposite electrode is a tungsten or silver rod, the tip of which has been machined to a 90° or 120° angled cone. Use either a 3 mm, 4 mm, or 5 mm (± 0.1 mm) analytical gap. Condition a fresh counter electrode with two burns to six burns using the operating conditions described in 11.1 and 11.2.

11.4 *Photomultiplier Potentials*—The sensitivities of the photomultipliers are normally established and set by the spectrometer manufacturer based on the particular wavelengths selected.

11.5 *Semiconductor Detector Array*—In newer instruments semiconductor detector arrays are replacing PMTs. The width of the individual pixels shall be similar to the width of the exit slits used in conventional instruments equipped with PMTs.

12. Calibration, Standardization, and Verification

12.1 *Calibration*—Using the conditions given in 11.1 – 11.3, measure calibrants and potential drift correction samples in a random sequence, bracketing these with measurements of any materials intended for use as verifiers. (A calibrant may be used later as a verifier. See 8.1.) There shall be at least three calibrants for each element, spanning the required mass fraction range. Measure each calibrant, drift correction sample, and verifier two times to four times and use the average value. If the spectrometer system and software permit, repeat with different random sequences at least two times. Using the averages of the data for each point, determine analytical curves as directed in the spectrometer manufacturer’s software or Practice E305.

12.2 *Standardization*—Following the manufacturer’s recommendations, standardize on an initial setup or anytime that it is known or suspected that readings have shifted. Make the necessary corrections either by adjusting the controls on the readout or by applying arithmetic corrections. Standardization shall be done anytime verification indicates that readings have gone out of statistical control. In the case of automatic

corrections conducted by the spectrometer software, observe the standardization factors or offsets, or both. The factors or offsets, or both, are often presented in the spectrometer software after standardization or stored in log files, or both. Refer to your instrument manual or instrument manufacturer for access to this information.

12.3 *Verification*—Verify that the instrument’s standardization is valid immediately after each standardization and as required in accordance with 12.3.2.

12.3.1 Analyze verifiers in accordance with Section 13. If results do not fall within the control limits established in 12.4, run another standardization or investigate why the instrument may be malfunctioning.

12.3.2 Each laboratory shall determine the necessary frequency of verification based on statistical analysis. Typically every 4 h to 8 h is practical and adequate. If the results are not within the control limits established in 12.4, perform a standardization and repeat verification. Repeat standardization as necessary so verification results are within control limits or investigate further for instrument problems.

12.4 *Quality Control*—Establish control limits in accordance with ASTM MNL 7, Practice E1329, or other equivalent quality control procedure.

13. Measurements

13.1 Place the prepared surface of the specimen on the sample stand so that the measurement will impinge on a location known to be representative of the entire specimen.

NOTE 8—With certain spectrometers, a properly burned specimen usually exhibits a dark ring around the pitted sparked area. With that equipment, a smooth texture, white burn without the characteristic dark ring indicates an improperly burned specimen. If boron nitride disks are used to mechanically restrict the burned area of the sample, a properly burned specimen may not exhibit a dark ring.

13.2 Measure specimens in duplicate and report the average of the duplicate results.

14. Calculation

14.1 Using the average results obtained in 13.2, calculate the mass fractions of the elements from the analytical curves developed in 12.1.

14.2 Rounding of test results obtained using this test method shall be performed in accordance with the Rounding Method of Practice E29, unless an alternative rounding method is specified by the customer or applicable material specification.

15. Precision and Bias

15.1 Supporting data have been filed at ASTM International Headquarters and may be obtained by requesting RR:E01-1122.² The interlaboratory test data summarized in Table 2 have been evaluated in accordance with Practice E1601.

15.2 *Precision*—Up to eight laboratories cooperated in performing this test method with thirteen unknown samples and obtained the statistical information summarized in Table 2.

15.3 *Bias*—Differences between average composition determined by this method and the certified compositions provides the bias found in the interlaboratory study. Bias information is found in Table 2.

TABLE 2 Statistical Information

Material	Number of Laboratories	Certified Value, %	Xbar	r	R	Bias
Aluminum						
Sample 1	7	0.016	0.0175	0.0034	0.0047	0.0015
Sample 2	7	0.041	0.0413	0.0037	0.0067	0.0003
Sample 3	8	0.015	0.0174	0.0013	0.0028	0.0024
Sample 4	8	0.018	0.0200	0.0006	0.0024	0.0020
Sample 5	8	0.062	0.0668	0.0042	0.0111	0.0048
Sample 6	8	0.0009	0.0023	0.0002	0.0022	0.0014
Sample 7	8	0.093	0.0890	0.0031	0.0152	-0.0040
Sample 8	7	0.021	0.0234	0.0015	0.0031	0.0024
Sample 9	7	0.03	0.0333	0.0017	0.0037	0.0033
Sample 10	8	0.024	0.0256	0.0009	0.0028	0.0016
Sample 11	8	0.027	0.0286	0.0010	0.0032	0.0016
Sample 12	8	0.017	0.0201	0.0012	0.0043	0.0031
Sample 13	7		0.0031	0.0010	0.0027	
Antimony						
Sample 1	7	0.027	0.0272	0.0026	0.0098	0.0002
Sample 2	5		0.0008	0.0007	0.0012	
Sample 3	7		0.0014	0.0009	0.0022	
Sample 4	6		0.0009	0.0007	0.0022	
Sample 5	6	0.0004	0.0011	0.0007	0.0024	0.0007
Sample 6	8	0.0025	0.0026	0.0011	0.0033	0.0001
Sample 7	8	0.006	0.0062	0.0011	0.0029	0.0002
Sample 8	7	0.002	0.0017	0.0007	0.0022	-0.0003
Sample 9	5		0.0006	0.0005	0.0011	
Sample 10	8	0.003	0.0019	0.0012	0.0037	-0.0011
Sample 11	6		0.0010	0.0008	0.0024	
Sample 12	8		0.0015	0.0008	0.0028	
Sample 13	7		0.0024	0.0010	0.0047	
Arsenic						
Sample 1	6	0.05	0.0459	0.0024	0.0116	-0.0041
Sample 2	6	0.003	0.0033	0.0008	0.0014	0.0003
Sample 3	7		0.0052	0.0008	0.0047	
Sample 4	7	(<0.005)	0.0004	0.0006	0.0009	
Sample 5	7	0.0035	0.0037	0.0008	0.0014	0.0002
Sample 6	7	0.0056	0.0071	0.0010	0.0021	0.0015
Sample 7	7	(0.005)	0.0065	0.0008	0.0048	
Sample 8	6	0.007	0.0074	0.0008	0.0037	0.0004
Sample 9	6		0.0041	0.0009	0.0034	
Sample 10	7	0.004	0.0053	0.0008	0.0035	0.0013
Sample 11	7	0.0044	0.0050	0.0009	0.0014	0.0006
Sample 12	7	0.007	0.0079	0.0012	0.0076	0.0009
Sample 13	6	0.07	0.0654	0.0077	0.0196	-0.0046
Boron						
Sample 1	4	(0.00002)	0.0002	0.0001	0.0002	
Sample 2	7	(0.0002)	0.0003	0.0001	0.0002	
Sample 3	8		0.0004	0.0001	0.0004	
Sample 4	8	(<0.0005)	0.0005	0.0001	0.0002	
Sample 5	7		0.0002	0.0001	0.0002	
Sample 6	8	0.0002	0.0003	0.0000	0.0002	0.0001
Sample 7	8	0.0047	0.0045	0.0004	0.0020	-0.0002
Sample 8	5		0.0002	0.0001	0.0005	
Sample 9	4		0.0003	0.0000	0.0004	
Sample 10	7		0.0002	0.0001	0.0002	
Sample 11	8		0.0003	0.0001	0.0004	
Sample 12	8	(0.0004)	0.0004	0.0001	0.0006	
Sample 13	7		0.0005	0.0001	0.0002	
Calcium						
Sample 1	3	(<0.0001)	0.0001	0.0001	0.0003	
Sample 2	4	(0.0005)	0.0001	0.0001	0.0001	
Sample 3	8		0.0003	0.0001	0.0003	
Sample 4	6	(0.001)	0.0002	0.0001	0.0002	
Sample 5	4	(<0.0005)	0.0001	0.0001	0.0002	
Sample 6	8	0.0012	0.0012	0.0003	0.0008	0.0000
Sample 7	8		0.0006	0.0001	0.0004	
Sample 8	7		0.0003	0.0002	0.0004	
Sample 9	7	0.002	0.0018	0.0004	0.0008	-0.0002
Sample 10	7	0.0003	0.0002	0.0001	0.0002	-0.0001
Sample 11	6		0.0001	0.0001	0.0002	
Sample 12	3	(0.0001)	0.0001	0.0000	0.0002	
Sample 13	7		0.0003	0.0003	0.0004	
Carbon						
Sample 1	7	0.211	0.2169	0.0073	0.0252	0.0059
Sample 2	7	0.142	0.1525	0.0084	0.0230	0.0105
Sample 3	8	0.13	0.1384	0.0072	0.0167	0.0084

TABLE 2 *Continued*

Material	Number of Laboratories	Certified Value, %	Xbar	r	R	Bias
Sample 4	8	0.658	0.6605	0.0075	0.0163	0.0025
Sample 5	8	0.483	0.4892	0.0092	0.0124	0.0062
Sample 6	8	0.457	0.4687	0.0110	0.0156	0.0117
Sample 7	8	0.332	0.3251	0.0202	0.0279	-0.0069
Sample 8	7	0.128	0.1305	0.0045	0.0076	0.0025
Sample 9	7	0.12	0.1196	0.0039	0.0142	-0.0004
Sample 10	8	1.03	1.024	0.0170	0.0227	-0.006
Sample 11	8	0.255	0.2530	0.0072	0.0137	-0.0020
Sample 12	8	0.107	0.1114	0.0040	0.0115	0.0044
Sample 13	7	0.376	0.3593	0.0280	0.0280	-0.0167
Chromium						
Sample 1	7	0.081	0.0797	0.0020	0.0066	-0.0013
Sample 2	7	0.044	0.0444	0.0009	0.0051	0.0004
Sample 3	8	4.22	4.209	0.0327	0.3209	-0.011
Sample 4	8	0.16	0.1564	0.0030	0.0133	-0.0036
Sample 5	8	0.021	0.0190	0.0012	0.0035	-0.0020
Sample 6	8	0.098	0.0973	0.0013	0.0079	-0.0007
Sample 7	8	5.11	5.086	0.0432	0.3534	-0.024
Sample 8	7	2.09	2.095	0.0169	0.0832	0.005
Sample 9	7	2.56	2.557	0.0167	0.1249	-0.003
Sample 10	8	1.36	1.356	0.0221	0.0911	-0.004
Sample 11	8	0.34	0.3334	0.0023	0.0255	-0.0066
Sample 12	8	8.22	8.143	0.0789	0.8918	-0.077
Sample 13	7	0.062	0.0650	0.0083	0.0083	0.0030
Cobalt						
Sample 1	7	0.19	0.1885	0.0011	0.0217	-0.0015
Sample 2	7	0.005	0.0033	0.0002	0.0028	-0.0017
Sample 3	8	0.011	0.0116	0.0006	0.0026	0.0006
Sample 4	8	0.0019	0.0020	0.0009	0.0024	0.0001
Sample 5	8	0.005	0.0038	0.0006	0.0025	-0.0012
Sample 6	8	0.0078	0.0072	0.0007	0.0032	-0.0006
Sample 7	8	0.006	0.0071	0.0009	0.0026	0.0011
Sample 8	7	0.01	0.0096	0.0006	0.0033	-0.0004
Sample 9	7		0.0088	0.0009	0.0029	
Sample 10	8	0.007	0.0068	0.0009	0.0028	-0.0002
Sample 11	8	0.01	0.0093	0.0009	0.0022	-0.0007
Sample 12	8	0.016	0.0160	0.0008	0.0041	0.0000
Sample 13	7		0.0061	0.0009	0.0032	
Copper						
Sample 1	7	0.023	0.0231	0.0011	0.0030	0.0001
Sample 2	7	0.03	0.0339	0.0014	0.0034	0.0039
Sample 3	8	0.11	0.1151	0.0045	0.0101	0.0051
Sample 4	8	0.151	0.1518	0.0070	0.0146	0.0008
Sample 5	8	0.015	0.0145	0.0007	0.0032	-0.0005
Sample 6	8	0.299	0.2993	0.0039	0.0254	0.0003
Sample 7	8	0.057	0.0569	0.0025	0.0058	-0.0001
Sample 8	7	0.177	0.1784	0.0034	0.0104	0.0014
Sample 9	7	0.08	0.0797	0.0034	0.0061	-0.0003
Sample 10	8	0.106	0.1068	0.0081	0.0130	0.0008
Sample 11	8	0.11	0.1094	0.0049	0.0103	-0.0006
Sample 12	8	0.115	0.1190	0.0050	0.0097	0.0040
Sample 13	7	0.051	0.0511	0.0031	0.0047	0.0001
Manganese						
Sample 1	7	0.316	0.3153	0.0033	0.0130	-0.0007
Sample 2	7	1.12	1.148	0.0200	0.0373	0.028
Sample 3	8	0.44	0.4549	0.0058	0.0161	0.0149
Sample 4	8	0.82	0.8319	0.0176	0.0336	0.0119
Sample 5	8	0.72	0.7330	0.0081	0.0265	0.0130
Sample 6	8	0.772	0.7825	0.0116	0.0298	0.0105
Sample 7	8	0.169	0.1713	0.0033	0.0091	0.0023
Sample 8	7	0.441	0.4437	0.0066	0.0168	0.0027
Sample 9	7	0.55	0.5584	0.0075	0.0226	0.0084
Sample 10	8	0.33	0.3340	0.0075	0.0182	0.0040
Sample 11	8	1.42	1.445	0.0132	0.0551	0.025
Sample 12	8	0.333	0.3374	0.0028	0.0131	0.0044
Sample 13	7	0.8	0.8070	0.0251	0.0402	0.0070
Molybenum						
Sample 1	7	0.05	0.0517	0.0016	0.0024	0.0017
Sample 2	7	0.008	0.0083	0.0002	0.0034	0.0003
Sample 3	8	0.47	0.4743	0.0095	0.0270	0.0043
Sample 4	8	0.019	0.0204	0.0009	0.0033	0.0014
Sample 5	8	0.005	0.0051	0.0004	0.0037	0.0001
Sample 6	8	0.0419	0.0417	0.0019	0.0026	-0.0002
Sample 7	8	1.28	1.307	0.0415	0.1200	0.027

TABLE 2 *Continued*

Material	Number of Laboratories	Certified Value, %	Xbar	r	R	Bias
Sample 8	7	0.89	0.9044	0.0163	0.0798	0.0144
Sample 9	7	1.02	1.035	0.0128	0.1032	0.015
Sample 10	8	0.044	0.0448	0.0013	0.0024	0.0008
Sample 11	8	0.42	0.4210	0.0103	0.0168	0.0010
Sample 12	8	0.9	0.8976	0.0215	0.0490	-0.0024
Sample 13	7	0.2	0.1978	0.0228	0.0228	-0.0022
Nickel						
Sample 1	7	0.43	0.4286	0.0044	0.0347	-0.0014
Sample 2	7	0.029	0.0316	0.0006	0.0027	0.0026
Sample 3	8	0.12	0.1290	0.0020	0.0091	0.0090
Sample 4	8	0.163	0.1682	0.0031	0.0116	0.0052
Sample 5	8	0.015	0.0146	0.0007	0.0031	-0.0004
Sample 6	8	0.154	0.1573	0.0018	0.0119	0.0033
Sample 7	8	0.445	0.4486	0.0074	0.0363	0.0036
Sample 8	7	0.197	0.1999	0.0021	0.0159	0.0029
Sample 9	7	0.25	0.2498	0.0033	0.0178	-0.0002
Sample 10	8	0.135	0.1356	0.0035	0.0115	0.0006
Sample 11	8	1.74	1.728	0.0237	0.0867	-0.012
Sample 12	8	0.123	0.1213	0.0023	0.0105	-0.0017
Sample 13	7	0.069	0.0694	0.0019	0.0071	0.0004
Niobium						
Sample 1	7	0.003	0.0031	0.0003	0.0013	0.0001
Sample 2	7	0.041	0.0415	0.0015	0.0074	0.0005
Sample 3	8	0.002	0.0064	0.0004	0.0028	0.0044
Sample 4	8	0.024	0.0245	0.0021	0.0048	0.0005
Sample 5	5	(<0.002)	0.0003	0.0002	0.0006	
Sample 6	7	0.0009	0.0005	0.0002	0.0007	-0.0004
Sample 7	8	0.122	0.1178	0.011	0.0215	-0.0042
Sample 8	7	(<0.003)	0.0036	0.0003	0.0018	
Sample 9	7		0.0044	0.0003	0.0018	
Sample 10	8		0.0021	0.0003	0.0015	
Sample 11	8		0.0013	0.0003	0.0022	
Sample 12	8	0.076	0.0839	0.0019	0.0139	0.0079
Sample 13	6		0.0005	0.0003	0.0006	0.0005
Nitrogen						
Sample 1	7	0.0099	0.0083	0.0008	0.0048	-0.0016
Sample 2	7		0.0078	0.0010	0.0050	
Sample 3	8	0.018	0.0228	0.00684	0.0684	0.0048
Sample 4	6		0.0011	0.0004	0.0029	
Sample 5	8	0.0056	0.0040	0.0009	0.0050	-0.0016
Sample 6	8	0.0106	0.0099	0.0010	0.0055	-0.0007
Sample 7	8	0.0076	0.0094	0.0013	0.0054	0.0018
Sample 8	7	0.0097	0.0102	0.0012	0.0056	0.0005
Sample 9	7		0.0101	0.0009	0.0064	
Sample 10	8	0.0084	0.0075	0.0007	0.0048	-0.0009
Sample 11	8	0.0066	0.0071	0.0008	0.0051	0.0005
Sample 12	6	0.055	0.0564	0.0026	0.0233	0.0014
Sample 13	7		0.0106	0.0010	0.0051	
Phosphorous						
Sample 1	7	0.018	0.0170	0.0007	0.0043	-0.0010
Sample 2	7	0.016	0.0183	0.0012	0.0053	0.0023
Sample 3	8	0.017	0.0159	0.0008	0.0047	-0.0011
Sample 4	8	0.011	0.0112	0.0010	0.0043	0.0002
Sample 5	8	0.01	0.0090	0.0007	0.0029	-0.0010
Sample 6	8	0.0096	0.0101	0.0018	0.0034	0.0005
Sample 7	8	0.006	0.0064	0.0008	0.0029	0.0004
Sample 8	7	0.012	0.0129	0.0008	0.0034	0.0009
Sample 9	7	0.01	0.0088	0.0006	0.0017	-0.0012
Sample 10	8	0.013	0.0164	0.00564	0.0567	0.0034
Sample 11	8	0.01	0.0087	0.0008	0.0019	-0.0013
Sample 12	8	0.008	0.0078	0.0008	0.0020	-0.0002
Sample 13	7	0.061	0.0584	0.0068	0.0105	-0.0026
Silicon						
Sample 1	7	0.015	0.0127	0.0010	0.0071	-0.0023
Sample 2	7	0.058	0.0668	0.0012	0.0088	0.0088
Sample 3	8	0.27	0.2787	0.0043	0.0117	0.0087
Sample 4	8	0.5	0.4959	0.0074	0.0153	-0.0041
Sample 5	8	0.24	0.2251	0.0036	0.0094	-0.0149
Sample 6	8	0.21	0.2111	0.0030	0.0111	0.0011
Sample 7	8	0.775	0.7541	0.0157	0.0330	-0.0209
Sample 8	7	0.255	0.2520	0.0022	0.0100	-0.0030
Sample 9	7	0.32	0.3211	0.0039	0.0117	0.0011
Sample 10	8	0.32	0.3217	0.0057	0.0116	0.0017
Sample 11	8	1.54	1.519	0.0217	0.0612	-0.021

TABLE 2 *Continued*

Material	Number of Laboratories	Certified Value, %	Xbar	r	R	Bias
Sample 12	8	0.327	0.3322	0.0043	0.0197	0.0052
Sample 13	7	0.31	0.3044	0.0082	0.0135	-0.0056
Sulfur						
Sample 1	7	0.005	0.0460	0.0024	0.3116	0.0410
Sample 2	7	0.008	0.0076	0.0008	0.0031	-0.0004
Sample 3	8	0.015	0.0146	0.0010	0.0021	-0.0004
Sample 4	8	0.012	0.0135	0.0018	0.0044	0.0015
Sample 5	8	0.025	0.0232	0.0039	0.0064	-0.0018
Sample 6	8	0.0234	0.0221	0.0035	0.0054	-0.0013
Sample 7	8	0.033	0.0321	0.0038	0.0063	-0.0009
Sample 8	7	0.026	0.0241	0.0026	0.0056	-0.0019
Sample 9	7	0.003	0.0013	0.0005	0.0014	-0.0017
Sample 10	8	0.014	0.0144	0.0032	0.0046	0.0004
Sample 11	8	0.004	0.0046	0.0007	0.0007	0.0006
Sample 12	8	0.008	0.0076	0.0005	0.0023	-0.0004
Sample 13	7	0.047	0.0454	0.0082	0.0112	-0.0016
Tin						
Sample 1	7	0.061	0.0588	0.0011	0.0079	-0.0022
Sample 2	7	0.002	0.0028	0.0008	0.0022	0.0008
Sample 3	8	0.008	0.0073	0.0003	0.0014	-0.0007
Sample 4	8	0.026	0.0263	0.0010	0.0028	0.0003
Sample 5	7	(0.0006)	0.0015	0.0002	0.0014	
Sample 6	8	0.0124	0.0127	0.0006	0.0016	0.0003
Sample 7	8	0.005	0.0049	0.0003	0.0023	-0.0001
Sample 8	7	0.013	0.0135	0.0003	0.0015	0.0005
Sample 9	7		0.0047	0.0002	0.0019	
Sample 10	8	0.006	0.0064	0.0010	0.0013	0.0004
Sample 11	8	0.006	0.0075	0.0003	0.0013	0.0015
Sample 12	8	0.009	0.0094	0.0021	0.0022	0.0004
Sample 13	7	0.054	0.0481	0.0043	0.0062	-0.0059
Titanium						
Sample 1	7	0.004	0.0036	0.0002	0.0008	-0.0004
Sample 2	7	0.008	0.0076	0.0003	0.0009	-0.0004
Sample 3	8	0.003	0.0036	0.0002	0.0005	0.0006
Sample 4	8	0.015	0.0156	0.0011	0.0016	0.0006
Sample 5	7	(0.001)	0.0011	0.0001	0.0006	
Sample 6	8	0.0009	0.0009	0.0001	0.0006	0.0000
Sample 7	8	0.034	0.0358	0.0028	0.0040	0.0018
Sample 8	7	(0.001)	0.0012	0.0001	0.0006	
Sample 9	7		0.0017	0.0001	0.0006	
Sample 10	8	0.003	0.0020	0.0010	0.0012	-0.0010
Sample 11	8	0.003	0.0037	0.0004	0.0007	0.0007
Sample 12	8	(0.002)	0.0027	0.0001	0.0006	
Sample 13	7	0.01	0.0112	0.0022	0.0028	0.0012
Vanadium						
Sample 1	7	0.01	0.0106	0.0008	0.0019	0.0006
Sample 2	7	0.012	0.0122	0.0007	0.0024	0.0002
Sample 3	8	0.016	0.0194	0.0008	0.0021	0.0034
Sample 4	8	0.012	0.0124	0.0010	0.0030	0.0004
Sample 5	7	(<0.002)	0.0012	0.0002	0.0014	
Sample 6	8	0.0295	0.0298	0.0009	0.0029	0.0003
Sample 7	8	0.802	0.8233	0.0205	0.0403	0.0213
Sample 8	7	0.003	0.0036	0.0005	0.0013	0.0006
Sample 9	7	0.015	0.0134	0.0009	0.0020	-0.0016
Sample 10	8	0.005	0.0046	0.0004	0.0015	-0.0004
Sample 11	8	0.003	0.0020	0.0003	0.0020	-0.0010
Sample 12	8	0.236	0.2366	0.0062	0.0163	0.0006
Sample 13	7		0.0036	0.0025	0.0031	
Zirconium						
Sample 1	6	(0.01)	0.0278	0.0096	0.0108	
Sample 2	6	0.022	0.0173	0.0055	0.0061	-0.0047
Sample 3	7		0.0005	0.0002	0.0019	
Sample 4	7		0.0003	0.0001	0.0006	
Sample 5	7		0.0002	0.0001	0.0004	
Sample 6	7	0.0007	0.0003	0.0001	0.0005	-0.0004
Sample 7	7	0.052	0.0513	0.0126	0.0190	-0.0007
Sample 8	5		0.0004	0.0001	0.0015	
Sample 9	6		0.0005	0.0006	0.0023	
Sample 10	7		0.0003	0.0002	0.0008	
Sample 11	7		0.0003	0.0001	0.0011	
Sample 12	7	(0.001)	0.0007	0.0001	0.0034	
Sample 13	6		0.0003	0.0001	0.0006	

16. Keywords

16.1 carbon steel; low-alloy steel; spark atomic emission; spectrometric analysis; spectrometry

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; <http://www.copyright.com/>