



# Standard Specification for Medical Screwdriver Bits<sup>1</sup>

This standard is issued under the fixed designation F116; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This specification covers the acceptable dimensions and tolerances for bits of screwdrivers to insert and remove metal screws used as surgical implants.

1.2 This specification is based, in part, upon ISO 8319–1, ISO 8319–2, and ISO 10664.

1.3 The screwdrivers with the bits described in this specification are suitable for use with screws described in Specification F543, ISO 5835, and ISO 9268.

1.4 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.5 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

## 2. Referenced Documents

### 2.1 ASTM Standards:<sup>2</sup>

E18 Test Methods for Rockwell Hardness of Metallic Materials

F90 Specification for Wrought Cobalt-20Chromium-15Tungsten-10Nickel Alloy for Surgical Implant Applications (UNS R30605)

F543 Specification and Test Methods for Metallic Medical Bone Screws

F565 Practice for Care and Handling of Orthopedic Implants and Instruments

F899 Specification for Wrought Stainless Steels for Surgical Instruments

F1744 Guide for Care and Handling of Stainless Steel Surgical Instruments

<sup>1</sup> This specification is under the jurisdiction of ASTM Committee F04 on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee F04.21 on Osteosynthesis.

Current edition approved Oct. 1, 2016. Published October 2016. Originally approved in 1969. Last previous edition approved in 2012 as F116 – 12. DOI: 10.1520/F0116-12R16.

<sup>2</sup> For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

### 2.2 ISO Standards:<sup>3</sup>

ISO 5832–5 Implants for Surgery—Metallic Materials—Part 5: Wrought Cobalt-Chromium-Tungsten-Nickel Alloy

ISO 5835 Implants for Surgery—Metal Bone Screws with Hexagonal Driver Connection, Spherical Under Surface of Head, Asymmetrical Thread—Dimensions

ISO 7153–1 Surgical Instruments—Metallic Materials—Part 1: Stainless Steel

ISO 8319–1 Orthopaedic Instruments—Drive Connections—Part 1: Keys for Use with Screws with Hexagon Socket Heads

ISO 8319–2 Orthopaedic Instruments—Drive Connections—Part 2: Screwdrivers for Single Slot Head Screws, screws with Cruciate Slot, and Cross-Recessed Head Screws

ISO 9268 Implants for Surgery—Metal Bone Screws with Conical Under-Surface of Head—Dimensions

ISO 10664 Hexalobular Internal Driving Feature for Bolts and Screws

## 3. Classification

3.1 This specification includes the following types of bits for medical screwdrivers:

3.1.1 *Type I*—Single-slot bit.

3.1.2 *Type II*—Cruciate-slot bit.

3.1.3 *Type III*—Cross-slot (Modified Phillips) bit.

3.1.4 *Type IV*—Hexagonal bit.

3.1.5 *Type V*—Square bit.

3.1.6 *Type VI*—Hexalobe bit.

## 4. Dimensions and Tolerances

4.1 Screwdriver bits conforming to this specification shall be fabricated in accordance with the dimensions and tolerances described below:

4.1.1 *Type I*—Single-slot screwdriver bits shall conform to the dimensions and tolerances provided in Table 1, and shown in Fig. 1.

4.1.2 *Type II*—Cruciate-slot screwdriver bits shall conform to the dimensions and tolerances provided in Table 2, and shown in Fig. 2.

<sup>3</sup> Available from American National Standards Institute (ANSI), 25 W. 43rd St., 4th Floor, New York, NY 10036, http://www.ansi.org.




FIG. 1 Dimensions of Single-Slot Screwdriver Bit

TABLE 1 Dimensions and Tolerances of Single-Slot Screwdriver Bit

Slot Thickness, <i>a</i> (mm)	Slot Width, <i>b</i> (mm; maximum)	Slot Angle, $\alpha$ (degrees)
1.10	+0.03 -0.07	5.0
1.10	+0.03 -0.07	5.0




FIG. 2 Dimensions of Cruciate-Slot Screwdriver Bit

TABLE 2 Dimensions and Tolerances of Cruciate-Slot Screwdriver Bit

Slot Thickness, <i>a</i> (mm)	Slot Width, <i>b</i> (mm; maximum)
1.10	+0.03 -0.07
1.10	+0.03 -0.07

4.1.3 *Type III*—Cross-slot (Modified Phillips) screwdriver bits shall conform to the dimensions and tolerances provided in Fig. 3.

4.1.4 *Type IV*—Hexagonal screwdriver bits shall conform to the dimensions and tolerances provided in Table 3, and shown in Fig. 4.

4.1.5 *Type V*—Square screwdriver bits shall conform to the general dimensions provided in Table 4, and shown in Fig. 5.

4.1.6 *Type VI*—Hexalobe screwdriver bits shall conform to the general dimensions provided in Table 5, and shown in Fig. 6.

## 5. Material Requirements

5.1 The bit and shaft portion of the screwdriver should be fabricated from one of the following materials:

5.1.1 Martensitic stainless steel (Specification F899 or ISO 7153-1).




FIG. 3 Dimensions and Tolerances of Cross-Slot (Modified Phillips) Screwdriver Bit

5.1.2 Cold worked Cobalt-Chromium-Tungsten-Nickel alloy (Specification F90 or ISO 5832-5).

5.2 The hardness of the material of the bit and shaft portion shall be 45-55 (stainless steel) or 45-50 (Cobalt-Chromium-Tungsten-Nickel alloy) when measured on the Rockwell C scale according to the procedures described in Test Methods E18.

5.3 The shaft shall be firmly fixed to a handle of appropriate material.

## 6. Finish and Marking

6.1 The bit portion of the screwdriver shall be free of nicks, dents, and scratches.

6.2 When size permits, the following information should be legibly marked on the shaft or handle of the screwdriver (in order of preference):

- 6.2.1 Manufacturer's name or logo.
- 6.2.2 Size (for Type IV through VI screwdriver bits).
- 6.2.3 Catalog number.
- 6.2.4 Manufacturing lot number.
- 6.2.5 Material (include ASTM designation, as appropriate).


6.3 The marking should be such that the mechanical integrity of the screwdriver is not compromised.

## 7. Care and Handling

7.1 Screwdrivers should be handled in accordance with Practices F565 and F1744, as appropriate.

## 8. Keywords

8.1 bone screw; orthopaedic medical devices; screwdriver; surgical instruments



NOTE 1—The hexagon length ( $L$ ) may be calculated from the following:  
 $L \approx l - R$  (measured with a radius gage).

FIG. 4 Dimensions and Tolerances of Hexagonal Screwdriver Bit

TABLE 3 Dimensions and Tolerances of Hexagonal Screwdriver Bit

Size	Hexagon Width Across Flats, $s$ (mm)		Hexagon Width Across Corners, $e$ (mm)		Hexagon Length, $L$ (mm; minimum)
1.5	1.500	+0.000 -0.025	1.670	+0.020 -0.020	2.0
2.5	2.500	+0.000 -0.030	2.820	+0.025 -0.025	3.0
3.5	3.500	+0.000 -0.030	3.956	+0.025 -0.025	4.0
4.5	4.500	+0.000 -0.030	5.106	+0.025 -0.025	5.0
5/32	3.950	+0.000 -0.050	4.471	+0.025 -0.025	5.0
3/16	4.735	+0.000 -0.050	5.373	+0.025 -0.025	6.0




FIG. 5 Dimensions of Square Screwdriver Bits

TABLE 4 Dimensions of Square Screwdriver Bits

Screwdriver Bit Size	Width Across Flats at Radius, $A$ (mm)	Width Across Corners at Radius, $B$ (mm)	Width Across Flats at Chamfer, $C$ (mm)	Width Across Corners at Chamfer, $D$ (mm)	Length, $E$ (mm)	Chamfer Angle, $F$ (°)	Chamfer Diameter, $G$ (mm)
1.0	0.84	1.19	0.67	0.95	1.25	24	0.40
1.5	1.16	1.64	0.96	1.38	1.57	28	0.48
2.0	2.03	2.87	1.20	1.70	2.69	26.5	0.76




FIG. 6 Dimensions of Hexalobe Screwdriver Bits

TABLE 5 Dimensions of Hexalobe Screwdriver Bits

Screwdriver Bit Size	Major Diameter A (mm)	Minor Diameter B (mm)	Internal Radius C (mm)	External Radius D (mm)	Length E (mm)	Chamfer Diameter F (mm)
4	1.27	0.89	0.31	0.08	1.52	<sup>A</sup>
5	1.40	1.00	0.34	0.10	1.52	<sup>A</sup>
6	1.68	1.19	0.39	0.12	1.52	<sup>A</sup>
8	2.27	1.60	0.55	0.15	1.78	<sup>A</sup>
10	2.69	1.91	0.62	0.20	2.03	<sup>A</sup>
15	3.23	2.29	0.74	0.24	2.16	<sup>A</sup>
20	3.81	2.71	0.87	0.28	2.29	2.96
25	4.36	3.07	0.96	0.34	2.54	3.33
30	5.45	3.86	1.22	0.42	3.18	4.11

<sup>A</sup> Chamfer not necessary.

## APPENDIX

### (Nonmandatory Information)

#### X1. RATIONALE

X1.1 This specification provides dimensions and tolerances for screwdrivers used to insert and remove medical bone screws. It is intended to provide assurances to the surgical staff that the specified screwdriver will fit and perform acceptably when used with the bone screws specified in Specification F543, ISO 5835, and ISO 9268.

X1.2 During a periodic review of the standard, the Subcommittee established that there are no consensus dimensions for

either of these bit types. However, an ISO standard relevant to Hexalobe screwdriver bits was identified and included. It was felt that deletion of the information on these bits from the standard would detract from the value of the standard as a whole.

*ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.*

*This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.*

*This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; http://www.copyright.com/*