



Standard Terminology for Surgical Suture Needles¹

This standard is issued under the fixed designation F1840; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This terminology covers general definitions for surgical needles.

2. Terminology

2.1 *Definitions of Terms Specific to the Instrument* (See Fig. 1):

attachment area, *n*—portion of the needle where the attachment of the suture takes place. For example, eyed, drilled, and channel.

body, *n*—central portion of the needle intended to be grasped by the needle holder.

chord length, *n*—the straight line distance between the two ends of a curved needle.

curvature, *n*—the shape of the needle viewed in profile. Some common shapes include, but are not limited to: straight, $\frac{1}{2}$ curve or “ski,” $\frac{1}{8}$ circle, $\frac{1}{4}$ circle, $\frac{3}{8}$ circle, $\frac{1}{2}$ circle, $\frac{5}{8}$ circle, and compound curvature (see Fig. 2).

cutting edge, *n*—cutting edges are made of various geometric shapes, that is, triangular, diamond, and hexagonal. The various edges may be sharpened by the manufacturer depending on the user performance.

needle length, *n*—the distance measured along the needle curvature from end to end.

needle radius, *n*—the radius of the uniformly curved portion or portions of the needle measured from the centerline of the needle body.

¹ This specification is under the jurisdiction of ASTM Committee F04 on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee F04.33 on Medical/Surgical Instruments.

Current edition approved Oct. 1, 2016. Published October 2016. Originally approved in 1997. Last previous edition approved in 2010 as F1840 – 10^{ε1}. DOI: 10.1520/F1840-10R16.

needle wire diameter, *n*—the gage or thickness of the needle wire, measured at a location between the needle body and the attachment area, where either no or minimal work has taken place.

point, *n*—portion of the needle intended to initiate tissue penetration.

point configuration, *n*—the shape of the point. Some common point configurations include, but are not limited to (see Fig. 3): taper, trocar, blunt, spatulated, conventional cutting edge, reverse cutting edge, cutting taper, and side cutting needle.

swage, *n*—the term used to describe any attachment method that uses mechanical force to crimp the end of the needle and firmly hold the suture in place.

2.2 *Definitions of Terms Specific to Mechanical Properties of a Surgical Needle:*

finish, *n*—final surface visual appearance classified as follows:
bright or mirror finish, *n*—highly reflective surface.

satın, matte, or black finish, *n*—reduced reflective surface (as compared to bright or mirror finish) varying from a dull appearance to a blackened surface.

maximum bend moment—the greatest moment applied to a needle during a bend test.

needle ductility—a measure of the amount of plastic bending a needle can withstand.

yield bend angle—the angle at which the yield bend moment occurs.

yield bend moment—the amount of moment required to initiate plastic deformation during a bend test.

3. Keywords

3.1 needle; stainless steel- surgical applications; surgical; suture needle




FIG. 1 Schematic of a Surgical Needle




FIG. 2 Typical Curvatures




FIG. 3 Typical Point Configurations

APPENDIX

(Nonmandatory Information)

X1. RATIONALE

X1.1 Because there is a clinical need for a variety of surgical suture needles for surgical procedures, they are manufactured in various configurations and from various materials. For practical purposes these devices supplied by different manufacturers necessitate a defined system of terms.

X1.2 The terms defined in this terminology are the most commonly used for surgical suture needles. However, the intent is not to prohibit technological innovation or to exclude surgical suture needles manufactured with other types of features.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; http://www.copyright.com/