



Standard Test Method for Dissolution Testing of Calcium Phosphate Granules, Fabricated Forms, and Coatings¹

This standard is issued under the fixed designation F1926/F1926M; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This test method covers calcium phosphate materials intended for use in surgical implant applications.

1.2 The material(s) shall be representative of that produced for sale. It shall have been produced and processed under standard manufacturing conditions.

1.3 The materials may be in the form of powders, granules, spall material, fabricated forms or coatings; and may be porous, nonporous, textured, and other implantable topographical substrate form representative of the end-use product.

1.4 The calcium phosphate material may constitute the only material in a substrate or it may be one of multiple materials so long as all other materials present do not dissolve under the test conditions described in this test method.

1.5 The values stated in either SI units or inch-pound units are to be regarded separately as standard. The values stated in each system may not be exact equivalents; therefore, each system shall be used independently of the other. Combining values from the two systems may result in nonconformance with the standard.

1.6 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

2. Referenced Documents

2.1 *ASTM Standards:*²

E691 Practice for Conducting an Interlaboratory Study to Determine the Precision of a Test Method

F1088 Specification for Beta-Tricalcium Phosphate for Surgical Implantation

¹ This test method is under the jurisdiction of ASTM Committee F04 on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee F04.13 on Ceramic Materials.

Current edition approved May 15, 2014. Published July 2014. Originally published in 1998. Last previous edition approved in 2010 as F1926/F1926M – 10. DOI: 10.1520/F1926_F1926M-14.

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

F1185 Specification for Composition of Hydroxylapatite for Surgical Implants

3. Terminology

3.1 *Definitions of Terms Specific to This Standard:*

3.1.1 *calcium phosphate, n*—any one of a number of inorganic chemical compounds containing calcium and phosphate ions as its principal constituents.

3.1.2 *coating, n*—layer of material mechanically or chemically adhering to the surface of a substrate.

4. Significance and Use

4.1 Aspects of the biological response to calcium phosphate materials in soft tissue and bone have been reported from laboratory studies and clinical use **(1-11)**.³

4.2 The requirements of this test method apply to calcium phosphate materials such as calcium hydroxyapatite (see Specification **F1185**), beta-tricalcium phosphate (see Specification **F1088**), and biphasic mixtures thereof with or without intentional addition of other minor components (<10 %).

4.3 This test method is limited to the laboratory evaluation of the dissolution rate of a calcium phosphate material. No correlation of the results to in vivo performance is implied. Therefore, it is recommended that a control material be included in the evaluation. The control material can be a standardized material such as NIST SRM 2910 or a historical control.

5. Dissolution Media

5.1 Water used for preparing reagents or dissolution media shall be degassed carbon dioxide free deionized or distilled water and have less than 0.1 ppm of residual Ca^{++} ion. Optionally, the water can be degassed *in situ*.

5.2 *Unbuffered Water Media*—Deionized or distilled water containing 8×10^{-5} M NaCl, 8×10^{-5} M CaCl_2 , and 5×10^{-5} M $\text{K}_3(\text{PO}_4)$.

³ The boldface numbers in parentheses refer to the list of references at the end of this standard.

5.3 *pH 5.5 MES Buffer Media*—1.0 M MES, [2-(N-morpholino)ethanesulfonic acid] having a pH of 5.5 at $37 \pm 0.5^\circ\text{C}$ and containing 8×10^{-5} M NaCl, 8×10^{-5} M CaCl₂, and 5×10^{-5} M K₃(PO₄).

5.3.1 A buffer concentration of 1.0 M will usually provide sufficient buffer capacity to keep the solution within ± 0.1 pH units of the initial value. If this is not the case, the buffer capacity should be adjusted accordingly.

5.3.2 The pH shall be adjusted to 5.5 at $37 \pm 0.5^\circ\text{C}$ using HCl or NaOH solutions.

5.4 *pH 7.4 TRIS Buffer Media*—1.0 M TRIS, [Tris(hydroxymethyl)aminomethane] having a pH of 7.4 at $37 \pm 0.5^\circ\text{C}$ and containing 8×10^{-5} M NaCl, 8×10^{-5} M CaCl₂, and 5×10^{-5} M K₃(PO₄).

5.4.1 A buffer concentration of 1.0 M will usually provide sufficient buffer capacity to keep the solution within ± 0.1 pH units of the initial value. If this is not the case, the buffer capacity should be adjusted accordingly.

5.4.2 The pH shall be adjusted to 7.4 at $37 \pm 0.5^\circ\text{C}$ using HCl or NaOH solutions.

6. Analytical Parameters

6.1 The following procedure should be performed with each of the media listed:

6.1.1 The dissolution rate shall be measured under the conditions of a constant ratio of initial material mass (mg) to total dissolution media volume (mL). The ratio of test material mass to dissolution media shall typically be between 1 to 4 mg/mL.

6.1.2 The dissolved Ca⁺⁺ concentration (± 1 ppm) shall be measured as soon as practical after the start of the experiment and at appropriate time intervals thereafter to allow the determination of their changes with time.

7. Analytical Procedures

7.1 Make pH measurements with an appropriately calibrated pH meter and probe.

7.2 Measure the Ca⁺⁺ concentrations potentiometrically. Ionic strength adjuster (ISA) must be added as required by the

electrode manufacturer. Other methods (for example, colorimetrically, atomic absorption (AA), inductively coupled plasma (ICP) spectroscopy, or inductively coupled plasma mass spectroscopy (ICP/MS)) may be used if equivalency can be demonstrated.

7.3 An appropriate bacteriostat (for example, 0.1 v/v % Hibiclens or 0.1 w/v % sodium azide) may be added to the dissolution media before the start of an experiment.

8. Dissolution Apparatus

8.1 The dissolution vessel shall be of such design to easily accommodate the test specimen, the stirrer, and the specific ion-electrode and reference electrode assemblies. It shall also be isolated from the atmosphere by an oxygen and carbon dioxide free inert gas purge.

8.1.1 A convenient apparatus (see Fig. 1) is a 100-mL jacketed beaker with circulating water from a thermostatically controlled vessel. A flat piece of polyethylene, or other inert plastic, with appropriate holes drilled to accommodate the probes, sample holder, and purge gas tube can serve as a lid.

8.2 It shall be of appropriate dimensions to contain the required volume of dissolution media at a level to keep the test material completely submerged during the test and facilitate sufficient stirring action from the magnetic stirrer bar.

8.3 The stirrer assembly shall be capable of maintaining a constant stirring rate of 100 ± 20 rpm.

8.3.1 *Magnetic Stirrer Bar*—Approximately 8-mm [0.31-in.] diameter, 28-mm [1.125-in.] length, polytetrafluoroethylene (PTFE)-coated).

8.3.2 *Overhead Stirrer*—An overhead stirrer is preferred especially when testing granulate, forms, or plates so as to avoid degradation of the materials by the stirring bar.

8.3.3 A different type of stirrer design and stirring rate may be used provided equivalence in experimental results can be demonstrated.

8.4 The dissolution vessel shall be thermostatically controlled at $37 \pm 0.5^\circ\text{C}$.




FIG. 1 Dissolution Apparatus

8.5 The dissolution apparatus may include various data recording and storage devices, strip chart recorders, computers, and so forth to facilitate continuous monitoring throughout the duration of the experiment.

9. Preparation of Test Specimens

9.1 Coatings:

9.1.1 The test specimen shall be manufactured from the same materials and processes as substrates produced for sale.

9.1.1.1 The standard test specimen is a flat plate about 3.5 by 3.5 cm [1.375 by 1.375 in.].

(1) The coated area on this plate would be about 25 by 30 mm [1 by 1.125 in.]. See Fig. 2.

(2) This test specimen may be either attached by screwing it to rod or placed on a polymeric mesh above the stir bar.

9.1.1.2 The alternative test specimen for evaluating coating materials is defined in Fig. 3.

9.1.1.3 The user may justify another test coupon geometry.

9.1.1.4 This can be conducted on particles scraped from the coating.

(1) The scraped particles shall be dispersed in the dissolution solution and stirred by an overhead stirrer to minimize degradation by the stirring mechanism.

9.1.2 The test specimen shall be manufactured from the same materials and processes as substrates produced for sale.

9.1.2.1 The test specimen shall have the upper shank threaded so as to mate with an appropriate supporting shaft as described in Fig. 1.

9.1.2.2 By appropriate masking, or other techniques, the coating shall be applied only to the central 1.3 ± 0.013 cm (0.5 ± 0.005 in.) of the test specimen.

9.2 Fabricated Forms—A fabricated form shall be tested in the form it is provided as a product. The form shall be placed on a polymeric screen over the stir bar or under a stirrer so that there is no contact with the stirring mechanism.

9.3 Powders and Granules—Powders and granules shall be tested in the form intended for use.

9.3.1 Fine powders under $250 \mu\text{m}$ shall be dispersed in the dissolution solution and stirred by an overhead stirrer to minimize degradation by the stirring mechanism.

9.3.2 Granulate and larger powders ($> 250 \mu\text{m}$) shall be spread on a polymeric screen so as to avoid degradation by the stirring bar or the overhead stirring mechanism.




FIG. 2 Flat Plate Test Specimen




FIG. 3 Test Specimen and Coating

10. Procedure for Monitoring Changes in pH and Calcium Concentrations

10.1 Prepare the test specimen as described in Section 8.

10.2 Procure the appropriate volume of dissolution media needed for the experiment and equilibrate at $37 \pm 0.5^\circ\text{C}$.

10.3 Sufficient sample for at least six replicate tests should be prepared.

10.4 Calibrate the pH probe and any other analytical instrumentation to be used immediately before starting the experiment or use as recommended by the manufacturers.

10.5 Adjust the stirrer assembly to 100 ± 20 rpm.

10.6 Assemble the dissolution apparatus (see Fig. 1). The dissolution media, all calibrated sensing electrodes, and the stirrer should be in place and operating. However, do not include the test specimen. Then equilibrate this assembly at $30 \pm 0.5^\circ\text{C}$.

10.7 Record the initial pH of the test media.

10.8 Calibrate the calcium specific ion electrode as recommended by the manufacturer except use the same test media for preparing the calibration standards as will be used in the experiment.

10.9 Set the experiment timing device to zero.

10.10 When ready to begin the experiment, add the test specimen to the dissolution vessel, and make any necessary adjustments to the equipment.

10.11 As soon as practical after the introduction of the test specimen to the dissolution media, record the initial pH and the initial dissolved Ca^{++} concentration.

10.12 Repeat the measurement of the dissolved Ca^{++} concentration at appropriate time intervals to define the dissolution rate curve.

10.12.1 A typical initial sampling rate is every three minutes or more frequently for the first 15 min. More frequent sampling may be necessary for some materials to determine accurately the slope of the initial linear portion of the dissolution curve.

10.12.2 A typical sampling rate for the remainder of the test is every hour or more frequently and then every 10 min for the last hour of the experiment.

10.13 The duration of a typical experiment shall be at least 24 h. Longer times may be needed for some materials if their measured concentrations of dissolved Ca^{++} continue to change by more than 10 % over a 1-h period.

10.14 Record the final pH of the test media. If the final pH differs by more than 0.2 pH units from the initial value, this entire experiment shall be considered invalid. Determine and correct the cause of the change in the pH of the buffered media and repeat the test.

10.15 Verify the calibration of the calcium electrode.

11. Report

11.1 Report the following results for each of the media systems used:

11.1.1 All procedural details that differ from those described in this test method.

11.1.2 The mass to dissolution volume ratio of the test.

11.1.3 The identity of the dissolution media, the substrate material(s).

11.1.4 The pH and concentration of dissolved Ca^{++} recorded at the start of the experiment (see 10.10).

11.1.5 The pH and concentration of dissolved Ca at the end of the experiment (see 10.13).

11.1.6 Plots of the concentrations of total dissolved Ca^{++} versus time data for the duration of the experiment.

11.1.7 The calculated initial dissolution rates as follows:

11.1.7.1 Calcium phosphates typically display two quasi-linear regions in their dissolution rate curves. The initial dissolution rate is usually distinguished from the final dissolution rate by a significantly higher rate of increase in the concentrations of dissolved Ca^{++} . The final dissolution rate is that rate observed in the quasi-equilibrium state obtained immediately before the termination of the experiment.

11.1.7.2 The initial dissolution rate (R_i) is expressed in terms of the initial changes in total Ca^{++} concentration with time. $(R_i)_{C_a}$ = initial slope of the Ca^{++} concentration versus time curve expressed in terms of the total dissolved Ca^{++} in milligrams of material per millilitre of dissolution media per hour.

11.1.8 The final dissolution rates (R_f) is expressed in terms of the final changes in total Ca^{++} concentration with time. $(R_f)_{C_a}$ = final slope of the Ca^{++} concentration versus time curve expressed in terms of the total dissolved Ca^{++} in milligrams of material per millilitre of dissolution media per hour.

12. Precision and Bias

12.1 Precision and bias of this test method will be determined after interlaboratory tests are carried out and the results tabulated. The interlaboratory tests will be carried out following Practice E691.

13. Keywords

13.1 calcium-phosphate; coating; dissolution rate; fabricated form; granule

APPENDIX

(Nonmandatory Information)

X1. RATIONALE

X1.1 Certain concentrations of TRIS have been reported to interfere with the performance of some Ca^{+2} ion specific electrodes. To minimize or eliminate this potential interference, it is important for the user of this test method to conduct all Ca^{+2} ion concentration measurements on calibration standard solutions and working solutions that have equivalent TRIS concentrations.

X1.2 Although this test method does not require the measurement, control, or reporting of the calcium phosphate material's surface area, porosity, coating thickness, or coating deposition method, all these parameters are known to affect its measured dissolution rate. Therefore, the user must carefully consider these parameters when attempting to compare different calcium phosphate materials.



REFERENCES

- (1) Cranin, A. N., Tobin, G., Gelbman, J., and Varjan, R., *Trans. Soc. Biomaterials*, 1986.
- (2) Kent, J. N., Quinn, J. H., Zide, M. F., Guerra, L. R., and Boyne, P. J., *Oral and Maxillofacial Surg.*, Vol 41, No. 10, 1983.
- (3) Yukna, R. A., Mayer, E. T., and Brite, D. V., *J. Periodontology*, Vol 55, No. 11, 1984.
- (4) Jarcho, M. P., Kay, J. F., Gumaer, K. I., Doremus, R. H., and Drobeck, H. P., *J. Bioengineering*, Vol 1, 1977.
- (5) Drobeck, H. P., Rothstein, S. S., Gumaer, K. I., Sherer, A. D., and Slighter, R. G., *J. Oral and Maxillofacial Surg.*, Vol 42, 1984.
- (6) Tracy, B. M. and Doremus, R. H., *J. Biomedical Materials Research*, Vol 18, 1984.
- (7) Cook, S. D., Kay, J. F., Thomas, K. A., and Jarcho, M., *Clin. Orthop.*, Vol 230, 1988.
- (8) Kay, J. F., Golec, T. S., and Riley, R. L., *J. Prosthet. Dent.*, Vol 58, 1987.
- (9) Jarcho, M., *Clin. Orthop.*, Vol 157, 1981.
- (10) Sendax, V., *Dent. Clin. N. Am.*, Vol 36, 1992.
- (11) Hergenrother, P. J., and Martin, S. F., *Anal. Biochem.*, Vol 251, 1997, p. 45.

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; <http://www.copyright.com/>