



Standard Specification for Medical Oxygen Delivery Systems for EMS Ground Vehicles¹

This standard is issued under the fixed designation F1949; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

INTRODUCTION

Within the United States, there are several widely recognized national and international standards organizations that have established standards and guidelines for oxygen delivery systems. These standards and guidelines were largely developed for intra-facility use. This standard, developed by ASTM Subcommittee F30.01, addresses the requirements for oxygen systems, both liquid and gaseous, for emergency medical services (EMS) ground vehicles.

1. Scope

1.1 This standard covers minimum requirements for primary medical oxygen delivery systems for EMS ground vehicles used in the following applications:

1.1.1 The transportation of the sick and injured to or from an appropriate medical facility while basic, advanced, or specialized life support services are being provided,

1.1.2 The delivery of interhospital critical transport care,

1.1.3 The delivery of nonemergency, medically required, transport services, and

1.1.4 The transportation and delivery of personnel and supplies essential for proper care of an emergent patient.

1.2 This standard establishes criteria to be considered in the performance, specification, purchase, and acceptance testing of ground vehicles for EMS use.

1.3 This entire standard should be read before ordering an ambulance in order to be knowledgeable of the types of equipment that are available and their performance requirements. Due to the variety of ambulance equipment or features, some options may be incompatible with all chassis manufacturers' models. Detailed technical information is available from the chassis manufacturers.

1.4 The sections in this standard appear in the following sequence:

	Section
Scope	1
Referenced Documents	2

¹ This specification is under the jurisdiction of ASTM Committee F30 on Emergency Medical Services and is the direct responsibility of Subcommittee F30.01 on EMS Equipment.

Current edition approved June 1, 2013. Published August 2013. Originally approved in 1999. Last previous edition approved in 2007 as F1949 – 99 (2007). DOI: 10.1520/F1949-99R13.

Terminology	3
Significance and Use	4
Design Requirements	5
Performance Requirements	6
Installation Requirements	7
GOX System Design Requirements	8
GOX System Installation Requirements	9
GOX System Test Requirements	10
LOX System Design Requirements	11
LOX System Installation Requirements	12
LOX System Test Requirements	13
Keywords	14

2. Referenced Documents

2.1 The following documents, of the issue currently in effect, form a part of this standard to the extent specified herein.

2.2 ASTM Standards:²

[F1177 Terminology Relating to Emergency Medical Services](#)

2.3 Military Standards:³

[MIL-STD-461 Requirements for the Control of Electromagnetic Interference Emissions and Susceptibility](#)

[MS 33584 Standard Dimensions for Flared Tubing End](#)

[MS 33611 End Bend Radii](#)

2.4 Federal Specifications:³

[RR-C-901 Cylinders, Compressed Gas: High Pressure, Steel DOT3AA, and Aluminum Applications, General Specification for](#)

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from Standardization Document Order Desk, 700 Robbins Ave., Building #4, Section D, Philadelphia, PA 19111-5094.

2.5 *ASME Standard*.⁴

B31.3 *Chemical Plant and Petroleum Refinery Piping*

2.6 *CGA Standards*.⁵

E-7 *Standard for Medical Gas Regulators and Flowmeters*

V-1 *Compressed Gas Cylinder Valve Outlet and Inlet Connections*

S-1.1 *Pressure Relief Device Standards Part 1 - Cylinders for Compressed Gases*

3. Terminology

3.1 *Definitions*—Specific terms used throughout this specification are defined in 3.2. Other applicable terms are contained in Terminology **F1177**.

3.2 *Definitions of Terms Specific to This Standard*:

3.2.1 *design operating pressure, n*—the nominal pressure at which the oxygen equipment or container is designed to operate during normal use.

3.2.2 *LOX container, n*—a vessel used to store or transport liquid oxygen.

3.2.3 *maximum allowable working pressure, n*—the maximum gage pressure to which the equipment or container can be subjected without exceeding the allowable design stress.

3.2.4 *maximum filling volume, n*—the maximum filling volume of liquid at its maximum permissible level.

3.2.5 *pressure relief device, n*—a device designed to open in order to prevent a rise of internal fluid pressure in excess of a specified value.

3.3 *Symbol*:

g = the normal or standard constant of gravity at sea level; approximately 32.2 ft/s/s (9.81 m/s/s)

3.4 *Acronyms*:

3.4.1 *GOX, n*—gaseous oxygen.

3.4.2 *LOX, n*—liquid oxygen.

4. Significance and Use

4.1 The intent of this standard is to establish minimum requirements, test parameters, and other criteria essential for oxygen system design, performance, and appearance, and to provide for a practical degree of standardization. The object is to provide oxygen systems that are properly constructed, and which, when properly serviced and maintained, will reliably function on an EMS ground vehicle.

5. Design Requirements

5.1 The medical oxygen delivery system may be either a gaseous oxygen (GOX) system, or a liquid oxygen (LOX) system.

5.2 The oxygen delivery system shall be a piped oxygen system designed and installed as follows:

5.2.1 *Capacity*—The oxygen system shall be capable of storing and supplying a minimum of 3000 L of gaseous medical oxygen.

5.2.2 *Components*—All oxygen delivery system components shall be approved by the manufacturer of the component for the intended service. The system shall include the following:

5.2.2.1 *Oxygen Piping System*, designed and sized to deliver the required flow rates at the utilization pressures. Piping and tubing shall be of non-ferrous or corrosion-resistant steel material and shall comply with the design requirements of ASME B31.3. Hose shall be electrically conductive and approved by the manufacturer for oxygen service at the pressure and temperatures the hose will be subjected to in service. Fittings shall be of non-ferrous or corrosion-resistant steel material and shall comply with the design requirements of ASME B31.3. Cast fittings shall not be used.

5.2.2.2 *Flow Control Device*, of a pressure-compensated type that includes a means to display and monitor delivered flow rate. It shall be continuously adjustable over a minimum range of 0 to 15 L/min, with a calibrated display resolution of at least 0.5 L/min. The flow control device shall be calibrated for 50 psig inlet pressure and be able to withstand a minimum inlet pressure of 200 psig without damage or failure. It shall incorporate an inlet filter and be electrically conductive from inlet to outlet. Flow control device accuracy shall be within $\pm 10\%$ of the indicated flow, or 0.25 L, whichever is greater.

5.2.2.3 *Oxygen Outlet*, piped to a self-sealing duplex oxygen outlet station. One of the outlets shall be for a flow control device or humidifier and the second oxygen outlet shall be for gas-specific, noninterchangeable, quick disconnect plug-in devices not requiring humidification. Outlets shall be marked and identified in accordance with CGA E-7.

5.2.2.4 *Shutoff Valve*, when specified, furnished in the 50-psig line and controlled and identified from the EMT panel. If a solenoid valve is utilized, a readily accessible, emergency bypass valve shall be furnished and identified.

5.2.2.5 *Secondary Oxygen Outlet*, when specified, of the self sealing, duplex wall outlet type. Additional outlets may also be specified. The outlets shall be marked and identified in accordance with CGA E-7 (see 6.1).

5.2.3 *Oxygen Compartment*—The oxygen compartment shall be provided with at least a 9-in.² cover device which will dissipate or vent leaking oxygen to the outside of the vehicle. The oxygen compartment shall not be utilized for the storage of any other equipment. No wiring or components shall terminate in the oxygen compartment except for the oxygen control solenoid, compartment light, switch plunger or trigger device, or other equipment that is integral to the oxygen system. Wiring passing through the oxygen compartment shall be routed in a metallic conduit.

6. Performance Requirements

6.1 *Delivery Flowrate*—The oxygen system shall be capable of delivering a minimum continuous gas flow of 100 L/min of gaseous oxygen, per patient, simultaneously, down to the 10 percent tank content level.

⁴ Available from American Society of Mechanical Engineers (ASME), ASME International Headquarters, Three Park Ave., New York, NY 10016-5990, <http://www.asme.org>.

⁵ Available from Compressed Gas Association (CGA), 4221 Walney Rd., 5th Floor, Chantilly, VA 20151-2923, <http://www.egonet.com>.

6.2 *Delivery Pressure*—The oxygen system shall provide a delivery pressure of 50 ± 5 psig, at the specified flowrate at each medical oxygen gas outlet.

6.3 *Delivery Temperature*—The temperature of the gaseous oxygen supplied from each medical oxygen gas outlet shall be within $+10$ or -20 °F ($+6$ or -11 °C) of ambient temperature when the oxygen delivery system is subjected to the continuous flow described in 7.1.

6.4 *Temperature Conditions*

6.4.1 *Storage Temperatures*—The oxygen system, when serviced and maintained in accordance with the manufacturer's recommendations, shall be capable of being stored without damage or deterioration in ambient temperatures of -30 °F (-34 °C) to 125 °F (52 °C).

6.4.2 *Cold Soak*—The oxygen system shall operate at a temperature of 0 °F (-18 °C) after being cold soaked for 6 h at -30 °F (-34 °C) followed by a 1 h cold soak at 0 °F (-18 °C).

6.4.3 *Heat Soak*—The oxygen system shall operate at a temperature of 110 °F (43 °C) after being heat soaked for 6 h at 125 °F (52 °C) followed by a 1 h heat soak at 100 °F (43 °C).

6.5 *Electromagnetic Interference*—Electrical or electronic componentry, or both, of the oxygen system shall meet the electromagnetic interference emissions and susceptibility requirements of MIL-STD-461 for ground vehicles.

6.6 *Structural Integrity*—Each oxygen system component, when mounted by its normal means of attachment and fitted with the equipment item(s) that are normally attached to it, shall have the structural integrity to withstand the vibration, acceleration, and shock environments described in 6.6.1-6.6.3.2. Pneumatic components may be non-operating but shall be pressurized to their maximum allowable working pressure when subjected to these environments. Electrical components may be non-operating but shall be “powered-up” when subjected to these environments. The vibration, acceleration, and shock environments shall produce no out-of-specification performance degradation or malfunctions except that allowed in 6.6.3.2.

6.6.1 *Vibration*—The component shall withstand sinusoidal vibration applied along each of three mutually perpendicular axes with the frequency range varying from 5 to 200 Hz with an amplitude of 1.0 in. from 5 to 5.5 Hz and an applied acceleration of 1.5 g from 5.5 to 200 Hz. The duration of applied vibration shall be 5.5 h per axis (a total time of 16.5 h). The frequency of applied vibration shall be swept over the specified range logarithmically with a sweep time of 12 min. The sweep time is that of an ascending plus a descending sweep.

6.6.2 *Acceleration*—The component shall withstand steady state acceleration loads applied along three mutually perpendicular axes in two opposite directions along each axis. The duration of applied acceleration in each direction shall be at least 1 min. The applied acceleration levels shall be 4 g laterally in all four directions, 9 g vertically downward, and 3 g vertically upward.

6.6.3 *Shock*—The component shall withstand the basic design and crash worthiness shock loads defined in 6.6.3.1 and

6.6.3.2 The applied shock pulse shall be of the amplitude specified, a half sine wave configuration, and 11 ms duration.

6.6.3.1 *Basic Design Shock Load*—The component shall withstand three shocks in each direction along three mutually perpendicular axes of the container (a total of 18 shocks). The peak value of the shock loads shall be 20 g.

6.6.3.2 *Crash Worthiness Shock Load*—The component shall withstand two shocks in each direction along three mutually perpendicular axes (a total of 12 shocks). The peak value of the shock loads shall be 60 g. The crash worthiness shock loads shall produce no failure of the mounting attachments. The component shall not break free from its mounting provisions or otherwise create a hazard. Permanent bending and distortion shall be permitted. The component need not be functional following application of the crash worthiness shock loads.

7. Installation Requirements

7.1 The installation of the oxygen system shall be in accordance with the following requirements:

7.1.1 *Oxygen Piping*—Oxygen piping shall be concealed and not exposed to the elements. The piping shall be accessible for inspection and replacement.

7.1.1.1 *Piping Routing and Mounting*—In routing the piping, the general policy shall be to keep total length to a minimum. Allowances shall be made for expansion, contraction, vibration, and component replacement. All piping shall be mounted to prevent vibration and chafing. This shall be accomplished by the proper use of rubberized or cushion clips installed at no greater than 20-in. intervals as close to the bends as possible. The piping, where passing through or supported by the vehicle structure, shall have adequate protection against chafing by the use of flexible grommets. The piping shall not strike against the vehicle during vibration and shock encountered during normal use of the vehicle.

7.1.1.2 *Flaring and Bending*—Tubing shall be single flared to conform with MS 33584. As an alternative, tubing may be welded, brazed or swaged using methods and quality controls that produce leakproof joints, provided there is no undue degradation of tubing strength, corrosion resistance, or fatigue life. Tubing systems having these permanent type of joints shall be designed for ease of fabrication, inspection, and installation in the vehicle. The system layout shall provide for rapid in-service repair and component replacement. Tubing bends shall be uniform, without kinks, and fit the span between fittings without tension. The minimum bend radius to tube center lines shall be in accordance with MS 33611.

7.1.2 *Flow Control Device*—Flow control devices shall be installed so that they are readable from the EMT seat and squad bench. Flowmeters shall be installed vertically.

7.1.3 *Oxygen Outlet Stations*—Oxygen outlet stations shall be provided with sufficient vertical space to permit connection of a humidifier or nebulizer, or both, directly to the oxygen flow control device.

7.1.3.1 *Primary Patient Oxygen Outlet*—The primary patient oxygen outlet shall be located within 35 in. from the center of the patient's head when in the normal supine position.

7.1.3.2 *Optional Oxygen Outlets*—When optional oxygen outlets are provided they shall be located at the patient’s head area. (See 6.1).

8. GOX System Design Requirements

8.1 In addition to the design requirements and components listed in Section 5 of this specification, GOX systems shall also meet the following design requirements and include the following components:

8.1.1 *Cylinders*, conforming to Federal Specification RR-C-901.

8.1.2 *Cylinder Changing Wrenches*, made of a non-sparking material.

8.1.3 *Cylinder Outlet Valve*, conforming to CGA V-1.

8.1.4 *Cylinder Pressure Gage*, having a range of 0 to 2500 psi with the gage scale graduated in no more than 100-psi increments.

8.1.5 *Pressure Regulator*, conforming to CGA E-7. The pressure regulator shall be preset to provide a line pressure of 50 ± 5 psig. The regulator shall supply a minimum flowrate of 100 L/min from the oxygen station outlet with a pressure of 150 psig applied to its inlet. The pressure regulator shall limit the line pressure to a maximum of 62 psig at lock-up with a pressure of 2000 psig applied to the regulator inlet. The pressure regulator shall be easy to connect and shall incorporate an inlet filter and a line relief valve set at 200 psig maximum. The pressure regulator shall be electrically conductive from inlet to outlet.

8.1.6 *Alternative or Optional Components*—When required or when specified, the GOX system shall also include the following components:

8.1.6.1 *Remote Cylinder Pressure Indicator*, such as a multi-light indicator showing “Full”, “ $\frac{3}{4}$ ”, “ $\frac{1}{2}$ ”, “ $\frac{1}{4}$ ”, and “Empty” or a digital or analog indicator, displaying cylinder contents in increments of at least 25 %. Pressure-driven remote pressure indicators shall not exceed 200 psig.

8.1.6.2 *Low Level Alarm*, actuating when the content of the oxygen system is down to between 10 and 25 %.

9. GOX System Installation Requirements

9.1 In addition to the requirements of Section 7, GOX systems shall comply with the following installation requirements:

9.1.1 *Cylinder*—Cylinders shall be installed such that the cylinder controls are accessible from both the inside and outside of the vehicle. The cylinder shall be accessible for changing from the exterior of the body and, on Type II vehicles, through a sealed door. Access to the cylinder, when located under the squad bench, shall be through the rear half of a two section squad bench cover. The oxygen cylinder shall be located in an area that provides for proper balance and weight distribution of the vehicle.

9.1.2 *Cylinder Retention Systems*—The cylinder retention system shall be mounted to the vehicle as required to comply with the requirements of 6.6.

9.1.3 *Cylinder Changing Wrench*—The cylinder changing wrench shall be chained and clipped within the oxygen cylinder compartment.

9.1.4 *Cylinder Outlet Valve*—The cylinder outlet valve shall be installed directly into the cylinder outlet port.

9.1.5 *Cylinder Pressure Gage*—The cylinder pressure gage shall be visible from the EMT’s seat or the squad bench, or both.

9.1.6 *Pressure Regulator*—The pressure regulator shall be mounted directly to the main oxygen cylinder at the outlet of the cylinder outlet valve and shall be accessible from both the inside and the outside of the vehicle.

9.1.7 *Remote Cylinder Pressure Indicator*—The remote cylinder pressure indicator shall be mounted to the EMT panel.

10. GOX System Test Requirements

10.1 *Pressure Test*—The installed oxygen piping and outlet system shall be leak tested at 200-psig pressure for 4 h. The system shall be tested with dry air or nitrogen gas, or equal, and kept decontaminated. After the successful completion of the test, the system shall be capped then tagged with the date and signature of the person and the firm performing the test.

11. LOX System Design Requirements

11.1 In addition to the design requirements and components listed in Section 5, LOX systems shall also meet the following design requirements and include the following components:

11.1.1 *LOX Container*:

11.1.1.1 *Design*—Containers shall be designed, fabricated, tested, and marked (or stamped) in accordance with the Regulations of the U.S. Department of Transportation (DOT) Specification 4L⁶ or the ASME rules for the construction of unfired pressure vessels applicable on the date of manufacture.⁷

11.1.1.2 *Reuse*—Containers complying with 11.1.1.1 shall be permitted to be reused, reinstalled, or continued in use. The user of the container shall be responsible for its suitability for continued service.

11.1.1.3 *Repair*—Repair or alteration of containers shall be in accordance with the code under which the container was fabricated.

11.1.1.4 *Maximum Filling Volume*—Containers shall be equipped to provide a signal to prevent them from becoming liquid full and to prevent liquid from covering the inlet of the relief devices when the pressure in the container reaches the set pressure of the pressure relief device.

11.1.1.5 *Pressure Control Valves and Pressure Relief Devices*—Containers shall be equipped with pressure control valves and pressure relief devices sized and selected as to type, location, and quantity and tested in accordance with CGA S-1.1. The pressure relief devices and pressure control valves shall communicate directly with the vapor space of the container. Pressure relief devices and pressure control valves shall be designed to minimize the possibility of tampering.

11.1.1.6 *Container Shutoff Valves*—A shutoff valve shall be installed directly on the tank at the vapor outlet nozzle with no intervening fittings, other than pressure relief devices, and shall

⁶ Available from Government Institutes, Inc., 4 Research Place, Suite 200, Rockville, MD 20850.

⁷ “Rules for the Construction of Unfired Pressure Vessels,” ASME Boiler and Pressure Vessel Code, available from ASME.⁴

be marked with the words “VAPOR SHUTOFF VALVE”. Another shutoff valve shall be installed directly on the tank at the liquid outlet nozzle and shall be marked with the words “LIQUID SHUTOFF VALVE”. (Decals or stencils are acceptable.) Normally closed automatic shutoff valves that are held open by electric current, pneumatic or hydraulic pressure, or a combination thereof, or manually operated shutoff valves shall be permitted to be used to meet this requirement.

11.1.2 *Quantity Gaging System*—The quantity gaging system shall provide an indication of the amount of oxygen remaining in the liquid oxygen container within an accuracy of $\pm 10\%$ of the capacity of the container. The gaging system shall display container contents in at least 25 % increments such as by a multi-light display showing “Full,” “ $\frac{3}{4}$,” “ $\frac{1}{2}$,” “ $\frac{1}{4}$ ” and “Empty” or a digital or analog display.

11.1.3 *Pressure Gage*—The system shall include a pressure gage designed for the pressure and temperature conditions to which it will be subjected with a minimum burst pressure safety factor of 4. The dial shall be graduated to indicate at least 1.2 times the marked set pressure of the container pressure relief device. The opening at the inlet connection shall not exceed 0.055 in.

11.1.4 *Fill Valve*—The fill valve shall be a quick disconnect device which shall open automatically when fully engaged with its mating filling connector and shall automatically and instantaneously close upon disengagement from its mating connector. The connection shall be leak-tight and shall not become disconnected by shocks or jars.

11.1.5 *Supply Heat Exchanger*—The supply heat exchanger shall have the capacity to completely vaporize the liquid oxygen supply and warm the vapor to the temperature required to comply with 6.3.

12. LOX System Installation Requirements

12.1 In addition to the requirements of Section 7, LOX systems shall also comply to the following installation requirements:

12.1.1 *LOX Container*—The LOX container shall be mounted to the vehicle as required to comply with the structural integrity requirements of 6.6 when filled to its maximum filling volume with LOX.

12.1.2 *LOX Contents Indicator*—The LOX contents indicator shall be mounted on the EMT Panel.

13. LOX System Test Requirements

13.1 *Pressure Tests*—The installed LOX system shall be leak tested at a pressure equivalent to the maximum allowable working pressure of the system for 4 h. The system shall be tested with dry air or nitrogen gas, or equal, and kept decontaminated. After the successful completion of the test, the system shall be tagged with the date and signature of the person and the firm performing the test.

14. Keywords

14.1 ambulance liquid oxygen delivery system; ambulance oxygen delivery system; medical gaseous oxygen delivery system; medical liquid oxygen delivery system; medical oxygen delivery system

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the ASTM website (www.astm.org/COPYRIGHT/).