



# Standard Specification for High Purity Calcium Sulfate Hemihydrate or Dihydrate for Surgical Implants<sup>1</sup>

This standard is issued under the fixed designation F2224; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon ( $\epsilon$ ) indicates an editorial change since the last revision or reapproval.

## 1. Scope

1.1 This specification covers material requirements for unfabricated and fabricated forms of hydrated calcium sulfate intended for surgical implants. Fabricated forms may include pressed and cast surgical implants in various geometric shapes. The calcium sulfate hemihydrate in the unfabricated form can be converted with the addition of water or other water-containing solutions to a fabricated calcium sulfate dihydrate form.

1.2 The requirements of this specification apply to calcium sulfate combined with two molecules of water or two calcium sulfate molecules sharing one water molecule.

Approximate chemical formulae:

Calcium Sulfate Dihydrate  
 $\text{CaSO}_4 \cdot 2\text{H}_2\text{O}$

Calcium Sulfate Hemihydrate  
 $\text{CaSO}_4 \cdot 1/2\text{H}_2\text{O}$  or  $\text{CaSO}_4 \cdot \text{H}_2\text{O} \cdot \text{CaSO}_4$

1.3 This specification specifically excludes calcium sulfate anhydrite and calcium sulfate forms that contain additives such as reinforcing phases, medicaments, biological agents, and so forth.

1.4 The presence of processing aids does not exclude a product from the physical and mechanical requirements of this specification.

1.5 Some provisions of Specification [C59/C59M](#) and Test Methods [C472](#) apply. Special requirements that are detailed in this specification are included to characterize the material which will be used in surgical implants.

1.6 The biological response to calcium sulfate in bone tissue has been well characterized by a history of clinical use ([1-14](#))<sup>2</sup> and by laboratory studies ([15-18](#)).

1.7 The following precautionary caveat pertains only to the test method portion, Sections [4](#), [5](#), and [6](#), of this specification.

<sup>1</sup> This specification is under the jurisdiction of ASTM Committee [F04](#) on Medical and Surgical Materials and Devices and is the direct responsibility of Subcommittee [F04.13](#) on Ceramic Materials.

Current edition approved March 1, 2014. Published March 2014. Originally approved in 2003. Last previous edition approved in 2009 as F2224 – 09. DOI: 10.1520/F2224-09 (2014).

<sup>2</sup> The boldface numbers in parentheses refer to the list of references at the end of this standard.

*This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory requirements prior to use.*

## 2. Referenced Documents

### 2.1 ASTM Standards:<sup>3</sup>

[C59/C59M](#) Specification for Gypsum Casting Plaster and Gypsum Molding Plaster

[C472](#) Test Methods for Physical Testing of Gypsum, Gypsum Plasters and Gypsum Concrete

[F648](#) Specification for Ultra-High-Molecular-Weight Polyethylene Powder and Fabricated Form for Surgical Implants

[F756](#) Practice for Assessment of Hemolytic Properties of Materials

[F763](#) Practice for Short-Term Screening of Implant Materials

[F813](#) Practice for Direct Contact Cell Culture Evaluation of Materials for Medical Devices

[F895](#) Test Method for Agar Diffusion Cell Culture Screening for Cytotoxicity

[F981](#) Practice for Assessment of Compatibility of Biomaterials for Surgical Implants with Respect to Effect of Materials on Muscle and Bone

[F1088](#) Specification for Beta-Tricalcium Phosphate for Surgical Implantation

[F1635](#) Test Method for *in vitro* Degradation Testing of Hydrolytically Degradable Polymer Resins and Fabricated Forms for Surgical Implants

### 2.2 Other Documents:

[BS 6463-102: 2001](#) Quicklime, Hydrated Lime and Natural Calcium Carbonate—Part 102: Methods for Chemical Analysis<sup>4</sup>

<sup>3</sup> For referenced ASTM standards, visit the ASTM website, [www.astm.org](http://www.astm.org), or contact ASTM Customer Service at [service@astm.org](mailto:service@astm.org). For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

<sup>4</sup> Available from the British Standards Institution, c/o IHS Engineering/IHS International, 15 Inverness Way East, Englewood, CO 80112, <http://www.global.ihs.com>.

US Pharmacopeia XXIV (USP 24) NF-19<sup>5</sup>  
 CFR Title 21, Part 820 Quality System Requirements<sup>6</sup>  
 Food Chemical Codex (FCC)<sup>7</sup>  
 European Pharmacopeia<sup>8</sup>  
 ISO 10993-1 Biological Evaluation of Medical Devices<sup>9</sup>

### 3. Terminology

#### 3.1 Definitions:

3.1.1 *calcium sulfate anhydrite*—a chemical substance having approximate molecular formula of  $\text{CaSO}_4$ .

3.1.2 *calcium sulfate dihydrate*—a chemical having the approximate molecular formula of  $\text{CaSO}_4 \cdot 2\text{H}_2\text{O}$ . This substance is also known as gypsum.

3.1.3 *calcium sulfate hemihydrate*—a chemical substance having approximate molecular formula of  $\text{CaSO}_4 \cdot 1/2\text{H}_2\text{O}$  or  $\text{CaSO}_4 \cdot \text{H}_2\text{O} \cdot \text{CaSO}_4$ . The mineral name of this substance is bassanite and the substance is also known as Plaster of Paris in the clinical literature.

3.1.4 *processing aids*—any constituent intentionally used in the processing of the raw material to fulfill a certain technological purpose during treatment or processing. Some examples would be: binders, lubricants, compaction aids, disintegrants, plasticizers, deflocculants, wetting agents, water retention agents, antistatic agents, antifoam agents, foam stabilizers, chelating or sequestering agents, phase stabilizers, and so forth.

3.1.4.1 *Discussion*—Use of a processing aid may result in the unintentional but technically unavoidable presence of residues of the substance or its derivatives in the final product.

3.1.5 *set time*—for a mixture of calcium sulfate hemihydrate and an aqueous solution, set time is defined as the elapsed time between the onset of mixing and the development of sufficient mechanical properties to meet a specific criteria (for example, hardness or resistance to indentation).

### 4. Chemical Requirements

4.1 Calcium sulfate for surgical implants (raw material) shall have a purity of not less than 98 % for calcium sulfate (absent of water) when measured by USP 24 NF 19. (This purity measurement method may not be applicable to the fabricated forms containing substantial quantities of additives.)

4.2 The total concentration of heavy metals (for example, lead, arsenic, cadmium, antimony, bismuth, and mercury) in the calcium sulfate raw material shall be limited to less than 10 ppm. Other trace elements, such as iron, may also affect implant performance and should be kept to a minimum. For example, for calcium sulfate to meet USP grade, the iron

concentration should not be higher than 100 ppm. Methods for measuring these trace elements are described in Specification **F1088** (Coupled Plasma—Atomic Absorption Spectrometry), the United States Pharmacopeia (USP), European Pharmacopeia, or Food Chemical Codex (FCC). A second method that may be used to analyze acid insoluble impurities is described in BS 6463-102.

4.2.1 When calcium sulfate dihydrate is converted into calcium sulfate hemihydrate, the mass of the material is reduced by approximately 15 % due to dehydration. Depending on the conversion process, the quantities (total mass) of most or all of the trace elements present in the dihydrate are not affected. Therefore, the concentration of those trace elements in the resulting hemihydrate material can be expected to increase by approximately 15 %. This should be taken into account when setting acceptance criteria for a calcium sulfate dihydrate raw material that will be used to produce a hemihydrate final product that is expected to conform to this specification.

### 5. Physical and Mechanical Characterization

5.1 The following physical and mechanical characterization may be applicable to calcium sulfate for surgical implant applications in either the fabricated form or intra-operative fabricated form. When characterization test results are reported in labeling, the test methods associated with these results shall be referenced. Labeling can be defined as but is not limited to the product label, brochures, technical monographs, and other related documentation.

5.2 *Set Time*—If set time is an applicable property, it should be reported along with the method by which it was determined in order to inform the final user. Test Methods **C472** as described in Specification **C59/C59M** can be used to define a typical set time. The actual method used for set time determination shall be described or referenced in labeling.

5.3 *Compressive Strength*—Calcium sulfate dihydrate in a fabricated final form is intended to be used in non-load bearing applications. If applicable to implant performance, documentation of typical compressive strength and the methods used to determine it should be reported in order to inform the final user. Test Methods **C472** as described in Specification **C59/C59M** can be used for the typical compression strength determination. The actual method used shall be described or referenced in labeling.

5.4 *In vitro Degradation*—For calcium sulfate dihydrate in a fabricated final form, weight loss from dissolution may be reported. If reported, the method used should be described or referenced. Additional information is given in Appendix **X1.4**.

### 6. Test Specimen Fabrication

6.1 Prepare test specimens from material(s) produced according to the same manufacturing procedures and processes employed in fabricating the implant device.

### 7. Quality Program Requirements

7.1 The manufacturer shall conform to Quality Systems Regulations (see Title 21, part 820, of the U.S. Code of Federal Regulations) or its equivalent.

<sup>5</sup> Available from U.S. Pharmacopeia (USP), 12601 Twinbrook Pkwy., Rockville, MD 20852, <http://www.usp.org>.

<sup>6</sup> Available from U.S. Government Printing Office Superintendent of Documents, 732 N. Capitol St., NW, Mail Stop: SDE, Washington, DC 20401, <http://www.access.gpo.gov>.

<sup>7</sup> Available from National Academies Press, 500 Fifth St., NW, Lockbox 285, Washington, DC 20055, <http://www.nap.edu>.

<sup>8</sup> Available from EDQM, European Pharmacopeia, Council of Europe, B.P. 907, F-67029, Strasbourg, France, <http://www.edqm.eu>.

<sup>9</sup> Available from American National Standards Institute (ANSI), 25 W. 43rd St., 4th Floor, New York, NY 10036, <http://www.ansi.org>.

## 8. Keywords

8.1 bone; calcium sulfate; gypsum; implant; plaster

## APPENDIX

### (Nonmandatory Information)

#### X1. RATIONALE

X1.1 The biocompatibility of calcium sulfate may depend upon processing conditions and source material, both which may not be identified by the compositional requirements of this specification. The biocompatibility of these products should be ensured by a combination of preclinical testing and process controls. Material derived under the desired process conditions should be tested in accordance with the recommendations of Practices **F756**, **F763**, **F813**, and **F981**; Test Methods **F895** and **F1635**; Specification **F648**; and ISO 10993-1, and manufacturing controls put in place to ensure that process variations outside of acceptable tolerances do not occur. Substantial changes in process conditions or source material parameters will necessitate additional biocompatibility or preclinical functional testing to ensure maintenance of an acceptable tissue response.

X1.2 X-ray powder diffraction (XRD) analysis of the material provides differentiation between calcium sulfate dihydrate and calcium sulfate hemihydrate. The International Centre for Diffraction Data (ICDD) designation of calcium sulfate hemihydrate for surgical implants is ICDD # 41-0224 (Bassanite, syn). A typical calcium sulfate dihydrate XRD pattern will show the peaks from ICDD #s 33-0311 (Gypsum, syn), 37-1496 (Anhydrite, syn), and 41-0224 (Bassanite, syn).

X1.3 The setting time measurement in Section 5 is useful for the calcium sulfate hemihydrate since the hemihydrate must be hydrated to a calcium sulfate dihydrate before surgical use. Likewise, the calcium sulfate hemihydrate and dihydrate raw material powders do not have the compressive strength or degradation attributes that are needed for surgical applications without being converted by hydration to calcium sulfate dihydrate or with the dihydrate form pressed into a geometric shape.

X1.4 The mechanism of resorption of calcium sulfate *in vivo* is complex and involves both physical dissolution and cell mediated resorption. As a result, *in vitro* dissolution experiments should not be expected to accurately predict *in vivo* resorption rates. In addition, it has been shown that the choice of simulated body fluid can radically alter *in vitro* resorption rates. For example, calcium sulfate dihydrate pellets immersed in a phosphate buffered saline solution (PBS) can develop a shell of recrystallized hydroxyapatite on their surface which significantly slows the process of dissolution (19). This behavior has not been observed *in vivo*. Similarly, the use of distilled or deionized water for *in vitro* testing has been shown to accelerate the rate of degradation relative to rates observed clinically and in pre-clinical animal studies (19).

#### REFERENCES

- (1) Alexander, D. I., et al, "Efficacy of Calcium Sulfate Plus Decompression Bone in Lumbar and Lumbosacral Spinal Fusion: Preliminary Results in 40 Patients," *Can. J. Surg.*, 44(4), 2001, pp. 262-266.
- (2) Kelly, C. M., et al, "The Use of a Surgical Grade Calcium Sulfate as a Bone Graft Substitute," *Clin. Orthop. Rel. Res.*, 382, 2001, pp. 42-50.
- (3) Mirzayan, R., et al, "The Use of Calcium Sulfate in the Treatment of Benign Bone Lesions," *J. Bone Joint Surg.*, 83A(3), 2001, pp. 355-358.
- (4) Peltier, L. F., "The Use of Plaster of Paris to Fill Defects in Bone," *Clin. Orthop. Rel. Res.*, 21, 1961, pp. 1-31.
- (5) Peltier, L. F. and Speer, D. P., "Calcium Sulfate," *Bone Grafts and Bone Substitutes*, Habal, M. B. and Reddi, A. H., eds., W. B. Saunders Company, 1993, pp. 243-246.
- (6) Damien, et al, "Bone Grafts and Bone Graft Substitutes," *J. Applied Biomat.*, 2, 1991, pp. 187-208.
- (7) Bahn, S., "Plaster: A Bone Substitute," *Oral Surg. Oral Med. & Oral Path.*, 21, 1966, pp. 672-681.
- (8) Coetzee, A. S., "Regeneration of Bone in the Presence of Calcium Sulfate," *Arch. Otolaryngol.*, 106, 1980, pp. 405-409.
- (9) Peltier, L. F., "The Use of Plaster of Paris to Fill Large Defects in Bone," *Am. J. Surg.*, 97, 1959, pp. 311-315.
- (10) Beeson, W. H., "Plaster of Paris as an Alloplastic Implant in the Frontal Sinus," *Arch. Otolaryngol.*, 107, 1981, pp. 664-669.
- (11) Shaffer, C. D. and App, G. R., "The Use of Plaster of Paris in Treating Infrabony Periodontal Defects in Humans," *J. Periodontol.*, 42, 1971, pp. 685-690.
- (12) Peltier, L. F. and Jones, R. H., "Treatment of Unicameral Bone Cysts by Curettage and Packing with Plaster of Paris Pellets," *J. Bone Joint Surg.*, 60A, 1978, pp. 820-822.
- (13) Edberg, E., "Some Experiences of Filling Osseous Cavities with Plaster," *Acta Chir. Scandinav.*, 67, 1931, pp. 313-319.
- (14) Nielson, A., "Filling of Infected or Sterile Bone Cavities by Means of Plaster of Paris," *Acta Chir. Scandinav.*, 91, 1944, pp. 17-24.
- (15) Sidqui, M., Collin, P., Vitte, C., and Forest, N., "Osteoblast Adherence and Resorption Activity of Isolated Osteoblast on Calcium Sulfate Hemihydrate," *Biomaterials*, 16, 1995, pp. 1327-1332.
- (16) DeWet, I. S. and Jansen, C., "The Use of Plaster of Paris to Fill Large Defects in Bone," *South African J. Surg.*, 11, 1973, pp. 1-8.
- (17) Bell, W. H., "Resorption Characteristics of Bone and Bone Substitutes," *Oral Surg. Oral Med. & Oral Path.*, 17, 1964, pp. 650-657.
- (18) Ricci, J. L., Rosenblum, S. F., Brezenoff, L., and Blumenthal, N. C., "Simulation of Bone Ingrowth into an Implantable Chamber through

the Use of Rapidly Resorbing Calcium Sulfate Hemihydrate,” *Proc. 4th World Biomaterials Congress*, 1992, p. 49.

Sulphate Bone Graft Substitute Materials,” *Trans. of the 7th World Biomaterials Congress*, 2004, p. 279.

(19) Moseley, et al, “ Dissolution Behavior of Two Types of Calcium

*ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.*

*This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.*

*This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the ASTM website (www.astm.org/COPYRIGHT).*