



Designation: G213 – 17

Standard Guide for Evaluating Uncertainty in Calibration and Field Measurements of Broadband Irradiance with Pyranometers and Pyrheliometers¹

This standard is issued under the fixed designation G213; the number immediately following the designation indicates the year of original adoption or, in the case of revision, the year of last revision. A number in parentheses indicates the year of last reapproval. A superscript epsilon (ϵ) indicates an editorial change since the last revision or reapproval.

1. Scope

1.1 This guide provides guidance and recommended practices for evaluating uncertainties when calibrating and performing outdoor measurements with pyranometers and pyrheliometers used to measure total hemispherical- and direct solar irradiance. The approach follows the ISO procedure for evaluating uncertainty, the Guide to the Expression of Uncertainty in Measurement (GUM) JCGM 100:2008 and that of the joint ISO/ASTM standard ISO/ASTM 51707 Standard Guide for Estimating Uncertainties in Dosimetry for Radiation Processing, but provides explicit examples of calculations. It is up to the user to modify the guide described here to their specific application, based on measurement equation and known sources of uncertainties. Further, the commonly used concepts of precision and bias are not used in this document. This guide quantifies the uncertainty in measuring the total (all angles of incidence), broadband (all 52 wavelengths of light) irradiance experienced either indoors or outdoors.

1.2 An interactive Excel spreadsheet is provided as adjunct, ADJG021317. The intent is to provide users real world examples and to illustrate the implementation of the GUM method.

1.3 The values stated in SI units are to be regarded as standard. No other units of measurement are included in this standard.

1.4 *This standard does not purport to address all of the safety concerns, if any, associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.*

1.5 *This international standard was developed in accordance with internationally recognized principles on standardization established in the Decision on Principles for the Development of International Standards, Guides and Recom-*

mendations issued by the World Trade Organization Technical Barriers to Trade (TBT) Committee.

2. Referenced Documents

2.1 *ASTM Standards:*²

[E772 Terminology of Solar Energy Conversion](#)

[G113 Terminology Relating to Natural and Artificial Weathering Tests of Nonmetallic Materials](#)

[G167 Test Method for Calibration of a Pyranometer Using a Pyrheliometer](#)

[Guide for Estimating Uncertainties in Dosimetry for Radiation Processing](#)

2.2 *ASTM Adjunct:*²

[ADJG021317 CD Excel spreadsheet- Radiometric Data Uncertainty Estimate Using GUM Method](#)

2.3 *ISO Standards:*³

[ISO 9060 Solar Energy—Specification and Classification of Instruments for Measuring Hemispherical Solar and Direct Solar Radiation](#)

[ISO/IEC Guide 98-3 Uncertainty of Measurement—Part 3: Guide to the Expression of Uncertainty in Measurement \(GUM:1995\)](#)

[ISO/IEC JCGM 100:2008 GUM 1995, with Minor Corrections, Evaluation of Measurement Data—Guide to the Expression of Uncertainty in Measurement](#)

3. Terminology

3.1 Standard terminology related to solar radiometry in the fields of solar energy conversion and weather and durability testing are addressed in ASTM Terminologies [E772](#) and [G113](#), respectively. Some of the definitions of terms used in this guide may also be found in ISO/ASTM 51707.

3.2 *Definitions of Terms Specific to This Standard:*

² For referenced ASTM standards, visit the ASTM website, www.astm.org, or contact ASTM Customer Service at service@astm.org. For *Annual Book of ASTM Standards* volume information, refer to the standard's Document Summary page on the ASTM website.

³ Available from International Organization for Standardization (ISO), ISO Central Secretariat, BIBC II, Chemin de Blandonnet 8, CP 401, 1214 Vernier, Geneva, Switzerland, <http://www.iso.org>.

¹ This test method is under the jurisdiction of ASTM Committee G03 on Weathering and Durability and is the direct responsibility of Subcommittee G03.09 on Radiometry.

Current edition approved Feb. 1, 2017. Published May 2017. DOI: 10.1520/G0213-17.

3.2.1 *aging (non-stability)*, *n*—a percent change of the responsivity per year; it is a measure of long-term non-stability.

3.2.2 *azimuth response error*, *n*—a measure of deviation due to responsivity change versus solar azimuth angle.

NOTE 1—Often cosine and azimuth response are combined as “Directional response error,” which is a percent deviation of the radiometer’s responsivity due to both zenith and azimuth responses.

3.2.3 *broadband irradiance*, *n*—the solar radiation arriving at the surface of the earth from all wavelengths of light (typically wavelength range of radiometers 300 to 3000 nm).

3.2.4 *calibration error*, *n*—the difference between values indicated by the radiometer during calibration and “true value.”

3.2.5 *cosine response error*, *n*—a measure of deviation due to responsivity change versus solar zenith angle. See **Note 1**.

3.2.6 *coverage factor*, *n*—numerical factor used as a multiplier of the combined standard uncertainty in order to obtain an expanded uncertainty.

3.2.7 *data logger accuracy error*, *n*—a deviation of the voltage or current measurement of the data logger due to resolution, precision, and accuracy.

3.2.8 *effective degrees of freedom*, *n*— v_{eff} for multiple (*N*) sources of uncertainty, each with different individual degrees of freedom, v_i that generate a combined uncertainty u_c , the Welch-Satterthwaite formula is used to compute:

$$v_{eff} = \frac{u_c^4}{\sum_{i=1}^N \frac{u_i^4}{v_i}} \quad (1)$$

3.2.9 *expanded uncertainty*, *n*—quantity defining the interval about the result of a measurement that may be expected to encompass a large fraction of the distribution of values that could reasonably be attributed to the measurand.

3.2.9.1 *Discussion*—Expanded uncertainty is also referred to as “overall uncertainty” (BIPM Guide to the Expression of Uncertainty in Measurement).⁴ To associate a specific level of confidence with the interval defined by the expanded uncertainty requires explicit or implicit assumptions regarding the probability distribution characterized by the measurement result and its combined standard uncertainty. The level of confidence that may be attributed to this interval can be known only to the extent to which such assumptions may be justified.

3.2.10 *leveling error*, *n*—a measure of deviation or asymmetry in the radiometer reading due to imprecise leveling from the intended level plane.

3.2.11 *non-linearity*, *n*—a measure of deviation due to responsivity change versus irradiance level.

3.2.12 *primary standard radiometer*, *n*—radiometer of the highest metrological quality established and maintained as an irradiance standard by a national (such as National Institute of Standards and Technology (NIST)) or international standards organization (such as the World Radiation Center (WRC) of the World Meteorological Organization (WMO)).

⁴ International Bureau of Weights and Measures (BIPM) Working Group 1 of the Joint Committee for Guides in Metrology (JCGM/WG 1).2008. “Evaluation of Measurement Data—Guide to the Expression of Uncertainty in Measurement (GUM).” JCGM 100:2008 GUM 1995 with minor corrections.

3.2.13 *reference radiometer*, *n*—radiometer of high metrological quality, used as a standard to provide measurements traceable to measurements made using primary standard radiometer.

3.2.14 *response function*, *n*—mathematical or tabular representation of the relationship between radiometer response and primary standard reference irradiance for a given radiometer system with respect to some influence quantity. For example, temperature response of a pyrheliometer, or incidence angle response of a pyranometer.

3.2.15 *routine (field) radiometer*, *n*—instrument calibrated against a primary-, reference-, or transfer-standard radiometer and used for routine solar irradiance measurement.

3.2.16 *sensitivity coefficient (function)*, *n*—describes how sensitive the result is to a particular influence or input quantity.

3.2.16.1 *Discussion*—Mathematically, it is partial derivative of the measurement equation with respect to each of the independent variables in the form:

$$y(x_i) = c_i = \frac{\delta y}{\delta x_i} \quad (2)$$

where $y(x_1, x_2, \dots, x_i)$ is the measurement equation in independent variables, x_i .

3.2.17 *soiling effect*, *n*—a percent change in measurement due to the amount of soiling on the radiometer’s optics.

3.2.18 *spectral mismatch error, radiometer*, *n*—a deviation introduced by the change in the spectral distribution of the incident solar radiation and the difference between the spectral response of the radiometer to a radiometer with completely homogeneous spectral response in the wavelength range of interest.

3.2.19 *temperature response error*, *n*—a measure of deviation due to responsivity change versus ambient temperature.

3.2.20 *tilt response error*, *n*—a measure of deviation due to responsivity change versus instrument tilt angle.

3.2.21 *transfer standard radiometer*, *n*—radiometer, often a reference standard radiometer, suitable for transport between different locations, used to compare routine (field) solar radiometer measurements with solar radiation measurements by the transfer standard radiometer.

3.2.22 *Type A standard uncertainty, adj*—method of evaluation of a standard uncertainty by the statistical analysis of a series of observations, resulting in statistical results such as sample variance and standard deviation.

3.2.23 *Type B standard uncertainty, adj*—method of evaluation of a standard uncertainty by means other than the statistical analysis of a series of observations, such as published specifications of a radiometer, manufacturers’ specifications, calibration, or previous experience, or combinations thereof.

3.2.24 *zero offset A*, *n*—a deviation in measurement output (W/m^2) due to thermal radiation between the pyranometer and the sky, resulting in a temperature imbalance in the pyranometer.

3.2.25 *zero offset B*, *n*—a deviation in measurement output (W/m^2) due to a change (or ramp) in ambient temperature.

NOTE 2—Both Zero Offset A and Zero Offset B are sometimes combined as “Thermal offset,” which are due to energy imbalances not directly caused by the incident short-wave radiation.

4. Summary of Test Method

4.1 The evaluation of the uncertainty of any measurement system is dependent on two specific components: a) the uncertainty in the calibration of the measurement system, and b) the uncertainty in the routine or field measurement system. This guide provides guidance for the basic components of uncertainty in evaluating the uncertainty for both the calibration and measurement uncertainty estimates. The guide is based on the International Bureau of Weights and Measures (acronym from French name: BIPM) Guide to the Uncertainty in Measurements, or GUM.⁴

4.2 The approach explains the following components; defining the measurement equation, determining the sources of uncertainty, calculating standard uncertainty for each source, deriving the sensitivity coefficient using a partial derivative approach from the measurement equation, and combining the

standard uncertainty and the sensitivity term using the root sum of the squares, and lastly calculating the expanded uncertainty by multiplying the combined uncertainty by a coverage factor (Fig. 1). Some of the possible sources of uncertainties and associated errors are calibration, non-stability, zenith and azimuth response, spectral mismatch, non-linearity, temperature response, aging per year, datalogger accuracy, soiling, etc. These sources of uncertainties were obtained from manufacturers’ specifications, previously published reports on radiometric data uncertainty, or experience, or combinations thereof.

4.2.1 Both calibration and field measurement uncertainty employ the GUM method in estimating the expanded uncertainty (overall uncertainty) and the components mentioned above are applicable to both. The calibration of broadband radiometers involves the direct measurement of a standard source (solar irradiance (outdoor) or artificial light (indoor)). The accuracy of the calibration is dependent on the sky condition or artificial light, specification of the test instrument (zenith response, spectral response, non-linearity, temperature




FIG. 1 Calibration and Measurement Uncertainty Estimation Flow Chart

Modified from Habte A., Sengupta M., Andreas A., Reda I., Robinson J. 2016. “The Impact of Indoor and Outdoor Radiometer Calibration on Solar Measurements,” NREL/PO-5D00-66668. <http://www.nrel.gov/docs/fy17osti/66668.pdf>.

response, aging per year, tilt response, etc.), and reference instruments. All of these factors are included when estimating calibration uncertainties.

NOTE 3—The calibration method example mentioned in [Appendix XI](#) is based on outdoor calibration using the solar irradiance as the source.

5. Significance and Use

5.1 The uncertainty in outdoor solar irradiance measurement has a significant impact on weathering and durability and the service lifetime of materials systems. Accurate solar irradiance measurement with known uncertainty will assist in determining the performance over time of component materials systems, including polymer encapsulants, mirrors, Photovoltaic modules, coatings, etc. Furthermore, uncertainty estimates in the radiometric data have a significant effect on the uncertainty of the expected electrical output of a solar energy installation.

5.1.1 This influences the economic risk analysis of these systems. Solar irradiance data are widely used, and the economic importance of these data is rapidly growing. For proper risk analysis, a clear indication of measurement uncertainty should therefore be required.

5.2 At present, the tendency is to refer to instrument datasheets only and take the instrument calibration uncertainty as the field measurement uncertainty. This leads to over-optimistic estimates. This guide provides a more realistic approach to this issue and in doing so will also assist users to make a choice as to the instrumentation that should be used and the measurement procedure that should be followed.

5.3 The availability of the adjunct (ADJG021317)⁵ uncertainty spreadsheet calculator provides real world example, implementation of the GUM method, and assists to understand the contribution of each source of uncertainty to the overall uncertainty estimate. Thus, the spreadsheet assists users or manufacturers to seek methods to mitigate the uncertainty from the main uncertainty contributors to the overall uncertainty.

6. Basic Uncertainty Components for Evaluating Measurement Uncertainty of Pyranometers and Pyrheliometers

6.1 As described in the BIPM GUM⁴ and summarized in Reda et al. 2008,⁶ and Reda 2011,⁷ the process for both calibration and field measurement uncertainty follows six basic uncertainty components:

6.1.1 *Determine the Measurement Equation for the Calibration Measurement System (or both)*—Mathematical description of the relation between sensor voltage and any other independent variables and the desired output (calibration response, or engineering units for measurements). [Eq 3](#) and [Eq 4](#) are

⁵ Available from ASTM International Headquarters. Order Adjunct No. [ADJG021317](#). Original adjunct produced in [ADJG021317](#). Adjunct last revised in 2017.

⁶ Reda, I.; Myers, D.; Stoffel, T. (2008). "Uncertainty Estimate for the Outdoor Calibration of Solar Pyranometers: A Metrologist Perspective. Measure." *NCSLI Journal of 100 Measurement Science*. Vol. 3(4), December 2008; 58-66

⁷ Reda, I. Technical Report NREL/TP-3B10-52194. *Method to Calculate Uncertainties in Measuring Shortwave Solar Irradiance Using Thermopile and Semiconductor Solar Radiometers*. 2011.

equations used for calculating responsivity or irradiance and they are used here for example purposes.

| Calibration Equation: | Field Measurement Equation: |
|--|--|
| $R = \frac{(V - R_{net} \times W_{net})}{G}$ | $G = \frac{(V - R_{net} \times W_{net})}{R}$ (3) |

$$G = N \times \cos(Z) + D$$

where R is the pyranometer's responsivity, in microvolt per watt per square meter $\mu\text{V}/(\text{Wm}^{-2})$,

V is the pyranometer's sensor output voltage, in μV

N is the beam irradiance measured by a primary or standard reference standard pyrheliometer, measuring the beam irradiance directly from the sun's disk in Wm^{-2} ,

Z is the solar zenith angle, in degrees

D is the diffuse irradiance, sky irradiance without the beam irradiance from the sun's disk, measured by a shaded pyranometer

G is the calculated irradiance, in Wm^{-2} ;

R_{net} is the pyranometer's net infrared responsivity, in $\mu\text{V}/(\text{Wm}^{-2})$, and

W_{net} is the net infrared irradiance measured by a collocated pyrgeometer, measuring the atmospheric infrared, in Wm^{-2} , if known. If not known, or not applicable, explicit magnitude (even if assumed to be zero, e.g., for a silicon detector radiometer) for the uncertainty associated with these terms must be stated. G is the calculated irradiance. The measurement equation with unknown or not applicable W_{net} and R_{net} is:

$$G = \frac{V}{R}$$
 (4)

6.1.2 *Determine Sources of Uncertainties*—Most of the sources of uncertainties (expanded uncertainties, denoted by U) were obtained from manufacturers' specifications, previously published reports on radiometric data uncertainty or professional experience. Some of the common sources of uncertainties are:

Solar Zenith Angle Response: pyranometer specification sheet

Spectral Response: user estimate/pyranometer specification sheet

Non-linearity: pyranometer specification sheet

Temperature response: pyranometer specification sheet

Aging per year: pyranometer specification sheet

Data logger accuracy: data logger specification sheet

Maintenance (e.g., soiling): user estimate

Calibration: calibration certificate

6.1.3 *Calculate the Standard Uncertainty, u* —calculate u for each variable in the measurement equation, using either statistical methods (Type A uncertainty component) or other than statistical methods (Type B uncertainty component), such as manufacturer specifications, calibration results, and experimental or engineering experience.

6.1.3.1 V : Sensor output voltage: from either the manufacturer's specifications of the data acquisition manual, specification data, or the most recent calibration certificate.

6.1.3.2 R_{net} : From the manufacturer's specifications, experimental data, or an estimate based on experience.

6.1.3.3 *W_{net}*: From an estimate based on historical net infrared at the site using pyrgeometer data and experience.

6.1.3.4 *N*: From the International Pyrheliometer Comparison (IPC) report described in reference or a pyrheliometer comparisons certificate based on annual calibrations or comparisons to primary reference radiometers traceable to the world radiometric reference, or combinations thereof.

6.1.3.5 *Z*: From a solar position algorithm for calculating solar zenith angle and a time resolution of 1 second.

6.1.3.6 *D*: From a diffuse pyranometer calibration described in Test Method G167.

6.1.3.7 *Discussion*—Type A and Type B classification are based on distribution of the measurement, and a requirement of the GUM approach is to associate each source of uncertainty to a specific distribution, either measured or assumed. See **Appendix X2** for a summary of typical distribution types (rectangular or uniform, Gaussian or normal, triangular, etc.) and the associated form of standard uncertainty calculation.

In the Type B, when the distribution of the uncertainty is not known, it is common to assume a rectangular distribution. In this case, the expanded uncertainty of a source of uncertainty with unknown distribution is divided by the square root of three.

$$u = \frac{U \times a}{\sqrt{3}} \quad (5)$$

where *U* is the expanded uncertainty of a variable, and *a* is the variable in a unit of measurement. For normal distribution, the equation is as follows:

$$u = \frac{U \times a}{k} \quad (6)$$

Type A standard uncertainty is calculated by taking repeated measurement of the input quantity value, from which the sample mean and sample standard deviation (SD) can be calculated. The Type A standard uncertainty (*u*) is estimated by:

$$SD = \sqrt{\frac{\sum_{i=1}^n (x_i - \bar{x})^2}{n - 1}} \quad (7)$$

where *X* represents individual input quantity, \bar{x} is the mean of the input quantity, and *n* equals the number of repeated measurement of the quantity value.

6.1.4 *Sensitivity Coefficient, c*—The GUM method requires calculating the sensitivity coefficients (*c_i*) of the variables in a measurement equation. These coefficients affect the contribution of each input factor to the combined uncertainty of the irradiance value. Therefore, the sensitivity coefficient for each

input is calculated by partial differentiation with respect to each input variable in the measurement equation. The respective sensitivity coefficient equations based on **Eq 3** are:

| Calibration Sensitivity Equations | Field Measurement Sensitivity Equations |
|--|---|
| $c_v = \frac{\delta R}{\delta V} = \frac{1}{N \cos(Z) + D}$ | $c_R = \frac{\delta G}{\delta R} = \frac{-(V - R_{net}) \times W_{net}}{R^2}$ |
| $c_{R_{net}} = \frac{\delta R}{\delta R_{net}} = \frac{-W_{net}}{N \cos(Z) + D}$ | $c_{R_{net}} = \frac{\delta G}{\delta R_{net}} = \frac{-W_{net}}{R}$ |
| $c_{W_{net}} = \frac{\delta R}{\delta W_{net}} = \frac{-R_{net}}{N \cos(Z) + D}$ | $c_{W_{net}} = \frac{\delta G}{\delta W_{net}} = \frac{-R_{net}}{R}$ |
| $c_N = \frac{\delta R}{\delta N} = \frac{-(V - R_{net}) \times W_{net} \cos(Z)}{(N \cos(Z) + D)^2}$ | $c_v = \frac{\delta G}{\delta V} = \frac{1}{R}$ |
| $c_z = \frac{\delta R}{\delta Z} = \frac{N \sin(Z) (V - R_{net}) \times W_{net}}{(N \cos(Z) + D)^2}$ | |
| $c_D = \frac{\delta R}{\delta D} = \frac{-(V - R_{net}) \times W_{net}}{(N \cos(Z) + D)^2}$ | |

6.1.5 *Combined Standard Uncertainty, u_c*—Calculate the combined standard uncertainty using the propagation of errors formula and quadrature (root-sum-of-squares) method.

6.1.5.1 The combined uncertainty is applicable to both Type A and Type B sources of uncertainties. Standard uncertainties (*u*) multiplied by their sensitivity factors (*c*) are combined in quadrature to give the combined standard uncertainty, *u_c*.

$$u_c = \sqrt{\sum_{i=0}^{n-1} (u_i \times c_i)^2} \quad (8)$$

6.1.6 *Calculate the Expanded Uncertainty (U₉₅)* by multiplying the combined standard uncertainty by a coverage factor, *k*, based on the equivalent degrees of freedom (see section 3.2.9).

$$U_{95} = u_c \times k \quad (9)$$

6.1.6.1 Typically, *k* = 1, 2, or 3 implies that the true value lies within the confidence interval defined by $y \pm U$ with confidence level of either 68.27 %, 95.45 %, or 99.73 % of the time, respectively. These ranges are meant to be analogous to the relation of the coverage of a normally distributed data set by numbers of standard deviations of such a data set. Thus *U* is often denoted as *U₉₅* or *U₉₉*.

7. Keywords

7.1 GUM; irradiance; pyranometers; pyrheliometers

APPENDIXES
(Nonmandatory Information)
X1. EXAMPLE OF CALIBRATION AND MEASUREMENT UNCERTAINTY ESTIMATION
X1.1 Overview

X1.1.1 This section provides examples of a) evaluating the uncertainty in the calibration of pyranometers for measuring total hemispherical solar radiation, and b) evaluating the uncertainty in a routine pyranometer field measurement system for measuring total hemispherical solar radiation. The examples follow the approach described in Reda et al. 2008⁶ for calibration, and Reda 2011,⁷ for measuring solar irradiance using thermopile or semiconductor radiometers.

X1.1.2 The examples provided here are generally applicable to evaluating the uncertainty in calibration results (instrument response functions, or responsivity) and routine field measurement data. Given the wide variety of instrumentation, radiometric reference (primary, transfer standard) radiometers used, and measurement techniques (indoor or outdoor calibration techniques) the guide cannot address every calibration and measurement system.

X1.1.3 The principles and essential components, including estimation of magnitudes and types of error (A or B), in conjunction with the documentation and reporting of the estimated uncertainties is the responsibility of the user of this guide. The absolutely critical aspect of this approach is to document the measurement equation, identified sources of uncertainty, the type of component (Type A or Type B), the basis for the estimates of magnitude for each variable, of the assumed sample distribution type, effective degrees of freedom, and associated coverage factor, standard uncertainties

and sensitivity functions for influencing quantities. Lastly, report the combined standard uncertainties and expanded uncertainty.

X1.2 Evaluating Field Measurement Uncertainty: As calibration uncertainty is propagated as an element of field measurement uncertainty; and that to start with a somewhat simpler example, looking at the field measurement uncertainty as an introduction is suggested because the calibration uncertainty is more complicated.

X1.2.1 Determine the measurement equation used to produce the engineering data, [Eq 3](#) and [Eq 4](#).

X1.2.2 Either a single responsivity value (example below is based on single responsivity value) or the responsivity as a function of solar zenith angle can be uniquely determined for an individual pyranometer or pyrliometer from calibration and used to compute global irradiance data. The uncertainty in the responsivity value can be reduced by as much as 50% if the responsivity as a function of solar zenith angle is used.⁷

X1.2.3 List sources of field measurement uncertainty: [Table X1.1](#) shows some of the sources as an example and depending on the type of radiometer, the lists could be different. Further, each source of uncertainty relates to a specific quantity or variable in the measurement equation. For example, calibration source of uncertainty relates to the responsivity quantity or variable in the measurement equation (see [Table X1.2](#)).

TABLE X1.1 List of Sources of Uncertainties and Standard Uncertainty Calculation

| Source of Uncertainty | Quantity | Statistical Distribution | Uncertainty Type | Standard Uncertainty (u) | Expanded Uncertainty (U) ^A |
|-----------------------------|----------|--------------------------|------------------|-----------------------------|---------------------------------------|
| Calibration | R | Normal | Type B | $\frac{U}{2} = 2.81$ | 5.62 % (calibration done at 45°) |
| Solar Zenith Angle Response | R | Rectangular | Type B | $\frac{U}{\sqrt{3}} = 1.15$ | 2 % (calibration done at 45°) |
| Spectral Response | R | Rectangular | Type B | $\frac{U}{\sqrt{3}} = 0.58$ | 1 % (calibration done at 45°) |
| Non-linearity | R | Rectangular | Type B | $\frac{U}{\sqrt{3}} = 0.29$ | 0.5 % |
| Temperature Response | R | Rectangular | Type B | $\frac{U}{\sqrt{3}} = 0.29$ | 1 % |
| Aging per Year | R | Rectangular | Type B | $\frac{U}{\sqrt{3}} = 0.58$ | 1 % |
| Datalogger Accuracy | V | Rectangular | Type B | $\frac{U}{\sqrt{3}} = 5.77$ | 10 μ V |
| Maintenance | R | Rectangular | Type B | $\frac{U}{\sqrt{3}} = 0.17$ | 0.3% |

^AExpanded uncertainty for each source of uncertainty could be obtained from manufacturer specification, calibration report, historical data, or professional judgment.

TABLE X1.2 Typical Type B Standard Uncertainties (u_B) for Pyranometer Measurement Equation

| Variable | Value Units | U% | U | Offset | a=U+ Offset | Distribution | Degrees of Freedom ⁴ | U_B |
|--------------------------------------|----------------------------------|-------|-----------------------------------|-------------------|-----------------------------------|--------------|---------------------------------|--------------------|
| V | 7930.3 μV | 0.001 | 0.079 μV | 1.0 μV | 1.079 μV | Rectangular | 1000 | 0.62 |
| R_{net} | 0.4 $\mu\text{V}/\text{Wm}^{-2}$ | 10 | 0.04 $\mu\text{V}/\text{Wm}^{-2}$ | ... | 0.04 $\mu\text{V}/\text{Wm}^{-2}$ | Rectangular | 1000 | 0.02 |
| W_{net} | -150 Wm^{-2} | 5 | 7.5 Wm^{-2} | ... | 7.5 Wm^{-2} | Rectangular | 1000 | 4.33 |
| N | 1000 Wm^{-2} | 0.4 | 4 Wm^{-2} | ... | 4 Wm^{-2} | Rectangular | 1000 | 2.31 |
| Z | 20° | ... | 2.10 ⁻⁵ | ... | 2.10 ⁻⁵ | Rectangular | 1000 | 1.10 ⁻⁵ |
| D | 50 Wm^{-2} | 3 | 1.5 Wm^{-2} | ... | 1.5 Wm^{-2} | Rectangular | 1000 | 1.44 |
| $R=8.0735\mu\text{V}/\text{Wm}^{-2}$ | | | | | | | | |

⁴Degrees of freedom assumed large based on the assumption of a typical (mean) values from a large number of samples for each specific variable resulting in the single reported value (as from the datalogger specifications, or zenith angle computations).

X1.2.4 For simplicity, the W_{net} and R_{net} variables of the measurement equation were not included in the example below for the measurement uncertainty estimation, therefore, Eq 3 is used.

X1.2.5 Compute or estimate the standard uncertainty for each variable in the measurement equation as it is described in Table X1.2. For this example $G = 1000 \text{ W/m}^2$ and $R = 15 \mu\text{V}/\text{Wm}^{-2}$.

$$u^2(V) = \sum_{i=1}^n u_i^2(V) = 5.77\mu\text{V}^2 = 33.33\mu\text{V} \quad (\text{X1.1})$$

$$u^2(R) = \sum_{i=1}^n u_i^2(R) \quad (\text{X1.2})$$

$$= \left(\frac{2.81}{100} \times 15\right)^2 + \left(\frac{1.15}{100} \times 15\right)^2 + \left(\frac{0.58}{100} \times 15\right)^2 + \left(\frac{0.29}{100} \times 15\right)^2 + \left(\frac{0.29}{100} \times 15\right)^2 + \left(\frac{0.58}{100} \times 15\right)^2 + \left(\frac{0.17}{100} \times 15\right)^2 = 0.22\mu\text{V}/\text{Wm}^{-2} \quad (\text{X1.3})$$

X1.2.6 Compute the sensitivity coefficients with respect to each variable in the measurement equation, for example:

$$c_V = \frac{\delta G}{\delta V} = \frac{1}{R} = \frac{1}{15} = 0.07\text{Wm}^{-2}/\mu\text{V} \quad (\text{X1.4})$$

$$c_R = \frac{\delta G}{\delta R} = \frac{-V}{R^2} \quad (\text{X1.5})$$

$$= \frac{\text{abs}\left(-1000 \text{ W m}^{-2} \times \frac{15\mu\text{V}}{\text{Wm}^{-2}}\right)}{\left(\frac{15\mu\text{V}}{\text{Wm}^{-2}}\right)^2} = 66.67\mu\text{V}^{-1}$$

X1.2.7 Using the sensitivity coefficients c_i compute the combined standard uncertainty, $c_i u_i$, associated with each variable, and the combined uncertainty is calculated using the root sum of the squares method, standard uncertainty and the respective sensitivity coefficient for individual variable.⁴ For this example, only Type B sources of uncertainties are considered.

$$u_c = \sqrt{\sum_{j=0}^{n-1} (u \times c)^2} \quad (\text{X1.6})$$

$$u_c = \sqrt{(u(V) \times C_V)^2 + (u(R) \times C_R)^2} \quad (\text{X1.7})$$

$$= \sqrt{(33.33 \times 0.07)^2 + (0.22 \times 66.67)^2}$$

$$= 14.85\text{Wm}^{-2}$$

Note that the computed irradiance according to the measurement equation would be 1000 Wm^{-2} . The resulting combined standard uncertainty is 14.85 Wm^{-2} . Because the irradiance is

computed “instantaneously” at these data points, the total combined uncertainty component u_A is zero (there is no standard deviation to compute) in the equation:

$$u_c = \sqrt{u_A^2 + u_B^2} \quad (\text{X1.8})$$

Note that the standard uncertainties are calculated at each data point, and R was considered constant. If the responsivity is corrected for zenith angle dependence (i.e. using it as a function of zenith angle) where u_R is usually only about 0.5%, or 50% smaller than the constant R s uncertainty, the combined standard uncertainty will be considerably reduced.

X1.2.8 The expanded uncertainty (U_{95}) was calculated by multiplying the combined uncertainty (u_c) by a coverage factor ($k=1.96$, for infinite degrees of freedom), which represents a 95% confidence level.

$$U_{95} = k u_c = 1.96 \times 14.85 \text{ Wm}^{-2}$$

$$= 29.1 \text{ Wm}^{-2} \text{ or } 2.9\% \text{ of the } 1000 \text{ Wm}^{-2} \text{ irradiance value} \quad (\text{X1.9})$$

X1.3 Outdoor Pyranometer Calibration Uncertainty

Evaluation: The components and principles described for the evaluation of measurement uncertainty are applied to the calibration uncertainty estimation. The example provided here is for a thermopile pyranometer using outdoor calibration methodology.

X1.3.1 Outdoor Thermopile Pyranometer Calibration—Measurement Equation.

X1.3.1.1 *Determine Measurement Equation (Eq 3)*, Each measurement data point consists of simultaneously recording the voltage output from the test pyranometer together with the output from a reference standard pyrheliometer, which measures the irradiance from the sun’s disk, a pyrgeometer, which is a thermopile instrument that measures the atmospheric infrared irradiance (if known or applicable), and a shaded pyranometer which determines the diffuse irradiance from the sky. The responsivity, R , of the test pyranometer is then calculated using Eq 3 (calibration equation).

NOTE X1.1— W_{net} is very often omitted from the measurement equation, in which case some estimation of the uncertainty contribution due to W_{net} should be made.

X1.3.1.2 All of the variables in the measurement equation are measured independently of each other, and there are no correlations or interdependence between the variables. For

measurement equations where there are variables that are correlated, the correlations between variables should be accounted for.

X1.3.1.3 Pyranometer Calibration Standard Uncertainties (Type B): Determine the standard uncertainty and associated distribution for each variable in measurement equation as described in section 6.1.3.

X1.3.2 Determine the degrees of freedom (DF) and distribution for each variable in Eq 3. The uncertainty from calibrating the measuring systems of the above listed variables is typically reported as U_{95} with no DF or identified distribution. Following the GUM, when the distribution of the uncertainty is not known, it is common to assume a rectangular distribution with infinite degrees of freedom. Here $DF = 1000$. **Table X1.2** presents representative values reported in Reda et al. 2008 based on calculating u with the assumption of a rectangular distribution, for which the standard uncertainty $u = a/\sqrt{3}$ for uncertainty bounds $\pm a$.⁶ The value for R at the bottom of the table is the nominal value of R for this one example data point.

X1.3.3 Calculate the sensitivity coefficient according to subsection 6.1.4, derived from the measurement equation (Type B source of uncertainty).

X1.3.4 Calculate Combined Standard Uncertainty: Entering values for the variables in **Table X1.2** into equations described in subsection 6.1.4 and computing c_i . The combined standard uncertainty u_c and effective degrees of freedom are shown in the last rows of **Table X1.3**. The effective degrees of freedom are computed according to Eq 1.

X1.3.5 Type B combined standard uncertainty (shown in **Table X1.3**) is the square root of the quadrature sum of the $c_i u_i$ product terms in the third column of **Table X1.3** = $0.022 \mu\text{V}/\text{Wm}^{-2}$. For the value of $R = 8.0735 \mu\text{V}/\text{Wm}^{-2}$ in **Table X1.3**, u_B is $100 \times (0.022/8.0735) = 0.27\%$ of the R value. For the large degrees of freedom, a coverage factor $k = 1.96$ should be used, and the total combined Type B uncertainty is $1.96 \times 0.27\% = 0.53\%$.

X1.3.6 Type A standard uncertainty (u_A) is calculated as the standard deviation (SD) of a data set or of a set of measured irradiance, then divide SD by the square root of the sample size. In the case of most pyranometers, the lack of uniform Lambertian (cosine) response results in a strong response function with respect to the angle of incidence of the direct beam, or solar zenith angle for a horizontal surface.

X1.3.6.1 In this example there is only one source of Type A uncertainty, u_A . This is the variation (variance) about the

specific responsivity chosen by the user. There are two choices available to the user: a) a single responsivity at a given solar zenith angle (for example, $Z = 45^\circ$), or b) a response function (or lookup table). The former is often used for simplicity, but leads to larger data uncertainty. The later will provide lower uncertainty, provided that the range of complex example of the responsivities observed during calibration exceeds the mean, or a representative single responsivity, by more than about 1.0%.

X1.3.6.2 Uncertainty in responsivity functions is calculated from the residuals of an interpolating function used to fit the response function, which is often different for the morning (AM) and afternoon (PM) outdoor calibration data (**Fig. X1.1**).

X1.3.6.3 AM and PM responsivities should be listed in the calibration report separately. The average responsivity is typically reported at some increment of Z ; it is calculated as the average of all readings in the range $\pm 0.3^\circ$ about the even Z . **Table X1.4** is a condensed example of reported morning and afternoon responsivities, R_{AM} and R_{PM} , and the Type B combined standard uncertainty, u_B from $Z = 18^\circ$ to 78° , as well as the value at $Z = 45^\circ$.

X1.3.7 Calculate the combined standard uncertainty for Type A.

X1.3.7.1 Type A standard uncertainty and the degrees of freedom resulting from interpolating the responsivities in **Table X1.4** are calculated using following steps:

X1.3.7.1.1 Use the tabulated responsivity versus zenith angle to calculate a fit to the calibration response function.

NOTE X1.2—Based on the calibration of many different pyranometers, the responsivity can be a complex function of the zenith angle. Therefore an interpolating polynomial, piecewise, may be needed. As long as the shape of the response function is such that the extremes of the data are larger than the scatter of the data about the mean value of the responsivity (over all or part of the range of zenith angles), this will minimize the final uncertainty in the calibration response function resulting from the interpolation function.

X1.3.7.1.2 Over the calibration zenith angle range, calculate the average of the squares of the residuals of all measured AM and PM responsivities from their calculated values using the interpolation functions:

$$r_{res}^2 = \frac{\sum_{i=1}^m (R_{i,m} - R_{i,AM})^2 + \sum_{i=1}^m (R_{i,m} - R_{i,PM})^2}{m+k} \quad (\text{X1.10})$$

Where $R_{i,m}$ is the i th average measured AM or PM responsivity, $R_{i,AM}$ and $R_{i,PM}$ are the interpolated or calculated AM and PM responsivity from the fitted response function. The degrees of freedom for the average r_{res}^2 is $DF = m + k - 2$.

X1.3.7.1.3 Calculate the standard deviation of residuals, σ_{res} , to obtain the Type A standard uncertainty from r_{res}^2 (the systematic term for the fit) and σ_{res} (the random term for the fit).

$$\sigma_{res} = \sqrt{\frac{\sum_{j=1}^{j+k-2} (r - r_j)^2}{j+k-2}} \quad (\text{X1.11})$$

Where r is the mean residual [$\sqrt{(r_{res}^2)}$] and r_j is each individual residual from the fitted response function at the j th data point.

The standard Type A uncertainty is then:

$$u_A = \sqrt{r_{res}^2 + \sigma_{res}^2} \quad (\text{X1.12})$$

For example, for typical values of the mean square of the

TABLE X1.3 Type B Standard Uncertainty Contributions for the Calibration Measurement Equation for Each Variable at $Z = 20^\circ$

| Variable | c_i | $c_i \times u_i, \mu\text{V}/\text{Wm}^{-2}$ |
|----------|-------------------------------------|--|
| V | 0.0001 (1/Wm ²) | 0.00063 |
| Rnet | 0.1516 | 0.00350 |
| Wnet | 0.0004 $\mu\text{V}/\text{Wm}^{-2}$ | 0.00175 |
| N | 0.0077 $\mu\text{V}/\text{Wm}^{-2}$ | 0.01770 |
| Z | 2.7901 $\mu\text{V}/\text{Wm}^{-2}$ | 0.0003 |
| D | 0.0082 $\mu\text{V}/\text{Wm}^{-2}$ | 0.0118 |
| u_c | | 0.022 $\mu\text{V}/\text{Wm}^{-2}$ |
| DF | | 1860 |

Two Degree Zenith Angle Bin Responsivity




FIG. X1.1 Example of Morning and Afternoon Responsivity Functions and Average Interpolated Values (dark line).

TABLE X1.4 Calibration Results and Type B Combined Standard Uncertainty by Zenith Angle

NOTE 1—Note the range of variation in R as a function of Z .

| Z | AM | | PM | |
|-----|--|-----------|--|-----------|
| | R_{AM} ($\mu\text{V}/\text{Wm}^{-2}$) | u_B (%) | R_{PM} ($\mu\text{V}/\text{Wm}^{-2}$) | u_B (%) |
| 18° | 8.076 | 0.18 | 8.070 | 0.18 |
| 20° | 8.074 | 0.19 | 8.075 | 0.19 |
| 22° | 8.071 | 0.19 | 8.044 | 0.19 |
| ... | ... | ... | ... | ... |
| 44° | 7.874 | 0.23 | 7.886 | 0.28 |
| ... | ... | ... | ... | ... |
| 76° | 7.449 | 0.50 | 7.365 | 0.52 |
| 78° | 7.339 | 0.53 | 7.229 | 0.55 |
| 45° | 7.886 | 0.24 | 7.876 | 0.26 |

the extremes of the R data from the selected R are used to represent the variance contributing to the Type A evaluation. Often, these extremes are different (unsymmetrical) for AM and PM data, resulting in asymmetrical uncertainty limits, depending on the time of day. The formulation of the unsymmetrical positive (U_{95}^+) and negative (U_{95}^-) Type A uncertainty limits depends on the extreme values of the calibration data and the selected responsivity, R .

$$U_{95}^+ = (R_{\max} - R) \times 100/R \quad (\text{X1.14})$$

$$U_{95}^- = (R_{\min} - R) \times 100/R \quad (\text{X1.15})$$

And the uncertainty in data becomes $R + U_{95}^+$, $R - U_{95}^-$ for the zenith angle range specified. Note that if morning and afternoon R_{\min} or R_{\max} are significantly different, the U_{95}^+ and U_{95}^- may need to be computed for each segment (AM or PM) of the day.

From the example in Table X1.4, if the mean R (from AM and PM data) at $Z = 45^\circ$ is selected by the user, then $R_{45} = 7.866 \mu\text{V}/\text{Wm}^{-2}$. Denote $R_{\min} = 8.076 \mu\text{V}/\text{Wm}^{-2}$ at minimum $Z = 18^\circ$ and $R_{\max} = 7.339$ at maximum $Z = 78^\circ$; from Eq X1.22 and Eq X1.23:

$$U_{95}^+ = 100(8.076 - 7.866)/7.866 = +2.67\% \quad (\text{X1.16})$$

$$U_{95}^- = 100(7.339 - 7.866)/7.866 = -6.70\% \quad (\text{X1.17})$$

residuals (in percent) of 0.10 ($r_{res} = 0.30\%$ of nominal value) and the standard deviation of the residuals of 1.0×10^{-3} (0.10%),

$$u_A = \sqrt{1.10^2 + 1.10^2} = 0.14\% \quad (\text{X1.13})$$

X1.3.7.1.4 Combined Type A Uncertainty for the Single R Case—Rather than using the variation in individual data points about a function through the calibration AM and PM data, the total deviation from a selected single responsivity function of

So the calibration component of uncertainty for a data collected using the $Z = 45^\circ$ Rs of $7.866 \mu\text{V}/\text{Wm}^2$ will be $+2.67\%$ to -6.70% for $18^\circ < Z < 78^\circ$.

X1.3.8 Calculate expanded uncertainty with 95% confidence level (U_{95}).

X1.3.8.1 Responsivity Function Expanded Uncertainty: The expanded uncertainty is the coverage factor, k , selected for the confidence interval and the type of distribution for the errors, multiplied by the Type A and Type B combined uncertainties, themselves combined in quadrature: $ku_c = k\sqrt{u_A^2 + u_B^2}$. From the example in X1.3.5, the combined Type B uncertainty u_B is 0.27%; the combined Type A uncertainty from the fitting of the responsivity function in X1.3.7.1.4 is $u_A = 0.14\%$. The combined uncertainties are $u_c = k\sqrt{u_A^2 + u_B^2} = 0.30\%$ and for U_{95} confidence interval, for a large number of (effective) degrees of freedom, $k = 1.96$, and $U_{95} = 1.96 \times 0.30\% = \pm 0.59\%$ for the Rs at each Z where the responsivity function is applied.

X1.3.8.2 Single Responsivity Expanded Uncertainty: From X1.3.7.1.4, the Type A expanded uncertainty for the single responsivity was derived from the calibration data extremes, and may be asymmetrical. Type B combined uncertainty is calculated from X1.3.5 (0.27% in the example). As in X1.3.8.1, Type A and Type B expanded uncertainties, (note the Type A expanded uncertainty is already calculated; the Type B expanded uncertainty must be calculated from $k \times u_B$) combined in quadrature: $u_{95}^{+/-} = k\sqrt{u_A^2 + u_B^2}$. Note that if the Type A uncertainties are asymmetrical, the calculation is performed for both the U_{95}^+ and U_{95}^- conditions.

For U_{95}^+ in the example of X1.3.5:

$$\begin{aligned} U_{95}^+ &= k\sqrt{u_A^{+2} + u_B^{+2}} \\ &= 1.96\sqrt{2.52^2 + 0.27^2} \\ &= +4.97\% \end{aligned} \quad (\text{X1.18})$$

$$\begin{aligned} U_{95}^- &= k\sqrt{u_A^{-2} + u_B^{-2}} \\ &= -1.96\sqrt{8.15^2 + 0.27^2} \\ &= -16.05\% \end{aligned} \quad (\text{X1.19})$$

X1.4 Report: See Appendix X4 for an example report. Generally, the reporting of uncertainty will include:

X1.4.1 The instrument owner(s), date(s), location(s); including latitude, longitude, and altitude above sea level, and operating agent(s) generating the report.

X1.4.2 If the report is generated in accordance with procedures certified by an internationally recognized accreditation body, state the accreditation body, standard (e.g., ISO 17025), and display the accrediting body logo or seal.

X1.4.3 Report environmental conditions in graphical or summary format, including the following.

X1.4.3.1 Ambient temperature,

X1.4.3.2 Relative humidity,

X1.4.3.3 Infrared sky conditions (if measured), and

X1.4.3.4 Other atmospheric conditions, (e.g., aerosol optical depth) if deemed appropriate.

X1.4.4 Cite specific standards or reference documents, or both, utilized in producing the report.

X1.4.5 The make, model, manufacturer, serial number, and type of detector for the radiometer.

X1.4.6 The explicit measurement equation(s) for the test (calibration or measurement, or both).

X1.4.7 The explicit sensitivity coefficient equations derived from the measurement equation(s).

X1.4.8 Whether a calibration, graphical or tabular presentation of the responsivity as a function of zenith angle, or other parameters used as independent variables for generating the responsivity or responsivity functions could be provided.

X1.4.9 Identified Type A (statistically derived from data, or data specification sheets for test equipment) standard uncertainties, their source and magnitude.

X1.4.9.1 Instrument make, model, manufacturer, and serial number (if appropriate; e.g., used in actual performance test to generate data).

X1.4.9.2 Relevant specifications for deriving the Type A standard uncertainty.

X1.4.9.3 Distribution type and degrees of freedom for combined standard uncertainty calculations.

X1.4.9.4 Evaluation of residuals from fitting functions as Type A contributions, if fitting functions (for any parameter) are computed.

X1.4.9.4.1 Standard error of estimate, mean square residuals, standard deviation of residuals, etc. are typical statistics that may be related to such a Type A evaluation, but the statistic used should be described and the magnitude computed and displayed.

X1.4.9.5 Combined Type A standard uncertainty calculations, u_A , using appropriate sensitivity coefficients.

X1.4.10 Identified Type B standard uncertainties, the source and magnitude of the Type B uncertainty.

X1.4.10.1 Instrument make, model, manufacturer, and serial number, if appropriate (used in actual performance test to generate data).

X1.4.10.2 Relevant specifications for deriving the Type B standard uncertainty.

X1.4.10.3 Distribution type, degrees of freedom, and appropriate coverage factor k selected for combined standard uncertainty calculations.

X1.4.10.4 Evaluation of residuals from fitting functions as Type B contributions, if fitting functions (for any parameter) are computed.

X1.4.10.5 Combined Type B standard uncertainty calculations, u_B , using appropriate sensitivity coefficients.

X1.4.10.6 Combined standard uncertainty.

X1.4.10.7 Expanded uncertainty (with coverage factor, k , and indicated confidence interval).

X2. STANDARD UNCERTAINTIES

X2.1 Standard uncertainties appropriate for selected typical distributions of errors (deviations from a reference value) according to the ISO Guide to Evaluation of the Uncertainty in Measurements.

TABLE X2.1 Standard Uncertainty Relations

| Type of Distribution | Parameters | Standard Uncertainty, u |
|--|--|----------------------------|
| Experimental data, Normal Distribution | Standard deviation s Sample size n | $u = \frac{s}{\sqrt{n-1}}$ |
| Rectangular | Data limits $-a$ to $+a$ | $u = \frac{a}{\sqrt{3}}$ |
| Triangular | Data limits $-a$ to $+a$ | $u = \frac{a}{\sqrt{6}}$ |
| Normal Distribution | Expanded uncertainty, U e.g., Calibration certificate | $u = \frac{U}{k}$ |

X3. EXAMPLE OF EXPANDED MEASUREMENT UNCERTAINTIES FOR VARIOUS SOLAR RADIOMETERS AS DESCRIBED IN REDA 2011

X3.1 **Table X3.1** presents measurement uncertainties derived from Reda 2011.⁷ Note the improvement (reduction) in uncertainty when responsivity as a function of solar zenith angle Z ($F(z)$) is used rather than a single constant responsivity at a given single Z . However, these results were obtained from few and specific instruments. The result cannot be construed to imply that these values would be the same for all instruments.

TABLE X3.1 Example Expanded Uncertainties for Semiconductor and Thermopile Solar Radiometers

| 95% Confidence Level Uncertainty (coverage factor $k=2$) U_{95} | Thermopile Pyranometer | Solid State Pyranometer | Thermopile Pyrheliometer | Solid State Pyrheliometer |
|--|---------------------------|----------------------------|-----------------------------|------------------------------|
| U_{95} , $R = \text{Constant}$ | 4.1% | 8.0% | 2.7% | 8.9% |
| U_{95} , $R = F(z)$ | 2.6% | 4.0% | 1.9% | 4.7% |
| Improvement due to $F(z)$ | 38% | 50% | 29% | 47% |

X4. EXAMPLE OF OUTDOOR SOLAR RADIOMETER CALIBRATION REPORT

Metrology Laboratory Calibration Certificate

| | | | |
|--------------------------|--------------------------------|----------------------------------|------------|
| Test Instrument: | Precision Spectral Pyranometer | Manufacturer: | Eppley |
| Model: | PSP | Serial Number: | 31257F3 |
| Calibration Date: | 5/5/2011 | Due Date: | 5/5/2012 |
| Customer: | | Environmental Conditions: | see page 4 |
| Test Dates: | 5/5 | | |

This certifies that the above product was calibrated in compliance with ISO/IEC 17025:2005. Measurement uncertainties at the time of calibration are consistent with the Guide to the Expression of Uncertainty in Measurement (GUM) using Reda et al., 2008. All nominal values are traceable to the International System (SI) Units of Measurement.

No statement of compliance with specifications is made or implied on this certificate. However the estimated uncertainties are the uncertainties of the calibration process; users must add other uncertainties that are relevant to their measuring system, environmental and sky conditions, outdoor set-up, and site location.

The Type-B Standard Uncertainty of using the responsivity at each even zenith angle is reported, and the Expanded Uncertainty of the calibration is reported using two methods:

1. The Expanded Uncertainty of using the responsivity at zenith angle = 45°, within the zenith angle range from 30.0° to 60.0°
2. The Expanded Uncertainty of using Spline Interpolating Functions for the responsivity versus zenith angle.

This certificate applies only to the item identified above and shall not be reproduced other than in full, without specific written approval from the calibration facility. Certificate without signature is not valid.

Table 1. Traceability

| Measurement Type | Instrument | Calibration Date | Calibration Due Date |
|-----------------------|--|------------------|----------------------|
| Beam Irradiance † | Eppley Absolute Cavity Radiometer Model HF, S/N 29219 | 10/03/2009 | 10/03/2011 |
| Diffuse Irradiance † | Eppley Black and White Pyranometer Model 8-48, S/N 32858 | 04/02/2011 | 04/02/2012 |
| Diffuse Irradiance † | Eppley Black and White Pyranometer Model 8-48, S/N 32871 | 04/02/2011 | 04/02/2012 |
| Date Acquisition | NREL Data Acquisition System Model RAP-DAQ, S/N 2005-998 | 04/17/2011 | 04/17/2013 |
| Date Acquisition | NREL Data Acquisition System Model RAP-DAQ, S/N 2005-999 | 04/17/2011 | 04/17/2013 |
| Infrared Irradiance ‡ | Eppley Downwelling Pyrgeometer Model PIR, S/N 31199F3 | 03/24/2011 | 03/24/2013 |

† Through the World Radiometric Reference (WRR)

‡ Through the World Infrared Standard Group (WISG)

Number of pages of certificate: 4

Calibration Procedure:

Setup: Radiometers are calibrated outdoors, using the sun as the source. Pyranometers are installed for horizontal measurements, with their signal connectors oriented north, if their design permits. The shading disk for the reference diffuse measurement subtends a solid angle of 5°. Pyrhemometers are installed on solar trackers.

Calibrated by:

Date

Calibration Results 31257F3 Eppley PSP

The responsivity of the test instrument during calibration is calculated using this Measurement Equation:

$$R = (V - R_{net} * W_{net}) / I \tag{1}$$

where,

- R = radiometer responsivity ($\mu\text{V}/\text{W}/\text{m}^2$),
- V = radiometer output voltage (microvolts),
- R_{net} = radiometer net infrared responsivity ($\mu\text{V}/\text{W}/\text{m}^2$), see Table 3,
- W_{net} = effective net infrared measured by pyrgeometer (W/m^2),
- I = reference irradiance (W/m^2), beam (B) or global (G)
 where, $G = B * \text{COS}(Z) + D$,
- Z = zenith angle (degrees),
- D = reference diffuse irradiance (W/m^2).

Figure 1. Responsivity vs Zenith Angle




Figure 2. Responsivity vs Local Standard Time




Table 2. Instrument Responsivity (R) and Calibration Type-B Standard Uncertainty, u(B)

| Zenith Angle (deg.) | AM | | | PM | | | Zenith Angle (deg.) | AM | | | PM | | |
|---------------------|---|----------------|---------------|---|----------------|---------------|---------------------|---|----------------|---------------|---|----------------|---------------|
| | R ($\mu\text{V}/\text{W}/\text{m}^2$) | u(B) \pm (%) | Azimuth Angle | R ($\mu\text{V}/\text{W}/\text{m}^2$) | u(B) \pm (%) | Azimuth Angle | | R ($\mu\text{V}/\text{W}/\text{m}^2$) | u(B) \pm (%) | Azimuth Angle | R ($\mu\text{V}/\text{W}/\text{m}^2$) | u(B) \pm (%) | Azimuth Angle |
| 0 | N/A | N/A | N/A | N/A | N/A | N/A | 46 | 7.9483 | 0.40 | 107.17 | 8.0687 | 0.42 | 252.97 |
| 2 | N/A | N/A | N/A | N/A | N/A | N/A | 48 | 7.9281 | 0.41 | 104.93 | 8.0344 | 0.44 | 255.18 |
| 4 | N/A | N/A | N/A | N/A | N/A | N/A | 50 | 7.9000 | 0.42 | 102.80 | 8.0470 | 0.43 | 257.33 |
| 6 | N/A | N/A | N/A | N/A | N/A | N/A | 52 | 7.8724 | 0.42 | 100.76 | 8.0014 | 0.45 | 259.39 |
| 8 | N/A | N/A | N/A | N/A | N/A | N/A | 54 | 7.8402 | 0.43 | 98.82 | 7.9736 | 0.48 | 261.39 |
| 10 | N/A | N/A | N/A | N/A | N/A | N/A | 56 | 7.8070 | 0.45 | 96.96 | 7.9560 | 0.48 | 263.18 |
| 12 | N/A | N/A | N/A | N/A | N/A | N/A | 58 | 7.7717 | 0.46 | 95.11 | 7.9277 | 0.51 | 265.04 |
| 14 | N/A | N/A | N/A | N/A | N/A | N/A | 60 | 7.7466 | 0.48 | 93.40 | 7.9099 | 0.51 | 266.76 |
| 16 | N/A | N/A | N/A | N/A | N/A | N/A | 62 | 7.7148 | 0.52 | 91.65 | 7.8836 | 0.53 | 268.45 |
| 18 | N/A | N/A | N/A | N/A | N/A | N/A | 64 | 7.7016 | 0.53 | 90.01 | 7.8031 | 0.58 | 270.15 |
| 20 | N/A | N/A | N/A | N/A | N/A | N/A | 66 | 7.7228 | 0.55 | 88.30 | 7.8698 | 0.61 | 271.82 |
| 22 | N/A | N/A | N/A | N/A | N/A | N/A | 68 | 7.7235 | 0.58 | 86.76 | 7.8816 | 0.67 | 273.42 |
| 24 | N/A | N/A | N/A | N/A | N/A | N/A | 70 | 7.7585 | 0.62 | 85.11 | 7.7543 | 0.72 | 275.05 |
| 26 | 8.1290 | 0.39 | 150.79 | N/A | N/A | N/A | 72 | 7.6029 | 0.67 | 83.47 | 7.6565 | 0.81 | 276.63 |
| 28 | 8.1139 | 0.37 | 141.87 | 8.1464 | 0.39 | 218.01 | 74 | 7.5413 | 0.73 | 81.93 | 7.5887 | 0.92 | 278.27 |
| 30 | 8.1006 | 0.38 | 135.14 | 8.1416 | 0.40 | 225.03 | 76 | 7.5115 | 0.82 | 80.31 | 7.5093 | N/A | 279.84 |
| 32 | 8.0857 | 0.40 | 129.84 | 8.1255 | 0.40 | 230.04 | 78 | N/A | N/A | N/A | N/A | N/A | N/A |
| 34 | 8.0811 | 0.38 | 125.43 | 8.1284 | 0.39 | 234.55 | 80 | N/A | N/A | N/A | N/A | N/A | N/A |
| 36 | 8.0521 | 0.38 | 121.55 | 8.1151 | 0.41 | 238.48 | 82 | N/A | N/A | N/A | N/A | N/A | N/A |
| 38 | 8.0387 | 0.39 | 118.10 | 8.0931 | 0.40 | 241.98 | 84 | N/A | N/A | N/A | N/A | N/A | N/A |
| 40 | 8.0113 | 0.40 | 115.05 | 8.0997 | 0.41 | 245.07 | 86 | N/A | N/A | N/A | N/A | N/A | N/A |
| 42 | 7.9968 | 0.39 | 112.32 | 8.0963 | 0.42 | 247.92 | 88 | N/A | N/A | N/A | N/A | N/A | N/A |
| 44 | 7.9670 | 0.40 | 109.64 | 8.0780 | 0.42 | 250.52 | 90 | N/A | N/A | N/A | N/A | N/A | N/A |

N/A - Not Available

Figure 3. Type-B Standard Uncertainty vs Zenith Angle




Figure 4. Residuals from Spline Interpolation




Table 3. Uncertainty using Spline Interpolation

| | |
|---|------------|
| Type-B Standard Uncertainty, $u(B)$ (%) | ± 0.92 |
| Type-A Interpolating Function, $u(int)$ (%) | ± 0.32 |
| Combined Standard Uncertainty, $u(c)$ (%) | ± 0.98 |
| Effective degrees of freedom, $DF(c)$ | 84544 |
| Coverage factor, k | 1.96 |
| Expanded Uncertainty, $U95$ (%) | ± 1.92 |
| AM Valid zenith angle range | 26° to 76° |
| PM Valid zenith angle range | 28° to 74° |

Table 4. Calibration Label Values


| | |
|---------------------------|--------------------------|
| R @ 45° ($\mu V/W/m^2$) | Rnet ($\mu V/W/m^2$) † |
| 8.0068 | 0.60000 |

† Rnet determination date: Estimated

Table 5. Uncertainty using R @ 45°

| | |
|---|----------------|
| Type-B Expanded Uncertainty, $U(B)$ (%) | ± 1.00 |
| Offset Uncertainty, $U(off)$ (%) | +1.68 / -3.25 |
| Expanded Uncertainty, U (%) | +2.69 / -4.25 |
| Effective degrees of freedom, DF | +Inf |
| Coverage factor, k | 1.96 |
| Valid zenith angle range | 30.0° to 60.0° |

Figure 5. History of instrument at Zenith Angle = 45°



References:

[1] Reda, I.; Hickey, J.; Long, C.; Myers, D.; Stoffel, T.; Wilcox, S.; Michalsky, J. J.; Dutton, E. G.; Nelson, D. (2005). "Using a Blackbody to Calculate Net Longwave Responsivity of Shortwave Solar Pyranometers to Correct for Their Thermal Offset Error During Outdoor Calibration Using the Component Sum Method." Journal of Atmospheric and Oceanic Technology, 2005; pp. 1531-1540; NREL Report No. JA-560-36646. doi:10.1175/JTECH1782.1

[2] Reda, I.; Myers, D.; Stoffel, T. (2008). "Uncertainty Estimate for the Outdoor Calibration of Solar Pyranometers: A Metrologist Perspective." Measure. (NCSLI Journal of Measurement Science). Vol.3(4), December 2008; pp. 58-66; NREL Report No. JA-581-4137

[3] Reda, I.; Andreas, A. (2004). "Solar Position Algorithm for Solar Radiation Applications." Solar Energy. Vol. 76(5), 2004; pp. 577-589; NREL Report No. JA-560-35518. doi:10.1016/j.solener.2003.12.003

[4] Stoffel, T.; Reda, I. (2009). "NREL Pyrheliometer Comparisons: 22 September – 3 October 2008 (NPC-2008)." 54 pp.; NREL Report No. TP-550-45016.

[5] Reda, I.; Stoffel, T.; Myers, D. (2003). "Method to Calibrate a Solar Pyranometer for Measuring Reference Diffuse Irradiance." Solar Energy. Vol. 74, 2003; pp. 103-112; NREL Report No. JA-560-35025. doi:10.1016/S0038-092X(03)00124-5

[6] Reda, I. (1996). Calibration of a Solar Absolute Cavity Radiometer with Traceability to the World Radiometric Reference. 79 pp.; NREL Report No. TP-463-20619.

[7] Reda, I.; Gröbner, J.; Stoffel, T.; Myers, D.; Forgan, B. (2008). Improvements in the Blackbody Calibration of Pyrheliometers. ARM 2008 Science Team Meeting (Poster).

Environmental and Sky Conditions

Calibration Facility:

Latitude: 39.742°N

Longitude: 105.180°W

Elevation: 1828.8 meters AMSL

Time Zone: -7.0

Reference Irradiance:

Figure 6. Reference Irradiance





Figure 7. Diffuse Ratios



Meteorological Observations:

Figure 8. Temperature




Figure 9. Humidity




Figure 10. Pressure




Figure 11. Effective Net Infrared




Figure 12. Estimated Broadband Aerosol Optical Depth




Table 6. Meteorological Observations

| Observations | Mean |
|---------------------------------|--------|
| Temperature (°C) | 15.30 |
| Humidity (%) | 19.97 |
| Pressure (mBar) | 814.3 |
| Est. Aerosol Optical Depth (BB) | 0.0630 |

ASTM International takes no position respecting the validity of any patent rights asserted in connection with any item mentioned in this standard. Users of this standard are expressly advised that determination of the validity of any such patent rights, and the risk of infringement of such rights, are entirely their own responsibility.

This standard is subject to revision at any time by the responsible technical committee and must be reviewed every five years and if not revised, either reapproved or withdrawn. Your comments are invited either for revision of this standard or for additional standards and should be addressed to ASTM International Headquarters. Your comments will receive careful consideration at a meeting of the responsible technical committee, which you may attend. If you feel that your comments have not received a fair hearing you should make your views known to the ASTM Committee on Standards, at the address shown below.

This standard is copyrighted by ASTM International, 100 Barr Harbor Drive, PO Box C700, West Conshohocken, PA 19428-2959, United States. Individual reprints (single or multiple copies) of this standard may be obtained by contacting ASTM at the above address or at 610-832-9585 (phone), 610-832-9555 (fax), or service@astm.org (e-mail); or through the ASTM website (www.astm.org). Permission rights to photocopy the standard may also be secured from the Copyright Clearance Center, 222 Rosewood Drive, Danvers, MA 01923, Tel: (978) 646-2600; <http://www.copyright.com/>