

BS EN ISO 1797-1:2011



BSI Standards Publication

# Dentistry — Shanks for rotary instruments

Part 1: Shanks made of metals (ISO 1797-1:2011)

**bsi.**

...making excellence a habit.™

**National foreword**

This British Standard is the UK implementation of EN ISO 1797-1:2011. It supersedes BS EN ISO 1797-1:1995 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/106/4, Dental Instruments and Equipment.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© BSI 2011

ISBN 978 0 580 65528 9

ICS 11.060.20

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 August 2011.

**Amendments issued since publication**

Date	Text affected
------	---------------

---

English Version

Dentistry - Shanks for rotary instruments - Part 1: Shanks made  
of metals (ISO 1797-1:2011)

Médecine bucco-dentaire - Queues pour instruments  
rotatifs - Partie 1: Queues en matériaux métalliques (ISO  
1797-1:2011)

Zahnheilkunde - Schäfte für rotierende Instrumente - Teil 1:  
Schäfte aus Metall (ISO 1797-1:2011)

This European Standard was approved by CEN on 14 August 2011.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

## Foreword

This document (EN ISO 1797-1:2011) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2012, and conflicting national standards shall be withdrawn at the latest by February 2012.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 1797-1:1995.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

### Endorsement notice

The text of ISO 1797-1:2011 has been approved by CEN as a EN ISO 1797-1:2011 without any modification.

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 1797-1 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 4, *Dental instruments*.

This second edition cancels and replaces the first edition (ISO 1797-1:1992) and ISO 1797-1:1992/Amd.1:1997, which has been technically revised. The following changes were made:

- a) addition of a maximum value for the end of the shank for Type 3;
- b) addition of a definition for shank;
- c) clarification of symbols and figures.

ISO 1797 consists of the following parts, under the general title *Dentistry — Shanks for rotary instruments*:

— *Part 1: Shanks made of metals*

— *Part 2: Shanks made of plastics*

## Introduction

This International Standard is one of a series of basic standards on dental rotary instruments. It constitutes an important link between the standards on dental rotary instruments and those on dental handpieces.

# Dentistry — Shanks for rotary instruments —

## Part 1: Shanks made of metals

### 1 Scope

This International Standard specifies shanks for rotary instruments used in dentistry and gives measurement methods for the verification of dimensions.

### 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 3274, *Geometrical Product Specifications (GPS) — Surface texture: Profile method — Nominal characteristics of contact (stylus) instruments*

ISO 4288, *Geometrical Product Specifications (GPS) — Surface texture: Profile method — Rules and procedures for the assessment of surface texture*

ISO 6507-1, *Metallic materials — Vickers hardness test — Part 1: Test method*

ISO 8325, *Dentistry — Test methods for rotary instruments*

### 3 Terms, definitions, symbols and abbreviated terms

#### 3.1 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

##### 3.1.1

##### **shank**

part of a rotary instrument used in dentistry which is designed to fit into the chuck of a straight or geared dental handpiece, a technical handpiece or a dental turbine

#### 3.2 Symbols and abbreviated terms

Symbols and terms are shown in Figures 1 to 8 with the following key:

$d_1$  shank diameter;

$d_2$  diameter in the groove;

- $s$  maximum distance from the circumference to the flat portion;
- $l_1$  fitting length;
- $l_2$  shoulder to end length;
- $l_3$  shoulder to groove length;
- $l_4$  width of groove;
- $l_5$  length of conical or rounded end;
- $\delta$  cylindricity of shank cylinder.

#### 4 Classification

Shanks for rotary instruments are classified into the following types, according to their diameters and designs:

- Type 1: diameter 2,35 mm with groove and flat (for contra-angle-connection);
- Type 2: diameter 2,35 mm cylindrical;
- Type 3: diameter 1,6 mm cylindrical with conical or rounded end;
- Type 4: diameter 3 mm cylindrical.




Figure 1 — Type 1 shank




Figure 2 — Type 2 and Type 4 shank




Figure 3 — Type 3 shank



## 5 Requirements

### 5.1 Material

Shanks shall be made of metallic materials (e.g. steel or tungsten carbide). The exact type of material and the treatment given to it are at the discretion of the manufacturer.

### 5.2 Dimensions

The dimensions and tolerances shall be as shown in Figures 4 to 7 and as given in Table 1.

Dimensions are given in millimetres, surface roughness in micrometres.

The end of the shank for Types 1, 2 and 4 shall be either flat, conical or rounded. The end of the shank for Type 3 shall be either conical or rounded.

The shape of the shank end shall be at the discretion of the manufacturer.

Dimensions in millimetres




Figure 4 — Dimensions Type 1 shank

Dimensions in millimetres




Figure 5 — Dimensions Type 2 shank

Dimensions in millimetres




Figure 6 — Dimensions Type 3 shank

Dimensions in millimetres




Figure 7 — Dimensions Type 4 shank

Table 1 — Fitting length of shank


Dimensions in millimetres

Shank	Diameter	Fitting length, $l_1$		
		miniature, short	standard, long	extra long
Type 1 <sup>a</sup>	2,35	—	11	12
Type 2	2,35	15	30	30
Type 3	1,6	9	11	12
Type 4	3	—	30	30

<sup>a</sup> The start of any enlargement on a Type 1 shank (e.g. by marking or the working head) shall be outside  $l_1 = 13,5$  mm.

### 5.3 Shank cylindricity

The shank cylindricity (see Figure 8) shall be within the tolerances of the acceptable shank diameter.



**Key**

- 1 upper limit of  $d_1$
- 2 lower limit of  $d_1$

**Figure 8 — Shank cylindricity**

The requirement for the shank cylindricity covers the fitting length,  $l_1$ , of the shank with the exception of the shank end geometry.

Test in accordance with 6.3.

#### **5.4 Surface roughness**

The surface roughness,  $R_a$ , shall be for

- a) Type 1, Type 2, and Type 4:  $\leq 1 \mu\text{m}$ ;
- b) Type 3:  $\leq 0,5 \mu\text{m}$ ,

as it is located and shown in Figure 4 to Figure 7.

Test in accordance with 6.4.

#### **5.5 Vickers hardness**

The Vickers hardness for shanks made of metallic material, shall be not less than 250 HV5.

Test in accordance with 6.5.

#### **5.6 Marking**

Marking, if envisaged, should be outside the fitting length,  $l_1$ .

If the marking is applied within the fitting length,  $l_1$ , the outside diameter of the shank with the marking shall not exceed the effective diameter of the shank.

Marking shall be resistant to reprocessing procedures as disinfection, cleaning and sterilization according to the manufacturer's instructions.

## 6 Test methods

### 6.1 Shank diameter

Measurements of the shank diameter shall be made with suitable measuring devices that have an accuracy of 0,001 mm. The accuracy of the measuring device shall be verified.

The shank diameter,  $d_1$ , shall be measured by traversing the length,  $l_1$ .

### 6.2 Other dimensions

Measurements shall be made in accordance with ISO 8325 with suitable measuring devices with an accuracy of 0,01 mm. The accuracy of the measuring device shall be verified.

### 6.3 Shank cylindricity

The test method is left to the discretion of the manufacturer.

### 6.4 Surface roughness

Test the surface roughness in accordance with ISO 3274 and ISO 4288.

### 6.5 Vickers hardness

Test the hardness in accordance with ISO 6507-1.

## 7 Quality control

### 7.1 Types of shank

For the purpose of quality control, the shanks of the instruments shall be classified as given in Clause 4.

### 7.2 Defects

#### 7.2.1 Major defects

Major defects shall be those deviations from the specifications listed in Table 2. Major defects include only those items which prevent an instrument from operating.

#### 7.2.2 Minor defects

Minor defects shall be all deviations in fitting dimensions not listed in Table 2. Minor defects include all other deviations from the specification which lower the quality of the instrument.

Acceptable quality levels are given in Annex A.

**Table 2 — Major defects**

<b>Shank</b>	$d_1$	$d_2$	$l_3$	$l_4$	$s$
Type 1	>2,35	>1,45	<1,80	<0,85	>1,80
Type 2	>2,35	—	—	—	—
Type 3	>1,60 <1,59	—	—	—	—
Type 4	>3	—	—	—	—

## Annex A (informative)

### Acceptable quality levels (AQL)

The acceptable quality level (AQL) expressed in terms of the number of defects per 100 pieces, for each type of instrument, should be as shown in Table A.1.

**Table A.1 — Acceptable quality level**

Shank	AQL	
	Major defects	Minor defects
Type 1	2,5	6,5
Type 2	2,5	6,5
Type 3	1,5	4,0
Type 4	2,5	6,5

## Bibliography

- [1] ISO 2859-1, *Sampling procedures for inspection by attributes — Part 1: Sampling plans indexed by acceptable quality level (AQL) for lot-by-lot inspection*







# British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

## About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

## Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at [bsigroup.com/standards](http://bsigroup.com/standards) or contacting our Customer Services team or Knowledge Centre.

## Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at [bsigroup.com/shop](http://bsigroup.com/shop), where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

## Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to [bsigroup.com/subscriptions](http://bsigroup.com/subscriptions).

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

**PLUS** is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit [bsigroup.com/shop](http://bsigroup.com/shop).

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email [bsmusales@bsigroup.com](mailto:bsmusales@bsigroup.com).

## BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

## Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

## Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

## Useful Contacts:

### Customer Services

**Tel:** +44 845 086 9001

**Email (orders):** [orders@bsigroup.com](mailto:orders@bsigroup.com)

**Email (enquiries):** [cservices@bsigroup.com](mailto:cservices@bsigroup.com)

### Subscriptions

**Tel:** +44 845 086 9001

**Email:** [subscriptions@bsigroup.com](mailto:subscriptions@bsigroup.com)

### Knowledge Centre

**Tel:** +44 20 8996 7004

**Email:** [knowledgecentre@bsigroup.com](mailto:knowledgecentre@bsigroup.com)

### Copyright & Licensing

**Tel:** +44 20 8996 7070

**Email:** [copyright@bsigroup.com](mailto:copyright@bsigroup.com)



...making excellence a habit.™