

BS EN ISO 3656:2011



BSI Standards Publication

Animal and vegetable fats and oils — Determination of ultraviolet absorbance expressed as specific UV extinction (ISO 3656:2011)

bsi.

...making excellence a habit.™

National foreword

This British Standard is the UK implementation of EN ISO 3656:2011. It supersedes BS EN ISO 3656:2002 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee AW/307, Oil seeds, animal and vegetable fats and oils and their by products.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© BSI 2011

ISBN 978 0 580 66514 1

ICS 67.200.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 28 February 2011.

Amendments issued since publication

Date	Text affected
------	---------------

English Version

Animal and vegetable fats and oils - Determination of ultraviolet absorbance expressed as specific UV extinction (ISO 3656:2011)

Corps gras d'origines animale et végétale - Détermination de l'absorbance dans l'ultraviolet, exprimée sous la forme d'extinction spécifique en lumière ultraviolette (ISO 3656:2011)

Tierische und pflanzliche Fette und Öle - Bestimmung der Ultraviolett-Absorption, ausgedrückt als spezifische UV-Extinktion (ISO 3656:2011)

This European Standard was approved by CEN on 22 January 2011.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Foreword

This document (EN ISO 3656:2011) has been prepared by Technical Committee ISO/TC 34 "Food products" in collaboration with Technical Committee CEN/TC 307 "Oilseeds, vegetable and animal fats and oils and their by-products - Methods of sampling and analysis" the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2011, and conflicting national standards shall be withdrawn at the latest by August 2011.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 3656:2002.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of ISO 3656:2011 has been approved by CEN as a EN ISO 3656:2011 without any modification.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 3656 was prepared by Technical Committee ISO/TC 34, *Food products*, Subcommittee SC 11, *Animal and vegetable fats and oils*.

This fourth edition cancels and replaces the third edition (ISO 3656:2002), which has been technically revised.

Introduction

This International Standard describes a method for the determination of the absorbance of light in the ultraviolet (UV) spectrum by fats and oils. Changes in absorption in the UV region are used as quality, purity, and authenticity criteria for fats and oils. Refining causes the formation of conjugated dienes and trienes and increased values of K_{232} and K_{268} , which then indicate the presence of refined oils. The oxidation of linoleic and linolenic acids results in the formation of hydroperoxides in which the double bonds become conjugated. Furthermore, the formation of either carbon-carbon bonds or carbon-oxygen bonds (α,β -unsaturated carbonyl compounds) as secondary autoxidation products are observed. These compounds all lead to an increase of the absorption in the region between 225 nm and 325 nm.

The third edition of this International Standard allowed the determination of the UV absorbance at 232 nm and 268 nm using three different solvents (isooctane, cyclohexane or *n*-hexane). However, it is known that the solvents themselves have an effect on the UV absorbance between 260 nm and 276 nm in vegetable oils. Recent investigations have shown that the measurement of K_{268} and K_{270} for the same oil in isooctane or in cyclohexane give significantly different results. In isooctane, the maxima appear at 267 nm to 268 nm whereas in cyclohexane, the maxima appear at 268 nm to 269 nm. In the IOC standards for the determination of the ultraviolet absorbance of (virgin) olive oils, the specified wavelengths are 232 nm and 270 nm.

Taking into account the above, the fourth edition of this International Standard specifies measurement at 268 nm when isooctane is used and measurement at 270 nm when cyclohexane is used. Moreover, the variation of the specific extinction, ΔK , for olive oils has been introduced. Precision data from a new collaborative trial have also been taken into consideration for this revision.

Animal and vegetable fats and oils — Determination of ultraviolet absorbance expressed as specific UV extinction

1 Scope

This International Standard specifies a method for the determination of the absorbance at ultraviolet wavelengths of animal and vegetable fats and oils.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 661, *Animal and vegetable fats and oils — Preparation of test sample*

3 Terms and definitions

NOTE The usual policy of ISO/TC 34/SC 11 of using symbols specified in International Standard 80000^[5] is not followed in this International Standard. The symbols are those used, for example, in Commission Regulation (EEC) No 2568/91^[6].

For the purposes of this document, the following terms and definitions apply.

3.1

specific extinction ($K_{232} - K_{268} - K_{270}$)

absorbance of 1 g sample, dissolved in 100 ml isooctane or cyclohexane, measured in a 10 mm cell at the specific wavelengths 232 nm, 268 nm, and 270 nm

4 Principle

A sample is dissolved in isooctane or cyclohexane and the absorbance is measured spectrophotometrically in a specified ultraviolet wavelength range. The specific absorbance at 232 nm and 268 nm in isooctane or 232 nm and 270 nm in cyclohexane for a concentration of 1 g per 100 ml in a 10 mm cell is calculated.

5 Reagents

WARNING — Attention is drawn to the regulations which specify the handling of hazardous substances. Technical, organizational and personal safety measures shall be followed.

During the analysis, unless otherwise stated, use only reagents of recognized analytical grade and distilled or demineralized water or water of equivalent purity.

5.1 Solvent: isooctane (2,2,4-trimethylpentane) for the measurement at 232 nm and 268 nm or **cyclohexane** for the measurement at 232 nm and 270 nm, having an absorbance less than 0,12 at 232 nm and less than 0,05 at 250 nm against water, measured in a 10 mm cell.

6 Apparatus

The glassware used for the determination shall be thoroughly cleaned and rinsed with the solvent (5.1) before use so that it is free from impurities having an absorbance within the wavelength range of 220 nm to 360 nm.

Usual laboratory apparatus and, in particular, the following.

6.1 One-mark volumetric flasks, capacity 25 ml, ISO 1042^[1] class A.

6.2 Analytical balance, capable of being read to the nearest 0,000 1 g.

6.3 Spectrometer, with quartz cells of pathlength 10 mm, suitable for measurements at ultraviolet wavelengths (220 nm to 360 nm).

6.3.1 General. Before use it is recommended that the wavelength and absorbance scales of the spectrometer be checked as specified in 6.3.2 and 6.3.3.

6.3.2 Wavelength scale. Check this using a reference material consisting of an optical glass filter containing holmium oxide which has distinct absorption bands. The reference material is designed for the verification and calibration of the wavelength scales of visible and ultraviolet spectrophotometers having nominal spectral bandwidths of 5 nm or less. The holmium glass filter is measured in the absorbance mode against an air blank, over the wavelength range 640 nm to 240 nm. For each spectral bandwidth (0,10, 0,25, 0,50, 1,00, 1,50, 2,00, and 3,00), a baseline correction is performed with an empty cell holder. The wavelengths of the spectral bandwidth are listed in the certificate of the reference material¹⁾.

6.3.3 Absorbance scale. This may be checked using a reference material consisting of four solutions of potassium dichromate in perchloric acid sealed in far UV quartz cells to measure the linearity and photometric accuracy reference in the UV. The potassium dichromate filled cells (40 mg/ml, 60 mg/ml, 80 mg/ml and 100 mg/ml) are measured against a perchloric acid blank (see Reference [7]). The net absorbance values are listed in the certificate of the reference material¹⁾.

7 Sampling

Sampling is not part of the method specified in this International Standard. A recommended sampling method is given in ISO 5555^[2].

It is important the laboratory receive a truly representative sample which has not been damaged or changed during transport or storage.

8 Preparation of test sample

Prepare the test sample in accordance with ISO 661.

Ensure that the test sample is perfectly homogeneous and contains no suspended impurities. Filter oils which are liquid at ambient temperature through paper at a temperature of approximately 30 °C. Homogenize and filter hard fats at a temperature of not more than 10 °C above their melting point.

1) Suitable holmium filters and potassium dichromate-sealed cells are available commercially, e.g. from Starna Scientific (www.starnascientific.com). This information is given for the convenience of users of this document and does not constitute an endorsement by ISO of this supplier.

9 Procedure

9.1 Test portion and preparation of the test solution

9.1.1 Weigh, to the nearest 0,1 mg (6.2), an amount of the test sample, generally 0,05 g to 0,25 g, necessary to obtain absorbance values between 0,2 and 0,8, into a 25 ml volumetric flask (6.1).

9.1.2 K_{232}/K_{268} . Dissolve the test portion at ambient temperature in a few millilitres of isooctane (5.1) for the determination of the specific absorbance at 232 nm and 268 nm and then make up to the mark with the same solvent. Mix thoroughly.

9.1.3 K_{232}/K_{270} . Dissolve the test portion at ambient temperature in a few millilitres of cyclohexane (5.1) for the measurement of the specific absorbance at 232 nm and 270 nm and then make up to the mark with the same solvent. Mix thoroughly.

9.1.4 Ensure that the solutions prepared in 9.1.2 and 9.1.3 are perfectly clear. If opalescence or turbidity is present, filter quickly through paper.

If the concentration of test sample in the test solution is greater than 1 g per 100 ml of solution, this shall be stated in the test report.

9.2 Determination

Rinse a quartz cell (6.3) three times with the test solution (9.1). Fill the cell with the test solution and measure the absorbance against the solvent used for dilution, by means of the spectrometer (6.3), over the wavelength range 220 nm to 360 nm, either continuously or at intervals of 1 nm or 2 nm, reducing the intervals to 0,5 nm in the regions of maximum and minimum absorbance.

NOTE It is possible that it is not necessary to measure the absorbance over the full wavelength range.

If the absorbance value obtained exceeds 0,8, dilute the test solution as appropriate and repeat the determination.

10 Expression of results

10.1 Specific extinction (extinction coefficients) at a specific wavelength

The specific extinction, K_λ (extinction coefficient) at a specific wavelength, λ , is calculated as follows:

$$K_\lambda = \frac{E_\lambda}{c s}$$

where

E_λ is the extinction measured at wavelength λ ;

c is the concentration, in grams per 100 ml, of the solution;

s is the pathlength, in centimetres, of the cuvette.

Express the results to two decimal places.

10.2 Variation of the specific extinction, ΔK

Spectrophotometric analysis of olive oil in accordance with Commission Regulation (EEC) No 2568/91^[6] also involves the determination of the variation of the specific extinction, ΔK , which is given by:

$$\Delta K = K_m - \frac{K_{m-4} + K_{m+4}}{2}$$

where K_m is the specific extinction at wavelength m , wavelength for maximum absorption which depends on the solvent used.

The results of this determination are expressed to two decimal places.

11 Precision

11.1 Interlaboratory test

Details of an interlaboratory test are summarized in Annex A. It is possible that the values derived from this interlaboratory test are not applicable to concentration ranges other than those given.

11.2 Repeatability

The absolute difference between two independent single test results, obtained using the same method on identical test material in the same laboratory by the same operator using the same equipment within a short interval of time, will in not more than 5 % of cases be greater than the overall repeatability limit, r .

11.3 Reproducibility

The absolute difference between two single test results, obtained using the same method on identical test material in different laboratories with different operators using different equipment, will in not more than 5 % of cases be greater than the overall reproducibility limit, R .

12 Test report

The test report shall contain at least the following information:

- a) all information necessary for the complete identification of the sample;
- b) the sampling method used, if known;
- c) the test method used, with reference to this International Standard (ISO 3656:2011);
- d) all operating details not specified in this International Standard, or regarded as optional, together with details of any incidents which may have influenced the test result(s);
- e) the test result(s) obtained;
- f) if the repeatability has been checked, the final quoted result obtained.

Annex A (informative)

Results of interlaboratory test

An interlaboratory test was carried out in 2009 in accordance with ISO 5725-1^[3] and ISO 5725-2^[4] by the International Olive Council. A total of 22 laboratories from seven countries participated. The results are summarized in Tables A.1 to A.6.

Table A.1 — UV extinction at 232 nm in isooctane

Parameter	Sample				
	A	B	C	D	E
No. participating laboratories, n_p	21	22	22	22	22
No. laboratories retained after eliminating outliers, n_p	20	18	18	21	17
No. test results in all laboratories, n_t	40	36	36	42	34
Mean specific extinction at 232 nm, \bar{K}_{232}	1,763	2,104	3,805	3,847	2,817
Repeatability standard deviation, s_r	0,026	0,013	0,016	0,036	0,019
Coefficient of variation of repeatability, $C_{V,r}$, %	1,5	0,6	0,4	0,9	0,7
Repeatability limit, r	0,072	0,035	0,043	0,101	0,054
Reproducibility standard deviation, s_R	0,077	0,069	0,016	0,036	0,069
Coefficient of variation of reproducibility, $C_{V,R}$, %	4,4	3,3	4,6	5,5	2,5
Reproducibility limit, R	0,216	0,194	0,488	0,582	0,194

Table A.2 — UV extinction at 232 nm in cyclohexane

Parameter	Sample				
	A	B	C	D	E
No. participating laboratories, n_p	21	21	21	21	21
No. laboratories retained after eliminating outliers, n_p	18	20	20	21	21
No. test results in all laboratories, n_t	36	40	40	42	42
Mean specific extinction at 232 nm, \bar{K}_{232}	1,758	2,122	3,827	3,863	2,789
Repeatability standard deviation, s_r	0,025	0,022	0,042	0,041	0,033
Coefficient of variation of repeatability, $C_{V,r}$, %	1,4	1,0	1,1	1,1	1,2
Repeatability limit, r	0,070	0,060	0,119	0,113	0,093
Reproducibility standard deviation, s_R	0,049	0,073	0,151	0,138	0,100
Coefficient of variation of reproducibility, $C_{V,R}$, %	2,8	3,4	4,0	3,6	3,6
Reproducibility limit, R	0,138	0,204	0,424	0,386	0,279

Table A.3 — UV extinction at 268 nm in isooctane

Parameter	Sample				
	A	B	C	D	E
No. participating laboratories, n_p	21	22	22	22	22
No. laboratories retained after eliminating outliers, n_p	20	18	20	17	20
No. test results in all laboratories, n_t	40	36	40	34	40
Mean specific extinction at 268 nm, \bar{K}_{268}	0,124	0,425	1,140	0,450	0,598
Repeatability standard deviation, s_r	0,005	0,005	0,016	0,007	0,007
Coefficient of variation of repeatability, $C_{V,r}$, %	4,0	1,2	1,4	1,5	1,1
Repeatability limit, r	0,014	0,014	0,043	0,018	0,018
Reproducibility standard deviation, s_R	0,010	0,016	0,030	0,013	0,034
Coefficient of variation of reproducibility, $C_{V,R}$, %	8,0	3,8	2,6	3,0	5,6
Reproducibility limit, R	0,028	0,045	0,083	0,038	0,094

Table A.4 — UV extinction at 270 nm in cyclohexane

Parameter	Sample				
	A	B	C	D	E
No. participating laboratories, n_p	21	21	21	21	21
No. laboratories retained after eliminating outliers, n_p	20	19	20	20	17
No. test results in all laboratories, n_t	40	38	40	40	34
Mean specific extinction at 270 nm, \bar{K}_{270}	0,128	0,430	1,120	0,450	0,589
Repeatability standard deviation, s_r	0,005	0,008	0,010	0,012	0,006
Coefficient of variation of repeatability, $C_{V,r}$, %	4,0	2,0	1,0	2,6	1,1
Repeatability limit, r	0,014	0,023	0,029	0,033	0,018
Reproducibility standard deviation, s_R	0,011	0,016	0,027	0,014	0,015
Coefficient of variation of reproducibility, $C_{V,R}$, %	8,5	3,7	2,4	3,2	2,5
Reproducibility limit, R	0,031	0,044	0,074	0,040	0,042

Table A.5 — Variation of the specific extinction ΔK at (270 ± 4) nm in cyclohexane

Parameter	Sample				
	A	B	C	D	E
No. participating laboratories, n_p	20	21	21	21	21
No. laboratories retained after eliminating outliers, n_p	19	20	19	20	18
No. test results in all laboratories, n_t	38	40	38	40	36
Mean specific extinction, $\overline{\Delta K}$	-0,002	0,002	0,085	0,035	0,047
Repeatability standard deviation, s_r	0,001	0,001	0,001	0,001	0,001
Coefficient of variation of repeatability, $C_{V,r}$, %	28,9	21,6	1,1	2,9	2,9
Repeatability limit, r	0,002	0,002	0,003	0,003	0,004
Reproducibility standard deviation, s_R	0,003	0,001	0,004	0,003	0,004
Coefficient of variation of reproducibility, $C_{V,R}$, %	147,5	52,0	5,1	7,6	8,1
Reproducibility limit, R	0,008	0,004	0,012	0,007	0,011

Table A.6 — Variation of the specific extinction ΔK at (268 ± 4) nm in isooctane

Parameter	Sample				
	A	B	C	D	E
No. participating laboratories, n_p	21	21	22	22	22
No. laboratories retained after eliminating outliers, n_p	21	18	21	20	20
No. test results in all laboratories, n_t	42	36	42	40	40
Mean specific extinction, $\overline{\Delta K}$	-0,002	0,000	0,082	0,031	0,042
Repeatability standard deviation, s_r	0,001	0,001	0,002	0,001	0,001
Coefficient of variation of repeatability, $C_{V,r}$, %	36,4	121,1	2,3	4,4	1,7
Repeatability limit, r	0,003	0,001	0,005	0,004	0,002
Reproducibility standard deviation, s_R	0,004	0,001	0,008	0,004	0,005
Coefficient of variation of reproducibility, $C_{V,R}$, %	148,2	234,8	10,0	12,6	10,6
Reproducibility limit, R	0,011	0,003	0,023	0,011	0,013

Bibliography

- [1] ISO 1042, *Laboratory glassware — One-mark volumetric flasks*
- [2] ISO 5555, *Animal and vegetable fats and oils — Sampling*
- [3] ISO 5725-1, *Accuracy (trueness and precision) of measurement methods and results — Part 1: General principles and definitions*
- [4] ISO 5725-2, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*
- [5] International Standard 80000 (all parts), *Quantities and units*
- [6] COMMISSION REGULATION (EEC) No 2568/91 of 11 July 1991. On the characteristics of olive oil and olive-residue oil and on the relevant methods of analysis. *Off. J.* 1991-09-05, **248**, p. 1. Available (viewed 2010-09-20) at:
<http://eur-lex.europa.eu/LexUriServ/LexUriServ.do?uri=CONSLEG:1991R2568:20080101:EN:PDF>
- [7] BURKE, R.W., MAVRODINEANU, R. *Certification and use of acidic potassium dichromate solutions as an ultraviolet absorbance standard SRM 935*. Washington, DC: National Bureau of Standards, 1977. 147 p. (*NBS Special Publication No. 260-54.*)

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email bsmusales@bsigroup.com.

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

Useful Contacts:

Customer Services

Tel: +44 845 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 845 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com



...making excellence a habit.™