

BRITISH STANDARD

**BS EN ISO
3830 : 1996
BS 2000 :
Part 270 : 1996
ISO 3830 : 1993**

Methods of test for

Petroleum and its products

**Part 270. Petroleum products - Determination
of lead content of gasoline - Iodine
monochloride method**

(Identical with IP 270/96)

The European Standard EN ISO 3830 : 1995 has the status of a
British Standard

BSi
British Standards

National foreword

This British Standard was published under the authority of the Materials and Chemicals Sector Board and comes into effect on 29 February 1996. It is the English language version of EN ISO 3830 : 1995 Petroleum products - Determination of lead content of gasoline - Iodine monochloride method, published by the European Committee for Standardization (CEN), which endorses ISO 3830 : 1993, published by the International Organization for Standardization (ISO).

This British Standard supersedes BS 2000 : Part 270 : 1994 and BS 5657 : 1988, which are withdrawn.

There are no technical differences between this British Standard and the previous edition of BS 2000 : Part 270. A new edition has been made available to fulfil BSI's obligation to publish all approved European Standards.

BS 2000 comprises a series of test methods for petroleum and its products that are published by the Institute of Petroleum (IP) and have been accorded the status of a British Standard. Each method should be read in conjunction with the preliminary pages of 'IP Standard methods for analysis and testing of petroleum and related products' which gives details of the BSI/IP agreement for publication of the series, provides general information on safety precautions, sampling and other matters, and lists the methods published as Parts of BS 2000.

The numbering of the Parts of BS 2000 follows that of the corresponding methods published in 'IP Standard methods for analysis and testing of petroleum and related products'. Under the terms of the agreement between BSI and the Institute of Petroleum, BS 2000 : Part 270/BS EN ISO 3830 will be published by the IP (in 'Standard methods for analysis and testing of petroleum and related products' and as a separate publication). BS 2000 : Part 270 : 1996 is thus identical with IP 270/96.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

© The Institute of Petroleum
& BSI 1996

The following BSI references
relate to the work of this
standard:

Committee reference PTI/13
Announcement in *BSI News*
March 1995



ISO 3830 : 1993



BS EN ISO 3830 : 1996
BS 2000 : Part 270 : 1996

EUROPEAN STANDARD

EN ISO 3830

NORME EUROPÉENNE

EUROPÄISCHE NORM

August 1995

ICS 75.160.20

Supersedes EN 23830: 1989

Descriptors: petroleum products, gasoline, chemical analysis, determination of content, lead, volumetric analysis

English version

Petroleum products – Determination of lead content of gasoline – Iodine monochloride method (ISO 3830: 1993)

Produits pétroliers – Détermination de la teneur en plomb de l'essence – Méthode au monochlorure d'iode (ISO 3830: 1993)

Mineralölerzeugnisse – Bestimmung des Bleigehaltes von Ottokraftstoffen – Iodmonochlorid-Verfahren (ISO 3830: 1993)

This European Standard was approved by CEN on 1995-07-30. CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

The European Standards exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

© 1995

All rights of reproduction and communication in any form and by any means reserved in all countries to CEN and its members.

Ref. no. EN ISO 3830: 1995 E

Foreword

The text of the International Standard from ISO/TC 28 "Petroleum products and lubricants" of the International Organization for Standardization (ISO) has been taken over as a European Standard by the Technical Committee CEN/TC 19 "Petroleum products, lubricants and related products".

This European Standard supersedes EN 23830: 1989

This European Standard shall be given the status of a National Standard, either by publication of an identical text or by endorsement, at the latest by February 1996, and conflicting national standards shall be withdrawn at the latest by February 1996.

According to CEN/CENELEC Internal Regulations, the following countries are bound to implement this European Standard: Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of the International Standard ISO 3830: 1993 has been approved by CEN as a European Standard without any modification.

Petroleum products — Determination of lead content of gasoline — Iodine monochloride method

WARNING — The use of this International Standard may involve hazardous materials, operations and equipment. This standard does not purport to address all of the safety problems associated with its use. It is the responsibility of the user of this standard to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

1 Scope

This International Standard specifies a method for the determination of total lead content in gasolines containing lead alkyls at concentrations between 0,026 g and 1,300 g of lead per litre.

This International Standard is not applicable to gasoline containing manganese anti-knock additives.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this International Standard. At the time of publication, the editions indicated were valid. All standards are subject to revision, and parties to agreements based on this International Standard are encouraged to investigate the possibility of applying the most recent editions of the standards indicated below. Members of IEC and ISO maintain registers of currently valid International Standards.

ISO 385-1:1984, *Laboratory glassware — Burettes — Part 1: General requirements.*

ISO 835-1:1981, *Laboratory glassware — Graduated pipettes — Part 1: General requirements.*

ISO 1042:1983, *Laboratory glassware — One-mark volumetric flasks.*

ISO 1770:1981, *Solid-stem general purpose thermometers.*

ISO 3007:1986, *Petroleum products — Determination of vapour pressure — Reid method.*

ISO 3170:1988, *Petroleum liquids — Manual sampling.*

ISO 3171:1988, *Petroleum liquids — Automatic pipeline sampling.*

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods.*

ISO 3839:1978, *Petroleum distillates and commercial aliphatic olefins — Determination of bromine number — Electrometric method.*

ISO 4788:1980, *Laboratory glassware — Graduated measuring cylinders.*

ISO 4800:1977, *Laboratory glassware — Separating funnels and dropping funnels.*

3 Principle

A known volume of the test sample is diluted with heavy distillate and shaken with aqueous iodine monochloride reagent. Any tetraalkyl lead compounds present react with the iodine monochloride and are extracted into the aqueous phase as the dialkyl lead compounds. The aqueous extract is separated from the gasoline and evaporated to low bulk to decompose free iodine monochloride. Any organic matter present is removed by oxidation with nitric acid, which also serves to convert the dialkyl lead compounds into inorganic lead compounds. The residue is dissolved in water and buffered to pH 5 with sodium acetate/acetic acid buffer. The lead content of the buffered solution is determined by titration with Na₂EDTA using xylenol orange as indicator.

4 Reagents

During the analysis described in this International Standard, use only reagents of recognized analytical reagent grade and only water complying with the requirements of Grade 3 of ISO 3696.

4.1 Nitric acid, concentrated [69 % (*m/m*) to 70,5 % (*m/m*)].

4.2 Hydrochloric acid, concentrated [35,4 % (*m/m*)].

4.2.1 Hydrochloric acid solution (1 + 1).

Mix one volume of concentrated hydrochloric acid (4.2) with one volume of water.

4.3 Ammonia solution (1 + 1).

Mix one volume of concentrated ammonia solution [35 % (*m/m*)] with one volume of water.

4.4 Heavy distillate.

A straight-run petroleum distillate having a maximum bromine number of 1,5 with approximately 10 % distilling at 205 °C and 90 % at 240 °C. It shall also be lead-free, having been, if necessary, previously extracted with the iodine monochloride solution (4.6).

4.5 Sodium acetate/acetic acid buffer solution (pH 5).

Dissolve 23,0 g of anhydrous sodium acetate (CH₃COONa) in approximately 500 ml of water. Using a burette or graduated pipette, add 7,2 ml of glacial acetic acid. Dilute to the mark with water in a 1 000 ml one-mark volumetric flask and shake to mix thoroughly.

4.6 Iodine monochloride reagent, 1,0 mol/l solution.

Dissolve 111,0 g of potassium iodide (KI) in approximately 400 ml of water. Add 445 ml of concentrated hydrochloric acid (4.2) and allow to cool to ambient temperature. Add 75,0 g of potassium iodate (KIO₃) slowly and with stirring, until all free iodine initially formed has just redissolved to give a clear, orange-red solution. Allow to cool to ambient temperature and dilute to 1 000 ml with water. Store in a glass-stoppered bottle.

WARNING — Iodine monochloride will react with ammonium ions under certain conditions to yield explosive nitrogen tri-iodide. Care shall be taken, therefore, that this reagent does not come into contact with ammonia or ammonium salts.

Never use rubber bungs to stopper vessels containing iodine monochloride solutions.

NOTE 1 The amounts of KI and KIO₃ are calculated to give a slight excess of iodate; if a greater excess is present, this will lead to precipitation of lead and indifferent end-points in the Na₂EDTA titration.

4.7 Lead nitrate, 0,005 mol/l standard solution.

Weigh, to the nearest 0,001 g, approximately 1,7 g of lead nitrate [Pb(NO₃)₂] that has been dried at 105 °C and allowed to cool in a desiccator. Dissolve it in water and add 10 ml of the concentrated nitric acid (4.1). Dilute to the mark with water in a 1 000 ml one-mark volumetric flask and shake thoroughly to mix.

Calculate the concentration, *c*₀, of the lead nitrate solution, in moles per litre, according to the equation:

$$c_0 = \frac{m}{331,23}$$

where *m* is the mass, in grams, of lead nitrate dissolved.

4.8 Disodium dihydrogen ethylenediaminetetraacetate, [−CH₂N(CH₂COOH)CH₂COONa]₂·2H₂O (Na₂EDTA), 0,005 mol/l standard volumetric solution.

4.8.1 Preparation

Dissolve approximately 3,75 g of the Na₂EDTA in 2 000 ml of water.

4.8.2 Standardization

Using a pipette, transfer 25,0 ml of the standard lead nitrate solution (4.7) to a 250 ml conical flask. Dilute to approximately 75 ml with water and add several drops of the bromothymol blue indicator solution (4.10). Titrate with the ammonia solution (4.3) until the colour of the solution just changes from yellow to blue, then add 10 ml of the sodium acetate/acetic acid buffer solution (4.5) and five drops of the xylenol orange indicator solution (4.9).

NOTE 2 In the presence of lead the solution will have a plum-red colour.

Titrate with the Na₂EDTA solution (4.8) until the end-point is reached. This is indicated by a sharp change from orange to a permanent bright lemon-yellow. Record the volume used and calculate the concentration of the Na₂EDTA solution.

NOTE 3 The addition of excess Na₂EDTA solution produces no further colour change at the end-point.

4.8.3 Calculation

Calculate the concentration, c_1 , of the Na_2EDTA solution to the nearest 0,000 01 mol/l according to the equation

$$c_1 = \frac{25c_0}{V}$$

where

c_0 is the concentration, in moles per litre, of the standard lead nitrate solution (4.7);

V is the volume, in millilitres, of the Na_2EDTA solution used in the standardization.

4.9 Xylenol orange indicator solution.

Dissolve 0,2 g of xylenol orange, sodium salt, in 100 ml of water and add 1 drop of 1 + 1 hydrochloric acid (4.2.1).

Prepare a fresh solution each week.

4.10 Bromothymol blue indicator solution.

Dissolve 0,1 g of bromothymol blue in 80 ml of 95 % (V/V) ethanol and dilute to 100 ml with water.

5 Apparatus

5.1 Burette, 25 ml capacity, Class A in ISO 385-1.

5.2 Pipette, 25 ml capacity, Class A in ISO 835-1.

5.3 Volumetric flasks, one-mark, 1 000 ml and 2 000 ml capacities, Class A in ISO 1042.

5.4 Conical flasks, wide neck, borosilicate glass, 250 ml and 500 ml capacities.

5.5 Measuring cylinders, 100 ml and 500 ml capacities, in accordance with ISO 4788.

5.6 Separating funnel, 250 ml capacity, borosilicate glass, glass-stoppered, to ISO 4800.

5.7 Watch glass, borosilicate glass, of a size sufficient to cover the mouth of the 500 ml conical flask (5.4).

NOTE 4 Ribbed watch glasses are recommended but may not be readily available in all countries. They have been found to significantly reduce the time required for evaporation of aqueous phases containing the extracted lead.

5.8 Thermometer, general purpose, nominal range 0 °C to 50 °C or 0 °C to 100 °C, in accordance with ISO 1770 Type L or Type M.

5.9 Balance, single or double pan, capable of weighing to $\pm 0,001$ g.

5.10 Desiccator, glass, of sufficient capacity, with desiccant.

6 Sampling

Test samples shall be taken in accordance with ISO 3170, ISO 3171 or an equivalent national standard. Gasoline samples having a Reid vapour pressure (ISO 3007) above 50 kPa shall be cooled in the sealed sample container to below 15 °C before removing the test sample for analysis. The sampling procedure shall be recorded in the test report.

7 Test procedure

7.1 Transfer 50 ml of the iodine monochloride reagent (4.6) and 25 ml of the heavy distillate (4.4) to the separating funnel (5.6). Measure the temperature of the test sample to the nearest 0,5 °C (see clause 6). Using the pipette (5.2), transfer 25 ml of the test sample of gasoline to the same separating funnel. Immediately stopper the funnel and shake the contents for 1 min. Allow the funnel to stand for several minutes, until the two phases have separated, and run the lower aqueous phase into the 500 ml conical flask (5.4). Wash the gasoline phase by shaking with three separate 20 ml portions of water and add the washings to the conical flask.

WARNING — No liquids should be sucked by mouth into a pipette. Release any pressure in the separating funnel during shaking.

7.2 Add a few glass beads, cover the conical flask with the watch glass (5.7) and bring the aqueous solution to a low-boiling condition on a hot-plate in a fume cupboard. When the volume of solution has been reduced to between 15 ml and 20 ml, slowly add, without removing the flask from the hot-plate, 5 ml of the nitric acid (4.1) down the side of the flask and evaporate the contents almost to dryness to oxidise any organic material present. Repeat the nitric acid treatment until a white residue remains. Finally, remove the watch glass and evaporate the solution to dryness. Remove the flask from the hot-plate and allow the contents to cool.

7.3 Add approximately 200 ml of water to the flask and swirl to dissolve the residue.

NOTE 5 The residue may be quickly dissolved by heating the solution.

If the solution has been heated, cool before proceeding. Add several drops of the bromothymol blue indicator solution (4.10) and titrate with the ammonia solution (4.3) until the colour just changes from straw yellow to blue, then add 10 ml of the sodium acetate/acetic acid buffer solution (4.5) and 5 drops of the xylenol orange indicator solution (4.9).

NOTE 6 After addition of xylenol orange, in the presence of lead, the solution will have a plum-red colour. At high lead concentrations, a 1 % (m/m) xylenol orange solution may give a better colour indication.

7.4 Titrate the resulting solution with the standard volumetric Na₂EDTA solution (4.8) until the end-point is reached. This is indicated by a sharp colour change from orange to a permanent bright lemon-yellow. Record the volume of Na₂EDTA solution used.

NOTE 7 The addition of excess standard volumetric Na₂EDTA solution produces no further colour change after the end-point.

7.5 Carry out a blank determination using the procedure given in 7.1 to 7.4, omitting the test portion, and record the volume, V_0 , of Na₂EDTA in ml.

8 Calculation

Calculate the concentration of lead, c_{lead} , in grams per litre at 15 °C, by means of the following equation (see note 8):

$$c_{\text{lead}} = 8,288(V_1 - V_0)c_1[1 + 0,001\ 2(t - 15)]$$

where

| | |
|---------|--|
| V_0 | is the volume, in millilitres, of standard volumetric Na ₂ EDTA solution used for the blank determination (7.5); |
| V_1 | is the volume, in millilitres, of standard volumetric Na ₂ EDTA solution used to titrate the test portion; |
| c_1 | is the concentration, in moles per litre, of the standard volumetric Na ₂ EDTA solution; |
| t | is the temperature, in degrees Celsius, of the gasoline when pipetting the sample; |
| 8,288 | is a constant relating the equivalency of Na ₂ EDTA to grams per litre of lead; |
| 0,001 2 | is a compromise coefficient of expansion of gasolines per 1 °C at 15 °C, and is intermediate between that of motor and aviation gasolines. |

NOTE 8 For gasoline containing only tetraethyl lead (TEL) or tetramethyl lead (TML), the grams of lead per unit volume can be converted to millilitres per unit volume by multiplying by the following factors:

for TEL = 0,946;

for TML = 0,648.

9 Expression of results

Report the concentration, c_{lead} , obtained to the nearest 0,002 g lead per litre at 15 °C.

10 Precision

The precision of this International Standard was obtained in an ISO cooperative test programme on samples covering a range of 0,3 to 1,0 g of lead per litre. In a subsequent ASTM testing programme, the range was extended down to 0,03 g of lead per litre with equal or better precision. The precision was established using dimethyl yellow indicator solution in place of bromothymol blue. Limited testing has shown that the change of indicator solution does not affect the precision.

10.1 Repeatability

The difference between two test results, obtained by the same operator with the same apparatus under constant operating conditions on identical test material, would in the long run, in the normal and correct operation of the test method, exceed the following value only in one case in twenty:

$$0,003\ 65 + 0,007\ 3\bar{c}_{\text{lead}}$$

where \bar{c}_{lead} is the average of the results being compared, in grams of lead per litre at 15 °C.

10.2 Reproducibility

The difference between two single and independent results obtained by different operators working in different laboratories on identical test material would, in the long run, in the normal and correct operation of the test method, exceed the following value only in one case in twenty:

$$0,013\ 5 + 0,027\bar{c}_{\text{lead}}$$

where \bar{c}_{lead} is the average of the results being compared, in grams of lead per litre at 15 °C.

11 Test report

The test report shall contain at least the following information:

- sufficient details for complete identification of the product tested;

LEAD IN GASOLINE ICL, IP 270

- b) a reference to this International Standard;
- c) the sampling procedure used (see clause 6);
- d) the result of the test (see clause 9);
- e) any deviation, by agreement or otherwise, from the procedure specified;
- f) the date of the test.

The Institute of Petroleum

61 New Cavendish Street
London
W1M 8AR

Tel: 0171 467 7100
Fax: 0171 255 1472

Buying Parts of BS 2000

Orders for BS 2000 publications should be addressed to the Library at the Institute of Petroleum.

Copyright

Copyright subsists in all BS 2000 publications. No part of this publication may be reproduced in any form without the prior permission in writing of BSI and the IP. Enquiries about copyright should be made to the Secretary of PTI/13 at the IP.



THE INSTITUTE
OF PETROLEUM