Water quality — Determination of free chlorine and total chlorine —

Part 2: Colorimetric method using *N*,*N*-diethyl-1,4-phenylenediamine, for routine control purposes

The European Standard EN ISO 7393-2:2000 has the status of a British Standard

ICS 13.060.50

BS EN ISO 7393-2:2000 BS 6068-2.26: 1986

Incorporating Amendment No. 1 to BS 6068-2.26:1986 (renumbers the BS as BS EN ISO 7393-2:2000)

> Confirmed July 2008



National foreword

This British Standard is the English language version of EN ISO 7393-2:2000. It is identical to ISO 7393-2:1985.

The UK participation in its preparation was entrusted by Technical Committee EH/3, Water quality, to Subcommittee EH/3/2, Physical, Chemical and Biochemical which has the responsibility to:

- aid enquirers to understand the text;
- present to the responsible international committee any enquiries on interpretation, or proposals for change, and keep UK interests informed;
- monitor related international and European developments and promulgate them in the UK

A list of organizations represented on this subcommittee can be obtained on request to its secretary.

BS EN ISO 7393-2 is one of a series of standards on water quality, others of which have been, or will be, published as sections of BS 6068. This standard has therefore been given the secondary identifier BS 6068-2.26. The various sections of BS 6068 are comprised within parts 1 to 7, which, together with part 0, are listed below.

- Part 0: Introduction;
- Part 1: Glossary;
- Part 2: Physical, chemicals and biochemical methods;
- Part 3: Radiological methods;
- Part 4: Microbiological methods;
- Part 5: Biological methods;
- Part 6: Sampling;
- Part 7: Precision and accuracy.

The British Standards which implement international and European publications referred to in this document may be found in the BSI Standards Catalogue under the section entitled "International Standards Correspondence Index" or by using the "Find" facility of the BSI Standards Electronic Catalogue.

The method referred to in 4.2, for disposal of solutions containing mercury given in ISO 5790 is technically equivalent to that given in the National appendix of BS 6070-2, Methods of sampling and test for sodium carbonate for industrial use — Part 2: Determination of chloride content.

A British Standard does not purport to include all the necessary provisions of a contract. Users of British Standards are responsible for their correct application.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

Summary of pages

This document comprises a front cover, an inside front cover, pages i and ii, the EN ISO title page, the EN ISO foreword page, pages 1 to 8, an inside back cover and a back cover.

The BSI copyright notice displayed in this document indicates when the document was last issued.

This British Standard, having been prepared under the direction of the Environment and Pollution Standards Committee, was published under the authority of the Board of BSI and comes into effect on 28 November 1986

 $\ensuremath{\mathbb{C}}$ BSI 05-2000

The following BSI references relate to the work on this standard:

Committee reference EPC/44 Draft for comment 84/52410 DC

ISBN 0 580 15453 X

Amendments issued since publication

Amd. No.	Date of issue	Comments
10870	May 2000	Renumbers BS 6068-2.26:1986 as BS EN ISO 7393-2:2000.

Contents

	I	Page
Nati	onal foreword Inside front c	over
0	Introduction	1
1	Scope and field of application	1
2	Definitions	1
3	Principle	1
4	Reagents	2
5	Apparatus	3
6	Procedure	3
7	Correction of interference due to the presence of oxidized manganese	4
8	Expression of results	4
9	Interferences	5
10	Test report	5
the r	ex A Separate determinations of combined chlorine of monochloramine type, combined chlorine of the dichloramine	C
	and of combined chlorine in the form of nitrogen trichloride	6
	ex B Preparation of water free from oxidizing and reducing substances	7
	iography	10
Fabl solut	e 1 — Terms and synonyms in relation to actual compounds in the tion	1
	e 2 — Analytical parameters from interlaboratory analysis for free lual chlorine	5

ii blank

EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

EN ISO 7393-2

January 2000

ICS 13.060

English version

Water quality — Determination of free chlorine and total chlorine — Part 2: Colorimetric method using N,N-diethyl-1,4-phenylenediamine for routine control purposes

(ISO 7393-2:1985)

Qualité de l'eau — Dosage du chlore libre et du chlore total — Partie 2: Méthode colorimétrique à la N,N-diéthylphénylène-1,4 diamine destinée aux contrôles de routine (ISO 7393-2:1985) Wasserbeschaffenheit — Bestimmung von freiem Chlor und Gesamtchlor —
Teil 2: Kolorimetrisches Verfahren mit N,N-Diethyl-1,4-Phenylendiam für Routinekontrollen (ISO 7393-2:1985)

This European Standard was approved by CEN on 20 January 2000.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CEN

European Committee for Standardization Comité Européen de Normalisation Europäisches Komitee für Normung

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

Foreword to EN ISO 7393-2:2000

The text of the International Standard from Technical Committee ISO/TC 147, Water quality, of the International Organization for Standardization (ISO) has been taken over as a European Standard by Technical Committee CEN/TC 230, Water analysis, the Secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2000, and conflicting national standards shall be withdrawn at the latest by July 2000.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

0 Introduction

ISO 7393 consists of the following parts.

Part 1: Titrimetric method using

N,*N*-diethyl-1,4-phenylenediamine.

Part 2: Colorimetric method using

N,N-diethyl-1,4-phenylenediamine, for routine control purposes.

Part 3: Iodometric titration method for the determination of total chlorine. 1)

1 Scope and field of application

This part of ISO 7393 specifies a method for the determination of free chlorine and total chlorine in water, readily applicable to field testing; it is based on measurement of the colour intensity by visual comparison of the colour with a scale of standards which is regularly calibrated.

Sea water and waters containing bromides and iodides comprise a group for which special procedures are required.^[2]

This method is applicable to concentrations, in terms of chlorine (Cl_2), from 0,000 4 to 0,07 mmol/l (0,03 to 5 mg/l) total chlorine. For higher concentrations the test portion must be diluted. If the speed of operation and the compactness of the equipment are not overriding requirements, spectrometric measurement is described as an alternative procedure.

In Annex A a procedure is presented for the differentiation of combined chlorine of the monochloramine type, combined chlorine of the dichloramine type and combined chlorine in the form of nitrogen trichloride.

Interferences are noted in clauses 7 and 9.

2 Definitions (see Table 1)

For the purpose of this part of ISO 7393, the following definitions apply.

2.1

free chlorine

chlorine present in the form of hypochlorous acid, hypochlorite ion or dissolved elemental chlorine

2.2

combined chlorine

the fraction of total chlorine present in the form of chloramines and organic chloramines

2.3

total chlorine

chlorine present in the form of "free chlorine" or "combined chlorine" or both

2.4

chloramines

derivatives of ammonia by substitution of one, two or three hydrogen atoms with chlorine atoms (monochloramine NH_2Cl , dichloramine $NHCl_2$, nitrogen trichloride NCl_3) and all chlorinated derivatives of organic nitrogen compounds as determined by the method specified in this part of ISO 7393

3 Principle

3.1 Determination of free chlorine

Direct reaction with the *N*,*N*-diethyl-1,4-phenylenediamine (DPD) and formation of a red compound at pH 6,2 to 6,5. Measurement of the colour intensity by visual comparison of the colour with a scale of permanent glass standards, or by spectrometry.

3.2 Determination of total chlorine

Reaction with DPD in the presence of an excess of potassium iodide then measurement as in 3.1.

Table 1 — Terms and synonyms in relation to actual compounds in the solution

Term	Synonym		Compounds		
Free chlorine	Free chlorine	Active free chlorine	Elemental chlorine, hypochlorous acid		
		Potential free chlorine	Hypochlorite		
Total chlorine	Total residual chlorine		Elemental chlorine, hypochlorous acid, hypochlorite, and chloramines		

¹⁾ At present at the stage of draft.

4 Reagents

During the analysis, use only reagents of recognized analytical grade, and only water as specified in 4.1.

4.1 *Water*, free from oxidizing and reducing substances.

Demineralized or distilled water of which the quality is checked as follows.

Into two 250 ml chlorine-demand-free conical flasks (clause **5**) place, in order,

a) in the first: 100 ml of the water to be checked and about 1 g of potassium iodide (4.4); mix and after 1 min add 5 ml of buffer solution (4.2) and 5,0 ml of DPD reagent (4.3);

b) in the second: 100 ml of the water to be checked and two drops of sodium hypochlorite solution (4.7); then, after 2 min, 5,0 ml of buffer solution (4.2) and 5 ml of DPD reagent (4.3).

No coloration should appear in the first flask whereas it is essential that a light pink coloration appears in the second flask.

In the case of demineralized or distilled water not having the desired quality it must be chlorinated. After a period of contact followed by dechlorination the quality finally has to be rechecked.

A procedure for the preparation of water free from oxidizing and reducing substances is given in Annex B.

4.2 Buffer solution, pH 6,5.

Dissolve in water (4.1) in this order: 24 g of anhydrous disodium hydrogen phosphate (Na₂HPO₄) or 60,5 g of the dodecahydrate form (Na₂HPO₄·12H₂O) and 46 g of potassium dihydrogen phosphate (KH₂PO₄). Add 100 ml of 8 g/l disodium

dihydrogenethylenedinitrilotetraacetate dihydrate (disodium EDTA dihydrate, $\rm C_{10}H_{14}N_2O_8Na_2\cdot 2H_2O)$ solution (or 0,8 g of the solid form).

If necessary, add 0.020~g of mercury(II) chloride (HgCl₂), to prevent mould growth and interference in the free chlorine test caused by any trace amounts of iodide in the reagents.

Dilute to 1 000 ml and mix.

NOTE Solutions containing mercury should be disposed of safely (for example a method is specified in ISO 5790, Inorganic chemical products for industrial use — General method for determination of chloride content — Mercurimetric method).

4.3 N,N-diethyl-1,4-phenylenediamine sulfate (DPD) [NH₂-C₆H₄-N(C₂H₅)₂·H₂SO₄], solution, 1,1 g/l.

Mix 250 ml water (4.1), 2 ml sulfuric acid (Q = 1.84 g/ml) and 25 ml of 8 g/l disodium EDTA dihydrate solution (or 0.2 g of the solid form). Dissolve in this mixture 1.1 g of anhydrous DPD or 1.5 g of the pentahydrate form, dilute to 1 000 ml and mix.

Store the reagent in a dark bottle protected from heat.

Renew the solution after 1 month or when it becomes discoloured.

4.4 *Potassium iodide*, crystals.

NOTE Reagents 4.2, 4.3 and 4.4 may be conveniently replaced by combined reagents commercially available in the form of stable powder or tablets.

4.5 Sulfuric acid, $c(H_2SO_4) \approx 1 \text{ mol/l}$.

Take 800 ml water (4.1) and add cautiously with continuous stirring 54 ml of sulfuric acid ($\varrho = 1,84$ g/ml). Cool to room temperature and transfer the solution to a 1 000 ml one-mark volumetric flask. Make up to the mark with water and mix well.

4.6 Sodium hydroxide, $c(NaOH) \approx 2 \text{ mol/l}$.

Weigh 80 g of sodium hydroxide pellets and add to 800 ml water (4.1) in a conical flask. Stir continuously until all pellets are dissolved. Wait until the solution has cooled to room temperature and transfer this solution to a 1 000 ml one-mark volumetric flask. Make up to the mark with water and mix well.

4.7 Sodium hypochlorite, solution, $\varrho(\text{Cl}_2)$ about 0,1 g/l.

Prepare by dilution of concentrated commercial sodium hypochlorite solution.

4.8 *Potassium iodate*, stock solution, $\rho(\text{KIO}_3) = 1,006 \text{ g/l}.$

Dissolve 1,006 g of potassium iodate (KIO_3) in about 250 ml water (4.1) in a 1 000 ml one-mark volumetric flask. Make up to the mark with water and mix.

4.9 Potassium iodate, standard solution, $\varrho(\text{KIO}_3) = 10,06 \text{ mg/l}.$

Take 10 ml of stock solution (4.8), place in a 1 000 ml one-mark volumetric flask, add about 1 g potassium iodide (4.4) and make up to the mark with water (4.1).

Prepare this solution on the day of use.

1 ml of this standard solution contains 10,06 μg of KIO $_{3}.$

10,06 μ g of KIO₃ is equivalent to 0,141 μ mol Cl₂.

 \odot BSI 05-2000

4.10 Sodium arsenite (NaAsO₂), solution, 2 g/l; or thioacetamide (CH₃CSNH₂), solution, 2,5 g/l.

5 Apparatus

Ordinary laboratory apparatus, and *Colorimetric equipment*, comprising one of the following:

- **5.1** *Comparator*, equipped with a scale of permanent glass colour standards specially set up for the DPD technique and suitable for concentrations from 0,000 4 to 0,07 mmol/l (0,03 to 5 mg/l) of chlorine.
- **5.2** *Spectrometer*, with a selector for continuous wavelength variation, suitable for use at 510 nm and equipped with rectangular cells with an optical path length of 10 mm or greater.
- **5.3** Spectrometer, with a selector for discontinuous wavelength variation, having its maximum transmission as near as possible to 510 nm and rectangular cells with an optical path length of 10 mm or greater.

NOTE ON THE PREPARATION OF GLASSWARE

Chlorine-demand-free glassware is obtained by filling with sodium hypochlorite solution (4.7) then, after 1 h, rinsing copiously with water (4.1). During the course of the analysis one set of glassware should be kept for the determination of free chlorine and another for the determination of total chlorine in order to avoid contamination of the free chlorine set.

6 Procedure

6.1 Test sample

Start determination immediately after taking samples. At all times avoid bright light, agitation and heat.

6.2 Test portions

Take two test portions, each of 100,0 ml (V_0) . If the concentration of total chlorine exceeds 70 μ mol/l (5 mg/l) it is necessary to take a smaller volume, V_1 , of test sample and to dilute with water (4.1) to 100,0 ml.

6.3 Calibration

Place in a series of 100 ml one-mark volumetric flasks, increasing quantities of the potassium iodate standard solution (4.9) in such a way as to set up a scale extending from $c(Cl_2) = 0.423$ up to 70,5 μ mol/l [ϱ (Cl₂) = 0,03 up to 5 mg/l; 0,3 up to 50 ml of standard solution (4.9)]. Add 1,0 ml sulfuric acid (4.5) and after 1 min, 1,0 ml of sodium hydroxide solution (4.6). Dilute to 100 ml with water (4.1). Transfer the contents of each flask, without rinsing, into a 250 ml conical flask containing 5 ml buffer solution (4.2) and 5 ml DPD reagent (4.3), added less than 1 min prior to the transfer, and mix (see the note). Then fill the measuring cell successively with each of the prepared standard matching solutions and measure within 2 min one of the following:

- the colour intensity with the comparator (5.1),
- the absorbance, against water in the reference cell, with a spectrometer (5.2 or 5.3).

As required, check and make any necessary corrections to the comparator scale of standards or prepare a calibration graph for the spectrometer. Carry out a calibration for each fresh preparation of DPD reagent and check daily one point on the scale or on the graph.

NOTE Prepare each standard matching solution separately to avoid the mixture of buffer and reagent standing too long in advance and the appearance of a false red colour.

6.4 Determination of free chlorine

Transfer the first test portion, without rinsing, to a 250 ml conical flask containing 5 ml of buffer solution (4.2) and 5 ml of DPD reagent (4.3) and mix (see the note to 6.3). Fill the measuring cell with this treated solution and immediately measure the colour under the same conditions as adopted for the calibration (6.3). Record c_1 , the concentration reading from the comparator scale or calibration graph (6.3).

In the case of an unknown water, possibly being very acid, or very alkaline or with a high concentration of salts, it is advisable to verify that the volume of buffer solution (4.2) added is sufficient to bring the water to pH 6,2 to 6,5. If not, use a greater volume of the buffer solution (4.2).

6.5 Determination of total chlorine

Transfer the second test portion, without rinsing, to a 250 ml conical flask containing 5 ml of buffer solution (4.2) and 5 ml of DPD reagent (4.3), add about 1 g of potassium iodide (4.4) and mix (see the note to 6.3). Fill the measuring cell with this treated solution and after 2 min measure the colour under the same conditions as adopted for the calibration (6.3). Record c_2 , the concentration reading from the comparator scale or calibration graph (6.3).

In the case of an unknown water, possibly being very acid, or very alkaline or with a high concentration of salts, it is advisable to verify that the volume of buffer solution (4.2) added is sufficient to bring the water to pH 6,2 to 6,5. If not, use a greater volume of the buffer solution (4.2).

7 Correction of interference due to the presence of oxidized manganese

Determine the effect of oxidized manganese by carrying out a supplementary determination on a further test portion (6.2) previously treated with the arsenite or thioacetamide solution (4.10) in order to neutralize all oxidizing compounds other than oxidized manganese.

Place this test portion in a 250 ml conical flask, add 1 ml of sodium arsenite solution (4.10) or thioacetamide solution (4.10) and mix. Again add 5,0 ml of buffer solution (4.2) and 5,0 ml of DPD reagent (4.3) and mix.

Fill the measuring cell with this treated solution and immediately measure the colour under the same conditions as adopted for the calibration (6.3). Record c_3 , the concentration reading from the comparator scale or calibration graph (6.3), corresponding to the oxidized manganese present.

In using comparators with permanent glass colour standards the arsenite or thioacetamide treated sample may be used as a blank to compensate for any interference colour so long as the time of addition of reagents is the same for both blank and sample.

8 Expression of results

8.1 Method of calculation

8.1.1 Calculation of the free chlorine concentration

The concentration of free chlorine, $c(C1_2)$, expressed in millimoles per litre, is given by the equation

$$c(Cl_2) = \frac{(c_1 - c_3) V_0}{V_1}$$

where

 c_1 is the concentration, expressed in millimoles of Cl_2 per litre, of chlorine as determined in **6.4**;

 c_3 is the concentration, expressed in millimoles of Cl_2 per litre, corresponding to the oxidized manganese present (see clause 7);

NOTE If oxidized manganese is absent, c_3 = 0. V_0 is the maximum volume, in millilitres, of the test portion (6.2) (V_0 = 100,0 ml);

 V_1 is the volume, in millilitres, of test sample in the test portion (6.2).

8.1.2 Calculation of the total chlorine concentration

The concentration of total chlorine, $c(C1_2)$, expressed in millimoles per litre, is given by the equation

$$c(Cl_2) = \frac{(c_2 - c_3) V_0}{V_1}$$

where

 c_2 is the concentration, expressed in millimoles of Cl_2 per litre, of chlorine as determined in **6.5**; c_3 , V_0 and V_1 are as defined in **8.1.1**.

8.2 Conversion of amount of substance concentration to mass concentration

The chlorine concentration expressed in moles per litre may be expressed in grams per litre by multiplying by a conversion factor of 70,91.

8.3 Repeatability and reproducibility

To obtain an indication of repeatability and reproductibility, figures are taken from measurements obtained by methods which are the same in principle as that specified in this part of ISO 7393.

The USA-EPA Environmental Monitoring and Support Laboratory^[1] evaluated the colorimetric method with the following results.

For distilled water samples at concentrations of $c(\text{Cl}_2)=5.5$ and $50.9~\mu\text{mol/l}$ [$\varrho(\text{Cl}_2)=0.39$, and 3.61~mg/l] total chlorine, the relative standard deviations were 3.1~%, and 3.2~%, respectively. Using drinking water containing $c(\text{Cl}_2)=13.3~\mu\text{mol/l}$ [$\varrho(\text{Cl}_2)=0.94~\text{mg/l}$] total chlorine, the relative standard deviation was 0.8~%. With river water at a concentration of $c(\text{Cl}_2)=12.1~\mu\text{mol/l}$ [$\varrho(\text{Cl}_2=0.86~\text{mg/l}]$ total chlorine and domestic sewage at a concentration of $c(\text{Cl}_2)=15.1~\mu\text{mol/l}$ [$\varrho(\text{Cl}_2)=1.07~\text{mg/l}$] total chlorine, the relative standard deviations were 1.9~% and 2.4~%, respectively.

Results published by the British Department of the Environment ^[2] based on data obtained at the Water Research Centre ^[3] showed for free chlorine concentrations of $c(\text{Cl}_2) = 1,4$ and 7,1 µmol/l $[\varrho(\text{Cl}_2) = 0,1$ and 0,5 mg/l], relative standard deviations of 4,0 % and 2,0 %, respectively.

The results presented in the preceding paragraphs relate to replicate determinations in the same laboratory and thus provide a measure of the repeatability of the method. Attempts in the past to quantify the reproducibility of the method by distribution of samples to different laboratories have produced unreliable results because of a general instability of solutions containing free and combined chlorine. More recently it has been found by the Quality Assurance Branch of EMSL-Cincinnati^[6] that a sealed vial of sodium hypochlorite in very pure water is remarkably stable when stored in the dark inside a mailing tube. Subsequent evaluation by various US federal and state laboratories has produced the analytical parameters listed in Table 2 for methods in current use.

9 Interferences

Two types of interference may be noted.

9.1 Interference by other chlorine compounds

A fraction of any chlorine dioxide that might be present is measured as total chlorine. This interference may be corrected by determining the chlorine dioxide in the water.^[2, 4, 5]

9.2 Interference by compounds other than chlorine compounds

Oxidation of DPD is not specifically caused by chlorine compounds. Depending on the concentration and the chemical oxidation potential, the reaction is effected by other oxidizing agents. The following substances may be mentioned in particular: bromine, iodine, bromamines, iodoamines, ozone, hydrogen peroxide, chromate, oxidized manganese, nitrite, iron(III) ions and copper ions. The interference is suppressed by the disodium EDTA in reagents **4.2** and **4.3** in the case of copper(II) ions (< 8 mg/l) and iron(III) ions (< 20 mg/l).

Interference by chromate may be eliminated by addition of barium chloride.^[7]

10 Test report

The test report shall include the following information:

- a) a reference to this part of ISO 7393;
- b) all information necessary for complete identification of the sample;
- c) the results and the method of expression used;
- d) details of any operations not included in this part of ISO 7393, or regarded as optional, together with any circumstance that may have affected the results.

Table 2 — Analytical parameters from interlaboratory analysis for free residual chlorine

True value			Number of				
$c(Cl_2)$	$[\varrho(\operatorname{Cl}_2)]$	Method code ^a	laboratories	Mean		Standard deviation	
2/	15(-2)1		(observations)				
μmol/l	(mg/l)			μmol/l	(mg/l)	μmol/l	(mg/l)
7	(0,5)	A	6	6,2	(0,44)	1,3	(0,09)
		В	7	6,8	(0,48)	1,8	(0,13)
11,3	(0,80)	A	10	10,9	(0,77)	1,1	(0,08)
		В	14	11,1	(0,79)	4,1	(0,29)
		C	6	11,6	(0,82)	1,3	(0.09)
15,5	(1,10)	A	10	15,5	(1,10)	2,0	(0,14)
		В	14	16,2	(1,15)	5,5	(0,39)
		C	6	16,5	(1,17)	1,1	(0,08)
18,2	(1,29)	A	6	18,6	(1,32)	1,3	(0.09)
,		В	7	19,9	(1,41)	5,4	(0,38)

^a A: Iodometric titration amperometric

B: DPD colorimetric

C: DPD titrimetric

Annex A Separate determinations of combined chlorine of the monochloramine type, combined chlorine of the dichloramine type and of combined chlorine in the form of nitrogen trichloride

A.1 Applicability

This annex specifies a method for the differentiation between combined chlorine of the monochloramine type, combined chlorine of the dichloramine type and combined chlorine in the form of nitrogen trichloride. The field of application of the method is the same as that for concentrations of free chlorine and total chlorine (see clause 1).

A.2 Principle

After determination of free chlorine and total chlorine, colorimetric measurement of two further test portions:

a) on the third test portion: reaction with DPD limited to free chlorine and to combined chlorine of the monochloramine type by the addition of a small quantity of potassium iodide;

b) on the fourth test portion, by addition of a small quantity of potassium iodide before the addition of buffer and DPD reagent: reaction with DPD by free chlorine, by combined chlorine of the monochloramine type and one half of the nitrogen trichloride.

Combined chlorine of the dichloramine type does not react in either of these two cases. Calculation of the concentration of combined chlorine of the monochloramine and dichloramine types and the concentration of nitrogen trichloride.

A.3 Reagents

The reagents given in clause 4 and

Potassium iodide, solution, 5 g/l.

Prepare this solution on the day of use and store in a brown bottle.

A.4 Apparatus

See clause 5.

A.5 Procedure

A.5.1 Test sample

See **6.1**.

A.5.2 Test portions

Work on two test portions similar to those in 6.2.

A.5.3 Calibration

See **6.3**.

A.5.4 Determination of free chlorine and combined chlorine of the monochloramine type

Place rapidly in a 250 ml conical flask in this order: 5,0 ml of buffer solution (4.2), 5,0 ml of DPD reagent (4.3), the third test portion and two drops (about 0,1 ml) of potassium iodide solution (clause A.3) or a very small crystal of potassium iodide (about 0,5 mg) and mix. Fill the measuring cell with this treated solution and immediately measure the colour under the same conditions as adopted for the calibration (A.5.3). Record c_4 , the concentration reading from the comparator scale or calibration graph (A.5.3).

A.5.5 Determination of free chlorine, combined chlorine of the monochloramine type and one half of the nitrogen trichloride

Place in a 250 ml beaker, the fourth test portion and two drops (about 0,1 ml) of potassium iodide solution (clause $\mathbf{A.3}$) or a very small crystal of potassium iodide (about 0,5 mg) and mix. Transfer the contents of the beaker to a 250 ml conical flask containing 5,0 ml of buffer solution (4.2) and 5,0 ml of DPD reagent (4.3) added less than 1 min prior to the transfer. Fill the measuring cell with this treated solution and immediately measure the colour under the same conditions as adopted for the calibration ($\mathbf{A.5.3}$). Record c_5 , the concentration reading from the comparator scale or calibration graph ($\mathbf{A.5.3}$).

A.6 Expression of results

A.6.1 Method of calculation

A.6.1.1 Calculation of the concentration of combined chlorine of the monochloramine type

The concentration of combined chlorine of the monochloramine type, $c(Cl_2)$, expressed in millimoles per litre, is given by the equation

$$c(Cl_2) = \frac{(c_4 - c_1)V_0}{V_1}$$

where

 c_4 is the concentration, expressed in millimoles of Cl_2 per litre, of chlorine as determined in **A.5.4**:

 c_1 , V_0 and V_1 are as defined in clause 8.1.

A.6.1.2 Calculation of the concentration of combined chlorine of the dichloramine type

The concentration of combined chlorine of the dichloramine type, $c(\operatorname{Cl}_2)$, expressed in millimoles per litre is given by the equation

$$I_2$$
) = $\frac{(c_2 - 2c_5 + c_4)}{V_1}$

where

 c_2 , V_0 and V_1 are as defined in 8.1;

 c_4 is as defined in **A.6.1.1**;

 c_5 is the concentration, expressed in millimoles of Cl_2 per litre, of chlorine as determined in **A.5.5**.

A.6.1.3 Calculation of the concentration of combined chlorine in the form of nitrogen trichloride

The concentration of combined chlorine in the form of nitrogen trichloride, $c(Cl_2)$, expressed in millimoles per litre, is given by the equation

$$c(Cl_2) = \frac{2(c_5 - c_4) V_0}{V_1}$$

where

 c_5 is as defined in **A.6.1.2**;

 c_4 is as defined in **A.6.1.1**;

 V_0 and V_1 are as defined in 8.1.

A.6.2 Conversion of amount of substance concentration to mass concentration

The chlorine concentration expressed in moles per litre may be expressed in grams per litre by multiplying by a conversion factor of 70,91.

Annex B Preparation of water free from oxidizing and reducing substances

To obtain dilution water of the desired quality, demineralized or distilled water is first chlorinated to a level of about 0,14 mmol/l (10 mg/l) and stored in a well-stoppered carboy for at least 16 h. The water is then dechlorinated by exposure to UV irradiation, sun-light for several hours or by contact with active carbon. Finally check the quality using the procedures as in **4.1**.

Bibliography

- [1] BENDER, D.F. Comparison of methods for the determination of total available residual chlorine in various sample matrices, Report No.
- EPA-600/4-78-019. Cincinnati, Ohio 45268, USA, US Environmental Protection Agency, 1978.
- [2] DoE. Chemical Disinfecting Agents in Waters, and Effluents, and Chlorine Demand, Methods for the Examination of Waters and Associated Materials. London, UK, HMSO, 1980.
- [3] NICOLSON N.J. An evaluation of the methods of determining residual chlorine in water. *Analyst* **90** 1965:187.
- [4] PALIN A.T. Methods for the determination in water of free and combined available chlorine, chlorine dioxide and chlorite, bromine, iodine and ozone, using diethyl-p-phenylenediamine. J. Inst. Water Eng. 21 1967:537
- [5] PALIN A.T. Analytical control of water disinfection with special reference to differential DPD methods for chlorine, chlorine dioxide, bromine, iodine and ozone. *J. Inst. Water Eng.* **28** 1974:139.
- [6] Studies WS007 and WS008, Cincinnati,
- Ohio 45268, USA, Quality Assurance Branch, Environmental Monitoring and Support Laboratory, Office of Research and Development, US Environmental Protection Agency, 1980.
- [7] PALIN A.T. New correction procedures for chromate interference in the DPD method for residual free and combined chlorine. *J. Inst. Water Eng. Sci.* **36** 1982:351.

BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: 020 8996 9000. Fax: 020 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: 020 8996 9001. Fax: 020 8996 7001.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre. Tel: 020 8996 7111. Fax: 020 8996 7048.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: 020 8996 7002. Fax: 020 8996 7001.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

If permission is granted, the terms may include royalty payments or a licensing agreement. Details and advice can be obtained from the Copyright Manager. Tel: 020 8996 7070.