

BS EN ISO 9173-1:2016



BSI Standards Publication

Dentistry — Extraction forceps

Part 1: General requirements (ISO 9173-1:2016)

National foreword

This British Standard is the UK implementation of EN ISO 9173-1:2016. It supersedes BS EN ISO 9173-1:2006 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/106/4, Dental Instruments and Equipment.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2016.
Published by BSI Standards Limited 2016

ISBN 978 0 580 85929 8

ICS 11.060.20

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 December 2016.

Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

English Version

Dentistry - Extraction forceps - Part 1: General requirements (ISO 9173-1:2016)

Médecine bucco-dentaire - Daviers - Partie 1:
Exigences générales (ISO 9173-1:2016)

Zahnheilkunde - Extraktionszangen - Teil 1: Allgemeine
Anforderungen (ISO 9173-1:2016)

This European Standard was approved by CEN on 20 August 2016.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

European foreword

This document (EN ISO 9173-1:2016) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2017, and conflicting national standards shall be withdrawn at the latest by May 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 9173-1:2006.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 9173-1:2016 has been approved by CEN as EN ISO 9173-1:2016 without any modification.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	2
4.1 Materials.....	2
4.1.1 Component parts and fastening components.....	2
4.2 Maximum overall length.....	2
4.3 Rockwell hardness.....	2
4.4 Surface finish.....	2
4.4.1 All surfaces.....	2
4.4.2 Beaks.....	3
4.5 Resistance to reprocessing.....	3
4.6 Joint opening/closing force.....	3
4.7 Joint movement.....	3
5 Test methods	3
5.1 Visual inspection.....	3
5.2 Resistance to reprocessing.....	3
5.3 Joint opening/closing force.....	3
5.4 Joint movement.....	3
6 Marking	4
Bibliography	5

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT) see the following URL: www.iso.org/iso/foreword.html.

The committee responsible for this document is ISO/TC 106, *Dentistry*, Subcommittee SC 4, *Dental instruments*.

This third edition cancels and replaces the second edition (ISO 9173-1:2006), which has been technically revised with the following changes:

- a) as reprocessing test, only the autoclave test was selected;
- b) the reprocessing cycles were increased to 100 cycles;
- c) the boiling water test was deleted.

ISO 9173 consists of the following parts, under the general title *Dentistry — Extraction forceps*:

- *Part 1: General requirements*
- *Part 2: Designation*
- *Part 3: Design*

Introduction

This revision of ISO 9173-1 is intended to cover all extraction forceps used in dentistry.

Dentistry — Extraction forceps —

Part 1: General requirements

1 Scope

This part of ISO 9173 specifies the general performance requirements for extraction forceps used in dentistry.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 6508-1, *Metallic materials — Rockwell hardness test — Part 1: Test method (scales A, B, C, D, E, F, G, H, K, N, T)*

ISO 17664, *Sterilization of medical devices — Information to be provided by the manufacturer for the processing of resterilizable medical devices*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

3.1

extraction forceps

type of pincers used for the extraction of teeth

3.2

beak

functional working end of forceps which enclose the teeth

3.3

facial beak

beak that is designed to be in contact with the facial surface of the tooth

3.4

lingual beak

beak that is designed to be in contact with the lingual surface of the tooth

3.5

beak separation

minimum gap between beak tips with the extraction forceps closed

3.6

overall beak length

distance from beak tip to pivot centre

3.7 fastening components

component of the extraction forceps used for fastening the pincers

EXAMPLE Pins, rivets and screws.

4 Requirements


4.1 Materials

4.1.1 Component parts and fastening components

All stainless steel components shall conform to the requirements of 4.2 to 4.6. Examples of suitable stainless steels can be found in ISO 7153-1.

4.2 Maximum overall length

The maximum overall length of extraction forceps shall not exceed 200 mm.



Dimensions in millimetres

Key

- 1 beak
- 2 joint
- 3 handle

Figure 1 — General description of extraction forceps

4.3 Rockwell hardness

The component parts of the extraction forceps, with the exception of the fastening components, shall be heat treated to attain a Rockwell hardness value of 40 HRC to 52 HRC when determined in accordance with ISO 6508-1.

Mating surfaces on the same extraction forceps, such as those on opposite beaks, shall not vary in hardness by more than 3 units on the Rockwell hardness scale C.

4.4 Surface finish

4.4.1 All surfaces

All surfaces shall be free from surface defects and residues and shall also be free from any non-functional sharp edges, including burrs or flash.

Test in accordance with 5.1.

4.4.2 Beaks

When the inner or crown space of the extraction forceps is serrated or textured, the serration or texture shall be consistent within any one pattern of extraction forceps.

Test in accordance with [5.1](#).

4.5 Resistance to reprocessing

The extraction forceps shall show no evidence of corrosion and shall comply with all the requirements in this part of ISO 9173 after 100 reprocessing cycles, as recommended by the manufacturer.

Test in accordance with [5.2](#).

NOTE Discolouration due to water marking does not constitute evidence of corrosion.

4.6 Joint opening/closing force

The joint of the extraction forceps shall be so constructed that the extraction forceps opens and closes smoothly.

The torque required to operate shall be between 0,20 N m and 0,68 N m.

Test in accordance with [5.3](#).

4.7 Joint movement

There shall be no perceptible sideways movement in any position of the joint between fully closed and a maximum handle opening of 50 mm when tested according to [5.4](#).

5 Test methods

5.1 Visual inspection

Visual inspection shall be carried out under normal vision, without magnification.

5.2 Resistance to reprocessing

Carry out 100 reprocessing cycles as recommended in the manufacturer's instructions, in accordance with ISO 17664. Visually inspect for any signs of corrosion. Retest for requirements as in [4.3](#) and [4.7](#). For the purpose of this test, a reprocessing cycle includes the methods of cleaning, disinfection and sterilization.

5.3 Joint opening/closing force

Mount the extraction forceps in a suitable apparatus to obtain a handle opening of 50 mm. Apply a force just sufficient to close the extraction forceps fully and record the moment of this force.

Compare the values obtained with the limits given in [4.6](#).

5.4 Joint movement

Grip each handle end between the index finger and thumb of each hand, and apply a force perpendicular to the hinge axis.

6 Marking

Each extraction forceps shall be permanently marked with the following information:

- a) name of manufacturer and/or brand name;
- b) pattern number;
- c) lot number;
- d) ISO 9173 (optional).

Bibliography

- [1] ISO 7153-1, *Surgical instruments — Metallic materials — Part 1: Stainless steel*
- [2] ISO 9173-2, *Dentistry — Extraction forceps — Part 2: Designation*
- [3] ISO 9173-3, *Dentistry — Extraction forceps — Part 3: Design*

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Copyright in BSI publications

All the content in BSI publications, including British Standards, is the property of and copyrighted by BSI or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use.

Save for the provisions below, you may not transfer, share or disseminate any portion of the standard to any other person. You may not adapt, distribute, commercially exploit, or publicly display the standard or any portion thereof in any manner whatsoever without BSI's prior written consent.

Storing and using standards

Standards purchased in soft copy format:

- A British Standard purchased in soft copy format is licensed to a sole named user for personal or internal company use only.
- The standard may be stored on more than 1 device provided that it is accessible by the sole named user only and that only 1 copy is accessed at any one time.
- A single paper copy may be printed for personal or internal company use only.

Standards purchased in hard copy format:

- A British Standard purchased in hard copy format is for personal or internal company use only.
- It may not be further reproduced – in any format – to create an additional copy. This includes scanning of the document.

If you need more than 1 copy of the document, or if you wish to share the document on an internal network, you can save money by choosing a subscription product (see 'Subscriptions').

Reproducing extracts

For permission to reproduce content from BSI publications contact the BSI Copyright & Licensing team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email subscriptions@bsigroup.com.

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Useful Contacts

Customer Services

Tel: +44 345 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 345 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK