

Dentistry — Extraction forceps

Part 2: Designation (ISO 9173-2:2010)

ICS 11.060.20

National foreword

This British Standard is the UK implementation of EN ISO 9173-2:2010.

The UK participation in its preparation was entrusted to Technical Committee CH/106/4, Dental Instruments and Equipment.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 June 2010

© BSI 2010

ISBN 978 0 580 62322 6

Amendments/corrigenda issued since publication

Date	Comments

ICS 11.060.20

English Version

Dentistry - Extraction forceps - Part 2: Designation (ISO 9173-2:2010)

Médecine bucco-dentaire - Daviers - Partie 2: Désignation
(ISO 9173-2:2010)

Zahnheilkunde - Extraktionszangen - Teil 2: Bezeichnung
(ISO 9173-2:2010)

This European Standard was approved by CEN on 5 April 2010.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Foreword

This document (EN ISO 9173-2:2010) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2010, and conflicting national standards shall be withdrawn at the latest by November 2010.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

Endorsement notice

The text of ISO 9173-2:2010 has been approved by CEN as a EN ISO 9173-2:2010 without any modification.

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 9173-2 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 4, *Dental instruments*.

ISO 9173 consists of the following parts, under the general title *Dentistry — Extraction forceps*:

- *Part 1: General requirements and test methods*
- *Part 2: Designation*

Introduction

Dental extraction forceps are used for the extraction of teeth. The majority of extraction forceps are for use on both sides and for more than one tooth, e.g. for all upper or lower premolars, for upper central incisors and upper canines (and larger lateral incisors), for upper lateral incisors and for upper deciduous incisors and canines. Some extraction forceps for upper molars and some extraction forceps for lower third molars are specially designed for the type of tooth and the side concerned. Some extraction forceps are only for roots.

In order that the dentist and/or his assistant can order the correct extraction forceps for an intended clinical purpose a global designation system is necessary.

General requirements and test methods for extraction forceps are specified in ISO 9173-1.

Dentistry — Extraction forceps —

Part 2: Designation

1 Scope

This part of ISO 9173 specifies the designation of dental extraction forceps.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 3166-1, *Codes for the representation of names of countries and their subdivisions — Part 1: Country codes*

ISO 3950, *Dentistry — Designation system for teeth and areas of the oral cavity*

ISO 9173-1, *Dentistry — Extraction forceps — Part 1: General requirements and test methods*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942, ISO 9173-1 and the following apply.

3.1

pattern number

specific brand number for the extraction forceps, defined by the manufacturer

3.2

functional designation

designation for the extraction forceps based on the intended clinical application

4 Designation

4.1 General

Extraction forceps shall be designated by their pattern number and their functional designation.

Designation = pattern number [functional designation]

4.2 Pattern number

The pattern number shall be the traditional figure (pattern) number (e.g. #210H) of the extraction forceps used in the region of origin of the manufacturer.

4.3 Functional designation

4.3.1 General information

The information provided by this part of ISO 9173 for functional designation of extraction forceps is intended as guidance for clinicians in the selection and application of dental extraction forceps.

The functional designation shall be in accordance with the digital designations of teeth and areas of the oral cavity as specified in ISO 3950, with the exception of quadrants or individual tooth numbers (see Figure 1).

00															Oral cavity	
01															Maxillary area	
10							20								Quadrant	
03					04					05					Sextant	
18	17	16	15	14	13	12	11	21	22	23	24	25	26	27	28	
			55	54	53	52	51	61	62	63	64	65				
			85	84	83	82	81	71	72	73	74	75				
48	47	46	45	44	43	42	41	31	32	33	34	35	36	37	38	
08					07					06					Sextant	
40							30								Quadrant	
02															Mandibular area	

Key

White area = used for extraction forceps.

Grey area = not used for extraction forceps.

Figure 1 — Designation of teeth and areas of the oral cavity

The remaining designation, used for extraction forceps, is shown in Table 1.

Table 1 — Remaining designation for areas of the oral cavity, used for extraction forceps

Area	Designation
Oral cavity (whole)	00
Maxillary area	01
Mandibular area	02
Upper right sextant	03
Upper anterior sextant	04
Upper left sextant	05
Lower left sextant	06
Lower anterior sextant	07
Lower right sextant	08

4.3.2 Specification of functional designation

The functional designation consists of the following three parts:

Functional designation = [First designator.Second designator.Third designator]

The first designator refers to the jaw as specified in Table 2.

The second designation refers to the sextant of the dental arch as specified in Table 3.

The third designator refers to the special use as specified in Table 4.

Table 2 — First designator used to describe the functional application of extraction forceps

Jaw	Common name	Digit
Universal	Both	00
Maxillary	Upper	01
Mandibular	Lower	02

Table 3 — Second designator used to describe the functional application of extraction forceps

Sextant	Digit
Anterior	04, 07
Maxillary posterior	03, 05
Mandibular posterior	06, 08

Table 4 — Third designator used to describe the functional application of extraction forceps

Special use	Digit
No special use	0
Pediatric forceps	1
Root forceps	2
Tooth splitting forceps	3
Premolar forceps	4
Molar forceps	5
Third molar forceps	6
Other forceps	7

When the extraction forceps are intended for use in more than one sextant, multiple sextants are separated by a slash mark as in the following example: second designator = [06/08].

For posterior universal forceps, separate the upper and lower sextants using a hyphen, e.g. second designator = [03/05-06/08].

4.3.3 Examples of designation for extraction forceps

The country codes used in the following examples are in accordance with ISO 3166-1, i.e. the alpha 2 codes for the countries (e.g. US, JP).

EXAMPLE 1 Forceps used for the extraction of upper anterior teeth in adults

#1 US Pattern = upper anterior forceps:

designation = #1 [01.04/07.0].

EXAMPLE 2 Two forceps with the same manufacturer pattern number that have different functional uses

#67 UK Pattern = upper molar forceps:

designation = #67 [01.03/05.5],

#67 JP Pattern = upper tooth splitting forceps:

designation = #67 [01.03/05.3].

EXAMPLE 3 Forceps used for the extraction of anterior teeth in children

#1S US Pattern = universal anterior pediatric forceps:

designation = #1S [00.04/07.1].

5 Marking, labelling and catalogue entries

5.1 Marking of extraction forceps

As a minimum requirement, extraction forceps shall be indelibly marked with their pattern number.

At the discretion of the manufacturer, extraction forceps should be indelibly marked with the designation as specified in 4.1.

5.2 Labelling on the primary packaging of extraction forceps

As a minimum requirement, the primary packaging shall be labelled with the pattern number.

The primary packaging of extraction forceps should be labelled with the designation as specified in 4.1.

5.3 Catalogue entries

Entries in electronic and/or printed catalogues of extraction forceps shall include the designation as specified in 4.1.

BSI - British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: +44 (0)20 8996 9000. Fax: +44 (0)20 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: +44 (0)20 8996 9001. Fax: +44 (0)20 8996 7001 Email: orders@bsigroup.com You may also buy directly using a debit/credit card from the BSI Shop on the Website <http://www.bsigroup.com/shop>

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact Information Centre. Tel: +44 (0)20 8996 7111 Fax: +44 (0)20 8996 7048 Email: info@bsigroup.com

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: +44 (0)20 8996 7002 Fax: +44 (0)20 8996 7001 Email: membership@bsigroup.com

Information regarding online access to British Standards via British Standards Online can be found at <http://www.bsigroup.com/BSOL>

Further information about BSI is available on the BSI website at <http://www.bsigroup.com>.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

Details and advice can be obtained from the Copyright and Licensing Manager. Tel: +44 (0)20 8996 7070 Email: copyright@bsigroup.com