

# Biological evaluation of medical devices —

## Part 8: Selection and qualification of reference materials for biological tests

The European Standard EN ISO 10993-8:2001 has the status of a  
British Standard

ICS 11.040.01; 11.100

## National foreword

This British Standard is the official English language version of EN ISO 10993-8:2001. It is identical with ISO 10993-8:2000.

The UK participation in its preparation was entrusted to Technical Committee CH/26, Biological testing of medical and dental materials and devices, which has the responsibility to:

- aid enquirers to understand the text;
- present to the responsible international/European committee any enquiries on the interpretation, or proposals for change, and keep the UK interests informed;
- monitor related international and European developments and promulgate them in the UK.

A list of organizations represented on this committee can be obtained on request to its secretary.

### Cross-references

The British Standards which implement international or European publications referred to in this document may be found in the BSI Standards Catalogue under the section entitled “International Standards Correspondence Index”, or by using the “Find” facility of the BSI Standards Electronic Catalogue.

A British Standard does not purport to include all the necessary provisions of a contract. Users of British Standards are responsible for their correct application.

**Compliance with a British Standard does not of itself confer immunity from legal obligations.**

### Summary of pages

This document comprises a front cover, an inside front cover, the EN ISO title page, EN ISO foreword page, ISO title page, pages ii to v, a blank page, pages 1 to 4, an annex ZA page and a back cover.

The BSI copyright notice displayed in this document indicates when the document was last issued.

### Amendments issued since publication

Amd. No.	Date	Comments

This British Standard, having been prepared under the direction of the Health and Environment Sector Committee, was published under the authority of the Standards Committee and comes into effect on 15 August 2001

© BSI 08-2001

ISBN 0 580 37872 1

ICS 11.100

English version

**Biological evaluation of medical devices - Part 8: Selection and qualification of reference materials for biological tests (ISO 10993-8:2000)**

Evaluation biologique des dispositifs médicaux - Partie 8:  
Sélection et qualification des matériaux de référence  
utilisés pour les essais biologiques (ISO 10993-8:2000)

Biologische Beurteilung von Medizinprodukten - Teil 8:  
Auswahl und Eignung von Referenzmaterialien für  
biologische Prüfungen (ISO 10993-8:2000)

This European Standard was approved by CEN on 1 September 2000.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**Central Secretariat: rue de Stassart, 36 B-1050 Brussels**

## Foreword

The text of the International Standard ISO 10993-8:2000 has been prepared by Technical Committee ISO/TC 194 "Biological evaluation of medical devices" in collaboration with Technical Committee CEN/TC 206 "Biocompatibility of medical and dental materials and devices", the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2001, and conflicting national standards shall be withdrawn at the latest by March 2001.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

## Endorsement notice

The text of the International Standard ISO 10993-8:2000 was approved by CEN as a European Standard without any modification.

NOTE: Normative references to International Standards are listed in annex ZA (normative).

---

---

**Biological evaluation of medical devices —**  
**Part 8:**  
**Selection and qualification of reference**  
**materials for biological tests**

*Évaluation biologique des dispositifs médicaux —*

*Partie 8: Sélection et qualification des matériaux de référence utilisés pour  
les essais biologiques*



Reference number  
ISO 10993-8:2000(E)

.....

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 3.

Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this part of ISO 10993 may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

International Standard ISO 10993-8 was prepared by Technical Committee ISO/TC 194, *Biological evaluation of medical devices*.

ISO 10993 consists of the following parts, under the general title *Biological evaluation of medical devices*:

- *Part 1: Evaluation and testing*
- *Part 2: Animal welfare requirements*
- *Part 3: Tests for genotoxicity, carcinogenicity and reproductive toxicity*
- *Part 4: Selection of tests for interactions with blood*
- *Part 5: Tests for in vitro cytotoxicity*
- *Part 6: Tests for local effects after implantation*
- *Part 7: Ethylene oxide sterilization residuals*
- *Part 8: Selection and qualification of reference materials for biological tests*
- *Part 9: Framework for identification and quantification of potential degradation products*
- *Part 10: Tests for irritation and sensitization*
- *Part 11: Tests for systemic toxicity*
- *Part 12: Sample preparation and reference materials*
- *Part 13: Identification and quantification of degradation products from polymeric medical devices*
- *Part 14: Identification and quantification of degradation products from ceramics*
- *Part 15: Identification and quantification of degradation products from metals and alloys*
- *Part 16: Toxicokinetic study design for degradation products and leachables*

## **EN ISO 10993-8:2000**

- *Part 17: Establishment of allowable limits for leachable substances using health-based risk assessment*
- *Part 18: Chemical characterization of materials*

Future parts will deal with other relevant aspects of biological testing.

.....



## **Introduction**

The information given in this part of ISO 10993 should be considered as a supplement to ISO 10993-12 which specifies requirements and gives guidance on procedures to be followed in the preparation of samples of medical devices for testing in biological systems in accordance with one or more parts of the ISO 10993 series. In clause 4 of ISO 10993-12:1996, there is a discussion of the use of reference materials as experimental controls. Annex A of ISO 10993-12:1996 provides specific information regarding current sources of commercially available reference materials.

# EN ISO 10993-8:2000

# Biological evaluation of medical devices —

## Part 8:

# Selection and qualification of reference materials for biological tests

## 1 Scope

This part of ISO 10993 specifies requirements on the use of reference materials or certified reference materials used to determine the biological response of a material. It specifies the selection and qualification of reference materials for biological tests and the characteristics of reference materials for the use of reference materials as experimental controls.

## 2 Normative references

The following normative documents contain provisions which, through reference in this text, constitute provisions of this part of ISO 10993. For dated references, subsequent amendments to, or revisions of, any of these publications do not apply. However, parties to agreements based on this part of ISO 10993 are encouraged to investigate the possibility of applying the most recent editions of the normative documents indicated below. For undated references, the latest edition of the normative document referred to applies. Members of ISO and IEC maintain registers of currently valid International Standards.

ISO 10993-12, *Biological evaluation of medical devices — Part 12: Sample preparation and reference materials*.

ISO Guide 30:1992, *Terms and definitions used in connection with reference materials*.

## 3 Terms and definitions

For the purposes of this part of ISO 10993, the terms and definitions given in ISO 10993-12, ISO Guide 30 and the following apply.

### 3.1

#### certified reference material

#### CRM

reference material, accompanied by a certificate, one or more of whose property values are certified by a procedure which establishes its traceability to an accurate realization of the unit in which the property values are expressed, and for which each certified value is accompanied by an uncertainty at a stated level of confidence

[ISO Guide 30:1992, 2.2]

NOTE Standard reference material (SRM) is a trademark-protected certification supplied by the National Institutes for Standards and Technology, Gaithersburg, MD, USA.

## 3.2

### reference material

#### RM

material or substance one or more of whose property values are sufficiently homogeneous and well established to be used for the calibration of an apparatus, the assessment of a measurement method, or for assigning values to materials

[ISO Guide 30:1992, 2.1]

## 3.3

### homogeneous

condition of being of uniform structure or composition with respect to the biological endpoint under study

NOTE 1 The RM is said to be homogeneous if the biological response to a specific test is found to lie within the specified uncertainty limits of the test, irrespective of the site in the batch or lot of material from where the test sample is taken.

NOTE 2 Adapted from ISO Guide 30:1992, 2.6.

## 3.4

### reference method

thoroughly investigated test method that clearly and exactly describes the necessary conditions and procedures for the evaluation of a specific biological endpoint, has been shown to have accuracy and precision commensurate with its intended use and can, therefore, be used to characterize a RM

NOTE Adapted from ISO Guide 30:1992, 3.10.

## 3.5

### stability of property values

ability of a material, when stored under specified conditions, to maintain the stated biological response, within specified limits, for a specific period of time

NOTE Adapted from ISO Guide 30:1992, 2.7.

## 4 Use of certified reference materials or reference materials

Reference materials (RM) or certified reference materials (CRM) shall be used in biological tests as control materials to qualify in-house tests and control materials. They demonstrate the suitability of the procedure to yield a reproducible, predictable response, e.g. positive or negative. Use of a reference material in this way will ensure the comparability of the response between laboratories.

The property values of any material used in this way shall be characterized with each biological test procedure for which the use of the material is desired. A material characterized and then certified for one reference test method or response, e.g. sensitization [6], shall not be used as a reference material for another, e.g. cytotoxicity [5], without additional characterization.

## 5 Characteristics of reference materials

### 5.1 One or more property values

RMs or CRMs used to determine the biological response of a material shall be evaluated with each biological test procedure for which the use of the material is desired. It is not sufficient to qualify a material for one type of reference test method or response, e.g. sensitization [6], and declare it a reference material for another, e.g. cytotoxicity [5], without additional qualification testing.

## **5.2 Long-term availability of the reference material**

To ensure the long-term availability of a reference material for determination of biological response, the user of the material should obtain a commitment as long as possible, preferably not less than five years, from the supplier of the RMs or CRMs.

A second, but less desirable, option is the publication by the reference material supplier of an "open formulation" for the material, i.e. publication of the source materials and details of the processing needed to ensure uniform batches of RM.

## **5.3 Certification of reference materials for biological safety testing**

The biological response of the certified reference material under specific test conditions shall be established through interlaboratory studies.

# **6 Use of reference materials as experimental controls**

## **6.1 Material screening versus biocompatibility of medical devices**

The use of the reference materials described in ISO 10993-12 and elsewhere in this part of ISO 10993 are limited to biological screening of the materials intended for use in the manufacture of medical devices. However, while not intended for the purpose, they are often used in the performance assessment of the finished medical device. The vertical standard for the device, when available, shall address the biological testing of the product in the performance environment of the device. Biological testing described in the vertical product standard takes precedence over testing performed to screen the materials for suitability.

## **6.2 Misuse of CRMs**

The attention of the users of this part of ISO 10993 is directed to the discussion of "proper use" and "misuse" of CRMs in the introduction to ISO Guide 33. This discussion points out areas of potential under- and over-utilization of RMs and CRMs.

Users of this part of ISO 10993 shall note that the use of calibration materials to evaluate the biological response of materials under investigation within a single laboratory is acceptable.

## Bibliography

The ISO Guides listed below provide material which will enable the reader to place the discussion in this International Standard into context.

- [1] ISO Guide 31:1981, *Contents of certificates of reference materials*.
- [2] ISO Guide 33:1989, *Uses of certified reference materials*.
- [3] ISO Guide 35:1989, *Certification of reference materials — General and statistical principles*.
- [4] ISO/IEC Directives, Part 2:1992, *Methodology for the development of International Standards*<sup>1)</sup>.

The following International Standards contain additional information on biological evaluation procedures for specific biological endpoints.

- [5] ISO 10993-5:1999, *Biological evaluation of medical devices — Part 5: Tests for in vitro cytotoxicity*.
- [6] ISO 10993-10:1995, *Biological evaluation of medical devices — Part 10: Tests for irritation and sensitization*.

---

1) At present under revision. It is anticipated that the contents will be divided between the revision of Part 1 and a new Part 2 comprising mainly the revision of present Part 3.

**Annex ZA (normative)**  
**Normative references to international publications**  
**with their relevant European publications**

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

<u>Publication</u>	<u>Year</u>	<u>Title</u>	<u>EN</u>	<u>Year</u>
ISO 10993-12	1996	Biological evaluation of medical devices – Part 12: Sample preparation and reference materials	EN ISO 10993-12	1996

---

---

## BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

### Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: 020 8996 9000. Fax: 020 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

### Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: 020 8996 9001. Fax: 020 8996 7001. Standards are also available from the BSI website at <http://www.bsi-global.com>.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

### Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre. Tel: 020 8996 7111. Fax: 020 8996 7048.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: 020 8996 7002. Fax: 020 8996 7001. Further information about BSI is available on the BSI website at <http://www.bsi-global.com>.

### Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

If permission is granted, the terms may include royalty payments or a licensing agreement. Details and advice can be obtained from the Copyright Manager. Tel: 020 8996 7070.

BSI  
389 Chiswick High Road  
London  
W4 4AL