

Packaging for terminally sterilized medical devices —

**Part 2: Validation requirements for
forming, sealing and assembly
processes**

ICS 11.080.30

National foreword

This British Standard is the UK implementation of EN ISO 11607-2:2006+A1:2014. It is identical to ISO 11607-2:2006, incorporating amendment 1:2014. It supersedes BS EN ISO 11607-2:2006, which is withdrawn.

The start and finish of text introduced or altered by amendment is indicated in the text by tags. Tags indicating changes to ISO text carry the number of the ISO amendment. For example, text altered by ISO amendment 1 is indicated by **A1** **A1**.

The UK participation in its preparation was entrusted to Technical Committee CH/198, Sterilization of medical devices.

A list of organizations represented on this committee can be obtained on request to its secretary.

The publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 May 2006

© The British Standards Institution 2014.
Published by BSI Standards Limited 2014

Amendments/corrigenda issued since publication

Date	Comments
30 September 2014	Implementation of ISO amendment 1:2014 with CEN endorsement A1:2014. Annex ZA amended

ICS 11.080.30

English Version

**Packaging for terminally sterilized medical devices - Part 2:
Validation requirements for forming, sealing and assembly
processes (ISO 11607-2:2006)**

Emballages des dispositifs médicaux stérilisés au stade
terminal - Partie 2: Exigences de validation pour les
procédés de formage, scellage et assemblage (ISO 11607-
2:2006)

Verpackungen für in der Endverpackung zu sterilisierende
Medizinprodukte - Teil 2: Validierungsanforderungen an
Prozesse der Formgebung, Siegelung und des
Zusammenstellens (ISO 11607-2:2006)

This European Standard was approved by CEN on 13 April 2006.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Foreword

This document (EN ISO 11607-2:2006) has been prepared by Technical Committee ISO/TC 198 "Sterilization of health care products" in collaboration with Technical Committee CEN/TC 102 "Sterilizers for medical purposes", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2006, and conflicting national standards shall be withdrawn at the latest by October 2006.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Endorsement notice

The text of ISO 11607-2:2006 has been approved by CEN as EN ISO 11607-2:2006 without any modifications.

Foreword to amendment A1

This document (EN ISO 11607-2:2006/A1:2014) has been prepared by Technical Committee ISO/TC 198 "Sterilization of health care products" in collaboration with Technical Committee CEN/TC 102 "Sterilizers for medical purposes" the secretariat of which is held by DIN.

This Amendment to the European Standard EN ISO 11607:2006 shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by January 2015, and conflicting national standards shall be withdrawn at the latest by January 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

For relationship with EU Directive(s), see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 11607-2:2006/Amd 1:2014 has been approved by CEN as EN ISO 11607-2:2006/A1:2014 without any modification.

Annex ZA
(informative)

Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on medical devices

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association to provide a means of conforming to Essential Requirements of the New Approach Directive 93/42/EEC on medical devices.

Once this standard is cited in the Official Journal of the European Union under that Directive and has been implemented as a national standard in at least one Member State, compliance with the clauses of this standard given in Table ZA.1 confers, within the limits of the scope of this standard, a presumption of conformity with the corresponding Essential Requirements of that Directive and associated EFTA regulations.

Table ZA.1 — Correspondence between this International Standard and Directive 93/42/EEC on medical device

Clause(s)/subclause(s) of this International standard	Essential Requirements (ERs) of EU Directive 93/42/EEC	Qualifying remarks/notes
4.1, 4.3, 5, 6, 7, 8	7.6	Partly addressed. To fully address the ER, requirements for materials, sterile barrier systems and packaging systems need to be addressed (ISO 11607-1).
4.1, 4.3, 5, 6, 7, 8	8.1, 8.3	Partly addressed. To fully address the ER, requirements for materials, sterile barrier systems and packaging systems need to be addressed (ISO 11607-1).
5	8.4	
6.1	8.5	Partly addressed. Standard does not mention or define values and/or limits.
6.3	8.6, 2nd part	Partly addressed. To fully address the ER, requirements for materials, sterile barrier systems and packaging systems need to be addressed (ISO 11607-1).

Contents

Page

Foreword.....	v
Introduction	vi
1 Scope	1
2 Normative references	1
3 Terms and definitions.....	1
4 General requirements.....	4
4.1 Quality systems	4
4.2 Sampling	4
4.3 Test methods.....	4
4.4 Documentation.....	5
5 Validation of packaging processes.....	5
5.1 General.....	5
5.2 Installation qualification (IQ)	6
5.3 Operational qualification (OQ).....	6
5.4 Performance qualification (PQ)	7
5.5 Formal approval of the process validation	8
5.6 Process control and monitoring	8
5.7 Process changes and revalidation.....	8
6 Packaging system assembly	8
7 Use of reusable sterile barrier systems.....	9
8 Sterile fluid-path packaging.....	9
Annex A (informative) Process development	10
Bibliography	11

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 11607-2 was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

ISO 11607-1 and ISO 11607-2 cancel and replace ISO 11607:2003, which has been technically revised.

ISO 11607 consists of the following parts, under the general title *Packaging for terminally sterilized medical devices*:

- *Part 1: Requirements for materials, sterile barrier systems and packaging systems*
- *Part 2: Validation requirements for forming, sealing and assembly processes*

Introduction

Medical devices delivered in a sterile state should be designed, manufactured and packed to ensure that they are sterile when placed on the market and remain sterile, under documented storage and transport conditions, until the sterile barrier system is damaged or opened. Additionally, medical devices delivered in a sterile state should have been manufactured and sterilized by an appropriate, validated method.

One of the most critical characteristics of a sterile barrier system and packaging system for sterile medical devices is the assurance of sterility maintenance. The development and validation of packaging processes are crucial to ensure that sterile barrier system integrity is attained and will remain so until opened by the users of sterile medical devices.

There should be a documented process validation program demonstrating the efficacy and reproducibility of all sterilization and packaging processes. Along with the sterilization process, some of the packaging operations that can affect sterile barrier system integrity are forming, sealing, capping or other closure systems, cutting and process handling. This part of ISO 11607 provides the framework of activities and requirements to develop and validate the process used to make and assemble the packaging system. ISO 11607-1 and ISO 11607-2 are designed to meet the Essential Requirements of the European Medical Device Directives.

One significant barrier to harmonization was terminology. The terms “package”, “final package”, “final pack”, “primary pack”, and “primary package” all have different connotations around the globe and choosing one of these terms to be the harmonized basis for this part of ISO 11607 was considered a barrier to successful completion of this document. As a result, the term “sterile barrier system” was introduced to describe the minimum packaging required to perform the unique functions required of medical packaging: to allow sterilization, to provide an acceptable microbial barrier, and to allow for aseptic presentation. “Protective packaging” protects the sterile barrier system, and together they form the packaging system. “Preformed sterile barrier systems” would include any partially assembled sterile barrier systems such as pouches, header bags or hospital packaging reels.

The sterile barrier system is essential to ensure the safety of terminally sterilized medical devices. Regulatory authorities recognize the critical nature of sterile barrier systems by considering them as an accessory or a component of a medical device. Preformed sterile barrier systems sold to healthcare facilities for use in internal sterilization are considered as medical devices in many parts of the world.

Packaging for terminally sterilized medical devices —

Part 2:

Validation requirements for forming, sealing and assembly processes

1 Scope

This part of ISO 11607 specifies the requirements for development and validation of processes for packaging medical devices that are terminally sterilized. These processes include forming, sealing, and assembly of preformed sterile barrier systems, sterile barrier systems and packaging systems.

This part of ISO 11607 is applicable to industry, to health care facilities, and wherever medical devices are packaged and sterilized.

This part of ISO 11607 does not cover all requirements for packaging medical devices that are manufactured aseptically. Additional requirements may also be necessary for drug/device combinations.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11607-1, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

expiry date

indication of the date, by which the product should be used, expressed at least as the year and month

3.2

installation qualification

IQ

process of obtaining and documenting evidence that equipment has been provided and installed in accordance with its specification

[ISO/TS 11139:2006]

3.3
labelling

written, printed, electronic or graphic matter affixed to a medical device or its packaging system; or accompanying a medical device

NOTE Labelling is related to identification, technical description and use of the medical device but excludes shipping documents.

3.4
operational qualification
OQ

process of obtaining and documenting evidence that installed equipment operates within predetermined limits when used in accordance with its operational procedures

[ISO/TS 11139:2006]

3.5
packaging system

combination of the sterile barrier system and protective packaging

[ISO/TS 11139:2006]

3.6
performance qualification
PQ

process of obtaining and documenting evidence that the equipment, as installed and operated in accordance with operational procedures, consistently performs in accordance with predetermined criteria and thereby yields product meeting its specification

[ISO/TS 11139:2006]

3.7
preformed sterile barrier system

sterile barrier system that is supplied partially assembled for filling and final closure or sealing

EXAMPLE Pouches, bags and open reusable containers



[ISO/TS 11139:2006]

3.8
process development

establishing the nominal values and limit(s) for critical process parameters

3.9
product

result of a process

[ISO9000:  2005 

NOTE For the purpose of sterilization standards, product is tangible and can be raw material(s), intermediate(s), sub-assembly(ies) and health care product(s).

[ISO/TS 11139:2006]

3.10
protective packaging

configuration of materials designed to prevent damage to the sterile barrier system and its contents until the point of use

[ISO/TS 11139:2006]

3.11

repeatability

closeness of the agreement between the results of successive measurements of the same particular quantity subject to measurement (measurand) carried out under the same conditions of measurement

NOTE 1 These conditions are called repeatability conditions.

NOTE 2 Repeatability conditions can include the following:

- the same measurement procedure;
- the same observer;
- the same measuring instrument, used under the same conditions;
- the same location; and
- repetition over a short period of time.

NOTE 3 Repeatability may be expressed quantitatively in terms of the dispersion characteristics of the results.

NOTE 4 Adapted from *International Vocabulary of Basic and General Terms in Metrology*, 1993, definition 3.6.

3.12

reproducibility

closeness of the agreement between the results of measurements of the same particular quantity subject to measurement (measurand) carried out under changed conditions of measurement

NOTE 1 A valid statement of reproducibility requires specification of the conditions changed.

NOTE 2 The changed conditions may include:

- principle of measurement;
- method of measurement;
- observer;
- measuring instrument;
- reference standard;
- location;
- conditions of use; and
- time.

NOTE 3 Reproducibility may be expressed quantitatively in terms of the dispersion characteristics of the results.

NOTE 4 Adapted from *International Vocabulary of Basic and General Terms in Metrology*, 1993, definition 3.7.

3.13

reusable container

rigid sterile barrier system designed to be repeatedly used

3.14

sterile barrier system

minimum package that prevents ingress of microorganisms and allows aseptic presentation of product at the point of use

[ISO/TS 11139:2006]

3.15

sterile fluid-path packaging

system of protective port covers and/or packaging designed to ensure sterility of the portion of the medical device intended for contact with fluids

NOTE An example of sterile fluid-path packaging would be the interior of the tubing for administration of an intravenous fluid.

3.16

validation

(process) documented procedure for obtaining, recording and interpreting the results required to establish that a process will consistently yield product complying with predetermined specifications

NOTE Adapted from ISO/TS 11139:2006.

4 General requirements

4.1 Quality systems

4.1.1 The activities described within this part of ISO 11607 shall be carried out within a formal quality system.

NOTE ISO 9001 and ISO 13485 contain requirements for suitable quality systems. Additional requirements may be specified by countries or regions.

4.1.2 It $\overline{A_1}$ shall not be $\overline{A_1}$ necessary to obtain third-party certification of the quality system to fulfil the requirements of this part of ISO 11607.

4.1.3 Health care facilities $\overline{A_1}$ shall consider using $\overline{A_1}$ the quality system required by their country or region.

4.2 Sampling

The sampling plans used for selection and testing of packaging systems shall be applicable to the process being evaluated. Sampling plans shall be based upon a statistically valid rationale.

NOTE Examples of suitable sampling plans are specified in ISO 2859-1 or ISO 186. Additional sampling plans may be specified by countries or regions.

4.3 Test methods

4.3.1 All test methods used to show compliance with this part of ISO 11607 shall be validated and documented.

NOTE Annex B in ISO 11607-1:2006 contains a list of suitable test methods.

4.3.2 Test method validation shall demonstrate the suitability of the method used. The following elements shall be included:

- establishment of a rationale for the selection of the appropriate tests for the packaging system;
- establishment of acceptance criterion;

NOTE Pass/fail is a type of acceptance criterion.

- determination of test method repeatability;
- determination of test method reproducibility;
- determination of test method sensitivity for integrity tests.

4.3.3 Unless otherwise specified in the test methods, test samples shall be conditioned at (23 ± 1) °C and (50 ± 2) % relative humidity for a minimum of 24 h.

4.4 Documentation

4.4.1 Demonstration of compliance with the requirements of this part of ISO 11607 shall be documented.

4.4.2 All documentation shall be retained for a specified period of time. The retention period shall consider factors such as regulatory requirements, expiry date and traceability of the medical device or sterile barrier system.

4.4.3 Documentation of compliance with the requirements may include, but is not limited to, performance data, specifications, test results from validated test methods, and protocols and results from IQ, OQ, PQ.

4.4.4 Electronic records, electronic signatures, and handwritten signatures executed to electronic records that contribute to validation, process control or other quality decision-making processes shall be reliable.

5 Validation of packaging processes

5.1 General

5.1.1 Preformed sterile barrier systems and sterile barrier system manufacturing processes shall be validated.

Examples of these processes include, but are not limited to:

- rigid and flexible blister forming;
- pouch, reel, or bag forming and sealing;
- form/fill/seal automated processes;
- kit assembly and wrapping;
- assembly of sterile fluid-path products ;
- tray/lid sealing;
- filling and closing of reusable containers;
- sterilization sheets folding and wrapping

5.1.2 Process validation shall include, at a minimum, an installation qualification, an operational qualification, and a performance qualification in this order.

5.1.3 Process development, while not formally part of process validation, should be considered as an integral part of forming and sealing (see Annex A).

5.1.4 Validation of existing products may rely on data from previous installation and operation qualification. That data can be used for determination of the tolerances for critical parameters.

5.1.5 When similar preformed sterile barrier systems and sterile barrier system manufacturing processes are validated, a rationale for establishing similarities and identifying the worst case configuration shall be documented. As a minimum, the worst case configuration shall be validated to determine compliance with this part of ISO 11607.

NOTE For example, similarity could be established by different sizes of preformed sterile barrier systems.

5.2 Installation qualification (IQ)

5.2.1 Installation qualification shall be performed.

Some installation qualification considerations are:

- equipment design features;
- installation conditions such as wiring, utilities, functionality, etc.;
- safety features;
- equipment operating within the stated design parameters;
- supplier documentation, prints, drawings and manuals;
- spare-parts lists;
- software validation;
- environmental conditions such as cleanliness, temperature, humidity;
- documented operator training;
- operating manual or procedure.

5.2.2 Critical process parameters shall be defined.

5.2.3 Critical process parameters shall be controlled and monitored.

5.2.4 Alarms, warning systems, or machine stops shall be challenged in the event that critical process parameters exceed predetermined limits.

5.2.5 Critical process instruments, sensors, displays, controllers, etc. shall be certified as calibrated and have written calibration schedules. Calibration should be performed before and after performance qualification.

5.2.6 There shall be written preventive maintenance and cleaning schedules.

5.2.7 The application of software systems, such as programmable logic controller, data collection, and inspection systems, shall be validated to ensure that they function as intended. Functional tests shall be performed to verify the correct functioning of the software and hardware, and especially the interfaces. The system shall be checked (e.g. by entering correct and incorrect data, by simulating a loss of electrical power) to detect the availability, reliability, identity, accuracy and traceability of data or records.

5.3 Operational qualification (OQ)

5.3.1 Process parameters shall be challenged to assure that they will produce preformed sterile barrier systems, and sterile barrier systems, that meet all defined requirements under all anticipated conditions of manufacturing.

5.3.2 Preformed sterile barrier systems, and sterile barrier systems, shall be produced at both the upper and lower parameter limits, and shall exhibit the properties that meet predefined requirements. The following quality properties shall be considered.

a) For forming/assembly:

- sterile barrier system completely formed/assembled;

- product fits into the sterile barrier system;
- essential dimensions are met.

b) For sealing:

- intact seal for a specified seal width;
- channels or open seals;
- punctures or tears;
- material delamination or separation.

NOTE **A1** See EN 868-5: 2009, 4.3.2 **A1** for an example of a seal width specification.

c) For other closure systems:

- continuous closure;
- punctures or tears;
- material delamination or separation.

5.4 Performance qualification (PQ)

5.4.1 The performance qualification shall demonstrate that the process will consistently produce acceptable preformed sterile barrier systems, and sterile barrier systems, under specified operating conditions.

5.4.2 Performance qualification shall include:

- the actual or simulated product;
- process parameters established in the operational qualification;
- verification of product/package requirements;
- assurance of process control and capability;
- process repeatability and reproducibility.

5.4.3 Challenges to the process shall include conditions expected to be encountered during manufacture.

NOTE These challenges can include, but are not limited to, machine set-up and change-over procedures; process start-up and restart procedures; power failure and variations, and multiple shifts, if applicable.

5.4.4 Challenges to the process shall include at least three production runs with adequate sampling to demonstrate variability within a run and reproducibility between different runs. The duration of a production run should account for process variables.

NOTE These variables include, but are not limited to, machine equilibrium, breaks and shift changes, normal starts and stops, and material lot-to-lot differences.

5.4.5 Documented procedures and specifications for the forming, sealing and assembly operations shall be established and incorporated into the performance qualification.

5.4.6 Essential process variables shall be monitored and recorded.

5.4.7 The process shall be under control and capable of consistently producing product according to predetermined requirements.

5.5 Formal approval of the process validation

5.5.1 Review and formal approval of the process validation shall be carried out and documented as a final step in the validation program.

5.5.2 The documentation shall summarize and reference all protocols and results, and state conclusions regarding the validation status of the process.

5.6 Process control and monitoring

5.6.1 Procedures shall be established to ensure that the packaging process is under control and within the established parameters during routine operation.

5.6.2 Critical process parameters shall be routinely monitored and documented.

5.7 Process changes and revalidation

5.7.1 Documents concerning packaging and sealing process shall be covered by a change-control procedure for documenting, verifying and authorizing change.

5.7.2 Processes shall be revalidated if changes are made to the equipment, product, packaging materials or packaging process, which compromise the original validation and affect the sterility, safety or efficacy of sterile medical devices.

NOTE The following is a list of changes which could affect the status of a validated process:

- raw material changes that would impact the process parameters;
- a new piece of equipment is installed;
- transfer of processes and/or equipment from one facility or location to another;
- sterilization-process changes;
- negative trends in quality or process control indicators.

5.7.3 The need for revalidation shall be evaluated and documented. If the situation does not require that all aspects of the original validation be repeated, this revalidation does not have to be as extensive as the initial validation.

5.7.4 Periodic revalidation or reviews should be considered since multiple minor changes could cumulatively affect the validation status of the process.

6 Packaging system assembly

6.1 The sterile barrier system shall be assembled under appropriate environmental conditions to minimize the risk posed by contaminants to the medical device.

6.2 The package system assembly process shall follow controlled labelling and processing procedures to prevent mislabelling.

NOTE Additional guidance can be found in DIN 58953-7 and DIN 58953-8.

6.3 Package systems shall be assembled and filled according to the instructions based on a validated process that assures sterilization in a defined sterilization process. These instructions should include configuration of contents and organizing inserts, total weight, inner wrapping, and absorbent materials.

7 Use of reusable sterile barrier systems

In addition to the requirements listed in Clause 6, instructions and restrictions for use as specified in 5.1.10 and 5.1.11 of ISO 11607-1:2006 shall be followed (e.g. assembly, disassembly, maintenance, repair, storage).

NOTE For additional guidance on reusable containers, see EN 868-8, DIN 58953-9, and AAMI/ANSI ST33. For additional guidance on reusable fabrics, see EN 13795-1 and ANSI/AAMI ST65.

8 Sterile fluid-path packaging

8.1 Assembly of sterile fluid-path components and closures shall meet the requirements of Clauses 5 and 6.

8.2 Medical devices labelled “sterile fluid path” shall maintain sterility of the fluid path by the construction of the device in combination with its closures.

NOTE 1 The requirements for microbial barrier properties and sterile barrier system integrity are provided in ISO 11607-1. The requirements apply to the device itself.

NOTE 2 For the purpose of interpreting the requirements of this part of ISO 11607, the device and its closures constitute the sterile barrier system.

Annex A

(informative)

Process development

Process development, while not a formal part of process validation, should be considered as an integral part of forming and sealing. Process development or process design requires an assessment to identify and evaluate critical parameters, along with their operating ranges, settings and tolerances.

A process assessment is conducted to establish appropriate and necessary upper and lower processing limits, as well as the expected normal operating conditions. These process limits should be sufficiently removed from failure or marginal conditions. One technique could be the creation of seal-strength curves with accompanying visual examples of seal results that could aid in the selection of an optimal process window.

Potential failure modes and action levels having the greatest impact on the process should be identified and addressed (failure mode and effects analysis, cause and effect analysis).

Statistically valid techniques, such as screening experiments and statistically designed experiments to optimize the process, should be used.

Essential processing parameters that are evaluated may include, but are not limited to:

- temperature;
- pressure/vacuum, including rate of change;
- dwell time (line speed);
- energy levels/frequency (radio frequency/ultrasonic);
- torque limits for lid/cap closure systems.

The selected essential parameters will be selected such that they will produce a process that is in control, and capable of yielding sterile barrier systems and packaging systems that meet predetermined design specifications.

Bibliography

- [1] ISO 186:2002, *Paper and board — Sampling to determine average quality*
- [2] ^{A1} ISO 2859-1:1999 (including Corrigendum 1:2001 + Amendment 1:2011), *Sampling procedures for inspection by attributes — Part 1: Sampling schemes indexed by acceptance quality limit (AQL) for lot-by-lot inspection* ^{A1}
- [3] ^{A1} ISO 9001:2008 (including Corrigendum 1:2009), *Quality management systems — Requirements* ^{A1}
- [4] ISO/TS 11139:2006, *Sterilization of health care products — Vocabulary*
- [5] ^{A1} ISO 13485:2003 (including Corrigendum 1:2009), *Medical devices — Quality management systems — Requirements for regulatory purposes* ^{A1}
- [6] ^{A1} ISO 9000:2005, *Quality management systems — Fundamentals and vocabulary* ^{A1}
- [7] ^{A1} EN 868-5:2009, *Packaging for terminally sterilized medical devices — Part 5: Sealable pouches and reels of porous materials and plastic film construction — Requirements and test methods* ^{A1}
- [8] ^{A1} EN 868-6:2009, *Packaging for terminally sterilized medical devices — Part 6: Paper for low temperature sterilization processes — Requirements and test methods* ^{A1}
- [9] ^{A1} EN 868-8:2009, *Packaging for terminally sterilized medical devices — Part 8: Re-usable sterilization containers for steam sterilizers conforming to EN 285 — Requirements and test methods* ^{A1}
- [10] ^{A1} EN 13795-1+A1:2009, *Surgical drapes, gowns and clean air suits, used as medical devices, for patients, clinical staff and equipment — Part 1: General requirements for manufacturers, processors and products* ^{A1}
- [11] ^{A1} AAMI/ANSI ST 65:2008, *Processing of reusable surgical textiles for use in health care facilities* ^{A1}
- [12] ^{A1} DIN 58953-7:2010, *Sterilization — Sterile supply — Part 7: Use of sterilization paper, nonwoven wrapping material, textile materials, paper bags and sealable pouches and reels* ^{A1}
- [13] ^{A1} DIN 58953-8:2010, *Sterilization — Sterile supply — Part 8: Logistics of sterile medical devices* ^{A1}
- [14] ^{A1} DIN 58953-9:2010, *Sterilization — Sterile supply — Part 9: Use of sterilization container* ^{A1}
- [15] GHTF Study Group 3, *Process validation guidance for medical device manufacturers*
- [16] *International Vocabulary of Basic and General Terms in Metrology*. 1993, BIPM, IEC, IFCC, ISO, IUPAC, IUPAP, OIML

BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover.
Tel: +44 (0)20 8996 9000. Fax: +44 (0)20 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: +44 (0)20 8996 9001.
Fax: +44 (0)20 8996 7001. Email: orders@bsi-global.com. Standards are also available from the BSI website at <http://www.bsi-global.com>.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre.
Tel: +44 (0)20 8996 7111. Fax: +44 (0)20 8996 7048. Email: info@bsi-global.com.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration.
Tel: +44 (0)20 8996 7002. Fax: +44 (0)20 8996 7001.
Email: membership@bsi-global.com.

Information regarding online access to British Standards via British Standards Online can be found at <http://www.bsi-global.com/bsonline>.

Further information about BSI is available on the BSI website at <http://www.bsi-global.com>.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

Details and advice can be obtained from the Copyright & Licensing Manager.
Tel: +44 (0)20 8996 7070. Fax: +44 (0)20 8996 7553.
Email: copyright@bsi-global.com.