

BS EN ISO 13212:2014



BSI Standards Publication

**Ophthalmic optics —  
Contact lens care products —  
Guidelines for determination of  
shelf-life**

**bsi.**

...making excellence a habit.™

**National foreword**

This British Standard is the UK implementation of EN ISO 13212:2014. It supersedes BS EN ISO 13212:2011 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CH/172/9, Contact lenses and contact lens care products.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2014. Published by BSI Standards Limited 2014

ISBN 978 0 580 79722 4

ICS 11.040.70

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 September 2014.

**Amendments issued since publication**

Date	Text affected
------	---------------

---

English Version

## Ophthalmic optics - Contact lens care products - Guidelines for determination of shelf-life (ISO 13212:2014)

Optique ophtalmique - Produits d'entretien pour lentilles de contact - Lignes directrices pour la détermination de la durée de conservation (ISO 13212:2014)

Augenoptik - Kontaktlinsenpflegemittel - Leitfaden für die Bestimmung der Lagerdauer (ISO 13212:2014)

This European Standard was approved by CEN on 14 May 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

## Foreword

This document (EN ISO 13212:2014) has been prepared by Technical Committee ISO/TC 172 "Optics and photonics" in collaboration with Technical Committee CEN/TC 170 "Ophthalmic optics" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2015, and conflicting national standards shall be withdrawn at the latest by March 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 13212:2011.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

### Endorsement notice

The text of ISO 13212:2014 has been approved by CEN as EN ISO 13212:2014 without any modification.

# Contents

Page

<b>Foreword</b> .....	<b>iv</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>1</b>
<b>3 Terms and definitions</b> .....	<b>1</b>
<b>4 General requirements</b> .....	<b>1</b>
<b>5 Determination of finished product stability</b> .....	<b>1</b>
5.1 Objective.....	1
5.2 Study methods.....	2
5.3 Description of the product under study.....	3
5.4 Characteristics.....	3
5.5 Evaluation methods.....	4
5.6 Presentation of results.....	4
5.7 Discussion, interpretation, and conclusions.....	4
5.8 Ongoing stability.....	5
<b>Annex A (informative) Example of a stability testing plan for the finished contact lens care product</b> .....	<b>6</b>
<b>Bibliography</b> .....	<b>7</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see [www.iso.org/directives](http://www.iso.org/directives)).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see [www.iso.org/patents](http://www.iso.org/patents)).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: Foreword - Supplementary information

The committee responsible for this document is ISO/TC 172, *Optics and photonics*, Subcommittee SC 7, *Ophthalmic optics and instruments*.

This third edition cancels and replaces the second edition (ISO 13212:2011), which has been technically revised.

## Introduction

The purpose of stability tests of contact lens care products is to obtain sufficient information to enable the manufacturer to establish an appropriate shelf-life and identify any unique storage conditions required to appear on the labelling of the product.

The quality of a contact lens care product is determined by its content of active ingredient(s), its purity, and its physicochemical and microbiological properties. It is important to take into account the possible interaction of the container/closure with the contents.

The stability studies are intended to ascertain how the quality of a product varies as a function of time and under the influence of a variety of environmental factors.

On the basis of the information obtained, storage conditions are recommended, which will guarantee the maintenance of the quality of the product in relation to its safety, performance, and acceptability throughout the proposed shelf-life.

The design of the finished product stability studies for a care product is based on the knowledge obtained from studies on the active ingredient(s) and from the development studies.





# Ophthalmic optics — Contact lens care products — Guidelines for determination of shelf-life

## 1 Scope

This International Standard provides guidance on the design of stability studies for use in gathering information to enable determination of the shelf-life of contact lens care products.

This International Standard does not address studies designed to obtain information to establish the in-use stability (i.e. notice of discard date) of contact lens care products.

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 14729, *Ophthalmic optics — Contact lens care products — Microbiological requirements and test methods for products and regimens for hygienic management of contact lenses*

ISO 14730, *Ophthalmic optics — Contact lens care products — Antimicrobial preservative efficacy testing and guidance on determining discard date*

ISO 18369-1, *Ophthalmic optics — Contact lenses — Part 1: Vocabulary, classification system and recommendations for labelling specifications*

## 3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 18369-1 apply.

## 4 General requirements

**4.1** The specified shelf-life of the contact lens care product shall be based on the evaluation of the results of stability studies.

**4.2** Analytical methods that have been validated and are stability indicating shall be used to assay for active ingredients. Validation includes, but is not limited to, being able to differentiate between the active ingredient and its degradation products. The test methods used shall be described in full.

## 5 Determination of finished product stability

### 5.1 Objective

The objective of stability testing on contact lens care products is to provide data for determining the time period during which the product performance characteristics are maintained and to define appropriate storage conditions.

The design of the stability tests is based on the known properties of the active ingredient(s), the properties of the chosen formulation, and the recommendations for use of the product.

The relevant assay methods shall be determined prior to the start of the stability testing.

The specifications proposed from the time of manufacture to the end of the proposed shelf-life shall reflect, as far as possible, the results of the stability studies, particularly in relation to any parameters which could have a bearing on performance and safety and on product acceptability.

## 5.2 Study methods

### 5.2.1 General

Before starting stability studies, a suitable testing plan should be set up, taking into consideration the properties of the active ingredient(s) as well as the proposed mode of action of the care product.

### 5.2.2 Real-time studies

These studies should be carried out under a range of controlled test conditions, when applicable, which will enable the shelf-life and the storage requirements which are to appear on the product container label/package insert to be defined. This will normally include studies which are intended to allow the properties of the product at temperatures between 20 °C and 30 °C to be evaluated. However, 25 °C ± 2 °C should be used as the mean kinetic testing temperature.

Relative humidity should be controlled for all products that are not in hermetically sealed containers. In the case of tablets, high humidity conditions should be considered. In the case of aqueous liquids, low humidity conditions should be considered (see [Annex A](#)).

For each study, the mean temperature, the ranges of temperature, and mean humidity, if applicable, shall be stated in the stability report.

These studies are intended to support the initial shelf-life request and, for shelf-life extensions, any changes that could significantly impact the safety and performance of the product (e.g. certain changes in formulation, packaging materials, or manufacturing methods).

NOTE Real-time studies are performed in conjunction with accelerated ageing studies to establish an initial shelf-life.

### 5.2.3 Studies under varying storage conditions

These studies shall be carried out to provide important additional information. They can fulfil a number of objectives, such as

- supporting the initial shelf-life request, by complementing the limited results of the early real-time studies because decomposition, if it occurs, is likely to be accelerated,
- producing useful data at an early stage of development, demonstrating the effects of adverse storage in the packaging and product, and enabling storage conditions and suitable labelling to be provided, and
- supporting a request to extend the shelf-life.

The various test conditions should be stated. Depending on the nature and objectives of the stability study, the following points might need to be considered:

- a) various test temperatures: three or more, particularly if long-term real-time data are unavailable. In addition, the effect of low temperatures might need to be considered, such as below -15 °C (freezer), 2 °C to 8 °C (refrigerator) and freeze-thaw cycling;
- b) high humidity: relative humidity up to 75 % ± 5 % (see [Annex A](#)). Storage under high humidity conditions applies particularly to solid dosage forms. For products such as solutions, suspensions, etc. contained in packs designed to provide a permanent barrier to water loss, storage under high humidity is not necessary. However, low humidity can have an adverse effect on products packaged in semi-permeable containers;

- c) elevated temperature and humidity in combination: e.g. temperature of 40 °C associated with a relative humidity of up to 75 %, possibly with the effects of cycling between different temperatures and humidities;
- d) since most contact lens care products are water-based, relative humidity of 40 % or less (for example, 25 °C and 40 % relative humidity) should be considered (see [Annex A](#));
- e) light: either natural daylight or defined artificial illumination (see [Annex A](#)).

### 5.3 Description of the product under study

#### 5.3.1 Number and nature of the batches tested

The number of batches tested shall be stated with the batch number, details of the composition, date of manufacture, size of the batch and the name of the manufacturer of the active ingredient(s) used. The material used for all packaging that can impact stability of the product shall be stated and supplier identified.

Normally, three batches of the finished product are studied. If the number of batches tested is less than three, it shall be justified.

Satisfactory performance of the product in the smallest size of container with the highest surface/volume ratio shall allow the extension of shelf-life to containers which are up to eight times larger in volume.

#### 5.3.2 Primary container

The product batches shall be packed in the primary containers proposed for marketing. The smallest primary container size should be tested. Satisfactory performance in the smallest size shall allow marketing of the product in containers with eight times the volume present in the smallest size.

Details of the packaging should be stated, including

- a) type(s) of container and closure and nature of the constituent material(s),
- b) nature of any desiccant if used, and
- c) the complete range of sizes of the product proposed for marketing.

### 5.4 Characteristics

#### 5.4.1 General

The following characteristics should be studied:

- a) those in the finished product specification that are likely to be affected by storage, and
- b) those not monitored routinely at the time of manufacture, but which can be indicative of the stability/instability of the particular product (e.g. dissolution of tablets).

#### 5.4.2 Physical characteristics of the finished product

The following physical characteristics should be tested:

- a) physical properties specific to the product, such as tablet hardness and hygroscopicity, or pH, colour, clarity, and viscosity for solutions;
- b) important quality parameters such as the *in vitro* dissolution, moisture content (e.g. in relation to any desiccant used in the packaging), and particle size;

- c) any other physical characteristics of the product that shall be known in order to assess product stability.

#### **5.4.3 Microbial characteristics**

The following microbial characteristics shall be tested:

- a) antimicrobial activity of finished products to be marketed for chemical disinfection of contact lenses shall be tested in accordance with ISO 14729, unless otherwise justified;
- b) preservative efficacy of preserved products shall be tested at the end of shelf-life, in accordance with ISO 14730, unless otherwise justified;
- c) sterility of sterile products (or provide valid data to demonstrate maintenance of package integrity);
- d) microbial limits of non-sterile products shall be given.

#### **5.4.4 Chemical characteristics of the finished product**

The following chemical characteristics should be determined:

- a) assay of the active ingredient(s), where possible;
- b) consideration of other agents (such as antimicrobial preservatives and antioxidants);
- c) any other chemical characteristics that shall be known in order to assess the quality of the product.

#### **5.4.5 Characteristics of primary container interactions**

If necessary, carry out a study of the interaction of the container and closure with the contents in any case where this is a risk.

#### **5.4.6 Performance characteristics**

If stability cannot be established by chemical method, it should be followed by relevant performance characteristics. Performance tests should mimic as closely as possible the in-use condition; otherwise, the rationale for the design of test(s) should be described.

### **5.5 Evaluation methods**

The test procedures applied to the stability tests on the finished product shall be described in full and validated.

### **5.6 Presentation of results**

The results shall be summarized (e.g. as tables and graphs). For each product batch tested, the initial results (at the time of manufacture) and the results obtained during storage should be given. Results of real-time data should be recorded as they become available, up to the proposed shelf-life.

### **5.7 Discussion, interpretation, and conclusions**

The discussion and conclusion shall provide a critical evaluation of the suitability of the test methods used, the results obtained, and the proposed shelf-life specification. This should take into account the safety and performance requirements of the product at the end of shelf-life.

If it was necessary to carry out any further studies due to significant changes in relevant properties, an explanation should be given, together with the results of these studies.

A minimum of three months' real-time data at 25 °C should be available, supported by data from accelerated stability studies. Such data would not normally be expected to be suitable for prediction of

a shelf-life in excess of two years. Any extension of the shelf-life should be based on additional real-time study results.

Studies under accelerated test conditions will increase the decomposition and can permit some extrapolation of the room temperature shelf-life from that which would otherwise be acceptable. However, such studies would always need to be supplemented by long-term real-time studies, and normally at least three months' real-time data should be available.

If product batches in test demonstrate a decreasing stability profile, the shelf-life proposed and any overage should be based on the stability of the least stable test result, unless an explanation can be given.

The shelf-life (expiration date) shall be proposed on the primary package to be used for sale.

If there is evidence that batches of the stored product as packed for sale are stable at temperatures up to 30 °C, the product need bear no special temperature storage instructions. However, if there is evidence that the product shall be stored under defined conditions of storage, this shall be stated on the container label and the package insert (if included) and the outer carton. The maximum (or minimum) storage temperature should be stated in degrees Celsius (e. g. store below 25 °C; store in a refrigerator at 2 °C to 8 °C; do not refrigerate, store above 8 °C). These storage recommendations on the label/package insert shall reflect conditions found in the country or countries in which the product is to be placed on the market.

Temperatures acceptable for accelerated extrapolation of the expiration date should maintain the same mechanism of decomposition. Generally, temperatures at or below 45 °C will be acceptable.

For every 10 °C increase in temperature, the rate of decomposition generally increases by a factor of two. Unless otherwise justified on the basis of kinetic evidence, this acceleration factor should be used.

## **5.8 Ongoing stability**

Where data on routine production batches are not provided, ongoing stability studies should be carried out on at least two of the first production batches and the results recorded.

## Annex A (informative)

### Example of a stability testing plan for the finished contact lens care product

The programme given in [Table A.1](#) or [Table A.2](#) can be suitable for stability testing of a finished contact lens care product (see [Clause 5](#)). A suitable number of replicate determinations should be made at each point in time.

**Table A.1 — Testing plan for finished contact lens care products in solid dosage form**

	Initial	1 month	2 months	3 months	6 months	12 months	18 months	24 months	36 months	Every 12 months up to the end of shelf-life claimed for products
25 °Ca, 60 % RH <sup>b</sup>	x			x	x	x	x	x	x	x
30 °Ca, 65 % RH <sup>b</sup>	x	x	x	x	x					
40 °Ca, 75 % RH <sup>b</sup>	x	x	x	x	x					
25 °Ca, light <sup>c</sup>	x			x	x					

<sup>a</sup> Controlled temperature  $\pm 2$  °C; humidity conditions monitored and reported.  
<sup>b</sup> Controlled relative humidity (RH)  $\pm 5$  %, where applicable.  
<sup>c</sup> Controlled and stated light conditions and duration of exposure, where applicable. ICH Q1B recommends not less than 1,2 million lux hours and integrated of not less than 200 W hours per square metre.

**Table A.2 — Testing plan for finished contact lens care products in liquid dosage form**

	Initial	1 month	2 months	3 months	6 months	12 months	18 months	24 months	36 months	Every 12 months up to the end of shelf-life claimed for products
25 °Ca, 40 % RH <sup>b</sup>	x			x	x	x	x	x	x	x
30 °Ca 35 % RH <sup>b</sup>	x	x	x	x	x					
40 °Ca, not more than 25 % RH <sup>b</sup>	x	x	x	x	x					
25 °Ca, light <sup>c</sup>	x			x	x					

<sup>a</sup> Controlled temperature  $\pm 2$  °C; humidity conditions monitored and reported.  
<sup>b</sup> Controlled relative humidity (RH)  $\pm 5$  %. A correction factor can be applied under ICH Q1A R2.  
<sup>c</sup> Controlled and stated light conditions and duration of exposure, where applicable. ICH Q1B recommends not less than 1,2 million lux hours and integrated of not less than 200 W hours per square metre.

## Bibliography

- [1] ISO 14534, *Ophthalmic optics — Contact lenses and contact lens care products — Fundamental requirements*
- [2] Stability Testing of New Drug Substances and Products, International Committee on Harmonization, Tripartite Guideline Q1A(R2)
- [3] Stability Testing; Photostability Testing of New Drug Substances and Products, International Committee on Harmonization, Tripartite Guideline Q1B







# British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

## About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

## Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at [bsigroup.com/standards](http://bsigroup.com/standards) or contacting our Customer Services team or Knowledge Centre.

## Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at [bsigroup.com/shop](http://bsigroup.com/shop), where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

## Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to [bsigroup.com/subscriptions](http://bsigroup.com/subscriptions).

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

**PLUS** is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit [bsigroup.com/shop](http://bsigroup.com/shop).

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email [bsmusales@bsigroup.com](mailto:bsmusales@bsigroup.com).

## BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

## Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

## Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

## Useful Contacts:

### Customer Services

**Tel:** +44 845 086 9001

**Email (orders):** [orders@bsigroup.com](mailto:orders@bsigroup.com)

**Email (enquiries):** [cservices@bsigroup.com](mailto:cservices@bsigroup.com)

### Subscriptions

**Tel:** +44 845 086 9001

**Email:** [subscriptions@bsigroup.com](mailto:subscriptions@bsigroup.com)

### Knowledge Centre

**Tel:** +44 20 8996 7004

**Email:** [knowledgecentre@bsigroup.com](mailto:knowledgecentre@bsigroup.com)

### Copyright & Licensing

**Tel:** +44 20 8996 7070

**Email:** [copyright@bsigroup.com](mailto:copyright@bsigroup.com)



...making excellence a habit.™