

BS EN ISO 16635-2:2014



BSI Standards Publication

Dentistry — Dental rubber dam instruments

Part 2: Clamp forceps

bsi.

...making excellence a habit.™

National foreword

This British Standard is the UK implementation of EN ISO 16635-2:2014.

The UK participation in its preparation was entrusted to Technical Committee CH/106/4, Dental Instruments and Equipment.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2014. Published by BSI Standards Limited 2014

ISBN 978 0 580 75169 1

ICS 11.060.20

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 September 2014.

Amendments issued since publication

Date	Text affected
------	---------------

ICS 11.060.20

English Version

Dentistry - Dental rubber dam instruments - Part 2: Clamp
forceps(ISO 16635-2:2014)

Médecine bucco-dentaire - Instruments de digue dentaire
en caoutchouc - Partie 2: Pincés porte-crampons (ISO
16635-2:2014)

Zahnheilkunde - Kofferdaminstrumente - Teil 2:
Klammerzangen (ISO 16635-2:2014)

This European Standard was approved by CEN on 28 June 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Foreword

This document (EN ISO 16635-2:2014) has been prepared by Technical Committee ISO/TC 106 “Dentistry” in collaboration with Technical Committee CEN/TC 55 “Dentistry” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2015, and conflicting national standards shall be withdrawn at the latest by March 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 16635-2:2014 has been approved by CEN as EN ISO 16635-2:2014 without any modification.

Contents

Page

Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements	2
4.1 General.....	2
4.2 Total length.....	2
4.3 Locking part.....	2
4.4 Distance between the forceps handles in the closed, passive state.....	2
4.5 Retaining pins.....	2
4.6 Spring and ability to return to the original state.....	3
4.7 Materials.....	3
4.8 Surface finish.....	3
4.9 Corrosion resistance and resistance to reprocessing.....	3
5 Test methods	3
5.1 Visual examination.....	3
5.2 Dimensions.....	3
5.3 Reprocessing.....	3
6 Marking	4

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 106, *Dentistry*, Subcommittee SC 4, *Dental instruments*.

ISO 16635 consists of the following parts, under the general title *Dentistry — Dental rubber dam instruments*:

- *Part 1: Hole punch*
- *Part 2: Clamp forceps*

Introduction

In order to support the increasing use of dental rubber dam the application of dental rubber dam should be supported by standardization of the required instruments and materials.

In dental practice and when used as intended, clamp forceps for dental rubber dam clamps come into contact with the patient.

Dentistry — Dental rubber dam instruments —

Part 2: Clamp forceps

1 Scope

This part of ISO 16635 specifies requirements and test methods for clamp forceps intended for the application of dental rubber dam clamps to teeth.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 6507-1, *Metallic materials — Vickers hardness test — Part 1: Test method*

ISO 6508-1, *Metallic materials — Rockwell hardness test — Part 1: Test method*

ISO 16635-1, *Dentistry — Dental rubber dam technique — Part 1: Hole punch*

ISO 15510:2010, *Stainless steels — Chemical composition*

ISO 17664, *Sterilization of medical devices — Information to be provided by the device manufacturer for the processing of reusable medical devices*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942, ISO 16635-1 and the following apply.

3.1

dental rubber dam

rubber dam

dental dam

sheet of elastomeric film composed of natural rubber latex or synthetic material that is used to isolate one or several teeth from the rest of the oral cavity during a dental restorative procedure

3.2

locking part

device located around the parts of the forceps handles near the joint, which locks automatically when the clamp forceps opens, thus fixating the forceps after it has picked up the dental dam clamp

3.3

clamp receiver

working end of the clamp forceps which serves for the picking up and bringing into position of the dental rubber dam clamp and is linked with the forceps joint by means of a connecting piece

Note 1 to entry: The working ends move diametrically opposed to the forceps handles as a result of the handles being guided in the joint.

3.4 retaining pins

pin-like ends of the clamp forceps receiver which are inserted into the holes of the clamp jaws of the dental rubber dam clamps

4 Requirements

4.1 General

In accordance with its intended use, a clamp forceps shall be designed such that it can be used to apply the dental rubber dam clamp across the row of teeth without the forceps coming into contact with the teeth.

This requires that the working end with the retaining pins is angled against the handle and/or stepped.

The retaining pins shall be provided with a depth-control stop to ensure that they can only be inserted to a certain depth into the hole in the clamp and that the clamp can be safely orientated in three spatial directions while the forceps is applied to the tooth.

4.2 Total length

The total length of the clamp forceps shall be ≤ 175 mm.

Test in accordance with [5.2](#).

4.3 Locking part

The clamp forceps shall be provided with a mechanism which locks and unlocks when the forceps is used, so that the clamp forceps holds the clamps in the tensioned state and releases them again following their application to the tooth.

Test in accordance with [5.1](#).

4.4 Distance between the forceps handles in the closed, passive state

The distance between the forceps handles in the closed, passive state shall not exceed 95 mm.

Test in accordance with [5.2](#).

4.5 Retaining pins

4.5.1 Length of retaining pins

The length of the retaining pins for the depth control stop shall be $(3,5 \pm 0,5)$ mm.

Test in accordance with [5.2](#).

4.5.2 Diameter of retaining pins

The diameter of the retaining pins shall be $(1,5 \pm 0,2)$ mm.

Test in accordance with [5.2](#).

4.5.3 Distance between the retaining pins

The distance between the retaining pins, when the forceps is opened to the maximum degree, shall be at least 20 mm when measured at the outside.

Test in accordance with [5.2](#).

4.6 Spring and ability to return to the original state

The clamp forceps shall be returned to a closed position by means of a spring.

Test in accordance with [5.1](#).

4.7 Materials

Clamp forceps shall be made of martensitic hardening corrosion-resistant stainless steel having the material numbers 4021-420-00-I (name of the steel designation: X20Cr13; S42020), or 4028-420-00-I (X30Cr13; S42030), 4034-420-00-I (X46Cr13, S42040) with a hardness of 42 HRC to 55 HRC or a Vickers hardness of 500 HV1 to 700 HV1, or of austenitic corrosion-resistant steel with the material number 4301-304-00-I (X5CrNi18-10; S30408) in accordance with ISO 15510:2010.

Same parts can be made of austenitic corrosion resistant steel having the material numbers 4301-304-00-I (X5CrNi18-10) or 4310-301-00-I (X10CrNi18-8; S30110) according to ISO 15510:2010.

Test the Rockwell hardness in accordance with ISO 6508-1, scale C, or test the Vickers hardness in accordance with ISO 6507-1.

4.8 Surface finish

The manufacturer may choose between a brightened and a matted surface.

All surfaces of the clamp forceps shall be free of pores, cracks and all residues, including abrasives and/or polishing agents.

Test in accordance with [5.1](#).

4.9 Corrosion resistance and resistance to reprocessing

The clamp forceps shall not show any signs of corrosion.

The clamp forceps shall withstand 100 reprocessing cycles, as defined by the manufacturer's instructions in accordance with ISO 17664, without deterioration or showing signs of corrosion.

Test in accordance with [5.3](#).

NOTE Discolorations due to water spots are not regarded signs of corrosion.

5 Test methods

5.1 Visual examination

Perform visual examination with normal visual acuity without any magnification.

5.2 Dimensions

The dimensions are tested by means of appropriate length measuring devices (e.g. micrometers, callipers) with a maximum permissible error of $\leq 0,1$ mm.

5.3 Reprocessing

Carry out 100 reprocessing cycles in accordance with the manufacturer's recommended methods of cleaning, disinfection and sterilization. Visually inspect the forceps for signs of corrosion or surface deterioration.

6 Marking

Clamp forceps shall be marked as follows:

- a) name of the manufacturer and/or brand;
- b) model number (reference number);
- c) lot number (batch designation).

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email bsmusales@bsigroup.com.

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

Useful Contacts:

Customer Services

Tel: +44 845 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 845 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com



...making excellence a habit.™