

Dentistry — Powered toothbrushes — General requirements and test methods

The European Standard EN ISO 20127:2005 has the status of a
British Standard

ICS 97.170

National foreword

This British Standard is the official English language version of EN ISO 20127:2005. It is identical with ISO 20127:2005.

The UK participation in its preparation was entrusted by Technical Committee CH/106, Dentistry, to Subcommittee CH/106/7, Oral hygiene products, which has the responsibility to:

- aid enquirers to understand the text;
- present to the responsible international/European committee any enquiries on the interpretation, or proposals for change, and keep UK interests informed;
- monitor related international and European developments and promulgate them in the UK.

A list of organizations represented on this subcommittee can be obtained on request to its secretary.

Cross-references

The British Standards which implement international or European publications referred to in this document may be found in the *BSI Catalogue* under the section entitled “International Standards Correspondence Index”, or by using the “Search” facility of the *BSI Electronic Catalogue* or of British Standards Online.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

Summary of pages

This document comprises a front cover, an inside front cover, the EN ISO title page, the EN ISO foreword page, the ISO title page, pages ii to iv, pages 1 to 6, an inside back cover and a back cover.

The BSI copyright notice displayed in this document indicates when the document was last issued.

Amendments issued since publication

Amd. No.	Date	Comments

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 2 November 2005

© BSI 2 November 2005

ICS 97.170

English version

**Dentistry - Powered toothbrushes - General requirements and
test methods (ISO 20127:2005)**

Art dentaire - Broses à dents électriques - Exigences
générales et méthodes d'essai (ISO 20127:2005)

Zahnheilkunde - Physikalische Eigenschaften von
elektrischen Zahnbürsten (ISO 20127:2005)

This European Standard was approved by CEN on 14 February 2005.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

EN ISO 20127:2005

Foreword

This document (EN ISO 20127:2005) has been prepared by Technical Committee ISO/TC 106 "Dentistry" in collaboration with Technical Committee CEN/TC 55 "Dentistry", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2005, and conflicting national standards shall be withdrawn at the latest by September 2005.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

Endorsement notice

The text of ISO 20127:2005 has been approved by CEN as EN ISO 20127:2005 without any modifications.

INTERNATIONAL
STANDARD

ISO
20127

First edition
2005-03-15

**Dentistry — Powered toothbrushes —
General requirements and test methods**

*Art dentaire — Broses à dents électriques — Exigences générales et
méthodes d'essai*



Reference number
ISO 20127:2005(E)

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 20127 was prepared by Technical Committee ISO/TC 106, *Dentistry*, Subcommittee SC 7, *Oral hygiene products*.

Introduction

Powered toothbrushes are used for the removal of dental plaque and oral debris in order to facilitate oral hygiene.

Dentistry — Powered toothbrushes — General requirements and test methods

1 Scope

This International Standard specifies requirements and test methods for the physical properties of powered toothbrushes in order to promote the safety of these products for their intended use.

Specifically excluded are other types of powered oral hygiene devices (such as powered interdental brushes) and manual toothbrushes.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 1942, *Dentistry — Vocabulary*

ISO 3696:1987, *Water for analytical laboratory use — Specification and test methods*

IEC 60335-1:2004, *Household and similar electrical appliances — Safety — Part 1: General requirements*

IEC 60335-2-52, *Household and similar electrical appliances — Safety — Part 2-52: Particular requirements for oral hygiene appliances*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 1942 and the following apply.

3.1

powered toothbrush

hand-held electrically-powered appliance, the brush head of which carries filaments, used primarily for cleaning surfaces within the oral cavity

3.2

brush head

fixed or removable working end of an electrically-powered toothbrush to which the filaments are attached

3.3

filament

single strand within the brush head

3.4

tuft

group of filaments gathered together and attached to the brush head

3.5

tuft removal force

force required to remove one tuft from the brush head

4 Requirements

4.1 Pass-fail criteria

Eight electrically-powered toothbrushes shall be tested. If none of the eight sampled powered toothbrushes fail, the powered toothbrush passes. If one toothbrush does not meet the minimum requirement, another eight powered toothbrushes shall be tested. If no more toothbrushes fail, the powered toothbrush passes. If a total of two or more powered toothbrushes of the sixteen fail, the powered toothbrush fails.

4.2 Physical inspection

The powered toothbrush, its components and all accessories shall be intact and free of visible contamination and sharp or rough surfaces when examined according to 5.3.

4.3 Electrical safety

All the powered toothbrushes and related accessories shall conform to the requirements described in IEC 60335-1 and IEC 60335-2-52.

4.4 Tuft retention

The tuft removal force shall not be less than 15 N when tested according to 5.4.

4.5 Mechanical strength

All powered toothbrushes and related accessories shall conform to the requirements described in IEC 60335-1:2004, Clause 21 and IEC 60335-2-52.

4.6 Resistance to chemical challenge

After being subjected to a chemical challenge test according to 5.5, the brush head shall conform to the requirements of the spring operated impact test as described in IEC 60335-1:2004, Clause 21, and the brush head shall be operational and free of sharp or rough surfaces when examined.

5 Test methods

5.1 Sampling

The powered toothbrushes obtained for testing shall be as manufactured and not modified in any way except as specified for this International Standard. Eight powered toothbrushes of each type shall be tested.

5.2 General test conditions

All tests shall be conducted using dry brushes at (23 ± 5) °C and relative humidity (50 ± 10) %.

5.3 Physical inspection

Inspection of the powered toothbrush and related accessories shall be performed using normal visual acuity without magnification. Tactile inspection shall also be performed to detect sharp or rough surfaces.

5.4 Tuft retention

5.4.1 Apparatus

5.4.1.1 Gripping unit, to secure the brush head, having a structure such that compressive force is not induced on the brush head (Figure A.1).

5.4.1.2 Clamp for securely holding all of the filaments in one tuft, for example a tuft gripping clamp (see Figure A.2). A collet chuck can also be used.

5.4.1.3 Apparatus for applying, measuring and indicating the removal force, a digital force gauge or a universal testing machine (force range from 5 N to 50 N, accurate to 0,1 N and a range of pulling speed from 20 mm/min to 100 mm/min).

5.4.2 Procedure

Place the brush in the gripping unit (5.4.1.1) and lock into place so that the clamp (5.4.1.2) pulls the tuft along the long axis of the tuft without any twisting. The brush head shall not be compressed during and after the placement.

Place the clamp on the filament tuft. Be sure to clamp all of the filaments from one tuft only; do not include filaments from the surrounding tufts. The filaments from one tuft should be secured at approximately the midpoint of the tuft length. Record the force required to pull out the tuft using the testing apparatus (5.4.1.3). Test two tufts of each tuft type (if available). The tufts should be non-adjacent.

5.5 Resistance to chemical challenge

5.5.1 Apparatus and chemicals

5.5.1.1 Apparatus for mixing the chemical challenge, for example stirring bar or mixer.

5.5.1.2 Container which can be sealed and is chemically inert, for example a polypropylene bottle.

5.5.1.3 Chemicals, of minimum purity and of the amounts listed in the Table 1.

Table 1 — Quantity and purity of products used

Chemicals	Minimum purity	Amount g
Ethanol	98,5 %	100
L-carvone	98,0 %	1,5
L-menthol	98,0 %	1,5
Sodium lauryl sulfate	95,0 %	15,0
Glycerine	98,0 %	200
Water	ISO 3696:1987, Grade 3	682

5.5.2 Procedure

Add 1,5 g L-carvone, 1,5 g L-menthol, 15 g sodium lauryl sulphate to 100 g ethanol in the container and stir well. Add 250 g water (ISO 3696:1987, Grade 3) and stir well to obtain a clear solution. Add 200 g glycerine and 432 g water (ISO 3696:1987, Grade 3) and stir to obtain a clear solution.

Place the removable brush head in the chemical-challenge solution. Ensure that the portion of the head containing filaments and at least 80 % of the total length of the removable head is immersed. If the head is not

removable, immerse the portion of the head containing filaments and an additional 80 % of the portion of the head exposed to a chemical challenge during normal toothbrushing. After immersion seal container. After 24 h without agitation, remove the toothbrush, rinse with water (ISO 3696:1987, Grade 3) and shake off excess water.

Perform the spring-operated impact test as described in IEC 60335-1:2004, Clause 21.

6 Test report

The test report shall contain the following information:

- a) identification of the powered toothbrush;
- b) results and the units of measure used;
- c) any unusual features noted during the test;
- d) reference to this International Standard (ISO 20127:2004);
- e) statement of conformance with this International Standard.

7 Marking and labelling

7.1 Powered toothbrush components

The powered toothbrush, all components and related accessories shall be marked in accordance with IEC 60335-1 and IEC 60335-2-52. The powered toothbrush shall also be marked with a manufacturer's tracking code.

7.2 Instructions for use

The product shall be supplied with instructions for use, maintenance and disposal and information on electrical safety (including warnings).

7.3 Marking and labelling

The packaging of powered toothbrushes shall be marked or labelled with the following information:


- a) name and address of manufacturer and/or responsible distributor;
- b) trade name;
- c) list of contents;
- d) tracking code.

8 Packaging

The packaging shall be such that it will neither contaminate nor permit contamination of the powered toothbrush when examined in accordance with 5.3.


Annex A
(informative)

Gripping devices



- Key**
1 brush head
2 brush-head gripping device

Figure A.1 — Brush-head gripping unit



Key

- 1 tuft
- 2 tuft gripping clamp

Figure A.2 — Tuft gripping clamp

BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: +44 (0)20 8996 9000. Fax: +44 (0)20 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: +44 (0)20 8996 9001. Fax: +44 (0)20 8996 7001. Email: orders@bsi-global.com. Standards are also available from the BSI website at <http://www.bsi-global.com>.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre. Tel: +44 (0)20 8996 7111. Fax: +44 (0)20 8996 7048. Email: info@bsi-global.com.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: +44 (0)20 8996 7002. Fax: +44 (0)20 8996 7001. Email: membership@bsi-global.com.

Information regarding online access to British Standards via British Standards Online can be found at <http://www.bsi-global.com/bsonline>.

Further information about BSI is available on the BSI website at <http://www.bsi-global.com>.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

Details and advice can be obtained from the Copyright & Licensing Manager. Tel: +44 (0)20 8996 7070. Fax: +44 (0)20 8996 7553. Email: copyright@bsi-global.com.