

Advanced technical ceramics — Methods of test for ceramic powders

Part 1. Determination of impurities in alumina

The European Standard EN 725-1 : 1997 has the status of a
British Standard

ICS 81.060.99

Committees responsible for this British Standard

The preparation of this British Standard was entrusted to Technical Committee RPI/13, Advanced technical ceramics, upon which the following bodies were represented:

AEA Technology
Aluminium Federation
British Ceramic Research Ltd.
British Industrial Ceramic Manufacturers' Association
Department of Trade and Industry (National Physical Laboratory)
Flat Glass Manufacturers' Association
GAMBICA (BEAMA Ltd.)
Institute of Refractories Engineers
Ministry of Defence
Refractories Association of Great Britain
Society of British Aerospace Companies Ltd.
University of Manchester

This British Standard, having been prepared under the direction of the Sector Board for Materials and Chemicals, was published under the authority of the Standards Board and comes into effect on 15 July 1997

© BSI 1997

Amendments issued since publication

Amd. No.	Date	Text affected

The following BSI references relate to the work on this standard:
Committee reference RPI/13
Draft for comment 93/311578 DC

ISBN 0 580 27918 9

Contents

	Page
Committees responsible	Inside front cover
National foreword	ii
Foreword	2
Text of EN 725-1	3

National foreword

This British Standard has been prepared by Technical Committee RPI/13 and is the English language version of EN 725-1 : 1997 *Advanced technical ceramics — Methods of test for ceramic powders — Part 1: Determination of impurities in alumina*, published by the European Committee for Standardization (CEN).

EN 725-1 : 1997 was produced as a result of international discussions in which the United Kingdom took an active part.

EN 725-1 : 1997 has been approved by CEN member bodies under the weighted voting procedures introduced in 1988 to coincide with the introduction of 'New Approach' Directives from the Commission of the European Community.

Cross-references

Publication referred to	Corresponding British Standard
ISO 3696	BS 3978 : 1995 <i>Water for analytical laboratory use — Specification and test methods</i>
ISO 5725	BS 5497 <i>Precision of test methods</i> Part 1 : 1987 <i>Guide for the determination of repeatability and reproducibility for a standard test method by inter-laboratory tests</i>

NOTE. International and European Standards as well as overseas standards, are available from Customer Services, BSI, 389 Chiswick High Road, London, W4 4AL.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

Summary of pages

This document comprises a front cover, an inside front cover, pages i and ii, the EN title page, pages 2 to 6, an inside back cover and a back cover.

ICS 81.060.99

Descriptors: Ceramics, powdery materials, impurities, aluminium oxide, chemical analysis, determination of content, sodium oxides, potassium oxides, iron oxides, magnesium oxides, calcium oxides, silicon oxides, atomic absorption spectrophotometry

English version

Advanced technical ceramics — Methods of test for ceramic powders — Part 1: Determination of impurities in alumina

Céramiques techniques avancées — Méthodes d'essai pour poudres céramiques — Partie 1: Détermination des impuretés dans l'alumine

Hochleistungskeramik — Prüfverfahren für keramische Pulver — Teil 1: Bestimmung von Verunreinigungen in Aluminiumoxidpulver

This European Standard was approved by CEN on 1997-02-24. CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Central Secretariat: rue de Stassart 36, B-1050 Brussels

Foreword

This European Standard has been prepared by Technical Committee CEN/TC 184, Advanced technical ceramics, the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 1997, and conflicting national standards shall be withdrawn at the latest by September 1997.

EN 725 consists of 11 parts:

Part 1 : *Determination of impurities in alumina*

Part 2 : *Determination of impurities in barium titanate (ENV)*

Part 3 : *Determination of oxygen content of non-oxides by thermal extraction*

Part 4 : *Determination of oxygen content of non-oxides by XRF analysis (ENV)*

Part 5 : *Determination of particle size distribution*

Part 6 : *Determination of specific area*

Part 7 : *Determination of absolute density*

Part 8 : *Determination of tapped bulk density*

Part 9 : *Determination of untamped bulk density*

Part 10 : *Determination of compaction properties*

Part 11 : *Determination of reactivity on sintering (ENV)*

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

Contents

	Page
Foreword	2
1 Scope	3
2 Normative references	3
3 Principle	3
4 Reagents	3
5 Apparatus	4
6 Test sample	4
7 Decomposition of the test sample	4
8 Calibration graph	5
9 Adjustment of the apparatus	5
10 Measurements	6
11 Expression of the results	6
12 Accuracy	6
13 Test report	6

1 Scope

This part of EN 725 specifies methods for the determination of elements of sodium, potassium, iron, silicon, calcium and magnesium present as impurities in alumina using atomic absorption (AAS) or inductively coupled plasma (ICP) instruments. For each element present as impurities, the methods are applicable to the following ranges, calculated as oxides:

Sodium oxide	20 ppm to 6000 ppm
Potassium oxide	20 ppm to 100 ppm
Ferric oxide	20 ppm to 300 ppm
Silica	50 ppm to 2000 ppm
Calcium oxide	20 ppm to 700 ppm
Magnesium oxide	5 ppm to 1000 ppm

2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and in the publications listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

ECSC/CI 9	<i>Chemical analysis of ferrous materials — Operational guidelines for the application of flame atomic absorption spectrometry in standard methods for the chemical analysis of iron and steel</i>
ISO 3696	<i>Water for analytical laboratory use — Specification and test methods</i>
ISO 5725	<i>Precision of test methods — Determination of repeatability and reproducibility for a standard test method by inter-laboratory tests</i>
ISO/DIS 13527	<i>Chemical analysis of ferrous materials — Guidelines on the use of inductively coupled plasma atomic emission spectroscopy</i>

3 Principle

The test sample is decomposed by using either a fusion method or an acid dissolution method. The acid dissolution method cannot be used for the determination of silicon. The solution is transferred to a volumetric flask and diluted to a known volume, and the elements are determined by AAS or ICP (see clause 1).

4 Reagents

4.1 General

During the analysis, use only reagents and calibration solutions of at least 99,99 % purity and only distilled water or water of equivalent purity (see ISO 3696).

4.2 Reagents for fusion

4.2.1 *Lithium carbonate*, Li_2CO_3

4.2.2 *Potassium carbonate*, K_2CO_3

4.2.3 *Boric acid*, H_3BO_3

4.2.4 *Sulfuric acid*, H_2SO_4 , ($\rho_{20} = 1,84$ g/ml)

4.2.5 *Lithium metaborate*, LiBO_2

4.2.6 *Nitric acid*, HNO_3 , ($\rho_{20} = 1,33$ g/ml)

4.2.7 *Phosphoric acid*, H_3PO_4 , ($\rho_{20} = 1,78$ g/ml)

4.3 Reagents for acid dissolution

4.3.1 *Sulfuric acid-phosphoric acid mixture (A)*

Pour 500 ml of phosphoric acid ($\rho_{20} = 1,78$ g/ml) into 500 ml of sulfuric acid ($\rho_{20} = 1,84$ g/ml)

4.3.2 *Sulfuric acid-phosphoric acid mixture (B)*

Pour 700 ml of phosphoric acid ($\rho_{20} = 1,78$ g/ml) into 300 ml of sulfuric acid ($\rho_{20} = 1,84$ g/ml).

4.4 Reagents for calibration

4.4.1 *Pure alumina*, of very low and known impurity levels.

4.4.2 *Sodium*, commercial solution or solution obtained by dissolution of pure chemical compound, concentration 1 g/l.

4.4.3 *Potassium*, commercial solution or solution obtained by dissolution of pure chemical compound, concentration 1 g/l.

4.4.4 *Iron (ferric)*, commercial solution or solution obtained by dissolution of pure chemical compound, concentration 1 g/l.

4.4.5 *Silicon*, commercial solution or solution obtained by dissolution of pure chemical compound, concentration 1 g/l.

4.4.6 *Calcium*, commercial solution or solution obtained by dissolution of pure chemical compound, concentration 1 g/l.

4.4.7 *Magnesium*, commercial solution or solution obtained by dissolution of pure chemical compound, concentration 1 g/l.

5 Apparatus

- 5.1** *Platinum crucible*, with a capacity of at least 50 ml.
- 5.2** *Gold-platinum crucible*, with a capacity of at least 50 ml.
- 5.3** *Vitreous carbon crucible*, with a capacity of at least 50 ml with lid and heating device.
- 5.4** *Muffle furnace*, suitable for operation in the range of 1000 °C to 1200 °C.
- 5.5** *Magnetic stirrer*, with heating.
- 5.6** *Atomic absorption spectrometer and/or inductively coupled plasma spectrometer*, in accordance with ECSC/CI 9 or ISO 13527 respectively.
- 5.7** *Laboratory glassware*.

6 Test sample

Use samples of approximately :

- 2 g for decomposition by fusion;
- 1 g for decomposition by acid dissolution.

Weigh them to 0,0005 g.

7 Decomposition of the test sample

7.1 General

Dissolve either by a fusion method (see 7.2 to 7.4) or an acid dissolution method (see 7.5 and 7.6)

7.2 Fusion 1

In a platinum crucible (see 5.1) weigh 1,5 g Li_2CO_3 (see 4.2.1) 5 g K_2CO_3 (see 4.2.2) and 2,5 g H_3BO_3 (see 4.3.3). Add the test sample of approximately 2 g (see clause 6) and mix intimately using a platinum spatula.

Place the crucible and contents into the muffle furnace (see 5.4), maintained at 1050 °C \pm 50 °C, for 30 min. Remove the crucible from the furnace and swirl the contents on the sides of the crucible, then allow to cool to room temperature.

Dissolve the fused product in a 400 ml beaker which contains 100 ml of water and 10 ml of sulfuric acid (see 4.2.4). Place the beaker, covered with a watch glass, on a hot plate and heat to boiling. Maintain at boiling point to obtain a complete dissolution. Remove the beaker from the hot plate. Allow to cool.

Transfer quantitatively the solution into a 200 ml volumetric flask. This procedure allows for a concentration of alumina up to 8 g/l but if needed, a dilution to a higher volume is possible. Allow to cool to room temperature and make up to the mark.

7.3 Fusion 2

In a platinum crucible (see 5.1) weigh 4 g of LiBO_2 (see 4.2.5) and 1 g of test sample. Mix intimately using a platinum spatula.

Place the crucible and contents into the muffle furnace (see 5.4) maintained at 1150 °C \pm 50 °C for 30 min (after the first 15 min, swirl the contents of the crucible for a few seconds). Remove the crucible from the

furnace and dip its base in water at ambient temperature (this procedure allows easy removal of the bead from the crucible). To prevent sticking of melt in the crucible, either use a new crucible or, with an old one, immerse it in the solution. Place the bead into a 400 ml beaker which contains 80 ml of water and 20 ml of nitric acid (see 4.2.6).

Place the beaker, covered with a watch glass on a magnetic stirrer with heating stirrer and maintain the agitation at approximately 80 °C \pm 10 °C until complete dissolution. Remove the beaker from the stirrer and allow to cool. Transfer quantitatively the solution into a 200 ml volumetric flask. This procedure allows for a concentration of alumina up to 8 g/l but if needed, a dilution to a higher volume is possible. Allow to cool to room temperature and make up to the mark.

7.4 Fusion 3

In a gold-platinum crucible (see 5.2), weigh 2,2 g Li_2CO_3 (see 4.2.1) and 5,5 g H_3BO_3 (see 4.2.3). Add the test sample (see clause 6). Mix intimately using a platinum spatula.

Place the crucible and contents into the muffle furnace (see 5.4), maintained at 1100 °C \pm 50 °C for 30 min. Remove the crucible from the furnace and swirl the contents on the sides of the crucible, then allow to cool to room temperature. Dissolve the fused product in a 400 ml beaker which contains 100 ml of water and 20 ml of sulfuric acid (see 4.2.4).

Place the beaker, covered with a watch glass, on a hot plate and heat to boiling. Maintain at boiling point to obtain complete dissolution. Remove the beaker from the hot plate. Allow to cool. Transfer quantitatively the solution into a 200 ml volumetric flask. This procedure allows for a concentration of alumina up to 8 g/l, but if needed a dilution to a higher volume is possible. Allow to cool to room temperature and make up to the mark.

7.5 Acid dissolution 1

Weigh the test sample (see clause 6) into a vitreous carbon crucible (see 5.3). Add carefully 15 ml of sulfuric acid-phosphoric acid mixture (see 4.3.1). Cover with a lid. Put the crucible with the lid into the heating device and maintain at boiling for 20 min. Remove the crucible from the heating device and allow to cool to room temperature.

Transfer quantitatively the contents into a 100 ml volumetric flask which contains 30 ml of water. Rinse the crucible and the lid with distilled water into the flask and after cooling, make up to the mark with water.

7.6 Acid dissolution 2

Weigh the test sample (see clause 6) into a gold-platinum crucible (see 5.2). Add carefully 12 ml of sulfuric acid-phosphoric mixture (see 4.3.2) and cover with a lid. Put the crucible with the lid on to the hot plate (see 5.5) and maintain at boiling for 12 min. Remove the crucible from the heating device and allow to cool.

Transfer quantitatively the contents into a 100 ml volumetric flask which contains 30 ml of water. Rinse the crucible and the lid with distilled water into the flask and after cooling, make up to the mark with distilled water.

8 Calibration graph

8.1 General

The optimum calibration graph is obtained using calibration solutions whose concentrations are compatible both with the analytical method (AAS or ICP) and with the impurity concentrations in the sample.

The following procedure is given as an example.

8.2 Fusion

Prepare five decompositions of pure alumina (see 4.4.1) in accordance with 7.2, 7.3 or 7.4. Transfer into five 200 ml volumetric flasks and dilute to 150 ml with water.

Add the quantities of solutions (see 4.4.2 and 4.4.4 to 4.4.7 for fusion 1 or 4.4.2 to 4.4.7 for fusion 2 and fusion 3) indicated in table 1. Make up to the mark with water.

Elements	1	2	3	4	5
Na	0 µl	1 ml	2 ml	4 ml	6 ml
K	0 µl	100 µl	200 µl	300 µl	400 µl
Ca	0 µl	250 µl	500 µl	750 µl	1000 µl
Fe	0 µl	100 µl	200 µl	300 µl	400 µl
Si	0 µl	200 µl	400 µl	800 µl	1600 µl
Mg	0 µl	250 µl	500 µl	750 µl	1000 µl

8.3 Acid dissolution

Prepare 5 dissolutions of pure alumina (see 4.4.1) in accordance with 7.5 or 7.6. Transfer into five 100 ml volumetric flasks and dilute to 50 ml with water. Add the quantities indicated in table 2. Make up to the mark with water.

Elements	1	2	3	4	5
Na	0 µl	500 µl	1000 µl	2000 µl	3000 µl
K	0 µl	50 µl	100 µl	150 µl	200 µl
Ca	0 µl	125 µl	250 µl	375 µl	500 µl
Fe	0 µl	50 µl	100 µl	150 µl	200 µl
Mg	0 µl	125 µl	250 µl	375 µl	500 µl

8.4 Drawing the calibration curve

8.4.1 Blank test

Prepare a blank test in accordance with 8.2, 8.3, 8.4, 8.5 or 8.6, using the same quantities of all reagents as for dissolution of the test sample, but using pure alumina (see 4.4.1) in place of the test sample.

8.4.2 Drawing the calibration curve

With the calibration solution prepared according to 8.3, curves can be drawn directly in ppm of impurity by using table 3 in which b_1 is the quantity of each oxide present as impurity in the solution obtained according to 8.4.1.

Oxides	1	2	3	4	5
Na ₂ O (ppm)	b_1	$674 + b_1$	$1348 + b_1$	$2695 + b_1$	$4044 + b_1$
K ₂ O (ppm)	b_2	$60 + b_2$	$120 + b_2$	$180 + b_2$	$240 + b_2$
CaO (ppm)	b_3	$175 + b_3$	$350 + b_3$	$525 + b_3$	$700 + b_3$
Fe ₂ O ₃ (ppm)	b_4	$71 + b_4$	$143 + b_4$	$214 + b_4$	$286 + b_4$
SiO ₂ (ppm)	b_5	$214 + b_5$	$428 + b_5$	$856 + b_5$	$1711 + b_5$
MgO (ppm)	b_6	$207 + b_6$	$414 + b_6$	$622 + b_6$	$829 + b_6$

9 Adjustment of the apparatus

9.1 Atomic absorption spectrometer

(see ECSC/CI9)

Follow the manufacturer's instructions for igniting and extinguishing the nitrous oxide-acetylene flame to avoid explosion, and ensure the safety screen is in place.

Set the wavelengths for the elements to be analyzed (see table 4) and adjust the apparatus so as to obtain maximum absorbance. Fit the correct burner and, in accordance with the manufacturer's instructions, light the flame. After 10 min preheating of the burner, adjust fuel and burner to obtain maximum absorbance while aspirating the highest calibration solution.

Aspirate water and set to give the zero absorbance, alternately aspirate the calibration solutions and water to establish that the absorbance reading is not drifting and draw the calibration graph.

9.2 Inductively coupled plasma spectrometer

(see ISO 13527)

Follow the manufacturer's instructions for igniting the plasma. Ensure the safety screen is in place.

The wavelengths in table 4 may be used for the analysis (according to analytical method and elements concentration, other wavelengths can also be used).

Wait until a stable signal is obtained. It is possible to use the Al signal as an internal standard to improve the precision of the results.

Elements	Atomic absorption (recommendations)	ICP (recommendations)
Na	589,0 nm	589,0 nm
K	766,5 nm	769,9 nm
Ca	422,7 nm	393,3 nm
Fe	248,3 nm	259,9 nm
Si	251,6 nm	251,6 nm
Mg	285,2 nm	279,5 nm

10 Measurements

Aspirate the calibration solutions and the final test solutions in order of increasing concentrations.

Aspirate water between each solution and record the readings when stable responses are obtained. Repeat the measurements at least twice more and calculate the average of the readings for each solution.

For each element, drawn the calibration curve by plotting the signal values of the calibration solution against the quantities of element oxide in ppm (see table 3). Convert the signal value of the test solution to ppm by mean of the calibration curves to obtain the raw result.

11 Expression of the results

Calculate the concentration of oxide in ppm by subtracting the from the raw result obtained according to clause 10, the blank value b_i (this value b_i is the concentration read at the intersection of the calibration curve with concentration axis).

12 Accuracy

Examples of the reproducibility (R) and repeatability (r) of these measurements, determined in accordance with ISO 5725, using a standard reference alumina (NBS 699) are given in table 5.

Oxides	Results number	Mean (ppm)	r (ppm)	R (ppm)	Certified value (ppm)
Na ₂ O	55	5778	228	763	5900 ± 100
K ₂ O	40	66	10	36	50 (not certified)
Fe ₂ O ₃	48	132	23	38	130 ± 10
SiO ₂	43	112	26	121	140 ± 20
CaO	51	343	18	88	360 ± 20
MgO	32	5	2	3	6 ± 2

13 Test report

The test report shall include the following information:

- a) name of testing establishment;
- b) place and date of test, report identification, signatory;
- c) reference to this European Standard (EN 725-1), including information on the method used (AAS or ICP);
- d) details of the equipment used, if needed;
- e) any information on the decomposition of the sample;
- f) calibration procedure;
- g) material type, manufacturing code, batch number;
- h) relevant test parameters;
- i) results of individual analyses;
- j) mean results and standard deviation;
- k) reproducibility and repeatability of the method;
- l) comments about the test and test results.

List of references

See national foreword.

BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: 020 8996 9000. Fax: 020 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: 020 8996 9001. Fax: 020 8996 7001.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre. Tel: 020 8996 7111. Fax: 020 8996 7048.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: 020 8996 7002. Fax: 020 8996 7001.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

If permission is granted, the terms may include royalty payments or a licensing agreement. Details and advice can be obtained from the Copyright Manager. Tel: 020 8996 7070.