

# Advanced technical ceramics — Methods of test for ceramic powders —

## Part 4: Determination of oxygen content in aluminium nitride by XRF analysis

The European Standard EN 725-4:2006 has the status of a  
British Standard

ICS 81.060.30

## National foreword

This British Standard was published by BSI. It is the UK implementation of EN 725-4:2006. It supersedes DD ENV 725-4:1994 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee RPI/13, Advanced technical ceramics.

A list of organizations represented on RPI/13 can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 November 2006

© BSI 2006

ISBN 0 580 49711 9

### Amendments issued since publication

Amd. No.	Date	Comments

English Version

**Advanced technical ceramics - Methods of test for ceramic  
powders - Part 4: Determination of oxygen content in aluminium  
nitride by XRF analysis**

Céramiques techniques avancées - Méthodes d'essai pour  
poudres céramiques - Partie 4: Détermination de la teneur  
en oxygène du nitrure d'aluminium par spectrométrie de  
fluorescence XRF

Hochleistungskeramik - Prüfverfahren für keramische  
Pulver - Teil 4: Bestimmung der Sauerstoffgehaltes in  
Aluminiumnitrid-Pulvern mittels Röntgenfluoreszenzanalyse  
(RFA)

This European Standard was approved by CEN on 13 March 2006.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**Management Centre: rue de Stassart, 36 B-1050 Brussels**

<b>Contents</b>		Page
<b>Foreword</b> .....		<b>3</b>
<b>1</b>	<b>Scope</b> .....	<b>5</b>
<b>2</b>	<b>Normative references</b> .....	<b>5</b>
<b>3</b>	<b>Principle</b> .....	<b>5</b>
<b>4</b>	<b>Apparatus and experimental conditions</b> .....	<b>5</b>
<b>4.1</b>	<b>Apparatus</b> .....	<b>5</b>
<b>4.2</b>	<b>Experimental conditions</b> .....	<b>5</b>
<b>5</b>	<b>Sample preparation</b> .....	<b>6</b>
<b>5.1</b>	<b>Granulometry</b> .....	<b>6</b>
<b>5.2</b>	<b>Test piece preparation</b> .....	<b>6</b>
<b>6</b>	<b>Calibration</b> .....	<b>6</b>
<b>6.1</b>	<b>Preparation of standards</b> .....	<b>6</b>
<b>6.2</b>	<b>Calibration curves</b> .....	<b>6</b>
<b>6.3</b>	<b>Recalibration</b> .....	<b>7</b>
<b>6.4</b>	<b>Sensitivity — limit of detection</b> .....	<b>7</b>
<b>7</b>	<b>Repeatability</b> .....	<b>7</b>
<b>8</b>	<b>Test report</b> .....	<b>8</b>
<b>Bibliography</b> .....		<b>10</b>

## Foreword

This European Standard (EN 725-4:2006) has been prepared by Technical Committee CEN/TC 184 "Advanced technical ceramics", the secretariat of which is held by BSI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2006, and conflicting national standards shall be withdrawn at the latest by October 2006.

This European Standard supersedes ENV 725-4:1994.

The main changes in the new edition are:

- revision of the foreword to reflect changes in the EN 725 series;
- changes to the normative references;
- modification of the test piece preparation requirements;
- modification of the test report requirements;
- addition of a bibliographical reference.

EN 725 *Advanced technical ceramics — Methods of test for ceramic powders* was prepared in Parts as follows:

- Part 1: *Determination of impurities in alumina*
- Part 2: *Determination of impurities in barium titanate*
- Part 3: *Determination of oxygen content of non-oxides by thermal extraction with a carrier gas*
- Part 4: *Determination of oxygen content in aluminium nitride by XRF analysis*
- Part 5: *Determination of particle size distribution*
- Part 6: *Determination of specific surface area* [withdrawn]
- Part 7: *Determination of absolute density* [withdrawn]
- Part 8: *Determination of tapped bulk density*
- Part 9: *Determination of un-tapped bulk density*
- Part 10: *Determination of compaction properties*
- Part 11: *Determination of densification on natural sintering*
- Part 12: *Chemical analysis of zirconia*

Parts 6 and 7 of the series were superseded in 2005 by EN ISO 18757 and EN ISO 18753 respectively.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Cyprus, Czech Republic,

## **EN 725-4:2006**

Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.

## 1 Scope

This European Standard specifies a method for the determination of oxygen contents of 2 % or less in aluminium nitride powder, by XRF analysis.

## 2 Normative references

The following referenced documents are indispensable for the application of this European Standard. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 725-5, *Advanced technical ceramics — Methods of test for ceramic powders — Part 5: Determination of particle size distribution*

EN ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories (ISO/IEC 17025:2005)*

EN ISO 18757, *Fine ceramics (advanced ceramics, advanced technical ceramics) — Determination of specific surface area of ceramic powders by gas adsorption using the BET method (ISO 18757:2003)*

## 3 Principle

The sample is compacted to obtain a pellet, and the net intensity of the oxygen  $K_{\alpha}$  characteristic ( $I_0 K_{\alpha}$ ) of the pellet is measured by XRF spectrometer. The net intensity is the background intensity subtracted from the peak intensity at the peak position. The oxygen content in the sample is determined from this net intensity and by reference to previously defined calibration graphs; the calibration graph expresses the direct relation between net intensity and oxygen concentration. Calibration standards are prepared from different mixtures of fine aluminium nitride powder and a fine alumina powder.

## 4 Apparatus and experimental conditions

### 4.1 Apparatus

Sequential X-ray fluorescence spectrometer, equipped with a multilayer pseudo crystal and a gas flow proportional counter.

NOTE For further details, see EN ISO 12677.

### 4.2 Experimental conditions

The experimental conditions depend on the apparatus, with an example of good practice being as follows:

- a) Mode: vacuum
- b) Excitation: tube voltage: 30 kV  
tube current: 95 mA
- c) Spectrometer: crystal: multilayer pseudo crystal  
2 d = 55 Å
- d) Peak: wavelength: 23,62 Å  
energy: 0,525 keV  
angular position: 51° 2θ (to be adjusted)

## EN 725-4:2006

- e) Detection: detector type: gas flow proportional counter Ar-CH<sub>4</sub> or He-CO<sub>2</sub>  
pulse height selection: the discriminate levels ( $V_{\min}$ ,  $V_{\max}$ ) should be set so as:  
 $V_{\min}, V_{\max} = V \pm 1,3 W$   
where  $W$  is the width at half height of the pulse amplitude distribution peaking at  $V$  (this allows more than 99,7 % of the pulse amplitude distribution to pass through the window)
- f) Background: Background intensity at the peak position is obtained using a linear regression between two background measurements before and after the fluorescence peak wavelength: 23,62 Å  
The recommended positions ( $2\theta_{\min}$ ,  $2\theta_{\max}$ ) are:
- $$2\theta_{\min}, \max = 2\theta_p \pm 4 W_p$$
- where  $2\theta_p$  is the peak position  
 $W_p$  is the peak width at half height of the peak
- g) Counting time: 100 s for both peak and background measurements; total counting time 300 s

## 5 Sample preparation

### 5.1 Granulometry

Powders shall be fine enough to obtain a good homogeneous pellet without a binder. The particle size distribution, determined in accordance with EN 725-5, shall be between 10 µm and 2 µm with an equivalent diameter corresponding to 50 % of the grain size.

NOTE Grinding the sample may increase the oxygen content of the powder.

### 5.2 Test piece preparation

**5.2.1** Where comparatively large (10 g) samples are available, pour the powder into a die (e.g. 40 mm in diameter) and compact it. The powder may be placed in an aluminium cup before pressing if required.

**5.2.2** For smaller samples (2 g), press the powder on to a backing plate made of cellulose or boric acid powder. The aluminium cup may again be used if required.

**5.2.3** Prepare two samples for each determination.

NOTE In most quantitative analysis, sample thickness is considered as infinite. If critical thickness is the thickness that would yield 99,9 % of the maximum intensity, calculation shows that critical thickness increased from 2,9 µm to 3,35 µm when oxygen concentration in AlN increases from 0,5 % to 10 %.

## 6 Calibration

### 6.1 Preparation of standards

Prepare calibration standards from mixtures of aluminium nitride powder in accordance with 5.1 and alumina powder with a specific surface area of about 6 m<sup>2</sup>/g according to EN ISO 18757 and an equivalent diameter corresponding to 50 % of the distribution, normally 1 µm or less. Mix the powders for at least 10 min in an agate mortar and pestle, and vibrate the mixture for 1 min.

### 6.2 Calibration curves

Establish curves which record the oxygen ( $K_a$ ) net intensity with the oxygen mass fraction in the different calibration mixtures.



As aluminium nitride powders used in calibration mixtures are not oxygen free, an addition method shall be used in order to determine the total oxygen concentration. In this method, a known amount of oxygen,  $\Delta[\text{O}]$ , is added to the sample. Fluorescence intensities  $I_1$  and  $I_2$ , before and after addition, are used to extrapolate to the unknown concentration. If linearity is assumed:

$$\frac{[\text{O}]}{\Delta[\text{O}]} = \frac{I_1}{I_1 - I_2}$$

The addition method is limited to fairly low concentrations, usually less than 1 %. Thereby, the use of an aluminium nitride powder with oxygen concentration  $\leq 1$  % is recommended for calibration purposes.

The shape of the calibration curve depends on the oxygen concentration range, but within a range of 0 % to 5 %, a linear regression is likely to be convenient.

### 6.3 Recalibration

Recalibration is needed in order to check long term instrumental drift. An aluminium nitride sintered pellet with a flat polished surface can be used to shift the calibration curve. The oxygen content of this reference standard should not be lower than 3 % in order to obtain a good signal intensity level.

Another solution, which is more time consuming, consists in the determination of a new calibration curve.

### 6.4 Sensitivity — limit of detection

#### EXAMPLE

Sensitivity (S): slope of the linear part of the calibration curve

expected sensitivity  $S = 80$  cps/% (counts per second per %).

Limit of detection for a confidence level of  $3\sigma$  ( $LD_{3\sigma}$ )

$$LD_{3\sigma} = \frac{3\sqrt{n/t}}{S}$$

where

$n$  is the background intensity at peak position (40 cps);

$t$  is the measuring time (100 s);

$LD_{3\sigma}$  is the oxygen content ( $240 \times 10^{-6}$  (240 ppm)).

## 7 Repeatability

This procedure has been applied to nine pellets of the same fine aluminium nitride powder, with the analytical results given in Table 1.

Table 1 — Example of repeatability for nine pellets from the same aluminium nitride powder

Pellet No	Oxygen concentration (wt%)
1	1,039
2	1,037
3	1,070
4	1,044
5	1,086
6	1,069
7	1,062
8	1,049
9	1,059
$\bar{x}$ (mean value)	1,057
$\sigma$ (standard deviation)	0,016
$\varepsilon = \frac{\sigma}{\bar{x}}$ (relative standard deviation)	1,55

NOTE 1 The relative error on oxygen concentration determination is

$$\pm \frac{t^*}{\sqrt{2^{**}}} \approx 2,5\%$$

\* with 95 % confidence level ( $p = 0,05$ ,  $n-1 = 8$ ), the Student factor = 2,306;

\*\* two pellets are pressed; one measurement is made on each pellet.

NOTE 2 The results obtained are not sufficient for establishing reproducibility values.

## 8 Test report

The report shall be in accordance with the reporting provisions of EN ISO/IEC 17025, and shall include at least following information:

- name and address of the testing establishment;
- date of test;
- on each page, a unique report identification and page number;
- customer name and address;
- reference to this European Standard, i.e. determined in accordance with EN 725-4;
- authorising signature;
- any deviation from the method described, with appropriate validation demonstration, to be acceptable for the parties involved;
- equipment used;

- i) calibration procedures;
- j) description of the powder (e.g. material type, manufacturer code, batch or code number);
- k) relevant test parameters;
- l) results of individual determinations, mean results and standard deviations;
- m) comments on the test or test results.

## Bibliography

- [1] EN ISO 12677, *Chemical analysis of refractory products by XRF — Fused cast bead method (ISO 12677:2003)*



---

---

## BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

### Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover.  
Tel: +44 (0)20 8996 9000. Fax: +44 (0)20 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

### Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: +44 (0)20 8996 9001.  
Fax: +44 (0)20 8996 7001. Email: [orders@bsi-global.com](mailto:orders@bsi-global.com). Standards are also available from the BSI website at <http://www.bsi-global.com>.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

### Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre.  
Tel: +44 (0)20 8996 7111. Fax: +44 (0)20 8996 7048. Email: [info@bsi-global.com](mailto:info@bsi-global.com).

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration.  
Tel: +44 (0)20 8996 7002. Fax: +44 (0)20 8996 7001.  
Email: [membership@bsi-global.com](mailto:membership@bsi-global.com).

Information regarding online access to British Standards via British Standards Online can be found at <http://www.bsi-global.com/bsonline>.

Further information about BSI is available on the BSI website at <http://www.bsi-global.com>.

### Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

Details and advice can be obtained from the Copyright & Licensing Manager.  
Tel: +44 (0)20 8996 7070. Fax: +44 (0)20 8996 7553.  
Email: [copyright@bsi-global.com](mailto:copyright@bsi-global.com).