

Non-biological systems for use in sterilizers

Part 1. General requirements

The European Standard EN 867-1 : 1997 has the status of a
British Standard

ICS 11.080

Committees responsible for this British Standard

The preparation of this British Standard was entrusted by Technical Committee LBI/35, Sterilizers, autoclaves and disinfectors, to Subcommittee LBI/35/3, Sterilization indicators, upon which the following bodies were represented:

ABHI Special Interest Section (Sterilizers and Disinfectors)
Association of British Healthcare Industries
BLWA Ltd. (the Association of the Laboratory Supply Industry)
Department of Health (Medical Devices Agency)
European Sterilization Packaging Association (UK)
Medical Sterile Products Association
Public Health Laboratory Service
Royal College of Pathologists
Royal Pharmaceutical Society of Great Britain

This British Standard, having been prepared under the direction of the Sector Board for Materials and Chemicals, was published under the authority of the Standards Board and comes into effect on 15 October 1997

© BSI 1997

Amendments issued since publication

Amd. No.	Date	Text affected

The following BSI references relate to the work on this standard:
Committee reference LBI/35/3
Draft for comment 92/57874 DC

ISBN 0 580 28130 2

National foreword

This Part of BS EN 867 has been prepared by Subcommittee LBI/35/3, and is the English language version of EN 867-1 : 1997 *Non-biological systems for use in sterilizers — Part 1: General requirements*, published by the European Committee for Standardization (CEN).

Cross-reference

Publication referred to	Corresponding British Standard
EN 28601	BS EN 28601 : 1992 <i>Specification for representation of dates and times in information interchange</i>

Attention is also drawn to BS 7720 : 1995 *Specification for non-biological sterilization indicators equivalent to the Bowie and Dick test*.

Compliance with a British Standard does not of itself confer immunity from legal obligations.

Summary of pages

This document comprises a front cover, an inside front cover, pages i and ii, the EN title page, pages 2 to 6, an inside back cover and a back cover.

ICS 11.080

Descriptors: Medical equipment, sterilizers, tests, sterilization, chemical indicators, classifications, specifications, labelling

English version

Non-biological systems for use in sterilizers — Part 1: General requirements

Systèmes non-biologiques destinés à être utilisés
dans des stérilisateurs —
Partie 1: Exigences générales

Nichtbiologische Systeme für en Gebrauch in
Sterilisatoren —
Teil 1: Allgemeine Anforderungen

This European Standard was approved by CEN on 1997-01-10. CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

The European Standards exist in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

CEN

European Committee for Standardization
Comité Européen de Normalisation
Europäisches Komitee für Normung

Central Secretariat: rue de Stassart 36, B-1050 Brussels

Foreword

This European Standard has been prepared by Technical Committee CEN/TC 102, Sterilizers for medical purposes, the secretariat of which is held by DIN.

This standard is one of a series of European Standards concerned with non-biological systems for testing sterilizers. These standards are:

- EN 867-1 *Non-biological systems for use in sterilizers — Part 1: General requirements*
- EN 867-2 *Non-biological systems for use in sterilizers — Part 2: Process indicators (Class A)*
- EN 867-3 *Non-biological systems for use in sterilizers — Part 3: Specification for Class B indicators for use in the Bowie and Dick test*

In addition, CEN/TC 102 Working Group 7 has prepared a series of European Standards describing biological indicators for use in sterilizers. These European Standards are:

- EN 866-1 *Biological systems for testing sterilizers and sterilization processes — Part 1: General requirements*
- EN 866-2 *Biological systems for testing sterilizers and sterilization processes — Part 2: Particular systems for use in ethylene oxide sterilizers*
- EN 866-3 *Biological systems for testing sterilizers and sterilization processes — Part 3: Particular systems for use in moist heat sterilizers*
- prEN 866-4 *Biological systems for testing sterilizers and sterilization processes — Part 4: Particular systems for use in irradiation sterilizers*
- prEN 866-5 *Biological systems for testing sterilizers and sterilization processes — Part 5: Particular systems for use in low temperature steam and formaldehyde sterilizers*
- prEN 866-6 *Biological systems for testing sterilizers and sterilization processes — Part 6: Particular systems for use in dry heat sterilizers*

- prEN 866-7 *Biological systems for testing sterilizers and sterilization processes — Part 7: Particular requirements for self-contained biological indicator systems for use in moist heat sterilizers*
- prEN 866-8 *Biological systems for testing sterilizers and sterilization processes — Part 8: Particular requirements for self-contained biological systems for use in ethylene oxide sterilizers*

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 1997, and conflicting national standards shall be withdrawn at the latest by August 1997.

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

Contents

	Page
Foreword	2
Introduction	3
1 Scope	3
2 Normative references	3
3 Definitions	3
4 Classification	4
5 General requirements	4
6 Labelling and instructions for use	5
Annex A (informative) Bibliography	6

Introduction

European Standards for sterilizers (EN 285 and other standards in preparation by CEN/TC 102) and for the validation and process control of sterilization (EN 550, EN 552 and EN 554) describe performance tests for sterilizers and methods of validation and routine control, respectively.

This standard specifies the general requirements for non-biological indicators and subsequent parts specify the particular requirements for non-biological indicators for particular applications and for defined tests of particular sterilization processes. The use of the indicators specified in this standard are described in EN 285, EN 550, EN 552 and EN 554.

The non-biological indicators specified in this standard are not intended for use in any process other than that specified. The use of an inappropriate indicator can give dangerously misleading results.

The performance of a non-biological indicator can be affected by the conditions of storage prior to use, the methods of use, and the conditions of storage after exposure to the process. For these reasons, the recommendations of the manufacturer for storage and use should be followed precisely. Non-biological indicators should not be used beyond any expiry date stated by the manufacturer.

When a physical and/or chemical variable of a sterilizing process is outside its specified limits, a sterilization cycle should always be regarded as unsatisfactory, irrespective of the results obtained from the non-biological indicators.

1 Scope

This European Standard specifies general requirements for indicators which are not dependent for their action on the presence or absence of living organisms, and which are used to monitor the presence or attainment of one or more of the variables required for a satisfactory sterilization process.

This standard does not specify requirements for the use of biological systems.

NOTE. Biological test systems are regarded as those tests which are dependent for their interpretation on the demonstration of viability of an organism. Test systems of this type are considered in EN 866 *Biological systems for testing sterilizers and sterilization processes*.

The requirements of this part of EN 867 apply to all indicator systems specified in subsequent parts of EN 867, unless the requirement is modified or added to by a subsequent part, in which case the requirement of the particular part will apply.

2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

EN 28601 *Data elements and interchange formats — Information interchange — Representation of dates and times* (ISO 8601 : 1988 and technical corrigendum 1 : 1991)

3 Definitions

For the purposes of this European Standard the following definitions apply:

3.1 graduated response

A progressive visible change occurring on exposure to one or more process variables allowing assessment of the level achieved.

3.2 defined end-point

The visible change occurring after exposure to the specified variable(s) at a level equal to or greater than that specified for the indicator.

3.3 indicator

The indicator system in the form in which it is intended to be used.

3.4 indicator reagent

The active ingredient or combination of ingredients before conversion into the indicator.

3.5 indicator system

The combination of the indicator reagent and its substrate.

4 Classification

NOTE. In the subsequent parts of this standard indicators are classified by their intended use and, within each of these classifications, by the process for which they are designed to be used.

4.1 Process indicators — Class A

Process indicators are intended for use with individual packs of product intended to be sterilized to demonstrate that the pack has been exposed to the process. They may be designed to react to one or more of the critical process variables but may be designed to achieve their end-point reaction after exposure to sub-optimal levels of the process variable.

Process indicators have a defined end-point reaction.

4.2 Indicators for use in specific tests — Class B

These indicators are designed for use in a specific test procedure defined in the relevant sterilizer/sterilization standard.

These indicators may have a graduated response or a defined end-point reaction.

4.3 'Single-variable' indicators — Class C

'Single-variable' indicators are designed to monitor the attainment of the required value of one critical variable in the sterilization process.

'Single-variable' indicators may have a graduated response or a defined end-point reaction.

4.4 'Multi-variable' indicators — Class D

'Multi-variable' indicators are designed to monitor the attainment of the required value of two or more critical variables in the sterilization process.

'Multi-variable' indicators may have a graduated response or a defined end-point reaction.

5 General requirements

5.1 General

5.1.1 The requirements of this section shall apply to all indicators unless specifically excluded or amended in a subsequent part of this standard.

5.1.2 The manufacturer shall establish, document and maintain a formal quality system to cover all operations required by this standard.

NOTE. EN ISO 9001 and EN ISO 9002 and of the EN 45000 series describe requirements for quality systems for manufacture and testing.

5.2 Labelling

Where a date is required on labelling or information supplied with the product, this shall be expressed in accordance with EN 28601.

5.3 Test methods

The test methods specified in this standard are reference methods. When alternative methods are used routinely these shall be defined, validated and have known correlation with the reference method.

5.4 Marking

5.4.1 Each indicator shall be clearly marked with the type of process for which it is intended to be used and the class of indicator as defined in this standard (see 4.1 to 4.4 inclusive).

If the use of the indicator is limited to specific sterilization cycles, this information shall also be stated or coded on the indicator, e.g. steam 15 min 121 °C.

Where the size or format of the indicator does not permit this information to be stated in a font of 6 characters per centimetre, or larger, the information shall be provided on the label and/or instructions for use.

5.4.2 Abbreviated descriptions of the process, if used, shall be in accordance with the following symbols:

STEAM

– all steam sterilization processes

DRY

– all dry heat sterilization processes

EO

– all ethylene oxide sterilization processes

IRRAD

– all ionizing radiation sterilization processes

FORM

– all steam/formaldehyde sterilization processes

5.5 Performance requirements

5.5.1 The change which occurs after exposure of the indicator to the specified conditions shall be clearly visible to the naked eye and shall be either from light to dark or shall be a change from one colour to a distinctly different colour.

5.5.2 The manufacturer shall retain documentary evidence that the indicator does not release any substance known as toxic, in sufficient quantity to cause a health hazard, before during or after the sterilization process for which it is designated.

NOTE. Until relevant European or International Standards are published National Regulations apply.

5.5.3 The condition of the indicator after exposure to a sterilization process, during which all the variables met or exceeded the specified level to produce a defined end-point or a known graduated response shall remain visually unchanged for a period of not less than 6 months, from the date of use, when stored under the conditions specified by the indicator manufacturer.

The manufacturer shall state any change which can occur on storage for periods greater than 6 months (see 6.2).

NOTE. Incompletely changed indicators can deteriorate on storage; either returning to the unchanged condition or slowly completing the change reaction. If this can occur this should be stated by the manufacturer (see 6.2). Such indicators can not be suitable for use as a permanent record.

6 Labelling and instructions for use

6.1 The label of each package of indicators, or the technical information leaflet supplied with the package, shall state:

- a) the class and process for which the indicator is designed;
- b) the storage conditions, before and after use;
- c) the expiry date under the specified storage conditions;
- d) a unique code from which the manufacturing history can be traced;
- e) any safety precautions required during use;
- f) the manufacturer's or supplier's name and address.

6.2 The label of each package of indicators, or the technical information leaflet supplied with the package or available to each customer, or user, on request shall state:

- a) the change that is intended to occur; and for colour change indicators where the colour change cannot be adequately described, examples of the expected colour range for both changed and unchanged indicators;
- b) the minimum conditions required to effect the change;
- c) any specific instructions for use essential to ensure correct functioning of the indicator;
- d) any interfering substances or conditions which are known to adversely affect the performance of the indicator;
- e) the manufacturer's or supplier's name and address;
- f) the storage conditions for the indicator after use;
- g) the nature of any change which can occur on storage for completely and incompletely changed indicators (see clause 5).

Annex A (informative)

Bibliography

- EN 285 *Sterilization — Steam sterilizers — Large sterilizers*
- EN 550 *Sterilization of medical devices — Validation and routine control of ethylene oxide sterilization*
- EN 552 *Sterilization of medical devices — Validation and routine control of sterilization by irradiation*
- EN 554 *Sterilization of medical devices — Validation and routine control of sterilization by moist heat*
- EN ISO 9001 *Quality systems — Model for quality assurance in design/development, production, installation and servicing (ISO 9001 : 1994)*
- EN ISO 9002 *Quality systems — Model for quality assurance in production, installation and servicing (ISO 9002 : 1994)*
- EN 45001 *General criteria for the operation of testing laboratories*
- EN 45014 *General criteria for suppliers declaration of conformity*
- prEN 1422 *Sterilizers for medical purposes — Ethylene oxide sterilizers — Specifications*

List of references

See national foreword.

BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: 020 8996 9000. Fax: 020 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: 020 8996 9001. Fax: 020 8996 7001.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre. Tel: 020 8996 7111. Fax: 020 8996 7048.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: 020 8996 7002. Fax: 020 8996 7001.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

If permission is granted, the terms may include royalty payments or a licensing agreement. Details and advice can be obtained from the Copyright Manager. Tel: 020 8996 7070.