

**Chemical analysis of  
ferrous materials —  
Determination of nickel  
in steels and irons —  
Flame atomic  
absorption  
spectrometric method**

This European Standard EN 10136:1989 has the status of a  
British Standard

## Cooperating organizations

The European Committee for Standardization, under whose supervision this European Standard was prepared, comprises the national standards organizations of the following Western European countries.

Austria	Oesterreichisches Normungsinstitut
Belgium	Institut belge de normalisation
Denmark	Dansk Standardiseringsraad
Finland	Suomen Standardisoimisiitto, r.y.
France	Association française de normalisation
Germany	Deutsches Institut für Normung e.V.
Greece	Hellenic Organization for Standardization
Iceland	Technological Institute of Iceland
Ireland	National Standards Authority of Ireland
Italy	Ente Nazionale Italiano di Unificazione
Luxemburg	Inspection du Travail et des Mines
Netherlands	Nederlands Normalisatie-instituut
Norway	Norges Standardiseringsforbund
Portugal	Instituto Português da Qualidade
Spain	Asociación Española de Normalización y Certificación
Sweden	Standardiseringskommissionen i Sverige
Switzerland	Association suisse de normalisation
United Kingdom	British Standards Institution

This British Standard was published under the authority of the Board of BSI and comes into effect on 31 January 1991

© BSI 12-1999

The following BSI references relate to the work on this standard:

Committee reference ISM/18  
Draft announced *BSI News*  
June 1989

### Amendments issued since publication

Amd. No.	Date of issue	Comments

ISBN 0 580 19300 4

---

---

# Contents

	Page
Cooperating organizations	Inside front cover
National foreword	ii
Brief history	2
Text of EN 10136	3
National appendix NA	Inside back cover

---

## National foreword

This British Standard has been prepared under the direction of the Iron and Steel Standards Policy Committee. It is the English language version of EN 10136:1989 “*Chemical analysis of ferrous materials. Determination of nickel in steels and irons. Flame atomic absorption spectrometric method*”, published by the European Committee for Standardization (CEN). It supersedes BS 6200-3.20.4:1986, which is withdrawn.

Other methods for the determination of nickel in ferrous materials have been published as BS 6200-3.20.1 and BS 6200-3.20.3.

NOTE The text of the EN contains the following minor typographical errors. These have been drawn to the attention of the appropriate Secretariat and will be corrected in the next edition of the EN.

In 5.1.1, lines 2 and 7, “absorbency” should read “absorbance”.

In 7.3.3, “Lamp current — Slit width” should be directly opposite “Follow manufacturer’s recommendations”.

In 7.3.5, lines 6 and 9, “absorbency” should read “absorbance”.

In 7.4, line 5, “velues” should read “values”.

A British Standard does not purport to include all the necessary provisions of a contract. Users of British Standards are responsible for their correct application.

**Compliance with a British Standard does not of itself confer immunity from legal obligations.**

### Summary of pages

This document comprises a front cover, an inside front cover, pages i and ii, the EN title page, pages 2 to 8, an inside back cover and a back cover.

This standard has been updated (see copyright date) and may have had amendments incorporated. This will be indicated in the amendment table on the inside front cover.

---

UDC 669.14:543.422:546.74

Key words: Iron- and steel products, steels, cast iron, chemical analysis, determination of content, nickel, atomic absorption spectrophotometry, flame photometry.

English version

Chemical analysis of ferrous materials  
Determination of nickel in steels and irons  
Flame atomic absorption spectrometric method

Analyse chimique des matériaux  
sidérurgiques — Dosage du nickel dans les  
aciers et les fontes Méthode par spectrométrie  
d'absorption atomique dans la flamme

Chemische Analyse von Eisenwerkstoffen —  
Bestimmung von Nickel in Stahl  
Flammenatomabsorptionsspektrometrisches  
Verfahren

This European Standard was accepted by CEN on 15 January 1989. CEN members are bound to comply with the requirements of the CEN/CENELEC Rules which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Central Secretariat has the same status as the official versions.

CEN members are the national standards organizations of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxemburg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

**CEN**

European Committee for Standardization  
Comité Européen de Normalisation  
Europäisches Komitee für Normung

**Central Secretariat: rue de Stassart 36, B-1050 Brussels**

**Brief history**

This European Standard takes over the content of EURONORM 136-85 "Chemical analysis of ferrous materials — Determination of nickel in steels and irons — Flame atomic absorption spectrometric method", prepared by ECISS/TC 20 "Methods of chemical analysis"; the Secretariat of which is allocated to the Dansk Standardiseringsrad (DS).

It has been submitted to the CEN Formal Vote following the decision of the Coordinating Commission (COCOR) of the European Committee for Iron and Steel Standardization on 1987-11-24/25.

It has been adopted and ratified by CEN BT on 1988-11-05.

According to the Common CEN/CENELEC Rules, following countries are bound to implement this European Standard:

Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxemburg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

Note in clauses 1 and 9 EURONORM shall read EUROPEAN STANDARD.

**Contents**

	Page
Brief history	2
1 Scope and field of application	3
2 Reference	3
3 Principle	3
4 Reagents	3
5 Apparatus	3
6 Sampling	4
7 Procedure	4
7.1 Test portion	4
7.2 Blank test	4
7.3 Determination	4
7.4 Establishment of the calibration graph	5
8 Expression of results	5
9 Test report	6
Annex Precision data	7

## 1 Scope and field of application

This EURONORM specifies a method for the determination of nickel in steels and irons by means of flame atomic absorption spectrometry.

The method is applicable to steels and irons with nickel contents of 0.003 to 2.0 % (m/m).

## 2 Reference

EURONORM 18 — Selection and preparation of samples and test pieces for steel and iron and steel products.

General guidelines for the application of flame atomic absorption spectrometric methods are in course of preparation.

## 3 Principle

Dissolution of a test portion in a mixture of appropriate acids and fuming with perchloric acid. Spraying of the solution into an air-acetylene flame. Determination of the nickel by means of the spectrometric measurement of the atomic absorption of the 232.0 nm or 352.5 nm line emitted by a nickel hollow cathode lamp.

The instrument is calibrated by addition of a nickel standard solution to a similar matrix to that of the test solution.

NOTE At the wavelength of 352.5 nm the signal-to-noise ratio is higher than at a wavelength of 232.0 nm. Generally, use of the 352.5 nm line will lead to a better reproducibility.

However, as the sensitivity at 352.5 nm is less than the sensitivity at 232.0 nm, with some instruments the use of the longer wavelength will be impossible when analysing low nickel contents.

## 4 Reagents

During the analysis use only reagents of recognized analytical reagent quality and having a very low nickel content, and only distilled water or water of equivalent purity.

Carefully check the nickel content of all reagents.

If possible, use only freshly prepared distilled or deionized water.

**4.1 Iron of high purity**, with a nickel content < 0.0005 % (m/m)

**4.2 Hydrochloric acid–nitric acid mixture**

Mix three volumes of hydrochloric acid,  $\rho$  1.19 g/ml approximately, one volume of nitric acid,  $\rho$  1.40 g/ml approximately and two volumes of water. This mixture is to be prepared immediately before use.

**4.3 Nitric acid–perchloric acid mixture**

Mix 100 ml of nitric acid,  $\rho$  1.40 g/ml approximately with 800 ml of perchloric acid,  $\rho$  1.54 g/ml approximately. Dilute to 1 l with water and mix.

NOTE Perchloric acid ( $\rho$  1.67 g/ml approximately) may also be used. 100 ml of perchloric acid ( $\rho$  1.54 g/ml approximately) is equivalent to 79 ml of perchloric acid ( $\rho$  1.67 g/ml approximately).

**4.4 Nickel stock solution**, corresponding to 1 mg of nickel per ml approximately

Weigh, to the nearest 0.001 g, about 0.5 g of high purity nickel ( $\geq$  99.9 % pure). Transfer the weighed mass to a 400 ml beaker and dissolve in 25 ml of nitric acid [ $\rho$  1.40 g/ml approximately diluted 1 + 1 (v/v)]. Boil to remove oxides of nitrogen. Cool and transfer the solution to a 500 ml volumetric flask, dilute to the mark with water and mix. Calculate the concentration of nickel in this stock solution, in mg/ml.

**4.5 Nickel reference solution**, corresponding to 40  $\mu$ g of nickel per ml approximately

Transfer 10.0 ml of nickel stock solution (4.4) to a 250 ml volumetric flask, dilute to the mark with water and mix. Calculate the concentration of nickel in the reference solution, in  $\mu$ g/ml.

## 5 Apparatus

Ordinary laboratory equipment and

**5.1 Atomic absorption spectrometer; a nickel hollow cathode lamp**; supplies of air and acetylene sufficiently pure to give a steady clear fuel-lean flame, free from water and oil, and free from nickel

The atomic absorption spectrometer used will be satisfactory if after optimization according to 7.3.4 the limit of detection and characteristic concentration are in reasonable agreement with the values given by the manufacturer and it meets the following performance requirements.

**5.1.1 Minimum precision**

The standard deviation of 10 measurements of the absorbency of the most concentrated solution shall not exceed 1.0 % of the mean absorbance.

The standard deviation of 10 measurements of the absorbance of the least concentrated calibration solution (excluding the zero calibration solution) shall not exceed 0.5 % of the mean absorbency of the most concentrated calibration solution.

For example, if the top and bottom calibration solutions represent 0.1 % and 0.01 % nickel in the sample, the precision called for (as two standard deviations) would be 0.002 % and 0.001 % nickel respectively, assuming curve linearity.

### 5.1.2 Additional performance requirements

It is also desirable that the instrument should conform to the following additional performance requirements.

- a) Characteristic concentration — The characteristic concentration for nickel in a matrix similar to the final test solution shall be better than 0.10 µg/ml.
- b) Limit of detection — The limit of detection of nickel in a matrix similar to the final test solution shall be better than 0.15 µg/ml.

### 5.2 Ancillary equipment

A strip chart recorder and/or digital readout device is recommended to evaluate the criteria in 5.1 and for all subsequent measurements.

## 6 Sampling

Sampling shall be carried out in accordance with EURONORM 18.

## 7 Procedure

**WARNING** — Perchloric acid vapours may cause explosions in the presence of ammonia, nitrous fumes or organic matter in general. Always use a specially designed fume-hood.

### 7.1 Test portion

Weigh the mass ( $m$ ) indicated below to the nearest 0.001 g:  $m = 1 \text{ g} \pm 5 \%$ .

### 7.2 Blank test

With each analytical run, carry out an analysis on a reagent blank in parallel with the test portion analysis, using identical reagents, conditions, analytical procedures and dilutions throughout.

### 7.3 Determination

#### 7.3.1 Preparation of the test solution

Place the test portion (7.1) in a 250 ml beaker. Add 10 ml of hydrochloric acid–nitric acid mixture (4.2). Heat gently until the reaction ceases. Add 15 ml of nitric acid–perchloric acid mixture (4.3) and heat until dense white fumes of perchloric acid appear. Continue heating for one minute and allow to cool.

**NOTE** If the test portion is readily soluble in the nitric acid–perchloric acid mixture (4.3), the addition of hydrochloric acid–nitric acid mixture (4.2) can be omitted. In that case the test portion is dissolved in the nitric acid–perchloric acid mixture (4.3) and the solution thus obtained is heated as described above.

Dissolve in 25 ml of water by heating gently. Cool again and transfer the solution to a 100 ml volumetric flask, dilute to the mark with water and mix. This is the test solution. If some residue has been left in the solution due to silicon, tungsten, niobium or tantalum, filter the solution through a dry, medium textured filter paper, discarding the first runnings.

If the expected nickel content of the test sample exceeds 0.1 % (m/m), dilute the solution as follows: Transfer 25.0 ml of the solution to a 500 ml volumetric flask, dilute to the mark with water and mix.

**NOTE** If the solution has to be diluted to give the test solution, the blank test solution (7.2) must be diluted in the same way.

#### 7.3.2 Preparation of the calibration solutions

Place 10 g of iron (4.1) in a 800 ml beaker. Add 100 ml of hydrochloric acid–nitric acid mixture (4.2) and heat gently to dissolve the iron. Add 150 ml of nitric acid–perchloric acid mixture (4.3) and heat until dense white fumes of perchloric acid appear. Continue heating for one minute and allow to cool. Dissolve in 100 ml of water by heating gently. Cool again and transfer the solution to a 250 ml volumetric flask, dilute to the mark with water and mix.

##### 7.3.2.1 Nickel content < 0.1 % (m/m)

Transfer a series of 25.0 ml aliquots of the iron solution to 100 ml volumetric flasks, add to each flask by means of a burette or pipette respectively 0, 2.5, 5.0, 10.0, 15.0, 20.0 and 25.0 ml of nickel reference solution (4.5), dilute to the mark with water and mix.

##### 7.3.2.2 Nickel content 0.1 to 2.0 % (m/m)

Transfer 25.0 ml of the iron solution to a 500 ml volumetric flask, dilute to the mark with water and mix. From this solution transfer 25.0 ml aliquots to 100 ml volumetric flasks.

Add to each flask by means of a burette or pipette respectively 0, 2.5, 5.0, 10.0, 15.0, 20.0 and 25.0 ml of nickel reference solution (4.5), dilute to the mark with water and mix.

**NOTE** 1 ml of nickel reference solution (4.5) is approximately equivalent to 0.004 % (m/m) in the sample in the case of 7.3.2.1 and 0.08 % (m/m) in the sample in the case of 7.3.2.2.



### 7.3.3 Adjustment of atomic absorption spectrometer (5.1)

Type of lamp:	Nickel hollow cathode.
Wavelength:	232.0 nm or 352.5 nm.
Flame:	Air-acetylene clear fuel-lean flame adjusted for maximum
Lamp current —	nickel response
Slit width	Follow manufacturer's recommendations.

NOTE The manufacturer's recommendations should be closely followed and particular attention is drawn to the following safety points:

- the explosive nature of acetylene and regulations concerning its use;
- the need to shield the eyes of the operator from UV radiation by means of tinted glass;
- the need to keep burner head clear of deposits. A badly clogged burner may cause a flash back;
- make sure that the liquid trap is filled with water.

### 7.3.4 Optimization of the atomic absorption spectrometer settings

Follow the manufacturer's instructions for preparing the instrument for use.

When the current to the lamp, the wavelength and flow of gas have been adjusted and the burner lit, spray water until the indication has stabilized. Set the absorbance value to zero. Choose a damping setting or integration time to give a signal steady enough to fulfil the precision requirements (5.1).

Adjust the flame to be clear fuel-lean and the burner height to about 1 cm below the light path. Spraying, alternately, the high and zero calibration solutions, adjust the gas flow and burner position (horizontally, vertically and rotationally) until the difference in absorbance between the calibration solutions is at a maximum.

Evaluate the criteria of 5.1.1 to ensure that the instrument is suitable for the determination.

### 7.3.5 Atomic absorption measurements

Set the scale expansion so that the top calibration solution gives nearly full scale deflection. Aspirate the calibration solutions in ascending order repetitively until each gives the acceptable precision (see 5.1.1), thus showing that the instrument has achieved stability. Select two calibration solutions, one having an absorbency just lower than the test solution and one just higher. Aspirate these first in ascending order, then in descending order, with the test solution as the middle solution in each case, measuring the absorbency in relation to water. Aspirate the complete range of calibration solutions again.

It is recognized that these procedures cannot be followed with automatic instruments which accept two calibration solutions only. In this case, it is suggested that the two "sandwiching" solutions should not be used for the primary calibration but should be analysed alternately with the test solution.

Obtain the net mean absorbance of each calibration solution by subtracting the mean absorbance of the zero calibration solution.

Obtain the mean absorbance of the test solution and the mean absorbance of the reagent blank solution.

Convert the mean absorbances of the test solution and of the reagent blank solution to micrograms of nickel per millilitre by means of the calibration graph (7.4).

### 7.4 Establishment of the calibration graph

It is necessary to draw up a new calibration graph for each series of determinations, and for the range of nickel contents expected.

Prepare a calibration graph by plotting the net mean absorbance values of the calibration solutions against micrograms of nickel per millilitre, calculated according to 4.5. Refer the net absorbances of the two adjacent calibration solutions to the graph. If these two calibration readings do not deviate from the graph by more than the permitted precision criteria then the test solution readings are acceptable.

## 8 Expression of results

The percentage by mass of nickel (Ni) is given by the expression

$$\text{Ni \% (m/m)} = \frac{(b - c) \cdot D \cdot 100}{10^6} \cdot \frac{100}{m} = \frac{(b - c) \cdot D}{100 m}$$

where:

- b* is the concentration of nickel in the test solution in  $\mu\text{g/ml}$ ;
- c* is the concentration of nickel in the reagent blank solution in  $\mu\text{g/ml}$ ;
- D* is the dilution factor in 7.3.1;
  - (*D* = 1 for samples with an expected nickel content of 0.1 % or less,
  - D* = 20 for samples with an expected nickel content of more than 0.1 %);
- m* is the mass of the test portion, in g.

## 9 Test report

The test report shall contain the following particulars:

- a) The method of analysis used by reference to this EURONORM;
- b) the results obtained, as well as the form in which they are expressed;
- c) any particular details which may have been noted during the determination;
- d) any operations not specified in this EURONORM or any optional operations which could have had an influence on the result;
- e) all indications necessary for the identification of the sample;
- f) the laboratory and the date of analysis.

## Annex Precision data

Planned trials of this method were carried out by analysts from 6 to 19 laboratories; 2–5 determinations were carried out by each analyst on each sample. From the results obtained the 95 % (2s) confidence limits have been calculated in accordance with ISO 5725 and are tabulated as follows:

CRM	Alloy type	Nickel %	Number of labs	Repeatability r	Reproducibility R
BCS 260/4	High purity iron	0.003	18	0.0013	0.0022
EURONORM 077-1	Plain carbon steel	0.031	6	0.0025	0.0031
BCS 431	Plain carbon steel	0.069	18	0.0037	0.0096
—	Cast iron	0.070	17	0.0024	0.0102
—	Low alloy iron	0.080	19	0.0076	0.0118
—	Low alloy iron	0.112	16	0.0073	0.0161
EURONORM 278-1	Cr-Mo steel	0.236	7	0.0221	0.0317
—	Low alloy iron	0.329	14	0.0132	0.0314
—	Low alloy iron	0.423	19	0.0146	0.0621
BCS 341	24 % Cr steel	0.56-	18	0.0241	0.0449
—	Ni-Cr-Mo-Cu iron	0.955	12	0.0187	0.0599
EURONORM 177-1	Low alloy steel	1.015	7	0.0195	0.0297
BCS 225/2	Ni-Cr-Mo steel	1.43-	18	0.0409	0.0513
BCS 406	Low alloy steel	1.69-	18	0.0464	0.0903
—	Low alloy iron	1.708	15	0.0418	0.0961
BCS 410	Low alloy steel	2.04-	18	0.0598	0.0784

### Repeatability

The difference between two single results found on identical material by one analyst using the same apparatus within a short time-interval will exceed the repeatability, r, not more than once in 20 cases in the normal and correct operation of the method.

### Reproducibility

The difference between two single and independent results found by two operators working in different laboratories on identical test material will exceed the reproducibility, R, on average not more than once in 20 cases in the normal and correct operation of the method.



## **National appendix NA**

The United Kingdom participation in the preparation of this European Standard was entrusted by the Iron and Steel Standards Policy Committee (ISM/-) to Technical Committee ISM/18 upon which the following bodies were represented:

BCIRA

British Forging Industry Association

British Steel Industry

Department of Trade and Industry (Laboratory of the Government Chemist)

Ferro Alloys and Metals Producers' Association

Ministry of Defence

---

---

## BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

### Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover.  
Tel: 020 8996 9000. Fax: 020 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

### Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: 020 8996 9001. Fax: 020 8996 7001.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

### Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre.  
Tel: 020 8996 7111. Fax: 020 8996 7048.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration.  
Tel: 020 8996 7002. Fax: 020 8996 7001.

### Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

If permission is granted, the terms may include royalty payments or a licensing agreement. Details and advice can be obtained from the Copyright Manager.  
Tel: 020 8996 7070.