BS EN 10211:2013



BSI Standards Publication

Chemical analysis of ferrous materials — Determination of titanium in steels and cast irons — Flame atomic absorption spectrometric method



BS EN 10211:2013 BRITISH STANDARD

National foreword

This British Standard is the UK implementation of EN 10211:2013. It supersedes BS EN 10211:1996 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee ISE/102, Methods of Chemical Analysis for Iron and Steel.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2013. Published by BSI Standards Limited 2013

ISBN 978 0 580 81798 4

ICS 77.040.30; 77.080.01

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 November 2013.

Amendments issued since publication

Date Text affected

EUROPEAN STANDARD NORME EUROPÉENNE EUROPÄISCHE NORM

EN 10211

November 2013

ICS 77.040.30

Supersedes EN 10211:1995

English Version

Chemical analysis of ferrous materials - Determination of titanium in steels and cast irons - Flame atomic absorption spectrometric method

Analyse chimique des produits ferreux - Détermination du titane dans les aciers et les fontes - Méthode par spectrométrie d'absorption atomique dans la flamme

Chemische Analyse von Eisenwerkstoffen - Bestimmung des Titananteils in Stahl und Eisen -Flammenatomabsorptionsspektrometrisches Verfahren

This European Standard was approved by CEN on 29 September 2013.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION COMITÉ EUROPÉEN DE NORMALISATION EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Con	tents	Page
Forew	ord	
1	Scope	4
2	Normative references	4
3	Principle	4
4	Reagents	4
5	Apparatus	6
6	Sampling	7
7	Procedure	7
8	Expression of results	10
9	Test report	10
Annex	A (informative) Precision data	12
A.1	General	12
A.2	Information on repeatability and reproducibilities	15
A.2.1	Repeatability, r	15
A.2.2	Within-laboratory reproducibility, R _w	15
A.2.3	Reproducibility, R	15
Annex	B (normative) Procedures for the determination of instrumental criteria	16
B.1	Determination of minimum precision	16
B.2	Determination of limit of detection	16
B.3	Linearity of calibration	16
B.4	Determination of characteristic concentration, $c_{\text{Ti,cc}}$	17
Biblio	graphy	18

Foreword

This document (EN 10211:2013) has been prepared by Technical Committee ECISS/TC 102 "Methods of chemical analysis for iron and steel", the secretariat of which is held by SIS.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2014, and conflicting national standards shall be withdrawn at the latest by May 2014.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 10211:1995.

In comparison with the previous edition of EN 10211, the following technical changes were made:

- Clause 1, "steel and iron" replaced by "non-alloyed and alloyed steels and cast irons";
- Clause 3, clarification regarding the use of aluminium chloride within both dissolutions described;
- Addition of 5.2.3 (information on the linearity criteria);
- Table 1, addition of terms corresponding to 0,025 % and 0,075 % titanium;
- Subclause 7.6 was merged in Subclause 7.5;
- Clause 8, formula: notation changed;
- Annex A, the notes were included in the text and the correlation coefficients corrected;
- Annex B: added.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

1 Scope

This European Standard specifies a flame atomic absorption spectrometric method for the determination of titanium in steels and cast irons. The method is applicable to non-alloyed and alloyed steels and cast irons with titanium contents of 0,01 % to 1,0 % (m/m).

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 648, Laboratory glassware - Single-volume pipettes (ISO 648)

EN ISO 1042, Laboratory glassware - One-mark volumetric flasks (ISO 1042)

EN ISO 14284, Steel and iron - Sampling and preparation of samples for the determination of chemical composition (ISO 14284)

3 Principle

Two dissolution procedures are possible and described:

- a) The first one is specific to cast irons and is based on dissolution of the test portion with nitric and perchloric acids, followed by evaporation with perchloric acid to dryness. The further steps being a separation of the insolubilised silica by filtration and collection of the filtrate.
- b) The second one, which can be used for steels as well as for cast irons is based on dissolution of the test portion with hydrochloric and nitric acid. This dissolution is followed by:
 - 1) filtration and ignition of the acid insoluble residue and removal of silica with hydrofluoric acid;
 - 2) fusion of the residue with a mixture of boric acid and sodium carbonate, dissolution of the melt with acid and addition of this solution to the reserved filtrate.

Addition of aluminium chloride as interaction buffer, in both cases.

Nebulisation of the test solution into a slightly fuel rich nitrous oxide/acetylene flame, of an atomic absorption spectrometer.

Spectrometric measurement of the atomic absorption of the 365,3 nm (or 364,3 nm) spectral line emitted by a titanium hollow-cathode lamp.

4 Reagents

During the analysis, use only reagents of recognised analytical grade and only distilled water or water of equivalent purity.

- **4.1** Nitric acid, HNO₃ ($\rho_{20} = 1,40 \text{ g/ml}$).
- **4.2** Hydrochloric acid, HCl ($\rho_{20} = 1,19 \text{ g/ml}$).

- **4.3** Perchloric acid, $HCIO_4$ ($\rho_{20} = 1,67$ g/ml).
- **4.4** Sulphuric acid, H_2SO_4 ($\rho_{20} = 1.84$ g/ml).
- **4.5** Hydrofluoric acid, HF ($\rho_{20} = 1,15 \text{ g/ml}$).
- 4.6 Nitric acid solution, 1 + 1.

Add 100 ml of nitric acid (4.1) to 100 ml water

4.7 Hydrochloric acid solution, 1 + 1.

Add 100 ml of hydrochloric acid (4.2) to 100 ml water

4.8 Aluminium chloride solution, equivalent to 10 g/l Al.

Dissolve 90 g of hydrated aluminium chloride (ALCl₃ . 6H₂O) in a 600 ml beaker with water and transfer the solution into a 1 000 ml one-mark volumetric flask. Dilute to the mark with water and mix well.

4.9 Fusion mixture.

Mix one part by mass of boric acid, H₃BO₃ and one part of sodium carbonate anhydrous, Na₂CO₃.

Store in a stoppered polyethylene container.

4.10 Iron-perchloric matrix solution, 20 g/l Fe.

Weigh, to the nearest 0,01 g, 10,00 g of pure iron free of titanium or of known low titanium content. Transfer into a 600 ml beaker and cover with a watch glass. Add in small portions, 150 ml of nitric acid (4.6). When effervescence ceases, add 200 ml of perchloric acid (4.3). Heat gently, until the iron is completely dissolved.

Continue heating at a temperature at which a steady reflux of white perchloric acid fumes is maintained on the walls of the beaker. Continue heating until there are no perchloric acid fumes visible inside the beaker and the iron perchlorate crystallises.

Allow to cool, add 150 ml of water, swirl and boil gently to dissolve the salts.

Transfer quantitatively into a 500 ml one-mark volumetric flask. Dilute to the mark with water and mix.

4.11 Iron-chloro nitric matrix solution, 20 g/l Fe.

Weigh, to the nearest 0,01 g, 20,00 g of pure iron free of titanium or of known low titanium content. Transfer into a 600 ml beaker, add 200 ml of water and then 220 ml of hydrochloric acid (4.2), in small portions. Cover with a watch glass and heat gently until effervescence ceases. Add gradually 60 ml of nitric acid (4.1). Heat for about 10 min at about 80 °C, to expel nitrous fumes.

Allow to cool and transfer quantitatively into a 1 000 ml one-mark volumetric flask. Add water until about 600 ml and add 20 g of fusion mixture (4.9). Dissolve, dilute to the mark with water and mix.

4.12 Titanium, 1 g/l standard solution.

Weigh $(1 \pm 0,001)$ g of titanium (99,99 % purity), transfer into a 400 ml beaker and cover with a watch glass. Dissolve it in 200 ml of hydrochloric acid solution (4.7) and add two drops of hydrofluoric acid (4.5).

Heat gently until the metal is dissolved. After cooling, transfer the solution quantitatively into a 1 000 ml one-mark volumetric flask, dilute to the mark with water and mix well.

1 ml of this solution contains 1 mg of titanium.

4.13 Titanium, 0,5 g/l standard solution [freshly prepared].

Transfer 50,0 ml of titanium standard solution (4.12) into a 100 ml one-mark volumetric flask. Dilute to the mark with water and mix well.

1 ml of this standard solution contains 0,5 mg of titanium.

4.14 Titanium, 0,05 g/l standard solution [freshly prepared].

Transfer 10,0 ml of titanium standard solution (4.12) into a 200 ml one-mark volumetric flask. Dilute to the mark with water and mix well.

1 ml of this standard solution contains 0,05 mg of titanium.

5 Apparatus

All volumetric glassware shall be Class A and calibrated, in accordance with EN ISO 648 or EN ISO 1042 as appropriate.

5.1 30 ml platinum crucibles.

5.2 Atomic absorption spectrometer.

This shall be equipped with a titanium hollow-cathode lamp and supplied with nitrous oxide and acetylene sufficiently pure to give a a clear, steady slightly fuel-rich flame, free from water and oil, and free from titanium.

The atomic absorption spectrometer used will be satisfactory if, after optimisation according to 7.5, the limit of detection and characteristic concentration are in reasonable agreement with the values given by the manufacturer and it meets the precision criteria given in 5.2.1 to 5.2.3.

It is also recommended that the instrument should conform to the additional performance requirements given in 5.2.4.

5.2.1 Minimum precision.

Calculate the standard deviation of 10 measurements of the absorbance of the most concentrated calibration solution. The standard deviation shall not exceed 1,5 % of the mean absorbance of this solution.

Calculate the standard deviation of 10 measurements of the absorbance of the least concentrated calibration solution (excluding the zero member). The standard deviation shall not exceed 0,5 % of the mean absorbance of the most concentrated calibration solution.

5.2.2 Limit of detection.

The limit of detection is a number, expressed in units of concentration (or amount) that describes the lowest concentration level (or amount) of an element that can be determined to be statistically different from an analytical blank.

The limit of detection of titanium in a matrix similar to that of the final test portion solution shall be better than 0,2 µg/ml for the 365,3 nm line and better than 0,1 µg/ml for the 364,3 nm line.

5.2.3 Linearity of calibration.

The value of the slope of the calibration graph covering the top 20 % of the concentration range (expressed as a change in absorbance) shall not be less than 0,7 times the value of the slope for the bottom 20 % of the concentration range (expressed as a change in absorbance) determined in the same way.

For instruments with automatic calibration using two or more standard solutions, it shall be established prior to the analysis, by obtaining absorbance readings, that the above requirements for graph linearity are fulfilled.

5.2.4 Characteristic concentration.

The characteristic concentration for titanium in a matrix similar to the final test portion solution shall be better than 1,9 μ g/ml for the 365,3 nm line and better than 1,4 μ g/ml for the 364,3 nm line.

5.3 Ancillary equipment.

A system for acquisition of the measured values, data processing and evaluation is necessary. Usually data processing and evaluation is performed with the appropriate software of the computer, connected to the atomic absorption spectrometer.

Scale expansion can be used until the noise observed is greater than the read-out error and is preferred for absorbances below 0,1. If scale expansion has to be used and the instrument does not have the means to read the value of the scale expansion factor, the value can be calculated by measuring the absorbances of a suitable solution with and without scale expansion and using the ratio of the signals obtained.

6 Sampling

Sampling shall be carried out in accordance with EN ISO 14284 or with an appropriate national standard for steels.

7 Procedure

7.1 Test portion

Weigh, to the nearest 0,001 g, 1 g of the test sample.

7.2 Blank test

In parallel with the determination and following the same procedure, carry out a blank test using the same quantities of all reagents as used for the determination in accordance with the chosen dissolution method, 7.3.1 or 7.3.2, including iron (solution 4.10 or 4.11 depending on the chosen dissolution).

7.3 Preparation of the test solution

7.3.1 Nitric perchloric dissolution

Transfer the test portion (7.1) into a 250 ml beaker. Cover the beaker with a watch glass and add 15 ml of nitric acid (4.6). When effervescence ceases, add 20 ml of perchloric acid (4.3). Heat gently until the reaction dissolution is achieved.

Continue heating at a temperature at which a steady reflux of white perchloric acid fumes is maintained on the walls of the beaker. Continue heating until there are no perchloric acid fumes visible inside the beaker and the iron perchlorate crystallises.

Allow to cool, add 25 to 30 ml of water, swirl and boil gently for a few minutes. Filter the silica through a medium-texture filter paper and collect the filtrate in a 100 ml volumetric flask. Wash the filter with hot water several times and collect the washings in the same 100 ml volumetric flask.

Allow to cool and add 5 ml of aluminium chloride solution (4.8). Dilute to the mark with water and mix.

7.3.2 Chloro-nitric dissolution with treatment of the residue

Transfer the test portion (7.1) into a 250 ml beaker. Add 20 ml of hydrochloric acid (4.7), cover the beaker with a watch glass and heat gently until effervescence ceases. Add gradually 3 ml of nitric acid (4.1). Heat for about 10 min, at about 80 °C to expel nitrous fumes. Allow to cool.

Filter through a dry filter and collect the filtrate in a 100 ml volumetric flask. Wash the residue on the filter with boiling water (about 40 ml).

Transfer the filter into a platinum crucible (5.1) and cover with a suitable platinum cover.

Place the covered crucible progressively into a muffle furnace set at 900 °C in order to ash the filter and the residue. Allow to stand at 900 °C for about 15 minutes, then remove and allow to cool. Add into the crucible 2 drops of sulphuric acid (4.4) and 2 ml of hydrofluoric acid (4.5). Evaporate to dryness on a hotplate and cool.

Add 1,0 g of the fusion mixture (4.9). Fuse the content of the crucible in a muffle furnace at 1 000 °C for 15 min to 20 min. Allow the crucible to cool and add about 10 ml of water and 1 ml of hydrochloric acid (4.2) to the solidified melt. Heat gently, in order to dissolve the fusion products. Allow the crucible to cool and add this solution quantitatively to the filtrate in the 100 ml volumetric flask. Dilute to the mark with water and mix.

7.4 Preparation of the calibration solutions

Transfer 50 ml of iron solution (4.10) or (4.11) according to the chosen method of dissolution (7.3.1 or 7.3.2) into each of a series of 100 ml volumetric flasks and add 5 ml of aluminium chloride solution (4.8). Using a burette or a pipette, add the volumes of titanium standard solutions (4.13 or 4.14) shown in Table 1. Dilute to the mark with water and mix.

The proposed calibration is given as a guide. It should be adapted in accordance with the sensitivity of the instrument used.

Table 1 — Calibration solutions

Titanium standard solution volume	Corresponding titanium mass	Iron mass	Corresponding content of titanium in the test sample
ml	mg	mg	% (<i>m/m</i>)
₀ a	0	1 000	0
2,0 (4.14)	0,10	1 000	0,010
5,0 (4.14)	0,25	1 000	0,025
10,0 (4.14)	0,50	1 000	0,050
15,0 (4.14)	0,75	1 000	0,075
20,0 (4.14)	1,00	1 000	0,100
40,0 (4.14)	2,00	1 000	0,200
6,0 (4.13)	3,00	1 000	0,300
10,0 (4.13)	5,00	1 000	0,500
14,0 (4.13)	7,00	1 000	0,700
20,0 (4.13)	10,00	1 000	1,000
a Zero member			

7.5 Adjustment of atomic absorption spectrometer

Fit the titanium hollow-cathode lamp into the atomic absorption spectrometer (5.2), switch on the current and allow it to stabilise. Adjust the wavelength in the region of 365,3 nm or 364,3 nm to minimum absorbance. Following the manufacturer's instructions, fit the correct burner, light the flame and allow the burner temperature to stabilise.

Adjust the flame to be slightly fuel-rich and the burner height to about 1 cm below the light path.

Aspirate the calibration solution containing the highest content of the analyte and adjust the burner configuration and gas flows to obtain the maximum absorption. It is important to ensure that the flow rate of the acetylene is greater than the minimum flow rate advised by the instrument manufacturer.

Evaluate the criteria of 5.2.1 to 5.2.4 to ensure that the instrument and the chosen wavelength are suitable for the determination by measuring the absorbance of the different solutions required by reference to water.

WARNING —The manufacturer's recommendations should be closely followed and particular attention is drawn to the following safety points, in particular for solutions containing perchloric acid:

- a) Aspirate all solutions for the shortest practical period;
- b) Always aspirate distilled water between each solution. Minimise the aspiration of air;
- c) Clean the burner frequently. Never allow the burner to clog and wash it thoroughly both inside and out;
- d) Ensure that all safety covers are in position and wear approved ear protectors and safety glasses.

7.6 Spectrometric measurements of the calibration solutions

Aspirate the series of calibration solutions (7.4) in succession, into the flame and measure the absorbance for each. Take care to keep the aspiration rate constant throughout the preparation of the calibration curve. Spray

water through the burner after each measurement Obtain the net mean absorbance of each calibration solution by subtracting the absorbance of the zero calibration solution.

7.7 Calibration curve

Establish the calibration curve using measured absorbances and corresponding analyte amounts, expressed as µg of titanium per ml. Use appropriate spectrometer software or an off-line computer for regression calculations or prepare a graphical representation.

7.8 Determination

7.8.1 Preliminary spectrometric measurement

Carry out an approximate measurement on the test solution (7.3) following the procedure specified in 7.5 at the same time as the spectrometric measurements are carried out on the calibration solutions (7.6). Estimate the approximate analyte amount by using the calibration curve (7.7).

7.8.2 Spectrometric measurements

Repeat the measurements and obtain the mean absorbance of each test solution and the mean absorbance of the blank solution.

Convert the mean absorbances of the test solution and of the blank solution to µg/ml of titanium by means of the calibration curve.

If pure metals and reagents have been used, the blank test and zero calibration member should give equal and negligibly small absorbances.

8 Expression of results

Calculate the titanium mass fraction, in percent (%), as follows:

$$W_{\text{Ti}} = \frac{(C_1 - C_0) \times V \times 100}{m \times 10^6} = \frac{(C_1 - C_0) \times V}{m \times 10^4}$$
(1)

where:

 W_{Ti} is the titanium mass fraction in per cent (%);

 C_1 is the titanium concentration in the test solution, in micrograms per millilitre (µg/ml);

 C_0 is the titanium concentration in the blank test solution, in micrograms per millilitre (μ g/ml);

V is the volume of the test solution (7.3). V is equal to 100 ml;

m is the mass of the test portion, in grams (g).

9 Test report

The test report shall contain the following information:

a) identification of the test sample;

- b) reference to this European Standard (EN 10211);
- c) method used;
- d) results;
- e) any unusual characteristics noted during the determination;
- f) any operation not included in this European Standard or in the document to which reference is made or regarded as optional;
- g) date of the test and/or date of preparation or signature of the test report;
- h) signature of the responsible person.

Annex A (informative)

Precision data

A.1 General

Sixteen laboratories in six European countries participated in planned trials of this method. Each laboratory carried out three determinations on each of 12 samples: four cast irons and eight steels.

Two of the three determinations were carried out under repeatability conditions as defined in ISO 5725:1986, i.e. one operator, same apparatus, identical operating conditions (same calibration) and a minimum period of time. The third determination was carried out at a different time (different day) by the same operator using the same apparatus with a new calibration.

The repeatability (r) and reproducibility (R) were calculated from the two values obtained on the first day (day 1) using the procedure specified in ISO 5725:1986. The within-laboratory reproducibility (R_w) was calculated from the first value obtained on day 1 and the value obtained on day 2.

The statistical data obtained are shown in:

- Table A.1 which gives the results obtained on the eight steel samples;
- Tables A.2 and A.3 which give the results obtained on the four cast iron samples.

Table A.1 — Results of precision tests — Dissolution 7.3.2 tested on eight steel samples

Sample label	Certified titanium value	Titanium content found (W _{Ti})	Repeatability	Reproducibility	
	m/m %	m/m %	r %	R _w %	R %
BCS 320	0,021	0,019 2	0,006 1	0,003 9	0,006 7
BCS 452/1	0,031	0,029 0	0,005 2	0,004 1	0,011 1
BCS 453/1	0,073	0,073 2	0,006 3	0,006 0	0,008 6
BCS 451/1	0,105	0,105 6	0,005 9	0,009 3	0,011 8
ECRM 281-1	0,216	0,206 8	0,009 9	0,020 7	0,049 9
IRSID 202-1	0,539	0,531 4	0,015 2	0,022 2	0,059 1
ECRM 128-1	0,890	0,893 8	0,021 4	0,043 3	0,066 7
5RA9OBrK1	1,05	0,994 0	0,015 4	0,093 0	0,186 0

Table A.2 — Results of precision tests — Dissolution 7.3.1 tested on four cast iron samples

Sample label	Certified titanium value	Titanium content found (<i>W_{Ti}</i>)	Repeatability	Reproducibility	
	m/m %	m/m %	r %	R _w %	R %
C83C	0,017	0,016 8	0,003 2	0,003 3	0,006 9
A84C (F 06-3)	0,041	0,039 9	0,002 2	0,003 8	0,003 6
A83C	0,060	0,058 4	0,003 5	0,004 5	0,004 4
C82C	0,085	0,084 4	0,001 2	0,003 7	0,006 1

NOTE 1 There were not enough results to allow the calculation of a logarithmic regression. However, an evaluation on the 0,017 % to 0,085 % content range gave mean values of 0,003 for r, 0,004 for R_w and 0,005 for R.

Table A.3 — Results of precision tests — Dissolution 7.3.2 tested on four cast iron samples

Sample label	Certified titanium value	Titanium content found (<i>W_{Ti}</i>)	Repeatability	Reproducibility	
	m/m %	m/m %	r %	R _w %	R %
C83C	0,017	0,017 3	0,001 5	0,003 3	0,002 6
A84C (F 06-3)	0,041	0,040 5	0,001 8	0,003 3	0,005 4
A83C	0,060	0,059 4	0,002 3	0,005 2	0,004 6
C82C	0,085	0,086 2	0,002 3	0,003 3	0,005 5

NOTE 2 The results shown in Table A.3 are slightly different from those obtained in accordance with dissolution 7.3.1 (see Table A.2): for the 0,017 to 0,085 % content range, the mean values for r, $R_{\rm w}$ and R are 0,002, 0,004 and 0,005 respectively.


Figure A.1 indicates the logarithmic relationships between the titanium content (W_{Ti}) and the repeatability (r) and reproducibility parameters (R_w and R) for the steel samples, plotted on a log-log scale.

The corresponding logarithmic regression formulae are:

$$\log r = 0.3339 \log W_{Ti} - 1.7630$$
 Correlation coefficient: 0,862] (A.1)

$$Log R_w = 0.731 \ 0 \ log W_{Ti} - 1.265 \ 1$$
 [Correlation coefficient: 0.924] (A.2)

$$Log R = 0.743 \ 2 log W_{Ti} - 0.977 \ 5$$
 [Correlation coefficient: 0,861] (A.3)



Key

- r repeatability
- R_w intra laboratory reproducibility
- R reproducibility
- a titanium content % (m/m)
- b precision (%)

Figure A.1 — Logarithmic relationships between titanium content, W_{T_i} , and repeatability (r) and reproducibility (R_w and R)

The predicted values for r, R_w and R derived from Formulae (A.1) to (A.3) are shown in Table A.4.

Table A.4 — Predicted values for r, R_w and R

Titanium content (W _{Ti})	Repeatability	Within-laboratory reproducibility	Reproducibility
%	r %	R _w %	R %
0,01	0,003 7	0,001 9	0,003 4
0,02	0,004 7	0,003 1	0,005 8
0,05	0,006 3	0,006 1	0,011 4
0,10	0,008 0	0,010 1	0,019 0
0,20	0,010 1	0,016 7	0,031 8
0,50	0,013 7	0,032 7	0,062 9
1,00	0,017 3	0,054 3	0,105 3

A.2 Information on repeatability and reproducibilities

A.2.1 Repeatability, r

The difference between two single results found on identical material by one analyst using the same apparatus within a short time interval will exceed the repeatability, r, not more than once in 20 cases in the normal and correct operation of the method.

A.2.2 Within-laboratory reproducibility, R_w

The difference between two single and independent results found by the same analyst working in the same laboratory on identical material on different occasions will exceed the within-laboratory reproducibility, $R_{\rm w}$, not more than once in 20 cases in the normal and correct operation of the method.

A.2.3 Reproducibility, R

The difference between two single and independent results found by two operators working in different laboratories on identical material will exceed the reproducibility, R, not more than once in 20 cases in the normal and correct operation of the method.

Annex B

(normative)

Procedures for the determination of instrumental criteria

B.1 Determination of minimum precision

Spray the most concentrated calibration solution 10 times to obtain 10 individual absorbance readings $A_{Ti,max,i}$ and calculate the corresponding mean value $A_{Ti,max}$ and standard deviation $s_{Ti,max}$.

Spray the least concentrated calibration solution (excluding the zero member) 10 times to obtain 10 individual absorbance readings $A_{Ti,min,i}$ and calculate the corresponding mean value $A_{Ti,min}$ and standard deviation $s_{Ti,min}$.

The minimum precisions of the most and least concentrated calibration solutions are obtained from $s_{Ti,max} \times 100/A_{Ti,max}$ and $s_{Ti,min} \times 100/A_{Ti,min}$, respectively.

B.2 Determination of limit of detection

Prepare two solutions each containing the same matrix concentration as the test solution, but with the element of interest at the following known concentrations:

- C_{Ti,A0.01} (μg/ml) to give an absorbance A_{Ti,A0.01} of approximately 0,01;
- the "zero member", giving an absorbance A_{Ti,0}

Alternatively, spray the "zero member" solution and the $C_{Ti,A0,01}$ solution 10 times, recording each reading for 10 s, and using sufficient scale expansion to make the fluctuations in signal clearly visible.

Calculate:

- the mean absorbance (A_{Ti,0}) and standard deviation (s_{Ti,0}) corresponding to the "zero member" solution readings;
- the mean absorbance (A_{Ti,A0,01}) corresponding to the C_{Ti,A0,01} solution readings;
- taking into account the statistical risks α and β each of them at the 95 % level (4,65), calculate the limit of detection (C_{lim}) according to the formula:

$$C_{lim} = \frac{4,65 \times ST_{i,0} \times CT_{i,A0,01}}{AT_{i,A0,01} - AT_{i,0}}$$
(B.1)

B.3 Linearity of calibration

Having established the calibration graph, before the application of any curve straightening device, obtain the net absorbance value A_A corresponding to the top 20 % of the concentration range and the net absorbance A_B corresponding to the bottom 20 % of the concentration range. Calculate A_A/A_B . This shall not be less than 0,7.

B.4 Determination of characteristic concentration, $c_{Ti,cc}$

Prepare a solution containing the same matrix concentration as the test solution, but with the element of interest at a known concentration $C_{Ti,A0,1}$ (µg/ml) to give an absorbance $A_{Ti,A0,1}$ of approximately 0,1.

Spray the $C_{Ti,A0,1}$ and the "zero member" solutions without scale expansion and measure the absorbances $A_{Ti,A0,1}$ and $A_{Ti,0}$.

The characteristic concentration $c_{\mathrm{Ti,cc}}$ is given by the formula:

$$C_{\text{Ti,cc}} = \frac{C_{\text{Ti,A0,1}} \times 0,0044}{A_{\text{Ti,A0,1}} \times A_{\text{Ti,0}}}$$
(B.2)

Bibliography

- [1] ISO 5725:1986, Precision of test methods Determination of repeatability and reproducibility for a standard test method by inter-laboratory tests
- [2] CR 10321, Chemical analysis of ferrous materials Recommendations for the drafting of standard methods of analysis employing flame atomic absorption spectrometry for the chemical analysis of iron and steel
- [3] CR 10322, Chemical analysis of ferrous materials Operational guidelines for the application of flame atomic absorption spectrometry in standard methods for the chemical analysis of iron and steel



British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards -based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email bsmusales@bsigroup.com.

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

Useful Contacts:

Customer Services

Tel: +44 845 086 9001

Email (orders): orders@bsigroup.com
Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 845 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070 Email: copyright@bsigroup.com

