

# Characterization of sludge — Determination of pH value

The European Standard EN 12176:1998 has the status of a  
British Standard

ICS 13.030.20

## National foreword

This British Standard is the English language version of EN 12176:1998.

The UK participation in its preparation was entrusted to Technical Committee EH/5, Sludge characterization, which has the responsibility to:

- aid enquirers to understand the text;
- present to the responsible European committee any enquiries on the interpretation, or proposals for change, and keep the UK interests informed;
- monitor related international and European developments and promulgate them in the UK.

A list of organizations represented on this committee can be obtained on request to its secretary.

### Cross-references

The British Standards which implement international or European publications referred to in this document may be found in the BSI Standards Catalogue under the section entitled “International Standards Correspondence Index”, or by using the “Find” facility of the BSI Standards Electronic Catalogue.

A British Standard does not purport to include all the necessary provisions of a contract. Users of British Standards are responsible for their correct application.

**Compliance with a British Standard does not of itself confer immunity from legal obligations.**

### Summary of pages

This document comprises a front cover, an inside front cover, the EN title page, pages 2 to 6, an inside back cover and a back cover.

This British Standard, having been prepared under the direction of the Health and Environment Sector Board, was published under the authority of the Standards Board and comes into effect on 15 June 1998

© BSI 1998

### Amendments issued since publication

Amd. No.	Date	Text affected

---

ICS 13.030.20

Descriptors: Sludges, sewage treatment sludges, tests, measurements, ph, samples, preparation, potentiometric methods

English version

## Characterization of sludge — Determination of pH-value

Caractérisation des boues — Détermination de la  
valeur du pH

Charakterisierung von Schlamm — Bestimmung  
des pH-Wertes

This European Standard was approved by CEN on 16 January 1998.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.

**CEN**

European Committee for Standardization  
Comité Européen de Normalisation  
Europäisches Komitee für Normung

**Central Secretariat: rue de Stassart 36, B-1050 Brussels**

## Foreword

This European Standard has been prepared by Technical Committee CEN/TC 308, Characterization of sludges, the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 1998, and conflicting national standards shall be withdrawn at the latest by August 1998.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

## Contents

	Page
Foreword	2
0 Introduction	3
1 Scope	3
2 Normative references	3
3 Definitions	3
3.1 pH value	3
3.2 Standard buffer solutions	3
4 Principle	3
5 Interferences	3
6 Hazards	4
7 Reagents	4
8 Apparatus	4
9 Preparation of samples	4
9.1 Sample preservation	4
9.2 Preparation of liquid sludge samples	4
9.3 Preparation of paste-like and solid sludges	4
10 Measurement of pH-value	4
10.1 pH meter set up and standardization	4
10.2 Measurement	4
11 Expression of results	5
12 Test report	5
Annex A (normative) Primary pH-standard reference solutions (according to ISO 10523)	6
Annex B (informative) Bibliographical references	6

## 0 Introduction

The pH value is a necessary parameter for monitoring the treatment of urban and comparable industrial sludges as well as of sludges from water supply treatment plants and to assess the suitability of the use of sludge in agriculture or for its disposal to land fill. The pH value of liquid sewage sludge may be used to monitor digester performance.

## 1 Scope

The scope of this standard is the sludges, i.e. and sludges products from:

- storm water handling;
- night soil;
- urban wastewater collecting systems;
- urban wastewater treatment plants;
- treating industrial wastewater similar to urban wastewater (as defined in Directive 91/271 EEC);
- water supply treatment plants;
- water distribution systems;
- but excluding hazardous sludges from industry.

This method is applicable to the determination of pH values of sludges which include liquid, paste-like or solid matter. If the sludge sample is sufficiently liquid, direct measurement should be made. If the sludge sample contains relatively high percentages of solids the measurement should be carried out after dilution with water.

## 2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendment to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

prEN 12880, *Characterization of sludges — Determination of dry residue and water content.*

ISO 3696, *Water for analytical laboratory use — Specification and test methods.*

ISO 10523:1994, *Water quality — Determination of pH.*

## 3 Definitions

For the purpose of this standard the following definitions apply:

### 3.1

#### pH-value

the pH value of an aqueous solution is theoretically a function of the electromotive force of a measuring cell (e.g. a glass electrode and a calomel electrode) which is defined by the equation given in annex A of ISO 10523:1994. The practical pH scale refers to exactly specified buffer solutions

### 3.2

#### standard buffer solutions

solutions of certain concentrations of pure substances or mixtures thereof with defined pH values

## 4 Principle

A suitable aqueous suspension of the solid or paste-like sludge is prepared and the pH value of the suspension is determined potentiometrically. In liquid samples the measurement is made directly.

## 5 Interferences

Substances present at their normal concentration in these types of samples do not cause significant interference. The pH value of suspensions is dependent, to a considerable extent, on the carbon dioxide equilibrium. Therefore manipulation after the sampling of liquid sludge and after the preparation of suspensions should be kept to a minimum. If oil or grease is visible in the sample, it should be removed by filtration through a pad of cotton wool (8.9) prior to carrying out the measurement. The presence of sulfides or arsenic may poison the glass interface of the electrode.

Electrodes with a low “alkali error” should be used for the measurement of sludges expected to be of a high pH.

NOTE 1 Glass electrodes normally respond to sodium ions at pH values above 9. This may cause errors in the measurement of the pH values of samples such as lime conditioned sludges.

NOTE 2 A film may form on the surface of the pH-electrodes due to oil and grease. This may lead to errors in the measuring results. Inorganic film-forming substances may also interfere if not removed from the pH-electrodes between determinations.

NOTE 3 If sludge samples are taken from a digester, digester gas may escape causing changes in the carbon dioxide content, which may lead to errors in the measured results.

## 6 Hazards

Samples of sewage sludge are liable to ferment. Do not store them in the open laboratory. If samples must be stored, then store them at (0 to 4) °C. When handling a stored sample wear gloves, face and eye protection and sufficient body protection to guard against bottles bursting. Use bottle carriers. The gas evolved is usually flammable so all equipment in the vicinity should be flame-proof and sources of ignition absent. Bursting glass bottles can produce micro-organism contaminated shrapnel. Plastic bottles can also burst and produce a hazardous spray and aerosol. Ensure that all spillages are cleared up. Do not handle broken glass directly.

There is also a danger of hydrogen sulfide formation during storage. Cap loosely to avoid pressure build-up. Sludge usually contains harmful organisms. Cleanliness is essential. Keep sludge away from food and eating. Protect all cuts.

## 7 Reagents

Analytical grade reagents shall be used.

### 7.1 Water

For the preparation of standard buffer solutions (see 3.2) freshly distilled or deionised water freshly taken from a column according to ISO 3696, Grade 3 shall be used. It shall be free from carbon dioxide and have a conductivity of  $\leq 2 \mu\text{S}/\text{cm}$ .

### 7.2 Standard buffer solutions

These shall be in accordance with ISO 10523 and annex A (normative). Such solutions are available commercially.

### 7.3 Ethanol, (CH<sub>3</sub>CH<sub>2</sub>OH), 70 % V/V.

### 7.4 Acetone, (CH<sub>3</sub>COCH<sub>3</sub>).

## 8 Apparatus

### 8.1 pH-meter

Temperature compensated input resistance  $\geq 10^{12} \Omega$  readable to at least  $\text{pH} = \pm 0,05$ , slope correction in  $\text{mV}/\text{pH}$ .

### 8.2 Glass electrode and reference electrode

The glass electrode measuring system shall have a zero voltage at  $\text{pH} = 7$ . This may consist of a glass electrode and a reference electrode, or a combination electrode system, see [2] and [3].

NOTE Some problems are associated with the reference electrodes and it is recommended that these should be tested before use. The use of a symmetrical pH-measuring system is suggested, i.e. glass and reference electrode of the same type, e.g. Ag/AgCl electrode system. Reference electrodes with gelled filling solutions should not be used.

### 8.3 Thermometer, capable of measuring to the nearest 0,5 °C.

### 8.4 Mechanical shaking apparatus

### 8.5 Sample bottles, plastics or borosilicate glass, wide mouthed.

### 8.6 Stoppered conical flask, Erlenmeyer type, 150 ml.

### 8.7 Balance, accuracy $\pm 0,01 \text{ g}$ .

### 8.8 Drying oven, (105 $\pm$ 5) °C.

### 8.9 Cotton pads

### 8.10 Soft tissue, and other usual laboratory devices.

## 9 Preparation of samples

### 9.1 Sample preservation

Sludge samples can change composition through biological or chemical activity. Therefore, they should be analysed as soon as possible after sampling. Liquid sludge samples will lose carbon dioxide on exposure to air. They shall therefore be kept in a closed partially filled container until the pH value is determined. Handling should be minimized to avoid loss of carbon dioxide. If prolonged storage of sludge samples is unavoidable they should be stored at (0 to 4) °C in a refrigerator.

### 9.2 Preparation of liquid sludge samples

These samples usually contain less than 50 g dry matter per kg sludge and do not require any preparation before pH-measurement.

NOTE If visible oil or grease is present it should be removed by filtration through a cotton wool pad prior to step 10.1. Use the filtrate at step 10.2.

### 9.3 Preparation of paste-like and solid sludges

Weigh an amount of sample equivalent to approximately 5,0 g dry matter into a conical flask (8.6). Add water (7.1) to make the total weight (100  $\pm$  1) g. Stopper the flask and shake mechanically (8.4) for at least 15 min, until the sample is thoroughly dispersed. In some cases it may be necessary to break down the solids prior to shaking. Proceed as in clause 10.

NOTE 1 Whilst it is preferable that precise quantities of sludge (expressed as dry matter) are used, this may not always be possible. Estimates of the dry matter content may be made in the interests of speed.

NOTE 2 Determine the dry solids content of the sludge in accordance with prEN 12880 by drying the sample at (105  $\pm$  5) °C.

NOTE 3 If visible oil or grease is present after shaking, follow the note given in 9.2.

## 10 Measurement of pH value

### 10.1 pH-meter set up and standardization

Measure the temperature of the sample and set the temperature control on the pH-meter (8.1) following the manufacturer's instructions. Standardize the pH-meter using standard pH-buffer solutions (7.2).

NOTE The range of pH values of the standard buffer solutions preferably includes that of the sample.

### 10.2 Measurement

Rinse the electrodes (8.2) thoroughly with water (7.1). Ensure that the sludge sample is well dispersed by swirling. Immediately, place the electrodes in the sample. Read the steady pH value from the pH-meter (8.1) after approximately 30 s to 60 s. Remove the electrodes and rinse with water (7.1).

At the end of the batch of samples clean the electrodes with a soft wet tissue (8.10) and rinse the electrodes thoroughly with water. If the samples contained organic impurities use ethanol (70 %) (7.3), acetone (7.4) or warm detergent solution for cleaning.

NOTE 1 Only use organic solvents for a short time as they decrease the sensitivity of glass electrodes by dewatering the leached layer of the glass membrane.

NOTE 2 If the suspended phase separates from the liquid phase, it is preferable that the electrodes are placed in the liquid phase.

## 11 Expression of results

The pH value shall be rounded to one decimal place.  
In addition, the temperature to which the pH-meter is compensated shall be specified.

### EXAMPLES

pH value at 25 °C:	6,3
pH value in the sludge suspension at 20 °C:	8,8
pH value in the filtrate at 20 °C:	8,4

## 12 Test report

The test report shall contain the following information:

- reference to this European Standard;
- all information necessary for complete identification of the sludge sample;
- details of sample pretreatment and preparation;
- results of the determination according to clause 11;
- any details not specified in this European Standard or which are optional and any other factors which may have affected the results.

Annex A (normative)

Primary pH-standard reference solutions (according to ISO 10523)

Table A.1 — pH values of standard buffer solutions

Temperature °C	Primary pH-standard buffer solutions				
	B	C	D	F	I
	Potassium hydrogen tartrate	Potassium hydrogen phthalate <sup>1)</sup>	Phosphate	Borax	Sodium carbonate/sodium hydrogen carbonate
0	—	4,000	6,984	9,464	10,317
5	—	3,998	6,951	9,395	10,245
10	—	3,997	6,923	9,332	10,179
15	—	3,998	6,900	9,276	10,118
20	—	4,001	6,881	9,225	10,062
25	3,557	4,005	6,865	9,180	10,012
30	3,552	4,011	6,835	9,139	9,966
35	3,549	4,018	6,844	9,102	9,926
40	3,547	4,027	6,838	9,068	9,889
45	3,547	4,038	6,834	9,038	9,856
50	3,549	4,050	6,833	9,011	9,828
Buffer value	0,027	0,016	0,029	0,020	0,029
$\Delta\text{pH}_{1/2}$	+ 0,049	+ 0,052	+ 0,080	+ 0,01	+ 0,079

<sup>1)</sup> This solution is called "reference value pH-standard".

Preparation of the standard solutions. Always use water free from carbon dioxide and:

B: Potassium hydrogen tartrate, saturated at  $(25 \pm 2)$  °C;

C: Potassium hydrogen phthalate,  $b = 0,05$  mol/l;

Dissolve 10,21 g of  $\text{C}_6\text{H}_4(\text{COOH})(\text{COOK})$ , that has been dried for 2 h at  $(120 \pm 5)$  °C, in water at  $(25 \pm 2)$  °C and bring to volume with water in a 1 l volumetric flask.

D: Disodium hydrogen phosphate,  $b = 0,025$  mol/l and potassium dihydrogen phosphate,  $b = 0,025$  mol/l.

Dissolve 4,45 g of  $\text{Na}_2\text{HPO}_4 \cdot 2 \text{H}_2\text{O}$  (or 3,55 g of  $\text{Na}_2\text{HPO}_4$ ) and 3,40 g of  $\text{KH}_2\text{PO}_4$  in water at  $(25 \pm 2)$  °C and bring to volume with water. The anhydrous salts are dried for 2 h at  $(120 \pm 5)$  °C;

F: Sodium tetraborate (borax),  $b = 0,01$  mol/l.

Dissolve 3,81 g of  $\text{Na}_2\text{B}_4\text{O}_7 \cdot 10 \text{H}_2\text{O}$  in water at  $(25 \pm 2)$  °C and bring to 1 l with water;

I: Sodium carbonate,  $b = 0,025$  mol/l, and sodium hydrogen carbonate,  $b = 0,025$  mol/l.

Dissolve 2,640 g of  $\text{Na}_2\text{CO}_3$ , that has been dried for 90 min at  $(250 \pm 5)$  °C and 2,092 g of  $\text{NaHCO}_3$ , that has been dried for 2 days  $\pm 2$  h over a molecular sieve, in water and bring to 1 l with water.

Annex B (informative)

Bibliographical references

- [1] NF X 31-117:1994, *Qualité du sol — Détermination du pH*
- [2] DIN 19263, *pH-measurement — Glass electrodes*
- [3] DIN 19264, *pH-measurement — Reference electrodes*
- [4] DIN 19266, *pH-measurement — Standard buffer solutions*
- [5] BS 1647-2, *Specification for reference value standard solutions and operational reference standard solutions*
- [6] ISO 10390:1994, *Soil quality — Determination of pH*





---

---

# BSI — British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

## Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: 020 8996 9000. Fax: 020 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

## Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: 020 8996 9001. Fax: 020 8996 7001.

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

## Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact the Information Centre. Tel: 020 8996 7111. Fax: 020 8996 7048.

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: 020 8996 7002. Fax: 020 8996 7001.

## Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

If permission is granted, the terms may include royalty payments or a licensing agreement. Details and advice can be obtained from the Copyright Manager. Tel: 020 8996 7070.