



BSI Standards Publication

Fertilizers — Determination of chelating agents in fertilizers by chromatography

Part 2: Determination of Fe chelated by o,o-EDDHA, o,o-EDDHMA and HBED by ion pair chromatography

National foreword

This British Standard is the UK implementation of EN 13368-2:2012. It supersedes BS EN 13368-2:2007 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee CII/37, Fertilisers and related chemicals.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2012. Published by BSI Standards Limited 2012

ISBN 978 0 580 73762 6

ICS 65.080

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 August 2012.

Amendments issued since publication

Date	Text affected
------	---------------

EUROPEAN STANDARD

EN 13368-2

NORME EUROPÉENNE

EUROPÄISCHE NORM

August 2012

ICS 65.080

Supersedes EN 13368-2:2007

English Version

Fertilizers - Determination of chelating agents in fertilizers by chromatography - Part 2: Determination of Fe chelated by o,o-EDDHA, o,o-EDDHMA and HBED by ion pair chromatography

Engrais - Détermination des agents chélatants dans les engrais par chromatographie - Partie 2 : Détermination du fer chélaté o,o-EDDHA, o,o-EDDHMA et HBED par chromatographie d'appariement d'ions

Düngemittel - Bestimmung von Chelatbildnern in Düngemitteln mit Chromatographie - Teil 2: Bestimmung von Fe chelatisiertem o,o-EDDHA, o,o-EDDHMA und HBED mit Ionen-Paarchromatographie

This European Standard was approved by CEN on 6 July 2012.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
1 Scope	4
2 Normative references	4
3 Terms and definitions	4
4 Principle.....	4
5 Interferences	5
6 Reagents.....	5
7 Apparatus	6
8 Sampling and sample preparation	7
9 Procedure	7
9.1 Preparation of the sample solution.....	7
9.2 Preparation of the calibration solutions.....	7
9.3 Chromatographic analysis.....	8
10 Expression of the results (Fe in Fe-chelates)	13
10.1 Fe chelated by [o,o] EDDHA or [o,o] EDDHMA.....	13
10.2 Fe chelated by HBED.....	13
11 Precision	13
11.1 Inter-laboratory test.....	13
11.2 Repeatability.....	14
11.3 Reproducibility.....	14
12 Test report	14
Annex A (informative) Statistical results of the inter-laboratory tests	16
A.1 Statistical results of the inter-laboratory test performed in 2003.....	16
A.1.1 Test samples	16
A.1.2 Inter-laboratory test procedure	16
A.1.3 Results and statistical interpretation.....	16
A.2 Results of the inter-laboratory test performed in 2010.....	17
A.2.1 Test samples	17
A.2.2 Inter-laboratory test procedure	17
A.2.3 Results and statistical interpretation.....	18
Annex B (informative) General procedure for the determination of the titrimetric purity of the chelating agents using a photometric automatic titrator	19
Annex C (informative) Complete names of chelating agents	20
Bibliography	21

Foreword

This document (EN 13368-2:2012) has been prepared by Technical Committee CEN/TC 260 "Fertilizers and liming materials", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2013, and conflicting national standards shall be withdrawn at the latest by February 2013.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 13368-2:2007.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

The following has been added to the former edition (EN 13368-2:2007) of this European Standard:

- a) new substance Fe-HBED added to the scope and to the title;
- b) requirements concerning the use of different columns in order to avoid overlapping of the *meso* isomer of Fe-[*o,o*] EDDHMA and Fe-HBED peaks added in Clause 5, Interferences;
- c) iron-nitrate solution and Fe-HBED solution added in Clause 6, Reagents;
- d) the sampling method is not part of the standard; informative reference to EN 1482-1 added;
- e) Table 1 enlarged by a column for total chelate;
- f) typical chromatogram of a commercial Fe-HBED sample added (Figure 4);
- g) Formula (2) on calculation of Fe-HBED added;
- h) Table 2 enlarged by the precision data concerning determination of Fe-HBED;
- i) A.2: Results of the inter-laboratory test performed in 2010 added;
- j) Annex C: Complete names of chelating agents technical revised;
- k) Bibliography revised.

This European Standard "*Fertilizers — Determination of chelating agents in fertilizers by ion chromatography*" consists of the following parts:

Part 1: Determination of EDTA, HEEDTA and DTPA by ion chromatography

Part 2: Determination of Fe chelated by o,o-EDDHA, o,o-EDDHMA and HBED by ion pair chromatography

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

1 Scope

This European Standard specifies a method for the chromatographic determination of the iron chelated by each individual *ortho*(hydroxy)-*ortho*(hydroxy) isomer of the chelating agents [o,o] EDDHA, [o,o] EDDHMA and by HBED in fertilizers containing one or more of these substances, except for [o,o] EDDHMA and HBED mixes. The method allows the identification and the determination of the total concentration of water soluble iron chelates of these chelating agents. It does not determine the free form of the chelating agents.

This method is applicable to EC fertilizers covered by Regulation (EC) No 2003/2003 [4]. It is applicable to a mass fraction of the metal chelated of at least 0,625 %.

NOTE 1 The substances EDDHA (ethylenediamine-N,N'-di[(hydroxyphenyl)acetic acid] and EDDHMA (ethylenediamine-N,N'-di[(hydroxymethylphenyl)acetic acid] exist as several different isomeric forms. Positional isomers for the hydroxyl or methyl groups (in *ortho*, *meta*, and *para* positions) as well as stereo isomers (*meso* and dl-racemic forms) are known. Both *meso* and dl-racemic forms of the *ortho,ortho*-EDDHA and *ortho,ortho*-EDDHMA are positional isomers for the hydroxyl groups allowed by the Regulation (EC) No 2003/2003. Since *para*, *meta* and *ortho* methyl positional isomers of the EDDHMA present quite similar stability, they could be grouped: in the method here described the *para*, *meta* and *ortho* methyl positional isomers of the [o,o] EDDHMA are considered together. HBED (N,N'-bis(2-hydroxybenzyl)-ethylenediamine-N,N'-diacetic acid) does not present isomeric forms.

NOTE 2 At present, analytically pure standards only exist for *ortho,ortho*-EDDHA, *ortho,ortho*-EDDHMA and HBED. All other substances being unavailable as a standard, the influence of their eventual presence in the samples (with respect to the sensitivity and the selectivity of this method) has not been studied.

NOTE 3 The *meso* and the dl-racemic forms of [o,o] EDDHA and [o,o] EDDHMA can be determined separately by this method.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 12944-1:1999, *Fertilizers and liming materials and soil improvers — Vocabulary — Part 1: General terms*

EN 12944-2:1999, *Fertilizers and liming materials and soil improvers — Vocabulary — Part 2: Terms relating to fertilizers*

EN 1482-2, *Fertilizers and liming materials — Sampling and sample preparation — Part 2: Sample preparation*

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 12944-1:1999 and EN 12944-2:1999 apply.

4 Principle

The iron chelates are separated and determined by isocratic ion-pair high-performance liquid chromatography. When an iron chelate (anion) is added to a polar fluid (eluent), containing a large cation, an ion pair is formed. This ion pair is retained by an apolar solid phase (stationary phase). The strength of the retention depends on the molecular size and its acidity. Then, each iron chelate presents a characteristic retention time depending on the chelating agent, and it is separated from the other substances present in the sample. The separation is carried out on a reverse phase silica column and an aqueous solution of TBA⁺ (tetrabutylammonium) and acetonitrile as eluent. The detection is based on photometry at 280 nm.

NOTE For additional information, see [5], [6] and [7].

5 Interferences

No interferences have been detected. Iron chelates with HBEP, EDDHSA, EDTA, DTPA, CDTA, HEEDTA, p,p-EDDHA, o,p-EDDHA, as well as the chelating agents do not interfere since they are separated from Fe-[o,o] EDDHA, Fe-[o,o] EDDHMA or Fe-HBED. Fe-[o,o] EDDHA does not interfere with Fe-[o,o] EDDHMA or Fe-HBED. In the case that Fe-[o,o] EDDHMA and Fe-HBED are present in the same sample, an overlapping of the *meso* isomer of the Fe-[o,o] EDDHMA and Fe-HBED peaks may occur with some equipment depending on the column used. The use of a different column of the same type can solve this problem.

NOTE For the complete names of the abbreviations of chelating agents used in this document, see Annex C.

6 Reagents

6.1 General

All reagents shall be of recognised analytical grade. All water used for the preparation of eluents, standards, and sample solutions shall conform to EN ISO 3696, grade 1 and shall be degassed, and free of organic contaminants. If products with a declared purity of less than 99 % are used for the preparation of standard solutions, a correction should be made in order to obtain exactly the required concentration in the solution.

If there is any doubt of the purity of the standard, it is necessary to determine it.

NOTE For this determination, a titrimetric method can be used. See Annex B for a general method using an automatic titrator. Manual titration could also be adequate.

6.2 Sodium hydroxide solution, $c(\text{NaOH}) = 0,1 \text{ mol/l}$.

Dissolve 4 g of NaOH in pellet form in a 1 l volumetric flask with water (6.1). Dilute to the mark and homogenise.

NOTE The incorporation of CO_2 from the atmosphere should be carefully avoided. Otherwise the dissolution of chelating agents (see 6.6, 6.7 and 6.8) can be incomplete.

6.3 Hydrochloric acid solution, $c(\text{HCl}) = 1,0 \text{ mol/l}$.

Dilute 88 ml of hydrochloric acid (mass fraction 35 % HCl) to 1 000 ml with water.

6.4 Hydrochloric acid solution, $c(\text{HCl}) = 0,1 \text{ mol/l}$.

Dilute 50 ml of hydrochloric acid 1,0 mol/l (6.3) to 500 ml with water.

6.5 Iron-nitrate solution, $\rho(\text{Fe}) = 1\,050 \text{ mg/l}$.

Dissolve 0,759 4 g of ferric nitrate 9-hydrate ($\text{Fe}(\text{NO}_3)_3 \cdot 9\text{H}_2\text{O}$) in 100 ml of water. Check (for example by AAS) that the Fe concentration in this solution is $1\,050 \text{ mg} \pm 30 \text{ mg/l}$.

NOTE As the $\text{Fe}(\text{NO}_3)_3 \cdot 9\text{H}_2\text{O}$ is deliquescent it will be added in solution of a known concentration.

6.6 Fe-[o,o] EDDHA solution, $\rho(\text{Fe}) = 100 \text{ mg/l}$.

Dissolve 0,322 1 g (see 6.1) of ethylenediamine-N,N'-di[*ortho*-hydroxyphenyl]acetic acid] in 350 ml of water and 27 ml of NaOH (6.2) in a 500 ml beaker. Add 50 ml of the Fe solution (6.5) to the chelating agent solution, stirring for about 5 min.

Adjust the solution to pH 7,0 with NaOH solution (6.2). Let the solution stand overnight in darkness to allow excess Fe to precipitate as oxide. Filter quantitatively through a cellulose filter and make up to volume (500 ml) in a volumetric flask with water. The standard obtained in this way can be stored in darkness for one year.

6.7 Fe-[o,o] EDDHMA solution, $\rho(\text{Fe}) = 100 \text{ mg/l}$.

Dissolve 0,347 1 g (see 6.1) of ethylenediamine-N,N'-di[(*ortho*-hydroxy-*para*-methylphenyl)acetic acid] [the *paramethyl* isomer of [o,o] EDDHMA, (see Note 1 in Clause 1)] in 350 ml of water and 27 ml of NaOH (6.2) in a 500 ml beaker. Add 50 ml of the Fe solution (6.5) to the chelating agent solution, stirring for about 5 min.

Adjust the solution to pH 7,0 with NaOH solution (6.2). Let the solution stand overnight in darkness to allow excess Fe to precipitate as oxide. Filter quantitatively through a cellulose filter and make up to volume (500 ml) in a volumetric flask with water. The standard obtained in this way can be stored in darkness for one year.

6.8 Fe-HBED solution, $\rho(\text{Fe}) = 100 \text{ mg/l}$.

Dissolve 0,347 1 g (see 6.1) of N,N'-bis(2-hydroxybenzyl)-ethylenediamine-N,N'-diacetic acid (HBED) or 0,379 7 g if HBED·HCl is used in 350 ml of water and 27 ml of NaOH (6.2) in a 500 ml beaker. Add 50 ml of the Fe solution (6.5) to the chelating agent solution, stirring for about 5 min.

Adjust the solution to pH 5,0 with HCl solution (6.4) or NaOH solution (6.2). Let the solution stand overnight in darkness to allow excess Fe to precipitate as oxide. Filter quantitatively through a cellulose filter and make up to volume (500 ml) in a volumetric flask with water. The standard obtained in this way can be stored in darkness for one year.

6.9 Eluent for the determination

Add 20 ml of TBAOH (mass fraction 40 % Tetrabutylammonium hydroxide solution in water) to 600 ml of water. Adjust to pH 6,0 with hydrochloric acid solution (6.3 and 6.4). Add 300 ml of acetonitrile (HPLC grade) and make up to volume in a 1 l volumetric flask with water. Filter through a 0,2 μm membrane filter (7.4 b)) and degas.

NOTE TBACl or TBABr can be used, providing that pH is adjusted to 6,0 with NaOH or HCl.

7 Apparatus

Usual laboratory equipment, glassware and the following:

7.1 Magnetic stirrer.

7.2 Chromatograph, equipped with:

- a) an isocratic pump delivering the eluent at a flow rate of 1,5 ml/min;
- b) an injection valve with a 20 μl injection loop;
- c) a C-18 column; internal diameter :3,9 mm; column length: 150 mm; $\text{dp} = 5 \mu\text{m}^1$);

1) SYMMETRY™ C18, from WATERS, LiChroCART® Purospher® RP-18, from MERCK or equivalent are examples of suitable products available commercially. This information is given for the convenience of users of this European Standard and does not constitute an endorsement by CEN of these products.

- d) a C-18 guard column (recommended);
- e) a UV/VIS-detector with a 280 nm-filter;
- f) an integrator.

7.3 Balance, capable of weighing to an accuracy of 0,1 mg.

7.4 Membrane filters, including:

- a) Micro membrane filters resistant to aqueous solutions, with porosity of 0,45 μm .
- b) Micro membrane filters resistant to organic solutions (e.g. polyamide 66 micro membrane filters), with porosity of 0,2 μm .

8 Sampling and sample preparation

Sampling is not part of the method specified in this document. A recommended sampling method is given in EN 1482-1.

Sample preparation shall be carried out in accordance with EN 1482-2.

For the size reduction of samples with a high amount of chelating agents, it is not recommended to use a high speed laboratory mill. It is more convenient to grind the sample to a particle size less than 1 mm.

9 Procedure

9.1 Preparation of the sample solution

Weigh, to the nearest 0,1 mg, 200 mg of the iron chelate into a 250 ml beaker. Add 200 ml of water. Stir using a magnetic stirrer for 1 h. Transfer quantitatively into a 250 ml volumetric flask. Dilute to the mark with water and homogenise.

For samples declaring more than 6 % of chelated Fe, use a 500 ml volumetric flask.

For liquid samples, weigh an amount equivalent to 200 mg of solid.

9.2 Preparation of the calibration solutions

Pipette V ml (see Table 1) of the Fe-[o,o] EDDHA (6.6) or Fe-[o,o] EDDHMA (6.7) or Fe-HBED (6.8) standard solution into six 100 ml volumetric flasks. Make up to volume with water and homogenise.

Table 1 — Composition of the calibration solutions

Solution	V ml	Fe concentration in the <u>rac</u> chelate mg/l	Fe concentration in the <i>meso</i> chelate mg/l	Total Fe concentration in chelate mg/l
1	5	2,5	2,5	5,0
2	10	5,0	5,0	10,0
3	20	10,0	10,0	20,0
4	30	15,0	15,0	30,0
5	40	20,0	20,0	40,0
6	50	25,0	25,0	50,0

NOTE 1 As the standard chelates solutions should be 50 % *meso*- and 50 % dl-racemic optical isomers, the concentration of the total iron chelate is the sum of the concentrations shown in Table 1.

NOTE 2 Since HBED does not present optical isomers the total concentrations should be used.

9.3 Chromatographic analysis


Immediately before injection, all solutions shall be filtered through a 0,45 µm membrane filter (7.4 a). Inject the standard solutions (see 9.2) into the chromatographic system (7.2). Measure the retention times and the areas of the two Fe-[*o,o*] EDDHA or Fe-[*o,o*] EDDHMA isomers (*meso* and dl-racemic) or Fe-HBED peaks for all solutions. For each appropriate iron chelate, draw two calibration graphs with the values of the peak areas of the standard solutions versus the iron concentration (mg/l) in the form of the corresponding chelating agent isomer. For Fe-HBED, only one graph is needed. See Figures 1 to 4.

NOTE 1 See NOTE 1 in 9.2.

NOTE 2 A statistically significant negative intercept can be caused by the contamination of the column with Fe oxide precipitates. In that case, the column should be replaced or cleaned.

Inject the sample solution (see 9.1). Identify the chelating agent by the retention time of the obtained peaks (see Figures 2, 3 and 4). Measure the areas of the isomer peaks for each chelating agent. Determine the concentration of the iron chelated by each optical isomer (mg Fe/l) using the corresponding calibration graph.


In Figure 3 the typical chromatogram of a commercial product containing Fe-[*o,o*] EDDHMA is shown. Besides *meso* and dl-racemic isomers of para-methyl ortho-hydroxyl Fe-[*o,o*] EDDHMA two other isomers, assigned to other methyl positional isomers, are found (peaks 3 and 4). Since they have similar stability to the para-methyl isomers (see NOTE 1 in Clause 1), their area should be added to the nearest (1 or 2) peak area.



Key

- 1 dl-racemic Fe-[o,o] EDDHA
- 2 meso Fe-[o,o] EDDHA
- X time in minutes


Figure 1 — Typical chromatogram of a Fe-[o,o] EDDHA standard solution with an Fe content of 25 mg/l /l



Key

- 1 dl-racemic Fe-[o,o] EDDHA
- 2 meso Fe-[o,o] EDDHA
- 3 ortho-hydroxy para-hydroxy isomer
- 4 para-hydroxy para-hydroxy isomer by-product
- X time in minutes


Figure 2 — Typical chromatogram of a commercial Fe-[o,o] EDDHA product



Key

- 1 dl-racemic Fe-[o,o] EDDHMA
- 2 meso Fe-[o,o] EDDHMA
- 3 and 4 methyl positional isomers of Fe-[o,o] EDDHMA
- X time in minutes

Figure 3 — Typical chromatogram of a commercial Fe-[o,o] EDDHMA product




Key

X time in minutes
Y absorbance units

Figure 4 — Typical chromatogram of a commercial Fe-HBED sample

If overlapping occurs, baseline correction in the integration should be done (see Figure 5).



Key

X time in minutes
Y absorbance units

Figure 5 — Example of baseline correction

10 Expression of the results (Fe in Fe-chelates)

10.1 Fe chelated by [o,o] EDDHA or [o,o] EDDHMA

Calculate, w_{Fe} , the mass fraction in percent of the sample of Fe chelated by [o,o] EDDHA or [o,o] EDDHMA in the fertilizer according to the following formula:

$$w_{\text{Fe}} = \frac{c_{\text{rac}} + c_{\text{m}}}{m} \times V \times 10^{-4} \quad (1)$$

where

c_{rac} and c_{m} are the Fe concentrations of the Fe chelated by the racemic or *meso* isomers respectively determined with the calibration graphs, in milligrams per litre;

m is the mass of the sample taken for analysis, in grams;

V is the volume of the sample solution (9.1), in millilitres.

If the sample contains both Fe-[o,o] EDDHA and Fe-[o,o] EDDHMA, the Fe chelated is the addition of the results obtained for each chelating agent.

10.2 Fe chelated by HBED

Calculate the mass fraction, w_{Fe} of Fe chelated by HBED in the fertilizer in percent of the sample according to the following formula:

$$w_{\text{Fe}} = \frac{c}{m} \times V \times 10^{-4} \quad (2)$$

where

c is the Fe concentration of the Fe chelated by HBED determined with the calibration graphs, in milligrams per litre;

m is the mass of the sample taken for analysis, in grams;

V is the volume of the sample solution (9.1), in millilitres.

11 Precision

11.1 Inter-laboratory test

An inter-laboratory test has been carried out in 2003 with fourteen participating laboratories and seven different samples (two standard solutions, two fertilizers containing [o,o] EDDHMA and three fertilizers containing [o,o] EDDHA as chelating agents). This test yielded the data given in CEN/TR 15106 and is summarised in A.1. Repeatability and reproducibility were calculated according to ISO 5725-2.

A new inter-laboratory test has been carried out in 2010 with ten participating laboratories and three different samples (one liquid solution and two solid fertilizers containing HBED as chelating agent). This test yielded the data is summarised in A.2. Repeatability and reproducibility were calculated according to ISO 5725-2.

The values derived from these inter-laboratory tests might not be applicable to concentration ranges and matrices other than those given in A.1 and A.2.

11.2 Repeatability

The absolute difference between two independent test results obtained with the same method on identical test material, in the same laboratory, by the same operator, using the same equipment within a short interval of time, will in not more than 5 % of the cases exceed the values of r given in Table 2.

11.3 Reproducibility

The absolute difference between two single test results obtained with the same method on identical test material, in different laboratories, by different operators, using different equipment, will in not more than 5 % of the cases exceed the values of R given in Table 2.

Table 2 — Mean values, repeatability and reproducibility limits

Sample	Mean value	Repeatability limit r	Reproducibility limit R
Solutions from standards			
EDDHA	24,91 mg/l	0,29 mg/l	1,81 mg/l
EDDHMA	25,42 mg/l	0,26 mg/l	2,07 mg/l
Solution from commercial fertilizers			
HBED	42,20 mg/l	0,21 mg/l	0,36 mg/l
Commercial fertilizers (Unit in Fe concentrations)			
EDDHA-1	3,08 g/100 g	0,05 g/100 g	0,38 g/100 g
EDDHA-2	4,98 g/100 g	0,09 g/100 g	0,43 g/100 g
EDDHA-3	3,80 g/100 g	0,07 g/100 g	0,36 g/100 g
EDDHMA-1	4,84 g/100 g	0,12 g/100 g	0,69 g/100 g
EDDHMA-2	4,22 g/100 g	0,04 g/100 g	0,62 g/100 g
HBED-1	10,57 g/100 g	0,24 g/100 g	0,36 g/100 g
HBED-2	6,56 g/100 g	0,21 g/100 g	0,28 g/100 g

12 Test report

The test report shall contain at least the following information:

- all information necessary for the complete identification of the sample;
- test method used with reference to this document, i.e. EN 13368-2;
- test results obtained expressed as percentage of Fe chelated by [o,o] EDDHA, [o,o] EDDHMA or HBED in the fertilizer;
- date of the sampling and sampling procedure (if known);
- date when the analysis was finished;

- f) whether the requirement of the repeatability limit has been fulfilled;
- g) all operating details not specified in this document, or regarded as optional, together with details of any incidents occurred when performing the method which might have influenced the test result(s).

Annex A (informative)

Statistical results of the inter-laboratory tests

A.1 Statistical results of the inter-laboratory test performed in 2003

A.1.1 Test samples

Seven different samples were provided to all the participants. Five samples were commercial iron chelates in solid form, three containing Fe-[o,o] EDDHA and two containing Fe-[o,o] EDDHMA. The other two samples were solutions prepared from standards of commercial sources, one containing Fe-[o,o] EDDHA and the other containing Fe-[o,o] EDDHMA.

A.1.2 Inter-laboratory test procedure

The test samples were sent to sixteen laboratories from eight countries, private ones as well as official ones, but only fourteen provided results.

The participating laboratories were requested to perform two replicates of each sample. The results of the two replicates of each sample were recorded to two decimal places.

Test results, observations and remarks were reported.

A.1.3 Results and statistical interpretation

Statistical calculations were run on all the results, according to ISO 5725-2.

Parameters of repeatability and reproducibility were evaluated for each sample (mean value, standard deviation of repeatability, standard deviation of reproducibility, repeatability, reproducibility, relative standard deviation of repeatability and relative standard deviation of reproducibility).

Table A.1 shows the statistical results of the inter-laboratory test.

Table A.1 — Statistical results of the inter-laboratory test

Parameter	Solutions from Standard		Commercial iron chelates (solid form)				
	EDDHA	EDDHMA	EDDHA-1	EDDHA-2	EDDHA-3	EDDHMA-1	EDDHMA-2
Number of laboratories	14	12	14	14	14	12	12
Number of outliers	0	2	0	1	0	0	1
Number of laboratories after elimination of outliers	14	10	14	13	14	12	11
Mean value (Fe concentrations)	24,91 mg/l	25,42 mg/l	3,08 g/100 g	4,98 g/100 g	3,80 g/100 g	4,84 g/100 g	4,22 g/100 g
Repeatability limit, r (Fe concentrations)	0,29 mg/l	0,26 mg/l	0,05 g/100 g	0,09 g/100 g	0,07 g/100 g	0,12 g/100 g	0,04 g/100 g
RSD_r %	0,42	0,37	0,57	0,65	0,61	0,89	0,38
Repeatability standard deviation s_r (Fe concentrations)	0,10 mg/l	0,09 mg/l	0,02 g/100 g	0,03 g/100 g	0,02 g/100 g	0,04 g/100 g	0,02 g/100 g
Reproducibility limit R (Fe concentrations)	1,81 mg/l	2,07 mg/l	0,38 g/100 g	0,43 g/100 g	0,36 g/100 g	0,69 g/100 g	0,62 g/100 g
RSD_R %	2,60	2,90	4,39	3,05	3,35	5,12	5,24
Reproducibility standard deviation s_R (Fe concentrations)	0,65 mg/l	0,74 mg/l	0,14 g/100 g	0,15 g/100 g	0,13 g/100 g	0,25 g/100 g	0,22 g/100 g
Horwitz Value R	9,9	9,8	3,4	3,1	3,3	3,2	3,2
Horrat R -index	0,26	0,30	1,30	0,97	1,02	1,62	1,63

A.2 Results of the inter-laboratory test performed in 2010

A.2.1 Test samples

Three different Fe-HBED samples were provided to all the participants. Two samples were commercial iron chelates in solid form. The other sample was a solution prepared dissolving a commercial fertilizer.

A.2.2 Inter-laboratory test procedure

The test samples were sent to ten laboratories from five countries, private ones as well as official ones and all of them provided results.

The participating laboratories were requested to perform three replicates of each sample. The results of the three replicates of each sample were recorded to two decimal places.

Test results, observations and remarks were reported.

A.2.3 Results and statistical interpretation

Statistical calculations were run on all the results, according to ISO 5725-2.

Parameters of repeatability and reproducibility were evaluated for each sample (mean value, standard deviation of repeatability, standard deviation of reproducibility, repeatability, reproducibility, relative standard deviation of repeatability and relative standard deviation of reproducibility).

Table A.2 shows the statistical results of the inter-laboratory test.

Table A.2 — Statistical results of the new inter-laboratory test

Parameter	Solution from a commercial fertilizer	Commercial iron chelates (solid form)	
	HBED	HBED-1	HBED-2
Sample			
Number of laboratories	10	10	10
Number of outliers	3	1	2
Number of laboratories after elimination of outliers	7	9	8
Mean value (Fe concentrations)	42,20 mg/l	10,57 g/100 g	6,56 g/100 g
Repeatability limit, r (Fe concentrations)	0,21 mg/l	0,24 g/100 g	0,21 g/100 g
RSD_r %	0,18	0,81	1,14
Repeatability standard deviation s_r (Fe concentrations)	0,08 mg/l	0,09 g/100 g	0,08 g/100 g
Reproducibility limit R (Fe concentrations)	0,36 mg/l	0,36 g/100 g	0,28 g/100 g
RSD_R %	0,31	1,21	1,52
Reproducibility standard deviation s_R (Fe concentrations)	0,13 mg/l	0,13 g/100 g	0,10 g/100 g
Horrat R -index	0,03	0,43	0,50

Annex B (informative)

General procedure for the determination of the titrimetric purity of the chelating agents using a photometric automatic titrator

Chelating agent solutions of approximately $1,00 \times 10^{-4}$ mol/l are prepared by dissolving the calculated amount of chelating agent in the necessary amount of NaOH solution. Then pH is adjusted to 6 by the addition of $2,00 \times 10^{-3}$ mol/l MES buffer [2-(N-morpholino) ethanesulfonic acid] and the volume made up to 50 ml or 100 ml.

The test solution is placed in a thermostated jacketed reaction vessel (100 ml or 150 ml capacity) with an airtight cap fitted with gas inlet and outlet tubes, the combined pH glass electrode, the photometric detector and two piston burettes, tip extending below the surface of the solution, and a magnetic stirrer. The photometric titration consists of the addition of $4,47 \times 10^{-4}$ mol/l Fe(III) standard solution to the chelating agent until the absorbance at 480 nm shows no change. End-point titration is calculated for two linear segments' intersection.

Annex C (informative)

Complete names of chelating agents

EDTA	ethylenediaminetetraacetic acid C ₁₀ H ₁₆ O ₈ N ₂	CAS-No. 60-00-4
HEEDTA	2-hydroxyethylethylenediaminetriacetic acid C ₁₀ H ₁₈ O ₇ N ₂	CAS-No. 150-39-0
DTPA	diethylenetriaminepentaacetic acid C ₁₄ H ₂₃ O ₁₀ N ₃	CAS-No. 67-43-6
[<i>o,o</i>] EDDHA	ethylenediamine-N,N'-di[<i>ortho</i> -hydroxyphenyl]acetic acid] C ₁₈ H ₂₀ O ₆ N ₂	CAS-No. 1170-02-1
[<i>o,o</i>] EDDHMA	ethylenediamine-N,N'-di[<i>ortho</i> -hydroxymethylphenyl] acetic acid] C ₂₀ H ₂₄ O ₆ N ₂	CAS-No. 641632-90-8
EDDHSA	ethylenediamine-N,N'-di-[(2-hydroxy-5-sulfophenyl)acetic acid] and its condensation products C ₁₈ H ₂₀ O ₁₂ N ₂ S ₂ + n*(C ₁₂ H ₁₄ O ₈ N ₂ S)	CAS-No. 57368-07-7 and 642045-40-7
NTA	nitrilotriacetic acid C ₆ H ₉ O ₆ N	CAS-No. 139-13-9
CDTA	1,2-cyclohexylenediaminetetraacetic acid C ₁₄ H ₂₂ O ₈ N ₂	CAS-No. 482-54-2
HBED	N,N'-bis(2-hydroxybenzyl)-ethylenediamine-N,N'-diacetic acid C ₂₀ H ₂₄ O ₆ N ₂	CAS-No. 35998-29-9
HBEP	N,N'-bis(2-hydroxybenzyl)-ethylenediamine-N,N'-dipropionic acid C ₂₂ H ₂₈ N ₂ O ₆	CAS- No. 788792-49-4
[<i>o,p</i>] EDDHA	ethylenediamine-N-[(<i>ortho</i> -hydroxyphenyl)acetic acid]-N'-[(<i>para</i> -hydroxyphenyl)acetic acid] C ₁₈ H ₂₀ O ₆ N ₂	CAS- No. 475475-49-1

Bibliography

- [1] EN 1482-1, *Fertilizers and liming materials — Sampling and sample preparation — Part 1: Sampling*
- [2] CEN/TR 15106, *Fertilizers — Determination of chelating agents EDDHA and EDDHMA by ion pair chromatography — Comparison of non-standardized Lucena method with EN 13368-2:2001*
- [3] ISO 5725-2:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*
- [4] *Regulation (EC) No 2003/2003 of the European Parliament and of the Council of 13 October 2003 relating to fertilisers*, Official Journal L 304, 21/11/2003, pp. 1-194
- [5] Lucena, J. J.; P. Barak and L. Hernández-Apaolaza. 1996. *Isocratic ion-pair high-performance liquid chromatographic method for the determination of various iron(III) chelates*; J. Chromatography A, 727, pp. 253-264
- [6] Hernández-Apaolaza, L.; P. Barak and J.J. Lucena. 1997. *Chromatographic determination of commercial Fe(III) chelates of ethylenediaminetetracetic acid, ethylenediaminedi(o-hydroxyphenylacetic) acid and ethylenediaminedi(o-hydroxy-p-methylphenylacetic) acid*; J. Chromatography A, 789, pp. 453-460
- [7] López Rayo, S., Hernández, D. and Lucena, J. J. 2009. *Chemical Evaluation of HBED/Fe³⁺ and the Novel HJB/ Fe³⁺ Chelates as Fertilizers to Alleviate Iron Chlorosis*; Journal of Agricultural and Food Chemistry 57, pp. 8504-8513

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email bsmusales@bsigroup.com.

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

Useful Contacts:

Customer Services

Tel: +44 845 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 845 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com



...making excellence a habit.™