

Characterization of waste and soil — Determination of elemental composition by X-ray fluorescence

ICS 13.030.10; 13.080.10

National foreword

This British Standard is the UK implementation of EN 15309:2007.

The UK participation in its preparation was entrusted to Technical Committee B/508/3, Characterization of waste.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 May 2009

© BSI 2009

ISBN 978 0 580 55631 9

Amendments/corrigenda issued since publication

Date	Comments

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 15309

May 2007

ICS 13.030.10; 13.080.10

English Version

Characterization of waste and soil - Determination of elemental composition by X-ray fluorescence

Caractérisation des déchets et du sol - Détermination de la composition élémentaire par fluorescence X

Charakterisierung von Abfällen und Böden - Bestimmung der elementaren Zusammensetzung durch Röntgenfluoreszenz-Analyse

This European Standard was approved by CEN on 22 March 2007.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: rue de Stassart, 36 B-1050 Brussels

Contents

	Page
Foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Safety remarks	7
5 Principle.....	7
6 Apparatus	7
7 Reagents.....	8
8 Interferences and sources of error	8
9 Sample preparation	9
9.1 General.....	9
9.2 Drying and determination of dry mass	9
9.3 Preparation of pressed pellet	9
9.4 Preparation of fused beads	10
10 Procedure	10
10.1 Analytical measurement conditions	10
10.2 Calibration	11
10.3 Analysis of the samples	17
11 Quality control.....	18
11.1 Drift correction procedure	18
11.2 Blank test.....	18
11.3 Reference materials.....	18
12 Calculation of the result.....	18
13 Test report	19
Annex A (informative) Semi-quantitative screening analysis of waste, sludge and soil samples	20
Annex B (informative) Examples for operational steps of the sample preparation for soil and waste samples.....	23
Annex C (informative) Suggested analytical lines, crystals and operating conditions	29
Annex D (informative) List of reference materials applicable for XRF-analysis	31
Annex E (informative) Validation	32
Bibliography	40

Foreword

This document (EN 15309:2007) has been prepared by Technical Committee CEN/TC 292 "Characterization of waste", the secretariat of which is held by NEN.

This document has been prepared in coordination with ISO/TC 190 "Soil quality".

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2007, and conflicting national standards shall be withdrawn at the latest by November 2007.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom

Introduction

X-ray fluorescence spectrometry is a fast and reliable method for the quantitative analysis of the total content of certain elements within different matrices.

The quality of the results obtained depends very closely on the type of instrument used, e.g. bench top or high performance, energy dispersive or wavelength dispersive instruments. When selecting a specific instrument several factors have to be considered, such as the matrices to be analyzed, elements to be determined, detection limits required and the measuring time. The quality of the results depends on the element to be determined and on the surrounding matrix.

Due to the wide range of matrix compositions and the lack of suitable reference materials in the case of inhomogeneous matrices like waste, it is generally difficult to set up a calibration with matrix-matched reference materials.

Therefore this standard describes two different procedures:

- a quantitative analytical procedure for homogeneous solid waste, soil and soil-like material in the normative part. The calibration is based on matrix-matched standards;
- an XRF screening method for solid and liquid material as waste, sludge and soil in the informative Annex A which provides a total element characterisation at a semi-quantitative level. The calibration is based on matrix-independent calibration curves, previously set up by the manufacturer.

1 Scope

This European Standard specifies the procedure for a quantitative determination of major and trace element concentrations in homogeneous solid waste, soil and soil-like material by energy dispersive X-ray fluorescence (EDXRF) spectrometry or wavelength dispersive X-ray fluorescence (WDXRF) spectrometry using a calibration with matrix-matched standards.

This European Standard is applicable for the following elements: Na, Mg, Al, Si, P, S, Cl, K, Ca, Ti, V, Cr, Mn, Fe, Co, Ni, Cu, Zn, As, Se, Br, Rb, Sr, Y, Zr, Nb, Mo, Ag, Cd, Sn, Sb, Te, I, Cs, Ba, Ta, W, Hg, Tl, Pb, Bi, Th and U. Concentration levels between approximately 0,000 1 % and 100 % can be determined depending on the element and the instrument used.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 14346, *Characterisation of waste — Calculation of dry matter by determination of dry residue or water content*

EN 15002, *Characterisation of waste — Preparation of test portions from the laboratory sample*

EN ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories (ISO/IEC 17025:1999)*

ISO 11464, *Soil quality — Pretreatment of samples for physico-chemical analysis*

ISO 11465, *Soil quality — Determination of dry matter and water content on a mass basis — Gravimetric method*

3 Terms and definitions

For the purpose of this document, the following terms and definitions apply.

NOTE See [13] and [10] for non specified terms.

3.1

absorption edge

jump of the mass absorption coefficient at a specific wavelength or energy

3.2

absorption of X-rays

loss of intensity of X-rays by an isotropic and homogenous material as described by the Bouger-Lambert law

3.3

analytical line

specific characteristic X-ray spectral line of the atom or ion of the analyte used for determination of the analyte content

3.4

Bremsstrahlung; continuous radiation

electromagnetic radiation produced by the acceleration of a charged particle, such as an electron, when deflected by another charged particle, such as an atomic nucleus

3.5
Compton-line
spectral line due to incoherent scattering (Compton-effect) occurring when the incident X-ray photon strike an atom without promoting fluorescence

NOTE Energy is lost in the collision and therefore the resulting scattered X-ray photon is of lower energy than the incident X-ray photon.

3.6
drift correction monitors
physically stable samples used to correct for instrumental drift

3.7
emitted sample X-rays
radiation emitted by sample consisting of X-ray fluorescence radiation and scattered primary X-rays

3.8
fused bead
analyte sample prepared by dissolution in a flux

3.9
liquid sample
analyte sample submitted as a solution for direct measurement in the sample cup

3.10
mass absorption coefficient
constant describing the fractional decrease in the intensity of a beam of X-radiation as it passes through an absorbing medium, expressed in units of cm^2/g . The mass absorption coefficient is a function of the wavelength of the absorbed radiation and the atomic number of the absorbing element

3.11
polarised excitation X-ray spectrometer
energy dispersive X-ray spectrometer where the excitation is performed by polarised radiation and the emitted X-ray fluorescence radiation is detected along the direction of polarisation

3.12
powder sample
analyte sample submitted as a powder for direct measurement in the sample cup

3.13
precision
closeness of agreement of results obtained by applying the method several times under prescribed conditions [ISO 5725-2:1994]

3.14
pressed pellet
analyte sample prepared by pressing milled material into a disk

3.15
primary X-rays
X-rays by which the sample is radiated

3.16
quality control sample
stable sample with known contents, e.g. certified reference material (CRM) used to monitor instrument and calibration performance

3.17

X-ray fluorescence radiation

emission of characteristic X-rays from a sample that has been bombarded by high-energy X-rays or gamma rays

4 Safety remarks

Anyone dealing with waste and sludge analysis has to be aware of the typical risks that this kind of material presents irrespective of the parameter to be determined. Waste and sludge samples may contain hazardous e.g. toxic, reactive, flammable, infectious substances, which could potentially undergo biological and/or chemical reaction. Consequently it is recommended that these samples should be handled with special care. The gases that may be produced by microbiological or chemical activity are potentially flammable and will pressurise sealed bottles. Bursting bottles are likely to result in hazardous shrapnel, dust and/or aerosol. National regulations should be followed with respect to all hazards associated with this method.

The X-ray fluorescence spectrometer shall comply with European and national regulations relevant to radiation protection.

The person responsible for managing or supervising the operation of X-ray equipment shall provide evidence of his knowledge of radiation protection according to national regulations.

5 Principle

After a suitable preparation, if necessary, the sample is introduced into a XRF-spectrometer and excited by primary X-rays. The intensities of the secondary fluorescent energy lines specific for each element are measured and the elemental composition of the sample is determined by reference to previously established calibration graphs or equations and applying corrections for inter-element effects. The calibration equations and inter-element corrections are established using pure reagents and/or series of internal or reference materials providing they meet all the requirements of the relevant preparation technique.

6 Apparatus

6.1 X-ray fluorescence spectrometer

The X-ray fluorescence spectrometer shall be able to analyse the elements according to the scope of this European Standard. The following types of X-ray fluorescence spectrometers are applicable:

- energy dispersive X-ray fluorescence (EDXRF) spectrometer that achieves the dispersion of the emitted X-ray fluorescence radiation by an energy dispersive detector;
- wavelength dispersive X-ray fluorescence (WDXRF) spectrometer that achieves the dispersion of the emitted X-ray fluorescence radiation by diffraction by a crystal or a synthetic multilayer.

The spectrometer consists of a number of components:

- primary X-ray source, an X-ray tube with a high voltage generator;
- a sample holder;
- detector unit including electronic equipment;
- source modifiers to modify the shape or intensity of the source spectrum or the beam shape (like source filters, secondary targets, polarising targets, collimators, focussing optics etc.).

The detector unit is different for WDXRF and for EDXRF spectrometers. WDXRF spectrometers take advantage of the dispersion of the emitted radiation by scattering by a crystal or a synthetic multilayer. The detector does not need to be capable of energy discrimination. EDXRF spectrometers use an energy dispersive detector. Pulses of current from the detector, which are a measure of the energy of the incoming X-rays, are segregated into channels according to energy using a Multi-Channel Analyser (MCA).

NOTE 1 The use of a high-energy X-ray tube increases the potential for losses of volatile analytes from samples by heating in the spectrometer during analysis.

NOTE 2 The new generation of EDXRF spectrometers takes advantage of the polarising target theory resulting in a significant decrease of the background scattering, and therefore lower limits of detection can be achieved (comparable to WDXRF).

6.2 Mill, preferable with walls made of agate, corundum or zircon.

6.3 Pellet preparation equipment: manual or automatic pellet press, capable of providing a pressure of at least 100 kN.

6.4 Aluminium cup: supporting backing cup for pressed pellets.

6.5 Fusion apparatus: electric, gas or high frequency induction furnace that can be heated up to a fixed temperature of between 1 050 °C and 1 250 °C.

6.6 Fusion crucibles: crucibles made of non-wetting platinum alloy (Pt 95 %; Au 5 % is suitable).

Lids, if used, shall be made from platinum alloy.

NOTE Certain metal sulphides (so called platinum poisons) affect the platinum crucibles in which the sample is melted.

6.7 Casting moulds: non-wetting platinum alloy (Pt 95 %; Au 5 % is suitable).

7 Reagents

The reagents mentioned are used as carrier material.

7.1 Binder: liquid or solid binder free of analytes of interest. Solid materials can contain a certain amount of moisture, which shall be compensated for.

NOTE Different type of binders may be used. A binder commonly used is wax.

7.2 Flux: solid flux free of analytes of interest. Solid materials can contain a certain amount of moisture, which shall be compensated for (see EN ISO 12677 for compensation for moisture in flux).

NOTE Different type of fluxes may be used. Fluxes commonly used are lithium metaborate, lithium tetraborate or mixtures of both.

8 Interferences and sources of error

The container in which the sample is delivered and stored can be a source of error. Its material shall be chosen according to the elements to be determined.

NOTE Elemental Hg can penetrate polyethylene walls very rapidly in both directions. In the case of glass containers, contamination may be observed for some elements e.g. Al, As, Ba, Ce, K, Na, Pb.

Interferences in X-ray fluorescence spectrometry are due to spectral line overlaps, matrix effects, spectral artefacts and particle size or mineralogical effects.

Spectral line overlaps occur when an analytical line cannot be resolved from the line of a different element. Corrections for these interferences are made using the algorithms provided with the software.

Matrix effects occur when the X-ray fluorescence radiation from the analyte element is absorbed or enhanced by other elements in the sample before it reaches the detector. In the case of complex matrices these effects generally have to be corrected.

Spectral artefacts e.g. escape peaks, sum peaks, pulse pile up lines, dead time, Bremsstrahlung correction, are accounted for by the provided software. Spectral artefacts differ for energy dispersive and wavelength dispersive XRF spectrometry.

Particle size effects can be reduced by milling the sample, and both particle size and mineralogical effects can be eliminated by preparing bead samples. It is vital for quantitative analysis that the same sample preparation procedure is applied to both the standards and the samples to be analysed.

9 Sample preparation

9.1 General

In analysis by XRF spectrometry the sample preparation step is crucial as the quality of the sample preparation will strongly influence the accuracy of the results.

For quantitative analysis of solid samples, pressed pellets or fused beads have to be prepared. The application of the pressed pellet method is recommended for the quantification of trace elements and mandatory for the quantification of volatile elements, and the fused bead method for the determination of non-volatile major and minor elements.

NOTE 1 The preparation of fused beads eliminates effects due to particle size and mineralogy.

The conditions of the preparation of fused beads shall be adapted to the matrix properties. Otherwise the preparation of fused beads may be difficult or cause problems in case of waste like matrices as sludges.

For a given calibration the same preparation method shall be used throughout, for both samples and standards.

NOTE 2 Depending on the sample type other sample preparation methods may be applied according to Annex B.

For precise quantitative measurements, homogeneous and representative test portions are necessary. Pre-treatment and preparation of test portions shall be carried out according to the appropriate clauses of ISO 11464 and EN 15002. The particle size of the sample may strongly affect the precision of the measurement. The particle size should preferably be smaller than 150 μm .

NOTE 3 Particle size smaller than 80 μm is recommended for the analysis of low atomic mass elements when using the pressed pellet method.

9.2 Drying and determination of dry mass

Prepare and dry the sample according to ISO 11464 or EN 15002. Determine the dry mass according to ISO 11465 or EN 14346.

9.3 Preparation of pressed pellet

After drying and milling or grinding the sample, a pellet is prepared in the pellet press (6.3). Before pressing, the sample shall be mixed and homogenised with a binder (7.1). For the preparation of 40 mm diameter pellets, about 10,0 g of sample is taken, for 32 mm diameter pellets about 4,5 g of sample is required. The amount of binder in the pellet shall be taken into account for the dilution factor. It is recommended to press the sample in an aluminium cup (6.4) as support.

NOTE 1 Different type of binders can be used. A binder commonly used is wax. In the case of a liquid binder the pellet is placed in an oven to evaporate organic solvent.

NOTE 2 Different dilution factors can be used. A proportion of sample: binder commonly used is 10:1 by weight.

9.4 Preparation of fused beads

After drying and milling or grinding the sample, a fused bead is prepared using the fusion apparatus (6.5).

Ignite the sample at an appropriate temperature until constant mass is reached. Determine the loss on ignition at the same temperature to correct for volatile elements and/or compounds being released during ignition of the sample.

NOTE 1 The ignition temperature can vary depending on the sample matrix. A temperature commonly used is $1025^{\circ}\text{C} \pm 25^{\circ}\text{C}$.

Because of the wide applicability of the fused bead technique, various fluxes and modes of calibration are permitted providing they have been demonstrated to be able to meet certain criteria of reproducibility, sensitivity and accuracy.

For application of alkaline fusion technique (e.g. selection of flux, fusion temperature, additives) ISO 14869-2 or CEN/TR 15018 should be used.

NOTE 2 Fluxes commonly used are lithium metaborate, lithium tetraborate or mixtures of both.

NOTE 3 Loss of volatile elements e.g. As, Br, Cd, Cl, Hg, I, S, Sb, Se, Tl may occur during the ignition and fusion processes. Also Cu may be volatile if a bromide releasing agent is used.

The flux (7.2) is added to the ignited material. For the preparation of 40 mm diameter beads, about 1,6 g of ignited sample is taken, for 32 mm diameter beads about 0,8 g of ignited sample is required. The amount of flux in the bead shall be taken into account for the dilution factor. The same sample preparation procedure and ratio of sample to flux shall be used for samples and standards. The beads produced should be visually homogeneous and transparent.

NOTE 4 Non ignited material may be used to prepare beads but, nevertheless, loss of ignition needs to be determined and needs to be taken into account in the calculation of the results. It should be noted that non ignited material may contain compounds that can damage the platinum crucibles during fusion.

NOTE 5 Different dilution factors may be used. A proportion of sample: flux commonly used is 1:5 by weight.

After fusion in a platinum-gold crucible (6.6) the melt is poured into a casting mould (6.7) to make a bead.

NOTE 6 Beads can deteriorate because of adverse temperature and humidity conditions, so it is recommended that beads are stored in desiccators.

10 Procedure

10.1 Analytical measurement conditions

10.1.1 Wavelength dispersive instruments

The analytical lines to be used and suggested operating conditions are given in Table C.1. The settings are strongly dependant on the spectrometer configuration, e.g. the type of X-ray tube (Rh, Cr), tube power, available crystals, type of collimators.

Intensities and background corrections

For the determination of trace elements the measured intensities have to be background corrected. The measured background positions should be free of spectral line interferences. The net peak intensity I , expressed as the number of counts per second of the element of interest, is calculated as the difference between the measured peak intensity of the element and the background intensity:

$$I = I_p - I_b \quad (1)$$

where

I_p is the count rate of the element i , expressed as the number of counts per second;

I_b is the background count rate of the element i , expressed as the number of counts per second.

Counting time

The minimum counting time is the time necessary to achieve an uncertainty ($2\sigma_{\%}$), which is less than the desired precision of the measurement. Choose a reference material with a concentration level in the middle of the working range and measure the count rate. The counting time for each element can be calculated according to:

$$t = \left(\frac{100}{2\sigma_{\%}} \cdot \frac{1}{\sqrt{I_p} - \sqrt{I_b}} \right)^2 \quad (2)$$

where

t is the total counting time for the peaks and background in seconds;

$2\sigma_{\%}$ is the relative target precision at a confidence level of 95 %, expressed as percentage.

10.1.2 Energy dispersive instruments

The analytical lines to be used and suggested operating conditions are given in Table C.2. The settings are strongly dependant on the spectrometer configuration, e.g. type of X-ray tube (Rh, Pd), tube power, available targets, type of filters.

Intensities and background corrections

Deconvolution of the spectra and background correction are needed when analysing samples with overlapping lines. Usually XRF-instruments are supplied with a specific software module for that purpose.

10.2 Calibration

10.2.1 General

The calibration procedure is similar for energy dispersive and wavelength dispersive techniques. In general calibration is established by using matrix-adapted reference materials. The calibration equations and inter-element corrections are calculated by the software of the instrument. An accuracy check is performed with CRMs or samples with known composition.

Different procedures for correcting matrix effects may be used according to the analytical accuracy required:

- the scattered radiation method is based on the principle that the intensities of the analyte line and of the Compton line are affected in the same proportion due to the overall mass absorption coefficient of the sample. This linear relationship holds when all analytes are at low concentrations (trace elements) and their absorption coefficients are not affected by an adjacent absorption edge. In this case an internal Compton correction can be used. Beside that, a correction method using the Compton intensity with Mass Absorption Coefficients (MAC) is also applicable. In this method, the intensities of the major elements are measured to apply a jump edge correction for the analysed trace elements;
- correction using the fundamental parameter approach;
- correction using theoretical correction coefficients (alphas) taking basic physical principles, instrumental geometry etc. into account;
- correction using empirical correction coefficients (alphas) based on regression analysis of standards with known elemental concentrations.

10.2.2 General calibration procedure

For calibration purposes the measurement of analyte lines of samples of known composition is needed. The basic equation implies a linear relationship between the intensity and the concentration.

$$C_i = a_{i,0} + a_{i,1} \cdot I_i \quad (3)$$

where

C_i is the concentration of the element of interest, expressed as mg/kg or percentage dry matter;

$a_{i,0}$ is the offset of the calibration curve;

$a_{i,1}$ is the slope of the calibration curve;

I_i is the net intensity of the element of interest, expressed as counts per second.

Matrix effects have to be taken into account in X-ray spectrometry according to the following equation:

$$C_i = (a_{i,0} + a_{i,1} \cdot I_i) \cdot M \quad (4)$$

where

M is the correction term due to the matrix effects.

The matrix effect correction term may consist of an internal standard Compton correction method or may be calculated from mathematical models.

10.2.3 Internal standard correction using Compton (incoherent) scattering method

The measured intensity of incoherent scattering may be used directly to compensate for matrix effects or indirectly for the determination of the effective mass absorption coefficient μ to correct for matrix effects. The compensation for matrix effects is based on a combination of sample preparation and experimental intensity data but not on fundamental and experimental parameters.

The Compton scatter method can be expressed as:

$$C_{i,u} = (C_{i,r} \cdot \frac{I_{inc,r}}{I_{i,r}}) \cdot (\frac{I_{i,u}}{I_{inc,u}}) \quad (5)$$

where

- $C_{i,u}$ is the concentration of the element of interest i of the sample, expressed as mg/kg or percentage dry matter;
- $C_{i,r}$ is the concentration of the element of interest i of the calibration reference material, expressed as mg/kg or percentage dry matter;
- $I_{inc,u}$ is the intensity of the incoherent Compton line of the sample, expressed as counts per second;
- $I_{inc,r}$ is the intensity of the incoherent Compton line element of the calibration reference material, expressed as counts per second;
- $I_{i,u}$ is the intensity of the element of interest i of the sample, expressed as counts per second;
- $I_{i,r}$ is the intensity of the element of interest i of the calibration reference material, expressed as counts per second.

10.2.4 Fundamental parameter approach

The fundamental parameter approach uses the physical processes forming the basis of X-ray fluorescence emission and scattering to construct a theoretical model for the correction of matrix effects in practice. The correction term M is calculated from first principle expressions. These are derived from basic X-ray physics and contain physical constants and parameters that include absorption and scattering coefficients, fluorescence yield, primary spectral distributions and spectrometry geometry. The use of scattered radiation (Compton and/or Rayleigh) allows the determination of matrix effects caused by sample elements that cannot be measured directly. The calculation of analyte concentrations in samples is based on making successively better estimates of composition by an iteration procedure. These iteration cycles are performed until the difference between the compared results is below a defined value.

NOTE The algorithm used for the procedure is usually implemented in the manufacturer's software.

10.2.5 Fundamental or theoretical influence coefficient method

The fundamental influence coefficient method encompasses any mathematical expression relating emitted intensities and concentrations in which the influence coefficients are defined and derived explicitly in terms of fundamental parameters.

The calculation of the concentration from the intensities is performed by linear regression whereby the net intensities are corrected for the present matrix effects. For each element the concentration is calculated according to the following equation:

$$C_{i,u} = \left(\frac{C_{i,r}}{I_{i,r} (1 + \sum_j \alpha_{ij} C_{jr})} \right) \cdot I_{i,u} \cdot M \quad (6)$$

$$C_{i,u} = \left(\frac{C_{i,r}}{I_{i,r} \left(1 + \sum_j \alpha_{ij} C_{j,r} \right)} \right) \cdot I_{i,u} \cdot \left(1 + \sum_j \alpha_{ij} C_{j,u} \right) \quad (7)$$

where

- $C_{i,u}$ is the concentration of the element of interest i of the sample, expressed as mg/kg or percentage dry matter;
- $C_{i,r}$ is the concentration of the element of interest i of the calibration reference material, expressed as mg/kg or percentage dry matter;
- $I_{i,r}$ is the intensity of the element of interest i of the calibration reference material, expressed as counts per second;
- $I_{i,u}$ is the intensity of the element of interest i of the sample, expressed as counts per second;
- $C_{j,r}$ is the concentration of the matrix element j of the calibration reference material, expressed as mg/kg or percentage dry matter;
- $C_{j,u}$ is the concentration of the matrix element j of the sample, expressed as mg/kg or percentage dry matter;
- M is the matrix correction term;
- α_{ij} is the correction coefficient α_{ij} (called alphas) calculated from theory, although some approximations are involved.

Different types of alpha coefficient exist, but all of them are calculated without reference to experimental data; they are calculated using intensity data resulting from a fundamental parameter expression. The alpha coefficients vary as a function of sample composition and are calculated by an iterative process.

10.2.6 Empirical alpha correction

Empirical alphas are obtained experimentally using regression analysis of data from reference materials in which the elements to be measured are known and the total concentration range is covered. Best results are achieved when the samples and reference materials are of similar composition. Thus, empirical alphas are based strictly on experimental data and do not take fundamental and instrumental parameters into account. Different models can be applied, but generally they are based on the above equation where the correction term for matrix effects is a function of concentrations.

The empirical alphas are only applicable for a limited concentration range and a well-defined analytical method where the matrices of samples and standards are similar. The reference materials used should contain each analyte together with fairly wide concentration ranges of each matrix element. Poor analytical results are obtained when inappropriate combinations of analytes are chosen. A large number of reference materials have to be analysed to define the alphas (rule of thumb: minimum of 3 times the number of parameters to be calculated).

10.2.7 Calibration procedure for trace elements using the pressed pellet method

The pressed pellet method is used to determine the concentrations of trace elements.

Select calibration standards with a similar composition as the samples under investigation containing the elements of interest and covering the concentration range of interest. The use of reference materials from different recognised producers is recommended (see Annex D) or synthetic mixtures of oxides may be

prepared. The element concentrations shall vary independently in the standards. If the calibration covers many elements in a wide range of concentrations, a large number of calibration samples may be necessary.

Prepare pressed pellets from the selected calibration standards according to 9.3.

Define the analytical measurement method for EDXRF or WDXRF as described in 10.1.

Start up the XRF equipment according to the instrument manufacturer's manual and measure the calibration standards using the defined measurement method. All measurements shall be performed under vacuum.

NOTE It is important to note that the pressed pellet method is not ideal for the determination of major elements, but these elements are measured so that alpha corrections can be applied to some elements of interest.

Follow the guidelines in the instrument manufacturer's manual to perform the regression, the background correction, the line overlap correction and the matrix corrections for all elements under consideration. In Table 1 the possible spectral line overlaps are indicated (dependant on the configuration of the instrument) and also the matrix correction method that can be applied. For trace elements with an absorption edge above the absorption edge of iron, a Compton internal standard correction can be applied. Otherwise a theoretical alpha correction or correction for the absorption edge should be performed (for these corrections all elements in the sample have to be analysed).

Depending on the type of instrument and the software programs available, alternative correction methods can be applied. Validation of the final calibration curves shall demonstrate the accuracy of the method.

Perform the regression calculation and verify that the correlation factors are within the limits of accuracy required.

Table 1 — Suggested analytical lines, spectral line overlaps and correction methods

Element	Line	Spectral line overlap	Type of matrix correction method
Na	K α	ZnL β	Alpha or FP
Mg	K α	AsL α	Alpha or FP
Al	K α	BrL α	Alpha or FP
Si	K α		Alpha or FP
P	K α		Alpha or FP
S	K α	CoK α PbM α NbL β	Alpha or FP or MAC
Cl	K α		Alpha or FP or MAC
K	K α		Alpha or FP
Ca	K α		Alpha or FP
Ti	K α	BaL α IL β	Alpha or FP
V	K α	Ti K β	Alpha or FP or MAC
Cr	K α	VK β PbL α	Alpha or FP or MAC
Mn	K α	CrK β	Alpha or FP
Fe	K α	MnK β	Alpha or FP
Co	K α	FeK β	Alpha or FP or MAC
Ni	K α	CoK β	Compton or FP or MAC
Cu	K α	TaL α ThL β	Compton or FP or MAC
Zn	K α	WL α	Compton or FP or MAC
As	K α K β	PbL α BrK α	Compton or FP or MAC
Se	K α		Compton or FP or MAC
Br	K α	AsK β	Compton or FP or MAC
Rb	K α	UL α BrK β	Compton or FP or MAC
Sr	K α	UL α	Compton or FP or MAC
Y	K α	RbK β	Compton or FP or MAC
Zr	K α	SrK β	Compton or FP or MAC
Nb	K α	YK β UL β	Compton or FP or MAC
Mo	K α	ZrK β UL β	Compton or FP or MAC
Ag	K α L α	CrK β	Compton or FP or MAC Alpha or FP
Cd	K α L α	AgL β	Compton or FP or MAC Alpha or FP
Sn	K α L α	CoK α	Compton or FP or MAC Alpha or FP or MAC
Sb	K α L β	CoK β	Compton or FP or MAC Alpha or FP or MAC
Te	K α L α	SnL β	Compton or FP or MAC Alpha or FP or MAC
I	K α L α		Compton or FP or MAC Alpha or FP or MAC
Cs	K α L α	ZnK α IL β	Compton or FP or MAC Alpha or FP or MAC
Ba	K α L α	TiK α IL β CuK β	Compton or FP or MAC Alpha or FP or MAC
Ta	L α	CuK α NiK β	Compton or FP or MAC
W	L α	TaLn	Compton or FP or MAC
Hg	L α	WL β	Compton or FP or MAC
Tl	L β	PbL β	Compton or FP or MAC

Element	Line	Spectral line overlap	Type of matrix correction method
Pb	L β	ThL α BiL β SnK α	Compton or FP or MAC
Bi	L α	TaL γ	Compton or FP or MAC
Th	L α	BiL β PbL β	Compton or FP or MAC
U	L α	BrK β RbK α	Compton or FP or MAC

10.2.8 Calibration procedure for major and minor oxides using the fused bead method

The fused bead method is used to determine the concentrations of major and minor elements.

Select calibration standards with a similar composition as the samples under investigation containing the elements of interest and covering the total concentration range of interest. The use of reference materials from different recognised producers is recommended (see Annex D) or synthetic mixtures of oxides may be prepared. The element concentrations shall vary independently in the samples. If the calibration covers many elements in a wide range of concentrations, a large number of calibration samples may be necessary.

Prepare fused beads from the selected calibration standards according to 9.4.

NOTE Due to a higher dilution factor for fused beads the limit of detection of the different elements will be higher than those for pressed pellets.

Define the analytical measurement method for EDXRF or WDXRF as described in 10.1.

Start up the XRF equipment according to the instrument manufacturer's manual and measure the calibration standards using the defined measurement method. All measurements shall be performed under vacuum.

In the calibration program all the elements of the reference materials has to be defined as oxides and the concentrations are reported as oxides.

Follow the guidelines in the instrument manufacturer's manual how to perform the regression, the background correction, the line overlap correction and the matrix corrections for all elements under consideration. In Table 1 the possible spectral line overlaps are indicated (dependant on the configuration of the instrument). For all elements an alpha correction method using theoretical alphas should be applied.

Depending on the type of instrument and the software programs available, alternative correction methods can be applied. Validation of the final calibration curves shall demonstrate the accuracy of the method.

Perform the regression calculation and verify that the correlation factors are within the limits of accuracy required.

10.3 Analysis of the samples

Follow the instrument manufacturer's instructions for set up, conditioning, preparation and maintenance of the XRF spectrometer.

Select the required preparation method and prepare the samples. For the quantification of trace elements the pressed pellet method is recommended and for the determination of major and minor elements the fused bead method should be used.

To analyse the prepared samples, an analytical measurement method has to be defined. The measurement method describes the analytical lines to be measured and the measurement parameters e.g. the XRF generator settings (tube voltage and current), selection of primary beam filters, targets and crystals, detector to be used, measurement time.

The same measurement parameters used for the calibration according to 10.2 are applied to the samples.

Before analysis, quality control samples have to be measured to check the instrument stability and the quality of the calibration, in accordance to the manufacturer's instructions.

Introduce the prepared sample into the XRF spectrometer and analyse it in accordance to the manufacturer's instructions.

11 Quality control

11.1 Drift correction procedure

XRF calibrations, once established, tend to be stable over long periods of time. Small amounts of instrumental drift can be corrected by analysing stable monitor samples as frequently as performance experience indicates.

Drift correction monitors are stable beads that should contain all the elements to be determined and at concentration levels comparable to or higher than those from the samples.

The monitor samples shall be measured together with the calibration samples in order to get the initial intensities stored. When drift correction is needed, they are measured again. The initial set and the actual set of intensities are used to adjust the calibration regression. The procedure described is usually part of the instruments software.

For EDXRF spectrometers, an additional energy calibration has to be performed on a regular basis, as defined by the manufacturer's instructions.

11.2 Blank test

To assess the level of impurities in the flux, carry out a blank test by preparing and analysing a blank sample using the same sample preparation procedure and the same quantities of reagents.

11.3 Reference materials

Verify the trueness of the results by applying the procedure to one or more reference materials not used for calibration and covering the concentration range of interest.

The element content of the reference material used shall be in accordance with the concentration range of interest.

12 Calculation of the result

Follow the guideline in the instrument manufacturer's manual how to perform the regression, the background correction and the overlap correction.

The concentrations of the analytes are calculated by the software program from the measured intensities using the calibrations curves previously set-up. The results are expressed as elements in terms of mg/kg dry matter for trace elements and as oxides in mass percentages dry matter for major and minor elements.

13 Test report

The work carried out by the testing laboratory shall be covered by a report which accurately, clearly and unambiguously presents the test results and all other relevant information as specified in EN ISO/IEC 17025.

In addition to test results the test report shall include at least the following information:

- a) description and identification of the laboratory sample;
- b) which processes, procedures and apparatus were used;
- c) results of the determination expressed as mg/kg dm or mass percentages dm;
- d) any details not specified in this European Standard or which are optional, and any other factors which may have affected the results;
- e) date of receipt of laboratory sample and date(s) of performance of test;
- f) a reference to this European Standard, i.e. EN 15309.

Corrections or additions to a test report after issue shall be made only by a further document suitably marked, e.g. "Amendment/Addendum to test report serial number (or as otherwise identified)", and shall meet the relevant requirements of the preceding paragraphs.

Annex A (informative)

Semi-quantitative screening analysis of waste, sludge and soil samples

A.1 Principle

The principle is identical as described in Clause 5, however the elemental composition of the sample is determined by reference to calibration curves, previously set up by the manufacturer. This procedure is often referred to as "standardless" analysis. The method is generally applicable for the semi-quantitative determinations of elements from sodium to uranium except noble gases at levels between approximately 0,01 % and 100 %, depending on the element and the instrument used.

During the evaluation and calculation of the element concentration of the sample the various interferences e.g. spectral line overlap, matrix effects, spectral artefacts and sample preparation are all accounted for with the provided analytical program.

A.2 Energy dispersive (ED) or wavelength dispersive (WD) X-ray fluorescence spectrometer

The same instruments as described in 6.1 may be used, however a specific software package is applied suitable to perform the XRF analysis without the use of calibration curves set up with dedicated reference samples. Most of the instruments available are delivered with pre-calibrated analytical methods. These calibrations are set up by the manufacturer with a suite of synthetic calibration samples to cover a wide concentration range on a broad spectrum of matrix types. Improvement of the accuracy can be obtained by additional analyses of sample specific reference materials and extending the calibration for the specific needs.

NOTE Because of the differences between various models of XRF instruments, no detailed operating instructions can be provided.

The validity of the programmed calibration curves can be checked and optimised by using reference materials of a similar composition as the samples under investigation.

A.3 Sample preparation

The sample preparation determines significantly the obtained quality of the XRF results. For detailed information on sample preparation procedures, refer to the flowcharts and the sample preparation techniques in Annex B.

A.4 Procedure

A.4.1 Analytical measurement conditions and calibration

All X-ray spectrometers are supplied with a spectrometer software program to operate the instrument. The software packages are manufacturer depended and contain 2 major modules:

- analytical measurement program for data collection. This module controls the measurement of a sample using a certain set of measurement parameters e.g. tube setting (kV, mA), targets and crystals, detectors, measurement times. The analytical program is always linked to a selected evaluation and calibration

program. Actually, the same measurement conditions have to be applied for both the standards of the calibration curve and the samples. Because in screening analysis the measurements will be performed with the predefined analytical programs, no further detailed descriptions will be given of the analytical measurement parameters. Follow the manufacturer's instruction for further operation and handling of the analytical software package;

- evaluation program for data processing. This module converts the measured intensities of the different element line to elemental concentrations taking all corrections into account. There are various types of evaluation programs available and each manufacturer has set up his own program for data processing based on the XRF principles.

Sensitivity, instrumental detection limits and precision are instrument dependent and should therefore be investigated and established for each individual analyte line on that particular instrument, and, if relevant, in function of matrix type and sample preparation.

A.4.2 Validation

Prior to analysis of a sample, the available pre-calibrated analytical method has to be validated by using reference samples with a similar composition as the samples under investigation. If no reference materials with a comparable matrix are available, only a qualitative analysis with indicative concentration values can be performed.

The reference sample can consist of:

- in-house or commercially available reference materials - if possible certified - with matrices similar to that of the sample;
- synthetic samples, made by weighing the appropriate amount of each pure reagent;
- site specific or batch specific samples, similar to the matrix of the sample;
- standard addition method or spiked samples may also be used to create standards for which appropriate reference materials are not available for an element of interest. The matrix material needs to match that of the sample.

The element concentrations of these reference samples have to be known by certification or by determination with a different analytical technique.

The reference samples have to be analysed under the same analytical conditions as the sample, meaning the same sample preparation (pellet, powder etc.), the same analytical measurement method etc.

A.4.3 Analysis

Follow the instrument's instructions for set up, conditioning, preparation and maintenance of the XRF spectrometer.

To analyse the prepared samples, an analytical measurement method has to be defined. Depending on the type of instrument, the analytical measurement method is immediately applicable, or can be modified starting from a 'master' measurement method or has to be set up using the available pre-calibrated spectral lines. The measurement method describes the analytical lines to be measured and the measurement parameters like the XRF generator settings (tube voltage and current), selection of primary beam filters, targets and crystals, detector to be used, measurement time etc.

The same measurement parameters of the analyte line used for the calibration have to be applied for the samples. The provided software program has to be capable of defining automatically for each analyte the required measurement parameters.

Before analysis, quality control samples have to be measured to check the instrument stability and the quality of the calibration, in accordance to the manufacturer's instruction.

Introduce the prepared sample into the XRF spectrometer and analyse it with the selected analytical measurement method, in accordance to the manufacturer's instruction.

A.5 Calculation and evaluation

The software program from the measured intensity calculates the element concentrations, expressed in mg/kg dry matter for trace and minor elements and m/m % dry matter for major elements.

The effective matrix corrections have to account for low atomic mass elements like hydrogen, carbon or oxygen. As they are easily absorbed they have a strong influence on the matrix effects. Different approaches are applied in commercial software packages for this effect. First of all, automated correction using mass absorption coefficients can be performed. Secondly, it may be possible to enter the matrix compounds e.g. CH₂ for polymers, H₂O for water samples, and C for coal samples, to optimise the calculation process. All these methods allow the use of one calibration for different matrices.

NOTE In some cases, it is possible to optimize the iterative calculation process for the element concentration by inserting concentration values of specific compounds, obtained with another analytical technique.

A.6 Quality control

The quality control has to be done according to Clause 11.

A.7 Safety remarks

The safety remarks are the same as described in Clause 4.

A.8 Test report

The test report has to be done according to Clause 13.

Annex B (informative)

Examples for operational steps of the sample preparation for soil and waste samples

As waste samples can consist of different material types, various sample preparations have to be applied prior to the XRF screening analysis. The operational steps for different matrices are given in the flow diagrams in Figures B.1 to B.5:

- Figure B.1: Operational steps for the preparation of soil, sediment, fly ash and sludge samples;
- Figure B.2: Operational steps for the preparation of samples consisting of carbon matrices;
- Figure B.3: Operational steps for the preparation of liquid samples;
- Figure B.4: Operational steps for the preparation of paste-like materials;
- Figure B.5: Operational steps for the preparation of scrap samples.

The procedures of the different sample preparation techniques as mentioned in the pictures are described in ISO 11464, EN 15002, explained in the normative part or otherwise given directly.

B.1 Preparation of liquid samples

If the liquid sample is homogeneous, it is poured directly into a sample cup. The bottom of the sample cup is closed by a thin-film support. Close the sample cup.

B.2 Preparation of powder samples

After milling/grinding of the sample to powder, the homogenised powder sample can be poured directly into a sample cup. The bottom of the sample cup is closed by a thin-film support. Close the sample cup. Press the powder slightly with a piston to form a flat and even surface and avoid any air bubbles inside the powder.

B.3 Impregnation on powders

B.3.1 Impregnation on aluminium oxide

Place $15,000 \text{ g} \pm 0,005 \text{ g}$ of aluminium oxide (precalcined at $1\ 500^\circ\text{C}$) into a HDPE bottle. Add $5,000 \text{ g} \pm 0,005 \text{ g}$ of the homogeneous sample and two mixing balls. Close the bottle and shake vigorously for approximately 30 s. Empty the aluminium oxide / sample mixture from the bottle to the sample cup. Press the powder slightly with a piston to form a flat and even surface and avoid any air bubbles inside the powder.

B.3.2 Impregnation on activated carbon

Weight $4,000 \text{ g} \pm 0,005 \text{ g}$ of sample, add $2,000 \text{ g} \pm 0,005 \text{ g}$ of active carbon and homogenise very well the mixture. Empty the active carbon / sample mixture from the bottle to the sample cup. The powder is slightly pressed with a piston to form a good surface to avoid any air holes on the bottom. High volatile liquids can be impregnated on active carbon.




Figure B.1 — Operational steps for the preparation of soil, sediment, fly ash and sludge samples




Figure B.2 — Operational steps for the preparation of samples consisting of carbon matrices




Figure B.3 — Operational steps for the preparation of liquid samples




Figure B.4 — Operational steps for the preparation of paste-like materials




Figure B.5 — Operational steps for the preparation of scrap samples

Annex C (informative)

Suggested analytical lines, crystals and operating conditions

Table C.1 — Suggested analytical lines, crystals and operating conditions for wavelength dispersive XRF

Optimum excitation for element analysis can be realised by applying a suitable crystal. The following table shows a list with potentially usable crystals. For all of these the tube voltage and current needs to be set carefully in order to get optimum results. The number of excitation conditions selected shall be optimised particularly with regard to detection sensitivity and required analysis time.

Element	Line	Crystals	kV/mA	Collimator	Detector
Na	K α	OVO-55/PX-1/ TlAp	30/100	coarse	FC
Mg	K α	OVO-55/PX-1/ TlAp	30/100	coarse	FC
Al	K α	OVO-55/PX-1/ PET	30/100	coarse	FC
Si	K α	OVO-55/PX-1/ PET	30/100	coarse	FC
P	K α	Ge/PET	30/100	coarse	FC
S	K α	Ge/PET	30/100	coarse	FC
Cl	K α	Ge/PET	30/100	coarse	FC
K	K α	LiF200	50/60	fine	FC
Ca	K α	LiF200	50/60	fine	FC
Ti	K α	LiF200	50/60	fine	FC
V	K α	LiF200	30/100	fine	FC
Cr	K α	LiF200	60/50	fine	FC
Mn	K α	LiF200	60/50	fine	FC
Fe	K α	LiF200	60/50	fine	FC
Co	K α	LiF200	60/50	fine	SC
Ni	K α	LiF200	60/50	fine	SC
Cu	K α	LiF200	60/50	fine	SC
Zn	K α	LiF200	60/50	fine	SC
As	K α	LiF200	60/50	fine	SC
	K β	LiF200	60/50	fine	SC
Se	K α	LiF200	60/50	fine	SC
Br	K α	LiF220	60/50	fine	FC
Rb	K α	LiF220	60/50	fine	FC
Sr	K α	LiF200	60/50	fine	FC
Y	K α	LiF220	60/50	fine	FC
Zr	K α	LiF220	60/50	fine	FC
Nb	K α	LiF220	60/50	fine	FC
Mo	K α	LiF220	60/50	fine	FC
Ag	K α	LiF220	60/50	fine	SC
	L α	PET	30/100	coarse	FC
Cd	K α	LiF220	60/50	fine	SC
	L α	PET	30/100	coarse	FC
Sn	L α	LiF200	30/100	fine	FC
Sb	L α	LiF200	30/100	fine	FC

Element	Line	Crystals	kV/mA	Collimator	Detector
Te	L α	LiF200	30/100	fine	FC
I	L α	LiF200	30/100	fine	FC
Cs	L α	LiF200	30/100	fine	FC
Ba	L α	LiF200	30/100	fine	FC
Ta	L α	LiF200	60/50	fine	SC
W	L α	LiF200	60/50	fine	SC
Hg	L α	LiF200	60/50	fine	SC
Tl	L β	LiF200	60/50	fine	SC
Pb	L β	LiF200	60/50	fine	SC
Bi	L α	LiF200	60/50	fine	SC
Th	L α	LiF220	60/50	fine	SC
U	L α	LiF220	60/50	Fine	SC
FC flow counter					
SC scintillation counter					

Table C.2 — Suggested analytical lines, targets and operating conditions for energy dispersive XRF

Optimum excitation for trace element analysis can be realised by applying secondary or polarisation targets in to the excitation beam of the spectrometer. The following table shows a list with potentially usable targets. For all of these the tube voltage and current needs to be set carefully in order to get optimum results. The number of excitation conditions selected shall be optimised particularly with regard to detection sensitivity and required analysis time.

Target	Target type	Elements
Mo	Secondary	Cr – Zr (K) Hf – U (L)
Zr	Secondary	Cr – Sr (K) Hf – Bi (L)
Al ₂ O ₃	Barkla	Mo - Ba (K)
CsI	Secondary	Mo – In (K)
Pd	Secondary	Fe – Mo (K) Hf – U (L)
B ₄ C	Barkla	Fe – Mo (K) Hf – U (L)
Ge	Secondary	Cr – Zn (K) Hf – Ta (L)
Zn	Secondary	Ti – Ni (K)
Co	Secondary	K – Mn (K) Cd – La (L)
Ti	Secondary	P – Ca (K)
Si	Secondary	Na – Al (K)
Al	Secondary	Na – Mg (K)
HOPG	Bragg	Na – Sr (K) Y – Hf (L)
(K) K α radiation (L) L α radiation		

Annex D (informative)

List of reference materials applicable for XRF-analysis

Reference materials should be selected containing the elements of interest and covering the total concentration range of interest. Additionally reference materials with a similar composition of the samples under investigation should be selected.

Table D.1 — Reference materials applicable for XRF-analysis

Identification		Matrix
BCR-141R	Institute for Reference Materials and Measurements	Calcareous loam soil
BCR-142R	Institute for Reference Materials and Measurements	Light sandy soil
BCR-143R	Institute for Reference Materials and Measurements	Sewage sludge soil
BCR-667	Institute for Reference Materials and Measurements	Estuarine sediment
ERM-CC690	Institute for Reference Materials and Measurements	Calcareous soil
BCR-144R	Institute for Reference Materials and Measurements	Sewage sludge
BCR-145R	Institute for Reference Materials and Measurements	Sewage sludge
BCR-146R	Institute for Reference Materials and Measurements	Sewage sludge
BCR-038	Institute for Reference Materials and Measurements	Fly ash
GSS-1 up to 16	National Research Centre for Certified Reference Materials (NRCCRM), China	Soil
SO-1 up to 4	Canadian Certified Reference Materials (CCRMP)	Soil
SARM 42	SACCRM, South Africa	Soil
GSD-1 up to 12	National Research Centre for Certified Reference Materials (NRCCRM), China	Sediments
SARM 51	SACCRM, South Africa	Sediment
SARM 52	SACCRM, South Africa	Sediment
LKSD-1 up to 4	Canadian Certified Reference Materials (CCRMP)	Sediment
STSD-1 up to 4	Canadian Certified Reference Materials (CCRMP)	Sediment
TILL-1 up to 3	Canadian Certified Reference Materials (CCRMP)	Soil
N2709	National Institute of Standards & Technology (NIST)	Soil
N2710	National Institute of Standards & Technology (NIST)	Soil
N2711	National Institute of Standards & Technology (NIST)	Soil
N1633b	National Institute of Standards & Technology (NIST)	Fly ash
N2689	National Institute of Standards & Technology (NIST)	Fly ash
N2690	National Institute of Standards & Technology (NIST)	Fly ash
N2691	National Institute of Standards & Technology (NIST)	Fly ash
188-WT-H	SPEX Certiprep, UK	City Treatment Sewage Sludge
188-WT-M	SPEX Certiprep, UK	City Treatment Sewage Sludge
188-WT-L	SPEX Certiprep, UK	City Treatment Sewage Sludge
BAM-U110	Federal Institute for Materials Research and Testing	Wetland Soil

NOTE This list may be incomplete.

Annex E (informative)

Validation

An inter-laboratory comparison supported by DIN was organized by CEN/TC 292 WG 3 in May 2006/August 2006 with participants from 7 member countries. For the inter-laboratory comparison two soil and three waste samples were selected and distributed to the participants. The samples are representing a wide spectrum of contents of elements and a broad variety of the matrix composition.

- Sample 1: soil sample of a sandy-loamy arable soil taken fifty meters beside a highway in Germany, homogenised and homogeneity tested by UBA (Federal Environment Agency).
- Sample 2: sample of contaminated soil of an area subject to periodic flooding of the Saale river in Germany, sandy silt, certified as soil reference material BAM-U110 [14].
- Sample 3: fly ash sample of a household waste incineration plant, homogenised and homogeneity tested by BAM (Federal Institute for Materials Research and Testing).
- Sample 4: ink waste sample, cited by (3) as SAMPLE CEN8/99, homogeneity tested by EU DG JRC (European Commission's DG Joint Research Centre) and used for validation of the digestion standards EN 13656 and EN 13657 [15].
- Sample 5: electronic sludge, cited by (3) as SAMPLE CEN9/99, homogeneity tested by EU DG JRC (European Commission's DG Joint Research Centre) and used for validation of the digestion standards EN 13565 and EN 13657 [15].

All the samples were sent to the participating laboratories as dried, fine grinded less than 90 µm, and homogenised material.

The samples were analysed using ED-XRF and WD-XRF instrumentation as well. As sample preparation methods the pressed pellet and the fused bead technique were applied depending on the available laboratory equipment. For each type of sample and each preparation method 2 test samples were prepared (pellets or beads) and analysed in duplicate. Finally, for each combination 4 results were reported.

Fifteen volunteer laboratories received the samples and all of them transmitted data. For sample 4 and 5 no data from the fused bead preparation method were reported because the commonly used fused bead method was not applicable for this type of samples.

The evaluation of the data was done according to ISO 5725-2, the principal steps of Figure 3 were followed strictly. According to this procedure, the data of two labs were removed. The maximum number of laboratories providing results for the statistical evaluation was therefore thirteen.

For the statistical evaluation the data of the ED-XRF and the WD-XRF technique were combined resulting in up to 14 different data sets for each element. For each sample and preparation method, i.e. pressed pellet or fused bead, the data were evaluated separately. Outliers were determined and eliminated according to ISO 5725-2 using the test of Mandel statistics combined with Grubbs test. The program used for statistical evaluation of the data and calculation of the performance characteristics was written in SPSS [16].

Tables E.1 through E.8 show the performance characteristics of the 5 samples. Results obtained by the pressed pellet preparation method are presented for sample 1 through 5, while for the fused bead method only results from sample 1, 2 and 3 are available. Repeatability and reproducibility were calculated according to the definitions given by ISO 5725-2.

Table E.1 — Validation data of sample 1 (soil) – pressed pellet method

Element	L	N	O	m [mg/kg]	S _R [mg/kg]	V _R [%]	S _r [mg/kg]	V _r [%]	m _{ICPMS} [mg/kg]
Na	12	40	4	8 182	2 402	29 %	492	6 %	8 100
Mg	12	40	4	3 930	947	24 %	238	6 %	3 800
Al	12	44	4	45 039	3 469	8 %	784	2 %	38 000
Si	13	46	0	317 432	44 283	14 %	4 204	1 %	n.d.
P	11	44	4	1 111	409	37 %	26	2 %	820
S	13	46	0	1 011	358	35 %	47	5 %	n.d.
Cl	12	29	4	103	53	51 %	25	24 %	n.d.
K	12	46	4	18 000	2 200	12 %	240	1 %	17 000
Ca	14	48	0	5 353	1 617	30 %	149	3 %	6 400
Ti	13	46	0	1 824	368	20 %	37	2 %	1 800
V	13	50	0	38	14	38 %	4	10 %	35
Cr	12	50	2	64	20	31 %	10	16 %	50
Mn	12	46	4	362	40	11 %	6	2 %	390
Fe	12	46	4	12 190	974	8 %	135	1 %	11 000
Co	12	27	4	6	4	62 %	2	27 %	n.d.
Ni	12	47	4	15	5	31 %	1	9 %	14
Cu	12	49	4	25	7	28 %	2	7 %	34
Zn	12	48	0	99	15	15 %	3	3 %	120
As	12	44	4	8	4	47 %	1	13 %	14
Br	12	42	4	11	3	31 %	2	21 %	n.d.
Rb	12	50	4	97	9	10 %	1	1 %	97
Sr	12	50	4	79	9	11 %	1	1 %	80
Y	13	48	0	16	4	27 %	1	7 %	9
Zr	12	50	4	205	19	9 %	8	4 %	320
Nb	11	44	8	8	3	35 %	1	8 %	18
Sn	12	28	4	6	2	35 %	1	25 %	n.d.
Cs	12	26	4	10	5	47 %	3	26 %	8
Ba	12	50	4	404	64	16 %	13	3 %	360
W	12	31	4	12	15	125 %	2	17 %	8
Pb	12	50	4	78	11	14 %	3	3 %	86
Th	12	40	4	8	3	37 %	1	13 %	n.d.

Table E.2 — Validation data of sample 1 (soil) – fused bead method

Element	L	N	O	m [mg/kg]	S _R [mg/kg]	V _R [%]	S _r [mg/kg]	V _r [%]	m _{ICPMS} [mg/kg]
Na	5	20	4	10 075	836	8 %	518	5 %	8 100
Mg	5	20	4	3 054	347	11 %	89	3 %	3 800
Al	5	20	0	42 233	980	2 %	775	2 %	38 000
Si	5	20	0	350 701	7 042	2 %	3 069	1 %	n.d.
P	5	20	0	707	203	29 %	58	8 %	820
K	5	20	4	19 282	464	2 %	187	1 %	17 000
Ca	6	24	4	4 952	215	4 %	67	1 %	6 400
Ti	5	20	0	1 849	109	6 %	45	2 %	1 800
V	5	10	2	35	2	6 %	4	11 %	35
Cr	5	11	0	67	15	22 %	5	7 %	50
Mn	5	16	0	362	50	14 %	14	4 %	390
Fe	5	20	0	12 013	2 058	17 %	275	2 %	11 000
Cu	5	10	0	69	14	21 %	5	7 %	34
Ba	5	12	4	344	73	21 %	25	7 %	360
Pb	5	10	2	82	1	1 %	2	2 %	86

Table E.3 — Validation data of sample 2 (contaminated soil) – pressed pellet method

Element	L	N	O	m [mg/kg]	S _R [mg/kg]	V _R [%]	S _r [mg/kg]	V _r [%]	m _{ICPMS} [mg/kg]	m _{CERT} [mg/kg]
Na	13	34	4	4 564	1 478	32 %	549	12 %	4 700	n.d.
Mg	14	44	0	10 338	1 812	18 %	963	9 %	11 000	8 380
Al	12	47	8	55 589	3 792	7 %	583	1 %	67 000	50 382
Si	13	51	4	239 921	26 488	11 %	1 720	1 %	n.d.	n.d.
P	12	47	4	5 305	1 078	20 %	78	1 %	4 800	n.d.
S	13	50	4	13 665	2 181	16 %	314	2 %	n.d.	n.d.
Cl	13	51	4	1 782	395	22 %	103	6 %	n.d.	n.d.
K	12	51	8	20 451	1 648	8 %	195	1 %	19 000	20 381
Ca	14	55	4	46 924	7 751	17 %	1 065	2 %	47 000	40 638
Ti	14	51	0	3 474	521	15 %	50	1 %	3 300	3 206
V	14	52	0	65	28	42 %	12	18 %	62	68
Cr	14	55	0	257	37	15 %	8	3 %	220	230
Mn	14	51	0	681	112	17 %	10	2 %	680	621
Fe	14	51	0	30 840	3 976	13 %	404	1 %	25 000	28 229
Co	14	43	0	23	9	39 %	3	13 %	14	16
Ni	13	55	4	108	14	13 %	4	3 %	87	101
Cu	13	55	4	265	28	10 %	7	3 %	240	263
Zn	12	51	4	992	52	5 %	14	1 %	1 000	1 000
As	12	44	8	14	4	30 %	2	11 %	20	16
Se	12	19	3	2	1	27 %	1	25 %	n.d.	n.d.
Br	12	48	5	17	4	25 %	1	5 %	n.d.	n.d.
Rb	13	52	4	92	6	7 %	2	2 %	91	n.d.
Sr	13	52	4	277	22	8 %	5	2 %	280	277
Y	13	52	4	24	4	16 %	1	4 %	17	n.d.
Zr	14	52	0	306	43	14 %	8	3 %	190	n.d.
Nb	13	44	4	14	4	30 %	1	5 %	23	n.d.
Mo	13	28	4	4	1	28 %	1	20 %	n.d.	n.d.
Ag	12	26	2	5	2	33 %	1	26 %	7	n.d.
Cd	14	24	0	8	2	29 %	1	17 %	n.d.	7
Sn	13	45	1	23	7	31 %	3	13 %	51	n.d.
Sb	12	27	4	7	2	30 %	2	26 %	n.d.	n.d.
I	13	16	1	7	2	29 %	4	62 %	n.d.	n.d.
Cs	13	20	4	15	4	28 %	6	40 %	8	n.d.
Ba	14	55	0	1 545	276	18 %	44	3 %	1 400	1 488
Hg	12	47	8	52	11	21 %	2	4 %	37	52
Tl	13	16	4	3	1	33 %			n.d.	n.d.
Pb	14	55	0	197	33	17 %	5	3 %	210	197
Bi	14	19	0	2	1	52 %	1	76 %	n.d.	n.d.
Th	12	43	8	10	2	20 %	1	15 %	n.d.	n.d.
U	12	31	7	7	3	46 %	2	30 %	6	n.d.

Table E.4 — Validation data of sample 2 (contaminated soil) – fused bead method

Element	L	N	O	m [mg/kg]	S _R [mg/kg]	V _R [%]	S _r [mg/kg]	V _r [%]	m _{ICPMS} [mg/kg]	m _{CERT} [mg/kg]
Na	4	19	3	5 958	582	10%	236	4 %	4 700	n.d.
Mg	5	19	0	8 658	789	9 %	340	4 %	11 000	8 380
Al	5	19	0	50 039	1 004	2 %	487	1 %	67 000	50 382
Si	5	19	0	257 753	3 623	1 %	1 435	1 %	n.d.	n.d.
P	4	19	3	3 555	177	5 %	47	1 %	4 800	n.d.
K	4	19	4	19 968	733	4 %	125	1 %	19 000	20 381
Ca	5	23	4	41 361	1 061	3 %	430	1 %	47 000	40 638
Ti	5	19	0	3 397	205	6 %	36	1 %	3 300	3 206
Cr	5	12	0	246	28	12 %	7	3 %	220	230
Mn	5	16	0	623	14	2 %	15	2 %	680	621
Fe	5	19	0	28 921	2 144	7 %	2 150	7 %	25 000	28 229
Cu	5	11	0	263	81	31 %	32	12 %	240	263
Zn	4	7	0	1 024	178	17 %	17	2 %	1 000	1 000
Ba	5	11	0	1 780	495	28 %	132	7 %	1 400	1 488

Table E.5 — Validation data of sample 3 (fly ash) – pressed pellet method

Element	L	N	O	m [mg/kg]	S _R [mg/kg]	V _R [%]	S _r [mg/kg]	V _r [%]	m _{ICPMS} [mg/kg]
Na	13	43	4	14 548	3 484	24 %	515	4 %	15 000
Mg	13	44	4	13 218	2 650	20 %	1 385	10 %	12 000
Al	13	47	4	53 651	6 836	13 %	680	1 %	48 000
Si	13	51	4	176 582	25 716	15 %	1 778	1 %	n.d.
P	12	47	4	5 257	1 036	20 %	112	2 %	4 100
S	12	47	8	14 592	2 483	17 %	185	1 %	n.d.
Cl	13	51	4	5 088	831	16 %	147	3 %	n.d.
K	14	51	0	13 109	1 486	11 %	202	2 %	12 000
Ca	14	55	4	124 013	14 792	12 %	3 144	3 %	100 000
Ti	14	51	0	5 829	829	14 %	99	2 %	5 100
V	13	45	4	46	14	29 %	6	13 %	46
Cr	12	55	7	291	31	11 %	8	3 %	190
Mn	14	51	0	1 161	189	16 %	23	2 %	1200
Fe	14	51	0	45 995	3 994	9 %	498	1 %	43 000
Co	13	42	4	28	14	49 %	6	20 %	17
Ni	13	55	4	104	19	19 %	4	4 %	100
Cu	12	55	8	1 528	114	7 %	17	1 %	1 600
Zn	11	51	8	4 342	273	6 %	81	2 %	4 400
As	13	22	4	12	8	62 %	3	27 %	16
Br	13	31	4	10	2	18 %	1	7 %	n.d.
Rb	13	52	4	51	5	10 %	2	4 %	50
Sr	12	52	8	337	23	7 %	7	2 %	350
Y	13	44	4	16	4	22 %	1	5 %	11
Zr	13	52	4	225	28	13 %	9	4 %	170
Nb	13	36	4	11	2	22 %	1	8 %	19
Mo	12	34	6	14	3	24 %	3	23 %	10
Ag	13	30	1	10	3	35 %	1	13 %	11
Cd	14	33	0	12	4	34 %	2	16 %	12
Sn	13	55	4	283	42	15 %	8	3 %	200
Sb	13	54	4	47	17	36 %	3	7 %	47
Ba	13	55	4	1 867	231	12 %	50	3 %	1 800
W	13	23	4	42	34	81 %	37	88 %	25
Pb	12	51	8	1 210	131	11 %	29	2 %	1 400
Bi	14	21	0	3	2	48 %	1	30 %	5
Th	13	34	2	14	6	44 %	3	19 %	n.d.

Table E.6 — Validation data of sample 3 (fly ash) – fused bead method

Element	L	N	O	m [mg/kg]	S _R [mg/kg]	V _R [%]	S _r [mg/kg]	V _r [%]	m _{ICPMS} [mg/kg]
Na	5	20	4	17 940	1 402	8 %	666	4%	15 000
Mg	5	20	0	12 848	744	6 %	400	3%	12 000
Al	5	20	0	54 578	2 426	4 %	1 786	3%	48 000
Si	5	20	0	197 167	5 562	3 %	3 194	2%	n.d.
P	5	20	4	4 695	220	5 %	89	2%	4 100
K	5	20	0	12 719	761	6 %	294	2%	12 000
Ca	6	24	4	114 789	5 223	5 %	1 961	2%	100 000
Ti	5	20	0	6 021	401	7 %	119	2%	5 100
V	5	11	0	47	26	54 %	8	18%	46
Cr	5	12	0	302	28	9 %	24	8%	190
Mn	5	16	0	1 240	41	3 %	20	2%	1 200
Fe	5	20	1	46 690	2 101	4 %	563	1%	43 000
Cu	5	12	0	1 631	226	14 %	58	4%	1 600

Table E.7 — Validation data of sample 4 (ink waste) – pressed pellet method

Element	L	N	O	m [mg/kg]	S _R [mg/kg]	V _R [%]	S _r [mg/kg]	V _r [%]	m _{ICPMS} [mg/kg]	m _{EUJRC} [mg/kg]
Na	13	32	6	6 119	2 146	35 %	341	6 %	4 200	6787
Mg	13	39	12	1 192	557	47 %	90	8 %	960	960
Al	13	41	9	2 536	683	27 %	194	8 %	3 200	2 111
Si	13	48	4	15 980	4 186	26 %	364	2 %	n.d.	n.d.
P	12	44	4	17 004	5 006	29 %	240	1 %	12 000	14 005
S	13	44	8	35 353	5 774	16 %	478	1 %	n.d.	33 032
Cl	13	48	4	4 630	1 961	42 %	1 436	31 %	n.d.	n.d.
K	13	44	8	1 032	388	38 %	44	4 %	510	965
Ca	14	52	4	124 307	57 376	46 %	5 685	5 %	110 000	107 315
Ti	13	44	0	318	112	35 %	11	3 %	340	232
V	13	32	0	23	13	55 %	5	23 %	20	15
Cr	13	52	4	5 005	1770	35 %	138	3 %	3 400	3 612
Mn	13	48	4	717	320	45 %	18	2 %	580	530
Fe	13	48	4	105 308	37 749	36 %	1 364	1 %	74 000	74 271
Co	13	38	8	30	14	49 %	5	19 %	13	14
Ni	13	37	6	25	12	50 %	3	12 %	19	23
Cu	13	52	4	16 530	6 586	40 %	163	1 %	11 000	12 429
Zn	12	48	4	1 504	584	39 %	28	2 %	1 200	1 203
Br	13	26	2	16	4	28 %	2	11 %	n.d.	n.d.
Rb	13	19	3	11	3	24 %	2	14 %	<5	n.d.
Sr	13	52	4	152	61	40 %	6	4 %	120	117
Y	13	14	2	10	2	18 %	3	31 %	<5	n.d.
Zr	13	47	4	30	13	43 %	5	16 %	41	n.d.
Mo	13	28	3	5	2	46 %	1	21 %	6	5
Sn	13	16	2	5	2	46 %	2	42 %	19	2
Ba	13	40	0	121	39	32 %	10	9 %	100	101
Pb	13	52	4	6 735	3 287	49 %	105	2 %	5 900	5 894
Th	13	20	4	26	17	67 %	3	10 %	n.d.	n.d.

Table E.8 — Validation data of sample 5 (electronic sludge) – pressed pellet method

Element	L	N	O	<i>m</i> [mg/kg]	<i>S_R</i> [mg/kg]	<i>V_R</i> [%]	<i>S_r</i> [mg/kg]	<i>V_r</i> [%]	<i>m</i> _{ICPMS} [mg/kg]	<i>m</i> _{EUJRC} [mg/kg]
Na	12	36	0	16 970	7 880	46 %	1 445	9%	19 000	13 207
Mg	12	32	4	2 712	1 649	61 %	1 564	58 %	2 000	2 163
Al	12	40	0	114 509	20 552	18 %	1 926	2 %	77 000	75 505
Si	12	44	4	26 278	5 565	21 %	598	2 %	n.d.	n.d.
P	11	40	0	5 776	1 383	24 %	186	3 %	4 100	4 864
S	12	40	4	73 081	12 698	17 %	1 114	2 %	n.d.	68 041
Cl	12	44	0	7 684	2 536	33 %	130	2 %	n.d.	n.d.
K	12	44	4	3 025	790	26 %	71	2 %	2 700	3 130
Ca	12	44	4	68 820	9 948	14 %	766	1 %	75 000	54 808
Ti	12	44	4	132	41	31 %	6	5 %	110	116
V	12	17	5	7			5	67 %	11	6
Cr	12	44	0	101	28	28 %	8	8 %	110	85
Mn	12	44	4	777	245	31 %	54	7 %	610	622
Fe	12	44	4	7 016	1 150	16 %	170	2 %	6 100	5 055
Ni	12	48	0	2 133	514	24 %	59	3 %	1 700	1 730
Cu	12	48	0	106 900	29 992	28 %	1 246	1 %	78 000	95 369
Zn	11	44	8	255	102	40 %	7	3 %	230	249
Br	12	36	4	91	15	17 %	5	6 %	n.d.	n.d.
Rb	12	33	0	28	11	40 %	1	4 %	20	n.d.
Sr	12	48	4	203	50	25 %	4	2 %	210	175
Zr	12	44	4	33	17	52 %	5	16 %	34	n.d.
Nb	12	19	0	23	27	115 %	1	5 %	5	n.d.
Mo	12	35	4	9	5	59 %	2	22 %	8	4
Cd	12	17	3	21	27	125 %	7	31 %	n.d.	1
Sn	12	48	4	22 385	4 661	21 %	397	2 %	17 000	16 565
I	12	17	1	77	73	95 %	17	23 %	n.d.	n.d.
Ba	12	39	0	78	31	40 %	8	10 %	86	68
Ta	12	21	5	244	135	55 %	26	11 %	n.d.	n.d.
Pb	12	48	0	9 754	2 646	27 %	171	2 %	10 000	9 380
Th	12	26	4	34	26	77 %	3	10 %	n.d.	n.d.

L number of accepted laboratories

N number of accepted results

O number of outliers removed

*m*_{XRF} mean content of elements specified calculated from N data sets, in mg/kg dry matter

S_R reproducibility standard deviation

S_r repeatability standard deviation

V_R relative reproducibility standard deviation

V_r relative repeatability standard deviation

*m*_{ICPMS} content of elements specified determined by ICP/MS technique after HNO₃/HCL/HF-digestion, in mg/kg dry matter

*m*_{cert} certified content of elements according [14], in mg/kg dry matter

*m*_{EUJRC} mean content of elements specified as given by [15], in mg/kg dry matter

n.d. not determined

NOTE Numbers given in italics indicate informative data according the certification report [2].

Discussion of the results

The validation data show that in general the standard is fit for purpose. In some cases, there are rather high standard deviations (up to 50 %), especially in case low elemental contents below 100 ppb. The fused bead sample preparation technique has been proved to be more accurate. Matrix-specific calibration will increase the accuracies of the results.

Comparison of the means of the two preparation methods of sample 1 and 2 (soil) and 3 (fly ash) show comparable results for most elements within the calculated reproducibility. The results of sample 2 obtained by method EN 15309 are in good agreement with the certified values [14].

Mean concentrations and reproducibility standard deviations of sample 4 (ink waste) are in most cases evidently higher than those reported in EN 13656 due to the results of three labs which have not been identified as outliers according to ISO 5725. The high loss on ignition (around 60 %) of this sample and the matrix properties are possible explanations.

The validation clearly indicates that for soil like matrices the fused bead sample preparation is the method of choice for main group elements, whereas the pressed pellet method is generally applicable with good results for all elements. On the other hand the validation proves that the fused bead sample preparation technique is not be used in case of waste like matrices as sludges.

Bibliography

- [1] ASTM D 6052-97, *Standard Test Method for Preparation and Elemental Analysis of Liquid Hazardous Waste by Energy-Dispersive X-Ray Fluorescence*
- [2] DIN 51418-1, *Röntgenspektralanalyse — Röntgenemissions- und Röntgenfluoreszenz-Analyse (RFA) — Teil 1: Allgemeine Begriffe und Grundlagen*
- [3] DIN 51418-2, *Röntgenspektralanalyse — Röntgenemissions- und Röntgenfluoreszenz-Analyse (RFA) — Teil 2: Begriffe und Grundlagen zur Messung, Kalibrierung und Auswertung*
- [4] EN 13656, *Characterization of waste — Microwave assisted digestion with hydrofluoric (HF), nitric (HNO₃) and hydrochloric (HCl) acid mixture for subsequent determination of elements*
- [5] EN 13657, *Characterization of waste — Digestion for subsequent determination of aqua regia soluble portion of elements*
- [6] IEC 111/24/CD, *Procedures for the Determination of Levels of Regulated Substances in Electrotechnical Products*
- [7] ISO 9516-1:2003, *Iron ores — Determination of various elements by X-ray fluorescence spectrometry — Part 1: Comprehensive procedure*
- [8] ISO 12980:2000, *Carbonaceous material used in the production of aluminium — Green coke and calcined coke for electrodes — Analysis using an X-ray fluorescence method*
- [9] ISO Guide 32, *Calibration in analytical chemistry and use of certified reference materials*
- [10] ISO Guide 33, *Use of certified reference materials*
- [11] ISO Guide 34, *General requirements for the competence of reference material producers*
- [12] ISO TC 183 N691 E, *XRF Analysis of Cu, Pb, Zn, Ni ores and their related products*
- [13] G.R. Lachance and F. Claisse, *Quantitative X-ray fluorescence analysis – Theory and application*, ISBN 0-471-95167-6, John Wiley & Sons, West Sussex, England, 1995.
- [14] Certification Report CRM BAM-U110; BERICHT zur Zertifizierung der Gesamtgehalte und der mit Königswasser extrahierbaren Gehalte der Elemente As, Cd, Co, Cr, Cu, Hg, Mn, Ni, Pb und Zn in einer Bodenprobe - Zertifiziertes Referenzmaterial BAM - U110), Bundesanstalt für Materialforschung und -prüfung (BAM) (http://www.bam.de/index_en.htm), FG I.1 - Anorganisch-chemische Analytik; Referenzmaterialien, Richard-Willstätter-Str. 11, D - 12489 Berlin, September 2006
- [15] H. Muntau, M.I. Kissler, C. Yuste, A. Giove and M. Bianchi, *The validation of CEN draft standards for the analysis of wastes*, JRC, EUR 19680 EN, 2000.
- [16] A. Bühl and P. Zöfel, *SPSS 11 Einführung in die moderne Datenanalyse unter Windows*, Scientific Tools, Pearson Studium, Munich, ed. 8, 2002.
- [17] CEN/TR 15018, *Characterisation of waste — Digestion of waste samples using alkali-fusion techniques*
- [18] EN ISO 12677, *Chemical analysis of refractory products by XRF — Fused cast bead method (ISO 12677:2003)*

- [19] ISO 5725-2:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*
- [20] ISO 14869-2:2002, *Soil quality — Dissolution for the determination of total element content — Dissolution of alkaline fusion*

BSI - British Standards Institution

BSI is the independent national body responsible for preparing British Standards. It presents the UK view on standards in Europe and at the international level. It is incorporated by Royal Charter.

Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover. Tel: +44 (0)20 8996 9000. Fax: +44 (0)20 8996 7400.

BSI offers members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

Buying standards

Orders for all BSI, international and foreign standards publications should be addressed to Customer Services. Tel: +44 (0)20 8996 9001. Fax: +44 (0)20 8996 7001 Email: orders@bsigroup.com You may also buy directly using a debit/credit card from the BSI Shop on the Website <http://www.bsigroup.com/shop>

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

Information on standards

BSI provides a wide range of information on national, European and international standards through its Library and its Technical Help to Exporters Service. Various BSI electronic information services are also available which give details on all its products and services. Contact Information Centre. Tel: +44 (0)20 8996 7111 Fax: +44 (0)20 8996 7048 Email: info@bsigroup.com

Subscribing members of BSI are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration. Tel: +44 (0)20 8996 7002 Fax: +44 (0)20 8996 7001 Email: membership@bsigroup.com

Information regarding online access to British Standards via British Standards Online can be found at <http://www.bsigroup.com/BSOL>

Further information about BSI is available on the BSI website at <http://www.bsigroup.com>.

Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI.

This does not preclude the free use, in the course of implementing the standard, of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained.

Details and advice can be obtained from the Copyright and Licensing Manager. Tel: +44 (0)20 8996 7070 Email: copyright@bsigroup.com