

BS EN 15938:2010  
BS 2000-586:2010



BSI Standards Publication

# Automotive fuels — Ethanol blending component and ethanol (E85) automotive fuel — Determination of electrical conductivity

NO COPYING WITHOUT BSI PERMISSION EXCEPT AS PERMITTED BY COPYRIGHT LAW

*raising standards worldwide*<sup>™</sup>



**National foreword**

This British Standard is the UK implementation of EN 15938:2010.

The UK participation in its preparation was entrusted to Technical Committee PTI/13, Petroleum Testing and Terminology.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© BSI 2010

ISBN 978 0 580 67064 0

ICS 75.160.20

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 October 2010.

**Amendments issued since publication**

| Date | Text affected |
|------|---------------|
|------|---------------|

---

EUROPEAN STANDARD

**EN 15938**

NORME EUROPÉENNE

EUROPÄISCHE NORM

September 2010

ICS 75.160.20

English Version

## Automotive fuels - Ethanol blending component and ethanol (E85) automotive fuel - Determination of electrical conductivity

Carburants pour automobiles - Éthanol comme base de mélange à l'essence et carburant éthanol (E85) pour automobiles - Détermination de la conductivité électrique

Kraftstoffe für Kraftfahrzeuge - Ethanol als Blendkomponente und Ethanolkraftstoff (E85) - Bestimmung der elektrischen Leitfähigkeit

This European Standard was approved by CEN on 7 August 2010.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

| <b>Contents</b>    |   | Page     |
|--------------------|---|----------|
| Foreword.....      |   | 3        |
| <b>1</b>           | <b>Scope .....</b>                                    | <b>4</b> |
| <b>2</b>           | <b>Normative references .....</b>                     | <b>4</b> |
| <b>3</b>           | <b>Terms and definitions .....</b>                    | <b>4</b> |
| <b>4</b>           | <b>Principle.....</b>                                 | <b>5</b> |
| <b>5</b>           | <b>Reagents.....</b>                                  | <b>6</b> |
| <b>6</b>           | <b>Apparatus .....</b>                                | <b>6</b> |
| <b>7</b>           | <b>Sampling.....</b>                                  | <b>6</b> |
| <b>8</b>           | <b>Procedure .....</b>                                | <b>7</b> |
| <b>8.1</b>         | <b>Preparation of the measuring device.....</b>       | <b>7</b> |
| <b>8.1.1</b>       | <b>General.....</b>                                   | <b>7</b> |
| <b>8.1.2</b>       | <b>Cleaning of the sample vessel .....</b>            | <b>7</b> |
| <b>8.1.3</b>       | <b>Cleaning of the measuring cell.....</b>            | <b>7</b> |
| <b>8.1.4</b>       | <b>Filling the sample into the sample vessel.....</b> | <b>7</b> |
| <b>8.2</b>         | <b>Calibration .....</b>                              | <b>7</b> |
| <b>8.2.1</b>       | <b>Interferences .....</b>                            | <b>7</b> |
| <b>8.2.2</b>       | <b>Checking the cell constant .....</b>               | <b>8</b> |
| <b>8.3</b>         | <b>Measurement procedure .....</b>                    | <b>8</b> |
| <b>9</b>           | <b>Expression of results .....</b>                    | <b>8</b> |
| <b>10</b>          | <b>Precision.....</b>                                 | <b>8</b> |
| <b>10.1</b>        | <b>General.....</b>                                   | <b>8</b> |
| <b>10.2</b>        | <b>Repeatability, <math>r</math>.....</b>             | <b>8</b> |
| <b>10.3</b>        | <b>Reproducibility, <math>R</math>.....</b>           | <b>9</b> |
| <b>11</b>          | <b>Test report .....</b>                              | <b>9</b> |
| Bibliography ..... |   | 10       |

## Foreword

This document (EN 15938:2010) has been prepared by Technical Committee CEN/TC 19 “Gaseous and liquid fuels, lubricants and related products of petroleum, synthetic and biological origin”, the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by March 2011, and conflicting national standards shall be withdrawn at the latest by March 2011.

This document was prepared by CEN/TC 19 — Ethanol Task Force under responsibility of its Working Group 21 and is based on DIN 51627-4 [1].

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

## 1 Scope

This European Standard specifies a test method for the determination of the electrical conductivity in ethanol and ethanol (E85) automotive fuel in the range from approximately (0,3 to 5)  $\mu\text{S}/\text{cm}$  at a temperature of 25 °C (see Clause 4). The electrical conductivity is determined from the measured electrical conductance.

The electrical conductivity is an important analytical criterion for the ascertainment and control of anionic and cationic components in ethanol and ethanol (E85) automotive fuel. Some of these components can exhibit corrosive properties.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3170, *Petroleum liquids — Manual sampling (ISO 3170:2004)*

EN ISO 3171, *Petroleum liquids — Automatic pipeline sampling (ISO 3171:1988)*

## 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

### 3.1 electrical conductance

$G$   
reciprocal value of the electrical resistance

NOTE The electrical conductance is expressed in siemens (S) or  $\Omega^{-1}$ .

### 3.2 electrical resistance

$R$   
measurement value of a sample, directly determined by means of the conductivity meter, which increases with the sample length  $l$  and decreases with the sample cross-section  $A$

NOTE The electrical resistance of a sample is expressed in ohms ( $\Omega$ ). The relation between  $R$  and the electrical conductance (3.1) is shown in Equation (1).

$$G = \frac{1}{R} \quad (1)$$

### 3.3 electrical conductivity

$\sigma_T$   
material-dependent measure of the electrical current conducted in an electrical field by the ions present in the solution

NOTE 1 The material-dependent electrical conductivity of a solution is related to the electrical current which is achieved by moving ions of the solution in an electrical field. It is therefore applicable as a method to limit ionic contamination in alcohols and alcohol based fuels. The electrical conductivity  $\sigma_T$  is determined in accordance with Equation (2) from the determined electrical conductance  $G$  (3.1) and the geometrical dimensions of the measuring cell (length and effective cross-section) which are added to form the cell constant  $K$  (3.4):

$$\sigma_T = G \cdot \frac{l}{A} = G \cdot K \quad (2)$$

where:

- $\sigma_T$  is the electrical conductivity of the sample, in  $\mu\text{S cm}^{-1}$  ( $1 \text{ S m}^{-1} = 10^4 \mu\text{S cm}^{-1}$ ), at the measurement temperature  $T$ , in  $^{\circ}\text{C}$ ;
- $l$  is the length of the measured section (geometrical electrode gap), in m or, following the respective unit conversion, in cm, with  $1 \text{ m} = 100 \text{ cm}$ ;
- $A$  is the effective cross-section of the measuring cell, in  $\text{m}^2$  or, following the respective unit conversion, in  $\text{cm}^2$ , with  $1 \text{ m}^2 = 10^4 \text{ cm}^2$ ;
- $K$  is the cell constant of the measuring cell in  $\text{m}^{-1}$  or, following the respective unit conversion, in  $\text{cm}^{-1}$ , with  $1 \text{ m}^{-1} = 0,01 \text{ cm}^{-1}$ .

NOTE 2 The electrical conductivity of a solution depends on the temperature; therefore the temperature is given together with the measurement value. Hence, the electrical conductivity, which is mostly determined at  $25^{\circ}\text{C}$ , is designated as  $\sigma_{25}$ .

NOTE 3 The electrical conductivity depends on ion concentration, ion type, temperature and viscosity of the solution. Therefore its measurement range can be large. "Ultrapure" (deionized and demineralised) water for example, due to its self-dissociation at  $25^{\circ}\text{C}$ , has an electrical conductivity of  $0,054 \text{ 83 } \mu\text{S cm}^{-1}$  ( $5,483 \mu\text{S m}^{-1}$ ).

### 3.4 cell constant

$K$   
geometrical dimension of the measuring cell used to form one value

NOTE The electrical conductivity  $\sigma$  (3.3) cannot be calculated simply from the electrical resistance  $R$  (3.2) and the geometrical cell dimensions because its functional relationship to the cell dimensions is very complex, especially in cases where the geometry is not cubical. Therefore, the measuring cell is calibrated by means of a calibration solution of known electrical conductivity  $\sigma^*$ . The cell constant  $K$  is determined in accordance with Equation (3) from the measured electrical conductance  $G^*$  caused by the calibration solution and its known electrical conductivity  $\sigma^*$ .

$$K = \frac{\sigma_{25}^*}{G_{25}^*} \quad (3)$$

where:

- $K$  is the cell constant of the measuring cell in  $\text{m}^{-1}$  or, following the respective unit conversion, in  $\text{cm}^{-1}$  ( $1 \text{ m}^{-1} = 0,01 \text{ cm}^{-1}$ );
- $G_{25}^*$  is the measured electrical conductance at  $25^{\circ}\text{C}$  in the measuring cell filled with the calibration solution, expressed in S;
- $\sigma_{25}^*$  is the electrical conductivity of the calibration solution at  $25^{\circ}\text{C}$ , in  $\text{S m}^{-1}$  or  $\mu\text{S cm}^{-1}$  ( $1 \text{ S m}^{-1} = 10^4 \mu\text{S cm}^{-1}$ ).

## 4 Principle

The determination of the electrical conductivity in ethanol or ethanol fuel is carried out by means of a direct conductometer (6.1) measurement, using a measuring cell (6.2) suitable for a measurement range of approximately  $(0,01 \text{ to } 5) \mu\text{S cm}^{-1}$ . The measurement is carried out at a sample temperature of  $(25 \pm 0,1)^{\circ}\text{C}$ . As electrical conductivity is strongly depending on temperature, selection of another measurement

temperature and the subsequent conversion of the electrical conductance to a target temperature of 25 °C is not allowed.

## 5 Reagents

**5.1 Ethanol**, absolute, of analytical grade or a similar quality, for cleaning and rinsing the measuring cell (6.2), the sample vessel (6.4) and all the other vessels/containers with which the sample can come into contact.

**5.2 Calibration solution**, KCl solution having a conductivity  $\sigma_{25}$  of max.100  $\mu\text{S cm}^{-1}$ .

**5.3 Water**, deionized or distilled with a maximum conductivity of 1  $\mu\text{S cm}^{-1}$ , for cleaning and rinsing the sample vessel (6.4) and the measuring cell (6.2).

## 6 Apparatus

**6.1 Conductometer**, suitable of being operated with measuring cells (6.2) having a cell constant of approximately (0,1 to 1)  $\text{cm}^{-1}$ , with an alternating voltage of a frequency between 50 Hz and 500 Hz in the measuring cell (6.2), directly displaying the measured conductivity to the nearest 0,01  $\mu\text{S cm}^{-1}$  at least.

A single-point calibration using a single KCl calibration solution (5.2) is sufficient.

**6.2 Measuring cell**, manufactured from a material inert to ethanol or ethanol fuel, for example high-grade steel.

The cell constant should be in the range (0,1 and 1)  $\text{cm}^{-1}$ , preferably close to 0,1  $\text{cm}^{-1}$ . The measuring cell is introduced from above into the sample vessel (6.4), through a cover, and fixed in a way it is completely immersed in the solution to be measured.

**6.3 Temperature sensor**, with an accuracy of  $\pm 0,1$  °C or better, suitable for recording the temperature of the solution subjected to measurement at approximately 25 °C.

This sensor can also be integrated in the measuring cell (6.2). The temperature signal is indicated on a display to the nearest  $\pm 0,1$  °C.

**6.4 Thermostatable glass sample vessel**, having a capacity of 20 ml to 500 ml, into which the sample to be measured is filled.

The sample vessel may be designed as a double-walled glass vessel, for example, which is thermostated by means of a circulating liquid. Alternatively, a thermostatable flow-through vessel may be used, provided it is ensured that the substance to be measured can be maintained at a temperature of  $(25 \pm 0,1)$  °C, the measuring cell (6.2) is completely immersed, and all requirements for cleaning are also complied with.

**6.5 Thermostat with thermometer**, suitable for thermostating the sample vessel (6.4) to  $(25 \pm 0,1)$  °C .

**6.6 Stirrer** for ensuring a homogeneous temperature in the sample by stirring it with a magnetic stirrer and a polytetrafluorethylene-coated stirring magnet.

## 7 Sampling

Unless specified otherwise, samples shall be taken in accordance with EN ISO 3170 or EN ISO 3171 and/or the requirements of national standards regarding the sampling of fuels.



Only glass bottles shall be used for samples. The glass bottles shall be cleaned very thoroughly and rinsed at least twice using the product to be sampled prior to the actual sampling. Avoid a stopper made up of aluminium, brass, copper, lead, plated steel, zinc, natural rubber, leather, cork, polyurethane, PVC and polyamides.

For any other handling of the samples particular attention shall be paid to avoid any risk of further contamination.

## **8 Procedure**

### **8.1 Preparation of the measuring device**

#### **8.1.1 General**

The device manufacturer's instructions shall be followed for the preparation and operation of the conductometer (6.1) and the measuring cell (6.2). The thermostat (6.5) is connected with the sample vessel (6.4) and set to achieve a temperature of  $(25 \pm 0,1)$  °C in the sample vessel.

#### **8.1.2 Cleaning of the sample vessel**

Before the measurement, the sample vessel (6.4) is rinsed several times with ethanol (5.1). The rest of the ethanol used for rinsing shall be dripped off.

If the sample vessel (6.4) has been used before for saline samples, it shall additionally be cleaned thoroughly with water (5.3) prior to using it for ethanol or ethanol fuel. Subsequently, it is dried in the drying oven at min. 80 °C. Afterwards it is rinsed once more using ethanol (5.1). If there are any visible contaminations (e.g. stains) on the inner glass surface of the sample vessel (6.4), the vessel shall not be used.

#### **8.1.3 Cleaning of the measuring cell**

The measuring cell (6.2) is also cleaned by rinsing it thoroughly with ethanol (5.1). If the measuring cell (6.2) had been used before for saline samples, it shall be cleaned thoroughly with water (5.3) prior to using it for ethanol or ethanol fuel. Subsequently, it is thoroughly rinsed with ethanol (5.1). If there are any visible contaminations (e.g. stains) on the inner surface of the measuring cell (6.2), the cell shall not be used.

#### **8.1.4 Filling the sample into the sample vessel**

The sample is filled into the sample vessel (6.4) by pouring it directly from the sample container into the vessel. Should an intermediate vessel be required for transferring, this vessel shall first be cleaned thoroughly using water (5.3), and dried in the drying oven at min. 80 °C. Before the intermediate vessel is used, it is rinsed with ethanol (5.1). Subsequently it shall be dripped off. Rubbing its inside with a fibrous cloth shall be avoided. The size of the intermediate vessel shall be chosen such that it is filled up to at least half its height. If there are any visible contaminations (e.g. stains) at the surface of the glass on the inside of the vessel, the vessel shall not be used.

### **8.2 Calibration**

#### **8.2.1 Interferences**

The measurement values of the electrical conductivity can very strongly be influenced by contaminations in the sample as well as by pollutions of the measuring cell (6.2) (suspended matter, grease or oil). Such interferences cannot easily be detected in routine operation; therefore, the cell constant should regularly be checked by means of a reference solution or the calibration solution (5.2).

### 8.2.2 Checking the cell constant

The cell constant is determined at  $(25 \pm 0,1)$  °C using the calibration solution (5.2), and stored in the device and recorded. In order to enable a fast measurement, the calibration solution (5.2) should be pre-adjusted to a temperature of 25 °C.

NOTE The cell constant is a parameter of the measuring cell. The conductometer (6.1) automatically works out this value with calibration solution (5.2).

The measuring cell (6.2) and the sample vessel (6.4) are cleaned of any adhering rinsing ethanol (5.1). To do so, the sample vessel (6.4) is filled with the calibration solution (5.2) and, after the measuring cell (6.2) has been immersed in the solution, the solution is stirred for at least one minute.

Subsequently, the sample vessel (6.4) is emptied and refilled with the calibration solution (5.2). The measuring cell (6.2) is immersed in the solution, and the lid is closed. In order to achieve a homogeneous temperature distribution, the solution is gently stirred. As soon as a constant temperature of  $(25 \pm 0,1)$  °C has been achieved in the solution, the calibration measurement can be started. The determined cell constant is stored in the device in accordance with the manufacturer's instructions and recorded rounded to the nearest  $0,001 \text{ cm}^{-1}$ .

Following the calibration of the conductometer (6.1) the first measurement using the calibration solution (5.2) is carried out. The sample vessel (6.4) is emptied and refilled with the calibration solution (5.2) without another intermediate cleaning step. The measurement result shall not deviate by more than 2 % relative from the conductivity  $\sigma_{25}^*$  of the calibration solution (5.2).

### 8.3 Measurement procedure

After cleaning the sample vessel (6.4) and the measuring cell (6.2) in accordance with 8.1.2 and 8.1.3, the sample to be measured is filled into the sample vessel (6.4) in accordance with the specifications of 8.1.4, the measuring cell (6.2) is immersed, the lid is closed, and the solution is stirred for at least one minute.

Once a constant temperature of  $(25 \pm 0,1)$  °C has been reached in the solution, the measured value,  $\sigma_T$ , is recorded.

## 9 Expression of results

Report the electrical conductivity of the sample at 25 °C, expressed in  $\mu\text{S cm}^{-1}$ , rounded to the nearest  $0,01 \mu\text{S cm}^{-1}$ .

## 10 Precision

### 10.1 General

The precision statements have been developed in a round robin study according to EN ISO 4259 [2].

NOTE This method has been tested only for ethanol 100 % and ethanol (E85) automotive fuel as a mixture of nominally 85 % (V/V) of ethanol and 15 % (V/V) of gasoline.

### 10.2 Repeatability, $r$

The difference between two test results, obtained by the same operator with the same apparatus under constant operating conditions on identical test material would in the long run, in the normal and correct operation of the test method, exceed the value given in Table 1 in only one case in twenty.

### 10.3 Reproducibility, $R$

The difference between two single and independent results, obtained by different operators working in different laboratories on identical test material, would in the long run, in the normal and correct operation of the test method, exceed the value given in Table 1 only in one case in twenty.

**Table 1 — Repeatability and reproducibility**

| <b>Repeatability</b><br>$\mu S/cm$                       | <b>Reproducibility</b><br>$\mu S/cm$ |
|--|--------------------------------------|
| $r = 0,034\ 2 + 0,012\ 9\ X$                             | $R = 0,068\ 5 + 0,119\ 1\ X$         |
| Where $X$ is the mean of the two results being compared. |                                      |

### 11 Test report

The test report shall contain the following information:

- a) Reference to this European Standard, i.e. EN 15938,
- b) All information required for the complete identification of the sample,
- c) Sampling procedure applied,
- d) Result of the test (see 9), i.e. the electrical conductivity,  $\sigma_{25}$ ,
- e) Any deviation, by agreement or otherwise, from the procedure specified,
- f) Date of the test.

## Bibliography

- [1] DIN 51627-4, *Automotive fuels — Test methods — Part 4: Determination of electrical conductivity in ethanol fuel*
- [2] EN ISO 4259, *Petroleum products — Determination and application of precision data in relation to methods of test (ISO 4259:2006)*

## Buying Parts of BS 2000

Orders for BS 2000 publications should be addressed to either:

### Energy Institute – Library and Information Service

61 New Cavendish Street  
London  
W1G 7AR

Tel: +44 (0)20 7467 7100

Fax: +44 (0)20 7255 1472

[www.energyinst.org.uk](http://www.energyinst.org.uk)

Order standards securely via:

[www.energyinstpubs.org.uk](http://www.energyinstpubs.org.uk)

or:

### British Standards Institution – Customer Services

389 Chiswick High Road  
London  
W4 4AL

Tel: +44 (0)20 8996 9001

Fax: +44 (0)20 8996 7001

[www.bsi-global.com](http://www.bsi-global.com)

Order hard copy standards securely via:

[www.bsi-global.com/bsonline](http://www.bsi-global.com/bsonline)

### Copyright

Copyright exists in all BS 2000 publications. No part of this publication may be reproduced in any form without the prior permission in writing of BSI and the Energy Institute. Enquiries about copyright should be made to the Secretary of PTI/13 at the Energy Institute.







# British Standards Institution (BSI)

BSI is the independent national body responsible for preparing British Standards and other standards-related publications, information and services.

It presents the UK view on standards in Europe and at the international level.

It is incorporated by Royal Charter.

## Revisions

British Standards are updated by amendment or revision. Users of British Standards should make sure that they possess the latest amendments or editions.

It is the constant aim of BSI to improve the quality of our products and services. We would be grateful if anyone finding an inaccuracy or ambiguity while using this British Standard would inform the Secretary of the technical committee responsible, the identity of which can be found on the inside front cover.

**Tel: +44 (0)20 8996 9001 Fax: +44 (0)20 8996 7001**

BSI offers Members an individual updating service called PLUS which ensures that subscribers automatically receive the latest editions of standards.

**Tel: +44 (0)20 8996 7669 Fax: +44 (0)20 8996 7001**

**Email: [plus@bsigroup.com](mailto:plus@bsigroup.com)**

## Buying standards

You may buy PDF and hard copy versions of standards directly using a credit card from the BSI Shop on the website [www.bsigroup.com/shop](http://www.bsigroup.com/shop). In addition all orders for BSI, international and foreign standards publications can be addressed to BSI Customer Services.

**Tel: +44 (0)20 8996 9001 Fax: +44 (0)20 8996 7001**

**Email: [orders@bsigroup.com](mailto:orders@bsigroup.com)**

In response to orders for international standards, it is BSI policy to supply the BSI implementation of those that have been published as British Standards, unless otherwise requested.

## Information on standards

BSI provides a wide range of information on national, European and international standards through its Knowledge Centre.

**Tel: +44 (0)20 8996 7004 Fax: +44 (0)20 8996 7005**

**Email: [knowledgecentre@bsigroup.com](mailto:knowledgecentre@bsigroup.com)**

Various BSI electronic information services are also available which give details on all its products and services.

**Tel: +44 (0)20 8996 7111 Fax: +44 (0)20 8996 7048**

**Email: [info@bsigroup.com](mailto:info@bsigroup.com)**

BSI Subscribing Members are kept up to date with standards developments and receive substantial discounts on the purchase price of standards. For details of these and other benefits contact Membership Administration.

**Tel: +44 (0)20 8996 7002 Fax: +44 (0)20 8996 7001**

**Email: [membership@bsigroup.com](mailto:membership@bsigroup.com)**

Information regarding online access to British Standards via British Standards Online can be found at [www.bsigroup.com/BSOL](http://www.bsigroup.com/BSOL)

Further information about BSI is available on the BSI website at [www.bsigroup.com/standards](http://www.bsigroup.com/standards)

## Copyright

Copyright subsists in all BSI publications. BSI also holds the copyright, in the UK, of the publications of the international standardization bodies. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. This does not preclude the free use, in the course of implementing the standard of necessary details such as symbols, and size, type or grade designations. If these details are to be used for any other purpose than implementation then the prior written permission of BSI must be obtained. Details and advice can be obtained from the Copyright & Licensing Manager.

**Tel: +44 (0)20 8996 7070**

**Email: [copyright@bsigroup.com](mailto:copyright@bsigroup.com)**

### BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

Tel +44 (0)20 8996 9001

Fax +44 (0)20 8996 7001

[www.bsigroup.com/standards](http://www.bsigroup.com/standards)