

BS EN 16166:2012



BSI Standards Publication

Sludge, treated biowaste and soil — Determination of adsorbable organically bound halogens (AOX)

bsi.

...making excellence a habit.™

National foreword

This British Standard is the UK implementation of EN 16166:2012.

The UK participation in its preparation was entrusted to Technical Committee H/-/4, Environmental testing programmes.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2012
Published by BSI Standards Limited 2012

ISBN 978 0 580 56586 1

ICS 13.030.01; 13.080.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 September 2012.

Amendments issued since publication

Date	Text affected
------	---------------

EUROPEAN STANDARD

EN 16166

NORME EUROPÉENNE

EUROPÄISCHE NORM

August 2012

ICS 13.030.01; 13.080.10

English Version

Sludge, treated biowaste and soil - Determination of adsorbable organically bound halogens (AOX)

Boues, biodéchets traités et sols - Détermination des composés organiques halogénés adsorbables (AOX)

Schlamm, behandelter Bioabfall und Boden - Bestimmung von adsorbierbaren organisch gebundenen Halogenen (AOX)

This European Standard was approved by CEN on 24 May 2012.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Principle.....	5
5 Interferences	5
6 Reagents.....	6
7 Apparatus	7
7.1 Apparatus for combustion and detection	7
7.2 Equipment for adsorption.....	7
7.3 Equipment for sample preparation	7
8 Sample storage and pretreatment.....	7
8.1 Sample storage	7
8.2 Sample pretreatment	8
9 Procedure	8
9.1 General.....	8
9.2 Adsorption and inorganic halide removal.....	8
9.3 Combustion	8
9.4 Initial calibration	9
9.4.1 General.....	9
9.4.2 Calibration of the entire system	9
9.4.3 Check of argentometric measuring device	10
9.5 Recalibration	10
9.6 Blank determination	10
10 Calculation.....	11
10.1 Method of calculation.....	11
10.1.1 Entire system calibration (according to 9.4.2).....	11
10.1.2 Direct argentometric measurement device calibration (according to 9.4.3)	11
10.2 Expression of results	12
11 Precision.....	12
12 Test report	12
Annex A (informative) Repeatability and reproducibility data.....	13
A.1 Materials used in the interlaboratory comparison study	13
A.2 Interlaboratory comparison results	14
Annex B (informative) Storage of activated carbon.....	15
Annex C (informative) Schematic diagram of an AOX apparatus	16
Bibliography	17

Foreword

This document (EN 16166:2012) has been prepared by Technical Committee CEN/TC 400 "Project Committee - Horizontal standards in the fields of sludge, biowaste and soil", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by February 2013, and conflicting national standards shall be withdrawn at the latest by February 2013.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

The preparation of this document by CEN is based on a mandate by the European Commission (Mandate M/330), which assigned the development of standards on sampling and analytical methods for hygienic and biological parameters as well as inorganic and organic determinants, aiming to make these standards applicable to sludge, treated biowaste and soil as far as this is technically feasible.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

This document is the result of a desk study "Horizontal European standard for determination of AOX in sewage sludge and comparable matrices" in the project "Horizontal" and aims at evaluating the latest developments in assessing AOX in sludge, soil, treated biowaste and neighbouring fields. After an evaluation study, in which the ruggedness of the method was studied, a European wide validation of the draft standard has taken place. The results of the desk studies as well as the evaluation and validation studies have been subject to discussions with all parties involved in the evaluation.

This European Standard is applicable and validated for several types of matrices as indicated in Table 1 (see also Annex A for the results of the validation).

Table 1 — Matrices for which this European Standard is applicable and validated

Matrix	Materials used for validation
Sludge	Municipal sludge
Compost	Fresh compost
	Compost
Soil	Sludge amended soil
	Agricultural soil

WARNING — Persons using this European Standard should be familiar with usual laboratory practice. This European Standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this European Standard be carried out by suitably trained staff.

1 Scope

This European Standard specifies an empirical method for the direct determination of organically bound chlorine, bromine and iodine (but not fluorine) adsorbed and occluded to the sample matrix. Non-volatile organically bound halogens adsorbable on activated carbon present in the aqueous phase of the sample prior to drying or adsorbed to sample surface are included in the determination.

This European Standard is intended for analysis of sludge, treated biowaste or soil in concentrations ranging from 5 mg/kg dry matter to approximately 6 g/kg dry matter. The exact concentration range covered depends on the instrument used for determination.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 16179, *Sludge, treated biowaste and soil — Guidance for sample pretreatment*

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696)*

EN ISO 5667-15, *Water quality — Sampling — Part 15: Guidance on the preservation and handling of sludge and sediment samples (ISO 5667-15)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

adsorbable organically bound halogens

AOX

equivalent amount of chlorine, bromine, and iodine contained in organic compounds, expressed as chloride when determined according to this European Standard

4 Principle

Activated carbon is added to a dried, homogenised solid sample. Inorganic halides are eluted and at the same time water soluble organic compounds are adsorbed onto the activated carbon by shaking with acidified nitrate solution.

The loaded carbon/sample mixture is combusted in an oxygen stream.

The hydrogen halides produced are absorbed followed by determination of the halide ions by an argentometric titration, such as microcoulometry. The result is expressed as the mass concentration of chloride.

5 Interferences

Sparingly soluble or occluded inorganic halides are included in the determination and may, if present, give a significant positive bias. Adequate washing is essential to remove inorganic interference.

NOTE Halogenated substances that volatilise at 105 °C are lost.

Organic bromine and iodine compounds may, during combustion, lead to the formation of elemental bromine or iodine respectively or to the formation of halogen oxides. The determination of these AOX fractions may be incomplete, thus leading to negative bias.

6 Reagents

Use only reagents of recognised analytical grade and water grade 1 in accordance with EN ISO 3696.

The AOX contribution from water, reagents and gases should be significantly lower than the lowest AOX content to be determined. The overall AOX content of water, chemicals, and gases shall be checked by measuring the total blank (see 9.6).

6.1 Activated carbon, grain size approximately 10 µm to 50 µm.

For the storage of activated carbon, see Annex B.

The blank value of the washed activated carbon shall be less than 15 µg of chloride equivalent per gram of activated carbon.

6.2 Nitric acid, HNO₃, ρ = 1,4 g/ml, 65 % (mass fraction) solution.

6.3 Hydrochloric acid, c(HCl) = 0,100 mol/l.

The molarity shall be precisely known, since the acid is used for checking the micro-titration (see 9.4.3).

6.4 Sulfuric acid, H₂SO₄, ρ = 1,84 g/ml.

6.5 Gases for combustion, e.g. oxygen (O₂), or a mixture of oxygen and an inert gas.

6.6 Sodium nitrate, NaNO₃, for the preparation of stock solution.

6.7 Nitrate stock solution, acidified, c(NaNO₃) = 0,2 mol/l.

Dissolve 17 g of sodium nitrate (6.6) in water in a 1 000 ml volumetric flask, add 15 ml of nitric acid (6.2), and make up to volume with water.

6.8 Nitrate washing solution, c(NaNO₃) = 0,01 mol/l.

Pipette 50 ml of the nitrate stock solution (6.7) in a 1 000 ml volumetric flask, and make up to volume with water.

6.9 Methanol, CH₃OH.

6.10 4-Chlorophenol stock solution, C₆H₅ClO, equivalent to AOX = 2,0 g/l.

Dissolve 0,725 g of 4-chlorophenol in methanol (6.9) in a 100 ml calibrated flask and make up to volume with methanol (6.9).

6.11 4-Chlorophenol working solutions, equivalent to AOX = 0,1 g/l and 0,5 g/l AOX, respectively.

Pipette 5 ml and 25 ml of 4-chlorophenol stock solution (6.10) into two separate 100 ml calibrated flasks, and make up to volume with methanol (6.9).

The stock solution (6.10) may be stored for at least one month and the working solutions (6.11) for one week in a refrigerator in glass bottles.

7 Apparatus

7.1 Apparatus for combustion and detection

7.1.1 Combustion apparatus, a furnace capable of being heated to at least 950 °C, equipped with a quartz tube approximately 30 cm long with an internal diameter of between 2 cm and 4 cm (see Annex C).

NOTE It is essential that the combustion temperature is sufficient. Temperatures below 950 °C are likely to result in poor recovery and increased variability.

7.1.2 Quartz sample boat, to fit in the quartz tube.

7.1.3 Argentometric measuring device for determining halide concentrations, e.g. a microcoulometer, capable of determining at least 0,03 µmol/l chloride with a repeatability variation coefficient of less than 10 %, or an equivalent device to determine chloride ions.

7.1.4 Absorber, filled with sulfuric acid (6.4), to dry the gas stream and designed so that the acid does not backflush into the furnace.

7.1.5 Syringe, to pipette volumes of 1 µl to 10 µl of hydrochloric acid (6.3) or 4-chlorophenol solutions (6.10 and 6.11).

7.2 Equipment for adsorption

7.2.1 Filtration apparatus, e.g. with a funnel capacity of 0,15 l and a filter diameter of 25 mm.

7.2.2 Low-halide polycarbonate membrane filter, to fit the filtration apparatus (7.2.1), with a pore size of 0,45 µm, or any equivalent filtration material, such as a dedicated quartz filter for AOX determination.

7.2.3 Conical flask (Erlenmeyer flask) of 25 ml capacity with ground glass stopper or 12 ml to 20 ml screw cap vial with polytetrafluoroethene (PTFE) lined cap.

7.2.4 Mechanical shaker.

7.3 Equipment for sample preparation

7.3.1 Porcelain evaporating dish.

7.3.2 Oven with forced ventilation or natural ventilation through adjustable vents adjustable to (105 ± 5) °C.

7.3.3 Desiccator, provided with a suitable desiccant.

7.3.4 Analytical mill or porcelain mortar.

7.3.5 Precision balance.

8 Sample storage and pretreatment

8.1 Sample storage

For the sampling and storage of sludge samples refer to EN ISO 5667-15.

Samples shall be stored in suitable containers with an appropriate closure material such as PTFE. Samples to be frozen may be stored in aluminium containers pre-cleaned by heating to 450 °C for minimum 4 h or by rinsing with a non-chlorinated solvent.

Samples shall be kept cold at a temperature $< 8\text{ }^{\circ}\text{C}$ and in the dark. The sample pretreatment should take place within 24 h of sampling. Alternatively, samples may be frozen ($-18\text{ }^{\circ}\text{C}$) directly after sampling and kept frozen for a maximum of one month before sample pretreatment.

8.2 Sample pretreatment

Pretreat the samples according to EN 16179, if not otherwise specified.

Cool the dried test sample in a desiccator, comminute and homogenize in an analytical mill or porcelain mortar (7.3.4) to a particle size of no more than 0,1 mm.

Store the ground material in a desiccator or a tightly closed glass container.

The homogenised wet sample portion may alternatively be freeze-dried, which in some cases makes the homogenizing of the dried sample easier.

9 Procedure

9.1 General

The test sample taken for analysis shall have an AOX value within the optimal working range of the instrument, which is generally between $1\text{ }\mu\text{g}$ to $30\text{ }\mu\text{g}$ (absolute amount).

9.2 Adsorption and inorganic halide removal

Ensure that the dried, ground test sample is homogenised by stirring or shaking before taking the test portion. Prepare the test portion for combustion as follows:

Transfer a homogenised test portion of 5 mg to 100 mg depending on the expected AOX content to a conical flask or screw cap vial (7.2.3).

Laboratories should evaluate the methods' repeatability for the test sample mass routinely used. Test sample masses below 25 mg often result in increased repeatability standard deviation.

Add 20 mg to 30 mg activated carbon (6.1) and 10 ml nitrate stock solution (6.7) to the sample. The amount of activated carbon shall be the same for all test samples, blank determinations etc.

Shake for 1 h using a mechanical shaker (7.2.4).

Filter the suspension through a filtration apparatus (7.2.1) using nitrate washing solution (6.8). Wash the filter cake with small amounts of nitrate washing solution. Use a total volume of 25 ml to ensure complete transfer.

Place the moist filter and the filter cake into the quartz sample boat (7.1.2) and proceed according to 9.3.

The laboratory should ensure that the nitrate washing procedure is sufficient to remove all inorganic halides typically found in the types of samples routinely analysed. If the concentration of AOX is found to decrease with additional washing stages, then the number of nitrate washes should be increased accordingly.

9.3 Combustion

The temperature in the combustion apparatus (7.1.1) shall be at least $950\text{ }^{\circ}\text{C}$; select other operating parameters in accordance with the manufacturer's instructions.

Connect the gas supply to the quartz tube (7.1.1) and the quartz tube to the absorber (7.1.4). Do not allow back-siphoning of the sulfuric acid (6.4) into the quartz tube; this can happen if a drop in temperature or pressure occurs.

Adjust the flow of gas (6.5) to approximately 150 ml/min or follow manufacturer's instructions.

Introduce the quartz sample boat (7.1.2) into the heated zone of the tube, following the manufacturer's instructions.

NOTE A number of experimental variables, e.g. the amount of activated carbon applied, the test portion size, the dimensions of the furnace, residence time, combustion temperature, and the adjustment of the gas flow depend on the type of apparatus and inappropriate selection may adversely affect the result.

It was demonstrated that carbon to sample ratios lower than 1 (1:1) may lead to incomplete combustion of the sample/carbon mixture (seen as tailing during the argentometric detection). If this is the case, the amount of activated carbon should be raised when test portions larger than 20 mg to 30 mg are analysed. The amount of activated carbon for the corresponding blank determinations should be raised as well.

9.4 Initial calibration

9.4.1 General

Calibrate the analytical instrument (7.1) using one of the two methods described below. The calibration of the entire system is the preferred method (see 9.4.2). The method described in 9.4.3 is useful for checking the microcoulometric cell.

9.4.2 Calibration of the entire system

Calibrate the system by analysing at least three different levels of AOX covering the whole of the relevant working range (e.g. 0,5 µg, 5,0 µg and 10,0 µg Cl⁻ or 1,0 µg, 10,0 µg and 20,0 µg Cl⁻) in the following way:

Transfer 20 mg to 30 mg activated carbon (6.1) to the filtration apparatus (7.2.1) and moisten it with a small amount of nitrate washing solution (6.8).

Place the moist filter and the activated carbon into the quartz sample boat (7.1.2).

Inject 5 µl or 10 µl of the 4-chlorophenol stock solution (6.10) or a 4-chlorophenol working solution (6.11) giving the desired amount of AOX (see Table 2) and proceed with combustion according to 9.3.

Table 2 — Suggested calibration solutions

Volume added µl	Concentration of solution µg/µl	Amount of AOX µg
5	0,1	0,5
10	0,1	1,0
10	0,5	5,0
5	2,0	10,0
10	2,0	20,0

Establish a linear relationship between the amount of AOX analysed and the instrument response according to Formula (1):

$$N_i = i + s \times m_{Cl} \quad (1)$$

where

N_i is the instrument response, in Coulomb (C) or instrument-dependent units;

i is the y-axis intercept, in Coulomb (C) or instrument-dependent units;

s is the slope, in Coulomb (C) or instrument-dependent units per microgram (μg) Cl;

m_{Cl} is the amount of AOX, in micrograms (μg) Cl.

Use the slope (s) for subsequent calculation of mass concentrations of AOX (see 10.1.1).

9.4.3 Check of argentometric measuring device

In the case of microcoulometric determination, check the instrument within the relevant working range, using at least one test solution as follows:

Inject directly with a syringe (7.1.5) a known volume within the range of 5 μl to 8 μl of the hydrochloric acid solution (6.3) into the titration cell.

Measure the quantity of charge transferred in this test.

Coulometry theory assumes a 100 % current yield. But for practical purposes, it is important to determine the actual current yield. Obtain the correction factor by using Formula (2):

$$Q = a \times Q_t \quad (2)$$

where

Q is the measured quantity of charge for the hydrochloric acid sample, in Coulomb (C);

Q_t is the theoretical quantity of charge for the hydrochloric acid sample, in Coulomb (C);

a is the correction factor.

Obtain the theoretical quantity of charge, Q_t , using Formula (3):

$$Q_t = V \times c_{\text{Cl}} \times F \quad (3)$$

where

V is the volume of the hydrochloric acid solution, in litres (l);

c_{Cl} is the chloride concentration of the hydrochloric acid solution, in mol per litre (mol/l);

F is the Faraday constant, $F = 96\,487$ C/mol.

The measuring device is suitable for the analysis, if the correction factor, a , is in the range of 0,97 to 1,03.

9.5 Recalibration

For every batch of samples, analyse at least two calibration standards with the concentrations of (20 ± 10) % and (80 ± 10) % of the established linear range and calculate the straight line from these measurements. If the straight line falls within 95 % of the initial calibration relationship according to Formula (1), the initial calibration line is assumed to be valid. If not, a new calibration line shall be established according to 9.4.

9.6 Blank determination

Suspend 20 mg to 30 mg of activated carbon (6.1) in 10 ml of nitrate stock solution (6.7) and proceed in the same way as described in 9.2.

The amount of activated carbon used for the blank determination should be equal to the amount used for the analysis of test samples.

Carry out at least two blank determinations in each series and use the average blank value for subsequent calculations.

10 Calculation

10.1 Method of calculation

10.1.1 Entire system calibration (according to 9.4.2)

Calculate the mass concentration of adsorbable organically bound halogens, expressed as chloride, using Formula (4):

$$Q_{\text{Cl(AOX)}} = \frac{N - N_0}{m_{\text{T}} \times s} \quad (4)$$

where

$Q_{\text{Cl(AOX)}}$ is the mass concentration of adsorbable organically bound halogens, expressed as chloride concentration on dry matter basis, in milligrams per kilogram (mg/kg);

N_0 is the average value for the blank, according to 9.6, in Coulomb (C) or instrument-dependent units;

N is the measured value for the test sample, in Coulomb (C) or instrument-dependent units;

s is the slope of the calibration curve as defined in Formula (1);

m_{T} is the mass of the dried test sample, in grams (g).

10.1.2 Direct argentometric measurement device calibration (according to 9.4.3)

Calculate the mass concentration of adsorbable organically bound halogens, expressed as chloride, using Formula (5):

$$Q_{\text{Cl(AOX)}} = \frac{N - N_0}{m_{\text{T}} \times s} \times \frac{M \times a}{F} \times 10^6 \quad (5)$$

where

$Q_{\text{Cl(AOX)}}$ is the mass concentration of adsorbable organically bound halogens, expressed as chloride concentration on dry matter basis, in milligrams per kilogram (mg/kg);

N_0 is the average value for the blank, according to 9.6, in Coulomb (C);

N is the measured value for the test sample, in Coulomb (C);

M is the molar mass of chloride, $M = 35,45$ g/mol;

m_{T} is the mass of the dried test sample, in grams (g);

a is the factor relating the theoretical to the actual quantity of charge (see Formula (2));

F is the Faraday constant, $F = 96\,487$ C/mol.

10.2 Expression of results

The result shall be expressed in milligrams per kilogram (mg/kg) chloride on dry matter basis and reported to two significant figures.

11 Precision

The performance characteristics of the method have been evaluated, see Annex A.

12 Test report

The test report shall contain at least the following information:

- a) a reference to this European Standard (EN 16166);
- b) complete identification of the sample;
- c) expression of results, according to 10.2;
- d) any details not specified in this European Standard or which are optional, as well as any factor which may have affected the results.

Annex A (informative)

Repeatability and reproducibility data

A.1 Materials used in the interlaboratory comparison study

The interlaboratory comparison of adsorbable organically bound halogens (AOX) in sludge, treated biowaste and soil was carried out with 13 to 15 European laboratories on six materials. Detailed information can be found in the final report on the interlaboratory comparison study mentioned in [6].

Table A.1 lists the types of materials tested.

Table A.1 — Materials tested in the interlaboratory comparison of the determination of adsorbable organically bound halogens (AOX) in sludge, treated biowaste and soil

Grain size	Sample	Material
Sludge (< 0,5 mm)	Sludge 1	Mix of municipal waste water treatment plant sludges from North Rhine Westphalia, Germany
	Sludge 2	Mix of municipal waste water treatment plant sludges from North Rhine Westphalia, Germany
Fine grained (< 2,0 mm)	Compost 1	Fresh compost from Vienna, Austria
	Compost 2	Compost from Germany
	Soil 4	Sludge amended soil from Hohenheim, Germany
	Soil 5	Agricultural soil from Reading, United Kingdom

A.2 Interlaboratory comparison results

The statistical evaluation was conducted according to ISO 5725-2. The average values, the repeatability standard deviation (s_r) and the reproducibility standard deviation (s_R) were obtained (Table A.2).

Table A.2 — Results of the interlaboratory comparison studies of determination of adsorbable organically bound halogens (AOX) in sludge, treated biowaste and soil

Matrix	l	n	n_o	$\bar{\bar{x}}$ mg/kg	s_R mg/kg	$C_{V,R}$ %	s_r mg/kg	$C_{V,r}$ %
Sludge 1	17	69	1	241	27,8	13,0	19,6	9,2
Sludge 2	15	45	4	188	15,2	8,1	8,7	4,6
Compost 1	6	27	0	42,6	11,5	27,0	3,3	7,8
Compost 2	14	54	1	46,4	11,8	25,3	5,0	10,7
Soil 4	14	62	0	26,7	9,1	34,2	2,6	9,9
Soil 5	13	62	0	21,5	6,8	31,7	3,0	13,8

Explanation of symbols	
l	number of participating laboratories
n	number of analytical results after outlier rejection
n_o	number of outliers
$\bar{\bar{x}}$	total mean of results (without outliers)
s_R	reproducibility standard deviation
$C_{V,R}$	coefficient of variation of reproducibility
s_r	repeatability standard deviation
$C_{V,r}$	coefficient of variation of repeatability

Annex B (informative)

Storage of activated carbon

Activated carbon of an adequate adsorption capacity and of low content of inorganic halogenide, suitable for the AOX determination, is commercially available. After opening of the original container, the activated carbon can become contaminated with adsorbable compounds (including organic halogens), coming from the air. The activated carbon can become inactive within five days.


In order to keep the carbon blank low, take small quantities sufficient for one day work, e.g. 1,5 g to 2 g, and transfer to suitable sealed glass containers.

Use the contents of the container on the day of its opening. Discard the unused remainder at the end of the day.

Keep the remaining stock sealed. If the carbon blank is too high, use another batch after checking its blank.

Annex C (informative)

Schematic diagram of an AOX apparatus



Key

- 1 sample inlet for AOX
- 2 AOX sample boat
- 3 furnace
- 4 quartz tube
- 5 absorber filled with sulfuric acid
- 6 titration cell
- 7 stirrer
- 8 control device for temperature, gas flow
- 9 combustion gas inlet

Figure C.1 — Schematic diagram of an AOX apparatus [EN ISO 9562]

Bibliography

- [1] EN 15934, *Sludge, treated biowaste, soil and waste — Calculation of dry matter fraction after determination of dry residue or water content*
- [2] EN ISO 5667-13, *Water quality — Sampling — Part 13: Guidance on sampling of sludges (ISO 5667-13)*
- [3] EN ISO 9562, *Water quality — Determination of adsorbable organically bound halogens (AOX) (ISO 9562)*
- [4] ISO 5725-2, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*
- [5] DIN 38414-18, *German standard methods for the examination of water, waste water and sludge; sludge and sediments (group S); determination of adsorbed organically bound halogens (AOX) (S 18)*
- [6] Horizontal Project Interlaboratory Comparison, www.horizontal.ecn.nl

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email bsmusales@bsigroup.com.

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

Useful Contacts:

Customer Services

Tel: +44 845 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 845 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com



...making excellence a habit.™