

BS EN 16175-1:2016



BSI Standards Publication

# Sludge, treated biowaste and soil — Determination of mercury

Part 1: Cold-vapour atomic absorption  
spectrometry (CV-AAS)

### National foreword

This British Standard is the UK implementation of EN 16175-1:2016.

BSI, as a member of CEN, is obliged to publish EN 16175-1 as a British Standard. However, attention is drawn to the fact that during the development of this European Standard, the UK committee voted against its approval as a European Standard.

The UK committee is concerned that some of the conditions attached to the validation trials lack clarity.

It is also the opinion of the UK committee that validation trials for horizontal methods should include conditions for the preparation and extraction of the original sample.

The UK participation in its preparation was entrusted by Technical Committee H/-, Health and Environment Sector Policy and Strategy Committee, to Subcommittee H/-/4, Environmental Testing Programmes.

A list of organizations represented on this subcommittee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2016.  
Published by BSI Standards Limited 2016

ISBN 978 0 580 91017 3

ICS 13.030.01; 13.080.10

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 November 2016.

### Amendments/corrigenda issued since publication

Date	Text affected
------	---------------

---

EUROPEAN STANDARD

**EN 16175-1**

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2016

ICS 13.030.01; 13.080.10

Supersedes CEN/TS 16175-1:2013

English Version

## Sludge, treated biowaste and soil - Determination of mercury - Part 1: Cold-vapour atomic absorption spectrometry (CV-AAS)

Boues, bio-déchets traités et sols - Détermination du mercure - Partie 1: Spectrométrie d'absorption atomique de vapeur froide (SAA-VP)

Schlamm, behandelter Bioabfall und Boden - Bestimmung von Quecksilber - Teil 1: Kaltdampf-Atomabsorptionsspektrometrie (CV-AAS)

This European Standard was approved by CEN on 19 March 2016.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

<b>Contents</b>	<b>Page</b>
<b>European foreword</b> .....	<b>3</b>
<b>Introduction</b> .....	<b>4</b>
<b>1 Scope</b> .....	<b>5</b>
<b>2 Normative references</b> .....	<b>5</b>
<b>3 Principle</b> .....	<b>5</b>
<b>4 Interferences</b> .....	<b>5</b>
<b>5 Reagents</b> .....	<b>6</b>
<b>6 Apparatus</b> .....	<b>7</b>
6.1 Usual laboratory apparatus.....	7
6.2 Atomic absorption spectrometer (AAS).....	7
6.3 Automated sample introduction system .....	8
6.4 Cold-vapour generator, batch system or an automated flow injection analysis system (FIA) .....	8
<b>7 Procedure</b> .....	<b>8</b>
7.1 Test solution .....	8
7.2 Test blank solution.....	8
7.3 Preparation of the calibration solutions .....	8
7.4 Calibration .....	8
7.5 Measurement of test sample .....	9
<b>8 Calculation and expression of results</b> .....	<b>9</b>
8.1 Calculation .....	9
8.2 Expression of results.....	9
<b>9 Performance data</b> .....	<b>9</b>
<b>10 Test report</b> .....	<b>10</b>
<b>Annex A (informative) Repeatability and reproducibility data</b> .....	<b>11</b>
A.1 Materials used in the interlaboratory comparison study .....	11
A.2 Interlaboratory comparison results .....	11
<b>Bibliography</b> .....	<b>13</b>

## European foreword

This document (EN 16175-1:2016) has been prepared by Technical Committee CEN/TC 444 “Test methods for environmental characterization of solid matrices”, the secretariat of which is held by NEN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2017, and conflicting national standards shall be withdrawn at the latest by April 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes CEN/TS 16175-1:2013.

The preparation of the previous edition of this analytical method by CEN is based on a mandate by the European Commission (Mandate M/330), which assigned the development of standards on sampling and analytical methods for hygienic and biological parameters as well as inorganic and organic determinants, aiming to make these standards applicable to sludge, treated biowaste and soil as far as this is technically feasible.

EN 16175, *Sludge, treated biowaste and soil — Determination of mercury* comprises the following parts:

- *Part 1: Cold-vapour atomic absorption spectrometry (CV-AAS);*
- *Part 2: Cold-vapour atomic fluorescence spectrometry (CV-AFS).*

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## Introduction

This European Standard is applicable and validated for several types of matrices as indicated in Table 1 (see Annex A for the results of validation).

**Table 1 — Matrices for which this European Standard is applicable and validated**

Matrix	Materials used for validation
Sludge	Municipal sludge
Biowaste	Compost
Soil	Soil

**WARNING — Persons using this European Standard should be familiar with usual laboratory practice. This European Standard does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.**

**IMPORTANT — It is absolutely essential that tests conducted according to this European Standard be carried out by suitably trained staff.**

## 1 Scope

This European Standard specifies a method for the determination of mercury in *aqua regia* or nitric acid digests of sludge, treated biowaste and soil, obtained according to EN 16173 or EN 16174 using cold-vapour atomic absorption spectrometry (CV-AAS). The lower working range limit is 0,03 mg/kg (dry matter basis).

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15934, *Sludge, treated biowaste, soil and waste — Calculation of dry matter fraction after determination of dry residue or water content*

EN 16173, *Sludge, treated biowaste and soil — Digestion of nitric acid soluble fractions of elements*

EN 16174, *Sludge, treated biowaste and soil — Digestion of aqua regia soluble fractions of elements*

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696)*

## 3 Principle

Mono- and divalent mercury is reduced to the elemental form by tin(II) chloride or sodium borohydride in acid medium. Elemental mercury is stripped off from the solution in a closed system. The mercury vapour, in the form of an atomic gas, passes through a cell positioned in the light path of an atomic absorption spectrometer. Its absorbance at a wavelength of 253,7 nm is measured. The absorbance signal is a function of mercury concentration, and the concentrations are calculated using a calibration curve.

NOTE The matrix of the solution analysed is dominated by the acids used in the digestion step. Tin(II) chloride as a reduction substance is recommended in this European Standard, because sodium borohydride reduces many elements commonly found in soil, sludge and waste extract solutions, to the elemental state, which may cause matrix problems under particular circumstances. However, it is still possible to use sodium borohydride as reduction agent. The concentration range 1 µg/l to 10 µg/l in the digested solution, corresponding to 0,03 µg/g to 0,3 µg/g of mercury when a 3,0 g sample has been digested, can be determined directly. Higher concentrations can be determined if the digested solution is diluted. Sensitivity can be increased by the amalgamation technique.

## 4 Interferences

Because the samples are digested with oxidizing, strong acids before determination of mercury starts, there are no interferences with organic substances in the gas phase that can absorb in the UV range. Neither are there any interferences with mercury compounds which cannot be fully reduced with tin(II) chloride alone. The interferences due to the presence of other elements in the matrix depend on their concentrations. Element concentrations exceeding those listed in Table 2 may cause negative bias.

Fewer interferences arise from heavy metals when tin(II) chloride is used rather than sodium borohydride. When flow systems are used, interference effects due to heavy metals may be less than indicated in Table 2.

**Table 2 — Tolerable concentrations of some matrix elements**

Element	Acceptable concentration in the test solution mg/l
Cu(II)	500
Ni(II)	500
Ag(I)	1

## 5 Reagents

For the determination of mercury at trace and ultra-trace level, the reagents shall be of adequate purity. The concentration of mercury or interfering substances in the reagents and the water should be negligible compared to the lowest concentration of mercury to be determined.

**5.1 Water** quality 2 according to EN ISO 3696 for all sample preparations and dilutions.

**5.2 Carrier gas**, argon or nitrogen, with a purity of  $\geq 99,99\%$ .

**5.3 Hydrochloric acid**, HCl,  $\rho(\text{HCl}) = 1,18 \text{ g/ml}$ ,  $c(\text{HCl}) = 12 \text{ mol/l}$ ,  $w(\text{HCl}) = 370 \text{ g/kg}$ .

The same batch of hydrochloric acid shall be used throughout the procedure.

**5.4 Nitric acid**,  $\text{HNO}_3$ ,  $\rho(\text{HNO}_3) = 1,4 \text{ g/ml}$ ,  $c(\text{HNO}_3) = 15 \text{ mol/l}$ ,  $w(\text{HNO}_3) = 650 \text{ g/kg}$ .

The same batch of nitric acid shall be used throughout the procedure.

**5.5 Nitric acid**, diluted solution.

Add 10 ml nitric acid (5.4) to 500 ml of water in a 1 000 ml volumetric flask, mix and fill to the mark with water.

**5.6 Nitric acid**,  $c(\text{HNO}_3) = 2 \text{ mol/l}$ , rinsing solution for glassware.

Add 150 ml of nitric acid (5.4) to about 500 ml of water, and dilute with water to 1 000 ml.

**5.7 Aqua regia**, diluted solution.

Add 21 ml hydrochloric acid (5.3) and 7 ml nitric acid (5.4) to 500 ml of water in a 1 000 ml volumetric flask, mix and fill to the mark with water.

**5.8 Tin(II) chloride solution**,  $\rho(\text{SnCl}_2 \cdot 2 \text{ H}_2\text{O}) = 100 \text{ g/l}$ .

Dissolve 10 g of  $\text{SnCl}_2 \cdot 2 \text{ H}_2\text{O}$  in 30 ml of hydrochloric acid (5.3), transfer to a 100 ml volumetric flask and fill to the mark with water. The blank concentration of mercury can be reduced by bubbling a stream of nitrogen through the solution for 30 min, if necessary. Prepare this solution on the day of use.

A solution of lower concentration, e. g. 0,5 g in 100 ml, may be used with flow systems. Prepare this latter solution freshly on the day of use from the more concentrated solution by diluting with water. Follow the recommendations of the manufacturer of the spectrometer.



### 5.9 Sodium borohydride solution, $\text{NaBH}_4$ , $\rho(\text{NaBH}_4) = 30 \text{ g/l}$ .

1 g sodium hydroxide,  $\text{NaOH}$ , is weighed into a 100 ml volumetric flask and dissolved in water. 3 g sodium borohydride,  $\text{NaBH}_4$ , are weighed into a 100 ml volumetric flask, dissolved and diluted to the mark with the sodium hydroxide solution.

A solution of lower concentration, e. g. 3 g/l may be used with flow systems. Prepare this latter solution freshly on the day of use from the more concentrated solution by diluting with water. Follow the recommendations of the manufacturer of the spectrometer.

### 5.10 Mercury standard stock solution, 1 000 mg/l.

Use a commercially available quantitative stock solution with a mercury concentration of  $(1\,000 \pm 2) \text{ mg/l}$ . This solution is considered to be stable for at least one year, but in reference to guaranteed stability, see the recommendations of the manufacturer.

### 5.11 Mercury, standard solution I, 100 mg/l.

Pipette 10 ml of the mercury standard stock solution (5.10) into a 100 ml volumetric flask, add 10 ml nitric acid (5.4), mix and fill to the mark with water.

This solution is stable for one month.

### 5.12 Mercury, standard solution II, 1 mg/l.

Pipette 1 ml of the mercury standard solution I (5.11) into a 100 ml volumetric flask, add 10 ml nitric acid (5.4), mix and fill to the mark with water.

This solution is stable for 7 days.

### 5.13 Mercury, standard solution III, 100 $\mu\text{g/l}$ .

Pipette 10 ml of the mercury standard solution II (5.12) into a 100 ml volumetric flask, add 10 ml nitric acid (5.4), mix and fill to the mark with water.

This solution shall be freshly prepared on the day of use.

## 6 Apparatus

### 6.1 Usual laboratory apparatus

All glassware shall be carefully cleaned for low trace element determinations, e. g. by immersion in nitric acid rinsing solution (5.6) for a minimum of 6 h, followed by rinsing with water before use. The nitric acid shall be replaced each week.

### 6.2 Atomic absorption spectrometer (AAS)

Instrument equipped with an appropriate monitoring system, with a cold-vapour generation system and a quartz or glass cuvette. As a radiation source use preferably an electrodeless discharge lamp (which gives a greater and more stable light intensity) operated at a current recommended by the lamp and instrument manufacturer, or a hollow cathode lamp. An AAS system with background correction device is recommended.

### 6.3 Automated sample introduction system

Automated mercury flow systems (flow injection systems or continuous flow systems) are very common for atomic absorption spectrometry. They allow a concentration range lower by about one order of magnitude. Manually operated systems or semi-automated batch systems are adequate as well.

### 6.4 Cold-vapour generator, batch system or an automated flow injection analysis system (FIA)

The system should be adaptable to the atomic absorption spectrometer (6.2). A flow-controlled nitrogen or argon stream (5.2) is used as an inert carrier to transport mercury vapour into the cell. Time-controlled addition of tin(II) chloride solution (5.8) in combination with automatic start of the read signal of the spectrometer is required. Systems with a 10 cm quartz cell with quartz windows adapted for atomic absorption spectrometry (6.2) are commercially available. In this case, the quartz cell shall be electrically heated to a temperature of 60 °C to 100 °C to avoid condensation of water.

If an automated system is used, where the reaction occurs in a continuously flowing system (e. g. FIA), the concentration of tin(II) chloride solution, reaction time and the gas-liquid separator configuration shall be optimised due to the slow reaction kinetics of the reducing system.

## 7 Procedure

### 7.1 Test solution

The test solution is an aliquot of the particle free digest or extraction solution prepared according to EN 16173 or EN 16174.

To dilute the test solution transfer 10 ml of the prepared digest solution to a 100 ml volumetric flask and fill to the mark with water.

### 7.2 Test blank solution

Prepare a test blank solution at the same time as the extraction of the test sample with *aqua regia* or nitric acid following the same procedure, using the same quantities of all reagents for the determination, but omitting the test solution.

Transfer 10 ml of the test blank solution to a 100 ml volumetric flask and fill to the mark with water.

### 7.3 Preparation of the calibration solutions

Before each batch of determinations, prepare a calibration blank solution and from the 100 µg/l mercury standard solution III (5.13) at least five calibration solutions covering the range of concentrations to be determined.

Pipette 0 ml, 1 ml, 2 ml, 5 ml, 10 ml and 20 ml of mercury standard solution III (5.13) into a series of 100 ml volumetric flasks. Fill to the mark with diluted nitric acid solution (5.5) or diluted *aqua regia* solution (5.7) and mix well. These solutions correspond to mercury concentrations of 0 µg/l, 1 µg/l, 2 µg/l, 5 µg/l, 10 µg/l and 20 µg/l, respectively.

### 7.4 Calibration

Set up the atomic absorption spectrometer (6.2) and the cold-vapour generator (6.4) according to the manufacturer's instructions, select a wavelength of 253,7 nm and adjust the quartz cell, the gas flow and the reaction time and flow rate of tin(II) chloride solution (5.8) or sodium borohydride solution (5.9) to get acceptable sensitivity and peak shape. Start the measurements with the test blank solution. The signal of the blank calibration solution should be negligible compared to the lowest calibration solution.

## 7.5 Measurement of test sample

Connect the cold-vapour generator to the atomic absorption spectrometer and start the time-controlled measurement programme. Analyse the test blank solution (7.2) and then each test solution (7.1), and record the signals.

If the mercury concentration of the test solution exceeds the range of calibration, the test sample solution shall be diluted with diluted nitric acid solution (5.5), or diluted *aqua regia* solution (5.7) accordingly.

NOTE The standard addition method can be used to check the absence of matrix effects. If the analytical results according to the standard addition method and the calibration curve method are equal, the calibration curve method can be applied.

## 8 Calculation and expression of results

### 8.1 Calculation

Calculate the mercury contents of solid samples, corrected to dry matter, according to Formula (1):

$$w_{\text{Hg}} = \frac{(\rho_1 - \rho_0) \cdot f \cdot V \cdot C}{m} \quad (1)$$

with

$$C = 100/w_{\text{dm}} \quad (2)$$

where

- $w_{\text{Hg}}$  is the mass fraction of mercury in the solid sample, expressed in milligrams per kilogram (mg/kg) corrected to dry matter at 105 °C;
- $\rho_1$  is the concentration of the element in the test solution (7.1), expressed in milligrams per liter (mg/l);
- $\rho_0$  is the concentration of the element in the test blank solution (7.2), expressed in milligrams per liter (mg/l);
- $f$  is the dilution factor of the test solution (7.1);
- $V$  is the volume of the digest solution prepared according to EN 16173 or EN 16174, expressed in liters (l);
- $m$  is the mass of the digested test portion, expressed in kilograms (kg);
- $C$  is the correction factor for dry mass;
- $w_{\text{dm}}$  is the dry matter fraction of the sample, determined according to EN 15934, in per cent (%).

### 8.2 Expression of results

Report as many significant figures as are acceptable according to the precision of the measuring values but not more than three significant figures.

## 9 Performance data

The performance characteristics of the method data have been evaluated (see Annex A).

## **10 Test report**

The test report shall contain at least the following information:

- a) a reference to this European Standard (EN 16175-1);
- b) all information necessary for identification of the sample;
- c) information about the pretreatment and method of digestion of the sample;
- d) results of the determination as indicated in 8.2;
- e) any details not specified in this European Standard or which are optional, as well as any factor which may have affected the results.

## Annex A (informative)

### Repeatability and reproducibility data

#### A.1 Materials used in the interlaboratory comparison study

The interlaboratory comparison on determination of mercury in *aqua regia* and nitric acid digests of sludge, treated biowaste and soil by cold-vapour atomic absorption spectrometry was carried out by 7 to 8 European laboratories on 5 materials. The materials selected for the interlaboratory comparison were chosen to represent soil, sludge and biowaste as broadly as possible, because the method will find general application across different types of soil and soil related materials.

Table A.1 lists the types of materials tested.

**Table A.1 — Materials tested in the interlaboratory comparison on the determination of mercury in *aqua regia* and nitric acid digests of sludge, treated biowaste and soil using CV-AAS**

Grain size	Sample	Material
Sludge (< 0,5 mm)	Sludge 1	Mix 1 of municipal waste water treatment plant sludges from North Rhine Westphalia, Germany
	Sludge 2	Mix 2 of municipal waste water treatment plant sludges from North Rhine Westphalia, Germany
Fine grained (< 2,0 mm)	Compost 2	Compost from Germany
	Soil 1	A sludge amended soil from Pavia, Italy
	Soil 2	A sludge amended soil from Düsseldorf, Germany

#### A.2 Interlaboratory comparison results

The statistical evaluation was conducted according to ISO 5725-2. The average values, the repeatability standard deviation ( $s_r$ ) and the reproducibility standard deviation ( $s_R$ ) were obtained (Table A.2).

**Table A.2 — Results of the interlaboratory comparison study on the determination of mercury in *aqua regia* and nitric acid digests – Cold-vapour atomic absorption spectrometry**

Matrix	$l$	$n$	$n_0$	$\bar{\bar{x}}$	$s_R$	$C_{V,R}$	$s_r$	$C_{V,r}$
Sludge 1	7	27	2	1,120	0,253	22,6	0,052	4,61
Sludge 2	8	41	0	0,660	0,156	23,6	0,053	8,05
Compost 2	7	32	1	0,118	0,014	12,0	0,005	3,88
Soil 1	7	38	0	0,070	0,021	29,7	0,008	11,8
Soil 2	7	38	0	0,052	0,008	14,7	0,005	8,71
Explanation of symbols								
$l$	number of participating laboratories							
$n$	number of analytical results							
$n_0$	number of rejected laboratories							
$\bar{\bar{x}}$	total mean of results (without outliers) in milligram per kilogram (mg/kg)							
$s_R$	reproducibility standard deviation in milligram per kilogram (mg/kg)							
$C_{V,R}$	coefficient of variation of reproducibility in percent (%)							
$s_r$	repeatability standard deviation in milligram per kilogram (mg/kg)							
$C_{V,r}$	coefficient of variation of repeatability in percent (%)							

## Bibliography

- [1] Umweltbundesamt, Texte 42/2015: Validierung von Analyseverfahren, die im Rahmen des Projektes HORIZONTAL entwickelt, jedoch nicht validiert wurden.  
Download: <http://www.umweltbundesamt.de/publikationen>
- [2] ISO 5725-2, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*







# British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

## About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

## Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at [bsigroup.com/standards](http://bsigroup.com/standards) or contacting our Customer Services team or Knowledge Centre.

## Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at [bsigroup.com/shop](http://bsigroup.com/shop), where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

## Copyright in BSI publications

All the content in BSI publications, including British Standards, is the property of and copyrighted by BSI or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use.

Save for the provisions below, you may not transfer, share or disseminate any portion of the standard to any other person. You may not adapt, distribute, commercially exploit, or publicly display the standard or any portion thereof in any manner whatsoever without BSI's prior written consent.

## Storing and using standards

Standards purchased in soft copy format:

- A British Standard purchased in soft copy format is licensed to a sole named user for personal or internal company use only.
- The standard may be stored on more than 1 device provided that it is accessible by the sole named user only and that only 1 copy is accessed at any one time.
- A single paper copy may be printed for personal or internal company use only.

Standards purchased in hard copy format:

- A British Standard purchased in hard copy format is for personal or internal company use only.
- It may not be further reproduced – in any format – to create an additional copy. This includes scanning of the document.

If you need more than 1 copy of the document, or if you wish to share the document on an internal network, you can save money by choosing a subscription product (see 'Subscriptions').

## Reproducing extracts

For permission to reproduce content from BSI publications contact the BSI Copyright & Licensing team.

## Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to [bsigroup.com/subscriptions](http://bsigroup.com/subscriptions).

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

**PLUS** is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit [bsigroup.com/shop](http://bsigroup.com/shop).

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email [subscriptions@bsigroup.com](mailto:subscriptions@bsigroup.com).

## Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

## Useful Contacts

### Customer Services

**Tel:** +44 345 086 9001

**Email (orders):** [orders@bsigroup.com](mailto:orders@bsigroup.com)

**Email (enquiries):** [cservices@bsigroup.com](mailto:cservices@bsigroup.com)

### Subscriptions

**Tel:** +44 345 086 9001

**Email:** [subscriptions@bsigroup.com](mailto:subscriptions@bsigroup.com)

### Knowledge Centre

**Tel:** +44 20 8996 7004

**Email:** [knowledgecentre@bsigroup.com](mailto:knowledgecentre@bsigroup.com)

### Copyright & Licensing

**Tel:** +44 20 8996 7070

**Email:** [copyright@bsigroup.com](mailto:copyright@bsigroup.com)

### BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK