

BS EN 16620:2015



BSI Standards Publication

Food analysis — Determination of furan in coffee and coffee products by headspace gas chromatography and mass spectrometry (HS GC-MS)

bsi.

...making excellence a habit.™

National foreword

This British Standard is the UK implementation of EN 16620:2015.

The UK participation in its preparation was entrusted to Technical Committee AW/275, Food analysis - Horizontal methods.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2015. Published by BSI Standards Limited 2015

ISBN 978 0 580 80860 9

ICS 67.140.20

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 April 2015.

Amendments issued since publication

Date	Text affected
------	---------------

EUROPEAN STANDARD

EN 16620

NORME EUROPÉENNE

EUROPÄISCHE NORM

April 2015

ICS 67.140.20

English Version

Food analysis - Determination of furan in coffee and coffee products by headspace gas chromatography and mass spectrometry (HS GC-MS)

Analyse des produits alimentaires - Dosage du furane dans le café et les produits à base de café par chromatographie en phase gazeuse avec espace de tête couplée à la spectrométrie de masse (ET-CPG-SM)

Lebensmittelanalytik - Bestimmung von Furan in Kaffee und Kaffee-Erzeugnissen mit Headspace-Gaschromatographie und Massenspektrometrie (HS GC-MS)

This European Standard was approved by CEN on 7 February 2015.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents		Page
Foreword		3
1	Scope	4
2	Normative references	4
3	Principle	4
4	Reagents	4
5	Apparatus	5
6	Procedure	6
7	Calculation	9
8	Precision	10
9	Test report	11
Annex A (informative) Typical chromatograms		12
Annex B (informative) Precision data		14
Bibliography		15

Foreword

This document (EN 16620:2015) has been prepared by Technical Committee CEN/TC 275 “Food analysis - Horizontal methods”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 2015 and conflicting national standards shall be withdrawn at the latest by October 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association.

WARNING — The use of this document can involve hazardous materials, operations and equipment. This document does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this document to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

1 Scope

This European Standard specifies a method for the determination of furan in coffee and coffee products with headspace-gas chromatography-mass spectrometry (HS-GC-MS), see [1] and [2]. Coffee products in the scope of this method are extracts which have been spray-dried, agglomerated or freeze-dried. The method has been validated in an interlaboratory study via the analysis of naturally contaminated samples of spray-dried coffee, freeze-dried coffee and ground roasted coffee ranging from 264 µg/kg to 2 840 µg/kg.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3696:1995, *Water for analytical laboratory use - Specification and test methods (ISO 3696:1987)*

3 Principle

Coffee and coffee products are weighed into headspace vials, diluted with water and incubated at 50 °C for 30 min. The furan content in the headspace is measured with HS-GC-MS. Quantification is done by using a standard addition curve, taking into account deuterated furan (*d4*-furan) as internal standard.

4 Reagents

Use only reagents of recognized analytical grade and water complying with grade 1 of EN ISO 3696:1995, unless otherwise specified. Solvents shall be of quality for HPLC analysis.

4.1 Furan, (CAS 110-00-9), mass concentration $\rho = 0,936$ g/ml.

4.2 *d4*-Furan, (CAS 6142-90-1), $\rho = 0,991$ g/ml.

4.3 Methanol.

4.4 Potassium hydroxide, aqueous solution, $\rho = 250$ g/l.

4.5 Stock solutions

4.5.1 Furan stock solution, ρ approximately 2,50 mg/ml.

Place 20 ml of methanol in a headspace vial and seal the vial. Weigh the sealed vial. Using a chilled syringe, transfer 50 µl of furan through the septum of the vial into the methanol. Reweigh the sealed vial. Mix the contents well. The solution is stable for 2 weeks at 4 °C. Pierced headspace vials shall be re-capped.

4.5.2 *d4*-Furan stock solution, ρ approximately 2,50 mg/ml.

Place 20 ml of methanol in a headspace vial and seal the vial. Weigh the sealed vial. Using a chilled syringe, transfer 50 µl of *d4*-furan through the septum of the vial into the methanol. Reweigh the sealed vial. Mix the contents well. The solution is stable for 2 weeks at 4 °C. Pierced headspace vials shall be re-capped.

4.6 Standard solutions

4.6.1 General

The amount of added solutions is calculated by differential weighing considering the densities. The solutions and additional dilutions should be kept cooled when working. Adapt the concentration of standard and internal standard solutions to the expected values in the samples. If necessary, prepare additional dilutions according to the working range. All furan and *d4*-furan solutions should be prepared in different rooms. Use separate syringes without plastic material.

4.6.2 Furan standard solution 1, ρ approximately 25 $\mu\text{g/ml}$.

Place 10 ml of water in a headspace vial and seal the vial. Weigh the sealed vial. Using a chilled syringe, transfer 100 μl of furan stock solution (4.5.1) through the septum of the vial into the water. Reweigh the sealed vial. Mix the contents well. Prepare the solution daily.

4.6.3 *d4*-Furan standard solution 1, ρ approximately 25 $\mu\text{g/ml}$.

Place 10 ml water in a headspace vial and seal the vial. Weigh the sealed vial. Using a chilled syringe, transfer 100 μl of *d4*-furan stock solution (4.5.2) through the septum of the vial into the water. Reweigh the sealed vial. Mix the contents well. Prepare the solution daily.

4.6.4 Furan standard solution 2, ρ approximately 250 ng/ml .

Place 10 ml water in a headspace vial and seal the vial. Weigh the sealed vial. Using a chilled syringe, transfer 100 μl of furan standard solution 1 (4.6.2) through the septum of the vial into the water. Reweigh the sealed vial. Mix the contents well. Prepare the solution daily.

4.6.5 *d4*-Furan standard solution 2, ρ approximately 250 ng/ml .

Place 10 ml water in a headspace vial and seal the vial. Weigh the sealed vial. Using a chilled syringe, transfer 100 μl of *d4*-furan standard solution 1 (4.6.3) through the septum of the vial into the water. Reweigh the sealed vial. Mix the contents well. Prepare the solution daily.

5 Apparatus

Usual laboratory glassware and equipment and, in particular, the following:

5.1 Positive displacement pipette.

5.2 Microlitre syringes.

5.3 Hand crimper and de-crimper.

5.4 Homogenizer or mill.

5.5 Test tube shaker.

5.6 Gas chromatograph for capillary gas chromatography, with split/splitless injector and headspace autosampler.

5.6.1 Fused silica capillary column, suitable for analysing volatile organic compounds e.g. non-polar polysiloxane with 5 % phenyl. The following column has been shown to be suitable: J&W DB-5 from Agilent¹⁾ length of 60 m, internal diameter (i.d.) of 0,32 mm and film thickness (d_f) of 1 μm .

5.6.2 PLOT column for separation of very volatile compounds, e.g. HP® Plot Q from Agilent¹⁾, length of 15 m, i.d. of 0,32 mm and d_f of 20 μm coupled to a restriction column HP-5MS from Agilent¹⁾, length of 30 m, i.d. of 0,25 mm and d_f of 0,25 μm . This column is an alternative for the column given under 5.6.1.

5.6.3 Rt®Q-Bond.²⁾

5.6.4 Carrier gas, helium.

5.6.5 Mass spectrometer with capability of single ion monitoring, coupled to gas chromatograph.

5.7 Headspace vials, 20 ml, with aluminium crimp seals and PTFE-faced silicone septa.

6 Procedure

6.1 Sample preparation

The samples shall be stored refrigerated in unopened original packages. Immediately prior to analysis the original packages should be opened and weighed with minimal delay to avoid losses of furan. Powdered samples should be weighed without further treatment. Coffee beans should be milled under conditions that do not raise the temperature and weighed. Although not tested in the interlaboratory study, the use of cooled milling equipment is recommended for milling in order to avoid losses of furan. Milling tests minimizing losses of furan should be performed ahead of the sample preparation.

6.2 Estimation of furan content

The furan content in a sample is quantified by means of standard addition.

The amount of furan and/or *d4*-furan which is added to the test portions is based on a rough estimation (x_0) in relation to the initial sample weight.

This rough estimation (x_0) is done as follows: to 0,5 g to 5 g of the sample (normally 1 g) water is added (normally 5 ml), an amount of *d4*-furan (internal standard) in ng is added (approximately corresponding to the expected amount of furan in the sample) and is analysed with HS-GC-MS as described in 6.4.

Using the peak area ratio furan/*d4*-furan ($m/z = 68/72$, see 7.2) and the amount of added internal standard in ng the furan content in relation to the initial weight of the sample solution can be estimated (equivalent to a simple isotope dilution analysis).

In order to avoid the forming of furan from precursors, the pH value is adjusted to ca. pH = 10. For this 5 drops to 6 drops of potassium hydroxide solution (4.4) are added into the headspace vial.

1) These are examples of suitable products available commercially. This information is given for the convenience of users of this European Standard and does not constitute an endorsement by CEN of these products.

2) Rt®-Q-Bond is an example of a suitable product available commercially. This information is given for the convenience of users of this European Standard and does not constitute an endorsement by CEN of this product.

6.3 Standard addition analysis

In 9 headspace vials each 0,5 g to 5 g of the sample (normally 1 g) are weighed in, water is added (normally 5 ml) and *d4*-furan standard solution in relation to the order of magnitude of the estimated amount of furan (see 6.2) into the solution. Three headspace vials are capped immediately. The others are fortified with 0,5, 1,0 and 2,0 times of estimated furan content. The fortifying volume should not exceed 10 % of the total volume. For each new sample the concentration of the furan- and *d4*-furan standard solution (see 4.6) shall be adapted.

For an estimated furan content in the sample of 2,5 mg/kg (2,5 µg/g) and an initial weight of 1 g, the furan standard solution 1 and *d4*-furan standard solution 1 (see 4.6) can be used. Table 1 gives an example of a standard addition for an estimated furan content of 2,5 µg/g.

Table 1 — Example of a standard addition for an estimated furan content x_0 of 2,5 µg/g

Sample solution	<i>d4</i> -Furan standard solution µl	<i>d4</i> -Furan standard solution µg	Furan standard solution µl	Furan standard solution µg
P ₁₁ (0 x_0)	100	2,5	-	-
P ₁₂ (0 x_0)	100	2,5	-	-
P ₁₃ (0 x_0)	100	2,5	-	-
P ₂₁ (0,5 x_0)	100	2,5	50	1,25
P ₂₂ (0,5 x_0)	100	2,5	50	1,25
P ₃₁ (1,0 x_0)	100	2,5	100	2,5
P ₃₂ (1,0 x_0)	100	2,5	100	2,5
P ₄₁ (2,0 x_0)	100	2,5	200	5,0
P ₄₂ (2,0 x_0)	100	2,5	200	5,0

With the above fortifying procedure the *d4*-furan is added to the sample solutions according to the estimated furan content of the sample (here 2,5 µg/kg).

If necessary more than 2 or 3 headspace vials can be prepared for each fortifying level.

6.4 Determination with headspace-GC-MS

The following parameters should be kept:

- Incubation temperature: 50 °C
- Incubation time: 30 min

Headspace parameters:

The following headspace parameters are examples and shall be adopted to the used device:

- Needle temperature: 52 °C to 55 °C
- Shaker speed: 300 min⁻¹

- Filling speed: 200 µl/s
- Pull up delay: 500 ms
- Injection speed: 500 µl/s
- Preinject delay: 500 ms
- Postinject delay: 500 ms
- Scavenging time: 3 min

GC-MS parameters:

In combination with the recommended column (5.6.1) the following parameter settings have been proven useful:

- GC-column gas flow: 1,7 ml/min Helium (constant flow)
- Gas saver: off
- Injector: split 2:1
- Injection temperature: 200 °C
- Injection volume: 1 000 µl
- Temperature program: 50 °C, 10 °C/min to 225 °C, 225 °C for 12,5 min, run time 30 min
- Coupling to MS: direct

In combination with the recommended column (5.6.2) the following parameters have been proven themselves:

- GC-column gas flow: 1,0 ml/min Helium (constant flow)
- Gas saver: off
- Injector: split or splitless
- Injection temperature: 200 °C
- Injection volume: 1 000 µl
- Temperature program: 50 °C, 10 °C/min to 225 °C, 225 °C for 6,5 min, run time 24 min
- Coupling to MS: direct

MS parameters:

- Ionization energy: 70 eV
- SIM (single ion monitoring)
- Mode:

- Furan: $m/z = 68$ (100 ms) and $m/z = 39$ (20 ms)
- *d4*-Furan: $m/z = 72$ (100 ms) and $m/z = 42$ (20 ms)
- Alternatively: Scan $m/z 35>75$
- MS source temperature: 230 °C
- MS quad temperature: 150 °C
- MS transfer line: 225 °C

Table 2 gives information on ions for identification and quantification of furan for confirmation purposes.

Table 2 — Ions for identification and quantification of furan (for more information see [3])

Name	Qualifier Ion Q_1 m/z	Qualifier Ion Q_2 m/z	Peak area ratio Q_1/Q_2	Peak area ratio - lower limit	Peak area ratio - upper limit
Furan	68	39	1,56	1,40	1,71

7 Calculation

7.1 General

Quantify the furan content with standard addition and the method of internal standard via integrated peak areas of $m/z = 68$ for furan and $m/z = 72$ for *d4*-Furan. Use at least 7 points of the standard addition (see 6.3) for the calculation.

7.2 For a multipoint measurement of the sample with internal standard correction, determine the area ratio of $m/z = 68$ for furan and $m/z = 72$ for the internal standard *d4*-furan using Formula (1):

$$Q = \frac{A_A}{A_S} \quad (1)$$

where

A_A is the peak area $m/z = 68$ furan;

A_S is the peak area $m/z = 72$ *d4*-furan.

7.3 For the standard calibration graph, plot the resulting peak area ratios (Q , see 7.2) against the fortified furan amount in ng according to the standard addition series.

Calibration graph $Q = bx + a$ or rather

$$x = \frac{(Q - a)}{b} \quad (2)$$

where

x is the mass of furan in ng in the sample;

a is the intercept of the calibration graph;

b is the slope of the calibration graph.

NOTE Calculation is done automatically by any modern chromatography software (mode: calculation with internal standard).

7.4 Calculate the furan content of the sample, in $\mu\text{g}/\text{kg}$, by determination of the point of intersection with the prolonged calibration graph and the axis of abscissae.

Determine the furan content in the sample in $\mu\text{g}/\text{kg}$ mathematically using Formula (3):

$$x = \frac{(Q - a)}{b} \quad (3)$$

with

$Q = 0$ becomes $x = |-a/b|$

$\rho = x/W$

where

ρ is the amount of furan in the sample ($\text{ng}/\text{g} = \mu\text{g}/\text{kg}$);

x is the mass of furan in ng in the sample (see 7.3);

W is the initial sample weight in g (see 6.1).

Result is given with 3 significant digits.

8 Precision

8.1 General

Details of the interlaboratory test of the precision of the method are summarized in Annex B. The values derived from the interlaboratory test may not be applicable to analyte concentration ranges and/or matrices other than those given in Annex B.

8.2 Repeatability

The absolute difference between two single test results found on identical test material by one operator using the same apparatus within the shortest feasible time interval will exceed the repeatability limit r in not more than 5 % of the cases.

The values for spray-dried coffee are:

$\bar{x} = 264 \mu\text{g}/\text{kg}$ $r = 53,5 \mu\text{g}/\text{kg}$ (naturally contaminated)

The values for freeze-dried coffee are:

$\bar{x} = 1\,140 \mu\text{g}/\text{kg}$ $r = 238 \mu\text{g}/\text{kg}$ (naturally contaminated)

The values for roasted coffee are:

$\bar{x} = 2\,840 \mu\text{g}/\text{kg}$ $r = 678 \mu\text{g}/\text{kg}$ (naturally contaminated)

8.3 Reproducibility

The absolute difference between two single test results found on identical test material reported by two laboratories will exceed the reproducibility limit R in not more than 5 % of the cases.

The values for spray-dried coffee are:

$$\bar{x} = 264 \mu\text{g/kg} \quad R = 142 \mu\text{g/kg} \quad (\text{naturally contaminated})$$

The values for freeze-dried coffee are:

$$\bar{x} = 1\,140 \mu\text{g/kg} \quad R = 517 \mu\text{g/kg} \quad (\text{naturally contaminated})$$

The values for roasted coffee are:

$$\bar{x} = 2\,840 \mu\text{g/kg} \quad R = 1\,530 \mu\text{g/kg} \quad (\text{naturally contaminated})$$


9 Test report

The test report shall contain the following data:


- a) all information necessary for the identification of the sample (type of sample, origin and designation of the sample);
- b) a reference to this European Standard;
- c) the date and type of sampling procedure (if known);
- d) the date of receipt;
- e) the date of test;
- f) the test results and the units in which they have been expressed;
- g) any operations not specified in the method or regarded as optional, which might have affected the results.

Annex A
(informative)

Typical chromatograms



a) Furan



b) *d*4-Furan

Key

t	time
Y	Abundance
1	m/z = 68
2	m/z = 39
3	m/z = 72
4	m/z = 42
Operating conditions	capillary column Rt [®] -Q-Bond ³⁾ (30 m, 0,32 mm, 10 µm film thickness)
Temperature programme	50 °C (0,1 min), 10 °C/min to 190 °C (0 min), 40 °C/min to 225 °C (20 min).
Flow	2,0 ml/min

Figure A.1 — Typical chromatogram of furan and d4-furan

3) Rt[®]-Q-Bond is an example of a suitable product available commercially. This information is given for the convenience of users of this European Standard and does not constitute an endorsement by CEN of this product.

Annex B (informative)

Precision data

The following data were obtained in an interlaboratory test according to ISO 5725-2 [4] for collaborative study procedures to validate characteristics of a method of analysis.

This method was worked out by the § 64 working group 'Furan-Analytik' of the 'Bundesamt für Verbraucherschutz und Lebensmittelsicherheit' and tested in an interlaboratory study with 15 participants.

Table B.1 — Precision data for furan in coffee with headspace GC-MS

Sample	Spray dried coffee	Freeze-dried coffee	Roasted coffee
Preparation of test material	naturally contaminated	naturally contaminated	naturally contaminated
Year of interlaboratory test	2008	2008	2008
Number of laboratories	15	15	15
Number of laboratories retained after eliminating outliers	13	14	15
Number of outliers (laboratories)	2	1	0
Number of accepted results	13	14	15
Mean value, \bar{x} , µg/kg	264	1 140	2 840
Repeatability standard deviation s_r , µg/kg	19,1	85,1	242,1
Repeatability relative standard deviation, RSD_r , %	7,2	7,5	8,5
Repeatability limit r [$r = 2,8 \times s_r$], µg/kg	53,5	238	678
Reproducibility standard deviation s_R , µg/kg	50,8	185	545
Reproducibility relative standard deviation, RSD_R , %	19,3	16,2	19,2
Reproducibility limit R [$R = 2,8 \times s_R$], µg/kg	142	517	1 530
HorRat value, according to [5] and [6]	0,99	1,04	1,4

Bibliography

- [1] FDA-METHOD. "Determination of furan in foods",
<http://www.fda.gov/Food/FoodborneIllnessContaminants/ChemicalContaminants/ucm078400.htm>
- [2] Method Ruggedness Testing, Determination of Furan in Foods,
<http://www.fda.gov/Food/FoodborneIllnessContaminants/ChemicalContaminants/ucm078408.htm>
- [3] Commission Regulation (EC) 2002/657 of 12 August 2002 implementing Council Directive 96/23/EC concerning the performance of analytical methods and interpretation of results
- [4] ISO 5725-2:1994, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*
- [5] HORWITZ W. A Protocol for the Design, Conduct and Interpretation of Collaborative Studies. *Pure Appl. Chem.* 1988, **60** pp. 855–864
- [6] HORWITZ W., KAMPS L.R., BOYER K.W. Quality Assurance in the analysis of foods and trace constituents. *J. Assoc. Off. Anal. Chem.* 1980, **63** pp. 1344–1354

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email bsmusales@bsigroup.com.

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

Useful Contacts:

Customer Services

Tel: +44 845 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 845 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com



...making excellence a habit.™