

BS EN 16801:2016



BSI Standards Publication

**Foodstuffs — Determination  
of elements and their chemical  
species — Determination of  
methylmercury in foodstuffs  
of marine origin by isotope  
dilution GC-ICP-MS**

**bsi.**

...making excellence a habit.™

**National foreword**

This British Standard is the UK implementation of EN 16801:2016.

The UK participation in its preparation was entrusted to Technical Committee AW/275, Food analysis - Horizontal methods.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2016.  
Published by BSI Standards Limited 2016

ISBN 978 0 580 79937 2

ICS 67.120.30

**Compliance with a British Standard cannot confer immunity from legal obligations.**

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 30 April 2016.

**Amendments/corrigenda issued since publication**

Date	Text affected
------	---------------

---

EUROPEAN STANDARD

**EN 16801**

NORME EUROPÉENNE

EUROPÄISCHE NORM

March 2016

ICS 67.120.30

English Version

## Foodstuffs - Determination of elements and their chemical species - Determination of methylmercury in foodstuffs of marine origin by isotope dilution GC-ICP-MS

Produits alimentaires - Détermination des éléments et de leurs espèces chimiques - Détermination de la teneur en méthylmercure dans les produits alimentaires d'origine marine par dilution isotopique GC-ICP-SM

Lebensmittel - Bestimmung von Elementen und ihren Verbindungen - Bestimmung von Methylquecksilber in Lebensmitteln marinen Ursprungs mit Isotopenverdünnung GC-ICP-MS

This European Standard was approved by CEN on 8 February 2016.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

<b>Contents</b>		Page
<b>European foreword</b> .....		<b>3</b>
<b>1</b>	<b>Scope</b> .....	<b>4</b>
<b>2</b>	<b>Normative references</b> .....	<b>4</b>
<b>3</b>	<b>Principle</b> .....	<b>4</b>
<b>4</b>	<b>Reagents</b> .....	<b>4</b>
<b>5</b>	<b>Apparatus and equipment</b> .....	<b>6</b>
<b>6</b>	<b>Procedure</b> .....	<b>7</b>
<b>7</b>	<b>Calculation</b> .....	<b>10</b>
<b>8</b>	<b>Precision</b> .....	<b>11</b>
<b>9</b>	<b>Test report</b> .....	<b>12</b>
<b>Annex A (informative) Precision data</b> .....		<b>13</b>
<b>Bibliography</b> .....		<b>14</b>

## European foreword

This document (EN 16801:2016) has been prepared by Technical Committee CEN/TC 275 “Food analysis - Horizontal methods”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by September 2016, and conflicting national standards shall be withdrawn at the latest by September 2016.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under mandate M 422 given to CEN by the European Commission and the European Free Trade Association.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

## 1 Scope

This European Standard describes a method for the determination of monomethylmercury (MMHg) in foodstuffs of marine origin. The method has been validated in an interlaboratory test on mussel tissue, squid muscle, crab claw muscle, dog fish liver, whale meat, cod muscle and Greenland halibut muscle (all freeze-dried) with mass fractions from 0,04 mg/kg to 3,6 mg/kg dry weight according to ISO 5725-2 [1].

Laboratory experiences have shown that this method is also applicable on fresh samples [2].

## 2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 13804, *Foodstuffs — Determination of elements and their chemical species — General considerations and specific requirements*

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696)*

## 3 Principle

The sample is spiked with an appropriate amount of Hg-isotope enriched MMHg and digested using tetramethylammonium hydroxide (TMAH). After pH adjustment, derivatisation and extraction, the organic phase is analysed using GC-ICP-MS. The GC separates the different mercury species before the derivatised species (ethylmethylmercury) is atomised and ionised in the high temperature by the ICP. The ions are extracted from the plasma by a set of sampler and skimmer cones and transferred to a mass spectrometer where the ions are separated by their mass/charge ratio and determined by a pulse-count and/or analogue detector. The result is calculated using the isotope dilution equation.

**WARNING — The use of this method may involve hazardous materials, operations and equipment. This method does not purport to address all the safety problems associated with its use. It is the responsibility of the user of this method to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.**

## 4 Reagents

### 4.1 General

The concentration of mercury species in the reagents and water used shall be low enough to not affect the results of the determination. When using a method of high sensitivity like ICP-MS, the control of the blank levels of water, acid and other reagents is very important. Generally ultra-pure water complying with ISO 3696 grade 1 (i.e. electrical conductivity below 0,1  $\mu\text{S}/\text{cm}$  at 25 °C) and acid of high purity is recommended, e.g. cleaned by sub-boiling distillation. Reagents should be of minimum p.a. quality where possible. Special facilities can be used in order to avoid contamination during the steps of preparation and measurement (e.g. uses of laminar flow benches or comparable clean room facilities).

## 4.2 Monomethylmercury stock solutions.

Commercially available MMHg standard enriched in the  $^{201}\text{Hg}$ -isotope with a mass fraction of  $5,5 \mu\text{g/g}$  (as Hg) is recommended, such as IES-MMHg201<sup>1)</sup>. Other MMHg Hg-isotope enriched standards may also be available in suitable mass fractions from other suppliers or may be prepared in-house. In this case, the method shall be adjusted accordingly. The quality of the standards should be designed to be used by isotope dilution methods. Stock solutions in diluted acid are preferred.

## 4.3 Monomethylmercury standard solution.

### 4.3.1 General

The mass fractions of the MMHg in the standard solutions shall be chosen in relation to the expected mass fraction of MMHg in samples. It is important that all dilutions are done by weighing so that their accurate mass fractions can be calculated. The following descriptions are given as examples.

### 4.3.2 MMHg approximately 500 ng/g (as Hg).

Dilute approximately 1 g, to the nearest milligram, of  $^{201}\text{Hg}$  enriched MMHg stock solution (4.2) with water up to 10 g. Calculate the exact mass fraction using the mass fraction of the stock solution and weight.

### 4.3.3 MMHg approximately 50 ng/g (as Hg).

Dilute approximately 1 g, to the nearest milligram, of the  $500 \text{ ng/g } ^{201}\text{Hg}$  enriched MMHg solution (4.3.2) with water up to 10 g. Calculate the exact mass fraction using the exact mass fraction of the  $500 \text{ ng/g}$  solution and weights.

### 4.3.4 MMHg approximately 5 ng/g (as Hg).

Dilute approximately 1 g, to the nearest milligram, of the  $50 \text{ ng/g } ^{201}\text{Hg}$  enriched MMHg solution (4.3.3) with water up to 10 g. Calculate the exact mass fraction using the exact mass fraction of the  $50 \text{ ng/g}$  solution and weights.

**4.4 Tetramethylammonium hydroxide (TMAH)**, mass fraction  $w = 25 \%$  in water, minimum synthesis quality.

**4.5 Acetic acid**, concentrated, mass concentration  $\rho = 1,05 \text{ g/ml}$ , minimum p.a. quality.

**4.6 Sodium hydroxide**, minimum p.a. quality.

**4.7 Sodium hydroxide solution**, substance concentration  $c(\text{NaOH}) = 0,1 \text{ mol/l}$ .

Transfer  $0,4 \text{ g}$  of sodium hydroxide to a  $100 \text{ ml}$  volumetric flask and add water to the mark.

**4.8 Sodium acetate**, minimum p.a. quality.

**4.9 Sodium acetate/acetic acid buffer (pH 5).**

Dissolve  $41 \text{ g}$  of sodium acetate in approximately  $0,5 \text{ l}$  of water. Adjust the pH of the solution to 5 by adding concentrated acetic acid (4.5) dropwise by using a pH-meter (5.4). Finally, dilute the solution to  $1 \text{ l}$  with water.

---

<sup>1)</sup> IES-MMHg201 is available from e.g. Innovative Solutions in Chemistry S.L., Edificio Científico-Tecnológico, Campus de "El Cristo", 33006, Oviedo, Spain <http://www.isc-science.com/> or Qmx Laboratories, bolford Street, Thaxted, Essex, CM6 2PY, UK <http://www.qmx.com/>. This is an example of a suitable product available commercially. This information is given for the convenience of the users of this International Standard and does not constitute an endorsement by CEN of this product.

**4.10 Nitric acid**, 65 %,  $\rho$  of approximately 1,4 g/ml. Hg-free quality. Other acid concentrations may be used if the volume added in 6.2 is adjusted accordingly.

**4.11 Sodium tetraethyl borate**, minimum synthesis quality (98 %).

**4.12 Sodium tetraethyl borate solution (2 %)**.

Dissolve 1 g of sodium tetraethyl borate in 0,1 mol/l of sodium hydroxide solution (4.7), transfer to a 50 ml-volumetric flask and fill up to the mark with 0,1 mol/l sodium hydroxide solution. Prepare freshly at each day of analysis or divide the solution into smaller amounts and store in the freezer at approximately  $-20\text{ }^{\circ}\text{C}$ . The solution may be stored at approximately  $-20\text{ }^{\circ}\text{C}$  for at least three months. The solution shall be used within the day after removal from the freezer.

**4.13 Hexane**, minimum HPLC-quality.

**4.14 Optimising solution for the ICP-MS**.

The optimising solution should contain elements that cover the whole mass range giving a high rate of oxides and doubly charged ions. Use the solutions recommended by the manufacturer of the ICP-MS instrument. A solution containing e.g. Li, Ce and Tl is suitable for those purposes. In this case, choose the concentration of these elements in order to achieve a count rate of  $> 10\ 000$  cps (counts per second).

## 5 Apparatus and equipment

All pieces of equipment described here are examples of suitable equipment and may be replaced by equivalent equipment unless otherwise stated. Generally, clean and rinse the vessels carefully according to the procedure in EN 13804. In addition to standard laboratory equipment, use the following:

**5.1 Analytical balance**, accuracy of 0,5 mg.

**5.2 Orbital/overhead rotator**, capable of approximately 0,04 g ( $20\text{ min}^{-1}$ ).

**5.3 Centrifuge**, capable of 1 200 g ( $4\ 000\text{ min}^{-1}$ ).

**5.4 pH-meter**.

**5.5 Laboratory ware**, volumetric flasks of glass, polypropylene tubes (10 ml) for samples, GC-vials, pH paper.

**5.6 Inductively Coupled Plasma Mass Spectrometer (ICP-MS)**.

Mass spectrometer with inductively coupled argon plasma operating in a mass range from 5 amu (atomic mass units) to 240 amu. Using routine settings the mass spectrometer shall be capable to resolve 1 amu peak width at 5 % peak height or better (resolution 300) with sufficient sensitivity to achieve the detection limits suitable for the analytical purpose.

**5.7 Argon**, purity  $\geq 99,99\%$ .

**5.8 Gas chromatograph (GC)**, with injector heating, programmable column heating and heating of transfer line to ICP-MS.

**5.9 GC-column**, capillary or preparative column capable of separating ethylmethylmercury from other mercury species (e.g. 30 m x 0,32 mm, analytical column with 5 % phenyl methyl siloxane; film thickness: 0,25  $\mu\text{m}$ ).



**5.10 Helium**, purity  $\geq 99,99$  %.

**5.11 Helium**, (5.10) with 1 % to 2 % added xenon for tuning of the GC-ICP-MS interface or some other tuning configuration capable of optimising the instrument parameters, optional.

**5.12 Oxygen**, optional, to prevent carbon deposition, according to manufacturer's instructions, e.g. 5 %.

## 6 Procedure

### 6.1 Calculation of optimal spike amount

The following description is given for the  $^{201}\text{Hg}$  enriched MMHg spike solution (4.2). To ensure that the measurement is within acceptable error limits, the abundance of the isotopes  $^{200}\text{Hg}$ ,  $^{201}\text{Hg}$  and  $^{202}\text{Hg}$  in the spiked sample should be as close as possible [3]. This is achieved when the amount of spike to analyte is 1 to 7 for the given isotope abundances in Table 1.

The ratio of 1 to 7 applies for the specified spike solution with the given isotopic composition. If another spike solution with another isotope composition is used, the factor should be adjusted accordingly.

**Table 1 — Isotopic composition of natural Hg and recommended  $^{201}\text{Hg}$  enriched spike**

Hg-isotope	% of isotope in natural Hg	% of isotope in enriched Hg
200	23,10	0,89
201	13,18	96,50
202	29,86	2,37

Calculate the appropriate spike amount in gram,  $m_{\text{Sp}}$ , to be added from either 4.3.2, 4.3.3 or 4.3.4 to the sample using Formula (1):

$$m_{\text{Sp}} = \frac{\left( \frac{w_{\text{Se}} \times m_{\text{S}}}{7} \right)}{w_{\text{Sp}}} \quad (1)$$

where

$w_{\text{Se}}$  is the estimated mass fraction of MMHg in the sample, in ng/g or  $\mu\text{g}/\text{kg}$ ;

$m_{\text{S}}$  is the mass of the sample to be analysed, in g;

7 is the factor needed to achieve approximate abundance matching of the isotopes in the spiked sample;

$w_{\text{Sp}}$  is the exact mass fraction from the diluted MMHg stock solutions 4.3.2, 4.3.3 or 4.3.4, in ng/g.

The amount of spike to be added should not be below 0,1 g and should not exceed 1 g. This is to ensure enough significant figures from the weighing of the spike and to ensure the complete dissolution of the sample after TMAH addition, respectively.

**EXAMPLE** For a sample with an estimated MMHg mass fraction,  $w_{\text{Se}}$  of 100 ng/g ( $\mu\text{g}/\text{kg}$ ) where the mass of the sample to be analysed,  $m_{\text{S}}$ , is 0,2 g;  $m_{\text{Sp}}$  would be approximately 0,006 g for solution 4.3.2, approximately 0,06 g for solution 4.3.3 and approximately 0,6 g for solution 4.3.4. This is an example with rounded numbers, use exact mass fractions for analysis.

If the sample has been analysed for total mercury content, this value can be used as,  $w_{Se}$ , the estimated mass fraction of MMHg in the calculation. A correction factor should be used depending on the amount of MMHg of the total Hg (e.g. muscle from fish would give a factor of around 1 to 0,8, whereas fish liver is usually in the range of 0,4 to 0,6 and mussels in the range of 0,2 to 0,6).

## 6.2 Sample preparation

Weigh a test portion of 0,2 g to the nearest milligram in a polypropylene tube and spike with the calculated amount to the nearest milligram of the appropriate  $^{201}\text{Hg}$  enriched MMHg solution (4.3.2, 4.3.3 or 4.3.4). This step is crucial for the correct result. Add 3 ml of TMAH solution (4.4) to the tube and mix the contents by gentle rotating (5.2) until the sample is completely dissolved, if necessary overnight. After digestion, add 1 ml of pH 5 sodium acetate/acetic acid buffer (4.9) followed by 600  $\mu\text{l}$  of nitric acid (4.10) and mix the solution using a swirl mixer. Add 1 ml of hexane (4.13) and then 500  $\mu\text{l}$  of sodium tetraethyl borate (4.12) to derivatise the sample. Cap the tube and mix the contents by rotating for 10 min before centrifugation at 1 200  $g$  for 5 min. Transfer the hexane layer to a GC-vial and analyse the sample on the GC-ICP-MS instrument.

When using other nitric acid concentrations than described in 4.10, the volume to be added shall be adjusted accordingly so that a pH of 5 is achieved after acid addition and mixing.

NOTE: When using isotopic dilution, the efficiency of both the derivatisation and extraction into hexane does not need to be taken into account. However, too low a derivatisation/extraction efficiency can limit the sensitivity for samples with low MMHg mass fractions.

Low level species transformation is known to take place during extraction with TMAH, pH adjustment and derivatisation [4]. Verify the trueness of the method by determining reference materials of similar matrix as the samples to be determined.

## 6.3 ICP-MS conditions

Use the instrument parameters described in the manufacturers' operating manual.

Different ranges of instrument settings used in the interlaboratory test are given in Table 2, see also [5].

**Table 2 — Example of instrument settings for ICP-MS**

Parameter	Setting
Radiofrequency (RF-)Power (W)	600 to 1 600
Carrier gas flow ( $\text{l min}^{-1}$ )	0,1 to 1,7
Plasma gas flow ( $\text{l min}^{-1}$ )	13 to 16
Make up gas flow ( $\text{l min}^{-1}$ )	0 to 1
Optional gas (oxygen)	5 %
Mass resolution (amu)	0,8
Integration time (ms)	25 to 50
Isotopes measured	$^{200}\text{Hg}$ , $^{201}\text{Hg}$ , $^{202}\text{Hg}$

The integration time should be short enough to provide enough points per chromatographic peak depending on the GC-column used. It is recommended to collect 12 points to 20 points per chromatographic peak in order to define it properly. Shorter or longer integration times on the isotopes may influence the sensitivity and accuracy of the determination.

## 6.4 GC conditions

Use the instrument parameters described in the manufacturers' operating manual.

An example of instrument settings is given in Table 3.

**Table 3 — Example of instrument settings for capillary GC**

Parameter	Setting
Injection mode	Splitless or on-column
Injection volume ( $\mu\text{l}$ )	0,2 to 3
Injection temperature ( $^{\circ}\text{C}$ )	100 to 180
Transfer line temperature ( $^{\circ}\text{C}$ )	200 to 250
Carrier gas flow ( $\text{ml min}^{-1}$ )	1,2 to 2,6
GC oven temperature program	$50\text{ }^{\circ}\text{C}(1\text{ min}) \xrightarrow{50\text{ }^{\circ}\text{Cmin}^{-1}} 200\text{ }^{\circ}\text{C}(1\text{ min}) \xrightarrow{50\text{ }^{\circ}\text{Cmin}^{-1}} 280\text{ }^{\circ}\text{C}(1,7\text{ min})$

## 6.5 Set up procedures for the ICP-MS and GC-ICP-MS

Before starting routine measurements, the following set up procedure should be applied: Warm up the ICP-MS in full running mode for a minimum of 20 min to 30 min.

Use the optimising solution (4.14) to check the mass resolution and mass calibration and to adjust maximum sensitivity at low rates of oxides and doubly charged ions. Use the optimising solution to tune the instruments plasma parameters, ion lenses, analyser parameters and detector parameters.

In the event of high MMHg mass fractions, specific instrumental setups shall be properly adjusted (e. g. dead time correction, dual detector calibration) or the samples should be diluted and reanalysed.

Tune the interface after the GC is connected to the ICP-MS. The GC-ICP-MS interface can be tuned using 1 % to 2 % xenon in the carrier gas (helium) for the GC (5.11). The plasma parameters such as: x,y,z position of the torch, carrier gas flow, plasma RF-power and auxiliary gas flow may be tuned using one or more Xe-masses (e.g.  $^{124}\text{Xe}$ ,  $^{129}\text{Xe}$  or  $^{132}\text{Xe}$ ). If a sufficient sensitivity cannot be achieved simply by tuning the plasma parameters, the ion lens parameters may also be tuned using the same Xe-masses. Other tuning setups (e.g.  $^{36}\text{Ar}^{40}\text{Ar}$  or  $^{202}\text{Hg}$ ) may also be used provided they are able to produce sufficient sensitivity for the determination.

## 6.6 Sample measurement

### 6.6.1 Preparation of the GC-ICP-MS for analysis

Run the sample with the estimated lowest level of MMHg, either a reference material (e.g. NIST SRM 1566b<sup>2</sup>, see also [5] or a derivatised standard solution (e.g. 10  $\mu\text{g/l}$ ) to check that the instrument sensitivity and the chromatographic separation of the system are satisfactory and fit for purpose.

2) NIST SRM 1566b is an example of a suitable product available commercially. This information is given for the convenience of the users of this European Standard and does not constitute an endorsement by CEN of this product.

## 6.6.2 Determination of samples and blank solution

Inject appropriate amounts of the instrument blank, reagent blank and samples and determine the peak areas for the  $^{200}\text{Hg}$ ,  $^{201}\text{Hg}$  and  $^{202}\text{Hg}$  masses. If a significant blank value for MMHg is identified, the source of this blank should be identified. The source should be eliminated and the analysis repeated.

NOTE By use of isotopic dilution the calibration of the instruments with known amounts of MMHg is not necessary.

## 6.6.3 Quality control

Analyse instrument and reagent blank and reference samples of comparable matrix having reliably known contents of the element species to be determined in parallel with all the series of samples for analytical quality control. Subject the reagent blank and reference samples to all the steps in the method.

Check regularly within suitable short intervals (e.g. after 5 samples), a blank sample. Memory effects in the sample delivery system may influence the results of samples analysed after measurement of samples with high mass fractions of MMHg. Prolonged washout times may have to be applied. Test the system for wash out times using a sample with high MMHg mass fraction.

If the MMHg mass fractions obtained using the two isotope ratios differ more than 10 %, apply corrective measures. Check that the sample has been spiked with the appropriate amount to achieve approximate abundance matching of the isotopes in the sample. In addition, check the integration and other potential sources of error.

Since there is no sample cleanup (i.e. removal of fat) pay attention to the condition of the GC-liner when running multiple samples with high fat content.

## 7 Calculation

Using the peak areas found in the analyses, calculate the mass fraction of MMHg as Hg in the sample,  $w_s$  in mg/kg with the isotope dilution in Formula (2) [6]. Other versions of the same formula may also be used for calculation [7] [8].

$$w_s = w_{Sp} \frac{m_{Sp}}{m_s} \frac{M_s}{M_{Sp}} \frac{A_{Sp}^b}{A_s^a} \left( \frac{R_m - R_{Sp}}{1 - R_m \cdot R_s} \right) \frac{1}{1000} \quad (2)$$

where

- $w_{Sp}$  is the MMHg mass fraction in the spike solution, in ng/g ( $\mu\text{g}/\text{kg}$ );
- $m_s$  is the mass of test portion, in g;
- $m_{Sp}$  is the mass of added spike solution, in g;
- $M_s$  is the molar mass of MMHg in sample, in g/mol;
- $M_{Sp}$  is the molar mass of MMHg in spike solution, in g/mol;
- $A_s^a$  is the percent of reference isotope a ( $^{200}\text{Hg}$  or  $^{202}\text{Hg}$ ) in sample;
- $A_{Sp}^b$  is the percent of reference isotope b ( $^{201}\text{Hg}$ ) in spike solution;
- $R_s$  is the isotope ratio of Hg-isotope *a* and *b* in sample, see Formula (3);
- $R_{Sp}$  is the isotope ratio of Hg-isotope *a* and *b* in spike solution, see Formula (4);
- $R_m$  is the measured isotope ratio between Hg-isotope *a* and *b* in spiked sample, see

Formula (5);

1/1000 is the calculation factor from  $\mu\text{g}/\text{kg}$  to  $\text{mg}/\text{kg}$ .

$$R_S = \frac{A_S^b}{A_S^a} \quad (3)$$

where

$A_S^b$  is the percent of reference isotope b ( $^{201}\text{Hg}$ ) in sample;

$A_S^a$  is the percent of reference isotope a ( $^{200}\text{Hg}$  or  $^{202}\text{Hg}$ ) in sample.

$$R_{Sp} = \frac{A_{Sp}^a}{A_{Sp}^b} \quad (4)$$

where

$A_{Sp}^a$  is the percent of reference isotope a ( $^{200}\text{Hg}$  or  $^{202}\text{Hg}$ ) in spike solution;

$A_{Sp}^b$  is the percent of reference isotope b ( $^{201}\text{Hg}$ ) in spike solution.

$$R_m = \frac{N_m^a}{N_m^b} \quad (5)$$

where

$N_m^a$  is the peak area of the Hg-isotope a ( $^{200}\text{Hg}$  or  $^{202}\text{Hg}$ ) in the spiked sample;

$N_m^b$  is the peak area of the Hg-isotope b ( $^{201}\text{Hg}$ ) in the spiked sample.

Use both the  $^{202}\text{Hg}/^{201}\text{Hg}$  and  $^{200}\text{Hg}/^{201}\text{Hg}$  ratios for calculation and report the mean value of MMHg from these calculations. If the MMHg mass fractions obtained using the two isotope ratios differ more than 10 %, apply corrective measures (6.6.3). A higher tolerance for the difference can be acceptable in respect to the variance of the results for the matrix.

If the sample is dried prior to analysis, the result should be corrected for the moisture content.

## 8 Precision

### 8.1 General

Results from an interlaboratory test are summarized in Annex A. The values derived from this interlaboratory test may not be applicable to concentration ranges and matrices other than those given in Annex A. Further information can be found in a report on the conduction and results from the interlaboratory test [5].

### 8.2 Repeatability

The absolute difference between two independent single test results obtained with the same test method on identical test material in the same laboratory by the same operator using the same

apparatus within a short time interval will exceed the repeatability limit  $r$  given in Table 4 in not more than 5 % of the cases.

### 8.3 Reproducibility

The absolute difference between two single test results obtained with the same test method on identical test material in different laboratories by different operators using different equipment will exceed the reproducibility limit  $R$  given in Table 4 in not more than 5 % of the cases.

**Table 4 — Mean values, repeatability ( $r$ ) and reproducibility ( $R$ ) limits**

Sample (all lyophilized)	$\bar{x}$ mg/kg	$r$ mg/kg	$R$ mg/kg
Mussel tissue	0,035	0,003	0,014
Squid muscle	0,185	0,008	0,011
Crab claw muscle	0,307	0,012	0,032
Dog fish liver <sup>a</sup>	1,329	0,061	0,104
Whale meat	1,454	0,074	0,140
Cod muscle	1,976	0,064	0,142
Greenland Halibut muscle	3,575	0,076	0,244
<sup>a</sup> The dog fish liver sample was NRCC DOLT 4 with a certified value of 1,33 mg/kg			

## 9 Test report

The test report should fulfil the requirements in EN ISO/IEC 17025 [9] and specify at least the following:

- a) all information necessary for the complete identification of the sample;
- b) the test method used and the elemental species to be determined with reference to this European Standard;
- c) the results obtained and the units in which they are specified;
- d) the date of sampling procedure (if known);
- e) the date when the analysis was finished;
- f) all operating details not specified in this European Standard or regarded as optional, together with details of any incidents occurred when performing the method which might have influenced the test result(s).

## Annex A (informative)

### Precision data

The precision and trueness of the method was established by the CEN TC 275 “Food analysis – Horizontal methods” Working Group 10 “Elements and their chemical species” in an interlaboratory test according to ISO 5725-2 [1] among eight laboratories using freeze-dried samples, performed in 2013 under a mandate given by the European Commission. The results are given in Table A.1 and A.2, see also [5].

**Table A.1 — Precision data**

Parameter	Mussel tissue	Squid muscle	Crab claw meat	Whale meat	Cod muscle	Greenland halibut muscle
Number of reporting laboratories	8	8	8	8	8	8
Number of laboratories after elimination of outliers	7 <sup>a</sup>	8	8	8	8	8
Number of outlying laboratories	1	0	0	0	0	0
Number of replicates after outlier rejection	14	16	16	16	16	15
Numbers of rejected replicates	2 <sup>a</sup>	0	0	0	0	1
Mean value $\bar{x}$ , mg/kg	0,035	0,185	0,307	1,454	1,976	3,575
Repeatability limit $r$ , mg/kg	0,003	0,008	0,012	0,074	0,064	0,076
Repeatability standard deviation $s(r)$ , mg/kg	0,003	0,008	0,012	0,074	0,064	0,076
RSD( $r$ ), %	8,67	4,53	3,97	5,11	3,24	2,14
Reproducibility limit $R$ , mg/kg	0,014	0,011	0,032	0,140	0,142	0,244
Reproducibility standard deviation $s(R)$ , mg/kg	0,014	0,011	0,032	0,14	0,142	0,244
RSD( $R$ ), %	41,73	5,75	10,54	9,66	7,17	6,9
Horwitz value according to Horwitz [10]	26,53	20,62	19,11	15,12	14,44	13,21
HorRat value according to Horwitz [10]	1,57	0,28	0,55	0,64	0,50	0,52
Horwitz value according to Thompson [11]	22,00	20,62	19,11	15,12	14,44	13,21
HorRat value according to Thompson [11]	1,90	0,28	0,55	0,64	0,50	0,52

<sup>a</sup> MMHg mass fraction in the sample was not quantifiable by one of the participants.

**Table A.2 — Trueness based on NRCC DOLT 4 (Dog fish liver)**

Parameter	MMHg
Found mean value (mg/kg)	1,329
Certified value (mg/kg)	1,33
$S_R$ (mg/kg)	0,104
Z-score	- 0,01

## Bibliography

- [1] ISO 5725-2, *Accuracy (trueness and precision) of measurement methods and results — Part 2: Basic method for the determination of repeatability and reproducibility of a standard measurement method*
- [2] POINT D., DAVIS W.C., ALONSO J.I.G., MONPERRUS M., CHRISTOPHER S.J., DONARD O.F.X. et al. Simultaneous determination of inorganic mercury, methylmercury, and total mercury concentrations in cryogenic fresh-frozen and freeze-dried biological reference materials. *Anal. Bioanal. Chem.* 2007, **389** (3) pp. 787–798
- [3] CATTERICK T., FAIRMAN B., HARRINGTON C. Structured approach to achieving high accuracy measurements with isotope dilution inductively coupled plasma mass spectrometry. *J. Anal. At. Spectrom.* 1998, **13** (9) pp. 1009–1013
- [4] CLEMENS S., MONPERRUS M., DONARD O.F.X., AMOUROUX D., GUERIN T. Mercury speciation in seafood using isotope dilution analysis: A review. *Talanta.* 2012, **89** pp. 12–20
- [5] VALDERSNES S., JULSHAMN K. Report of the collaborative study on the quantitative determination of mono methylmercury in food of marine origin by GC-ICP-ID-MS. Accessible at: <http://nifes.no/wp-content/uploads/2015/05/reportofthemmhgcollaborativestudygicpms2.pdf>
- [6] RODRIGUEZ-GONZALEZ P., MARCHANTE-GAYON J.M., ALONSO J.I.G., SANZ-MEDEL A. Isotope dilution analysis for elemental speciation: A tutorial review. *Spectrochim. Acta B At. Spectrosc.* 2005, **60** (2) pp. 151–207
- [7] FASSETT J.D., PAULSEN P.J. Isotope-Dilution Mass-Spectrometry for Accurate Elemental Analysis. *Anal. Chem.* 1989, **61** (10) p. A643
- [8] LAMBERTSSON L. Mercury species transformations in marine and biological systems studied by isotope dilution mass spectrometry and stable isotope tracers. PhD Thesis 2005
- [9] EN ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories (ISO/IEC 17025)*
- [10] HORWITZ W., ALBERT R. The Horwitz Ratio (HorRat): A Useful Index of Method Performance with Respect to Precision. *J. AOAC Int.* 2006, **89** pp. 1095–1109
- [11] THOMPSON M. Recent Trends in Inter-Laboratory Precision at ppb and sub-ppb Concentrations in Relation to Fitness for Purpose Criteria in Proficiency Testing. *Analyst (Lond.)*. 2000, **125** pp. 385–386





# British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

## About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

## Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at [bsigroup.com/standards](http://bsigroup.com/standards) or contacting our Customer Services team or Knowledge Centre.

## Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at [bsigroup.com/shop](http://bsigroup.com/shop), where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

## Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to [bsigroup.com/subscriptions](http://bsigroup.com/subscriptions).

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

**PLUS** is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit [bsigroup.com/shop](http://bsigroup.com/shop).

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email [bsmusales@bsigroup.com](mailto:bsmusales@bsigroup.com).

## Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

## Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

## Useful Contacts:

### Customer Services

**Tel:** +44 845 086 9001

**Email (orders):** [orders@bsigroup.com](mailto:orders@bsigroup.com)

**Email (enquiries):** [cservices@bsigroup.com](mailto:cservices@bsigroup.com)

### Subscriptions

**Tel:** +44 845 086 9001

**Email:** [subscriptions@bsigroup.com](mailto:subscriptions@bsigroup.com)

### Knowledge Centre

**Tel:** +44 20 8996 7004

**Email:** [knowledgecentre@bsigroup.com](mailto:knowledgecentre@bsigroup.com)

### Copyright & Licensing

**Tel:** +44 20 8996 7070

**Email:** [copyright@bsigroup.com](mailto:copyright@bsigroup.com)

## BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

