



BSI Standards Publication

Foodstuffs - Determination of calcium, copper, iron, magnesium, manganese, phosphorus, potassium, sodium, sulfur and zinc by ICP-OES

National foreword

This British Standard is the UK implementation of EN 16943:2017.

The UK participation in its preparation was entrusted to Technical Committee AW/275, Food analysis - Horizontal methods.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2017
Published by BSI Standards Limited 2017

ISBN 978 0 580 91792 9

ICS 67.050

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 July 2017.

Amendments/corrigenda issued since publication

Date	Text affected

EUROPEAN STANDARD

EN 16943

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2017

ICS 67.050

English Version

**Foodstuffs - Determination of calcium, copper, iron,
magnesium, manganese, phosphorus, potassium, sodium,
sulfur and zinc by ICP-OES**

Produits alimentaires - Dosage du calcium, du cuivre,
du fer, du magnésium, du manganèse, du phosphore,
du potassium, du sodium, du soufre et du zinc par ICP-
OES

Lebensmittel - Bestimmung von Calcium, Kupfer, Eisen,
Magnesium, Mangan, Phosphor, Kalium, Natrium,
Schwefel und Zink mit ICP OES

This European Standard was approved by CEN on 20 February 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

	Page
European foreword.....	3
1 Scope.....	4
2 Normative references.....	4
3 Principle	4
4 Reagents	5
5 Apparatus.....	7
6 Procedure.....	8
6.1 Sample preparation.....	8
6.2 Optical emission spectrometry with inductively coupled plasma	8
6.2.1 ICP-OES working conditions.....	8
6.2.2 Determination with ICP-OES.....	8
6.3 Quality control of the analysis.....	10
7 Calculation.....	10
8 Precision.....	11
8.1 General.....	11
8.2 Repeatability.....	11
8.3 Reproducibility	11
9 Test report.....	14
Annex A (informative) Spectral interferences.....	15
Annex B (informative) Precision Data	18
B.1 Details on the inter-laboratory study	18
Bibliography.....	30

European foreword

This document (EN 16943:2017) has been prepared by Technical Committee CEN/TC 275 "Food analysis - Horizontal methods", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2017, and conflicting national standards shall be withdrawn at the latest by November 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

1 Scope

This European Standard describes a method for the determination of calcium, copper, iron, magnesium, manganese, phosphorus, potassium, sodium, sulfur and zinc in different foodstuffs using optical emission spectrometry with inductively coupled plasma (ICP-OES) after pressure digestion.

This method has been validated in an interlaboratory study according to ISO 5725 [1] on infant formula soya based, cheese, chicken meat, wheat flour, apple juice, lobster and milk (see elements ranges Table 1 and validation data in Annex B).

HorRat values greater than 2 have been observed for certain analyte/matrix combinations during the validation study.

Table 1 — Validated element ranges

Element	Range mg/kg
Calcium	70 to 7178
Copper	0,60 to 16,40
Iron	0,88 to 77
Magnesium	45 to 1 174
Manganese	0,44 to 5,12
Phosphorus	72 to 9 708
Potassium	605 to 14 312
Sodium	11 to 2 220
Sulfur	26 to 8 542
Zinc	0,16 to 43,5

At European or International level, vertical standards for the determination of specific minerals can exist, e.g. for milk and milk products or for animal and vegetable fats and oils [2].

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 13804, *Foodstuffs - Determination of elements and their chemical species - General considerations and specific requirements*

EN 13805, *Foodstuffs - Determination of trace elements - Pressure digestion*

3 Principle

After digestion of the sample with the pressure digestion process described in EN 13805, calcium, copper, iron, magnesium, manganese, phosphorus, potassium, sodium, sulfur and zinc are determined quantitatively with the ICP-OES [2]. The digestion solution is nebulized, the aerosol is directed into an inductively coupled argon plasma, where the elements are atomized and excited for radiation. The emission radiation is resolved spectrally and its intensity determined at element-specific wavelengths with a detector system. Ionization interference can be minimized using an ionization buffer.

4 Reagents

The concentration of the elements to be determined shall be low enough in the reagents and water not to affect the results.

4.1 Nitric acid, minimum $w = 65\%$ ¹, density about 1,4 g/ml

4.2 Hydrochloric acid, $w = 30\%$, density = 1,15 g/ml

4.3 Hydrogen peroxide, $w = 30\%$

Hydrogen peroxide (4.3) can contain phosphoric acid as a stabilizing agent. Hydrogen peroxide stabilized with phosphoric acid contaminates the samples with phosphorus and thus leads to incorrect results.

4.4 Caesium chloride solution as ionization buffer, e.g. $w = 10\%$ (see 6.2.2)

4.5 Internal standard, e.g. scandium, ytterbium, yttrium

Depending on the matrix, e.g. for samples with high total salt content, the application of an internal standard can be advantageous in order to reduce physical interference. In doing so, it shall be ensured that no faults from line interference result due to the internal standard.

NOTE It has been shown that if an internal standard is used, it is applicable to all elements mentioned in the scope.

4.6 Stock solutions, $\rho = 1\,000 \text{ mg/l}$ ² per element

Calcium, copper, iron, magnesium, manganese, phosphorus, potassium, sodium, sulfur and zinc individual element stock solutions and internal standard used.

The individual element stock solutions may not contain any of the elements to be determined here as accompanying substances or impurities, which could affect the correctness of the results.

Alternatively, commercial mixed stock solutions can also be used.

4.7 Standard solutions, $\rho = 100 \text{ mg/l}$ per element

Copper, iron, manganese and zinc standard solution, produced by dilution of the respective stock solution (4.6).

Pour 10 ml of water into a 50-ml volumetric flask, add 2 ml of nitric acid (4.1) and mix. After cooling to room temperature, add exactly 5 ml respectively of the copper, iron, manganese and zinc stock solution (4.6) by pipette and fill up with water to the mark. These standard solutions are stable for at least six months.

4.8 Element reference solutions

4.8.1 General

The concentrations of the reference solutions indicated in 4.8.3 to 4.8.6 serve as examples and shall be modified corresponding to the instrument sensitivity, the direction of viewing (axial or radial) and the concentration range to be examined, if necessary. If possible, the calibration shall be performed in the

¹ w = mass fraction

² ρ = mass concentration

linear range. Minimum 3, in the nonlinear range 5, reference solutions of different concentrations should be prepared for the calibration. The concentration of the reagents (acids, if applicable ionization buffer and internal standard) shall correspond to the sample measuring solution.

For reasons of stability and purity, the phosphorus and sulfur reference solutions should be prepared separately from the multi-element reference solutions of the cations.

4.8.2 Preparation of the flask

Pour 10 ml to 20 ml of water into a 100-ml volumetric flask, add 10 ml of nitric acid (4.1) and 2 ml of hydrochloric acid (4.2) and mix. After cooling to room temperature, add stock (4.6) and standard solutions (4.7) by pipette. If using an ionization buffer, add 10 ml of 10 % caesium chloride solution (4.4). If using an internal standard (4.5), add the corresponding amount. Alternatively, the ionization buffer and/or the internal standard can be pumped via a Y-piece into the sample flow during the measurement.

These reference solutions are stable for a month.

The following data under 4.8.3 to 4.8.6 serve as examples and shall be adapted corresponding to the device sensitivity and the emission wavelength selected, and the direction of viewing (axial or radial), if necessary.

4.8.3 Reference solution 1

Pipette the following amounts of stock solutions (4.6) into the prepared 100-ml volumetric flask (4.8.2):

- 2,0 ml calcium →³⁾ $\rho = 20 \text{ mg/l}$ calcium;
- 1,0 ml magnesium → $\rho = 10 \text{ mg/l}$ magnesium;
- 2,0 ml potassium → $\rho = 20 \text{ mg/l}$ potassium;
- 1,0 ml sodium → $\rho = 10 \text{ mg/l}$ sodium;

Furthermore, pipette the following amounts of standard solutions (4.7) into the same flask and fill up to the mark with water:

- 0,50 ml copper → $\rho = 0,5 \text{ mg/l}$ copper;
- 0,50 ml iron → $\rho = 0,5 \text{ mg/l}$ iron;
- 0,50 ml manganese → $\rho = 0,5 \text{ mg/l}$ manganese;
- 0,50 ml zinc → $\rho = 0,5 \text{ mg/l}$ zinc.

4.8.4 Reference solution 2

Pipette the following amounts of stock solutions (4.6) into the prepared 100-ml volumetric flask (4.8.2):

- 10,0 ml calcium → $\rho = 100 \text{ mg/l}$ calcium;
- 5,0 ml magnesium → $\rho = 50 \text{ mg/l}$ magnesium;
- 10,0 ml potassium → $\rho = 100 \text{ mg/l}$ potassium;
- 5,0 ml sodium → $\rho = 50 \text{ mg/l}$ sodium.

3) the sign "→" means: "which leads to a mass concentration of"

Furthermore, pipette the following amounts of standard solutions (4.7) into the same flask and fill up to the mark with water:

- 2,0 ml copper → $\rho = 2,0 \text{ mg/l}$ copper;
- 2,0 ml iron → $\rho = 2,0 \text{ mg/l}$ iron;
- 2,0 ml manganese → $\rho = 2,0 \text{ mg/l}$ manganese;
- 2,0 ml zinc → $\rho = 2,0 \text{ mg/l}$ zinc.

4.8.5 Reference solution 3

Pipette the following amounts of stock solutions (4.6) into the prepared 100-ml volumetric flask (4.8.2):

- 20,0 ml calcium → $\rho = 200 \text{ mg/l}$ calcium;
- 10,0 ml magnesium → $\rho = 100 \text{ mg/l}$ magnesium;
- 20,0 ml potassium → $\rho = 200 \text{ mg/l}$ potassium;
- 10,0 ml sodium → $\rho = 100 \text{ mg/l}$ sodium.

Furthermore, pipette the following amounts of standard solutions (4.7) into the same flask and fill up to the mark with water:

- 5,0 ml copper → $\rho = 5,0 \text{ mg/l}$ copper;
- 5,0 ml iron → $\rho = 5,0 \text{ mg/l}$ iron;
- 5,0 ml manganese → $\rho = 5,0 \text{ mg/l}$ manganese;
- 5,0 ml zinc → $\rho = 5,0 \text{ mg/l}$ zinc.

4.8.6 Reference solutions phosphorus and sulfur

Pipette the following amounts of stock solutions (4.6) into the prepared 100-ml volumetric flask (4.8.2) and fill up to the mark with water:

- 1,0 ml respectively phosphorus and sulfur → $\rho = 10,0 \text{ mg/l}$ phosphorus and sulfur;
- 5,0 ml respectively phosphorus and sulfur → $\rho = 50,0 \text{ mg/l}$ phosphorus and sulfur;
- 10,0 ml respectively phosphorus and sulfur → $\rho = 100,0 \text{ mg/l}$ phosphorus and sulfur.

4.9 Zero value solution

The zero value solution contains water, 10 ml of nitric acid (4.1), 2 ml of hydrochloric acid (4.2), if applicable 10 ml caesium chloride solution (4.4) and a corresponding amount of internal standard (4.5) in 100 ml.

5 Apparatus

5.1 General

All equipment and labware that come into direct contact with the sample and the solutions used shall be carefully pretreated/cleaned corresponding to EN 13804 to minimize the blank value.

5.2 ICP-OES, optical emission spectrometer with inductively coupled argon plasma, sample supply and atomization system as well as device control and evaluation unit.

6 Procedure

6.1 Sample preparation

Mineralize the sample in a pressure digestion corresponding to EN 13805.

In the case of dry, powdery or free-flowing materials (water content below 20 %), add 2 ml of water in relation to a test portion of 400 mg, and mix intensively so that the sample is well suspended in the water. Add nitric acid (4.1) and mix. For the determination of iron in order to avoid adsorption losses, add 0,5 ml to 1 ml of hydrochloric acid (4.2) to the digestion vessel, in addition to the nitric acid (4.1). The amount of hydrochloric acid depends on the amount of nitric acid used. Do not add the hydrochloric acid before/until the termination/expiration of the spontaneous reaction caused by the nitric acid. Seal the digestion vessel immediately after addition of the hydrochloric acid. The pressure digestion should be started shortly afterwards. Use digestion conditions according to the device manufacturer, the reactivity of the sample, the maximum pressure stability of the digestion vessel and the attainable temperature.

NOTE The trueness of the determination of the other elements is not impaired by the addition of hydrochloric acid during the digestion.

Fill up the digestion solution obtained according to EN 13805 after the pressure digestion to a definite volume, e.g. 20 ml. It can be used directly or after dilution for the following element determinations. The same concentration of acids, if applicable internal standard and ionization buffer, shall be present in all measuring solutions as in the reference solutions.

6.2 Optical emission spectrometry with inductively coupled plasma

6.2.1 ICP-OES working conditions

Ensure that the device is set corresponding to the manufacturer's data and the plasma is ignited. After sufficient heating and stabilization of the device, optimize the settings [3], [4].

6.2.2 Determination with ICP-OES

Start the measurements after optimizing the device. The following wavelengths (Table 2), which have been tested for possible interference (see Table A.1), are recommended for determination of elements and for internal standards (Table 3). The wavelength data in the device-specific spectral libraries can deviate. Further wavelengths can be used, although these shall be tested for potential interference in individual cases.

Table 2 — Recommended wavelengths for element determination

Element	Wavelengths nm	Possible interference
Calcium	317,93 315,89 422,67	Ce, Fe, Mo, V Co, Mo, OH band Mo, V, Zr
Copper	324,75 327,40	Cr, Fe, Mo, Nb, Ti, U Co, Nb, Ti, U
Iron	259,94 238,20	Co, Mn, Nb Co
Magnesium	279,08 285,21	Fe, Nb, Rh Cr, Fe, Mo, W
Manganese	257,61 293,31	Cr, Fe, Mo, W Al, Cr, Fe, Ti
Phosphorus	213,62 214,91 177,43	Co, Cu, Fe, Mo, Zn Al, Cu, Fe, Mg Cu
Potassium	766,49 769,90 404,72	Ar, Ba, Mg, Ba Ba
Sodium	589,59 330,24	Ar, Ba, V Zn
Sulfur	181,97 180,67 182,56	B, Nb, Ni, Pb As, Ca, Co, Ni, Si B, Cu, Mg
Zinc	213,86 206,20	Al, As, Cu, Fe, Mg, Ni Cr, Nb

Table 3 — Recommended wavelengths for internal standards

Element	Wavelength nm
Scandium	361,38
Ytterbium	328,94
Yttrium	371,03

When determining sodium and potassium, ionization interferences are generally present. In general, an ionization buffer, e.g. caesium chloride ($w = 1\%$ in the measurement solution), should be used for the axial recording of the plasma emission. In the case of radial recording of the emission intensity, the user of this standard shall establish whether the use of an ionization buffer can be omitted by optimizing the observation height and the gas settings. If element lines below 190 nm are applied, ensure sufficient inert gas purging between the plasma and inlet gap as well as the spectrometer so that oxygen absorption cannot cause any drifting of intensities.

If the ionization buffer is added via a Y-piece, adjust volume ratios so that the solution that enters the atomizer or plasma contains an excess of caesium chloride.

Measure zero value solution (4.9) and reference solutions (4.8) and prepare a calibration curve using emission intensities or peak areas and concentrations. In the case of complex matrices, the standard addition method may be necessary.

Aspirate sample solution and measure it. The determined peak areas or intensities are converted via calibration curve using linear regression into concentration units.

Ensure sufficient time for rinsing before the next measurement in case of samples with high element concentrations. Rinse performance can be monitored with zero value solution (4.9).

When making dilutions, ensure that the diluted measuring solutions have the same concentrations of reagents as the original measuring solutions.

6.3 Quality control of the analysis

For quality control, analyse control samples with reliably known contents of the elements to be determined in parallel to every measurement series. The control samples should pass through all analytical steps, beginning with the digestion. Likewise, prepare reagent blank solutions and measure for every digestion series, taking into account all process steps.

To check for matrix effect spiking experiments should be performed by adding a known concentration of the elements to the measurement solution.

7 Calculation

Calculate the mass fraction w in milligrams per kilogram or litre of sample according to Formula (1):

$$w = \frac{a \cdot V \cdot F}{m} \quad (1)$$

where

a is the mass fraction of the element in the sample solution in milligrams per litre;

V is the volume of the sample solution after digestion in millilitres;

F is the dilution factor of the sample measuring solution;

m is the mass of the test portion in grams or the sample volume in millilitres used for digestion.

Causes of increased element contents in the blank value solution shall be clarified and, if necessary, deducted from the calculation of the results.

8 Precision

8.1 General

Details of the inter-laboratory test of the precision of the methods are summarized in Annex B. The values derived from this test may not be applicable to analyte concentration ranges and matrices other than given in Annex B.

8.2 Repeatability

The absolute difference between two independent single test results obtained using the same method on identical test material in the same laboratory by the same operator using the same equipment within a short time interval, will in no more than 5 % of the cases exceed the values of r given in Table 4.

8.3 Reproducibility

The absolute difference between two single test results obtained using the same method on identical test material in different laboratories by different operators using the equivalent equipment will in no more than 5 % of the cases exceed the values of R given in Table 4.

Table 4 — Precision data

Sample/elements	\bar{x} mg/kg	r mg/kg	R mg/kg
Infant formula (soya based)			
Calcium	6 191	591	1 381
Copper	4,51	0,54	1,40
Iron	77,0	5,9	15,0
Magnesium	599	72	128
Manganese	2,19	0,29	0,84
Phosphorus	4 129	399	909
Potassium	6 733	769	1 035
Sodium	2 220	228	304
Sulfur	1 234	133	370
Zinc	43,5	3,2	8,4
Cheese			
Calcium	7 178	394	930
Copper	14,66	0,96	2,53
Iron	2,3	0,4	1,1
Magnesium	587	58	104
Manganese	0,2	-	-
Phosphorus	9 708	1 080	782
Potassium	8 397	879	1 293
Sodium	2 039	132	235

Sample/elements	\bar{x} mg/kg	r mg/kg	R mg/kg
Sulfur	5 764	504	1 176
Zinc	38,4	3,5	7,1
Chicken meat			
Calcium	186	18	33
Copper ^a	1,50	0,70	1,35
Iron	24,0	4,4	7,7
Magnesium	1 147	116	242
Manganese	1,35	0,17	0,46
Phosphorus	9 034	1 022	1 911
Potassium	14 312	473	2 459
Sodium	1 303	124	145
Sulfur	8 542	465	1 464
Zinc	24,8	3,0	4,6
Wheat flour			
Calcium	305	23	34
Copper	2,51	0,33	0,90
Iron	15,3	1,9	3,9
Magnesium	397	27	74
Manganese	5,12	0,40	0,92
Phosphorus	1 466	119	258
Potassium	1 325	163	203
Sodium ^a	11	3	8
Sulfur	1 694	149	320
Zinc	11,2	1,2	1,5
Apple juice			
Calcium	70	6	9
Copper	0,04	-	-
Iron	0,88	0,11	0,19
Magnesium	45	3	10
Manganese	0,44	0,05	0,09
Phosphorus	72	7	15
Potassium	1 090	99	196

Sample/elements	\bar{x} mg/kg	r mg/kg	R mg/kg
Sodium	19	4	9
Sulfur	26	6	13
Zinc	0,16	0,04	0,09
Lobster			
Calcium	183	25	32
Copper	16,40	2,63	3,13
Iron	12,1	2,2	2,9
Magnesium	85	9	20
Manganese	1,20	0,16	0,27
Phosphorus	973	86	194
Potassium	871	88	153
Sodium	1 175	9	217
Sulfur	876	77	185
Zinc	13,9	1,8	3,1
Milk			
Calcium a	516	45	209
Copper	0,60	0,12	0,21
Iron	7,7	0,7	1,4
Magnesium	53	4	11
Manganese	0,04	-	-
Phosphorus a	319	30	130
Potassium	605	85	148
Sodium	186	14	34
Sulfur	177	13	44
Zinc	5,1	0,4	0,9
a This element/matrix combination lead to a Horrat value of greater than 2 in the validation study, see Annex B.			

9 Test report

The test report should comply with EN ISO/IEC 17025 [5] and shall contain at least the following data:

- a) all information necessary for the complete identification of the sample;
- b) the test method used and the elements to be determined with reference to this European standard;
- c) the results obtained and the units in which they are specified;
- d) the date of the sampling procedure (if known);
- e) the date when the analyses was finished;
- f) all operating details not specified in this European Standard or regarded as optional, together with details of any incidents occurred when performing the method which might have influenced the test results.

Annex A (informative)

Spectral interferences

The line interferences were checked with three device systems of different design and resolution capacity (Perkin Elmer Optima 3000, Spectro Ciros, Varian Vista). The interferences listed in the following Table A.1 have been checked with the concentrations indicated.

Line interference is generally affected by the resolution of the spectrometer used and the background correction and shall therefore always be checked on the spectrometer used for the measurement. Table A.1 can thus only serve as a starting point for the level of an interference.

Table A.1 — Line interferences

Element (Analyte)	Wavelength nm	Interfering element	Concentration of interfering element mg/l	Apparent analyte concentration mg/l
Calcium	317,93	Ce	10	> 0,01
		Fe	50	< 0,01
		Mo	10	< 0,01
		V	10	> 0,01
	315,89	Co	10	> 0,01
		Mo	10	> 0,01
	422,67	Mo	10	< 0,01
		V	10	> 0,01
		Zr	10	< 0,01
Copper	324,75	Cr	10	< 0,01
		Fe	50	< 0,01
		Mo	10	< 0,01
		Nb	10	> 0,1
		Ti	10	< 0,01
		U	10	< 0,01
	327,40	Co	10	< 0,01
		Nb	10	> 0,01
		Ti	10	< 0,01
		U	10	> 0,01
Iron	259,94	Co	10	< 0,01
		Mn	10	< 0,01
		Nb	10	< 0,01
	238,20	Co	10	< 0,01

Element (Analyte)	Wavelength nm	Interfering element	Concentration of interfering element mg/l	Apparent analyte concentration mg/l
Magnesium	279,08	Fe	50	> 0,01
		Nb	10	> 0,1
		Rh	10	> 0,1
	285,21	Cr	10	< 0,01
		Fe	50	< 0,01
		Mo	10	< 0,01
		W	10	< 0,01
Manganese	257,61	Cr	10	< 0,01
		Fe	50	< 0,01
		Mo	10	< 0,01
		W	10	< 0,01
	293,31	Al	50	< 0,01
		Cr	10	< 0,01
		Fe	50	< 0,01
		Ti	10	< 0,01
Phosphorus	213,62	Co	10	< 0,01
		Cu	10	> 0,2
		Fe	50	< 0,01
		Mo	10	> 0,01
		Zn	50	> 0,01
	214,91	Al	50	< 0,01
		Cu	10	> 0,05
		Fe	50	> 0,05
		Mg	200	< 0,01
Potassium	177,43	Cu	10	> 0,01
	766,49	Ba	10	< 0,01
		Mg	200	< 0,01
	769,90	Ba	10	< 0,01
Sodium	404,72	Ba	10	< 0,01
	589,59	Ba	10	> 1
		V	10	< 0,01
	330,24	Zn	10	> 5

Element (Analyte)	Wavelength nm	Interfering element	Concentration of interfering element mg/l	Apparent analyte concentration mg/l
Sulfur	181,97	B	10	< 0,1
		Nb	10	< 0,1
		Ni	10	< 0,1
		Pb	10	< 0,1
	180,67	As	10	< 0,1
		Ca	200	> 2
		Co	10	< 0,1
		Ni	10	< 0,1
		Si	50	< 0,1
	182,56	B	10	> 1
		Cu	10	< 0,1
		Mg	200	< 0,1
Zinc	213,86	Al	50	< 0,01
		As	10	< 0,01
		Cu	10	> 0,01
		Fe	50	> 0,01
		Mg	200	< 0,01
		Ni	10	> 0,05
	206,20	Cr	10	> 0,2
		Nb	10	< 0,01

Annex B
(informative)

Precision Data

B.1 Details on the inter-laboratory study

The data given in Table B.1 to Table B.10 were obtained in an inter-laboratory study [2] organized in 2011 by the Working Group “Elementanalytik” of the Federal Office for Consumer Protection and Food Safety (Bundesamt für Verbraucherschutz und Lebensmittelsicherheit — BVL) according to the German Food and Feed Act, Paragraph 64 (Lebensmittel- und Futtermittelgesetzbuch-LFGB, § 64) in accordance with ISO 5725-2, -4 and -6 with 13 participating laboratories. Each quantitative value of a laboratory results from a double determination or, in the case of cheese (double blind), from a fourfold determination. The materials used in the interlaboratory study have been fresh materials (milk, apple juice and lobster) and dry materials (wheat flour, infant formula, cheese and chicken meat). Four materials were reference materials (NCS ZC73009 wheat, NCS ZC73016 chicken, NRC LUTS-1 non defatted lobster hepatopancreas (approximately 8 % fat) and dried low fat fresh cheese (approximately 4,5 % fat, 23 % carbohydrates and 67 % protein) as a former proficiency test material). The infant formula soya based material contained 58 % carbohydrates, 24 % fat and 13 % protein. The apple juice was a commercial available product mainly containing 10 % carbohydrates.

The milk with a typical fat content of 4 %, 2,5 % protein, a dry weight of 12 % and a carbohydrate content of 8,5 % was enriched with minerals and proteins. For calcium and phosphorous the reproducibility was relatively high. The laboratories did not use the same procedure for homogenization which led to insufficiently homogenous partition of the casein fraction in the sample during sample preparation.

In Table B.11 the data concerning trueness determined by the use of reference material is given.

Table B.1 — Validation data for calcium

Statistical parameter	Calcium						
	Infant formula soya based	Cheese	Chicken meat	Wheat flour	Apple juice	Lobster	Milk
Number of participating laboratories	13	13	13	13	13	13	13
Number of laboratories with quantitative values	13	13	13	13	13	13	13
Number of laboratories after elimination of outliers	13	12	12	12	12	13	13
Number of outlier laboratories	0	1	1	1	1	0	0
Mean value $\bar{x} \pm$ confidence interval, mg/kg	6 191 ±261	7 178 ±176	186 ±6	305 ±6	70 ±2	183 ±5	516 ±41
Reproducibility standard deviation s_R , mg/kg	493	332	12	12	3	12	75
Relative reproducibility standard deviation $s_{R,rel}$	7,97 %	4,62 %	6,24 %	3,97 %	4,67 %	6,31 %	14,50 %
Reproducibility limit R , mg/kg	1 381	930	33	34	9	32	209
Relative reproducibility limit R_{rel}	22,31 %	12,95 %	17,48 %	11,10 %	13,06 %	17,67 %	40,60 %
Repeatability standard deviation s_r , mg/kg	211	141	6	8	2	9	16
Relative repeatability standard deviation $s_{r,rel}$	3,41 %	1,96 %	3,37 %	2,70 %	3,03 %	4,90 %	3,09 %
Repeatability limit r , mg/kg	591	394	18	23	6	25	45
Relative repeatability limit r_{rel}	9,55 %	5,48 %	9,45 %	7,56 %	8,49 %	13,71 %	8,64 %
Relative Horwitz standard deviation	4,30 %	4,20 %	7,28 %	6,76 %	8,45 %	7,30 %	6,25 %
HorRat	1,85	1,10	0,86	0,59	0,55	0,86	2,32

Table B.2 — Validation data for copper

Statistical parameter	Copper						
	Infant formula soya based	Cheese	Chicken meat	Wheat flour	Apple juice^a	Lobster	Milk
Number of participating laboratories	13	13	13	13	13	13	13
Number of laboratories with quantitative values	13	13	11	13	5	13	13
Number of laboratories after elimination of outliers	13	12	11	12	4	13	13
Number of outlier laboratories	0	1	0	1	1	0	0
Mean value $\bar{x} \pm$ confidence interval, mg/kg	4,51 ±0,27	14,66 ±0,48	1,50 ±0,27	2,51 ±0,18	0,04	16,40 ±0,50	0,60 ±0,04
Reproducibility standard deviation s_R , mg/kg	0,50	0,90	0,48	0,32		1,12	0,07
Relative reproducibility standard deviation $s_{R,rel}$	11,06 %	6,15 %	32,25 %	12,76 %		6,82 %	12,20 %
Reproducibility limit R , mg/kg	1,40	2,53	1,35	0,90		3,13	0,21
Relative reproducibility limit R_{rel}	30,98 %	17,23 %	90,31 %	35,74 %		19,11 %	34,15 %
Repeatability standard deviation s_r , mg/kg	0,19	0,34	0,25	0,12		0,94	0,04
Relative repeatability standard deviation $s_{r,rel}$	4,30 %	2,34 %	16,75 %	4,67 %		5,72 %	7,04 %
Repeatability limit r , mg/kg	0,54	0,96	0,70	0,33		2,63	0,12
Relative repeatability limit r_{rel}	12,03 %	6,54 %	46,89 %	13,08 %		16,01 %	19,70 %
Relative Horwitz standard deviation	12,75 %	10,68 %	15,06 %	13,93 %		10,50 %	17,27 %
HorRat	0,87	0,58	2,14	0,92		0,65	0,71

^a Only some of the participating laboratories have determined quantitative values for the copper content in the apple juice. As less than 7 values were available for determination of the precision data after eliminating the outliers, precision data will not be indicated at this point. The copper content of this sample was in the range of the quantification limit at 0,04 mg/kg.

Table B.3 — Validation data for iron

Statistical parameter	Iron						
	Infant formula soya based	Cheese	Chicken meat	Wheat flour	Apple juice	Lobster	Milk
Number of participating laboratories	13	13	13	13	13	13	13
Number of laboratories with quantitative values	13	12	13	13	12	13	13
Number of laboratories after elimination of outliers	13	11	13	12	12	13	13
Number of outlier laboratories	0	1	0	1	0	0	0
Mean value $\bar{x} \pm$ confidence interval, mg/kg	77,0 ±2,9	2,3 ±0,2	24,0 ±1,4	15,3 ±0,8	0,88 ±0,04	12,1 ±0,5	7,7 ±0,3
Reproducibility standard deviation s_R , mg/kg	5,4	0,4	2,8	1,4	0,07	1,0	0,5
Relative reproducibility standard deviation $s_{R,rel}$	6,98 %	16,83 %	11,49 %	9,19 %	7,81 %	8,59 %	6,59 %
Reproducibility limit R , mg/kg	15,0	1,1	7,7	3,9	0,19	2,9	1,4
Relative reproducibility limit R_{rel}	19,55 %	47,13 %	32,17 %	25,73 %	21,86 %	24,05 %	18,44 %
Repeatability standard deviation s_r , mg/kg	2,1	0,1	1,6	0,7	0,04	0,8	0,3
Relative repeatability standard deviation $s_{r,rel}$	2,75 %	6,13 %	6,61 %	4,48 %	4,38 %	6,45 %	3,47 %
Repeatability limit r , mg/kg	5,9	0,4	4,4	1,9	0,11	2,2	0,7
Relative repeatability limit r_{rel}	7,69 %	17,15 %	18,50 %	12,54 %	12,27 %	18,06 %	9,72 %
Relative Horwitz standard deviation	8,32 %	14,10 %	9,92 %	10,62 %	16,31 %	11,00 %	11,78 %
HorRat	0,84	1,19	1,16	0,87	0,48	0,78	0,56

Table B.4 — Validation data for magnesium

Statistical parameter	Magnesium						
	Infant formula soya based	Cheese	Chicken meat	Wheat flour	Apple juice	Lobster	Milk
Number of participating laboratories	13	13	13	13	13	13	13
Number of laboratories with quantitative values	13	13	13	13	13	13	13
Number of laboratories after elimination of outliers	13	13	13	13	13	13	13
Number of outlier laboratories	0	0	0	0	0	0	0
Mean value $\bar{x} \pm$ confidence interval, mg/kg	599 ±23	587 ±18	1 174 ±45	397 ±14	45 ±2	85 ±4	53 ±2
Reproducibility standard deviation s_R , mg/kg	46	37	86	26	4	7	4
Relative reproducibility standard deviation $s_{R,rel}$	7,64 %	6,35 %	7,35 %	6,67 %	7,97 %	8,63 %	7,68 %
Reproducibility limit R , mg/kg	128	104	242	74	10	20	11
Relative reproducibility limit R_{rel}	21,38 %	17,79 %	20,59 %	18,67 %	22,31 %	24,18 %	21,50 %
Repeatability standard deviation s_r , mg/kg	26	21	41	10	1	3	1
Relative repeatability standard deviation $s_{r,rel}$	4,30 %	3,55 %	3,52 %	2,44 %	2,29 %	3,73 %	2,55 %
Repeatability limit r , mg/kg	72	58	116	27	3	9	4
Relative repeatability limit r_{rel}	12,05 %	9,95 %	9,87 %	6,84 %	6,42 %	10,44 %	7,13 %
Relative Horwitz standard deviation	6,11 %	6,13 %	5,52 %	6,50 %	9,01 %	8,20 %	8,81 %
HorRat	1,25	1,04	1,33	1,03	0,88	1,05	0,87

Table B.5 — Validation data for manganese

Statistical parameter	Manganese						
	Infant formula soya based	Cheese^a	Chicken meat	Wheat flour	Apple juice	Lobster	Milk^b
Number of participating laboratories	13	13	13	13	13	13	13
Number of laboratories with quantitative values	13	7	11	13	13	13	6
Number of laboratories after elimination of outliers	13	6	11	13	13	13	6
Number of outlier laboratories	0	1	0	0	0	0	0
Mean value $\bar{x} \pm$ confidence interval, mg/kg	2,19 ±0,16	0,2	1,35 ±0,10	5,12 ±0,17	0,44 ±0,02	1,20 ±0,05	0,04
Reproducibility standard deviation s_R , mg/kg	0,30		0,16	0,33	0,03	0,10	
Relative reproducibility standard deviation $s_{R,rel}$	13,71 %		12,15 %	6,39 %	6,85 %	7,95 %	
Reproducibility limit R , mg/kg	0,84		0,46	0,92	0,09	0,27	
Relative reproducibility limit R_{rel}	38,38 %		34,03 %	17,90 %	19,17 %	22,26 %	
Repeatability standard deviation s_r , mg/kg	0,10		0,06	0,14	0,02	0,06	
Relative repeatability standard deviation $s_{r,rel}$	4,67 %		4,46 %	2,81 %	3,71 %	4,74 %	
Repeatability limit r , mg/kg	0,29		0,17	0,40	0,05	0,16	
Relative repeatability limit r_{rel}	13,09 %		12,48 %	7,88 %	10,38 %	13,28 %	
Relative Horwitz standard deviation	14,21 %		15,29 %	12,51 %	18,07 %	15,57 %	
HorRat	0,96		0,79	0,51	0,38	0,51	

^a Only some of the participating laboratories have determined quantitative values for the manganese content in the cheese. As less than 7 values were available for determination of the precision data after eliminating the outliers, precision data will not be indicated at this point. The manganese content of this sample was in the range of the quantification limit at 0,2 mg/kg.

^b Only some of the participating laboratories have determined quantitative values for the manganese content in milk. As less than 7 values were available for determination of the precision data, precision data will not be indicated at this point. The manganese content of this sample was in the range of the quantification limit at 0,04 mg/kg.

Table B.6 — Validation data for phosphorus

Statistical parameter	Phosphorus						
	Infant formula soya based	Cheese	Chicken meat	Wheat flour	Apple juice	Lobster	Milk
Number of participating laboratories	13	13	13	13	13	13	13
Number of laboratories with quantitative values	13	13	13	13	13	13	13
Number of laboratories after elimination of outliers	13	13	13	13	13	13	13
Number of outlier laboratories	0	0	0	0	0	0	0
Mean value $\bar{x} \pm$ confidence interval, mg/kg	4 129 ±171	9 708 ±300	9 034 ±351	1 466 ±48	72 ±3	973 ±37	319 ±25
Reproducibility standard deviation s_R , mg/kg	325	636	683	92	5	69	46
Relative reproducibility standard deviation $s_{R,rel}$	7,87 %	6,55 %	7,56 %	6,28 %	7,31 %	7,13 %	14,58 %
Reproducibility limit R , mg/kg	909	1 782	1 911	258	15	194	130
Relative reproducibility limit R_{rel}	22,03 %	18,35 %	21,16 %	17,59 %	20,48 %	19,98 %	40,84 %
Repeatability standard deviation s_r , mg/kg	143	386	365	42	2	31	11
Relative repeatability standard deviation $s_{r,rel}$	3,45 %	3,97 %	4,04 %	2,89 %	3,34 %	3,16 %	3,38 %
Repeatability limit r , mg/kg	399	1 080	1 022	119	7	86	30
Relative repeatability limit r_{rel}	9,66 %	11,12 %	11,32 %	8,09 %	9,35 %	8,84 %	9,46 %
Relative Horwitz standard deviation	4,57 %	4,02 %	4,06 %	5,34 %	8,40 %	5,68 %	6,72 %
HorRat	1,72	1,63	1,86	1,18	0,87	1,26	2,17

Table B.7 — Validation data for potassium

Statistical parameter	Potassium						
	Infant formula soya based	Cheese	Chicken meat	Wheat flour	Apple juice	Lobster	Milk
Number of participating laboratories	13	13	13	13	13	13	13
Number of laboratories with quantitative values	13	13	13	13	13	13	13
Number of laboratories after elimination of outliers	13	13	13	11	13	13	13
Number of outlier laboratories	0	0	0	2	0	0	0
Mean value $\bar{x} \pm$ confidence interval, mg/kg	6 733 ±174	8 397 ±207	14 312 ±441	1 325 ±36	1 090 ±36	871 ±28	605 ±27
Reproducibility standard deviation s_R , mg/kg	370	462	878	72	70	55	53
Relative reproducibility standard deviation $s_{R,rel}$	5,49 %	5,50 %	6,14 %	5,46 %	6,43 %	6,27 %	8,72 %
Reproducibility limit R , mg/kg	1 035	1 293	2 459	203	196	153	148
Relative reproducibility limit R_{rel}	15,37 %	15,39 %	17,18 %	15,28 %	18,01 %	17,55 %	24,41 %
Repeatability standard deviation s_r , mg/kg	275	314	526	58	35	32	30
Relative repeatability standard deviation $s_{r,rel}$	4,08 %	3,74 %	3,68 %	4,39 %	3,26 %	3,63 %	5,01 %
Repeatability limit r , mg/kg	769	879	1 473	163	99	88	85
Relative repeatability limit r_{rel}	11,43 %	10,47 %	10,29 %	12,28 %	9,12 %	10,15 %	14,03 %
Relative Horwitz standard deviation	4,25 %	4,11 %	3,79 %	5,42 %	5,58 %	5,78 %	6,10 %
HorRat	1,29	1,34	1,62	1,01	1,15	1,09	1,43

Table B.8 — Validation data for sodium

Statistical parameter	Sodium						
	Infant formula soya based	Cheese	Chicken meat	Wheat flour	Apple juice	Lobster	Milk
Number of participating laboratories	13	13	13	13	13	13	13
Number of laboratories with quantitative values	13	13	13	11	12	13	13
Number of laboratories after elimination of outliers	13	12	13	10	12	13	13
Number of outlier laboratories	0	1	0	1	0	0	0
Mean value $\bar{x} \pm$ confidence interval, mg/kg	2 220 ±51	2 039 ±42	1 303 ±23	11 ±2	19 ±2	1 175 ±40	186 ±7
Reproducibility standard deviation s_R , mg/kg	108	84	52	3	3	78	12
Relative reproducibility standard deviation $s_{R,rel}$	4,89 %	4,11 %	3,97 %	26,32 %	16,94 %	6,60 %	6,60 %
Reproducibility limit R , mg/kg	304	235	145	8	9	217	34
Relative reproducibility limit R_{rel}	13,68 %	11,50 %	11,10 %	73,71 %	47,44 %	18,47 %	18,48 %
Repeatability standard deviation s_r , mg/kg	82	47	44	1	2	39	5
Relative repeatability standard deviation $s_{r,rel}$	3,67 %	2,30 %	3,40 %	8,15 %	8,17 %	3,31 %	2,71 %
Repeatability limit r , mg/kg	228	132	124	3	4	109	14
Relative repeatability limit r_{rel}	10,28 %	6,45 %	9,52 %	22,82 %	22,87 %	9,27 %	7,58 %
Relative Horwitz standard deviation	5,02 %	5,08 %	5,44 %	11,10 %	10,31 %	5,52 %	7,29 %
HorRat	0,97	0,81	0,73	2,37	1,64	1,19	0,91

Table B.9 — Validation data for sulfur

Statistical parameter	Sulfur						
	Infant formula soya based	Cheese	Chicken meat	Wheat flour	Apple juice	Lobster	Milk
Number of participating laboratories	13	13	13	13	13	13	13
Number of laboratories with quantitative values	11	11	11	11	10	11	11
Number of laboratories after elimination of outliers	11	11	10	11	10	11	11
Number of outlier laboratories	0	0	1	0	0	0	0
Mean value $\bar{x} \pm$ confidence interval, mg/kg	1 234 ±77	5 764 ±235	8 542 ±322	1 694 ±65	26 ±3	876 ±38	177 ±9
Reproducibility standard deviation s_R , mg/kg	132	420	523	114	5	66	16
Relative reproducibility standard deviation $s_{R,rel}$	10,71 %	7,29 %	6,12 %	6,75 %	17,85 %	7,54 %	8,94 %
Reproducibility limit R , mg/kg	370	1 176	1 464	320	13	185	44
Relative reproducibility limit R_{rel}	29,98 %	20,40 %	17,14 %	18,89 %	49,98 %	21,13 %	25,02 %
Repeatability standard deviation s_r , mg/kg	48	180	166	53	2	27	5
Relative repeatability standard deviation $s_{r,rel}$	3,86 %	3,12 %	1,94 %	3,15 %	7,76 %	3,13 %	2,59 %
Repeatability limit r , mg/kg	133	504	465	149	6	77	13
Relative repeatability limit r_{rel}	10,81 %	8,74 %	5,44 %	8,81 %	21,73 %	8,75 %	7,26 %
Relative Horwitz standard deviation	5,48 %	4,35 %	4,10 %	5,23 %	9,80 %	5,77 %	7,34 %
HorRat	1,95	1,68	1,49	1,29	1,82	1,31	1,22

Table B.10 — Validation data for zinc

Statistical parameter	Zinc						
	Infant formula soya based	Cheese	Chicken meat	Wheat flour	Apple juice^a	Lobster	Milk
Number of participating laboratories	13	13	13	13	13	13	13
Number of laboratories with quantitative values	13	13	13	13	8	13	13
Number of laboratories after elimination of outliers	13	12	13	13	8	13	12
Number of outlier laboratories	0	1	0	0	0	0	1
Mean value $\bar{x} \pm$ confidence interval, mg/kg	43,5 ±1,6	38,4 ±1,3	24,8 ±0,8	11,2 ±0,3	0,16 ±0,02	13,9 ±0,6	5,1 ±0,2
Reproducibility standard deviation s_R , mg/kg	3,0	2,5	1,7	0,5	0,03	1,1	0,3
Relative reproducibility standard deviation $s_{R,rel}$	6,89 %	6,61 %	6,66 %	4,88 %	19,80 %	7,90 %	6,52 %
Reproducibility limit R , mg/kg	8,4	7,1	4,6	1,5	0,09	3,1	0,9
Relative reproducibility limit R_{rel}	19,29 %	18,50 %	18,65 %	13,66 %	55,45 %	22,12 %	18,25 %
Repeatability standard deviation s_r , mg/kg	1,1	1,2	1,1	0,4	0,02	0,6	0,2
Relative repeatability standard deviation $s_{r,rel}$	2,60 %	3,25 %	4,25 %	3,87 %	9,83 %	4,63 %	2,96 %
Repeatability limit r , mg/kg	3,2	3,5	3,0	1,2	0,04	1,8	0,4
Relative repeatability limit r_{rel}	7,27 %	9,10 %	11,89 %	10,84 %	27,53 %	12,96 %	8,28 %
Relative Horwitz standard deviation	9,07 %	9,24 %	9,86 %	11,12 %	21,11 %	10,77 %	12,52 %
HorRat	0,76	0,72	0,68	0,44	0,94	0,73	0,52

^a The zinc content of the apple juice is in the range of the quantification limit.

Table B.11 — Data concerning trueness determined by the use of reference material

Reference material/ element	Mean mg/kg	Uncertainty^a of mean mg/kg	Reference value mg/kg	Uncertainty^a of reference value mg/kg
Dried fresh cheese				
copper	14,66	0,48	14,49	0,31
zinc	38,4	1,3	37,56	1,13
NRCC LUTS-1				
calcium	183	5	203	33
copper	16,40	0,50	15,9	1,2
iron	12,1	0,5	11,6	0,9
magnesium	85	4	89,5	4,1
manganese	1,20	0,05	1,20	0,13
potassium	871	28	948	72
zinc	13,9	0,6	12,4	0,8

^a 95% confidence interval.

Bibliography

- [1] ISO 5725 (all parts), *Accuracy (trueness and precision) of measurement methods and results*
- [2] ISO 21033:2016, *Animal and vegetable fats and oils - Determination of trace elements by inductively coupled plasma optical emission spectroscopy (ICP-OES)*
- [2] Official collection of analysis methods according to § 64 of the German Food and Feed Code (LFGB), L 00.00-144: Bestimmung der Mineralstoffe Calcium, Kalium, Magnesium, Natrium, Phosphor und Schwefel sowie der Spurenelemente Eisen, Kupfer, Mangan und Zink in Lebensmitteln mit der optischen Emissionsspektrometrie mit induktiv gekoppeltem Plasma (ICP-OES): 01-2013
- [3] NOLTE J. *ICP Emission Spectrometry: A Practical Guide*. Wiley-VCH, 2003
- [4] NOLTE J. *Fehlerfrei durch die ICP Emissionsspektrometrie*. Wiley-VCH, 2012
- [5] EN ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories (ISO/IEC 17025:2005)*

This page deliberately left blank

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Copyright in BSI publications

All the content in BSI publications, including British Standards, is the property of and copyrighted by BSI or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use.

Save for the provisions below, you may not transfer, share or disseminate any portion of the standard to any other person. You may not adapt, distribute, commercially exploit, or publicly display the standard or any portion thereof in any manner whatsoever without BSI's prior written consent.

Storing and using standards

Standards purchased in soft copy format:

- A British Standard purchased in soft copy format is licensed to a sole named user for personal or internal company use only.
- The standard may be stored on more than 1 device provided that it is accessible by the sole named user only and that only 1 copy is accessed at any one time.
- A single paper copy may be printed for personal or internal company use only.

Standards purchased in hard copy format:

- A British Standard purchased in hard copy format is for personal or internal company use only.
- It may not be further reproduced – in any format – to create an additional copy. This includes scanning of the document.

If you need more than 1 copy of the document, or if you wish to share the document on an internal network, you can save money by choosing a subscription product (see 'Subscriptions').

Reproducing extracts

For permission to reproduce content from BSI publications contact the BSI Copyright & Licensing team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email subscriptions@bsigroup.com.

Rewrites

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Useful Contacts

Customer Services

Tel: +44 345 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 345 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK