

BS EN 60195:2016



BSI Standards Publication

Method of measurement of current noise generated in fixed resistors

National foreword

This British Standard is the UK implementation of EN 60195:2016. It is identical to IEC 60195:2016. It supersedes BS 4119:1967 which is withdrawn.

The UK participation in its preparation was entrusted to Technical Committee EPL/40X, Capacitors and resistors for electronic equipment.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2016.

Published by BSI Standards Limited 2016

ISBN 978 0 580 86205 2

ICS 31.040.10

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 31 July 2016.

Amendments/corrigenda issued since publication

| Date | Text affected |
|------|---------------|
|------|---------------|

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 60195

July 2016

ICS 31.040.10

English Version

**Method of measurement of current noise generated in fixed resistors
(IEC 60195:2016)**

Méthode pour la mesure du bruit produit en charge par les résistances fixes
(IEC 60195:2016)

Messverfahren für das Stromrauschen in Festwiderständen
(IEC 60195:2016)

This European Standard was approved by CENELEC on 2016-05-12. CENELEC members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration.

Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CENELEC member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CENELEC member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CENELEC members are the national electrotechnical committees of Austria, Belgium, Bulgaria, Croatia, Cyprus, the Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, the Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



European Committee for Electrotechnical Standardization
Comité Européen de Normalisation Electrotechnique
Europäisches Komitee für Elektrotechnische Normung

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

European foreword

The text of document 40/2431/FDIS, future edition 2 of IEC 60195, prepared by IEC/TC 40 "Capacitors and resistors for electronic equipment" was submitted to the IEC-CENELEC parallel vote and approved by CENELEC as EN 60195:2016.

The following dates are fixed:

- latest date by which the document has to be implemented at national level by publication of an identical national standard or by endorsement (dop) 2017-02-12
- latest date by which the national standards conflicting with the document have to be withdrawn (dow) 2019-05-12

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CENELEC [and/or CEN] shall not be held responsible for identifying any or all such patent rights.

Endorsement notice

The text of the International Standard IEC 60195:2016 was approved by CENELEC as a European Standard without any modification.

In the official version, for Bibliography, the following notes have to be added for the standards indicated:

IEC 60027 (series)

NOTE Harmonized as EN 60027 (series).

Annex ZA (normative)

Normative references to international publications with their corresponding European publications

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

NOTE 1 When an International Publication has been modified by common modifications, indicated by (mod), the relevant EN/HD applies.

NOTE 2 Up-to-date information on the latest versions of the European Standards listed in this annex is available here: www.cenelec.eu.

| <u>Publication</u> | <u>Year</u> | <u>Title</u> | <u>EN/HD</u> | <u>Year</u> |
|--------------------|-------------|---|--------------|-------------|
| IEC 60068-1 | 2013 | Environmental testing -- Part 1: General and guidance | EN 60068-1 | 2014 |

CONTENTS

| | |
|--|----|
| FOREWORD..... | 4 |
| 1 Scope..... | 6 |
| 2 Normative references..... | 6 |
| 3 Terms and definitions | 6 |
| 4 Method of measurement | 7 |
| 4.1 Noise basics | 7 |
| 4.1.1 Noise | 7 |
| 4.1.2 Thermal noise | 8 |
| 4.1.3 Current noise | 8 |
| 4.2 Measurement principle | 9 |
| 4.3 Measurement system | 10 |
| 4.3.1 Proposal of a suitable measuring system | 10 |
| 4.3.2 Alternative measuring systems | 11 |
| 4.4 Measurement system requirements | 11 |
| 4.4.1 Input circuit..... | 11 |
| 4.4.2 Isolation resistor R_M | 12 |
| 4.4.3 DC voltage source | 12 |
| 4.4.4 DC electronic voltmeter | 12 |
| 4.4.5 Calibration resistor R_{Cal} | 12 |
| 4.4.6 Calibration source | 13 |
| 4.4.7 Determination of the calibration voltage | 13 |
| 4.4.8 AC band-pass amplifier | 15 |
| 4.4.9 AC r.m.s. meter..... | 16 |
| 4.4.10 Test fixture | 16 |
| 4.5 Verification of the measuring system | 17 |
| 4.5.1 Performance check by measurement of instrument and thermal noise | 17 |
| 4.5.2 Performance check by comparison of repeated measurements..... | 17 |
| 5 Measurement procedure | 18 |
| 5.1 Ambient conditions | 18 |
| 5.2 Preparation of specimen..... | 18 |
| 5.3 Procedure | 18 |
| 5.3.1 General | 18 |
| 5.3.2 Calibration | 18 |
| 5.3.3 Measurement of system noise S | 18 |
| 5.3.4 Measurement of total noise T | 19 |
| 5.4 Precautions..... | 22 |
| 6 Evaluation of measurement results..... | 22 |
| 6.1 Term for the contribution of system noise | 22 |
| 6.2 Determination of the current-noise index A_1 | 24 |
| 6.3 Determination of the current-noise voltage ratio CNR_U | 25 |
| 6.4 Accuracy..... | 26 |
| 6.5 Requirements..... | 26 |
| 7 Information to be given in the relevant component specification | 26 |
| Annex A (informative) Letter symbols and abbreviations | 27 |
| A.1 Letter symbols | 27 |

A.2 Abbreviations27

Annex X (informative) Cross-reference for references to the prior revision of this standard28

Bibliography29

Figure 1 – Block schematic of a suitable measuring system 11

Figure 2 – Typical transfer function of the band-pass amplifier 16

Figure 3 – Contribution of system noise, $f(T - S)$ 23

Table 1 – Permissible limits of system noise..... 17

Table 2 – Recommended operating conditions (1 of 2).....20

Table 3 – Numeric values of the contribution of system noise, $f(T - S)$24

Table X.1 – Cross reference for references to the 1st edition of this standard28

INTERNATIONAL ELECTROTECHNICAL COMMISSION

**METHOD OF MEASUREMENT OF CURRENT
NOISE GENERATED IN FIXED RESISTORS**

FOREWORD

- 1) The International Electrotechnical Commission (IEC) is a worldwide organization for standardization comprising all national electrotechnical committees (IEC National Committees). The object of IEC is to promote international co-operation on all questions concerning standardization in the electrical and electronic fields. To this end and in addition to other activities, IEC publishes International Standards, Technical Specifications, Technical Reports, Publicly Available Specifications (PAS) and Guides (hereafter referred to as "IEC Publication(s)"). Their preparation is entrusted to technical committees; any IEC National Committee interested in the subject dealt with may participate in this preparatory work. International, governmental and non-governmental organizations liaising with the IEC also participate in this preparation. IEC collaborates closely with the International Organization for Standardization (ISO) in accordance with conditions determined by agreement between the two organizations.
- 2) The formal decisions or agreements of IEC on technical matters express, as nearly as possible, an international consensus of opinion on the relevant subjects since each technical committee has representation from all interested IEC National Committees.
- 3) IEC Publications have the form of recommendations for international use and are accepted by IEC National Committees in that sense. While all reasonable efforts are made to ensure that the technical content of IEC Publications is accurate, IEC cannot be held responsible for the way in which they are used or for any misinterpretation by any end user.
- 4) In order to promote international uniformity, IEC National Committees undertake to apply IEC Publications transparently to the maximum extent possible in their national and regional publications. Any divergence between any IEC Publication and the corresponding national or regional publication shall be clearly indicated in the latter.
- 5) IEC itself does not provide any attestation of conformity. Independent certification bodies provide conformity assessment services and, in some areas, access to IEC marks of conformity. IEC is not responsible for any services carried out by independent certification bodies.
- 6) All users should ensure that they have the latest edition of this publication.
- 7) No liability shall attach to IEC or its directors, employees, servants or agents including individual experts and members of its technical committees and IEC National Committees for any personal injury, property damage or other damage of any nature whatsoever, whether direct or indirect, or for costs (including legal fees) and expenses arising out of the publication, use of, or reliance upon, this IEC Publication or any other IEC Publications.
- 8) Attention is drawn to the Normative references cited in this publication. Use of the referenced publications is indispensable for the correct application of this publication.
- 9) Attention is drawn to the possibility that some of the elements of this IEC Publication may be the subject of patent rights. IEC shall not be held responsible for identifying any or all such patent rights.

International Standard IEC 60195 has been prepared by IEC technical committee 40: Capacitors and resistors for electronic equipment.

This second edition cancels and replaces the first edition published in 1965 and constitutes a technical revision.

This edition includes the following significant technical changes with respect to the previous edition:

- harmonization of the allocation of isolation resistors R_M in the recommended operating conditions given in Table 2;
- correction of erroneous numeric values of the contribution of system noise, $f(T - S)$ in Table 3;
- addition of advice on the prescription of requirements in a relevant component specification;
- addition of a set of recommended measuring conditions for specimens with a rated dissipation of less than 100 mW;

- complete editorial revision.

The text of this standard is based on the following documents:

| FDIS | Report on voting |
|--------------|------------------|
| 40/2431/FDIS | 40/2458/RVD |

Full information on the voting for the approval of this standard can be found in the report on voting indicated in the above table.

This publication has been drafted in accordance with the ISO/IEC Directives, Part 2.

The committee has decided that the contents of this publication will remain unchanged until the stability date indicated on the IEC website under "<http://webstore.iec.ch>" in the data related to the specific publication. At this date, the publication will be

- reconfirmed,
- withdrawn,
- replaced by a revised edition, or
- amended.

METHOD OF MEASUREMENT OF CURRENT NOISE GENERATED IN FIXED RESISTORS

1 Scope

This International Standard specifies a method of measurement and associated test conditions to assess the "noisiness", or magnitude of current noise, generated in fixed resistors of any given type. The method applies to all classes of fixed resistors. The aim is to provide comparable results for the determination of the suitability of resistors for use in electronic circuits having critical noise requirements.

The current noise in resistive materials reflects the granular structure of the resistive material. For some resistor technologies utilizing homogenous layers it is regarded as providing an indication of defects, which are considered as a root cause for abnormal ageing of the component under the influence of temperature and time.

The method described in this International Standard is not a general specification requirement and therefore is applied if prescribed by a relevant component specification, or, if agreed between a customer and a manufacturer.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

IEC 60068-1:2013, *Environmental testing – Part 1: General and guidance*

3 Terms and definitions

For the purposes of this document the following terms and definitions apply.

3.1

current-noise

combination of all random fluctuations of current flow in a resistor which are not attributed to thermal agitation of the charge carriers (thermal noise) and which depend on the applied direct current

3.2

current-noise index

A_1

logarithmic index of the ratio of the open circuit r.m.s. current-noise voltage in a frequency decade, in μV , over the d.c. voltage applied under test, in V, used to express the "noisiness" of an individual resistor

Note 1 to entry: The current-noise index is expressed in dB. The ratio between μV and V is not considered in this index, leading to its value being 120 dB less than the mathematical current-noise index A_1' . This practical index follows the history of prior revisions of this method.

3.3 mathematical current-noise index

A_1'

logarithmic index of the ratio of the open circuit r.m.s. current-noise voltage in a frequency decade over the d.c. voltage applied under test, established in consistent units and their multiples

Note 1 to entry: The mathematical current-noise index is expressed in dB. This index has been introduced for the mathematical derivation of the considered parameters.

3.4 current-noise voltage ratio

CNR_U

ratio of the open circuit r.m.s. current-noise voltage in a frequency decade over the d.c. voltage applied under test, established in $\mu V/V$, used to express the "noisiness" of an individual resistor

3.5 flicker noise pink noise

random fluctuation present in most electronic devices and typically related to internal properties of the respective device, which depends on direct current and has a power spectral density inversely proportional to the frequency

3.6 noise

random fluctuation in an electrical signal having instantaneous amplitude values which, due to their distribution in a random manner, can only be predicted in terms of probability statements

3.7 shot noise

random fluctuation in electric current due to the flowing current consisting of discrete charges, which is independent of temperature and has nearly constant power spectral density throughout the frequency spectrum

3.8 thermal noise

random fluctuation generated by the thermal agitation of the charge carriers (usually the electrons) inside an electrical conductor at equilibrium, which is independent of any applied voltage and has nearly constant power spectral density throughout the frequency spectrum

Note 1 to entry: Thermal noise is also referred to as Johnson noise or as Nyquist noise.

4 Method of measurement

4.1 Noise basics

4.1.1 Noise

Noise appears as a spontaneous fluctuating voltage $e_n(t)$ with instantaneous amplitude values.

Noise voltage is a statistically independent random variable, where for most kinds of noise the frequency distribution of amplitudes follows a Gaussian distribution curve. Therefore noise voltage cannot be predicted except in terms of probability statements.

Usually the characteristic of principal interest is not the instantaneous amplitude value but the "time-averaged" value.

The measurement of amplitude commonly used and adopted for this International Standard is the effective (r.m.s.) voltage E_n observed in a particular frequency pass-band.

4.1.2 Thermal noise

The thermal noise of a resistor is a fluctuating voltage caused by the random motion of thermally agitated charges, which is present in all resistors. The root mean-square value of the fluctuating voltage appearing at the open-circuit terminals of a resistor, which would be indicated by the measuring system, may be calculated using Nyquist's equation:

$$E_{th} = \sqrt{e_{th}^2} = \sqrt{4 \cdot k \cdot T \cdot R \cdot \Delta f}$$

where

E_{th} is the effective voltage (r.m.s. voltage) of the thermal noise in a given bandwidth;

e_{th} is the momentary voltage of the thermal noise in a given bandwidth;

k is the Boltzmann constant, $k \approx 1,38 \times 10^{-23}$ J/K;

T is the absolute temperature;

R is the resistance;

Δf is the bandwidth of the effective band-pass filter of the measuring system.

The presence of thermal noise cannot be ignored because the thermal noise of the resistor under test is frequently a major source of interference in the measurement.

4.1.3 Current noise

The presence of direct current in a fixed resistor causes an increase in the observed total noise above the level attributed to thermal noise. Regardless of its originating nature, this excess noise is referred to as current noise.

$$E_t^2 = E_{th}^2 + E_c^2$$

where

E_t is the effective voltage of the total noise in a given bandwidth;

E_{th} is the effective voltage of the thermal noise in a given bandwidth;

E_c is the effective voltage of the current noise in a given bandwidth.

Hence, the current noise is the geometric difference between the total noise and the thermal noise

$$E_c^2 = E_t^2 - E_{th}^2$$

The effective current-noise voltage per 1 Hz bandwidth is substantially inversely proportional to frequency

$$[e(f)]^2 \sim \frac{I^2}{f}$$

where

$e(f)$ is the momentary voltage of the current noise as a function of frequency;

I is the d.c. current passing through the resistor;

f is the frequency for which the current noise voltage is considered.

The effective current noise voltage for a given bandwidth is calculated by integrating the current noise voltage over the frequency band

$$\begin{aligned} E_c^2 &= \int_{f_1}^{f_2} [e(f)]^2 df \\ &\sim \int_{f_1}^{f_2} \frac{I^2}{f} df \\ &\sim I^2 \ln\left(\frac{f_2}{f_1}\right) \end{aligned}$$

where

E_c is the effective voltage of the current noise in a given bandwidth;

f_1 is the lower cut-off frequency of the ideal band-pass;

f_2 is the upper cut-off frequency of the ideal band-pass.

If the mean-square voltage is inversely proportional to frequency, then ideal rectangular pass-bands having equal ratios of upper to lower band-pass limits transmit equal amounts of noise voltage from a given noise source.

A resistor exhibiting current noise may be represented as a noise source having a zero-impedance current-noise voltage generator connected in series with an independent thermal-noise voltage generator and with a noise-free resistor.

4.2 Measurement principle

The current noise voltage E_c is, in general, closely proportional to the applied d.c. test voltage U_T . It is recommended, however, to apply a harmonized set of operating conditions in order to ensure the most comparable measurements for all resistors.

Table 2 gives a set of operating conditions recommended for the testing of resistors with resistances in the range of 100 Ω to 22 M Ω . The values given therein also avoid overloading the specimen and the input circuit.

The frequency dependence of noise voltages requires the prescription of a frequency pass-band to be used in this measurement, which is an ideal rectangular pass-band of one frequency decade, geometrically centered at 1 000 Hz.

The measurement results in the mathematical current noise index in a frequency decade, A_1' , as follows:

$$A_1' = 20 \lg \left(\frac{E_c'}{U_T} \right) \text{ dB}$$

where

E_c' is the effective open circuit current-noise voltage in a frequency decade, given in V;

U_T is the d.c. voltage applied to the resistor under test, given in V.

The typical magnitude of the current-noise voltage being in the microvolt range rather than in a volt range is reflected in the prevalent current noise index in a frequency decade, A_1 ,

$$A_1 = 20 \lg \left(\frac{E_C}{U_T} \right) \text{ dB}$$

where

E_C is the effective open circuit current-noise voltage in a frequency decade, given in μV ;

U_T is the d.c. voltage applied to the resistor under test, given in V.

The ratio between μV and V, which results in an offset of 120 dB, is neglected in the traditional definition of the current noise index A_1 , hence the following relationship applies:

$$A_1 = A_1' - 120 \text{ dB} .$$

Since the current-noise power spectrum approximates to a $1/f$ frequency characteristic, the index and the ratio provides an estimate of current noise in any frequency decade.

4.3 Measurement system


4.3.1 Proposal of a suitable measuring system

Figure 1 shows a block schematic of a suitable measuring system.

A three-position switch may be used to access any of the three modes of operation normally followed in the measurement procedure:

- calibration;
- measurement of system noise;
- measurement of total noise.

The input circuit consists of the resistor under test R_T , the isolation resistor R_M and the calibration resistor R_{Cal} , where the isolation resistor R_M is required to reduce the shunting effect of the d.c. supply system on the noise generated in the resistor under test.



IEC

Key

| | | | |
|-------------|--|-------------|--------------|
| P | DC voltage source | | |
| G | Calibration source, $f = 1 \text{ kHz}$ | | |
| S | Three position switch | Position 1: | Calibration |
| | | Position 2: | System noise |
| | | Position 3: | Total noise |
| R_M | Isolation resistor | | |
| R_{Cal} | Calibration resistor, $R_{Cal} = 1 \Omega$ | | |
| R_T | Resistor under test | | |
| U_T | Test voltage, d.c. | | |
| BPA | Band-pass amplifier with adjustable gain | | |
| $U_{N rms}$ | Noise voltage, a.c. r.m.s. | | |

Figure 1 – Block schematic of a suitable measuring system

The following content of this International Standard refers to this suitable measuring system, unless otherwise specified.

4.3.2 Alternative measuring systems

The proposal of a measuring system in 4.3.1 intends to unify the test and measurement procedures used for the assessment of the current noise generated in fixed resistors. This system, however, is not necessarily the only system which can be used, except when specifically designated as referee or reference methods.

The provider and user of any alternative measuring system shall demonstrate that such system will give results equivalent to those obtained by the proposed system.

4.4 Measurement system requirements**4.4.1 Input circuit**

The input impedance of the measurement system is influenced by the impedance of the d.c. electronic voltmeter, which is in parallel with the isolation resistor R_M and also with the resistor under test, and thereby attenuates the noise signal generated in the specimen.

The input impedance of the d.c. electronic voltmeter shall meet the impedance requirement given in 4.4.4 in order to avoid any detrimental influence on the measurement.

4.4.2 Isolation resistor R_M

A number of current noise free isolation resistors R_M will be needed to cover the range of resistance values of the specimen, which may be switched into the circuit as required. The isolation resistor shall be current noise-free (for example, good quality wirewound resistors). Each isolation resistor shall have a rated dissipation of at least 1 W and the resistance tolerance shall be ± 1 %.

At least four isolation resistors R_M are required if the range of specimen resistance extends from 100 Ω to 22 M Ω . Examples for suitable values of R_M are 1 k Ω ; 10 k Ω ; 100 k Ω and 1 M Ω . These values are used for establishing the test conditions in Table 2.

4.4.3 DC voltage source

The d.c. voltage source shall be capable of supplying a suitable range of voltages, which depends on the specimen resistance R_T , on the required test voltage U_T , and on the isolation resistor R_M . The adjusted d.c. test voltage shall be maintained sufficiently stable throughout a measurement.

Table 2 provides recommended operating conditions for the specimen resistance R_T in the range from 100 Ω to 22 M Ω , leading to test voltages U_T in the range from 2,2 V to 250 V. In order to achieve this, the d.c. voltage source is required to provide a voltage adjustable in the range of 14 V through 500 V.

There may be some hum and noise interference introduced by the d.c. voltage source when it drives a current through the resistor under test. The influence of this on the observed noise index shall not exceed 0,5 dB, when the connected test resistor is known not to generate any current noise itself (e.g. a good quality wirewound or metal foil resistor).

4.4.4 DC electronic voltmeter

The voltmeter used for measuring the d.c. test voltage U_T shall have a constant impedance of at least 4 M Ω in the frequency range from 0 Hz to 1 600 Hz.

The meter, in conjunction with a step attenuator, shall be capable of indicating the required d.c. test voltages with an accuracy of ± 3 %. The time constant shall be less than 0,5 s.

The meter shall support the reading of the d.c. test voltage U_T in volt, and the reading of the d.c. test voltage index D in dB, which is determined by

$$D = 20 \lg \left(\frac{U_T}{1 \text{ V}} \right) \text{ dB}$$

There may be some interference introduced by the voltmeter when it is connected to the input circuit. The influence of this on the observed noise index shall not exceed 0,2 dB.

4.4.5 Calibration resistor R_{Cal}

The calibration resistor R_{Cal} shall meet the following specification details:

$$R_{Cal} = 1,00 \Omega$$

$$P_r \geq 0,5 \text{ W}$$

The calibration resistor shall be selected for the lowest possible generation of current noise (e.g. a good quality wirewound or metal foil resistor).

4.4.6 Calibration source

The calibration source shall be a stable sine-wave generator with a fixed frequency within the range of 980 Hz to 1 020 Hz. Its output shall supply a voltage across the calibration resistor R_{Cal} , which is adjustable within a range from 0,6 mV to 0,7 mV, where the actual required calibration voltage is determined in 4.4.7. The stability of the adjusted calibration voltage shall be better than $\pm 2\%$.

The calibration source is connected to the measuring system only in calibration mode.

4.4.7 Determination of the calibration voltage

The calibration voltage U_{Cal} is determined to produce a noise meter reading equal to that produced by a current-noise voltage having an r.m.s. value of 1 mV in a frequency decade.

In 4.1.3 it has been shown that the effective current noise voltage depends of the d.c. current and of the cut-off frequencies of the ideal band-pass like

$$E_c^2 \sim I^2 \ln\left(\frac{f_2}{f_1}\right)$$

where

E_c is the effective voltage of the current noise in a given bandwidth;

I is the d.c. current passing through the resistor;

f_1 is the lower cut-off frequency of the ideal band-pass;

f_2 is the upper cut-off frequency of the ideal band-pass.

For a frequency decade and an ideal band-pass the relationship of the two cut-off frequencies is

$$f_2 = 10 f_1$$

For the considered reference condition with

$$E_c = 1 \text{ mV}$$

the above relationship is

$$(1 \text{ mV})^2 \sim I^2 \ln(10)$$

where

I is the d.c. current passing through the resistor.

For this method an ideal band-pass filter of 1 kHz bandwidth, geometrically centered at 1 kHz shall be used, with a lower cut-off frequency $f_1 = 618$ Hz and an upper cut-off frequency $f_2 = 1\,618$ Hz, see 4.4.8.

For this condition applies

$$E_c = U_{\text{Cal}}$$

and hence the above relationship is

$$(U_{\text{Cal}})^2 \sim I^2 \ln\left(\frac{f_2}{f_1}\right)$$

where

U_{Cal} is the calibration voltage required to achieve a 1 mV per frequency decade reading;

I is the d.c. current passing through the resistor;

f_1 is the lower cut-off frequency of the ideal band-pass;

f_2 is the upper cut-off frequency of the ideal band-pass.

Dividing $(U_{\text{Cal}})^2$ by $(1 \text{ mV})^2$ results in

$$\frac{(U_{\text{Cal}})^2}{(1 \text{ mV})^2} = \frac{\ln\left(\frac{f_2}{f_1}\right)}{\ln(10)}$$

and finally in the determination of the calibration voltage as

$$U_{\text{Cal}} = \sqrt{\frac{\ln\left(\frac{f_2}{f_1}\right)}{\ln(10)}} \times 1 \text{ mV}$$

For practical use this equation may be simplified to

$$U_{\text{Cal}} = \sqrt{A} \cdot 0,659 \text{ mV}$$

where

A is a non-dimensional value representing the area under the pass-band curve

$$A = \ln\frac{f_2}{f_1}$$

where

f_1 is the lower cut-off frequency of the ideal band-pass;

f_2 is the upper cut-off frequency of the ideal band-pass.

For the prescribed ideal band-pass filter of 1 kHz bandwidth, geometrically centered at 1 kHz, the calculation results in $A_i = 0,962$.

Hence, for this case the required calibration voltage is

$$U_{\text{Cal}} = \sqrt{0,962} \times 0,659 \text{ mV} = 0,646 \text{ mV}$$

For any particular non-ideal band-pass filter, value A can be computed as follows.

- a) The voltage gain is measured throughout the pass-band of the system versus frequency, where the voltage gain is the ratio of the output voltage, as indicated by the output meter, to the input voltage, applied across the terminals of the calibration resistor.
- b) The power gain is calculated by squaring of the voltage gain for each frequency.
- c) Each power gain value is divided by its respective frequency and plotted against frequency in a linear diagram.
- d) The area under the resulting curve is planimetrically measured to give the value A for the respective non-ideal band-pass filter, where the accuracy of this determination shall be within $\pm 2,5$ % of the result.

Where there is the result of a numerical simulation available of the gain over frequency of a designed band-pass filter, with the gain typically given in dB, the procedure described above can be executed numerically, e.g. using a calculation spreadsheet.

The resulting value A should be close to the result for an ideal band-pass, $A_i = 0,962$.

NOTE The numerical determination of the area under the pass-band shown in Figure 2 results in a value $A = 0,959$, which leads to a required calibration voltage $U_{\text{Cal}} = 0,645$ mV.

The calibration source shall be adjusted to provide the r.m.s. voltage U_{Cal} at the terminals of the calibration resistor R_{Cal} with a relative accuracy of better than ± 2 %.

4.4.8 AC band-pass amplifier


The amplifier gain shall be sufficient to measure circuit noise with the input terminals shorted and with the adjustable gain control, described below, set near its minimum gain position.

In a verification with $R_M \geq 100$ k Ω and no d.c. current present, the circuit noise shall be no greater than that equivalent to the thermal noise of a 6,2 k Ω resistor. Therefore, the increase in the output reading when the short is replaced by a resistor of 6,2 k Ω with a relative tolerance of ± 5 % shall be at least 3 dB with the gain control setting kept unchanged.

The amplifier shall be capable of measuring input signals up to 650 μ V. This signal amplitude gives a scale reading of approximately 60 dB when the system is calibrated.

A continuously adjustable gain control shall be provided for maintaining a fixed overall system gain which would otherwise vary with input conditions listed in Table 2. The necessary gain control range is approximately 33 dB.

The pass-band shall be flat, shall have a fixed half-power pass-band of approximately 1 000 Hz within the limits ± 50 Hz and shall be geometrically centered at 1 000 Hz ± 50 Hz. Ripple in the flat top of the pass-band shall not exceed $\pm 0,2$ dB. Figure 2 shows the transfer function of a band-pass meeting these requirements.



IEC

Figure 2 – Typical transfer function of the band-pass amplifier

These requirements shall be satisfied for all measurement conditions listed in Table 2. Neither the area under the pass-band, A , as determined in 4.4.7, nor the half-power pass-band shall vary with respect to measurement condition by more than $\pm 4\%$ for any recommended measurement condition. Compliance at $100\ \Omega$ and at $22\ \text{M}\Omega$ is considered sufficient.

The a.c. amplifier shall respond to noise signals without introducing a significant error due to clipping. This requires the dynamic range to extend at least 10 dB beyond the indicated a.c. r.m.s. value.

4.4.9 AC r.m.s. meter

The a.c. measuring system shall be calibrated in dB from $-20\ \text{dB}$ to at least $+60\ \text{dB}$ with $0\ \text{dB}$ being $1\ \mu\text{V}$ in a frequency decade. The accuracy of the a.c. r.m.s. meter shall be $\pm 0,4\ \text{dB}$. The time constant shall be in the range of $0,8\ \text{s}$ to $1,5\ \text{s}$.

4.4.10 Test fixture

The test fixture for the resistor under test, R_T , shall be capable of providing a safe electrical connection and sufficient shielding from any external fields.

The lead-to-lead and lead-to-ground capacitances of the resistor under test in its test fixture and of the leads to the input of the band-pass amplifier shall be minimized, e.g. by the use of short leads, adequate spacing and careful mounting.

Good shielding practice shall be adopted in the construction of the measurement system. The input circuit operates at extremely low signal levels, which makes it necessary that all parts and leads in the input circuit be very well shielded. Components carrying large signals should not be located near the input circuit.

4.5 Verification of the measuring system

4.5.1 Performance check by measurement of instrument and thermal noise

It is recommended to verify the performance of the measurement system by checking the level of system noise, including thermal noise, without involving any specimen.

For a measurement system as proposed in 4.3, the following procedure should be applied:

- a) turn the function switch to “calibration” and short-circuit the terminals for the specimen R_T ;
- b) adjust the gain of the band-pass amplifier to the calibrate line on the a.c. r.m.s. meter;

NOTE The calibration line typically is a line centred on the a.c. r.m.s. meter scale. With the measuring system set in calibration mode the meter is connected without an attenuator network.

- c) turn the function switch to “system noise” and read the noise index S_k ;
- d) remove the short circuit from the terminals for the specimen R_T ;
- e) read the noise index S_o for each isolation resistor R_M .

The readings of the noise index should fall within the limits given in Table 1, unless other recommendations are given for a specific measuring system.

Table 1 – Permissible limits of system noise

| R_M | R_T | Permissible limits of system noise reading dB |
|----------------|----------------------|--|
| Any | Short circuit | $S_k \leq -5$ dB |
| 1 k Ω | None, terminals open | -9 dB $\leq S_o \leq -4,5$ dB |
| 10 k Ω | None, terminals open | -5 dB $\leq S_o \leq -2$ dB |
| 100 k Ω | None, terminals open | 5 dB $\leq S_o \leq 7,5$ dB |
| 1 M Ω | None, terminals open | 14 dB $\leq S_o \leq 18$ dB |

The first two readings are essentially measurements of the noise in the amplifier, while the last two readings are essentially measurements of thermal noise of R_M in the pass-band of the instrument. The third reading is influenced by both factors.

There may be different performance check procedures and permissible limits prescribed for practical realizations of the measurement system proposed in 4.3.1, or for alternative measuring systems as suggested in 4.3.2.

4.5.2 Performance check by comparison of repeated measurements

It is recommended to verify the performance of the measurement system by checking the current noise of specific resistor specimens after repeated measurements.

A practical means of monitoring the stability of the measurement system is to keep a record of the measurements made on a set of specific control resistors, where it is desirable for the set of control resistors to consist of different types of resistors and to represent a large range of resistance and current-noise values.

Plotting the data against time in the form of a control chart for each specimen is suggested as a simple and effective means for detecting any irregularity within the measurement system.

5 Measurement procedure

5.1 Ambient conditions

The measurement shall preferably be made under standard atmosphere for referee measurements and test as given in IEC 60068-1:2013, 4.2.

- Temperature: 23 °C ± 2 °C
- Relative humidity: 45 % to 55 %
- Air pressure: 86 kPa to 106 kPa

A relevant specification may prescribe other ambient conditions for this measurement.

NOTE The generally applied standard atmospheric conditions for testing with their wider permissible temperature range are not recommendable for this test due to the influence of temperature on the measurement, e.g. by means of thermal noise.

5.2 Preparation of specimen

The specimen shall be stored at the ambient conditions prescribed in 5.1 for at least 24 h before a measurement is made.

5.3 Procedure

5.3.1 General

The measurement system shall be stored at the ambient conditions prescribed in 5.1 for at least 24 h before a measurement is made.

The complete operation consists of three consecutive steps:

- a) calibration;
- b) measurement of system noise;
- c) simultaneous measurement of total noise and of the d.c. voltage across the specimen.

Considerable saving of time can be achieved when groups of similar resistors are to be measured by taking advantage of the stability of the particular test equipment. The intervals with which the calibration and system noise measurements shall be repeated will be dependent on the stability of the equipment and on the accuracy required.

5.3.2 Calibration

The resistor to be tested, R_T , shall be inserted into the test fixture, and the appropriate isolation resistor R_M shall be switched into the circuit. The allocation given in Table 2 shall apply, unless other provisions are made by the relevant specification.

With the measuring system set into configuration for calibration, the isolation resistor R_M is connected to ground instead of to the d.c. voltage source, and the 1 000 Hz calibration voltage is connected to the calibration resistor R_{Cal} .

The gain control of the band-pass amplifier shall be adjusted so that the noise voltage meter indicates +60 dB, or its equivalent 1 mV, in a frequency decade.

5.3.3 Measurement of system noise S

The connection of the specimen R_T and of the isolation resistor R_M shall remain as set in 5.3.2.

With the measuring system configured to measure system noise, the isolation resistor R_M is connected to ground instead of to the d.c. voltage source, and the calibration resistor R_{Cal} is disconnected from the calibration voltage.

The system noise index S is read on the noise voltage meter after a minimum delay of 5 s, allowing the meter to reach a representative mean value.

5.3.4 Measurement of total noise T

The connection of the specimen R_T and of the isolation resistor R_M shall remain as set in 5.3.2.

With the measuring system configured to measure total noise, the isolation resistor R_M is connected to the d.c. voltage source in order to pass a current through the resistor under test, R_T . The calibration resistor R_{Cal} is disconnected from the calibration voltage.

The d.c. voltage U_T shall be set according to the prescription given in Table 2, unless other provisions are made by the relevant specification, e.g. in consideration of the dissipation rating of a particular style of resistors. For values of R_T not contained in Table 2, it is suitable to apply the parameters for the next lower given value.

The d.c. test voltage index D , is read on the d.c. voltmeter, and the total noise index T is read on the noise voltage meter after a minimum delay of 5 s, allowing the meter to reach a representative mean value.

Table 2 – Recommended operating conditions (1 of 2)

| R_T | R_M | $P_r \geq 0,5 \text{ W}$ | | | $0,5 \text{ W} > P_r \geq 0,1 \text{ W}$ | | | $0,1 \text{ W} > P_r$ | | |
|----------------|----------------|--------------------------|-----------|-------------|--|-----------|-------------|-----------------------|-----------|-------------|
| | | U_T V | D dB | P_T mW | U_T V | D dB | P_T mW | U_T V | D dB | P_T mW |
| 100 Ω | 1,0 k Ω | 3,2 | 10,0 | 100 | 3,2 | 10,0 | 100 | 2,2 | 7,0 | 50 |
| 120 Ω | 1,0 k Ω | 3,8 | 11,6 | 120 | 3,5 | 10,8 | 100 | 2,4 | 7,8 | 50 |
| 150 Ω | 1,0 k Ω | 4,7 | 13,5 | 150 | 3,9 | 11,8 | 100 | 2,7 | 8,8 | 50 |
| 180 Ω | 1,0 k Ω | 5,7 | 15,1 | 180 | 4,2 | 12,6 | 100 | 3,0 | 9,5 | 50 |
| 220 Ω | 1,0 k Ω | 7,0 | 16,8 | 220 | 4,7 | 13,4 | 100 | 3,3 | 10,4 | 50 |
| 270 Ω | 1,0 k Ω | 8,2 | 18,3 | 250 | 5,2 | 14,3 | 100 | 3,7 | 11,3 | 50 |
| 330 Ω | 1,0 k Ω | 9,1 | 19,2 | 250 | 5,7 | 15,2 | 100 | 4,1 | 12,2 | 50 |
| 390 Ω | 1,0 k Ω | 9,9 | 19,9 | 250 | 6,2 | 15,9 | 100 | 4,4 | 12,9 | 50 |
| 470 Ω | 1,0 k Ω | 10,8 | 20,7 | 250 | 6,9 | 16,7 | 100 | 4,8 | 13,7 | 50 |
| 560 Ω | 1,0 k Ω | 11,8 | 21,5 | 250 | 7,5 | 17,5 | 100 | 5,3 | 14,5 | 50 |
| 680 Ω | 1,0 k Ω | 13,0 | 22,3 | 250 | 8,2 | 18,3 | 100 | 5,8 | 15,3 | 50 |
| 820 Ω | 1,0 k Ω | 14,3 | 23,1 | 250 | 9,1 | 19,1 | 100 | 6,4 | 16,1 | 50 |
| 1,0 k Ω | 1,0 k Ω | 15,8 | 24,0 | 250 | 10,0 | 20,0 | 100 | 7,1 | 17,0 | 50 |
| 1,2 k Ω | 1,0 k Ω | 17,3 | 24,8 | 250 | 11,0 | 20,8 | 100 | 7,7 | 17,8 | 50 |
| 1,5 k Ω | 1,0 k Ω | 19,4 | 25,7 | 250 | 12,2 | 21,8 | 100 | 8,7 | 18,8 | 50 |
| 1,8 k Ω | 1,0 k Ω | 21,2 | 26,5 | 250 | 13,4 | 22,6 | 100 | 9,5 | 19,5 | 50 |
| 2,2 k Ω | 1,0 k Ω | 23,5 | 27,4 | 250 | 14,8 | 23,4 | 100 | 10,5 | 20,4 | 50 |
| 2,7 k Ω | 10 k Ω | 26,0 | 28,3 | 250 | 16,4 | 24,3 | 100 | 11,6 | 21,3 | 50 |
| 3,3 k Ω | 10 k Ω | 28,7 | 29,2 | 250 | 18,2 | 25,2 | 100 | 12,8 | 22,2 | 50 |
| 3,9 k Ω | 10 k Ω | 31,2 | 29,9 | 250 | 19,7 | 25,9 | 100 | 14,0 | 22,9 | 50 |
| 4,7 k Ω | 10 k Ω | 34,3 | 30,7 | 250 | 21,7 | 26,7 | 100 | 15,3 | 23,7 | 50 |
| 5,6 k Ω | 10 k Ω | 37,4 | 31,5 | 250 | 23,7 | 27,5 | 100 | 16,7 | 24,5 | 50 |
| 6,8 k Ω | 10 k Ω | 41,2 | 32,3 | 250 | 26,1 | 28,3 | 100 | 18,4 | 25,3 | 50 |
| 8,2 k Ω | 10 k Ω | 45,3 | 33,1 | 250 | 28,6 | 29,1 | 100 | 20,2 | 26,1 | 50 |
| 10 k Ω | 10 k Ω | 50,0 | 34,0 | 250 | 31,6 | 30,0 | 100 | 22,4 | 27,0 | 50 |
| 12 k Ω | 10 k Ω | 54,8 | 34,8 | 250 | 34,6 | 30,8 | 100 | 24,5 | 27,8 | 50 |
| 15 k Ω | 10 k Ω | 61,2 | 35,7 | 250 | 38,7 | 31,8 | 100 | 27,4 | 28,8 | 50 |
| 18 k Ω | 10 k Ω | 67,1 | 36,5 | 250 | 42,4 | 32,6 | 100 | 30,0 | 29,5 | 50 |
| 22 k Ω | 10 k Ω | 74,2 | 37,4 | 250 | 46,9 | 33,4 | 100 | 33,2 | 30,4 | 50 |
| 27 k Ω | 100 k Ω | 82,2 | 38,3 | 250 | 52,0 | 34,3 | 100 | 36,7 | 31,3 | 50 |
| 33 k Ω | 100 k Ω | 90,8 | 39,2 | 250 | 57,4 | 35,2 | 100 | 40,6 | 32,2 | 50 |
| 39 k Ω | 100 k Ω | 98,7 | 39,9 | 250 | 62,4 | 35,9 | 100 | 44,2 | 32,9 | 50 |
| 47 k Ω | 100 k Ω | 108 | 40,7 | 250 | 68,6 | 36,7 | 100 | 48,5 | 33,7 | 50 |
| 56 k Ω | 100 k Ω | 118 | 41,5 | 250 | 74,8 | 37,5 | 100 | 52,9 | 34,5 | 50 |
| 68 k Ω | 100 k Ω | 130 | 42,3 | 250 | 82,5 | 38,3 | 100 | 58,3 | 35,3 | 50 |
| 82 k Ω | 100 k Ω | 143 | 43,1 | 250 | 90,6 | 39,1 | 100 | 64,0 | 36,1 | 50 |

Table 2 (2 of 2)

| R_T | R_M | $P_r \geq 0,5 \text{ W}$ | | | $0,5 \text{ W} > P_r \geq 0,1 \text{ W}$ | | | $0,1 \text{ W} > P_r$ | | |
|----------------|----------------|--------------------------|-----------|-------------|--|-----------|-------------|-----------------------|-----------|-------------|
| | | U_T V | D dB | P_T mW | U_T V | D dB | P_T mW | U_T V | D dB | P_T mW |
| 100 k Ω | 100 k Ω | 158 | 44,0 | 250 | 100 | 40,0 | 100 | 70,7 | 37,0 | 50 |
| 120 k Ω | 100 k Ω | 173 | 44,8 | 250 | 110 | 40,8 | 100 | 77,5 | 37,8 | 50 |
| 150 k Ω | 100 k Ω | 194 | 45,7 | 250 | 122 | 41,8 | 100 | 86,6 | 38,8 | 50 |
| 180 k Ω | 100 k Ω | 212 | 46,5 | 250 | 134 | 42,6 | 100 | 94,9 | 39,5 | 50 |
| 220 k Ω | 100 k Ω | 235 | 47,4 | 250 | 148 | 43,4 | 100 | 105 | 40,4 | 50 |
| 270 k Ω | 100 k Ω | 250 | 48,0 | 231 | 164 | 44,3 | 100 | 116 | 41,3 | 50 |
| 330 k Ω | 100 k Ω | 250 | 48,0 | 189 | 182 | 45,2 | 100 | 129 | 42,2 | 50 |
| 390 k Ω | 100 k Ω | 250 | 48,0 | 160 | 197 | 45,9 | 100 | 140 | 42,9 | 50 |
| 470 k Ω | 100 k Ω | 250 | 48,0 | 133 | 217 | 46,7 | 100 | 153 | 43,7 | 50 |
| 560 k Ω | 100 k Ω | 250 | 48,0 | 112 | 237 | 47,5 | 100 | 167 | 44,5 | 50 |
| 680 k Ω | 100 k Ω | 250 | 48,0 | 92 | 250 | 48,0 | 92 | 184 | 45,3 | 50 |
| 820 k Ω | 100 k Ω | 250 | 48,0 | 76 | 250 | 48,0 | 76 | 202 | 46,1 | 50 |
| 1,0 M Ω | 100 k Ω | 250 | 48,0 | 63 | 250 | 48,0 | 63 | 224 | 47,0 | 50 |
| 1,2 M Ω | 100 k Ω | 250 | 48,0 | 52 | 250 | 48,0 | 52 | 245 | 47,8 | 50 |
| 1,5 M Ω | 100 k Ω | 250 | 48,0 | 42 | 250 | 48,0 | 42 | 250 | 48,0 | 42 |
| 1,8 M Ω | 1,0 M Ω | 250 | 48,0 | 35 | 250 | 48,0 | 35 | 250 | 48,0 | 35 |
| 2,2 M Ω | 1,0 M Ω | 250 | 48,0 | 28 | 250 | 48,0 | 28 | 250 | 48,0 | 28 |
| 2,7 M Ω | 1,0 M Ω | 250 | 48,0 | 23 | 250 | 48,0 | 23 | 250 | 48,0 | 23 |
| 3,3 M Ω | 1,0 M Ω | 250 | 48,0 | 19 | 250 | 48,0 | 19 | 250 | 48,0 | 19 |
| 3,9 M Ω | 1,0 M Ω | 250 | 48,0 | 16 | 250 | 48,0 | 16 | 250 | 48,0 | 16 |
| 4,7 M Ω | 1,0 M Ω | 250 | 48,0 | 13 | 250 | 48,0 | 13 | 250 | 48,0 | 13 |
| 5,6 M Ω | 1,0 M Ω | 250 | 48,0 | 11 | 250 | 48,0 | 11 | 250 | 48,0 | 11 |
| 6,8 M Ω | 1,0 M Ω | 250 | 48,0 | 9,2 | 250 | 48,0 | 9,2 | 250 | 48,0 | 9,2 |
| 8,2 M Ω | 1,0 M Ω | 250 | 48,0 | 7,6 | 250 | 48,0 | 7,6 | 250 | 48,0 | 7,6 |
| 10 M Ω | 1,0 M Ω | 250 | 48,0 | 6,3 | 250 | 48,0 | 6,3 | 250 | 48,0 | 6,3 |
| 12 M Ω | 1,0 M Ω | 250 | 48,0 | 5,2 | 250 | 48,0 | 5,2 | 250 | 48,0 | 5,2 |
| 15 M Ω | 1,0 M Ω | 250 | 48,0 | 4,2 | 250 | 48,0 | 4,2 | 250 | 48,0 | 4,2 |
| 18 M Ω | 1,0 M Ω | 250 | 48,0 | 3,5 | 250 | 48,0 | 3,5 | 250 | 48,0 | 3,5 |
| 22 M Ω | 1,0 M Ω | 250 | 48,0 | 2,8 | 250 | 48,0 | 2,8 | 250 | 48,0 | 2,8 |

The values for this table are established under the following prerequisites:

- The voltage to be provided by the internal d.c. voltage source shall not exceed 400 V.
- The voltage U_T at the resistor specimen shall not exceed 250 V.
- The dissipation on the isolation resistor R_M shall not exceed 1 W.
- The allocation of isolation resistors R_M to specimen resistance R_T shall apply to all specimen styles.

Key

P_r is the rated dissipation of the resistor specimen under test

R_T is the resistance of the resistor specimen under test

R_M is the isolation resistor

U_T is the d.c. voltage applied to the specimen

D is the d.c. test voltage index, $D = 20 \lg (U_T / 1 \text{ V})$ dB

P_T is the power dissipated in the specimen

5.4 Precautions

Reasonable precautions such as are commonly associated with sensitive measurements should be followed when operating the test set. The operating location should be free of strong magnetic and electric fields and of sources of electro-magnetic radiation. Ordinarily, it need not be operated in a screened room. The location should be free from strong mechanical vibrations and from sources of loud sound. These precautions are mentioned to serve as a guide in selecting suitable locations. The suitability of a location can be determined by comparing test set performance in the selected location with that obtained in a "quiet" location. Usually, sources of interference are readily identifiable.

6 Evaluation of measurement results

6.1 Term for the contribution of system noise

The total noise voltage is the geometric sum of the system noise voltage and the current-noise voltage of the resistor under test

$$E_t^2 = E_s^2 + E_c^2$$

where

E_t is the effective total noise voltage in a given bandwidth;

E_s is the effective system noise voltage in a given bandwidth;

E_c is the effective current noise voltage in a given bandwidth.

NOTE For practical reasons, these noise voltages are generally given in μV .

This equation can be rearranged for E_c

$$E_c^2 = E_t^2 - E_s^2$$

Each noise voltage can be replaced by a term based on its logarithmic index,

$$C = 20 \lg \left(\frac{E_c}{1 \mu\text{V}} \right) \text{ dB} \quad \text{hence} \quad E_c = 10^{\left(\frac{C}{20 \text{ dB}} \right)} \times 1 \mu\text{V}$$

$$T = 20 \lg \left(\frac{E_t}{1 \mu\text{V}} \right) \text{ dB} \quad \text{hence} \quad E_t = 10^{\left(\frac{T}{20 \text{ dB}} \right)} \times 1 \mu\text{V}$$

$$S = 20 \lg \left(\frac{E_s}{1 \mu\text{V}} \right) \text{ dB} \quad \text{hence} \quad E_s = 10^{\left(\frac{S}{20 \text{ dB}} \right)} \times 1 \mu\text{V}$$

where

C is the logarithmic index of the current noise voltage, scaled at 0 dB = 1 μV ;

T is the logarithmic index of the total noise voltage, scaled at 0 dB = 1 μV ;

S is the logarithmic index of the system noise voltage, scaled at 0 dB = 1 μV .

This leads to the equation

$$\left(10^{\frac{C}{20 \text{ dB}}}\right)^2 = \left(10^{\frac{T}{20 \text{ dB}}}\right)^2 - \left(10^{\frac{S}{20 \text{ dB}}}\right)^2$$

Solving this equation for C and isolating T results in the following expression,

$$C = T + 10 \lg \left(1 - 10^{-\left[\frac{T-S}{10 \text{ dB}}\right]} \right) \text{ dB}$$

which permits to isolate a term for the contribution of system noise in relationship to the total noise, $f(T-S)$

$$C = T - f(T-S).$$

This function of the difference between total noise and system noise itself hence is defined as

$$f(T-S) = -10 \lg \left(1 - 10^{-\left[\frac{T-S}{10 \text{ dB}}\right]} \right) \text{ dB}$$




Figure 3 – Contribution of system noise, $f(T-S)$

Figure 3 shows the result of $f(T-S)$ plotted over $(T-S)$ and illustrates the special considerations discussed below.

- The accuracy of the determination of the current-noise deteriorates if the total noise approaches the system noise, which is the case for $(T - S) < 1,0$ dB. For such a case the reporting of seemingly accurate values for the current-noise index A_1 is no longer appropriate.
- The contribution of the system noise may be ignored if $f(T - S)$ does not provide a result which, rounded to one decimal, is at least 0,1 dB. This is the case for $(T - S) > 19,5$ dB.

For the reader's convenience, Table 3 provides numerical values of the contribution of system noise, rounded to one decimal.

Table 3 – Numeric values of the contribution of system noise, $f(T - S)$

| $T - S$ dB | $f(T - S)$ dB | $T - S$ dB | $f(T - S)$ dB | $T - S$ dB | $f(T - S)$ dB |
|---------------|------------------|---------------|------------------|---------------|------------------|
| <1,0 | 7,0 ^a | 2,9 | 3,1 | 5,0 | 1,7 |
| 1,0 | 6,9 | 3,0 | 3,0 | 5,1 | 1,6 |
| 1,1 | 6,5 | 3,1 | 2,9 | 5,2 | 1,6 |
| 1,2 | 6,2 | 3,2 | 2,8 | 5,3 | 1,5 |
| 1,3 | 5,9 | 3,3 | 2,7 | 5,4 | 1,5 |
| 1,4 | 5,6 | 3,4 | 2,7 | 5,5 to 5,7 | 1,4 |
| 1,5 | 5,3 | 3,5 | 2,6 | 5,8 to 6,0 | 1,3 |
| 1,6 | 5,1 | 3,6 | 2,5 | 6,1 to 6,3 | 1,2 |
| 1,7 | 4,9 | 3,7 | 2,4 | 6,4 to 6,6 | 1,1 |
| 1,8 | 4,7 | 3,8 | 2,3 | 6,7 to 7,0 | 1,0 |
| 1,9 | 4,5 | 3,9 | 2,3 | 7,1 to 7,5 | 0,9 |
| 2,0 | 4,3 | 4,0 | 2,2 | 7,6 to 7,9 | 0,8 |
| 2,1 | 4,2 | 4,1 | 2,1 | 8,0 to 8,5 | 0,7 |
| 2,2 | 4,0 | 4,2 | 2,1 | 8,6 to 9,2 | 0,6 |
| 2,3 | 3,9 | 4,3 | 2,0 | 9,3 to 10,0 | 0,5 |
| 2,4 | 3,7 | 4,4 | 2,0 | 10,1 to 11,1 | 0,4 |
| 2,5 | 3,6 | 4,5 | 1,9 | 11,2 to 12,5 | 0,3 |
| 2,6 | 3,5 | 4,6 | 1,8 | 12,6 to 14,6 | 0,2 |
| 2,7 | 3,3 | 4,7 | 1,8 | 14,7 to 19,4 | 0,1 |
| 2,8 | 3,2 | 4,8 | 1,7 | ≥19,5 | 0 |
| | | 4,9 | 1,7 | | |

^a The value $f(T - S) = 7$ dB given for $(T - S) < 1,0$ dB should be applied only for the determination of a maximum current noise index $A_{1 \max}$.

6.2 Determination of the current-noise index A_1

The current noise index in a frequency decade, A_1 , is determined by the quotient of the current-noise voltage in a frequency decade, E_C , over the applied d.c. voltage U_T .

The mathematical index A_1' is based on dividing the current noise voltage and d.c. voltage with proper consideration of the multiples of their unit volt, which with their logarithmic indexes C' and D is presented as

$$A_1' = C' - D$$

where

C' is the logarithmic index of the current noise voltage, scaled at 0 dB = 1 V;

D is the logarithmic index of the d.c. voltage U_T applied to the specimen, scaled at 0 dB = 1 V.

The practical index A_1 is based on dividing the current noise voltage and d.c. voltage in their individual units and multiples, which with their logarithmic indexes C and D is presented as

$$A_1 = C - D$$

where

C is the logarithmic index of the current noise voltage, scaled at 0 dB = 1 μ V.

The difference of 120 dB between the two indexes A_1' and A_1

$$A_1 = A_1' - 120 \text{ dB}.$$

is generally neglected.

Applying the definition of $f(T - S)$ from above (see 6.1), leads to the following standard equation for the determination of the current noise index A_1

$$A_1 = T - f(T - S) - D$$

The readings of the total noise index T , the system noise index S , both scaled at 0 dB = 1 μ V, and the index D of the applied d.c. voltage U_T , scaled at 0 dB = 1 V, shall be used to calculate the current noise index A_1 .

The following special consideration applies.

If the measured total noise approaches the measured system noise, hence if

$$(T - S) < 1,0 \text{ dB}$$

and thus the accuracy of the determination deteriorates, it is not recommended to use this method for the determination of a current-noise index A_1 with a pretended accuracy.

In such case the current-noise should only be reported as being less than a maximum current-noise index $A_{1 \text{ max}}$, which is determined by

$$A_{1 \text{ max}} = T - 7 \text{ dB} - D$$

6.3 Determination of the current-noise voltage ratio CNR_U

The current-noise voltage ratio in a frequency decade CNR_U

$$CNR_U = \frac{E_C}{U_T}$$

is derived from an established current-noise index A_1 by calculating

$$CNR_U = 10^{\frac{A_1}{20 \text{ dB}}} \cdot \frac{\mu\text{V}}{\text{V}}$$

6.4 Accuracy

The accuracy of the noise voltage measurement shall be $\pm 0,75$ dB. The accuracy of determinations of the current-noise index shall be ± 1 dB when the current noise is large compared to the system noise, i.e. $(T - S)$ is greater than 15 dB.

It is not uncommon for certain resistors to exhibit noise measurement variations greater than 0,75 dB. It should therefore be recognized that lack of agreement of repeated measurements on such resistors does not necessarily reflect a loss of accuracy of the measuring system, but is an indication of a noise property of the resistor.

6.5 Requirements

Acceptance criteria for the current noise of tested products shall be given with reference to a required maximum current-noise index A_1 in the relevant component specification.

Such acceptance criteria should be stated through a fixed maximum value, typically given as function of the specimen resistance.

7 Information to be given in the relevant component specification

When this test is included in a relevant component specification, the following details shall be given as far as they are required or applicable:

| | Subclause |
|---|----------------|
| The environmental conditions for this measurement | 5.1 |
| The dissipation to be provided through the applied d.c. voltage | 5.3.4, Table 2 |
| A limitation to the applied d.c. voltage, if applicable | 5.3.4, Table 2 |
| The relevant component specification shall specify for its own purpose: | |
| | Subclause |
| Acceptance criteria to the current noise index A_1 | 6.5 |

Annex A (informative)

Letter symbols and abbreviations

A.1 Letter symbols

| | | |
|--------------------|--|---------------|
| A | Non-dimensional figure representing the area of a pass-band in a gain-over-frequency diagram | 1 |
| A_1 | Current-noise index in a frequency decade (see 3.2) | dB |
| C | Logarithmic index of the current noise voltage | dB |
| D | Logarithmic index of the d.c. bias voltage | dB |
| $e(f)$ | Momentary current noise voltage as a function of frequency | μV |
| e_{th} | Momentary thermal noise voltage | μV |
| E_c | Effective (r.m.s) current noise voltage | μV |
| E_s | Effective (r.m.s) system noise voltage | μV |
| E_t | Effective (r.m.s) total noise voltage | μV |
| E_{th} | Effective (r.m.s) thermal noise voltage | μV |
| f | Frequency | Hz |
| f_1 | Lower cut-off frequency of a band-pass filter | Hz |
| f_2 | Upper cut-off frequency of a band-pass filter | Hz |
| f_c | Centre frequency of a band-pass filter | Hz |
| Δf | Pass-band of a band-pass filter | Hz |
| I | Current | A |
| k | Boltzmann constant, $k \approx 1,38 \times 10^{-23}$ J/K | J/K |
| R | Resistance | Ω |
| R_{Cal} | Resistance of the calibration resistor | Ω |
| R_M | Resistance of the isolation resistor | Ω |
| R_T | Resistance of the specimen | Ω |
| S | Logarithmic index of the system noise voltage | dB |
| S_k | Logarithmic index of the system noise voltage with shorted test terminals | dB |
| S_o | Logarithmic index of the system noise voltage with open test terminals | dB |
| T | Logarithmic index of the total noise voltage | dB |
| T | Absolute temperature | K |
| U_{Cal} | Calibration voltage | V |
| $U_{\text{N rms}}$ | AC r.m.s. noise voltage | μV |
| U_T | DC test voltage applied to the specimen | V |

A.2 Abbreviations

| | |
|-----|------------------------------------|
| BPA | Band-pass amplifier |
| G | Calibration source, a.c. generator |
| P | DC power supply |
| S | Switch |

Annex X (informative)

Cross-reference for references to the prior revision of this standard

The revision of this standard has resulted in a new clause numbering. Table X.1 provides cross-references between the clause numbering of this edition compared to the first edition of this standard.

Table X.1 – Cross reference for references to the 1st edition of this standard

| IEC 60195:1965 1 st edition | IEC 60195:2016 2 nd edition | Notes |
|---|---|---|
| Clause | Clause | |
| 1 | 1 | Object and scope are merged into one new clause. |
| 2 | | |
| 3 | 3 4.1 | – |
| 4 | 4.3 4.4 | – |
| 4.1 | 4.4.1 4.4.2 | – |
| 4.2 | 4.2 4.4.3 4.4.4 | – |
| 4.3 | 4.4.6 4.4.8 4.4.9 | – |
| 4.4 | 4.3.1 | – |
| 5 | 4.4.6 | – |
| 6 | 5 | – |
| 6.1 | 5.3.2 | – |
| 6.2 | 5.3.3 | – |
| 6.3 | 5.3.4 | – |
| 6.4 | 6 | – |
| 7 | 6.3 | – |
| 8 | – | The content of the prior Clause 8 is spread into new subclauses, see below. |
| 8.1 | 4.4 | – |
| 8.2 | 5.4 | – |
| 8.3 | 4.5 | – |
| Table I | Table 2 | Table 2 is modified for a harmonized allocation of the suggested R_M to the indicated values of R_T . Table 2 is amended with a new set of recommended operating conditions for specimen with $P_r < 0,1$ W. |
| Table II | Figure 3 Table 3 | Table 3 is re-calculated based on the given equation. |

Bibliography

IEC 60027 (all parts), *Letter symbols to be used in electrical technology*

IEC 60617, *Graphical symbols for diagrams*

MIL-STD-202G, Method 308 “*Current-Noise Test for Fixed Resistors*”, 1961

G. T. Conrad, Jr., N. Newman, A.P. Stansbury, “*A Recommended Standard Resistor-Noise Test System*”, IRE Transactions on Component Parts, Volume CP-7, pp. 1-18; September 1960

Hameg Instruments, Professional Article “*What is noise?*”, 2009

N. Newman, G. T. Conrad, Jr., “*Discussion of errors of a Recommended Standard Resistor-Noise Test System*”, IRE Transactions on Component Parts, Volume CP-9, pp. 180-192; December 1962

NoiseKen Noise Laboratory Co., Ltd, “*Resistor Current Noise Tester RCN-2011*”, Datasheet, 2012

Quan-Tech, Division of KMS Industries, Inc., “*Model 315B Resistor Noise Test Set*”, Instruction Manual, 1973

Quan-Tech, Division of KMS Industries, Inc., “*Model 315C Resistor Noise Test Set*”, Instruction Manual

F. Zandman, P.-R. Simon, J. Szwarc, “*Resistor Theory and Technology*”, ISBN 189112112X, 2002

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Copyright in BSI publications

All the content in BSI publications, including British Standards, is the property of and copyrighted by BSI or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use.

Save for the provisions below, you may not transfer, share or disseminate any portion of the standard to any other person. You may not adapt, distribute, commercially exploit, or publicly display the standard or any portion thereof in any manner whatsoever without BSI's prior written consent.

Storing and using standards

Standards purchased in soft copy format:

- A British Standard purchased in soft copy format is licensed to a sole named user for personal or internal company use only.
- The standard may be stored on more than 1 device provided that it is accessible by the sole named user only and that only 1 copy is accessed at any one time.
- A single paper copy may be printed for personal or internal company use only.

Standards purchased in hard copy format:

- A British Standard purchased in hard copy format is for personal or internal company use only.
- It may not be further reproduced – in any format – to create an additional copy. This includes scanning of the document.

If you need more than 1 copy of the document, or if you wish to share the document on an internal network, you can save money by choosing a subscription product (see 'Subscriptions').

Reproducing extracts

For permission to reproduce content from BSI publications contact the BSI Copyright & Licensing team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email subscriptions@bsigroup.com.

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Useful Contacts

Customer Services

Tel: +44 345 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 345 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK