

BS ISO/CIE 11664-6:2014



BSI Standards Publication

Colorimetry

Part 6: CIEDE2000 Colour-difference formula

bsi.

...making excellence a habit.™

National foreword

This British Standard is the UK implementation of ISO/CIE 11664-6:2014.

The UK participation in its preparation was entrusted to Technical Committee STI/14, Colour measurement and schedules.

A list of organizations represented on this committee can be obtained on request to its secretary.

This publication does not purport to include all the necessary provisions of a contract. Users are responsible for its correct application.

© The British Standards Institution 2014. Published by BSI Standards Limited 2014

ISBN 978 0 580 82701 3

ICS 17.180.20

Compliance with a British Standard cannot confer immunity from legal obligations.

This British Standard was published under the authority of the Standards Policy and Strategy Committee on 28 February 2014.

Amendments issued since publication

| Date | Text affected |
|------|---------------|
|------|---------------|

INTERNATIONAL
STANDARD

BS ISO/CIE 11664-6:2014

ISO/CIE
11664-6

First edition
2014-02-01

Colorimetry —

Part 6:

CIEDE2000 Colour-difference formula

Colorimétrie —

Partie 6: Formule d'écart de couleur CIEDE2000



Reference number
ISO/CIE 11664-6:2014(E)

© ISO/CIE 2014



COPYRIGHT PROTECTED DOCUMENT

© ISO/CIE 2014

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized otherwise in any form or by any means, electronic or mechanical, including photocopying, or posting on the internet or an intranet, without prior written permission. Permission can be requested from either ISO or CIE at the respective addresses below.

ISO copyright office
Case postale 56 • CH-1211 Geneva 20
Tel. + 41 22 749 01 11
Fax + 41 22 749 09 47
E-mail copyright@iso.org
Web www.iso.org

CIE Central Bureau
Babenbergerstraße 9/9A • A-1010 Vienna
Tel. + 43 1 714 3187

E-mail ciecb@cie.co.at
Web www.cie.co.at

Published in Switzerland

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

ISO/CIE 11664-6 was prepared by CIE Technical Committee TC 1-57 of Division 1 *Vision and colour*, as CIE S 014-6.

ISO/CIE 11664 consists of the following parts, under the general title *Colorimetry*:

- *Part 1: CIE standard colorimetric observers*
- *Part 2: CIE standard illuminants*
- *Part 3: CIE tristimulus values*
- *Part 4: CIE 1976 L*a*b* Colour space*
- *Part 5: CIE 1976 L*u*v* Colour space and u', v' uniform chromaticity scale diagram*
- *Part 6: CIEDE2000 Colour-difference formula*



COMMISSION INTERNATIONALE DE L'ÉCLAIRAGE
INTERNATIONAL COMMISSION ON ILLUMINATION
INTERNATIONALE BELEUCHTUNGSKOMMISSION

International Standard

CIE S 014-6/E:2013

Colorimetry – Part 6: CIEDE2000 Colour-Difference Formula

Colorimétrie – Partie 6: Formule d'écart de couleur CIEDE2000

Farbmessung – Teil 6: CIEDE2000-Farbabstandsformel

CIE International Standards are copyrighted and shall not be reproduced in any form, entirely or partly, without the explicit agreement of the CIE.

CIE Central Bureau, Vienna
Babenbergerstraße 9/9A • A-1010 Vienna

CIE S 014-6/E:2013

UDC: 535.65:006
535.643.2

Descriptor: Standardisation of colour measurement
Standard colorimetric systems

© CIE 2013

This document is a CIE International Standard and is copyright-protected by CIE.

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from CIE Central Bureau at the address below.

CIE Central Bureau
Babenbergerstraße 9/9A
A-10F0 Vienna
Austria
Tel.: +43 1 714 3187
e-mail: ciecb@cie.co.at
www.cie.co.at

Foreword

International Standards produced by the Commission Internationale de l'Eclairage are concise documents on aspects of light and lighting that require a unique definition. They are a primary source of internationally accepted and agreed data which can be taken, essentially unaltered, into universal standard systems.

This CIE International Standard has been prepared by CIE Technical Committee TC 1-57*. It has been approved by the Board of Administration and Division 1 "Vision and Colour" of the Commission Internationale de l'Eclairage and the CIE National Committees.

The following ISO and IEC committees and working groups co-operated in the preparation of this International Standard:

IEC TC100/TA2 (Audio, video and multimedia systems)

ISO TC6 (Paper, board and pulps)

ISO TC35/SC9/WG22 (Paints and varnishes)

ISO TC38/SC1/WG7 (Textiles)

ISO TC42 (Photography)

ISO TC130 (Graphic technology)

ISO/IEC/JTC1/SC28 (Office systems)

* The chairperson of this TC was A.R. Robertson (CA), members were: P.J. Alessi (US), M. Brill (US), J. Campos Acosta (ES), E. Carter (US), R. Connelly (US), J.-F. Decarreau (FR), R. Harold (US), R. Hirschler (HU), B. Jordan (CA), C. Kim (KR), D. McDowell (US), P. McGinley (AU), M. Melgosa (ES), Y. Ohno (US), M.R. Poin ter (GB), K. Richter (DE), G. Rösler † (DE), J.D. Schanda (HU), R. Sève (FR), K. Witt (DE), H. Yaguchi (JP), J. Zwinkels (CA).

CONTENTS

| | |
|---|-----|
| Foreword | vii |
| Introduction | 1 |
| 1 Scope | 1 |
| 2 Normative References | 1 |
| 3 Definitions, Symbols and Abbreviations | 1 |
| 4 Reference Conditions | 3 |
| 5 Calculation Method..... | 3 |
| 6 Parametric Factors | 6 |
| Annex A (Informative) Three-Component Micro-Spaces..... | 7 |
| Bibliography | 8 |

Colorimetry – Part 6: CIEDE2000 Colour-Difference Formula

Introduction

The three-dimensional colour space produced by plotting CIE tristimulus values (X , Y , Z) in rectangular coordinates is not visually uniform, nor is the (x , y , Y) space nor the two-dimensional CIE (x , y) chromaticity diagram. Equal distances in these spaces and diagrams do not represent equally perceptible differences between colour stimuli. For this reason the CIE has standardized two more-nearly uniform colour spaces (known as CIELAB and CIELUV) whose coordinates are non-linear functions of X , Y and Z . Numerical values representing approximately the relative magnitude of colour differences can be described by simple Euclidean distances in these spaces or by more sophisticated colour-difference formulae that improve the correlation with the relative perceived size of differences. The purpose of this CIE International Standard is to define one such formula, the CIEDE2000 formula. The Standard is based on CIE Technical Report 142-2001.

The formula is an extension of the CIE 1976 $L^*a^*b^*$ colour-difference formula (ISO 11664-4:2008(E)/CIE S 014-4/E:2007) with corrections for variation in colour-difference perception dependent on lightness, chroma, hue and chroma-hue interaction. Reference conditions define material and viewing environment characteristics to which the formula applies.

1 Scope

This CIE International Standard specifies the method of calculating colour differences according to the CIEDE2000 formula.

The Standard is applicable to input values of CIELAB L^* , a^* , b^* coordinates calculated according to ISO 11664-4:2008(E)/CIE S 014-4/E:2007. The Standard may be used for the specification of the colour difference between two colour stimuli perceived as belonging to reflecting or transmitting objects. This includes displays, if they are being used to simulate reflecting or transmitting objects and if the tristimulus values representing the stimuli are appropriately normalized. The Standard does not apply to colour stimuli perceived as belonging to areas that appear to be emitting light as primary light sources, or that appear to be specularly reflecting such light.

2 Normative References

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

CIE 142-2001. *Improvement to industrial colour-difference evaluation*, 2001.

CIE S 017/E:2011. *ILV: International Lighting Vocabulary*, 2011.

ISO 11664-4:2008(E)/CIE S 014-4/ E:2007. Joint ISO/CIE Standard: *Colorimetry – Part 4: 1976 $L^*a^*b^*$ Colour Space*, 2008.

3 Definitions, Symbols and Abbreviations

For the purposes of this International Standard, the terms and definitions given in CIE S 017/E:2011 (International Lighting Vocabulary), and the following symbols and abbreviations apply.

| | |
|-----------------|--|
| L^* | CIELAB lightness |
| a^*, b^* | CIELAB a^*, b^* coordinates |
| C_{ab}^* | CIELAB chroma |
| h_{ab} | CIELAB hue angle |
| L' | CIEDE2000 lightness |
| \bar{L}' | arithmetic mean of the CIEDE2000 lightnesses of two colour stimuli |
| a', b' | CIEDE2000 a', b' coordinates |
| C' | CIEDE2000 chroma |
| \bar{C}' | arithmetic mean of the CIEDE2000 chromas of two colour stimuli |
| h' | CIEDE2000 hue angle |
| \bar{h}' | arithmetic mean of the CIEDE2000 hue angles of two colour stimuli |
| G | switching function used in the modification of a^* |
| $\Delta L'$ | CIEDE2000 lightness difference |
| $\Delta C'$ | CIEDE2000 chroma difference |
| $\Delta h'$ | CIEDE2000 hue-angle difference |
| $\Delta H'$ | CIEDE2000 hue difference |
| ΔE_{00} | CIEDE2000 colour difference |
| S_L | lightness weighting function |
| S_C | chroma weighting function |
| S_H | hue weighting function |
| T | T -function for hue weighting |
| R_T | rotation function |
| $\Delta\theta$ | hue dependence of rotation function |
| R_C | chroma dependence of rotation function |
| k_L | lightness parametric factor |
| k_C | chroma parametric factor |
| k_H | hue parametric factor |

4 Reference Conditions

The CIEDE2000 formula is intended to be applicable to objects viewed under the following reference conditions:

- Illumination: source simulating the relative spectral irradiance of CIE Standard Illuminant D65.
- Illuminance: 1 000 lx.
- Observer: normal colour vision.
- Background field: uniform, neutral grey with $L^* = 50$.
- Viewing mode: object.
- Sample size: sample pair subtending a visual angle greater than 4° .
- Sample separation: minimum sample separation achieved by placing the sample pair in direct edge contact.
- Sample colour-difference magnitude: 0 to 5 CIELAB units.
- Sample structure: homogeneous colour without visually apparent pattern or non-uniformity.

When conditions of use deviate appreciably from the reference conditions, parametric factors may be used to correct for the effects of material or experimental variables, as described in Clause 6.

NOTE 1 A proposed Technical Report of CIE T C 1-63 "Validity of the Range of CIE DE2000" will discuss and may show some application limitations of CIEDE2000 for CIELAB colour differences < 1 unit or > 5 units and for the CIE 1931 standard colorimetric observer (CIE 2 degree observer).

NOTE 2 A proposed Technical Report of CIE T C 1-81 "Validity of Formulae for Predicting Small Colour Differences" will discuss and may show some application limitations of CIEDE2000 for CIELAB colour differences < 2 units.

5 Calculation Method

All angular quantities in this Standard shall be evaluated in degrees.

CIELAB L^* , a^* , b^* , and C_{ab}^* coordinates of the two samples shall be calculated according to ISO 11664-4:2008(E)/CIE S 014-4/E:2007.

Modified CIELAB coordinates shall be calculated according to equations (1) to (7).

$$L' = L^* \tag{1}$$

$$a' = (1 + G) a^* \tag{2}$$

$$b' = b^* \tag{3}$$

$$C' = (a'^2 + b'^2)^{1/2} \tag{4}$$

$$h' = \begin{cases} \arctan\left(\frac{b'}{a'}\right) & \text{if } a' > 0 \text{ and } b' \geq 0 \\ \arctan\left(\frac{b'}{a'}\right) + 360^\circ & \text{if } a' > 0 \text{ and } b' < 0 \\ \arctan\left(\frac{b'}{a'}\right) + 180^\circ & \text{if } a' < 0 \\ 90^\circ & \text{if } a' = 0 \text{ and } b' > 0 \\ 270^\circ & \text{if } a' = 0 \text{ and } b' < 0 \end{cases} \quad (5)$$

$$h' = 0^\circ \quad \text{if } a' = 0 \text{ and } b' = 0 \quad (6)$$

where

$$G = 0,5 \left(1 - \frac{\sqrt{\left(\overline{C_{ab}^*}\right)^7}}{\sqrt{\left(\overline{C_{ab}^*}\right)^7 + 25^7}} \right) \quad (7)$$

and $\overline{C_{ab}^*}$ is the arithmetic mean of the C_{ab}^* values for the two samples of the colour-difference pair.

Equation (5) ensures that h' is the angular position of the point a', b' in the range from 0° to 360° measured from the positive a' axis in the a', b' plane. In the case where $a' = b' = 0$, h' is indeterminate and shall be assigned a value of zero as indicated in equation (6).

NOTE 1 The L', a', b', C' and h' values should be used only for the calculation of colour difference and should not be used as an alternative uniform colour space. When reporting CIELAB colour space coordinates, L^*, a^*, b^*, C_{ab}^* and h_{ab} values should be used.

Differences between two samples denoted by subscripts 0 (usually the reference) and 1 (usually the test) shall be calculated as follows:

$$\Delta L' = L'_1 - L'_0 \quad (8)$$

$$\Delta C' = C'_1 - C'_0 \quad (9)$$

$$\Delta H' = 2 \left(C'_0 C'_1 \right)^{1/2} \sin(\Delta h' / 2) \quad (10)$$

where

$$\Delta h' = 0^\circ \quad \text{if } C'_0 C'_1 = 0 \quad (11)$$

$$\Delta h' = h'_1 - h'_0 \quad \text{if } C'_0 C'_1 \neq 0 \text{ and } |h'_1 - h'_0| \leq 180^\circ \quad (12)$$

$$\Delta h' = h'_1 - h'_0 - 360^\circ \quad \text{if } C'_0 C'_1 \neq 0 \text{ and } (h'_1 - h'_0) > 180^\circ \quad (13)$$

$$\Delta h' = h'_1 - h'_0 + 360^\circ \quad \text{if } C'_0 C'_1 \neq 0 \text{ and } (h'_1 - h'_0) < -180^\circ \quad (14)$$

NOTE 2 Equations (11) to (14) avoid possible computational difficulties when h'_0 and h'_1 are in different quadrants or when one of the chromas is zero. They are based on Sharma et al., 2005.

NOTE 3 In information technology and other fields the subscripts r (for reference) and t (for test) are sometimes used instead of 0 and 1, respectively. Similarly in industrial evaluation of small colour differences s (for standard) and b (for batch) are sometimes used.

The CIEDE2000 colour difference, ΔE_{00} between the two samples shall be calculated by

$$\Delta E_{00} = \left[\left(\frac{\Delta L'}{k_L S_L} \right)^2 + \left(\frac{\Delta C'}{k_C S_C} \right)^2 + \left(\frac{\Delta H'}{k_H S_H} \right)^2 + R_T \left(\frac{\Delta C'}{k_C S_C} \right) \left(\frac{\Delta H'}{k_H S_H} \right) \right]^{1/2} \quad (15)$$

where

$$S_L = 1 + \frac{0,015 (\bar{L}' - 50)^2}{\sqrt{20 + (\bar{L}' - 50)^2}} \quad (16)$$

$$S_C = 1 + 0,045 \bar{C}' \quad (17)$$

$$S_H = 1 + 0,015 \bar{C}' T \quad (18)$$

$$T = 1 - 0,17 \cos(\bar{h}' - 30^\circ) + 0,24 \cos(2\bar{h}') + 0,32 \cos(3\bar{h}' + 6^\circ) - 0,20 \cos(4\bar{h}' - 63^\circ) \quad (19)$$

$$R_T = -\sin(2\Delta\theta)R_C \quad (20)$$

$$\Delta\theta = 30^\circ \exp\{-[(\bar{h}' - 275^\circ)/25^\circ]^2\} \quad (21)$$

$$R_C = 2 \sqrt{\frac{(\bar{C}')^7}{(\bar{C}')^7 + 25^7}} \quad (22)$$

k_L , k_C and k_H are parametric factors explained in Clause 6.

NOTE 4 The \bar{L}' , \bar{C}' and \bar{h}' values used in equations (16) to (22) are the arithmetic means of the corresponding values of the colour-difference pair. A consequence of this is that the total colour difference is reversible, that is, the total colour difference between a pair is the same whether the first or second sample is used as the standard for calculation of colour-difference components.

NOTE 5 The locus of points of equal total colour difference from a standard is not an exact ellipsoid and is not exactly centred on the standard.

Users should take care in calculating the mean hue angle if the colour-difference pair has samples in different quadrants. For example, if a colour-difference pair has hue angles of 30° and 300°, the simple mean, 165°, is incorrect, the correct value being 345°. To determine the mean correctly, the following equations (Sharma et al., 2005) shall be used.

$$\bar{h}' = (h'_0 + h'_1) / 2 \quad \text{if } |h'_0 - h'_1| \leq 180^\circ \text{ and } C'_0 C'_1 \neq 0 \quad (23)$$

$$\bar{h}' = (h'_0 + h'_1 + 360^\circ) / 2 \quad \text{if } |h'_0 - h'_1| > 180^\circ \text{ and } (h'_0 + h'_1) < 360^\circ \text{ and } C'_0 C'_1 \neq 0 \quad (24)$$

$$\bar{h} = (h'_0 + h'_1 - 360^\circ) / 2 \quad \text{if } |h'_0 - h'_1| > 180^\circ \text{ and } (h'_0 + h'_1) \geq 360^\circ \text{ and } C'_0 C'_1 \neq 0 \quad (25)$$

$$\bar{h} = h'_0 + h'_1 \quad \text{if } C'_0 C'_1 = 0 \quad (26)$$

NOTE 6 Some worked examples of the calculations are given in CIE 142-2001, Luo et al., 2001 and Sharma et al., 2005. Sharma et al. also give some useful implementation notes and mathematical observations.

6 Parametric Factors

Experimental observation and material variables can have parametric effects that influence the visual colour-difference results (CIE, 1993). The parametric factors k_L , k_C and k_H may be used to correct for these effects.

Under the reference conditions the parametric factors have assigned values of unity and do not affect the total colour difference.

The parametric factors provide a method to correct for deviation in experimental conditions from the defined reference conditions. Users are cautioned against indiscriminate use of these parametric factors without thorough experimental validation. Industry groups may define parametric factors to correspond to typical experimental conditions for that industry.

NOTE In the textile industry it is common practice to set the lightness parametric factor to 2. However, the experimental conditions leading to this parametric correction to lightness-difference sensitivity are not yet well understood.

Annex A (Informative) Three-Component Micro-Spaces

For certain applications, such as apportioning a colour difference into lightness, chroma and hue components for shade sorting and attributing size and direction to a specific difference in recipe prediction, it is useful to have a three-term version of equation (15) valid in a micro-space around the reference. The following equations accomplish this (Nobbs, 2002).

$$\tan(2\varphi) = R_T \frac{(k_C S_C)(k_H S_H)}{(k_H S_H)^2 - (k_C S_C)^2} \quad (\text{A.1})$$

with φ taken to be between -90° and 90° .

NOTE 1 If $k_H S_H = k_C S_C$, then 2φ is equal to 90° and φ is equal to 45° .

$$\Delta C'' = \Delta C' \cos(\varphi) + \Delta H' \sin(\varphi) \quad (\text{A.2})$$

$$\Delta H'' = \Delta H' \cos(\varphi) - \Delta C' \sin(\varphi) \quad (\text{A.3})$$

$$S_C'' = (k_C S_C) \sqrt{\frac{2(k_H S_H)}{2(k_H S_H) + R_T(k_C S_C) \tan(\varphi)}} \quad (\text{A.4})$$

$$S_H'' = (k_H S_H) \sqrt{\frac{2(k_C S_C)}{2(k_C S_C) - R_T(k_H S_H) \tan(\varphi)}} \quad (\text{A.5})$$

$$\Delta L_{00} = \frac{\Delta L'}{k_L S_L} \quad (\text{A.6})$$

$$\Delta C_{00} = \frac{\Delta C''}{S_C''} \quad (\text{A.7})$$

$$\Delta H_{00} = \frac{\Delta H''}{S_H''} \quad (\text{A.8})$$

$$\Delta E_{00} = [(\Delta L_{00})^2 + (\Delta C_{00})^2 + (\Delta H_{00})^2]^{1/2} \quad (\text{A.9})$$

NOTE 2 Equation (A.9) gives a value of colour difference ΔE_{00} identical to that calculated using equation (15). However, the definition of the three-dimensional "micro-space" is different for each colour centre.

Bibliography

CIE 101-1993. *Parametric Effects in Colour-Difference Evaluation*, 1993.

LUO, M.R., CUI, G., RIGG, B. The development of the CIE 2000 colour-difference formula: CIEDE2000, *Color Res. Appl.*, **26**, 340-350, 2001.

NOBBS, J.H., A Lightness, Chroma and Hue Splitting Approach to CIEDE2000 Colour Differences, *Advances in Colour Science and Technology*, **5**, 46-53, 2002.

SHARMA, G., WU, W., DALAL, E.N. The CIEDE2000 Color-Difference Formula: Implementation Notes, Supplementary Test Data, and Mathematical Observations, *Color Res. Appl.*, **30**, 21-30, 2005.

Copies of CIE publications are available from the National Committees of most CIE member countries or from the CIE web shop (www.cie.co.at).

CIE Central Bureau
Babenbergerstraße 9/9A
A-10F0 Vienna
Austria
Tel.: +43 1 714 3187
e-mail: ciecb@cie.co.at
<http://www.cie.co.at/>

British Standards Institution (BSI)

BSI is the national body responsible for preparing British Standards and other standards-related publications, information and services.

BSI is incorporated by Royal Charter. British Standards and other standardization products are published by BSI Standards Limited.

About us

We bring together business, industry, government, consumers, innovators and others to shape their combined experience and expertise into standards-based solutions.

The knowledge embodied in our standards has been carefully assembled in a dependable format and refined through our open consultation process. Organizations of all sizes and across all sectors choose standards to help them achieve their goals.

Information on standards

We can provide you with the knowledge that your organization needs to succeed. Find out more about British Standards by visiting our website at bsigroup.com/standards or contacting our Customer Services team or Knowledge Centre.

Buying standards

You can buy and download PDF versions of BSI publications, including British and adopted European and international standards, through our website at bsigroup.com/shop, where hard copies can also be purchased.

If you need international and foreign standards from other Standards Development Organizations, hard copies can be ordered from our Customer Services team.

Subscriptions

Our range of subscription services are designed to make using standards easier for you. For further information on our subscription products go to bsigroup.com/subscriptions.

With **British Standards Online (BSOL)** you'll have instant access to over 55,000 British and adopted European and international standards from your desktop. It's available 24/7 and is refreshed daily so you'll always be up to date.

You can keep in touch with standards developments and receive substantial discounts on the purchase price of standards, both in single copy and subscription format, by becoming a **BSI Subscribing Member**.

PLUS is an updating service exclusive to BSI Subscribing Members. You will automatically receive the latest hard copy of your standards when they're revised or replaced.

To find out more about becoming a BSI Subscribing Member and the benefits of membership, please visit bsigroup.com/shop.

With a **Multi-User Network Licence (MUNL)** you are able to host standards publications on your intranet. Licences can cover as few or as many users as you wish. With updates supplied as soon as they're available, you can be sure your documentation is current. For further information, email bsmusales@bsigroup.com.

BSI Group Headquarters

389 Chiswick High Road London W4 4AL UK

Revisions

Our British Standards and other publications are updated by amendment or revision.

We continually improve the quality of our products and services to benefit your business. If you find an inaccuracy or ambiguity within a British Standard or other BSI publication please inform the Knowledge Centre.

Copyright

All the data, software and documentation set out in all British Standards and other BSI publications are the property of and copyrighted by BSI, or some person or entity that owns copyright in the information used (such as the international standardization bodies) and has formally licensed such information to BSI for commercial publication and use. Except as permitted under the Copyright, Designs and Patents Act 1988 no extract may be reproduced, stored in a retrieval system or transmitted in any form or by any means – electronic, photocopying, recording or otherwise – without prior written permission from BSI. Details and advice can be obtained from the Copyright & Licensing Department.

Useful Contacts:

Customer Services

Tel: +44 845 086 9001

Email (orders): orders@bsigroup.com

Email (enquiries): cservices@bsigroup.com

Subscriptions

Tel: +44 845 086 9001

Email: subscriptions@bsigroup.com

Knowledge Centre

Tel: +44 20 8996 7004

Email: knowledgecentre@bsigroup.com

Copyright & Licensing

Tel: +44 20 8996 7070

Email: copyright@bsigroup.com



...making excellence a habit.™